

	MW	$m/z$	Calibration factor ncps ppbv <sup>-1</sup>
Formaldehyde	30	31	1.36 ± 21 %
Acetaldehyde	44	45	19.81 ± 6 %
Acetone	58	59	24.02 ± 7 %
Isoprene	68	69	8.84 ± 17 %
Benzene	78	79	17.15 ± 6 %
Toluene	92	93	19.87 ± 6 %
<i>m</i> -Xylene	106	107	19.78 ± 8 %