

		May	August
O <sub>3</sub>	Zero offset (mV)	-34.6417	-42.7629
	Sensitivity (mV ppb <sup>-1</sup> )	0.6404	0.2997
CO	Zero offset (mV)	108.9770	89.5812
	Sensitivity (mV ppb <sup>-1</sup> )	1.2192	1.0301
NO	Zero offset (mV)	-14.2030	-17.7801
	Sensitivity (mV ppb <sup>-1</sup> )	1.5758	1.2972
NO <sub>2</sub>	Zero offset (mV)	-13.7159	-6.0649
	Sensitivity (mV ppb <sup>-1</sup> )	0.4842	0.3843