

Effect of Dobson characteristics measurements within ATMOZ

Absorption coefficients	A	C	D	AD	CD
Dobson/Komhyr (nominal)	1.806	0.833	0.374	1.432	0.459
D064					
Ratio EAC / nominal	0.997	0.993	0.972	1.003	1.011
EAC	1.800	0.828	0.364	1.436	0.464
Relative difference in % TOC				-0.310	-1.060
D074					
Ratio EAC / nominal	0.993	1.006	0.989	0.994	1.020
EAC	1.794	0.838	0.370	1.424	0.468
Relative difference in % TOC				0.559	-1.961
D083					
Ratio EAC / nominal	0.997	0.998	0.983	1.001	1.011
EAC	1.800	0.832	0.367	1.433	0.464
Relative difference in % TOC				-0.060	-1.120
2010-results	1.805	0.830	0.376	1.429	0.454