

| AVODOR | Standard | Correlation fraction | | |
|---|----------|----------------------|--------|--------------|
| | | uncertainty | Full | Unfavourable |
| Source of uncertainty in measured $E(\lambda)$ | % | | | |
| Radiometric calibration | 2.50 | 1/2 | 1/2 | $1/\sqrt{2}$ |
| 250 W lamp stability | 0.14 | 1 | 0 | 0 |
| Non-linearity | 0.50 | 0 | 1 | 0 |
| Stability | 0.60 | 1 | 0 | 0 |
| Temperature dependence | 0.20 | 1 | 0 | 0 |
| Measurement noise | 1.30 | 0 | 0 | 1 |
| Wavelength shift | 0.10 | 0 | 1 | 0 |
| Combined uncertainty ($k = 1$) | 2.94 % | 1.41 % | 1.35 % | 2.19 % |