

Supplement of Atmos. Meas. Tech., 13, 273–285, 2020
<https://doi.org/10.5194/amt-13-273-2020-supplement>
© Author(s) 2020. This work is distributed under
the Creative Commons Attribution 4.0 License.




Supplement of

**Quantification of nitrous acid (HONO) and nitrogen dioxide (NO₂)
in ambient air by broadband cavity-enhanced absorption
spectroscopy (IBBCEAS) between 361 and 388 nm**

Nick Jordan and Hans D. Osthoff

Correspondence to: Hans D. Osthoff (hosthoff@ucalgary.ca)


The copyright of individual parts of the supplement might differ from the CC BY 4.0 License.



10

Figure S1 Absorption spectroscopy in the 350–400 nm spectral window. The LED output (blue area) was measured by directing the LED output through the cavity sample cell in the absence of the broadband mirrors and peaked at 367.8 nm. A bandpass filter (green dashed line) was used to cut out stray light and unwanted emission from the LED outside the maximum mirror reflectivity range (i.e., > 390 nm). The mirror reflectivity (see section 2.2 of the main manuscript) is shown in red. The cavity output spectrum (black) is a convolution between the LED output, the mirror curve and the bandpass filter function; its intensity is scaled to that of the LED output for presentation purposes. The absorption cross-sections of HONO (blue dashed line) and NO₂ (solid blue line) are superimposed. The yellow shaded region shows the spectral fit window used in the data retrieval.

15



20

Figure S2 An illustration of the Ne calibration lamp emission spectrum between 340 and 390 nm. The grey trace shows the raw spectrum recorded by the spectrograph while the black trace is that of the calibrated spectrum using a 5th degree polynomial fit to the raw lines in Table S1. The relative intensities of the Ne spectral lines are indicated on the graph (in red), as well as the relative difference (+/-) between the raw and calibrated spectrum (in black). Dotted red lines are used as a visual guide.

25

Table S1 Ne lamp spectral line intensities at the raw, calibrated and literature (Bartmess, 2011) wavelengths.

Raw line (nm)	Calibrated line (nm)	Ne lamp emission line (nm)	Difference (+/- nm)
341.80	341.79	341.79	0.00
346.93	347.27	347.26	+0.01
351.70	352.03	352.05	-0.02
359.34	359.36	359.35	+0.01
375.65	375.42	375.42	0.00
386.94	386.87	386.88	-0.01

30




Figure S3 Literature and convolved absorption cross-sections of NO₂ and HONO. The absorption cross-section of NO₂ is from Vandaele et al. (1998). The absorption cross-section of HONO is that by Stutz et al. (2000).





Figure S4 Measurement of HONO mixing ratios by TD-CRDS. **a)** Sample inlet temperature scan. The heated red (“hot”) channel detects NO + O₃, i.e. NO₂ on top of a NO_x background, while the blue channel measures NO₂ at a reference temperature. The black trace shows the quartz inlet temperature, which was decremented by 10 °C every 10 s. **b)** Mixing ratio observed in the NO₂+HONO channel as a function of inlet temperature. The contribution from HONO (ΔNO_2) is measured from the flat (“cold”) portion of the curve to the vertex (“hot”) end of the ramp.



45 **Figure S5** Determination of the cavity effective absorption path length: **a**) as R_L using O_4 from the ratio of the $[O_2]_{on}$ and $[O_2]_{off}$ using Eq. (4). Both sets were carried out at ambient pressure (~ 880 hPa). Blue shaded areas give $R_L \pm 2\sigma$; **b**) in a varied sample flow rate. At a higher volumetric flow rate (**c**), d_0 converges to d due to diffusion of the sample gas into the purge volumes of the instrument.




Figure S6 Scatter plots for: **a)** NO₂ and **b)** HONO. Error bars give the $\pm 1\sigma$ uncertainty in each measurement method.




Figure S7 Time series of IBBCEAS HONO data at 1 s time resolution. The grey shaded regions represent the zero air intervals used in the data reduction. The onset and end of the actual zero intervals coincides with the sharp changes in the mixing ratio of either absorber. The rise (0% to 100%) time between sample and zero is ~ 5.5 s, consistent with the sample flow rate of 2 slpm for which the sample residence time is 5.5 s.

References

65

Bartmess, J. E.: NIST Chemistry WebBook, NIST Standard Reference Database Number 69, in, edited by: Linstrom, P. J., and Mallard, W. G., National Institute of Standards and Technology, Gaithersburg MD <http://webbook.nist.gov> (retrieved June 16, 2011). 2011.

70

Stutz, J., Kim, E. S., Platt, U., Bruno, P., Perrino, C., and Febo, A.: UV-visible absorption cross sections of nitrous acid, *J. Geophys. Res.*, 105, 14585-14592, 10.1029/2000JD900003, 2000.

Vandaele, A. C., Hermans, C., Simon, P. C., Carleer, M., Colin, R., Fally, S., Merienne, M. F., Jenouvrier, A., and Coquart, B.: Measurements of the NO₂ absorption cross-section from 42 000 cm⁻¹ to 10 000 cm⁻¹ (238-1000 nm) at 220 K and 294 K, *Journal of Quantitative Spectroscopy & Radiative Transfer*, 59, 171-184, 10.1016/s0022-4073(97)00168-4, 1998.

75