



Supplement of

The new instrument using a TC–BC (total carbon–black carbon) method for the online measurement of carbonaceous aerosols

Martin Rigler et al.

Correspondence to: Martin Rigler (martin.rigler@aerosol.eu)

The copyright of individual parts of the supplement might differ from the CC BY 4.0 License.


1. Example of in-house intercomparison results of TCA08 and AE33 instruments after their assembly to the reference set of instruments

Table S1. Example of intercomparison results of TCA08 and AE33 after their assembly to the reference set of instruments as one of the tests during final inspection procedure. In-house defined requirements for successful intercomparison between new and reference set of instruments are: 1. TCA08: TC concentration range up to 75.000 ng/m³, slope between 0.95-1.05, R² above 0.98 ; 2. AE33: eBC concentrations up to 25.000 ng/m³; slope between 0.95-1.05, R² above 0.98.

	TCA08 serial number	Slope	R ²	N	TC _{max} (ng/m ³)	TC _{min} (ng/m ³)
1	TCA08-S00-0131	1.02	1.00	65	45856	2732
2	TCA08-S00-0132	1.01	1.00	65	45803	2732
3	TCA08-S00-0133	1.00	1.00	19	67274	6286
4	TCA08-S00-0134	1.00	1.00	19	67274	6286
5	TCA08-S00-0135	1.00	1.00	19	67274	6286
6	TCA08-S00-0136	0.99	1.00	139	76684	2498
7	TCA08-S00-0137	1.00	1.00	65	51119	2592
8	TCA08-S00-0138	0.99	1.00	65	51119	2592
9	TCA08-S00-0139	0.96	1.00	65	51119	2592
10	TCA08-S00-0140	1.02	1.00	62	73190	2644
11	TCA08-S00-0141	1.00	1.00	22	64579	3336
12	TCA08-S00-0142	1.01	1.00	136	76466	2823
13	TCA08-S00-0143	1.00	1.00	22	64579	3336
14	TCA08-S00-0144	1.00	1.00	22	64609	3336
15	TCA08-S00-0145	0.97	1.00	137	72992	2823

	AE33 serial number	Slope	R ²	N	eBC _{max} (ng/m ³)	eBC _{min} (ng/m ³)
1	AE33-S08-01036	1.01	1.00	6368	19666	265
2	AE33-S08-01037	1.00	1.00	6368	19478	220
3	AE33-S08-01038	1.02	1.00	6368	21988	139
4	AE33-S08-01039	1.02	1.00	6368	20126	198
5	AE33-S08-01040	1.00	1.00	6368	21338	224
6	AE33-S08-01041	0.99	1.00	6368	20384	231
7	AE33-S08-01042	1.04	1.00	6320	22141	123
8	AE33-S08-01043	1.04	1.00	6368	21380	286
9	AE33-S08-01044	1.03	1.00	6368	21249	223
10	AE33-S08-01045	1.02	1.00	6368	21050	172
11	AE33-S08-01046	0.99	1.00	6368	20433	267
12	AE33-S08-01047	1.03	1.00	6368	20583	230
13	AE33-S08-01048	1.03	1.00	6368	21061	237
14	AE33-S08-01049	1.03	1.00	6320	20699	249
15	AE33-S08-01050	1.02	1.00	6320	20064	243

2. TCA08 setup for semi-continuous denuder breakthrough determination



5

Figure S1. TCA08 setup when (a) sampling and (b) performing semi-continuous denuder breakthrough measurement. Note that the tubing length is identical in both setups. This permits the test to be performed at a permanent installation without disturbing the inlet plumbing.

10


15

20

25

30

3. Evaluation of denuder breakthrough contribution to the TC measurement uncertainty with inline Teflon filter method



5

Figure S2. Example of evaluation of denuder breakthrough contribution to the TC measurement uncertainty with inline Teflon filter method (see Fig. S1 (b) for TCA08 setup). Denuder breakthrough (TC_{DB}) is measured in chamber 1 where sample air stream passes denuder and inline Teflon filter. The TCA08 was operated on a 1-hour time-base, sampling PM_{2.5} fraction at 16.7 LPM. The measurement campaign was conducted between 18 December 2019 and 4 January 2020 at the urban background air quality monitoring of Aerosol d.o.o. company at 46.0715°N, 14.5018°E, elevation 302 m.

10