



Corrigendum to **“Performance and methodological evaluation of a quantum-cascade-laser photoacoustic aerodynamic gradient system for field-scale NH₃ flux measurements” published in Atmos. Meas. Tech., 19, 4989–5006, 2026**

János Fekete¹, Zoltán Bozóki^{1,3}, Csilla Gombi¹, László Horváth³, Zoltán Nagy², Krisztina Pintér²,
Tamás Weidinger^{4,5}, Anna Szabó^{1,3}, and Helga Huszár^{1,3}

¹Department of Optics and Quantum Electronics, Institute of Physics, University of Szeged, 6720 Szeged, Hungary

²Department of Plant Physiology and Plant Ecology, Institute of Agronomy, Hungarian University for Agriculture
and Life Sciences, 2100 Gödöllő, Hungary

³HUN-REN-SZTE Research Group for Photoacoustic Monitoring of Environmental Processes,
University of Szeged, 6720 Szeged, Hungary

⁴Department of Meteorology, Institute of Geography and Earth Sciences, Eötvös Loránd University, 1117 Budapest, Hungary

⁵Institute of Ecology and Botany, HUN-REN Centre for Ecological Research, 2163 Vácrátót, Hungary

Correspondence: László Horváth (horvathl@titan.physx.u-szeged.hu)

Published: 2 October 2026

The incorrect unit was already included in the preprint and was inadvertently inserted during the preparation of the manuscript. This was an error on our side and we sincerely apologise for the mistake.

In the third paragraph of the right-hand column on p. 4991, the sensitivity is incorrectly reported as $17.9 \mu\text{V ppb}^{-1}$. The correct unit is 17.9 nV ppb^{-1} .

This was a typographical error. The reported minimum detectable mixing ratio (detection limit, DL) was calculated using the correct value and is therefore unaffected. This error does not affect any of the results or calculations presented in the article.