

A cavity ring down/cavity enhanced absorption device for measurement of ambient NO₃ and N₂O₅

G. Schuster¹, I. Labazan^{1,*}, and J. N. Crowley¹

¹Max-Planck-Institut für Chemie, Mainz, Germany

*now at: NXP Semiconductors, Nijmegen, The Netherlands

Received: 12 August 2008 – Accepted: 1 September 2008 – Published: 9 September 2008

Correspondence to: J. N. Crowley (crowley@mpch-mainz.mpg.de)

Published by Copernicus Publications on behalf of the European Geosciences Union.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion




Abstract

An inexpensive, compact instrument for the sensitive measurement of NO_3 and N_2O_5 in ambient air at high time resolution is described. Light from a red-emitting laser diode (≈ 662 nm) is coupled off-axis into an optical resonator defined by two highly reflective mirrors to achieve effective absorption paths exceeding 20 km. Light leaking from the cavity is detected either as single ring-down events over ≈ 200 μs following rapid switching of the laser intensity at 200 Hz (Cavity Ring Down mode), or as an integrated intensity (Cavity Enhanced Absorption mode). The operational conditions, detection limits and total uncertainty for the prototype device for NO_3 and N_2O_5 detection/monitoring are assessed and the first measurements in ambient air and from an environmental chamber are described.

1 Introduction

NO_3 and N_2O_5 are important atmospheric trace gases (Wayne et al., 1991) which have a significant impact on the night-time oxidation of some organic compounds (e.g. aldehydes and dimethylsulphide), the partitioning of nitrogen oxides between the NO_x ($\text{NO} + \text{NO}_2$) and NO_y families (where NO_y is the sum of all nitrogen oxides and includes long lived reservoir species such as HNO_3) and also, via heterogeneous reactions, the partitioning of NO_y between the gas and particulate phases. NO_3 is formed predominantly in the reaction of NO_2 with ozone (R1) and is converted to N_2O_5 via further reaction with NO_2 (R2). Due to the thermal decomposition of N_2O_5 , the relative concentrations NO_3 and N_2O_5 are closely linked (Brown et al., 2003) and are a strong function of temperature and $[\text{NO}_2]$. NO_3 absorbs red light efficiently (see Sect. 2.1) and is rapidly photolysed with a lifetime of only a few seconds during daytime, so that NO_3 and (due to the equilibrium R2, R-2) N_2O_5 concentrations are very low. At night, NO_3 can react with NO (R4) or with gas-phase organic species, RH (R5).



AMTD

1, 67–102, 2008

CRD/CEAS measurement of NO_3 and N_2O_5

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion





NO₃ has been measured in the atmosphere on several occasions: by differential optical absorption spectroscopy (DOAS), dating back several decades (see e.g. Platt et al., 1981), matrix isolation ESR (Geyer et al., 1999) and more recently, by laser cavity absorption techniques (e.g. King et al., 2000; Brown et al., 2002) and laser induced fluorescence (e.g. Matsumoto et al., 2005; Wood et al., 2005). N₂O₅ measurements are less common, though techniques using thermal conversion to NO₃ which can be detected using optical (Brown et al., 2001; Simpson, 2003) and mass-spectrometer based methods (Slusher et al., 2004) have been described. Most of the available instruments include expensive, sensitive and bulky components (e.g. pulsed lasers, spectrographs, vacuum pumps) and are not always suitable for research platforms with space/weight constraints.

Depending on air mass, NO₃ concentrations may exceed 100 ppt, and N₂O₅ may be up to a factor of ≈10 larger. Frequently however, low production rates of NO₃ (due e.g. to low [NO₂] and large loss rates of NO₃ (due e.g. to gas phase reactions with organic species) or N₂O₅ (due to heterogeneous processing) result in lifetimes of the order of just a few minutes and levels of less than 10 ppt. Ideally, an instrument for measurement of atmospheric NO₃ or N₂O₅ should be able to detect ppt levels of the radical on timescales of a few seconds in a well defined air-mass (point measurement). In addition, it should be lightweight, have a small footprint and low power consumption and, to facilitate widespread usage, should be made from relatively inexpensive, robust components. In its present state, our instrument is characterised by a total weight of <40 kg, power consumption of <500 W and a total cost of close to 20 kEuro.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO₃
and N₂O₅**G. Schuster et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

The instrument we describe is an optical-cavity/laser absorption device. It is operated either in “cavity ring-down” (CRD) mode or “cavity enhanced absorption” mode (CEAS), which is closely related to “integrated cavity output spectroscopy” (ICOS). The basic principles of CRD and CEAS are described in recent, comprehensive review articles (Berden et al., 2000; Brown, 2003; Mazurenka et al., 2005). Here, we present a detailed description of the prototype device, results of laboratory characterisation, first measurement of N₂O₅ (+NO₃) mixing ratios outside the Max Planck Institute building in Mainz and some results from the first deployment in a multi-instrument NO₃/N₂O₅ inter-comparison campaign. The inter-comparison was carried out at the SAPHIR environmental chamber (Rohrer et al., 2005) in Jülich, Germany in the summer of 2007 and the results will be presented in detail in forthcoming publications (Apodaca et al., 2008¹; Dorn et al., 2008²).

2 Experimental set up

The instrument makes use of the well characterised, strong absorption band of NO₃ between circa 610 and 665 nm. Recent developments in laser optics, including the increasing availability of powerful, cheap, red emitting diode lasers and mirrors with

¹Apodaca, R. L., Simpson, W. R., Brauers, T., Brown, S. S., Cohen, R. C., Crowley, J., Dorn, H. P., Dubé, W. P., Fry, J., Fuchs, H., Haseler, R., Heitmann, U., Kato, S., Kajii, Y., Kiendler-Scharr, A., Labazan, I., Matsumoto, J., Nishida, S., Tillmann, R., Rohrer, F., Rollings, A. W., Schlosser, E., Schuster, G., Tillmann, R., Wahner, A., Wegener, R., and Wooldridge, P. J.: Intercomparison of N₂O₅ sensors using SAPHIR reaction chamber, Atmos. Chem. Phys. Discuss., in preparation, 2008.

²Dorn, H. P., Apodaca, R. L., Ball, S. M., Brauers, T., Brown, S. S., Crowley, J. N., Dube, W. P., Fuchs, H., Häseler, R., Heitmann, U., Jones, R. L., Labazan, I., Langridge, J., Meinen, J., Platt, U., Pöhler, D., Rohrer, F., Ruth, A. A., Schlosser, E., Schuster, G., Shillings, A., Simpson, W., Thieser, J., Varma, R., Venables, D., and Wahner, A.: Intercomparison of NO₃ radical detection instruments in the Atmosphere Simulation Chamber SAPHIR., Atmos. Chem. Phys. Discuss., in preparation, 2008.

extremely high reflectivity at appropriate wavelengths can be combined with the large absorption cross section of NO_3 close to 660 nm to make cavity based detection methods a powerful tool for ambient $\text{NO}_3/\text{N}_2\text{O}_5$ measurements.

The central component of our instrument is an optical cavity (resonator length ≈ 70 cm) flushed with ambient air which can be operated either in cavity ring down (CRD) or cavity enhanced absorption (CEAS) mode. NO_3 , which possesses a distinct and intensive absorption feature at ≈ 662 nm is measured directly, whereas N_2O_5 ($+\text{NO}_3$) is detected by heating part of the inlet and cavity to 95°C to dissociate it to NO_3 (-R2). Using a value of $k_{-2}=32\text{ s}^{-1}$ at this temperature and 1 atmosphere pressure (Sander et al., 2006), the lifetime of N_2O_5 w.r.t. dissociation to NO_3 is less than 0.1 s. By comparison, the lifetime at room temperature is ≈ 100 s. Assuming no loss of NO_3 via thermal dissociation, operation at 95°C means that the sum of NO_3 and N_2O_5 concentrations is measured.

A schematic diagram illustrating the various optical, gas-flow and electronic components is given in Fig. 1. In this paper the instrument is described for the first time

2.1 Optical configuration

The high finesse optical cavity consists of two, 1 inch diameter dielectric mirrors (radius of curvature=1 m, reflectivity ≈ 99.998 , Los Gatos Research) located ≈ 70 cm apart in home built mounts (see inset to Fig. 1). The laser is coupled off-axis into the cavity to spatially separate the multiple reflections within the cavity and increase the number of passes before the re-entrant condition is fulfilled (i.e. when the beam retraces its original path). Compared to on-axis alignment, this reduces optical feedback and cavity intensity fluctuations, removes the need for fine mechanical stability of the cavity, and increases the density of the cavity mode spectrum and free spectral range (Paul et al., 2001; Kasyutich et al., 2002; Bakhirkin et al., 2004; Ayers et al., 2005). Adjustment of the cavity is achieved by monitoring the cavity output for maximum ring down time, but also avoiding spurious, high-intensity cavity emissions due to incomplete suppression of cavity resonances (mode co-incidence between laser and unstabilised cavity). The

CRD/CEAS measurement of NO_3 and N_2O_5

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO₃
and N₂O₅**G. Schuster et al.

cavity gain under such conditions is similar to that achieved when the laser is coupled to a single cavity mode, but the transmitted power level is drastically reduced (Paul et al., 2001). The laser current is modulated to slightly broaden the laser spectral bandwidth ensuring that the re-entrance length was longer than the laser coherence length. In CEAS mode, current modulation of the laser improves the signal/noise ratio (S/N) of the cavity emission by \approx an order of magnitude but, as the wavelength jitter was small, does not reduce the effective absorption cross section. A similar improvement in S/N was observed in an earlier version of the instrument in which the cavity length was dithered at 50 Hz by use of a piezo-electric transducer attached to one mirror mount e.g. (Kasyutich et al., 2002). The combination of both laser wavelength jitter and cavity dither did not improve the S/N compared to the use of either. The cavity material is Pyrex glass (ID 15 mm), coated with a thin film of Teflon (DuPont FEP 121 a). The laser used is a current modulated, 120 mW, CW laser diode emitting close to 662 nm (Mitsubishi ML101 J27) with current and temperature control (34.70°C) using a ThorLabs diode laser controller ITC 502.

Operation in both CEAS and CRD modes is possible with simultaneous measurement of integrated light intensity or ring down decay constants, respectively. In CRD mode, laser modulation is at 200 Hz (100% amplitude, square-wave) and at 200 kHz (10% amplitude, sine-wave) for CEAS. The two modes of operation result in different laser mode structures (integrated over many modulation cycles) which were measured at a resolution of 0.03 nm using a 1 m monochromator (3600 lines mm⁻¹ grating) diode array (Instaspec II) set-up. Wavelength calibration was with three Ne lines and is accurate to \approx 0.1 nm. The laser emission spectra are illustrated in Fig. 2, along with the NO₃ absorption spectrum (Yokelson et al., 1994). In both CEAS and CRD modes, the major emission is between 661.7 and 662 nm, though CEAS shows more complex structure with an additional major mode at 662.06 nm and several, weak, blue shifted modes. The spectra were observed to be unchanged over a 3 month period prior to and after the NO₃/N₂O₅ inter-comparison we report on below. In addition, following coupling of the cavity output to the monochromator, we obtained identical spectra, indicating no

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

significant, preferential excitation of some cavity modes.

The effective cross sections calculated from the overlap of laser emission and NO₃ absorption spectrum were 2.17×10^{-17} and 2.20×10^{-17} cm² molecule⁻¹ for CEAS and CRD, respectively at 298 K. The temperature dependence of the NO₃ cross section was taken from Orphal et al. (2003) the absolute values from Yokelson et al. (1994). A broadening of the NO₃ absorption spectrum leads to a reduction of the σ_{\max} at ≈ 662 nm from 2.21×10^{-17} at 22°C to 1.65×10^{-17} at 95°C.

A small fraction of the laser intensity enters the cavity through a 662 nm coated dielectric mirror and exits via an identical mirror to be measured by a photomultiplier (PMT) located behind a 662 nm interference filter. The distance between the two dielectric mirrors is 70 cm, curtain flows of zero-air purge the volume extending ≈ 5 cm in front of the mirrors, protecting them from aggressive trace gases and aerosol deposition.

2.2 Data acquisition and analysis

In CRD mode (laser fully modulated at 200 Hz) the pre-amplified PMT signal is digitised and averaged with a 100 MHz, 9 bit oscilloscope (Tektronix TDS3014B). The sliding mean of 64 ring-down profiles (each 10 000 data points) was transferred to a PC at 3 Hz. The effective time resolution was ≈ 5 s. The cavity loss due to absorption at 662 nm is calculated from the change in ring down constant in the presence of an absorber and converted to a concentration of NO₃ using the effective cross section of NO₃ at the experimental temperature and the established relation (Berden et al., 2000; Mazurenka et al., 2005).

$$[\text{NO}_3] = \frac{\Delta k_{rd} L}{\sigma_{\text{NO}_3} c d} \quad (1)$$

where [NO₃] is the concentration of NO₃ (molecule cm⁻³) Δk_{rd} is the difference in the ring-down decay constant with and without NO₃, L is the distance between the cavity

CRD/CEAS measurement of NO₃ and N₂O₅

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.

mirrors (70 cm), d is the length of the cavity which is filled with absorber (and c is the speed of light $\approx 2.998 \times 10^{10}$ cm s⁻¹). By variation of the mirror purge gas flow whilst monitoring the CEAS signal due to a constant flow of NO₂ into the cavity (NO₂ also absorbs at 662 nm), we were able to calculate a value of L/d of 1.05 ± 0.03 . Note that the ring-down time constant, τ , is equal to $1/k_{rd}$ and for a cavity free of absorbing species was typically close to $100 \mu\text{s}$, indicating an effective optical path length of ≈ 27 km.

In CEAS mode, the amplified PMT output is averaged over 200 ms, digitised and transferred to the PC. The change in integrated signal intensity in the presence of an absorber is converted to a concentration of NO₃ using Eq. (2) (Berden et al., 2000; Mazurenka et al., 2005).

$$[\text{NO}_3] = \frac{1 - R}{d\sigma_{\text{NO}_3}} \left(\frac{I_0}{I} - 1 \right) \quad (2)$$

Where R is the mirror reflectivity (obtained from ring-down measurements in the absence of absorbing species) and I_0 and I are the light intensities exciting the cavity without and with absorber, respectively. The reflectivity is related to k_{rd} and the optical path-length (l) by:

$$k_{rd} = \frac{c(1 - R)}{L} \text{ or } l = \frac{d}{(1 - R)} \quad (3)$$

Note that operation in CEAS mode thus requires knowledge of the reflectivity (R) or the path length (l). Both can be calculated by measuring k_{rd} in the absence of absorber or by adding a known concentration of absorber (with known cross section at the laser wavelength) to the cavity.

In most experiments reported here, although the laser was modulated at 200 Hz (i.e. for CRD mode), both the CRD and CEAS signal were recorded simultaneously. Spot checks showed the agreement between concentrations of NO₃ made using CRD and CEAS modes to be generally excellent (<5 % deviation) though sometimes larger deviations were observed (see later).

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



2.3 Gas flows and titration of NO₃

Air samples are drawn into the instrument through a ≈ 50 cm long piece of PFA tubing of 9.5 mm internal diameter, at a (typical) flow rate of 12 litres per minute (SLM). Of this flow, 8 SLM are drawn through a further 18 cm piece of PFA tubing and Teflon (FEP) coated aluminium incorporating a 22 mm diameter Teflon membrane filter (pore size 1 or 2 μm) to eliminate particles from the air stream. Following the Teflon filter, the air sample passes into a Teflon coated (FEP 121a), inverted, thermally insulated glass T-piece (i.d. 1.8 cm) which can be heated to 95°C using resistive heating wire wrapped at close spacing around the glass (for measurement of NO₃+N₂O₅) or maintained at ambient temperature (for measurement of NO₃ only). The horizontal section of the glass T-piece ($l=61$ cm) defines the gas flow within the optical cavity, in which the average residence time is ≈ 1 s. The vertical section is long enough to ensure complete titration of NO₃ with NO and/or thermal conversion of N₂O₅ to NO₃. Temperatures in both vertical and horizontal sections are monitored with thermistors with an accuracy of 0.2°C. The cavity mirrors are purged with 2 \times 150 sccm of bottled, filtered, zero air.

In both CEAS and CRD modes of operation, the ring down constant or integrated output intensity in the absence of absorbing species (chemical “zero”) is measured by adding $\approx 10^{12}$ molecule cm⁻³ of NO to the inlet upstream of the Teflon filter. Based on the evaluated rate coefficient for Reaction (R4) ($k_4=1.8 \times 10^{-11} \exp(110/T) \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$) (Atkinson et al., 2007) this amount of NO should result in complete titration of NO₃ within ≈ 0.1 s (3–half lives). This method of determining the baseline signal allows separation of the NO₃ signal from other atmospheric absorbers such as NO₂, O₃ and H₂O and is superior to e.g. flushing the cavity with zero air. The importance of this is illustrated in Fig. 3, which shows the optical density (per cm) of a number of atmospheric trace gases at typical mixing ratios. Clearly, fluctuating H₂O mixing ratios could represent a problem if the chemical zero is not measured frequently enough and at the same H₂O mixing ratio as the NO₃ measurement. In addition, we note that adding NO titrant to air samples containing O₃ at typical tropospheric mixing

AMTD

1, 67–102, 2008

CRD/CEAS measurement of NO₃ and N₂O₅

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



ratios (>20 ppb) results in removal of O_3 and formation of NO_2 . The effect on the total absorption at 662 nm is generally small and corrections can be accurately applied (see Sect. 3.3).

2.4 Raw data in CEAS/CRD modes

5 A typical set of raw data obtained whilst sampling NO_3 -containing air from the SAPHIR chamber in CEAS mode is given in Fig. 4 (upper panel). The titration was carried out over a 20 s period every 100 s following mixing of NO_2 (≈ 3 ppb) and O_3 (≈ 100 ppb) in the chamber. As the signal integration time is less than the residence time of gas in the cavity, the response of NO_3 to addition of NO is rapid (≈ 5 s for these experiments
10 at a flow rate of 3 SLM) compared to the duty cycle. The raw CEAS signal as shown in Fig. 4 is an average over 200 ms and the random noise levels ($\Delta I/I_0$ of $5-10 \times 10^{-4}$) associated with this signal are equivalent to ≈ 1 ppt of NO_3 . We note that this noise limited detection limit is superior to that reported for modern pulsed CRD field devices which report e.g. 2σ detection limits of 0.5 ppt in 5 s (Brown et al., 2002). On the
15 other hand we note that increasing the flow rate through the inlet and cavity to 8 SLM resulted in a significant increase in the signal noise, presumably due to turbulence and pressure fluctuations in the cavity. As we show later, operation at 8 SLM is desirable as heterogeneous losses of NO_3 can be minimised.

The CEAS signal is dramatically reduced in quality by removal of the particle filter
20 (lower panel) as particle induced noise of ≈ 100 ppt NO_3 equivalents is apparent. Note that the noise is not randomly distributed around the mean signal, but is biased to low values, indicating that individual particles on passage through the cavity cause substantial intensity fluctuations due to light extinction at 662 nm. We note that the air sample extracted from the SAPHIR chamber was for this experiment nominally “particle
25 free” indicating extreme sensitivity of this method to particles.

Raw data obtained in CRD mode is displayed in Fig. 5 which presents a ring-down profile (upper panel) with $k_{rd} \approx 10\,563 \pm 5 \text{ s}^{-1}$ (errors are 2σ statistical) or a ring-down time of $\approx 95 \mu\text{s}$ corresponding to an optical path length of ≈ 27 km (Eq. 3). As men-

CRD/CEAS measurement of NO_3 and N_2O_5

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



tioned above, the data were obtained by averaging (in the oscilloscope) 64 ring down profiles each of 10 000 data points, resulting in a sliding average value over ≈ 5 s. The first $\approx 5\%$ of the ring-down profile were often non-exponential and were rejected online during data transfer from the oscilloscope and the last 2000 data points were used to establish the signal baseline, which may be distinct from zero due to electrical/optical offsets. Once the baseline was subtracted, ≈ 2000 data points were analysed by least squares regression to extract k_{rd} before the next (5 s averaged) ring down profile was transferred from the scope. The resulting data density was ≈ 3 ring down constants per second. When working in CRD mode, a 50% titration duty cycle was generally used with 1 min measurement followed by 1 min of titration. When performing optimally, 1σ random noise levels of $\delta k_{rd} = 20 \text{ s}^{-1}$ corresponding to minimum detectable changes in the ring-down time ($\Delta\tau_{\min}$) of $\approx 0.2 \mu\text{s}$ (in ≈ 5 s averaging) were achieved, corresponding to NO_3 concentrations of ≈ 2 ppt and defining the detection limit of the device in CRD mode. Our detection limit is thus comparable to the value of 1.4 ppt in 4.6 s reported by Ayers et al. (2005) for a similar off axis, CRD set up using a CW diode laser. More generally, the detection limit can be expressed as the minimum detectable extinction α (where $\alpha = \Delta k_{rd} L / cd$) per integration interval as $= 1 \times 10^{-9} \text{ cm}^{-1} \text{ Hz}^{-1/2}$

Although the noise limited detection limit in CEAS operation mode was (at low flows) better than in CRD mode, CEAS suffers from the disadvantage that the optical path length can vary by several percent over a period of hours and thus needs to be determined frequently to avoid systematic error. In addition, an increase in noise in the CEAS signal was observed when the instrument was operated at high flows (> 3 SLM). In contrast, the CRD mode is insensitive to laser intensity drifts and the minimum measurable Δk_{rd} did not deteriorate significantly when flow rates were increased between 3 and 8 SLM.

**CRD/CEAS
measurement of NO_3
and N_2O_5** G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



3 Data correction, assessment of systematic errors and detection limit

The use of an extractive approach for measurement of ambient NO₃ or N₂O₅ requires knowledge of the transmission efficiency of both species through the inlet lines and the particle filter. We also need to assess systematic errors resulting e.g. from the procedure to titrate NO₃. These are discussed in detail below and used to derive the overall uncertainty of the device for measurement of NO₃ and N₂O₅.

3.1 Inlet/cavity transmission of NO₃ and N₂O₅

Estimation of the loss rates of NO₃ during transmission through the inlet tubing and the cavity was made on four separate occasions during the SAPHIR inter-comparison by variation of the flow rate of air drawn through the instrument and thus the residence time. The tests were carried out when stable concentrations of NO₃ were observed in the chamber (small corrections were applied to take into account mixing ratio drifts) and with different NO₃ concentrations and NO₃/N₂O₅ ratios in the chamber. Results from two days are displayed in Fig. 6, in which two different filters were used. From the data displayed, the decay constants obtained were $(8.2 \pm 1.0) \times 10^{-2} \text{ s}^{-1}$ and $(9.1 \pm 2.0) \times 10^{-2} \text{ s}^{-1}$, which, within experimental uncertainty, are identical. Although the inlet and cavity are constructed from Teflon and Teflon-coated glass tubing of slightly different diameter (and thus different flow velocities in different parts of the system) to a first approximation we can use the average loss rate constant (k_{obs}) of NO₃ ($8.2 \times 10^{-2} \text{ s}^{-1}$) to derive a sticking coefficient (γ , the per collision efficiency of removal of NO₃ from the gas-phase) for NO₃ on Teflon tubing via:

$$\gamma = \frac{2rk_{\text{obs}}}{\bar{c}} \quad (4)$$

where r is the average radius of the tubing and \bar{c} the mean molecular velocity of NO₃ ($32\,000 \text{ cm s}^{-1}$) as derived from the Maxwell expression. With $r=0.75 \text{ cm}$, we derive

AMTD

1, 67–102, 2008

CRD/CEAS
measurement of NO₃
and N₂O₅

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



$\gamma=4\times 10^{-6}$, indicating that only 4 collisions per million result in removal from the gas-phase.

With a total flow of 8 SLM (corresponding to an average residence time of ≈ 1.5 s), the associated correction factor from these two days is 1.14. Correction factors for all 4 days are listed in Table 1. By comparison, use of a 3 SLM flow would result in a correction factor of ≈ 1.4 .

From Table 1 we see that the NO_3 inlet loss was fairly constant over a period of 7 days, with an average correction factor of about 1.13 when data from the 14th, 15th and 21st are taken into consideration. The exception is the 20th of June when the flow test was conducted with a large $[\text{N}_2\text{O}_5]/[\text{NO}_3]$ ratio in the chamber. On this day, the apparent NO_3 loss rate constant was much smaller. This is due to N_2O_5 thermal dissociation within the residence time in inlet and cavity (1.5 s) which disguises the true NO_3 loss. At low N_2O_5 the rate of formation of NO_3 via N_2O_5 thermal decomposition is small and we measure the “true” NO_3 wall loss rate. The appropriate correction factor is thus the larger one. Numerical simulations (Curtis and Sweetenham, 1987) of the $\text{NO}_3/\text{N}_2\text{O}_5$ equilibrium with different $\text{N}_2\text{O}_5/\text{NO}_3$ ratios/amounts were thus performed to assess the size of this effect. The simulations were initiated by reacting NO_2 with O_3 and allowing the system to approach steady NO_3 and N_2O_5 concentrations before introducing a NO_3 loss. The relative $\text{NO}_3/\text{N}_2\text{O}_5$ concentration was varied by changing the NO_2 concentration and the temperature. The simulation results (Table 2) show that, as expected, at high N_2O_5 , the decay of NO_3 due to wall loss is not exponential (NO_3 is reformed on the time scale of its wall loss) whereas at low N_2O_5 it is. This is illustrated by the simulation results displayed in Fig. 7. However, at our short residence times the simple analytical correction gives (within a few percent) the same factor as the full simulation (accounting for N_2O_5 thermal dissociation). So, to a good approximation, as long as the residence time is short and temperatures are close to ambient, a single correction factor can be applied. The error on this correction factor is increased to 1.13 ± 0.1 to take into account effects such as those discussed above.

CRD/CEAS measurement of NO_3 and N_2O_5

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO₃
and N₂O₅**G. Schuster et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

It has frequently been shown (see e.g. Simpson, 2003; Aldener et al., 2006) that the loss of N₂O₅ through Teflon piping/filters is minor compared to NO₃. Nonetheless, when operating in N₂O₅ mode (i.e. heated cavity) flow tests similar to those described above showed that the N₂O₅ signal was lower at long residence times. Recall that, in this system, N₂O₅ is measured as NO₃ following thermal conversion before flowing into the hot cavity and thus provides a measure of the sum of N₂O₅+NO₃. Losses of NO₃ on the heated inlet and cavity walls therefore have to be taken into account to derive a corrected N₂O₅+NO₃ summed mixing ratio. Using the same approach as for NO₃, the apparent loss rate constant for N₂O₅ was determined to be $5.5 \times 10^{-2} \text{ s}^{-1}$, resulting in a correction factor of 1.1 at a flow rate of 8 SLM. This correction is only rigorous in the limit of N₂O₅/NO₃ → infinity, as it does not take into account the different rate of loss of NO₃ and N₂O₅ to the walls, or the filter loss of NO₃. The correction factor for the measurement of N₂O₅+NO₃ was thus determined as 1.1 (±0.1).

3.2 Filter losses of NO₃ and N₂O₅

Loss of NO₃ during transmission through the filter was tested in the laboratory. Teflon filters with 1 μm (Millipore FALPO2500) and 2 μm (Pal Teflo R2PJ047) pore sizes were used. Due to flow restriction, the 1 μm filters caused a pressure drop of ≈30 mbar when sampling 8 SLM air, whilst use of the 2 μm filters resulted in a pressure drop of just ≈5 mbar. A 40 L glass vessel (coated internally with a very thin PFA film and blackened outside) was used as reaction volume in which ≈100 ppb O₃ and 10–20 ppb NO₂ were mixed and reacted over a period of ≈40 min. The exhaust (250 sccm) passed into a heated (≈80°C) Teflon pipe where it resided for ≈1.5 s before being dynamically diluted by a flow of 8000 sccm air and transported into the cavity at room temperature. The final flow contained about 100–200 ppt NO₃ at atmospheric pressure (and <10 ppt N₂O₅). The NO₃ mixing ratio was measured with and without a filter several times to derive the transmission. Summary data are shown in Fig. 8. Drifts in the NO₃ signal, presumably caused by pressure drops/increases when the filters were removed/added and contamination of the inlet by laboratory air whilst open precluded very accurate

measurement of the filter transmission. However, it is clear that the filter has a transmission of between 80 and 90%. This is consistent with previously measured NO₃ transmission efficiencies of Teflon membrane filters (Aldener et al., 2006). It is also apparent that there is no significant difference between the transmission of the 1 μm and 2 μm filters, though the latter is preferable as the pressure drop across the filter is smaller. NO₃ data sets thus require correction by a factor of 1.18±0.1 to take into account loss at the filter.

Loss of N₂O₅ during transmission through the filter was tested in a similar manner, the difference being that the exhaust of the reaction volume was not heated and that N₂O₅ (with a minor amount of NO₃) passed through the filter before entering the hot inlet to be converted to NO₃. A filter test dataset is presented in Fig. 8. Within the experimental scatter no loss of N₂O₅ at the filter could be observed. From the statistical noise we estimate a transmission of 100±4%. A “correction factor” of 1±0.04 may be applied to all N₂O₅ data sets to account for filter loss. Note that the filter-transmission efficiencies determined above refer to use of clean filters which had not prior exposure to polluted air. In the course of the SAPHIR inter-comparison it was noted that large aerosol loadings could reduce the filter transmission over the course of less than 1 h (Dorn et al., 2008), suggesting that frequent changing of the filter is key to accurate concentration measurement.

3.3 NO₂ impurities and NO+O₃ reaction during titration

The derivation of the “chemical zero” was conducted by titrating NO₃ with NO. Typically, 3 sccm of a 100 ppm mix of NO/N₂ was added to the total flow of 8000 sccm to result in a concentration of NO of approximately 10¹² molecule cm⁻³ (40 ppb). NO₂ impurity at the 1% level (this is easily achieved by trap-to-trap distillation of the NO sample) would then result in 1×10¹⁰ molecule cm⁻³ NO₂ or 1×10⁶ NO₃ (0.05 ppt) equivalents in the cavity. This amount of NO₂ is thus too low to warrant correction. Similarly, although the reaction of NO₃ with NO generates two NO₂ molecules, the ratio of cross sections

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



of NO_3 and NO_2 ($\approx 10^4$) means that this effect is not significant. NO can however be converted to NO_2 if the ambient air contains sufficient O_3 . Taking $\text{O}_3=100$ ppb and $k(\text{NO}+\text{O}_3)=1.9\times 10^{-14}$ cm^3 molecule $^{-1}$ s $^{-1}$ and an average reaction time of 1.5 s we can calculate that only 3 ppb of O_3 are removed (and 3 ppb of NO_2 are generated).
5 As the NO_2 cross section is larger (factor 2) than the O_3 cross section at 662 nm, the overall correction (in terms of NO_3 equivalents) is 0.3 ppt which can be a significant correction only at low NO_3 mixing ratios.

3.4 Estimation of detection limit and total uncertainty of prototype device

When working in CRD mode, random fluctuations in the ring down constant result in a
10 detection limit of 2 ppt (5 s averaged data) for both NO_3 and $\text{NO}_3+\text{N}_2\text{O}_5$. The accuracy of the measurement is controlled by systematic errors in the absorption cross section of NO_3 , errors in the correction for filter loss, inlet transmission and effective absorption path (taking purged volumes into account). The correction factors (and associated errors), which are different for NO_3 and $\text{N}_2\text{O}_5+\text{NO}_3$ are listed in Table 3. The overall
15 uncertainty of the NO_3 measurement using CRD mode is estimated at about 14%, though at very low NO_3 concentrations and high ambient O_3 concentrations the effect of NO_2 formation in the reaction of $\text{NO}+\text{O}_3$ needs to be taken into account. Typically this will be less than 1 ppt as shown in Table 3 and discussed elsewhere in the manuscript. The major source of systematic error in the measurement of both ambient NO_3 and
20 N_2O_5 is thus the transmission of both trace gases through the inlet and filter, both of which can become chemically reactive due to deposition of aerosols.

We have indicated that the random noise associated with the CEAS measurement is less than that for CRD so that, hypothetically, a detection limit of 1 ppt in less than
25 1 s could be achieved. However, this precision was only observed when the instrument was operated at low flow rates, which resulted in significant inlet loss of the NO_3 radical and loss of accuracy. In addition, the CEAS measurement can suffer from systematic bias due to drift in laser intensity between titrations and we frequently observed spo-

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



radic drifts in the cavity output intensity that corresponded to ≈ 5 – 10 ppt equivalents of NO_3 over periods of a minute. Fluctuations in both stray light reaching the PMT and spurious laser emissions can contribute to intensity fluctuations in CEAS mode. These issues do not impact CRD measurements as ring down profiles are measured over very short timescales ($200 \mu\text{s}$) and reject the first few μs in which light of inappropriate wavelengths is rejected from the cavity. In addition, when operated in pure CEAS mode (10% current modulation) we do not measure the ring-down time and thus the path length frequently, so that drifts in this parameter propagate directly into the concentration derived.

For the reasons outlined above, despite its lower sensitivity for this instrument, the datasets described below were obtained in CRD mode (i.e. 100 % current modulation at 200 Hz). For comparison we also took the CEAS signal, which is still available in this mode of operation, but with a reduced duty cycle of 50%.

4 Exemplary measurements of NO_3 and N_2O_5

The performance of the device is best illustrated with NO_3 measurements taken either in ambient air or at the SAPHIR chamber Fig. 9 shows a ≈ 4 h dataset taken on the 13th of June at the SAPHIR chamber. Note that the dataset has been corrected for inlet and filter loss as described above. The formation of NO_3 was initiated by the mixing of ≈ 200 ppb of O_3 and 0.5 ppb of NO_2 in the dark chamber at $\approx 09:05$ UTC. NO_3 increased over the next ≈ 40 min before production and loss rates roughly balanced to result in a quasi stable mixing ratio. At this point the inlet and cavity flow rates were varied to check the transmission of the sampling lines for NO_3 . The large variations (reductions) in mixing ratio between $\approx 10:05$ and $10:20$ highlight the importance of characterising inlet losses and maintaining high flow rates (short residence times). The chamber roof was open or partially open between $11:15$ and $11:24$ so that no NO_3 was observed. After closing the roof and refuelling with 6 ppb of NO_2 at $\approx 11:40$ an strong increase in NO_3 was observed, with maximum values of ≈ 250 ppt reached at $12:06$, after which the

CRD/CEAS measurement of NO_3 and N_2O_5

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



chamber was opened again. Although the device was operated in CRD mode (200 Hz laser modulation) both CRD and CEAS signals were acquired and evaluated according to Eqs. (1) and (2). The measurement/titration cycle was carried out at an interval of 1 min, so that 5 s averaged data are generated at a rate of 1 data point per 2 min. The ring down time constant in the absence of NO_3 was used to calculate the optical path length in the CEAS measurements. There is clearly very good agreement between the two datasets, though there are indications of a slight deviation at higher mixing ratios whereby the CRD analysis results in the larger number. Analysis of several other datasets, revealed that when deviations between CRD and CEAS occurred, the CRD result was always higher, though the difference never exceeded 10%. One possible explanation is the detection of amplified spontaneous emission from the diode laser which (depending on wavelength) may transmit the cavity efficiently and contribute to the overall integrated intensity signal in CEAS. In CRD mode light of the “wrong” wavelength will leave the cavity with a ring down time on much shorter time-scale than 662 nm light and as we reject the first few percent of the decay for analysis) would not influence the measurement of k_{rd} . Such CEAS-specific problems can be minimised by use of a narrow band interference filter in front of the photomultiplier which is matched to the mirror transmission curve. We also note that varying electrical offsets (caused e.g. by changing temperature of the PMT or other data acquisition electronics) may result in falsification of the CEAS data analysis if the zero signal (i.e. when the laser light is blocked) is not measured occasionally. This would appear to be the most likely explanation for the differences in CEAS and CRD sometimes observed in our datasets. We note that a comparison of our CRD data for NO_3 and N_2O_5 with those of several other instruments operated at the SAPHIR chamber showed very good agreement, as will be discussed in publications in preparation (Apodaca et al., 2008¹; Dorn et al., 2008²).

A dataset (this time N_2O_5 measurements) obtained in ambient air is displayed in Fig. 10. The measurements were conducted on the night of the 5th–6th of October 2007 by sampling air (8 SLM) through 1/4” Teflon tubing outside of the laboratory win-

**CRD/CEAS
measurement of NO_3
and N_2O_5** G. Schuster et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

dow on the 2nd storey (height ≈ 15 m) at the Max-Planck building. The appropriate correction factor for transmission through the Teflon was calculated from the measured loss rate in PFA tubing described above. Measurements were actually performed on a number of nights from the Max-Planck building, but N_2O_5 was seen only occasionally.

5 The proximity of the institute to several busy motorways results in large local NO mixing ratios, which frequently titrate O_3 at night so that NO_3 (and N_2O_5) are not detectable. NO_2 and O_3 mixing ratios on the 5th–6th of October are also plotted in Fig. 10. They were taken from the closest air quality monitoring station (≈ 5 km away from the MPI) and indicate that conditions were favourable for NO_3 formation. Relatively high NO_2 (10–30 ppb) resulted in large rates of NO_3 formation (O_3 was present at 20 ± 10 ppb) and a shift in the $\text{NO}_3/\text{N}_2\text{O}_5$ equilibrium towards N_2O_5 which was detectable shortly prior to local sunset and which reached a maximum mixing ratio of almost 100 ppt at 02:00 on the 6th. The large variability in the N_2O_5 signal (especially between 22:00 and 24:00) may be due to slight changes in wind direction resulting in sampling of air masses with different chemical history/aerosol loading or due to variable loss rates to e.g. the MPI building (the inlet protruded only ≈ 50 cm from the outside wall) and surrounding architecture. The high N_2O_5 mixing ratio did not persist through the night, but decreased between 03:00 a.m. and sunrise. The accompanying growth in NO (and reduction in O_3) during this time suggests that titration of NO_3 in Reaction (R4) was responsible for removal of N_2O_5 .

5 Conclusions and outlook

We have designed, constructed and tested a new instrument for ambient $\text{NO}_3/\text{N}_2\text{O}_5$ measurements. The first field deployment and participation at an $\text{NO}_3/\text{N}_2\text{O}_5$ inter-comparison have shown that the device can be operated at detection limits of a few ppt in a few seconds, which is adequate for measurement of these trace gases under most conditions. Future improvements to the device will include the capability to measure both NO_3 and N_2O_5 simultaneously in two independent detection axes using a single

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



5 laser diode. Data acquisition rates (and instrument weight) will be improved by replacement of the digital oscilloscope with a PC data-acquisition card, resulting in improved detection limits. The laser emission (both before entering and exiting the cavity) will be checked in the field by use of a small, high resolution spectrograph in order to reduce uncertainty associated with e.g. cavity adjustment dependent changes in the effective NO₃ cross-section. The irises in front of the laser (to prevent reflected light from first cavity mirror entering the laser diode) will be replaced by an optical isolator to increase the power throughput. It is also planned to work with shorter gas residence times in the inlet and cavity to reduce NO₃ losses (Fuchs et al., 2008) and to improve the duty cycle so that at only 10 s of each minute sampling time is dedicated to the titration of NO₃. A major limitation at present is the requirement to manually change the aerosol particle filter at regular intervals through the night. This task will in future be automated (Dubé et al., 2006) to preserve high transmission efficiencies of N₂O₅ and NO₃. Indeed, we foresee that the instrument will operate in fully automated mode to allow prolonged measurements at remote locations and implementation in space/operator limited aircraft platforms.

20 *Acknowledgements.* We thank T. Brauers and H.-P. Dorn for organising the NO₃-N₂O₅ inter-comparison at the SAPHIR environmental chamber in Jülich. We also thank the staff at Jülich and the other participants at the inter-comparison for their technical support and encouragement. We also extend our thanks to G. W. Harris, who encouraged us to construct the instrument.



25 MAX-PLANCK-GESELLSCHAFT

The publication of this article is
financed by the Max Planck Society.

**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



References

- Aldener, M., Brown, S. S., Stark, H., Williams, E. J., Lerner, B. M., Kuster, W. C., Goldan, P. D., Quinn, P. K., Bates, T. S., Fehsenfeld, F. C., and Ravishankara, A. R.: Reactivity and loss mechanisms of NO_3 and N_2O_5 in a polluted marine environment: Results from in situ measurements during New England Air Quality Study 2002, *J. Geophys. Res.*, 111, D23S73, doi:10.1029/2006JD007252, 2006.
- Atkinson, R., Baulch, D. L., Cox, R. A., Crowley, J. N., Hampson, R. F., Hynes, R. G., Jenkin, M. E., Kerr, J. A., Rossi, M. J., and Troe, J.: IUPAC Subcommittee for gas kinetic data evaluation. Evaluated kinetic data: <http://www.iupac-kinetic.ch.cam.ac.uk/>, 2007.
- Ayers, J. D., Apodaca, R. L., Simpson, W. R., and Baer, D. S.: Off-axis cavity ringdown spectroscopy: application to atmospheric nitrate radical detection, *Appl. Optics*, 44, 7239–7242, 2005.
- Bakhrkin, Y. A., Kosterev, A. A., Roller, C., Curl, R. F., and Tittel, F. K.: Mid-infrared quantum cascade laser based off-axis integrated cavity output spectroscopy for biogenic nitric oxide detection, *Appl. Optics*, 43, 2257–2266, 2004.
- Berden, G., Peeters, R., and Meijer, G.: Cavity ring-down spectroscopy: Experimental schemes and applications, *Int. Rev. Phys. Chem.*, 19, 565–607, 2000.
- Brown, S. S.: Absorption spectroscopy in high-finesse cavities for atmospheric studies, *Chem. Rev.*, 103, 5219–5238, 2003.
- Brown, S. S., Stark, H., Ciciora, S. J., McLaughlin, R. J., and Ravishankara, A. R.: Simultaneous in situ detection of atmospheric NO_3 and N_2O_5 via cavity ring-down spectroscopy, *Rev. Sci. Inst.*, 73, 3291–3301, 2002.
- Brown, S. S., Stark, H., Ciciora, S. J., and Ravishankara, A. R.: In-situ measurement of atmospheric NO_3 and N_2O_5 via cavity ring-down spectroscopy, *Geophys. Res. Lett.*, 28, 3227–3230, 2001.
- Brown, S. S., Stark, H., and Ravishankara, A. R.: Applicability of the steady state approximation to the interpretation of atmospheric observations of NO_3 and N_2O_5 , *J. Geophys. Res.*, 108, 4539, doi:10.1029/2003JD003407, 2003.
- Curtis, A. R. and Sweetenham, W. P.: Facsimile, AERE, Report R-12805, 1987.
- Dubé, W. P., Brown, S. S., Osthoff, H. D., Nunley, M., Ciciora, S. J., Paris, M. W., McLaughlin, R. J., and Ravishankara, A. R.: Aircraft instrument for simultaneous, in situ measurement of NO_3 and N_2O_5 via pulsed cavity ring-down spectroscopy, *Rev. Sci. Inst.*, 77, doi:

AMTD

1, 67–102, 2008

CRD/CEAS
measurement of NO_3
and N_2O_5

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



10.1063/1.2176058, 2006.

Fuchs, H., Dubé, W. P., Ciciora, S. J., and Brown, S. S.: Determination of inlet transmission and conversion efficiencies for in situ measurements of the nocturnal nitrogen oxides, NO₃, N₂O₅ and NO₂ via pulsed cavity ring-down spectroscopy, *Anal. Chem.*, 80, 6010–6017, 2008.

Geyer, A., Aliche, B., Mihelcic, D., Stutz, J., and Platt, U.: Comparison of tropospheric NO₃ radical measurements by differential optical absorption spectroscopy and matrix isolation electron spin resonance, *J. Geophys. Res.*, 104, 26 097–26 105, 1999.

Kasyutich, V. L., Canosa-Mas, C. E., Pfrang, C., Vaughan, S., and Wayne, R. P.: Off-axis continuous-wave cavity-enhanced absorption spectroscopy of narrow-band and broadband absorbers using red diode lasers, *Appl. Phys. B.*, 75, 755–761, 2002.

King, M. D., Dick, E. M., and Simpson, W. R.: A new method for the atmospheric detection of the nitrate radical (NO₃), *Atmos. Environ.*, 34, 685–688, 2000.

Matsumoto, J., Imai, H., Kosugi, N., and Kajii, Y.: In situ measurement of N₂O₅ in the urban atmosphere by thermal decomposition/laser-induced fluorescence technique, *Atmos. Environ.*, 39, 6802–6811, 2005.

Mazurenka, M., Orr-Ewing, A. J., Peverall, R., and Ritchie, G. A. D.: Cavity ring-down and cavity enhanced spectroscopy using diode lasers, *Annu. Rep. Prog. Chem., Sect C.*, 101, 100–142, 2005.

Orphal, J., Fellows, C. E., and Flaud, P. M.: The visible absorption spectrum of NO₃ measured by high-resolution Fourier transform spectroscopy, *J. Geophys. Res.*, 108, 4077, doi:10.1029/2002JD002489, 2003.

Paul, J. B., Lapson, L., and Anderson, J. G.: Ultrasensitive absorption spectroscopy with a high-finesse optical cavity and off-axis alignment, *Appl. Optics*, 40, 4904–4910, 2001.

Platt, U., Perner, D., Schröder, J., Kessler, C., and Toennissen, A.: The diurnal variation of NO₃, *J. Geophys. Res.*, 86, 11 965–11 970, 1981.

Rohrer, F., Bohn, B., Brüning, D., Johnen, F.-J., Wahner, A., and Kleffmann, J.: Characterisation of the photolytic HONO-source in the atmosphere simulation chamber SAPHIR, *Atmos. Chem. Phys.*, 5, 2189–2201, 2005, <http://www.atmos-chem-phys.net/5/2189/2005/>.

Sander, S. P., Friedl, R. R., Golden, D. M., Kurylo, M. J., Huie, R. E., Orkin, V. L., Moortgat, G. K., Ravishankara, A. R., Kolb, C. E., Molina, M. J., and Finlayson-Pitts, B. J.: Chemical kinetics and photochemical data for use in atmospheric studies: Evaluation Number 15, Jet Propulsion Laboratory, National Aeronautics and Space Administration/Jet Propulsion Laboratory/California Institute of Technology, Pasadena, CA, 2006.

AMTD

1, 67–102, 2008

**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Simpson, W. R.: Continuous wave cavity ring-down spectroscopy applied to in situ detection of dinitrogen pentoxide (N_2O_5), *Rev. Sci. Inst.*, 74, 3442–3452, 2003.

Slusher, D. L., Huey, L. G., Tanner, D. J., Flocke, F. M., and Roberts, J. M.: A thermal dissociation-chemical ionization mass spectrometry (TD-CIMS) technique for the simultaneous measurement of peroxyacyl nitrates and dinitrogen pentoxide, *J. Geophys. Res.*, 109, D19315, doi:10.1029/2004JD004670, 2004.

Wayne, R. P., Barnes, I., Biggs, P., Burrows, J. P., Canosa-Mas, C. E., Hjorth, J., Le Bras, G., Moortgat, G. K., Perner, D., Poulet, G., Restelli, G., and Sidebottom, H.: The nitrate radical: Physics, chemistry, and the atmosphere, *Atmos. Environ.*, 25A, 1–206, 1991.

Wood, E. C., Bertram, T. H., Wooldridge, P. J., and Cohen, R. C.: Measurements of N_2O_5 , NO_2 and O_3 east of the San Francisco Bay, *Atmos. Chem. Phys.*, 5, 483–491, 2005, <http://www.atmos-chem-phys.net/5/483/2005/>.

Yokelson, R. J., Burkholder, J. B., Fox, R. W., Talukdar, R. K., and Ravishankara, A. R.: Temperature-dependence of the NO_3 absorption-Spectrum, *J. Phys. Chem.*, 98, 13 144–13 150, 1994.

AMTD

1, 67–102, 2008

**CRD/CEAS
measurement of NO_3
and N_2O_5**

G. Schuster et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.

Table 1. Correction for NO₃ transmission through the inlet and cavity.

Day (in June)	correction factor	[N ₂ O ₅] ^a	[NO ₃] ^a
14th	1.11±0.05	120	60
15th	1.14±0.05	110	60
20th	1.002±0.05	3550	110
21st	1.14±0.05	47	61

Notes:

^a Mixing ratio (ppt) in the SAPHIR chamber.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Table 2. Simulation of correction for NO₃ loss.

Case	[N ₂ O ₅] ^a	[NO ₃] ^a	simple loss ^b @ 0.082 s ⁻¹	numerical (true) loss	δf (percent)
High N ₂ O ₅ High NO ₃ 295 K	5800	650	650→575 (f=1.13) (f=1.11)	650→585	2
High N ₂ O ₅ low NO ₃ 285 K	5200	132	132→117 (f=1.13) (f=1.10)	132→120	3
low N ₂ O ₅ medium NO ₃ 295 K	50	145	145→128 (f=1.13) (f=1.13)	145→128	0

Notes:

^a Mixing ratio (ppt) in the SAPHIR chamber.^b After 1.5 s using a correction factor of 1.13 assumes exponential decay with [NO₃]_t=[NO₃]₀ exp (-k_{obs}t). f=correction factor at t=1.5 s.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.

Table 3. Corrections to CRD data and assessment of uncertainty.

Correction/errors for:	NO ₃	NO ₃ + N ₂ O ₅
Filter loss ^a	1.18±0.1 (8.5%)	1.00±0.04 (4%)
Inlet/cavity loss ^b	1.13±0.1 (8.9%)	1.1±0.1 (9.1%)
Cavity length	1.05±0.03 (2.9%)	1.05±0.03 (2.9%)
Cross section of NO ₃	5% ^c	8% ^d
Reaction of NO with O ₃	<1 ppt	<1 ppt
total	14%	13%

Notes:

^a Only applicable to a fresh filter. The transmission of a contaminated filter was observed to decrease to as low as ≈50%.

^b At a flow rate of 8 SLM and cavity residence time of about 1.5 s,

^c at 295 K,

^d at 370 K.

[Title Page](#)
[Abstract](#)
[Introduction](#)
[Conclusions](#)
[References](#)
[Tables](#)
[Figures](#)
[Back](#)
[Close](#)
[Full Screen / Esc](#)
[Printer-friendly Version](#)
[Interactive Discussion](#)


CRD/CEAS measurement of NO_3 and N_2O_5

G. Schuster et al.




Fig. 1. Schematic/technical diagram of the instrument. F1–4: flow controllers (10, 10, 1000, and 10 000 sccm, respectively), M: 662 nm coated dielectric cavity mirrors, LD: laser diode, IF: interference filter centred at 662 nm, Syn.air: bottled synthetic air for mirror purging, PMT: photomultiplier, Osci: digital oscilloscope for signal acquisition, PA: signal pre-amplifier, NO/air: bottled mixture of 100 ppm NO in air for NO_3 titration. The lower section of the figure gives a detailed view of the purge gas in- and outlets and the mirror mounts as enclosed in the dotted circle of the upper section.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO_3
and N_2O_5**

G. Schuster et al.




Fig. 2. Laser emission spectra when used in CEAS (black line) or CRD mode (blue line). The spectra are normalised to have an integrated intensity of unity and are compared with a small section of the NO_3 absorption spectrum (red curve, right y-axis) at 298 K (Yokelson et al., 1994).

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**CRD/CEAS
measurement of NO_3
and N_2O_5**

G. Schuster et al.




Fig. 3. Upper: NO_3 spectrum compared to other “red” absorbers. The absorption coefficients of NO_2 , O_3 and H_2O have been scaled upwards as shown. Lower: Result of folding the absorption spectra of NO_3 , NO_2 , O_3 and H_2O with representative ambient concentrations.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.




Fig. 4. Upper panel: Raw data obtained in CEAS mode. NO₃ was detected following mixing of NO₂ (≈ 3 ppb) and O₃ (≈ 100 ppb) in the SAPHIR chamber at 10:30 UT. The difference in signals with and without NO titration correspond to concentrations of NO₃ of between ≈ 10 ppt (at 10:40) and 21 ppt of NO₃ (at 10:49) as indicated. The total flow was 3 SLM. Lower Panel: As above, data obtained after removal of filter from inlet. The NO₃ concentration of ≈ 50 ppt is barely detectable above a particle induced “noise” of ≈ 100 ppt NO₃ equivalents. Note that these concentrations have not been corrected for e.g. inlet/filter loss.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.




Fig. 5. Upper panel: Raw data obtained in CRD mode. 64 ring-downs were averaged to yield a decay constant of $13\,732 \pm 5$ (2σ). The ring-down signal extends to time=0, but is obscured by the width of the regression line. Lower Panel: NO₃ Data obtained whilst sampling ≈ 8 SLM from the SAPHIR chamber. Values of δk_{rd} obtained with and without NO (see Eq. 2) indicate NO₃ is present in the cavity at levels of ≈ 110 ppt. The ring down constant of $\approx 10\,200$ s⁻¹ during the titration phase indicates optical path length of ≈ 26 km.

[Title Page](#)
[Abstract](#)
[Introduction](#)
[Conclusions](#)
[References](#)
[Tables](#)
[Figures](#)
[◀](#)
[▶](#)
[◀](#)
[▶](#)
[Back](#)
[Close](#)
[Full Screen / Esc](#)
[Printer-friendly Version](#)
[Interactive Discussion](#)


**CRD/CEAS
measurement of NO_3
and N_2O_5** G. Schuster et al.




Fig. 6. Determination of NO_3 loss in the inlet lines (PFA-Teflon) and cavity (Teflon FEP 121a coated glass). The two datasets displayed were obtained on separate days (open circles on the 21st solid circles on the 15th) and using two different filters (see text for details). Within experimental uncertainty the decay coefficients were the same.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.




Fig. 7. Simulated losses of NO₃ in inlet and cavity. Chemistry initiated with NO₂ and O₃. The gas mixture enters the inlet after 7000 s in the chamber. The data points are taken from the analytical expression $[\text{NO}_3]_t = [\text{NO}_3]_0 \exp(-k_{\text{obs}}t)$. The solid line is a simulation of the NO₃ loss using the same value of k_{obs} at different N₂O₅ mixing ratios.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

**CRD/CEAS
measurement of NO_3
and N_2O_5** G. Schuster et al.




Fig. 8. N_2O_5 (upper panel) and NO_3 (lower panel) filter transmission tests: Black data points: filter in place, red data points: no filter. The red and black solid lines through the NO_3 data are polynomial fits which guide the eye through the slow changes in the initial NO_3 concentration. For NO_3 , filters with both $1\ \mu\text{m}$ and $2\ \mu\text{m}$ pores were used.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

**CRD/CEAS
measurement of NO₃
and N₂O₅**

G. Schuster et al.




Fig. 9. Comparison of NO₃ measurements at the SAPHIR chamber using CRD analysis (red data points) and CEAS analysis (black data points). Error bars are 1 σ statistical, based on uncertainty in measurements of I , I_0 and Δk_{rd} (see Eqs. 2 and 3).

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

CRD/CEAS
measurement of NO_3
and N_2O_5

G. Schuster et al.




Fig. 10. Upper panel: O_3 , NO and NO_2 mixing ratios reported by the local air monitoring site (≈ 5 km distance from the MPI building). Lower panel: Data points (left y-axis): N_2O_5 (+ NO_3) mixing ratios outside the MPI-building on the night of 5th–6th of October 2007. The solid lines (right y-axis) are global radiation taken from the measurement station of the local (in Hessen) environmental and geological agency.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)