Atmos. Meas. Tech. Discuss., 2, 1725–1770, 2009 www.atmos-meas-tech-discuss.net/2/1725/2009/ © Author(s) 2009. This work is distributed under the Creative Commons Attribution 3.0 License.



Atmospheric Measurement Techniques Discussions is the access reviewed discussion forum of Atmospheric Measurement Techniques

Minimizing light absorption measurement artifacts of the Aethalometer: evaluation of five correction algorithms

M. Collaud Coen¹, E. Weingartner², A. Apituley³, D. Ceburnis⁴, H. Flentje⁵, J. S. Henzing⁶, S. G. Jennings⁴, M. Moerman³, A. Petzold⁷, R. Schmidhauser², O. Schmid^{8,9,10}, and U. Baltensperger²

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Abstract Introduction

Conclusions References

Tables Figures







Full Screen / Esc

Printer-friendly Version



¹MeteoSwiss, Aerological Station, Les Invuardes, 1530 Payerne, Switzerland

²Laboratory of Atmospheric Chemistry, Paul Scherrer Institut, 5232 Villigen PSI, Switzerland

³National Institute for Public Health and the Environment, Bilthoven, The Netherlands

⁴School of Physics/Environmental Change Institute, National University of Ireland, Galway, Ireland

⁵Deutscher Wetterdienst (DWD), Meteorologisches Observatorium (MOHP), Albin-Schwaiger-Weg 10, 82383 Hohenpeissenberg, Germany

⁶Netherlands Organisation for Applied Scientific Research, TNO, 80015 Utrecht, The Netherlands

⁷Institut für Physik der Atmosphäre, Deutsches Zentrum für Luft- und Raumfahrt, 82234 Wessling, Germany

⁸Helmholtz Zentrum München, German Research Center for Environmental Health, Institute of Lung Biology and Disease, 85758 Neuherberg/Munich, Germany

⁹Missouri University of Science and Technology, Center of Excellence for Aerospace Particulate Emissions Reduction Research, Rolla, MO 65409, USA

¹⁰formerly at: Max Planck Institute for Chemistry, Biogeochemistry Department, 55020 Mainz, Germany

Received: 1 July 2009 - Accepted: 6 July 2009 - Published: 28 July 2009

Correspondence to: M. Collaud Coen (martine.collaud@meteoswiss.ch)

Published by Copernicus Publications on behalf of the European Geosciences Union.

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien



Abstract

The aerosol light absorption coefficient is an essential parameter involved in atmospheric radiation budget calculations. The Aethalometer (AE) has the great advantage of measuring the aerosol light absorption coefficient at several wavelengths, but the derived absorption coefficients are systematically too high when compared to reference methods. Up to now, four different correction algorithms of the AE absorption coefficients have been proposed by several authors. A new correction scheme based on these previously published methods has been developed, which accounts for the optical properties of the aerosol particles embedded in the filter. All the corrections have been tested on six datasets representing different aerosol types and loadings and include multi-wavelength AE and white-light AE. All the corrections have also been evaluated through comparison with a Multi-Angle Absorption Photometer (MAAP) for four datasets lasting between 6 months and five years. The modification of the wavelength dependence by the different corrections is analyzed in detail. The performances and the limits of all AE corrections are determined and recommendations are given.

1 Introduction


The single scattering albedo ω_0 and the extinction Ångström exponent of atmospheric aerosol particles are needed in models calculating aerosol radiative forcing. These parameters can be determined from concomitant multi-wavelength measurements of aerosol scattering and absorption coefficients. Instruments and methods to measure the light absorption by atmospheric particles have been described in detail elsewhere (Bohren and Huffman, 1983; Horvath, 1993; Heintzenberg et al., 1997; Moosmüller et al., 1997; Bond and Bergstrom, 2006). Among the direct measurement methods, filter-based instruments have been widely used both at ground sites and on airborne platforms due to their ease of operation. However, most of the filter-based absorption techniques, which determine the aerosol absorption coefficient from the attenuation of

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.





Close

Introduction

References

Figures



Printer-friendly Version



light passing through an aerosol-laden filter, suffer from various systematic errors that need to be corrected (Liousse et al., 1993; Petzold et al., 1997; Bond et al., 1999): firstly, attenuation is enhanced by multiple scattering by the filter fibers which increases the optical path (multiple scattering correction); secondly, light attenuation is further enhanced due to scattering of aerosols embedded in the filter (scattering correction); and thirdly, attenuation is gradually increased by the light absorbing particles accumulating in the filter thus reducing the optical path for a loaded filter (filter-loading correction).

The most frequently used filter-based commercial instruments to measure real-time black carbon (BC) mass concentrations are the Aethalometer (AE) and the Particle Soot Absorption Photometer (PSAP). The multi-wavelength AE measures at seven wavelengths covering the ultra-violet to the near-infrared wavelength range (the AE-31 measures in the range from λ =370 to 950 nm), the multi-wavelength PSAP has only recently become commercially available and measures at three wavelengths (λ =467 to 660 nm). The absorption data from both instruments need to be corrected for the above mentioned artifacts in the filter matrix and these corrections require concomitant scattering measurements. In contrast to the AE and PSAP, the more recently developed Multi-Angle Absorption Photometer (MAAP) detects not only the transmitted, but also the backscattered light at two angles to resolve the influence of light-scattering aerosol components on the angular distribution of the backscattered radiation. The absorption coefficient at λ =630 nm is thereafter obtained from a radiative transfer scheme (Petzold and Schönlinner, 2004; Petzold et al., 2005). This technique treats the multiple scattering in the filter and the scattering effect of the particles embedded on the filter. Hence, the MAAP instrument does not use any empirically determined, aerosol-related correction factors.

Various correction schemes have been published for the PSAP (Bond et al., 1999; Virkkula et al., 2005a) taking into account the above mentioned artifacts, and particularly the scattering correction. Similarly, AE correction methods presented in the literature also take these effects into account (Weingartner et al., 2003; Arnott et al., 2005; Schmid et al., 2006; Virkkula et al., 2007).

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.


Title Page


Introduction

References

Figures

Close







Using the above mentioned correction methods, high instrument correlation but with highly variable regression slopes were found for intercomparison measurements with various absorption instruments both in the laboratory and under atmospheric conditions (Arnott et al., 2005; Saathoff et al., 2003; Schmid et al., 2006; Wallace, 2005; Rice, 2004; Petzold et al., 2005; Virkkula et al., 2005a, b; Schnaiter et al., 2005; Park et al., 2006; Slowik et al., 2007). For example, the intercomparison of continuously operated Aethalometers, MAAPs and Photoacoustic spectrometers (PAS) at the Fresno Supersite (Park et al., 2006) resulted in regression slopes between 0.2 and 2. The intercomparison also pointed out differences between winter and summer measurements indicating that the aerosol composition also plays an important role for instrument correlation.

It is widely accepted (Arnott et al., 2005; Schmid et al, 2006; Rice, 2004) that the uncorrected AE measures too high absorption coefficients. Weingartner et al. (2003), Arnott et al. (2005) and Schmid et al. (2006) published AE corrections, taking into account either results of chamber experiments involving extinction and scattering coefficient measurements or comparison with a PAS. The ability to correct the AE and/or PSAP for all of the above mentioned instrumental artifacts is important: firstly, to derive climatically important aerosol parameters more accurately from a simple instrument; secondly, to take advantage of already existing long-term data sets (such as the 13year AE dataset from the JFJ, or the 15-year AE dataset from Mace Head (MHD, see Junker et al., 2006); and thirdly, to perform multi-wavelength measurements providing spectral information on absorption and single scattering albedo, which is not yet available from more reliable instruments.

Two new AE corrections are developed in this paper, using already published AE corrections schemes. These new corrections as well as all the previously published ones were tested on six datasets from different sites, of which four also included a MAAP (Table 1). The analyzed absorption coefficients include aerosol measured in the Alps (Jungfraujoch, Hohenpeissenberg – HOP), in a flat region near populated and industrialized areas (Cabauw - CAB), at a coastal site (Mace Head), on a pasture site

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page

Introduction

References

Figures

Close



Tables

I◀

Back



Interactive Discussion

Full Screen / Esc

affected by biomass burning (Amazon Basin – AMA) and in a city (Thessaloniki – THE). Therefore these sites represent free tropospheric, continental, maritime, biomass burning and heavily polluted environments and are characterized by a annual mean single scattering albedo (ω_0) between 0.65 and 0.90 (at λ =660–840 nm). Comparisons with ₅ a MAAP have been performed on datasets lasting between six months and five years, so that the correlation between both absorption measurements can be established on real atmospheric aerosols during a time period long enough to study the performance of MAAP and AE instruments and associated corrections in long-term monitoring programs.

Experimental

Measurement sites and instrumentation

Table 1 gives the main characteristics of the used datasets. The scattering and backscattering coefficients of all stations were measured by Integrating Nephelometers and corrected for the truncation error and for the non-idealities in the angular intensity distribution of the light inside the instrument according to Anderson and Ogren (1998) and Nessler et al. (2005). All the data were aggregated to hourly means.

The multi-wavelength AE's cover the 370-950 nm wavelength range, and the whitelight AE's have a broad spectral range from 500 nm to 1100 nm with a peak sensitivity in the near IR at about 840 nm (Weingartner et al., 2003).

2.2 Aethalometer corrections

All the already published corrections are summarized in this section and the mentioned equations report only the final applied corrections. To enable a better comprehension of all corrections proposed, Table 2 lists all the used parameters, their units, a brief description and the corresponding parameters used in the previously published correc-

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page

Introduction

References

Figures

Close





Printer-friendly Version

tion schemes, and Fig. 1 schematically describes the new correction scheme. In the following, some formal definitions are first given.

The light attenuation (ATN) through the aerosol-laden section of a filter spot is defined as

$$5 \quad ATN = \ln\left(\frac{I_0}{I}\right) \tag{1}$$

where I_0 is the intensity of light passing through a pristine portion of the filter and I the intensity passing through the loaded filter. The particles embedded in the filter during a time interval Δt will increase ATN, so that the *n*-th measure of the aerosol attenuation coefficient (b_{ATN}) of the filtered aerosol particles is obtained from

$$b_{ATN,n}(\lambda) = \frac{(ATN_n(\lambda) - ATN_{n-1}(\lambda))}{\Lambda t} \frac{A}{V}$$
 (2)

where A is the area of the sample spot and V the volumetric flow rate. The corrections discussed below are then applied to infer the true aerosol absorption coefficient $b_{abs,n}$ of airborne particles from $b_{ATN,n}$.

Both the absorption Ångström exponent \mathring{a}_{abs} and the scattering Ångström exponent $\mathring{a}_{\text{scat}}$, which refer to the wavelength dependence of the respective coefficients, were determined by fitting the measured absorption or scattering coefficients with a wavelength power-law dependence $(b \sim \lambda^{-\dot{a}})$. These Ångström exponents were used to calculate coefficients at other wavelengths, such as the absorption coefficient at λ =630 nm to allow comparison with the MAAP.

2.2.1 The Weingartner correction

Weingartner et al. (2003) proposed an empirical correction R_{w} for the attenuation effect due to the filter-loading and determined the calibration constant C_{ref} for different aerosol types produced in the AIDA aerosol chamber (at FZ Karlsruhe, Germany) to correct for

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page

Abstract Conclusions I◀ Back



Introduction

References

Figures







Full Screen / Esc

Printer-friendly Version



the multiple scattering in the filter matrix. The resulting n-th absorption coefficient $b_{abs,n}$ is given by:

$$b_{\text{abs},n} = \frac{b_{\text{ATN},n}}{C_{\text{ref}} R_{w,n}} = \frac{b_{\text{ATN},n}}{C_{\text{ref}} \left[\left(\frac{1}{f} - 1 \right) \frac{\ln \text{ATN}_n - \ln(10\%)}{\ln(50\%) - \ln(10\%)} + 1 \right]}$$
(3)

where C_{ref} is determined by first correcting b_{ATN} for the filter-loading correction and then comparing it with the absorption coefficient measured simultaneously with a reference instrument ($b_{\text{abs,ref}}$). Since the Weingartner filter-loading correction $R_{\scriptscriptstyle W}$ takes ATN=10% as a reference point, C_{ref} relates in this case to ATN=10%:

$$C_{\text{ref}} = \frac{b_{\text{ATN}}}{R_W b_{\text{abs,ref}}} \tag{4}$$

A parameter $f(\lambda)$ is introduced which characterizes the slope between $b_{\text{ATN},n}$ and $\ln(\text{ATN}_n)$ and parameterizes the filter-loading correction R_w . A clear dependence of f on $1-\omega_0$ was also observed for pure, internally and externally mixed diesel soot particles, which leads to the following quasi-linear relation

$$f = m\left(1 - \omega_0\right) + 1,\tag{5}$$

where m is nearly constant (0.87 to 0.85) for λ =470 to 660 nm.

The Weingartner correction parameter $R_{\scriptscriptstyle W}$ can depend on the light wavelength, but $C_{\rm ref}$ does not. Therefore, a constant $C_{\rm ref}$ value over the wide spectral range (370–950 nm) was chosen. Weingartner et al. (2003) also determined f values for different aerosol types. In this paper, $R_{\scriptscriptstyle W}$ will always correspond to the Weingartner correction with fixed f values chosen for each dataset depending on the aerosol type, as it is presently applied by most users, so that the $R_{\scriptscriptstyle W}$ correction does not need concomitant $b_{\rm scat}$ measurements once f has been determined. It is evident from Eqs. (3) through (5) that if ω_0 =1, both f and $R_{\scriptscriptstyle W}$ are equal to unity. Since the aerosol measured at the high alpine site JFJ was aged aerosol with ω_0 values mostly close to unity, $R_{\scriptscriptstyle W}$ can be taken as unity.

AMTD

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

I₫











Full Screen / Esc

Printer-friendly Version



2.2.2 The Arnott correction

Arnott et al. (2005) proposed a theoretically well documented correction, which includes an explicit scattering correction similarly to the correction commonly applied to the PSAP (Bond et al., 1999). The form of this scattering correction ($-\alpha$ b_{scat}) was deduced from the non-zero b_{ATN} measured by an AE with a purely scattering aerosol. The scattering coefficient b_{scat} weighted by the α values was therefore subtracted from b_{ATN} to correct for the scattering artifact (Eq. 6). The Arnott filter-loading correction R_A was derived from multiple scattering theory, which shows that the exponential behavior of light absorption in the strong multiple scattering limit scales as the square root of the total absorption optical depth. The corrected $b_{\text{abs},n}$ is given by :

$$b_{\text{abs},n} = \frac{b_{\text{ATN},n} - \alpha \ b_{\text{scat},n}}{C_{\text{ref}} R_{A,n}} = \frac{b_{\text{ATN},n} - \alpha \ b_{\text{scat},n}}{C_{\text{ref}}} \sqrt{1 + \frac{\left(\frac{V \ \Delta t}{A}\right) \sum_{i=1}^{n-1} b_{\text{abs},i}}{\tau_{a,f,x}}}$$
(6)

$$\alpha(\lambda) = \beta_{\text{scat,non-abs}}^{d-1} c \lambda^{-\mathring{a}_{\text{scat,non-abs}}} (d-1)$$
 (7)

where n, V, Δt , and A were introduced in Eq. (2), $\beta_{\text{scat},\text{non-abs}}$ and $\mathring{a}_{\text{scat},\text{non-abs}}$ are obtained from measurements of ammonium sulfate aerosol and correspond to the power-law fit of the wavelength dependence of $b_{\text{scat},\text{non-abs}}$:

$$b_{\text{scat},\text{non-abs}} = \beta_{\text{scat},\text{non-abs}} \lambda^{-\hat{a}_{\text{scat},\text{non-abs}}}$$
 (8)

c and d are also obtained from the measured non-zero b_{ATN} in the presence of a non-absorbing aerosol and correspond to the relation between $b_{\text{scat non-abs}}$ and b_{ATN} :

$$b_{\mathsf{ATN}} = c \ b_{\mathsf{scat}\,\mathsf{non-abs}}^d \tag{9}$$

and $au_{a,f_X}(\lambda)$ is the filter absorption optical depth for the filter fraction x that has particles embedded in it, and C_{ref} is obtained by comparison between AE and PAS absorption coefficients. $au_{a,f_X}(\lambda)$ and $C_{\text{ref}}(\lambda)$ were determined from kerosene soot measurements during the Reno Aerosol Optics Experiment, under the condition that $b_{\text{abs}}(\lambda)$,

AMTD

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.



Back

Full Screen / Esc

Close

Printer-friendly Version

Interactive Discussion



1733

when extrapolated to λ =532 nm with \mathring{a}_{abs} =1, were all equal to PAS measurements at λ =532 nm. In this Arnott correction, all the reported calibration constants depend on the wavelength of the light. Six weeks of ambient measurements at an urban site allowed testing of the proposed correction, which showed that different values of the above parameters are needed for ambient and laboratory generated aerosol.

2.2.3 The Schmid correction

Schmid et al. (2006) proposed a correction that includes firstly the filter-loading correction R_W with constant values of f and the multiple scattering correction $C_{\rm ref}$ developed by Weingartner et al. (2003). Secondly, comparing Arnott and Weingartner methods, they derived a new scattering correction depending on ω_0 and the $\alpha(\lambda)$ constants derived by Arnott, which replaces the scattering correction introduced by Arnott. Instead of subtracting a part of $b_{\rm scat}$ in the nominator like Arnott, the Schmid scattering correction adds a term $C_{\rm scat}$ to $C_{\rm ref}$, leading to a correction that includes both the multiple scattering and the scattering corrections in the denominator:

$$b_{\text{abs},n}(\lambda) = \frac{b_{\text{ATN},n}}{\left(C_{\text{ref}} + C_{\text{scat},n}\right) R_{W,n}} = \frac{b_{\text{ATN},n}}{\left(C_{\text{ref}} + \alpha \frac{\omega_{0,n}}{1 - \omega_{0,n}}\right) \left[\left(\frac{1}{f} - 1\right) \frac{\ln \text{ATN}_{n} - \ln(10\%)}{\ln(50\%) - \ln(10\%)} + 1\right]}$$
(10)

The corrected $b_{\rm abs}$ measured either by an AE or a PSAP were compared to the $b_{\rm abs}$ measured by a PAS at λ =532 nm for some days of measurements in AMA, leading to high correlations and ratios between AE and PAS $b_{\rm abs}$ of between 0.94 and 1.03.

2.2.4 Virkkula filter-loading correction

Virkkula et al. (2007) proposed a filter-loading correction R_V very close to that of the PSAP, assuming that the three last values measured on the filter spot i and the three first values measured on the next filter spot i+1 should be equal, and that the values

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Abstract Introduction

Conclusions References

ables Figures

15 21

Close

Full Screen / Esc

Back

Printer-friendly Version



measured on lightly loaded filters are the closest to the real concentration:

$$b_{\text{ATN,R_corrected},n} = \frac{b_{\text{ATN},n}}{R_V} = (1 + k_i \text{ ATN}) b_{\text{ATN},n}$$
 (11)

$$k_{i} = \frac{BC_{0}(t_{i+1,first}) - BC_{0}(t_{i,last})}{ATN(t_{i,last}) BC_{0}(t_{i,last}) - ATN(t_{i+1,first})BC_{0}(t_{i+1,first})}$$
(12)

where $t_{i,\text{last}}$ is the time of the last measurement on the filter spot i and $t_{i+1,\text{first}}$ is the time of the first measurement on the next filter spot i+1. A k_i value is therefore determined for each filter spot and applied to all n measurements on the i-th filter spot. The correction was validated by a comparison with simultaneous aerosol volume concentration measurements at three sites with different aerosol types.

Virkkula et al. (2007) chose to set the scattering correction to zero, since a lot of AE users do not have concomitant scattering measurements. In addition, they did not introduce a multiple scattering correction. Henceforth, the Virkkula correction will be considered as a filter-loading correction only and not as a total correction of the attenuation coefficient, since the multiple scattering correction is not negligible as will be shown below.

2.2.5 The new correction

The necessity of a further development of the above described corrections became obvious when the corrections were applied to the JFJ dataset (see Sects. 3 and 4). Firstly, the Weingartner filter-loading correction R_W results in $1/R_W < 1$ for ATN < 10%. Considering that a pristine filter should produce no artifact, $1/R_W$ should be equal to 1 for ATN=0. The subtraction of ln(10%) was therefore removed for the new filter-loading correction. Weingartner et al. (2003) found a linear relationship between $b_{\rm ATN}$ and ln(ATN), which leads to a m value (Eq. 5) independent of the wavelength. Recent experiments with aged diesel soot showed a linear relationship between $b_{\rm ATN}$ and ATN (Steiger, 2008). Investigations of the JFJ and THE datasets show that the regressions between $b_{\rm ATN}$

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Abstract Introduction

Conclusions References

Tables Figures













Full Screen / Esc

Printer-friendly Version



and ATN are statistically better than between b_{ATN} and In(ATN). The direct proportionality is therefore chosen for the modified Weingartner filter-loading correction (Eq. 13). A new m value was consequently calculated similarly to that of Weingartner et al. (2003) and with the same datasets from the AIDA chamber experiment but by fitting a linear relationship between $b_{\Delta TN}$ and ATN. This m value has a mean value m=0.74 when averaged over the experiments and the wavelengths, but is wavelength dependent. It was verified for all datasets that a change of the *m* value induces a change in the calculated $C_{\rm ref}$ but insignificant changes in the final $b_{\rm abs}$ values. The new filter-loading correction is:

$$R_{\text{new},s,n} = \left(\frac{1}{m\left(1 - \bar{\omega}_{0,s,n}\right) + 1} - 1\right) \frac{\text{ATN}_n}{50\%} + 1$$
 (13)

where ATN is given in %, $\bar{\omega}_{0,s,n}$ is the mean of the *n* single scattering albedo measured since the filter spot change, where the subscript s indicates that the mean optical properties of the aerosol particles embedded in the filter spot and not only of the n-th measurement are taken into account.

Secondly, the new corrections take explicitly into account the fact that the AE measures a non-zero b_{abs} when loaded with non-absorbing aerosol. Either the Arnott or the Schmid scattering correction schemes can be applied. Both corrections use the Arnott $\alpha(\lambda)$ parameters to weight the scattering correction. Following the Arnott assumption of α being constant for a defined λ , the ratio b_{abs}/b_{scat} for non-absorbing aerosols should be constant, independently of the value of b_{scat} . The 2007 EUSAAR intercomparison campaign in Leipzig showed however that the ratio between the absorption and scattering optical depth increased with decreasing scattering optical depth for all filter-based instruments (PSAP, MAAP and AE) (Müller et al., 2008). The scattering coefficient and its wavelength dependence (Eq. 8) should therefore match the measured aerosol and not the ammonium sulfate experiment involving high ω_0 and high $\mathring{a}_{\text{scat}}$ ($\mathring{a}_{\text{scat},\text{non-abs}}$ =3). The Arnott $\beta_{\text{scat},\text{non-abs}}$ and $\mathring{a}_{\text{scat},\text{non-abs}}$ constants have therefore to correspond to β_{scat} and $\mathring{a}_{\text{scat}}$ of the aerosol particles embedded in the filter spot. Therefore, the main modification introduced by the new correction is that the constant

AMTD

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Introduction

References

Figures

Close

Abstract Conclusions

I◀

Back



Full Screen / Esc

Printer-friendly Version

 $eta_{
m scat,non-abs}$ and $\mathring{a}_{
m scat,non-abs}$ of Eq. (8) are replaced by a mean scattering coefficient $ar{eta}_{
m scat,s,n}$ and a mean scattering Ångström exponent $\ddot{a}_{
m scat,s,n}$ of the total aerosol load in the filter spot. The power-law regression between $b_{
m ATN}$ and $b_{
m scat}$ for non-absorbing aerosols (Eq. 9) can on the other hand be considered to be universal, so that the constants c and d can be taken from Arnott et al. (2005). This new $\alpha_{
m new,n}(\lambda)$ (see Eq. 15 below) allows one to take the real scattering properties of the measured aerosol into account and is therefore applied to both the Arnott and Schmid scattering correction schemes, leading to two new corrections. Similarly, instead of using $b_{
m scat,n}$ or $\omega_{0,n}$ measured simultaneously with the n-th absorption measurement $b_{
m abs,n}$ since a filter spot change, the mean scattering coefficient $\bar{b}_{
m scat,s,n}$ or the mean single scattering albedo $\bar{\omega}_{0,s,n}$ of the total aerosol loading in the filter spot are used for the Arnott scattering and for the Schmid scattering corrections, respectively.

Finally the new AE correction is described in Fig. 1 and by Eqs. (14a) or (14b) depending if the Arnott or the Schmid scattering correction is applied:

$$b_{\text{abs},n}(\lambda) = \frac{b_{\text{ATN},n} - \alpha_{\text{new},s,n} \, \bar{b}_{\text{scat},s,n}}{C_{\text{ref}} \, R_{\text{new},s,n}} \quad \text{similar to Arnott}$$
(14a)

$$b_{\text{abs},n}(\lambda) = \frac{b_{\text{ATN},n}}{\left(C_{\text{ref}} + C_{\text{scat},s,n}\right) R_{\text{new},s,n}}$$

$$= \frac{b_{\text{ATN},n}}{\left(C_{\text{ref}} + \alpha_{\text{new},s,n} \frac{\bar{\omega}_{0,s,n}}{1 - \bar{\omega}_{0,s,n}}\right) R_{\text{new},s,n}}$$
 similar to Schmid (14b)

with the new $\alpha_{\text{new},s,n}(\lambda)$ given by

The C_{ref} is determined by comparing the b_{ATN} already corrected for the filter-loading with the new R_{new} correction to $b_{\text{abs,ref}}$ measured by a reference instrument, as described in Eq. (4).

AMTD

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.



Abstract

Introduction

Conclusions

References

Tables

Figures













Full Screen / Esc

Printer-friendly Version



The wavelength dependent $\omega_0(\lambda)$ (Eq. 16) is obtained from $\omega_{0,\text{ref}}$ that can be calculated with the scattering coefficient taken at one of the AE wavelengths λ_{ref} and the first estimation of $b_{\text{abs}} = b_{\text{ATN}}/\text{SG}$ (Schmid et al., 2006), where SG is the mass specific attenuation cross-section proposed by the manufacturer to be $14.625/\lambda$ [m² g⁻¹], λ in [nm]. It was verified with the JFJ and THE datasets that the choice of the initial wavelength λ_{ref} is not important for the final result.

$$\omega_{0}(\lambda) = \frac{\omega_{0,\text{ref}} \left(\frac{\lambda}{\lambda_{\text{ref}}}\right)^{-\hat{a}_{\text{scat}}}}{\omega_{0,\text{ref}} \left(\frac{\lambda}{\lambda_{\text{ref}}}\right)^{-\hat{a}_{\text{scat}}}} + (1 - \omega_{0,\text{ref}}) \left(\frac{\lambda}{\lambda_{\text{ref}}}\right)^{-\hat{a}_{\text{abs}}}$$
(16)

3 Results

3.1 Effects of each partial correction corresponding to a systematic artifact

3.1.1 The filter-loading correction

Figure 2 shows 1/R calculated by the methods by Weingartner et al. (2003), Arnott et al. (2005), and Virkkula et al. (2007) as well as the new one described in Sect. 2.2.5 for one month of measurements at the JFJ, and Table 3 summarizes the results of the four R corrections applied to all the datasets. The constant f of the Weingartner $1/R_w$ correction was estimated from Weingartner et al. (2003), with f=1.025 corresponding to aged mixed aerosols (JFJ, HOP) and f=1.2 for aerosols near pollution sources (THE), while intermediate values of f=1.05 and f=1.10 were taken for MHD and CAB, respectively. The form of the Weingartner filter-loading correction results in $1/R_w$ < 1 for ATN<10%, which leads to minimal $1/R_w$ values of 0.77<1/ R_w <0.93 depending on the station and consequently to an increase of h_{abs} measured on lightly loaded filters. For all the analyzed datasets, the maximum of $1/R_w$ amounts to 1.06, whereas the mean of $1/R_w$ varies between 1.01 and 1.03. The Arnott filter-loading correction leads to $1/R_A$

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.











Full Screen / Esc

Printer-friendly Version



values between 1 and 1.81 with mean values between 1.00 and 1.08. The new correction has a minimum $1/R_{\rm new}$ value of 1 similar to the Arnott correction and a maximum value of 2.56. The mean $1/R_{\rm new}$ values vary between 1.05 and 1.37. The Weingartner, Arnott and the new filter-loading corrections have a clear wavelength dependence with a greater 1/R at lower wavelengths, following the wavelength dependence of ATN and $b_{\rm abs}$. The Virkkula filter-loading correction is highly non-stable, leading to large negative and positive $1/R_V$ outliers. Moreover, its wavelength dependence varies. The difficulty of applying the Virkkula correction is due to the natural high variability of $b_{\rm ATN}$ as a function of time, which is for most of the time greater than the $b_{\rm ATN}$ decrease induced by filter changes.

The filter-loading artifact can be clearly visualized after a filter change in a distinct step in $b_{\rm ATN}$ during chamber studies, when the aerosol type and concentration remains fairly constant. In an ambient environment however, steps in $b_{\rm ATN}$ are not only due to the filter-loading artifact but also due to the natural variability of the aerosol properties. The efficacy of the four filter-loading corrections to smooth steps in $b_{\rm ATN}$ due to the filter-loading artifact was tested on the THE and JFJ datasets, leading to no significant differences between the filter-loading corrections.

Since the filter-loading correction minimizes the attenuation effect due to the filter-loading, the ratio $b_{\text{ATN},R_\text{corrected}}/b_{\text{abs,ref_MAAP}}$ between the filter-loading corrected absorption and a reference absorption measurement should no longer depend on ATN. Taking the MAAP as a reference absorption measurement instrument, these $b_{\text{ATN},R_\text{corrected}}/b_{\text{abs,ref_MAAP}}$ ratios for all filter-loading corrections are plotted as a function of ATN in Fig. 3 for JFJ, CAB, MHD and HOP datasets, and the mean and the standard deviation of C_{ref} are given in Table 4. At JFJ and HOP, the new correction leads to the flattest curve between the ratio $b_{\text{ATN},R_\text{corrected}}/b_{\text{abs,ref_MAAP}}$ versus ATN, leading to the lowest standard deviations in Table 4. At CAB, the Arnott correction leads to the flattest curve, the new correction having however the smallest variability with ATN for the smallest ATN. At MHD, the Weingartner and Arnott corrections are the flattest ones, the new correction becoming too large at high ATN.

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.





Printer-friendly Version

The multiple scattering correction

For the evaluation of $C_{\rm ref}$ of the multi-wavelength AE, the wavelength of λ =660 nm, which is nearest to the MAAP wavelength (λ =630 nm), was chosen. The comparison of the AE and MAAP for the four datasets leads to average C_{ref} values between 2.9 and 5 4.3 (Table 4) with the new filter-loading correction algorithm, while the results found with the Arnott and Weingartner filter-loading corrections are lower. Due to the particularly strong seasonal cycle of aerosol optical properties at the JFJ, an estimation of C_{ref} for each month was also performed. $C_{\rm ref}$ is lower from April to September with a mean value of 2.83, and higher during the October-March period with a mean of 3.24. The CAB dataset is too short to analyze a complete seasonal cycle. However, C_{ref} is clearly higher during the May-July period than for the rest of the dataset, either because of seasonal variation of the aerosol composition or due to a modification of the inlet during March 2008 (see Table 1). HOP and MHD have less distinct seasonal cycles. Figure 4 shows C_{ref} as a function of ω_0 for the four complete datasets (squares) as well as for the JFJ and CAB datasets divided into 2 seasons. A clear correlation between $C_{\rm ref}$ and ω_0 exists, C_{ref} becoming greater for lower ω_0 . Figure 4 shows C_{ref} for datasets with ω_0 <0.7. Further analysis with other ambient aerosol types, particularly in very polluted environments with lower $\omega_0(<0.7)$ such as big cities, should enable an extension of the results to a more universal $C_{\text{ref}}(\omega_0)$ curve. It is very important to establish such correlation, since the multiple scattering correction is the most important one as will be discussed later.

According to the correction schemes, the multiple scattering artifact should depend only on the filter properties and not on the embedded aerosol. However, the calculated $C_{\rm ref}$ is not always constant, even at the same station. This variation of $C_{\rm ref}$ could be caused by semi-volatile organic compounds (VOCs) as well as water vapor condensing on the filter fibers (Weingartner et al., 2003). Firstly, the scattering of the filter fibers can be enhanced by these compounds, leading to a longer optical path length; secondly, the scattering phase function can also be modified leading to a modification of the

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page











Full Screen / Esc

Printer-friendly Version



mean filter reflectance; thirdly, the sticking coefficient (probability to stick on a surface) of the aerosol on the fiber and the possible change of the inter-fiber spacing can change the depth of aerosol deposition in the filter leading to a change in its optical properties. The experimental setup used in this study does not allow us to further investigate these potential influences.

3.1.3 The scattering correction

The loading of the filter with scattering aerosol leads to two different artifacts: the aerosol particles scatter light in all directions, leading firstly to an increase of backscattered light and consequently to an apparent greater reflectance of the filter, and secondly to an increased light optical path and consequently to a higher probability of encountering an embedded absorbing particle. Due to its form, the applied scattering correction (subtraction of an amount proportional to the scattering coefficient) clearly corrects for an increase in the reflectance. The reflectance depends on the aerosol asymmetry parameter, which is lower for longer wavelengths, so that the proportionality factor α should increase with increasing wavelength. The second scattering artifact, as described above, leads to a higher probability of encountering an embedded particle, which is presently not compensated by any correction. The MAAP measurement technique does not correct for this second artifact either, but shows close agreement with a reference absorption measurement through the difference between light extinction and scattering (Petzold et al., 2005). This second scattering effect is therefore probably not very pronounced.

Contrary to the Arnott scattering correction, but according to recent results (Müller et al., 2008) (see Sect. 2.2.5), the new correction introduces an $\alpha_{\text{new}}(\lambda)$ that depends on $\bar{\beta}_{\rm scat}$ and $a_{\rm scat}$ (Eq. 15). Figure 5a shows $\alpha_{\rm new}$ at λ =660 nm as a function of $b_{\rm scat}$ for the JFJ, CAB, AMA and THE datasets. $\alpha_{\text{new}}(\lambda)$ is lower than about 0.08 for the CAB, AMA and THE datasets. When applied to the JFJ dataset, $\alpha_{\text{new}}(\lambda)$ ranges between 0 and 0.20, but the great majority of $\alpha_{\text{new}}(\lambda)$ values are smaller than 0.1. As expected from the wavelength dependence of the asymmetry parameter, α_{new} is greater

AMTD

2, 1725-1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page **Abstract** Introduction Conclusions References **Tables Figures** I◀

Back

Full Screen / Esc

Close

Printer-friendly Version



at longer wavelengths. Since $\mathring{a}_{\text{scat}}$ at AMA and THE are always near 2 (standard deviation < 0.15 for the two stations), α_{new} clearly increases with decreasing b_{scat} . $\mathring{a}_{\text{scat}}$ has a lower mean for the JFJ and CAB datasets (1.64±1.25 and 1.56±0.72, respectively), and is much more variable, particularly at the JFJ, due firstly to the longer dataset and secondly to the presence of aged accumulation mode aerosol (Weingartner et al., 1999), which are coarser and lead to low $\mathring{a}_{\text{scat}}$ and sometimes even to negative $\mathring{a}_{\text{scat}}$ in presence of mineral particles. α_{new} is therefore not always increasing with b_{scat} , but tends towards zero for very low $\mathring{a}_{\text{scat}}$.

Figure 5b shows the $C_{\rm scat}$ dependence on ω_0 for the JFJ, CAB, AMA and THE datasets. As can be deduced from Eq. (14b), $C_{\rm scat}$ increases with increasing ω_0 . While the $C_{\rm scat}$ dependence on ω_0 is well defined for the high aerosol loading at Thessaloniki, it becomes less sharp at lower aerosol concentrations, particularly at the JFJ, due firstly to a greater uncertainty in the measurement of very low $b_{\rm abs}$ and $b_{\rm scat}$ and secondly to a broader range of $\alpha_{\rm new}$ values (see Fig. 5a). $C_{\rm scat}$ maximum values can be equal to 2 for ω_0 near one at the JFJ, but remains normally below 0.5 for most values of ω_0 <0.95. The $C_{\rm scat}$ correction is usually greater at longer wavelengths, but inversion of this wavelength dependence is also observed.

3.2 Applicability of the AE corrections

The very low aerosol concentrations, particularly at the JFJ and MHD, induce some instrumental difficulties. Due to the fact that the measurements result from the difference between the last and the last before the last measured light attenuation through a loaded filter, very low concentrations sometimes induce negative $b_{\rm abs}$ due to electronic noise in the raw signals (Petzold and Schönlinner, 2004). Even if negative $b_{\rm abs}$ are not real, these values have to be kept, because they are necessary to avoid a bias in the calculated averages. The amount of negative $b_{\rm abs}$ can be lowered by use of larger flow rates or by longer integrating times. Long-term monitoring sites may however have constraints in not allowing variable integrating times. In addition, long integrating times may induce a loss of information such as, for example, diurnal cycles. Correction

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.



Full Screen / Esc

Close

Back

Printer-friendly Version



algorithms that can be applied to negative values are therefore necessary.

Since the Weingartner correction consists of a simple multiplication with constant factors, it can be applied to the negative values of $b_{\rm abs}$, it does not create new negative values as other algorithms do (see below), and has no impact on $\mathring{a}_{\rm abs}$ (the difference between the measured and the Weingartner corrected $\mathring{a}_{\rm abs}$ is less than 0.01%), as long as f is taken as a constant for a given dataset.

The Arnott correction is able to treat negative values of $b_{\rm abs}$. Due to the subtraction of the scattering correction, the Arnott correction often creates new negative values of $b_{\rm abs}$. This artifact is found to be small for the shorter wavelengths (2.9% and 0.3% at λ =370 nm), but the created negative values reach up to 9.6% and 3.4% at λ =950 nm at JFJ and CAB, respectively. The given constants $C_{\rm ref}(\lambda)$ also yield too high values of $b_{\rm abs}$. Finally, the Arnott correction has a broad impact on $\mathring{a}_{\rm abs}$. As can be seen in Fig. 6, the Arnott correction produces values of $\mathring{a}_{\rm abs}$ that on average are 25% higher than the initially measured ones for both datasets.

The Schmid correction depends on ω_0 ; negative values of $b_{\rm abs}$ lead to ω_0 values greater than 1 and therefore to negative $C_{\rm scat}$ values for the scattering correction. The use of $\mathring{a}_{\rm abs}$ to calculate ω_0 prevents its application to negative $b_{\rm abs}$. This produces new missing values of $b_{\rm abs}$ (2.3% and 0.9% at JFJ and CAB, respectively) and a few new negative values of $b_{\rm abs}$ (<0.1% for both datasets). The absorption wavelength dependence is also clearly modified, leading to $\mathring{a}_{\rm abs}$ on average 10–15% higher than the initially measured ones (Fig. 6) for both datasets.

The new AE corrections cannot be used for negative $b_{\rm abs}$, similarly to the Schmid correction. They therefore generate missing values unless $\mathring{a}_{\rm abs}$ values are averaged in the case of negative $b_{\rm abs}$. If the new correction similar to the Arnott method (Eq. 14a) is used, new negative $b_{\rm abs}$ values are generated (2.1% and <0.1% at λ =370 nm and 2.8% and <0.1% at λ =950 nm at JFJ and CAB, respectively) due to the scattering correction subtraction, which is rather less than use of the Arnott correction at high wavelength values. It also introduces more outliers than the new method, similar to the Schmid method. Finally this new correction mainly preserves $\mathring{a}_{\rm abs}$ (Fig. 6) with 58% and 46%

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien


M. Collaud Coen et al.


Title Page

Introduction

References

Figures







of the exponents remaining constant, with 18% and 42% of the exponents having a difference of only 5%, and 7% and 1% of the exponents having a 10% difference from the initially measured exponent at JFJ and CAB, respectively.

3.3 Correlation of b_{abs} values derived from AE and MAAP

To evaluate the AE corrections, the slopes of the linear regression between the corrected and the MAAP b_{abs} values are reported in Fig. 7 for each month of simultaneous AE and MAAP measurements at the four stations. The averages of the monthly slopes with their standard deviations are given in Table 5. As already mentioned the Arnott correction produces a lot of outliers and has consequently very large mean slopes and is therefore not reported in Fig. 7. As can be seen in Table 5, the new correction like the Arnott and like the Schmid methods both lead to very similar results for all the stations but for JFJ, so that only the new correction like the Schmid method is reported in Fig. 7. Finally, the errors on the slope determination are always far smaller than the slope fluctuation between months, so that they have not been reported for clarity purposes.

At the JFJ, the Weingartner correction leads to higher slopes and the Schmid correction to lower slopes than the new correction. At CAB and MHD, similar slopes are found for the Weingartner, the Schmid and the new corrections. At HOP the Weingartner correction leads to the slope nearest to one, followed by the new corrections and finally the Schmid correction. At HOP and JFJ, the Weingartner, Schmid and the new correction similar to Schmid provide similar standard deviations. At MHD, the Weingartner correction leads to the lowest standard deviation of the slopes, and the new correction to greatest ones. At CAB, the Schmid correction and the new correction similar to Schmid have the lowest standard deviations. The new correction similar to ²⁵ Arnott leads sometimes to higher standard deviations due to the presence of outliers.

None of the tested AE corrections are able to get rid of some time dependence of the monthly slopes such as the increase from October 2006 to February 2007 at the JFJ or the decrease in April 2008 at CAB and in October 2005 at HOP. Monthly slopes

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page **Abstract** Introduction Conclusions References **Tables** I◀

Back

Full Screen / Esc

Figures

Close

Printer-friendly Version



between the MAAP $b_{\rm abs}$ output and the $b_{\rm abs}$ measured by the MAAP in transmission only (thus using the MAAP in a similar way as the AE, i.e., without the backscattering measurements and the radiative transfer scheme, and then treated similarly to AE measurements), are also plotted in Fig. 7 (black dashed line) for the JFJ dataset. Some features such as the slope increase at the end of 2006 followed by a sharp decrease at the beginning of 2007 are also present. The variability of the monthly slopes as a function of time is therefore not only due to AE instrumental non-idealities, but also due to variations in aerosol optical properties. Part of this variability might also be attributed to the condensation artifact described above. It was also checked that monthly slopes are not correlated with the asymmetry parameter or with ω_0 at the JFJ.

To evaluate the influence of the filter-loading, the scattering and the multiple scattering corrections individually, the new correction was applied to the four datasets while removing successively the different components of the correction. It was found that less than three percents of the total correction is due to the filter-loading correction R, which was expected due to the very low mean values of 1/R (Table 3). The greatest part of the correction is due to the multiple scattering correction, so that the $C_{\rm ref}$ determination is the most important one to ensure a good correlation with a reference method. The difference between the Weingartner, the Schmid and the new corrections can be mainly attributed to the scattering correction. For all the four stations, the scattering correction does not have a great impact on the slope between AE and MAAP $b_{\rm abs}$.

4 Discussion

4.1 Criteria for a good correction

Before recommendations are formulated on the best way to correct the AE data, criteria for a good correction procedure should be stated. Firstly, the correction has to take into account all known artifacts occurring during measurements. Secondly, the correction

AMTD

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.



Full Screen / Esc

Printer-friendly Version



should be applicable to all kind of datasets. Thirdly, the correction should introduce few outliers, or new missing or negative values. Fourthly, the correction should lead to the best b_{ahs} correlation with a "true" reference method. Fifthly, the correction should also lead to a real wavelength dependence of the b_{abs} that is real \mathring{a}_{abs} . Finally, the ease of use of the correction may also be a consideration for choosing a correction method.

Modification of the absorption Angström exponent by AE corrections


The lack of a reference measurement for \mathring{a}_{abs} requires us to develop some considerations on the effects of each part of the correction on the absorption wavelength dependence. Firstly, the light absorption by aerosol loading in a filter decreases the optical path length. Since aerosol light absorption increases with decreasing wavelength, the filter-loading artifact is expected to be larger at shorter wavelengths and will decrease the absorption wavelength dependence. Secondly, the aerosol light scattering increases but the backscattered fraction decreases with decreasing wavelength. This results in a greater scattering coefficient but a lower reflectance at shorter wavelengths. It is therefore difficult to determine if the scattering correction for increased reflectance should modify the absorption wavelength dependence or not, and, in case of modification, if the \mathring{a}_{abs} will be increased or decreased. Thirdly, due to increasing scattering from the particles with decreasing wavelength, the increase of the absorption due to increased probability of encountering an embedded aerosol particle should lead to an increased absorption wavelength dependence. Fourthly, since the filter fibers are nonabsorbing and large (typically 1 micrometer for the guartz fiber and 10 micrometer for the cellulose fiber; Arnott et al., 2005), geometric scattering occurs from the cylindrical fibers, so that the multiple scattering correction can be expected in a first approximation to be wavelength independent and to induce no change in the absorption wavelength dependence. Weingartner et al. (2003) concluded similarly. Considering the effects of all parts of the correction on the absorption wavelength dependence, it is not possible to make precise conclusions on the expected changes of å_{abs} induced by the AE correction.

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.





Introduction

References

Figures









Full Screen / Esc

Printer-friendly Version



Weingartner et al. (2003) show a high correlation between AE \mathring{a}_{abs} and a reference method (measurement of the difference between extinction and scattering) for "pure" Diesel and Palas soot particles as well as soot particles externally mixed with secondary organic aerosol or ammonium sulfate; a lower correlation was found for coated 5 soot particles. Virkkula et al. (2005a) showed a high agreement between å_{ahs} from a 3λ -PSAP calculated with the Virkkula method and a reference value for \mathring{a}_{abs} between 1.0 and 1.3. However Virkkula et al. (2005b) also concluded that the correction algorithm could still be improved regarding the wavelength dependence. A. Petzold (personal communication, 2008) showed that the 3λ-PSAP Virkkula correction (Virkkula et al., 2005b) leads to minor modification of \mathring{a}_{abs} for values around 1 but to up to 2.5 times larger \mathring{a}_{abs} for values around 4 corresponding to desert dust. The wavelength dependence of the PSAP filter transmission is far greater than that of the AE filter, which has a low wavelength dependence (see Fig. 6 in Arnott et al., 2005). Since the reflectance of an AE filter is greater than of a PSAP filter (Arnott et al., 2005), its dependence on the scattering artifact is lower (Lindberg et al., 1999). The discrepancy between real and measured \mathring{a}_{abs} is therefore probably smaller for an AE than for PSAP measurements. Reviewing literature and taking into account all considerations about the modification of the absorption wavelength dependence induced by measurement artifacts, an AE correction that minimizes modifications of \mathring{a}_{abs} is preferred.

Both the Arnott and the Schmid corrections induce important \mathring{a}_{abs} modifications. A finer analysis of these wavelength dependent modifications shows that they are present mostly for low \mathring{a}_{scat} and are greater for low positive \mathring{a}_{abs} . They are in fact due to the wavelength dependence of the scattering correction constants $\alpha(\lambda)$: since the α wavelength dependence is fixed for both Arnott and Schmid corrections ($\alpha \sim \lambda^{+1.3}$), the subtracted amount of the scattering correction increases with increasing wavelength for low \mathring{a}_{scat} . The large \mathring{a}_{abs} increase (Fig. 6) is therefore determined by the scattering properties of both non-absorbing and Diesel-soot aerosols measured by Arnott in a chamber study and cannot be explained by considerations about the wavelength dependence modification due to the scattering correction. The Weingartner correction

AMTD

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Introduction





Printer-friendly Version



keeps the measured \mathring{a}_{abs} absolutely unmodified for all aerosol types. The new correction similar to the Arnott or to the Schmid method mainly preserves \mathring{a}_{ahs} , inducing a small increase of å_{abs} in some cases, particularly in presence of mineral dust as determined for the JFJ dataset. Since no å_{abs} reference measurements are available to determine the true absorption coefficient wavelength dependence, the new correction probably offers a median solution between no modification and large modifications that are not based on theoretically or experimentally founded reasons.

4.3 Evaluation of each AE correction

Taking into account the criteria for a good correction procedure defined in Sect. 4.1, Table 6 summarizes the applicability and the performances of each AE correction scheme.

The AE correction proposed by Virkkula et al. (2007) does not take into account all the known artifacts and produces a lot of outliers. It is therefore not recommended to use it on atmospheric aerosol long-term datasets.

The Arnott correction (Arnott et al., 2005) proves to have technical limits mainly due to the generation of new negative b_{abs} values, that are greatly enhanced at lower aerosol concentration and particularly at higher wavelengths. It introduces large adhoc modifications of the absorption wavelength dependence, which are not grounded in theory. The Arnott method remains therefore difficult to apply to all atmospheric aerosol datasets.

The empirical Weingartner correction (Weingartner et al., 2003) is easy to use, since it does not directly need b_{scat} measurements, even if an evaluation of ω_0 can influence the filter-loading correction. However the determination of the f constant used in the Weingartner filter-loading correction is not clearly defined. It does not modify the \mathring{a}_{abs} and shows a very good agreement with the MAAP.

The Schmid correction (Schmid et al., 2006) needs concomitant scattering coefficient measurements and cannot cope with negative values of b_{abs} . It also introduces artifacts in the absorption wavelength dependence like the Arnott method, but to a

AMTD

2, 1725-1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page **Abstract** Introduction Conclusions References **Tables Figures** I◀











Full Screen / Esc

Printer-friendly Version



lesser extent. It shows good agreement with the MAAP $b_{\rm abs}$, leading however to lower slopes than the Weingartner scheme as well as the new corrections for two stations out of the four.

Both new corrections need concomitant scattering coefficient measurements. Similarly to their originals, the new correction like Arnott generates some new negative $b_{\rm abs}$ values and the new correction like Schmid cannot cope with negative values of $b_{\rm abs}$. Both new corrections introduce a small modification of $\mathring{a}_{\rm abs}$ and show a very good agreement with the MAAP $b_{\rm abs}$.

All the corrections apart from the Arnott one lead to pretty good agreements with the MAAP $b_{\rm abs}$. The small differences in the agreement with the MAAP between the Schmid, the Weingartner and both new corrections leads to the conclusion that $C_{\rm ref}$ is indeed the most important parameter in order to obtain a good agreement with another instrument. The scattering correction, which is achieved differently for the Weingartner, Schmid and the new corrections, has therefore a far lower impact than the multiple scattering correction. For stations where no $b_{\rm abs}$ reference measurements are available, Fig. 4 yields an estimate of $C_{\rm ref}$ if ω_0 is known.

The monthly slopes between AE and MAAP $b_{\rm abs}$ vary as a function of time for all datasets with the standard deviations of the monthly slope reaching 8% to 25%. Similar variations are also found as a function of other extensive and intensive aerosol parameters such as $b_{\rm scat}$, $\mathring{a}_{\rm scat}$, $\mathring{a}_{\rm abs}$ and ω_0 . The slope between the MAAP $b_{\rm abs}$ calculated by the MAAP software including the backscatter measurement and from the MAAP transmission measurement only (similarly to AE) are also not constant as a function of time (Fig. 7). The inter-comparison of six carbon measurement methods at the Fresno Supersite during a year (Park et al., 2006) also resulted in standard deviations of monthly values ranging between 6% and 19% of the average of all monthly values. Better results are only met for similar instruments such as two AE's. The comparison of $b_{\rm abs}$ measured by different types of instruments seems to depend always on the aerosol properties despite the presently applied corrections algorithms. Lack et al. (2008) showed that there is a correlation between the aerosol organic content (OC),

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Introduction

References



Tables Figures

I ✓ ►I

✓ ►

Back Close

Full Screen / Esc

Printer-friendly Version



ranged between 0 and $17 \,\mu \mathrm{g} \, \mathrm{m}^{-3}$, and the ratio between the b_{abs} measured by a PSAP and a PAS. The aerosol OC was measured during two campaigns of 2.5 months duration at the JFJ (Cozic et al., 2007). Since the daily mean OC showed small variations and did not exceed $1.5 \,\mu \mathrm{g} \, \mathrm{m}^{-3}$ at the JFJ, no correlation between the measured OC and the ratio between the AE and MAAP b_{abs} was found in this dataset. However, several results presented in this paper, particularly the monthly slopes presented in Fig. 7, indicate that the variability in aerosol composition is a main clue to the variable slopes between the absorption coefficients measured by the MAAP and AE instruments at the four sites.

5 Conclusions and recommendations

Corrections developed by Weingartner, Arnott, Schmid and Virkkula have been applied to datasets obtained from four stations with various aerosol types and loadings, two stations having a multi-wavelength AE and two having white-light AE. Two new corrections based on previously published ones are also presented and are applied to the four datasets as well as to two datasets measured in the Amazon basin and in Thessaloniki. The main modifications introduced by these new corrections are firstly a new parameterization of the scattering correction, which depends on the scattering coefficient and Angström exponent, and secondly the consideration of the optical properties of all aerosol embedded in the filter for the filter loading and for the scattering corrections. Comparisons with MAAP b_{abs} were performed for all correction schemes and all four datasets. Principal criteria for a sound AE correction scheme were determined as to provide a good agreement with a reference instrument and to preserve \mathring{a}_{ahs} , which is an important parameter measured by the multi-wavelength AE. The Arnott correction generates many new negative $b_{\rm abs}$ at low aerosol concentrations, and it does not preserve \mathring{a}_{abs} . The Schmid correction leads to a good agreement with the MAAP, with slopes between AE and MAAP $b_{
m abs}$ somewhat lower than the other corrections and it does not preserve \mathring{a}_{abs} either. The empirical Weingartner correction does not need

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien





the simultaneous $b_{\rm scat}$ measurement; it has no application restriction for lower aerosol concentrations; it has good agreement with MAAP and it induces no modification of $\mathring{a}_{\rm abs}$. The f constant determining the filter-loading correction is however not well defined. The new correction scheme has also good agreement with MAAP and it mainly preserves $\mathring{a}_{\rm abs}$ allowing however a modification of $\mathring{a}_{\rm abs}$, for example in the case of mineral dust events. The new correction similar to the Schmid method leads to a better performance than the one similar to the Arnott method.

The results and recommendations for the AE measurements and correction algorithms are summarized as follows:

- Measured attenuation values need to be stored in databases, since they are used for all AE corrections.
- Concomitant scattering coefficient measurements are highly recommended, since they are used in all the corrections, and an estimation of the single scattering albedo ω_0 is important in the determination of several correcting constants.
- The determination of the multiple scattering constant C_{ref} is crucial to allow good agreement with MAAP data. The analysis of data from four stations permits a first estimation of C_{ref} as a function of the single scattering albedo ω_0 . Further analyses, particularly at stations with very high BC load, are however recommended prior to determining a universal correlation between C_{ref} and ω_0 .
- The new filter-loading correction R_{new} is the most appropriate algorithm at two stations because it leads to the most constant ratio between AE and MAAP absorption coefficients as a function of ATN.
- Modification of the absorption coefficient wavelength dependence is a very important factor issue for global radiation transfer assessments. Considerations on the wavelength dependence of each AE artifact does not allow to determine if an increase or a decrease of å_{abs} occurs. The use of multi-wavelength reference methods (when available) should resolve this issue.

25

AMTD

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Abstract Introduction

Conclusions References

Tables Figures

|**4** | **>**|

Back Close

Full Screen / Esc

Printer-friendly Version



- If no scattering coefficient measurements are available, the new correction with α_{new} =0, which corresponds to the Weingartner correction apart from the R calculation, or the Weingartner correction are recommended.
- If scattering coefficient measurements are available, the new correction procedure similar to the Schmid method is recommended.

Acknowledgements. We thank the International Foundation High Altitude Research Stations Jungfraujoch and Gornergrat (HFSJG), which made it possible for PSI to carry out the experiments at the High Altitude Research Station at the Jungfraujoch. This work was supported by MeteoSwiss within the Swiss program to the Global Atmosphere Watch (GAW) of the World Meteorological Organization. Measurements at Jungfraujoch, Mace Head and Cabauw were also supported by the EC project EUSAAR (European Supersites for Atmospheric Aerosol Research). For the measurements at Mace Head we acknowledge the Irish Environmental Protection Agency for a Senior Researcher Fellowship under the STRIVE 2007-2013 program, and for the measurements at Cabauw, we acknowledge the support of the Dutch BSIK-program. Financial support from the EC-Project SCOUT-O3 is also acknowledged. The measurements at the Amazon Basin were performed within the framework of the Smoke, Aerosols, Clouds, Rainfall, and Climate (SMOCC) project, a European contribution to the Large-Scale Biosphere-Atmosphere Experiment in Amazonia (LBA), under the leadership of M. O. Andreae and P. Artaxo. It was financially supported by the Max Planck Society (MPG) and the Environmental and Climate Program of the European Commission (contract NEVK2-CT-2001-00110 SMOCC).

References

Anderson, T. L. and Ogren, J. A.: Determining aerosol radiative properties using the TSI 3563 integrating nephelometer, Aerosol Sci. Tech., 29(1), 57–69, 1998.

Andreae, M. O., Rosenfeld, D., Artaxo, P., Costa, A. A., Frank, G. P., Longo, K. M., and Silva-Dias, M. A. F.: Smoking rain clouds over the Amazon, Science, 303(5662), 1337-1342, 2004.

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.



Introduction References

Figures

Close

Full Screen / Esc

Printer-friendly Version



- Arnott, W., Hamasha, K., Moosmüller, H., Sheridan, P. J., and Ogren, J. A.: Towards aerosol light-absorption measurements with a 7-wavelength aethalometer: Evaluation with a photoacoustic instrument and 3-wavelength nephelometer, Aerosol Sci Tech., 39(1), 17–29, 2005.
- Baltensperger, U., Gäggeler, H. W., Jost, D. T., Lugauer, M., Schwikowski, M., Weingartner, E., and Seibert, P.: Aerosol climatology at the high-alpine site Jungfraujoch, Switzerland, J. Geophys. Res., 102(D16), 19707-19715, 1997.
- Bohren, C. F. and Huffman, D. R.: Absorption and Scattering by small Particles, Wiley, New York, 1983.
- Bond, T. and Bergstrom, R.: Light absorption by carbonaceous particles: An investigative review, Aerosol Sci. Tech., 40, 1-41, 2006.
- Bond, T. C., Anderson, T. L., and Campbell, D.: Calibration and intercomparison of filter-based measurements of visible light absorption by aerosols, Aerosol Sci. Tech., 30(6), 582-600, 1999.
- Cozic, J., Verheggen, B., Weingartner, E., Crosier, J., Bower, K. N., Flynn, M., Coe, H., Henning, S., Steinbacher, M., Henne, S., Collaud Coen, M., Petzold, A., and Baltensperger, U.: Chemical composition of free tropospheric aerosol for PM1 and coarse mode at the high alpine site Jungfraujoch, Atmos. Chem. Phys., 8, 407–423, 2008, http://www.atmos-chem-phys.net/8/407/2008/.
- Heintzenberg, J., Charlson, R. J., Clarke, A. D., Liousse, C., Ramaswamy, V., Shine, K. P., Wendisch, M., and Helas, G.: Measurements and modeling of aerosol single-scattering albedo: progress, problems and prospects, Contributions to atmospheric physics, 70(4), 249–263, 1997.
- Horvath, H.: Atmospheric light absorption: a review, Atmos. Environ., 27(3), 293-317, 1993.

20

- Jennings, S. G., Kleefeld, C., O'Dowd, C. D., Spain, T. G., O'Brien, P., Roddy, A. F., and O' Connor, T. C.: Mace Head Atmospheric Research Station - characterization of aerosol radiative parameters, Boreal Environ. Res., 8, 303–314, 2003.
- Junker, C., Jennings, S. G., and Cachier, H.: Aerosol light absorption in the North Atlantic: trends and seasonal characteristics during the period 1989 to 2003, Atmos. Chem. Phys., 6, 1913-1925, 2006, http://www.atmos-chem-phys.net/6/1913/2006/.
- 30 Kaminski, U., Fricke, M., Wilhelm, R., Plörer, E., Hofmann, M., and Schafranek, R.: Aerosol Measurements at the GAW Global Station Hohenpeissenberg, in: The German Contribution to the GAW Programme: Upon the 225th Anniversary of GAW Hohenpeissenberg Observatory (WMO TD No. 1336), edited by: Fricke, W., Barrie, L. A., and Schleyer, R., GAW report

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Introduction

References

Figures

Close





No. 167, 72–83, 2006.

20

- Lack, D. A., Cappa, C. D., Baynard, T., Massoli, P., Covert, D. S., Sierau, B., Bates, T. S., Quinn, P. K., Lovejoy, E. R., and Ravishankara, A. R.: Bias in filter-based aerosol absorption measurements due to organic aerosol loading: Evidence from ambient sampling, Aerosol Sci. Tech., 42(12), 1033–1041, 2008.
- Lindberg, J. D., Douglass, R. E., and Garvey, D. M.: Atmospheric particulate absorption and black carbon measurement, Appl. Optics, 38, 2369–2376, 1999.
- Liousse, C., Cachier, H., and Jennings, S. G.: Optical and thermal measurements of black carbon aerosol content in different environments: Variation of the specific attenuation crosssection, sigma (σ), Atmos. Environ., 27(8), 1203–1211, 1993.
- Lugauer, M., Baltensperger, U., Furger, M., Gäggeler, H. W., Jost, D. T., Schwikoswki, M., and Wanner, H.: Aerosol transport to the high Alpine sites Jungfraujoch (3454 m a.s.l.) and Colle Gnifetti (4452 m a.s.l.), Tellus B, 50, 76–92, 1998.
- Moosmüller, H., Arnott, W. P., and Rogers, C. F.: Methods for real-time, in situ measurement of aerosol light absorption, J. Air Waste Manage., 47(2), 157–166, 1997.
- Müller, T., Henzing, B., and Wiedensohler, A.: Intercomparison of three types of absorption photometers, European Aerosol Conference EAC-2008, Thessaloniki, Greece, 23.29 August 2008, T04A05P, 2008.
- Nessler, R., Weingartner, E., and Baltensperger, U.: Adaptation of dry nephelometer measurements to ambient conditions at the Jungfraujoch, Environ. Sci. Technol, 39(7), 2219-2228, 2005.
- Park, K., Chow, J. C., Watson, J. G., Trimble, D. L., Doraiswamy, P., Park, K., Arnott, W. P., Stroud, K. R., Bowers, K., and Bode, R.: Comparison of continuous and filter-based carbon measurements at the Fresno Supersite, J. Air Waste Manage., 56(4), 474-491, 2006.
- 25 Petzold, A., Kopp, C., and Niessner, R.: The dependence of the specific attenuation crosssection on black carbon mass fraction and particle size, Atmos. Environ., 31(5), 661-672, 1997.
 - Petzold, A., Schloesser, H., Sheridan, P., Arnott, W., Ogren, J., and Virkkula, A.: Evaluation of multiangle absorption photometry for measuring aerosol light absorption, Aerosol Sci. Tech., 39(1), 40-51, 2005.
 - Petzold, A. and Schönlinner, M.: Multi-angle absorption photometry a new method for the measurement of aerosol light absorption and atmospheric black carbon, J. Aerosol Sci., 35(4), 421–441, 2004.

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Introduction

References

Figures

Close





- Rice, J.: Comparison of integrated filter and automated carbon aerosol measurements at Research Triangle Park, North Carolina, Aerosol Sci. Tech., 38(S2), 23–36, 2004.
- Russchenberg, H., Bosveld, F., Swart, D., Ten Brink, H., de Leeuw, G., Uijlenhoet, R., Arbesser-Rastburg, B., Van der Marel, H., Ligthart, L., Boers, R., and Apituley, A.: Ground-based atmospheric remote sensing in the Netherlands: European outlook, IEICE T. Commun., E88-B(6), 2252-2258, 2005.
- Saathoff, H., Naumann, K. H., Schnaiter, M., Schöck, W., Weingartner, E., Baltensperger, U., Krämer, L., Bozoki, Z., Pöschl, U., and Niessner, R.: Carbon mass determinations during the AIDA soot aerosol campaign 1999, J. Aerosol Sci., 34(10), 1399–1420, 2003.
- Schaap, M., Apituley, A., Timmermans, R. M. A., Koelemeijer, R. B. A., and de Leeuw, G.: Exploring the relation between aerosol optical depth and PM2.5 at Cabauw, the Netherlands, Atmos. Chem. Phys., 9, 909-925, 2009, http://www.atmos-chem-phys.net/9/909/2009/.
 - Schmid, O., Artaxo, P., Arnott, W. P., Chand, D., Gatti, L. V., Frank, G. P., Hoffer, A., Schnaiter, M., and Andreae, M. O.: Spectral light absorption by ambient aerosols influenced by biomass burning in the Amazon Basin. I: Comparison and field calibration of absorption measurement techniques, Atmos. Chem. Phys., 6, 3443-3462, 2006, http://www.atmos-chem-phys.net/6/3443/2006/.
 - Schnaiter, M., Schmid, O., Petzold, A., Fritzsche, L., Klein, K. F., Andreae, M. O., Helas, G., Thielmann, A., Gimmler, M., and Möhler O.: Measurement of wavelength-resolved light absorption by aerosols utilizing a UV-VIS extinction cell, Aerosol Sci. Tech., 39(3), 249-260, 2005.

20

- Seinfeld, J. H. and Pandis, S. N.: Atmospheric Chemistry and Physics, From Air Pollution to Climate Change, John Wiley & Sons, Inc., New York, 1998.
- Slowik, J. G., Cross, E. S., Han, J. H., Davidovits, P., Onasch, T. B., Jayne, J. T., Williams, L. R., Canagaratna, M. R., Worsnop, D. R., and Chakrabarty, R. K.: An inter-comparison of instruments measuring black carbon content of soot particles, Aerosol Sci. Tech., 41(3), 295–314, 2007.
- Steiger, M.: An overview of Diesel and wood burning soot chamber experiments and a loading effect correction for Aethalometer measurements, Diploma thesis, ETH Zürich, Switzerland, 2008.
- Virkkula, A., Ahlquist, N., Covert, D., Arnott, W., Sheridan, P., Quinn, P., and Coffman, D.: Modification, calibration and a field test of an instrument for measuring light absorption by particles, Aerosol Sci. Tech., 39(1), 68-83, 2005a.

AMTD

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Introduction

References

Figures

Close





- Virkkula, A., Ahlquist, N., Covert, D., Sheridan, P., Arnott, W., and Ogren, J.: A three-wavelength optical extinction cell for measuring aerosol light extinction and its application to determining light absorption coefficient, Aerosol Sci. Tech., 39(1), 52–67, 2005b.
- Virkkula, A., Makela, T., Hillamo, R., Yli-Tuomi, T., Hirsikko, A., Hameri, K., and Koponen, I. K.: A simple procedure for correcting loading effects of aethalometer data, J. Air Waste Manage., 57(10), 1214–1222, 2007.
- Wallace, L.: Real-time measurements of black carbon indoors and outdoors: A comparison of the Photoelectric Aerosol Sensor and the Aethalometer, Aerosol Sci. Tech., 39(10), 1015–1025, 2005.
- Weingartner, E., Nyeki, S., and Baltensperger, U.: Seasonal and diurnal variation of the aerosol size distribution (10<D<750 nm) at a high alpine site (Jungfraujoch 3580 m asl), J. Geophys. Res., 104, 26 809–26 820, 1999.</p>
 - Weingartner, E., Saathof, H., Schnaiter, M., Streit, N., Bitnar, B., and Baltensperger, U.: Absorption of light by soot particles: Determination of the absorption coefficient by means of Aethalometers. J. Aerosol Sci., 34, 1445–1463, 2003.

15

AMTD

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien




Table 1. Description of the used datasets, including the measuring sites, the instruments, time periods and brief site characteristics.

ID	site	description	inlet	AE	MAAP/PAS	Neph	period, project	reference
JFJ	Jungfraujoch 46° N, 8° E 3580 m a.s.l.	Alps, free troposphere, PBL influence	Dried TSP	AE-31 λ: 370–950 nm Filter: Pallflex Q250F	MAAP λ: 630 nm	TSI 3563 λ: 450–700 nm	3.2003–12.2007 GAW aerosol monitoring program	Baltensperger et al. (1997), Lugauer et al. (1998)
CAB	Cabauw 51° N, 4° E -0.7 m a.s.l.	Plane, moderate maritime, near populated and industrialized areas	Before March 2008, ambient air PM ₁₀ inlet placed near the instruments on a platform at 60 m. After March 2008, 60 m inlet pipe, PM ₁₀ , followed by a bundle of nafion dryers.	AE-31 \(\alpha\): 370–950 nm Filter: paperbacked quartz filter	MAAP λ: 630 nm	TSI 3563 λ: 450–700 nm	2.2008-7.2008 CESAR	Schaap et al. (2009), Russchenberg et al. (2005)
MHD	Mace Head 53° N, 10° W 5 m a.s.l.	Coastal site, with prevailing marine (North Atlantic) air masses >50% of the time	PM ₁ , 15 m sampling line with inlet at 10 m	AE-16 White light Filter: glass fiber filter GF10	MAAP λ: 630 nm	TSI 3563 λ: 450–700 nm	1.2007–12.2007	Jennings et al. (2003), Junker et al. (2006)
HOP	Hohenpeissenberg 48° N, 11° E 985 m a.s.l.	Alpine Foothill, PBL influenced, rural background	PM ₁₀ , heated 2° C above ambient air	AE-10-IM λ: white light Filter: Pallflex Q250F	MAAP λ: 630 nm	TSI 3563 λ: 450–700 nm	1.2005–12.2005	Kaminski et al. (2006)
AMA	Amazon basin 11° S, 62° W 235 m a.s.l.	pasture site frequently affected by bio- mass burning haze from fire-assisted land clearing	5 m inlet tubing, Non-dried PM ₁₀ for AE, and dried PM _{1.5} for PAS	AE-30 λ: 370–950 nm Filter: quartz fiber	PAS λ: 532 nm	TSI 3563 λ: 450–700 nm	9.2002–11.2002 LBA-SMOCC	Andreae et al. (2004), Schmid et al. (2006)
THE	Thessaloniki 40° N, 22° W 25 m a.s.l.	City, polluted	Ambient TSP	AE-31 λ: 370–950 nm Filter: Pallflex Q250F	not available	TSI 3563 λ: 450–700 nm	7.2006 SCOUT-O3	

2, 1725-1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page

Abstract Conclusions Tables I◀



Introduction













Full Screen / Esc

Printer-friendly Version



Table 2. Symbols, units, description of the parameters used in this paper as well as the corresponding symbols used in the papers describing the previously published correction schemes. The used nomenclature follows symbols commonly used by Seinfeld and Pandis (1998).

Symbol	Units	Name and function	Corresponding symbols in other papers
α		Parameter for the Arnott scattering correction	α in Arnott m_s in Schmid s in Virkkula
α_{new}		Parameter for the new scattering correction	
å _{abs}		Absorption Ångström exponent. It describes the wavelength dependence of $b_{\rm abs}$.	$lpha$ in Weingartner $lpha_a$ in Schmid
$\mathring{a}_{\text{scat}}$		Scattering $\mathring{\text{A}}$ ngström exponent. It describes the wavelength dependence of b_{scat} .	$lpha_s$ in Schmid
å _{scat,non-abs}		Scattering Ångström exponent for Arnott ammonium sulfate aerosol experiment	b in Arnott
$\bar{\mathring{a}}_{\mathrm{scat},s,n}$		$\mathring{a}_{\text{scat}}$ mean over the n first measurements since a filter spot (s) change	
Α	m ²	Area of the sample spot	A in Weingartner, Arnott, Schmid and Virkkula
ATN		Filter attenuation measured and recorded by AE.	ATN in Weingartner, Arnott Schmid and Virkkula
$oldsymbol{eta}_{scat}$	m^{-1}	Proportionality constant of the wavelength power law dependence of b_{scat}	A in Arnott
$eta_{ extsf{scat}, extsf{non-abs}}$	m ⁻¹	Proportionality constant of the wavelength power law dependence of $b_{\rm scat}$ for Arnott ammonium sulfate aerosol experiment.	
$ar{eta}_{scat,s,n}$	m^{-1}	$eta_{ m scat}$ mean over the n first measurements since a filter spot change.	
$b_{ m abs}$	m^{-1}	Absorption coefficient	$b_{ m abs}$ in Weingartner $B_{ m abs}$ in Arnott $\sigma_{ m aeth}$ in Schmid $\sigma_{ m abs}$ in Virkkula
b _{ATN}	m ⁻¹	Attenuation coefficient measured and recorded by AE.	$b_{ m ATN}$ in Weingartner $B_{ m aeth}$ in Arnott $\sigma_{ m ATN}$ in Schmid σ_0 in Virkkula
b _{ATN,R_corrected}	m^{-1}	Attenuation coefficient corrected only for the filter-loading artifact R.	
b _{abs,ref_MAAP}	m^{-1}	Reference absorption coefficient from MAAP.	$b_{ m abs}$ in Weingartner
BC_0	g/m ³	Black carbon concentration measured and recorded by AE.	BC ₀ in Virkkula
$b_{ m scat}$	m^{-1}	Scattering coefficient	b_s in Weingartner ${\sf B}_{\sf sca}$ in Arnott σ_s in Schmid σ_{SP} in Virkkula
$\bar{b}_{scat,s,n}$	m^{-1}	$b_{\rm scat}$ mean over the n first measurements since a filter spot change	

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

Title	Page
Abstract	Introdu
Conclusions	Refere
Tables	Figu
I⁴	<u> </u>
4	-
Back	Clo
Full Scr	een / Esc
Printer-frie	ndly Versi
Interactive	Discussion



Table 2. Continued.

Symbol	Units	Name and function	Corresponding symbols in other papers
С	m ⁻¹	Proportionality constant of the power-law dependence of b_{ATN} and b_{scat} for Arnott non-absorbing aerosol experiment	c in Arnott
C_{ref}		Multiple scattering correction constant	C in Weingartner
$C_{\rm scat}$		Scattering correction in the Schmid method	M in Arnott C in Schmid
d		Exponent of the power-law dependence of b_{ATN} and b_{scat} for Arnott non-absorbing aerosol experiment	d in Arnott
f		Parameter estimating the slope of the $b_{\rm ATN}$ versus ATN curve. It parameterizes the Weingartner filter-loading correction.	f in Weingartner and Schmid
i		Counter for the i th number of filter spots.	
<i>I</i> ₀		Light intensity through a pristine portion of the filter	I ₀ in Weingartner, Schmid and Virkkula
Ī		Light intensity through the loaded filter	/ in Weingartner, Schmid and Virkkula
k		Variable defining the proportionality between the last point of a filter spot and the first points of the next filter spot. It parameterizes the Virkkula filter-loading correction	<i>k</i> in Virkkula
m		Slope of the f versus $1-\omega_0$ curve. It parameterizes the new filter-loading correction	a in Weingartner
n		Counter for the <i>n</i> -th measurement since a filter change	
R_A		Arnott filter-loading correction	
R_{new}		New filter-loading correction	
R_V		Virkkula filter-loading correction	
R_w		Weingartner filter-loading correction, with a fixed chosen f value chosen for the whole dataset.	R in Weingartner and Schmid
SG	$m^2 g^{-1}$	Spectral mass specific attenuation cross-section proposed by the manufacturer	$\sigma_{ m abs}$ in Weingartner SG in Arnott $lpha_{ m ATN}$ in Schmid
t	s	Time	t in Weingartner, Arnott, Schmid and Virkkula
$\tau_{a,fx}$		Filter absorption optical depth for the filter fraction <i>x</i> that has particles embedded in it.	τ_{a,f_X} in Arnott
V	m ³ /s	Volumetric flow rate	Q in Weingartner, Schmid and Virkkula V in Arnott
ω ₀		Single scattering albedo	ω_0 in Weingartner and Schmid ω in Arnott
ω _{0.ref}		Single scattering albedo at a chosen reference wavelength $\lambda_{\rm ref}$	$\omega_{0,\mathrm{ref}}$ in Schmid
$\bar{\omega}_{0,s,n}$		ω_0 mean over the n first measurements since a filter spot change	
λ	nm	Wavelength	λ in Weingartner, Arnott, Schmid and Virkkula

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page

Introduction

References

Figures

M

Close

Abstract
Conclusions
Tables

I

Back
Full Sc



Full Screen / Esc

Printer-friendly Version

Table 3. Use of the scattering coefficient, minimum, maximum and mean of 1/R values, for all four filter-loading corrections and for the four measuring stations. For the Weingartner correction f = 1.025, f = 1.1, f = 1.05, f = 1.025, f = 1.025 and f = 1.2 were taken for JFJ, CAB, MHD, HOP, AMA and THE datasets, respectively. R values were taken at λ =660 nm for multi-wavelength AE (JFJ, CAB, AMA and THE).

			JFJ			CAB			MHD			HOP			AMA			THE	
Filter-loading correction	Use of $b_{\rm scat}$	Min (1/ <i>R</i>)	Max (1/ <i>R</i>)	Mean (1/ <i>R</i>)	Min (1/ <i>R</i>)	Max (1/ <i>R</i>)	Mean (1/ <i>R</i>)	Min (1/R)	Max (1/ <i>R</i>)	Mean (1/ <i>R</i>)	Min (1/ <i>R</i>)	Max (1/ <i>R</i>)	Mean (1/ <i>R</i>)	Min (1/R)	Max (1/R)	Mean (1/R)	Min (1/ <i>R</i>)	Max (1/ <i>R</i>)	Mean (1/ <i>R</i>)
Weingartner R _W	no	0.87	1.02	1.01	0.77	1.06	1.01	0.84	1.06	1.03	0.91	1.02	1.01	0.93	1.03	1.01	0.94	1.02	1.01
Arnott R _A	no	1.00	1.07	1.00	1.00	1.15	1.01	1.00	1.40	1.03	1.00	1.22	1.03	1.00	1.81	1.08	1.00	1.30	1.03
Virkkula R _V	no	-250	1803	2.23	-3.60	6.99	1.05	-1.18	10.08	1.14	-87	223	1.31	-1.61	8.73	1.07	-134	182	0.92
New correction R _{new}	yes	1.00	1.36	1.05	1.00	1.23	1.05	1.00	2.56	1.12	1.00	1.37	1.09	1.00	2.05	1.29	1.00	1.27	1.09

2, 1725–1770, 2009

Correction algorithms for **Aethalometer light** absorption coefficien

M. Collaud Coen et al.

Title Page Introduction Abstract Conclusions References **Tables Figures** Ī◀ Back



M

Full Screen / Esc

Printer-friendly Version Interactive Discussion



Table 4. Mean $C_{\rm ref}$ constants with standard deviations for all four filter-loading corrections for JFJ, CAB, MHD and HOP stations. The correlation with the MAAP was done with λ =660 nm at JFJ and CAB.

	JFJ	CAB	MHD	HOP
Weingartner R _W	2.81±0.11	4.09±0.12	3.05±0.10	2.81±0.10
Arnott R_A	2.80 ± 0.13	4.12±0.06	3.08 ± 0.11	2.78±0.09
Virkkula R_V	7.77±4.15	4.57 ± 0.49	3.83 ± 0.41	3.16 ± 0.43
New correction R _{new}	2.88±0.08	4.26±0.11	3.51 ± 0.23	2.89±0.03

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien



Table 5. Averages and standard deviations of the monthly slopes between MAAP (λ =630 nm) and AE (reported at λ =630 nm) absorption coefficients measured at the four stations for manufacturer estimation and each AE correction.

	AE	Weingartner	Arnott	Schmid	New correction similar to			
	manufacturer				Arnott (Eq. 14a)	Schmid (Eq. 14b)		
JFJ	2.73±0.41	1.00±0.15	-inf	0.90±0.14	1.09±1.35	0.97±0.19		
CAB	3.64 ± 0.93	0.84 ± 0.34	1.77±0.59	0.88 ± 0.26	0.85 ± 0.35	0.84 ± 0.24		
MHD	2.85±0.24	0.97 ± 0.08	15.2±43.7	1.04±0.11	0.98 ± 0.15	0.99 ± 0.15		
HOP	2.40±0.20	0.85 ± 0.24	2.48±0.71	0.67±0.21	0.78±0.23	0.76±0.23		

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien




Table 6. Evaluation of the applicability and the performance of AE corrections based on the criteria defined under Sect. 4.1.

Correction Scheme	All artifacts	Applicable to $b_{\rm abs}$ <0	Need $b_{\rm scat}$ measurement	Outliers and new $b_{\rm abs}$ <0	Agreement with MAAP	$\mathring{a}_{\mathrm{abs}}$ modification
Virkkula	no	yes	no	yes	_	_
Weingartner	yes	yes	no	no	very good	none
Arnott	yes	yes	yes	yes, a lot	not good	large
Schmid	yes	no	yes	no	good	large
New like Arnott	yes	no	yes	yes, few	very good	small
New like Schmid	yes	no	yes	no	very good	small

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien







Fig. 1. Description of the sequence of steps to apply the new correction scheme.

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien






Fig. 2. 1/R for the four filter-loading corrections as a function of time for about one month measured at the JFJ at $\lambda = 370$ nm.

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien








Fig. 3. $C_{\rm ref}$ calculated from $b_{\rm abs}$ corresponding to a defined ATN as a function of ATN for the four filter-loading corrections and for the JFJ, CAB, MHD and HOP datasets as well as for the $b_{\rm abs}$ AE output calculated as recommended by the manufacturer. For JFJ and CAB, $b_{\rm abs}$ at λ =660 nm are represented. For JFJ, the Virkkula filter-loading correction leads to too high $C_{\rm ref}$ values varying between 3.7 and 17, and for HOP, the AE $b_{\rm abs}$ calculated as recommended by the manufacturer leads to around 0.6; these extreme values are not shown for clarity purpose.

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.

Title Page






Fig. 4. $C_{\rm ref}$ as a function of the single scattering albedo ω_0 for the JFJ, CAB, MHD and HOP datasets. For the JFJ and CAB datasets, $C_{\rm ref}$ and ω_0 are also separated between summer (S) and winter (W).

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien






Fig. 5. Dependence of α_{new} at λ =660 nm as a function of scattering coefficient at λ =550 nm and of C_{scat} as a function of ω_0 at λ =660 nm for the four datasets. α_{new} (Fig. 5a) and C_{scat} (Fig. 5b) were calculated with the new correction scheme and represent the multiplying factor of the scattering correction (Eq. 14a), and the scattering correction in Schmid method (Eq. 14b), respectively.

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien

M. Collaud Coen et al.






Fig. 6. Histogram of the relative change (in percent) of AE derived \mathring{a}_{abs} values using Arnott, Schmid and the new corrections, the method similar to Schmid being represented here. The JFJ and CAB datasets are presented in light and dark colors, respectively. The numbers of cases represented for $\pm 100\%$ correspond to \mathring{a}_{abs} changes equal or greater than $\pm 100\%$, with values reaching sometimes up to $\pm 200\%$ or greater.

2, 1725–1770, 2009

Correction algorithms for Aethalometer light absorption coefficien






Fig. 7. Monthly regression slopes between the absorption coefficients at λ =630 nm from the MAAP and the AE (interpolated to λ =630 nm) at the four stations for all the different correction schemes as well as between the absorption coefficients given by the MAAP and those measured by the MAAP in transmission only (black dashed line). For all the corrections schemes, the $C_{\rm ref}$ values from Table 4 are applied. The lack of measurements from September 2005 to March 2006 at the JFJ is due to a leak in the MAAP inlet.

2, 1725-1770, 2009

Correction algorithms for Aethalometer light absorption coefficien



