

We would like to thank Dr. Wolff for his question. The ozone instrument was running continuously from March through May, which is inconsistent with what is shown in the power table, as pointed out. The discrepancy is due to our deviation from the power schedule during this test deployment stage. For future deployments all instruments will follow their pre-designed schedule. However, we will have the ability to alter the schedule at any time should that be necessary. This has been clarified in the text: pp. 9-10 lines 235-240.