

Interactive comment on “Chemical ionization mass spectrometer (CIMS) for ambient measurements of ammonia” by D. R. Benson et al.

c. Percival

carl.percival@manchester.ac.uk

Received and published: 25 March 2010

The authors state that there have only been 3 uses of CIMS for the measurement of ammonia. This statement is incorrect there is at least one more, see for example Bobutzki, K., Blom, M., Braban, C., Coe, H., Famulari, D., Gallagher, M. W., Geddes, J., Ghalaieny, M., Jones, S., McGillen, M.R., Mohacsi, A., Murphy, J., Percival, C. J., Pogany, A., Popescu, R., Rantanen, S., Whitehead, J. M., Sutton, M.A., Nemitz, E. G. (2010) Field inter-comparison of eleven atmospheric ammonia measurement techniques Atmospheric Measurement Techniques, 3, 91-112.

Interactive comment on Atmos. Meas. Tech. Discuss., 3, 1133, 2010.