



Interactive  
Comment

## ***Interactive comment on “Operational considerations to improve total ozone measurements with a Microtops II ozone monitor”***

**by J. L. Gómez-Amo et al.**

**Anonymous Referee #2**

Received and published: 24 January 2012

The added plots certainly show that the linear interpolation method is clearly the proper technique to be applied to the data. This reviewer suggests that 3 figures should be included into the final paper showing the differences between the Brewer & MicroTOPS (i.e figures 1,4,7 or 2,5,8 or 3,6,9). A short paragraph discussing the fact that the other methods were tried and discarded would be helpful as well as a statement that this TOPS instrument seems to have degraded/changed in a more or less linear manner and that this technique could/should be applied to other instruments that have had calibration issues.

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Discussion Paper



---

Interactive  
Comment

C2702

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Discussion Paper

