

**An integrated flask
sample collection
system**

J. Turnbull et al.

An integrated flask sample collection system for greenhouse gas measurements

**J. Turnbull^{1,2}, D. Guenther^{2,3}, A. Karion^{2,3}, C. Sweeney^{2,3}, E. Anderson^{2,3},
A. Andrews³, J. Kofler^{2,3}, N. Miles⁴, T. Newberger^{2,3}, S. Richardson⁴, and
P. Tans³**

¹National Isotope Center, GNS Science, Lower Hutt, New Zealand

²CIRES, University of Colorado at Boulder, Boulder, CO, USA

³NOAA/ESRL, Boulder, CO, USA

⁴Pennsylvania State University, University Park, PA, USA

Received: 24 May 2012 – Accepted: 27 May 2012 – Published: 8 June 2012

Correspondence to: J. Turnbull (j.turnbull@gns.cri.nz)

Published by Copernicus Publications on behalf of the European Geosciences Union.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Abstract

A one hour integrated flask sampling system to collect air in automated NOAA/ESRL 12-flask packages is described. The integrating compressor system uses a mass flow controller to regulate the flow of air through a 15 l volume, thus providing a mixture of air collected over an hour-long period. By beginning with a high flow rate of 3800 standard liters per minute and gradually decreasing the flow rate over time to 290 standard liters per minute it is possible to obtain a nearly uniformly time averaged sample of air and collect it into a pressurized 0.7l flask. The weighting function determining the air mixture obtained is described in detail. Laboratory and field tests demonstrate that the integrated sample approximates a simple mean of air collected during the one-hour sampling time.

1 Introduction

Atmospheric measurements of greenhouse and other trace gases are important to understanding the global carbon cycle, climate change and air quality problems. Due to the extreme accuracy required to quantify global, regional and local impacts on the mixing ratios of these gases, especially in the case of long-lived greenhouse gases, flask measurements can play an essential role. Flask samples allow a single instrument with calibration standards tied directly to the World Meteorological Organization (WMO) to measure samples from many different locations (e.g. Conway et al., 2011; Brand, 2009). With flasks, small changes observed between one location and the next can be easily detected, whereas it may be difficult to keep multiple in situ instruments calibrated. These flask samples must be collected using a method that allows accurate, unbiased determination of the species of interest, and the samples must be collected in such a way that the results can be interpreted in terms of the regional and/or global atmosphere. Flask samples are typically collected as nearly instantaneous grab samples, whereby a flask is filled over a period of less than one minute into a pre-flushed

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



or pre-evacuated flask (e.g. Conway et al., 2011; Brand, 2009; Conway et al., 1994). However, there can be substantial short-term variability in mixing ratios of greenhouse gases at many locations, particularly at urban and other polluted sites. Grab samples may therefore also be difficult to interpret, particularly when the interpretation requires convolution with models and meteorological data which is typically available with time steps of one to three hours. Ongoing comparison of flask and in situ measurements is also a key part of data quality assurance for both types of measurement, yet comparisons are often compromised by the difficulty in matching the exact timing of the flask air collection with in situ measurements. Integrated flask samples can dramatically reduce the impact of small timing errors on these comparisons.

This paper describes an integrating compressor (IC) flask sample collection system to obtain air samples integrated over one hour. The IC is designed to be as simple as possible, while achieving the goal of collecting an integrated sample that is not contaminated, fractionated or otherwise altered during the collection process. The flasks are measured for more than 50 gases and isotopes, and the sample air must remain unaltered for all of these species. Therefore, wherever possible, parts that have already been tested for use with these gases are selected. Flexibility in the design, so that the integration time can be varied easily, is also an important consideration, as is the ability to remotely and automatically trigger sampling.

2 Sampler design and description

First we describe the principles and overview of the IC, and then describe each part in detail. Figure 1 shows the overall system. The air enters from the inlet through Synflex 1300 tubing (Eaton Corporation, Ohio, USA), and then passes through a dryer unit, removing the bulk of the water from the air. Then the air passes through a 15l spherical mixing chamber – the heart of the IC. Two pumps and a mass flow controller are used to regulate the flow of air through the mixing chamber. A bypass valve can be used to override the mass flow controller and allow a higher flow rate. All tubing

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



within the IC is 1/8" stainless steel. During the hour-long integration period, the air is flushed out through flasks contained in the NOAA/ESRL 12-flask Programmable Flask Package (PFP (<http://www.esrl.noaa.gov/gmd/ccgg/aircraft/index.html>; Sweeney et al., 2012), and at the end of the integration time, the flasks are filled and pressurized with air from the mixing chamber. In standby mode, when no sample is being collected, air is flushed through the mixing chamber and pumps and vented to atmosphere via a standby vent valve.

This design was chosen over several possible alternatives, such as: (1) a large volume or bag that is evacuated and then air is allowed to gradually enter over the designated time period; (2) a syringe like system where the integrating volume is initially small, and the volume is increased linearly through the integrating time period. Although these alternative designs may be conceptually simpler, they use components that have not been tested for the suite of gases of interest and extensive testing would be needed to ensure that the trace gas composition was not altered. This is of particular concern for bags, in which the trace gas composition of air could be altered by absorption/desorption due to the evacuation procedure or the bag material.

2.1 Mixing chamber

The mixing chamber is a ~15l spherical volume of electro-polished stainless steel made by Restek Corporation (Pennsylvania, USA). A 1/8" OD dip tube is inserted into the volume, so that air enters the volume at the base, and flows out from the top, thus ensuring well-mixed conditions inside the volume. First, we tested the mixing chamber to ensure that mixing is nearly complete throughout the fill time and thus can be predicted using a simple e-folding model. The e-folding model of a well-mixed volume predicts a mean lifetime (residence time, τ) of any given parcel of air in the mixing chamber of

$$\tau = 1/k \quad (1)$$

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



where k is the rate constant for the removal of the air from the volume; k is determined by the flow rate f and the volume of the mixing chamber v such that

$$k = \frac{f}{v}. \quad (2)$$

When an air parcel with an initial mass $m(t_0)$ enters the volume at time t_0 , the amount of that air parcel remaining at a later time t is determined according to the e-folding model as

$$m(t) = m(t_0)e^{-k(t-t_0)}. \quad (3)$$

To test the mixing inside the volume, two cylinders, one of dried natural air (“natural air”) with 378.95 ppm (moles CO_2 per mole dry air) of CO_2 , the other of air with no CO_2 (“zero air”), were connected to the mixing chamber. In this arrangement, the CO_2 mixing ratio is a proxy for $m(t)$. The regulators were set to release cylinder air at slightly over atmospheric pressure with the excess pressure vented to the room, so that the air entering the IC was at room pressure (850 mbar at 1600 m a.s.l. in Boulder, CO). The air then passed through the mixing chamber and pumping system, and the outlet was connected to a cavity ring-down spectrometer (Picarro Model G1301; Crosson, 2008) continuously monitoring the CO_2 mixing ratio. Initially, natural air flowed through the mixing chamber at the maximum flow rate of 3800 standard liters per minute (SPLM) until the Picarro measurement stabilized at 378.95 ppm. The incoming air was then switched to the zero air cylinder, and we observed the decay of the CO_2 mixing ratio while pumping at a known, constant flow rate of 3800 SPLM (Fig. 2). The mixing ratio decay fit an exponential decay curve extremely well, and we determined $k = 0.00428 \text{ s}^{-1}$, which implies a mean lifetime $\tau = 3.89 \text{ min}$ and volume v of 14.8 l, approximately equal to the nominal volume of the mixing chamber. This experiment confirms that air mixes well in the spherical integration volume, and follows a theoretical e-folding decay, dependent on flow rate and the size of the mixing chamber.

An integrated flask sample collection system

J. Turnbull et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

2.2 Weighting function

Ideally, the final mixture of air in the flask would have a flat, constant weighting function, such that at the end of a one-hour sampling interval, the mixing chamber would contain an equal amount of air from each minute of the hour, with 1.67 % of the total air in the chamber having entered it in each minute of the hour. The 15 l mixing chamber would therefore contain 0.25 l of air from each of the 60 min of the hour (Fig. 3). However, the actual weighting function is determined by the flow rate and mixing chamber volume. If the flow rate is kept constant, the weighting function can be calculated from Eqs. (2) and (3). In this constant flow rate case, the final air mixture is dominated by air from the last part of the hour. The higher the flow rate, the steeper the weighting function, so that the final air mixture is more strongly dominated by the last part of the hour (4000 SPLM case in Fig. 3). Lowering the flow rate results in a more even weighting function, but when the flow rate drops too low, some air from prior to the 60-min interval will remain inside the flask (1000 SPLM case in Fig. 3).

A perfectly even weighting function can be achieved if the flow rate is infinitely adjustable, starting with a very high flow rate, and gradually dropping the flow rate through time. As the following equations cannot be integrated analytically, we demonstrate the concept using a time step of one minute. Of the air that entered the flask during the last minute of the one-hour interval ($m_1(t_1)$), the amount remaining inside the flask at the end of the one-hour interval (final time t_f) is

$$m_1(t_f) = m_1(t_1)e^{-k_1(t_f-t_1)} \quad (4)$$

and the rate constant for the last time step k_1 depends on the flow rate f_1 during that last time step and the mixing chamber volume

$$k_1 = \frac{f_1}{V}. \quad (5)$$

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



The initial amount of air $m_1(t_1)$ introduced during this last time step is a function of the flow rate and length of the time step Δt (equal to $t_f - t_1$), such that

$$m_1(t_1) = f_1 \Delta t \quad (6)$$

and therefore

$$5 \quad m_1(t_f) = f_1 \Delta t e^{-\frac{f_1}{V} \Delta t}. \quad (7)$$

Of the air that entered during the preceding minute t_{1-1} , the amount of air remaining at time t_f is

$$m_{1-1}(t_f) = f_{1-1} \Delta t e^{-\frac{\Delta t}{V} f_{1-1}} e^{-\frac{\Delta t}{V} f_1} \quad (8)$$

and more generally,

$$10 \quad m_{1-n}(t_f) = f_{1-n} \Delta t e^{-\frac{\Delta t}{V} \sum f_{1-i}}. \quad (9)$$

That is, the amount of air remaining from a given time step at the end of the hour is a function of the flow rate during that time step and of the flow rate in all the following time steps.

To achieve a perfectly linear weighting function using a 15 l volume, we would need to adjust the flow rate from an initial rate of more than 100 000 SPLM to less than 10 SPLM at the end of the hour, which is not feasible with available equipment. Nor is it desirable, since pressures throughout the system would change dramatically, especially if the change in flow rate occurred throughout the entire length of sample tubing. We use a flow controller that adjusts between 100 and 1000 SPLM, and by opening the bypass valve (Fig. 1), we can obtain a higher flow rate of 3800 SPLM. In future upgrades to the system, we plan to use a 200–2000 SPLM flow controller, and note that although flow controllers with larger dynamic range are available, the control is typically accurate to within only a few percent, so that accurate control at the low end

An integrated flask sample collection system

J. Turnbull et al.

Title Page	
Abstract	Introduction
Conclusions	References
Tables	Figures
⏪	⏩
◀	▶
Back	Close
Full Screen / Esc	
Printer-friendly Version	
Interactive Discussion	



**An integrated flask
sample collection
system**J. Turnbull et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

of the range becomes problematic. We aim for a near-linear weighting function, changing the flow rate every 12 min, with the initial flow rate at 3800 SPLM, dropping down to 1000 SPLM, 550 SPLM, 380 SPLM and finally 290 SPLM in the last 12 min. This weighting function is shown as the red saw-tooth curve in Fig. 3. Each 12-min segment has equal weight (20 %), and air from prior to the start of the hour contributes 0.7 % to the total. The weighting of individual minutes during the hour ranges from 0.2–3.8 % with the largest variability in the weighting during the first 12 min of the hour.

The simple pumping system and long (typically 100 m) inlet line means that when the flow rate changes, the lag time for air reaching the mixing chamber from the top of the inlet also changes. Air reaching the mixing chamber near the end of the hour has a residence time of 15 min. Minimizing the residence time is thought to reduce the potential for wall effects, so future versions of the system will be built with modifications to keep the flow in the inlet line constant. This will be achieved by installing a constant flow rate flush pump upstream of the mixing chamber and venting the excess air. In the current configuration, we account for the lag time in the weighting function calculation.

We tested the integration parameters and weighting function in the laboratory, using the same setup described in Sect. 2.1.1. In this experiment, the flow rate was varied according to our protocol, stepping down the flow rate every 12 min of inlet time. The air from a natural air cylinder with 363.5 ppm CO₂ (measured on the Picarro instrument) was flushed through the system at 3800 SPLM for 20 min prior to the start of the experiment. The incoming air was switched between the natural and zero air cylinders at several arbitrary times, shown as grey bars in Fig. 4. We then used our theoretical weighting function (Table 1, Fig. 3) to predict the CO₂ mixing ratio at the end of our sampling time, and compared with the measured CO₂ mixing ratios. The measured value on the Picarro instrument at the end of 60 min was 337.92 ppm; the expected value was 337.98 ppm.

2.3 Drying of air

The air from the IC is eventually collected and pressurized to 40 psia in 0.7l PFP flasks (Sect. 2.4). As pressurization may exacerbate problems with water vapor, the air is dried prior to entering the mixing chamber (Fig. 1). A thermo-electric chiller with two trapping loops and a temperature of 5 °C (M&C products ECP-20-2) is used. The air flows through both loops and a peristaltic pump removes condensed water. Sufficient water is removed to result in a maximum specific humidity of 0.0014 mole H₂O per mole of dry air in the pressurized flasks. During high humidity periods, occasional problems with the chiller icing up and blocking air flow have been encountered.

The IC was tested for any change in CO₂ mixing ratio during drying. Using the same setup described in Sect. 2.1.1 (including the full integrating system), natural cylinder air was flowed through and measured using the Picarro cavity ring down spectrometer. For this test, we added a magnesium perchlorate (Mg(ClO₄)₂) trap just prior to the Picarro inlet so that no water correction was needed in the instrument's CO₂ mixing ratio determination. Although it is possible that the Mg(ClO₄)₂ trap may introduce artifacts, any such bias is expected to be consistent across all the tests. Four tests were performed: (1) dry natural cylinder air was flowed through the IC with the thermo-electric chiller completely removed; (2) dry natural cylinder air was flowed through the IC with the chiller included; (3) natural cylinder air was wetted by bubbling through a flask of distilled water and flowed through the IC with the chiller in place but not switched on; (4) same as (3) but with the chiller switched on. In these tests, the air was flowed at a constant rate of either the high 3800 SPLM or low 290 SPLM flow rate, and the CO₂ mixing ratio entering the Picarro was monitored through time. The CO₂ mixing ratio in all these tests varied by less than 0.06 ppm.

2.4 Simultaneous filling of multiple flasks

In order to provide enough air for measurement of ~50 trace gases and isotopes including ¹⁴CO₂, two 0.7l PFP flasks filled to 40 psia are collected as a pair. The PFP

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



firmware was initially designed to fill a single flask at a time, and in many cases, the paired flasks for $^{14}\text{CO}_2$ have been collected sequentially (Turnbull et al., 2006; Miller et al., 2012), assuming that variability in the air during the few minutes between flask fills is negligible relative to the $^{14}\text{CO}_2$ measurement uncertainty. However, the IC allows the opportunity to override the standard flask filling firmware, and simultaneously flush and fill two flasks. During the one-hour integration period, both flasks are opened, and the air exiting the IC is flushed through the flasks and out into the room. Once the one-hour integration period is complete, the back valves on the flasks are closed and the pumps pressurize both flasks to 40 psia. The IC remains open to the inlet and at the last flow rate of 290 SPLM during the pressurization so that the air collected in the flasks actually contains a small amount of air (1 % of the total) from the 61st minute of the interval. This is necessary with the pumps currently used, as they are unable to pressurize the flasks sufficiently when the pressure inside the 15L spherical volume drops too low. The impact of this on the final mixture is negligible across a reasonable range of mixing ratios.

For most samples, one flask of the pair is reserved for $^{14}\text{CO}_2$ measurement, and the second flask is measured for all other species. However, for a small subset of 12 samples where $^{14}\text{CO}_2$ measurement was not needed, both flasks of the pair were measured for all other species (Fig. 5). Pair agreement is excellent for all species measured and is as good as or better than the acceptance criteria used for the NOAA/ESRL Cooperative Sampling Network (Montzka et al., 2007; Zhao and Tans, 2006; Conway et al., 1994; Komhyr et al., 1985; B. R. Miller, personal communication, 2012). We calculate the mean pair difference as the mean of the absolute values of the pair differences, and the uncertainty as the one-sigma scatter of those pair differences. For CO_2 , the mean pair difference is 0.07 ± 0.08 ppm, versus the acceptable pair difference cutoff of 0.1 ppm; mean CO pair difference is 0.5 ± 0.4 ppb versus acceptable pair difference of 3 ppb, and mean CH_4 pair difference is 1 ± 1 ppb versus the acceptable pair difference of 3 ppb.

An integrated flask sample collection system

J. Turnbull et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

2.5 Remote flask triggering

The IC is controlled by a CR1000 Campbell Scientific Datalogger and Loggernet software (Utah, USA), which control sampling times, changes in flow rate, and other sampling parameters. The datalogger is remotely operated and monitored via cellular modem. In the current configuration, the software is set to automatically sample beginning at 03:00 p.m. local standard time (when the boundary layer is well mixed and most easily modeled) each day. Flasks are over-sampled the next day if a post-sampling analysis of meteorological conditions does not meet pre-defined criteria. Over-sampling involves expunging the existing sample and flushing and filling with the new sample as usual. We examined the agreement between in situ and flask measurements for flasks that were oversampled and those that were not oversampled, and found no discernable difference. One limitation of the oversampling strategy is that if a flask flushing problem occurred, it would be difficult to identify the problem (versus a more typical strategy whereby residual fill gas in the flasks can be identified by its abnormal mixing ratios; Conway et al., 2011).

2.6 Field testing

Two IC units were deployed near Indianapolis, USA on cell phone towers, alongside cavity ring-down spectrometers that continuously measure CO₂ in situ. The two systems are designed to complement each other, with the continuous measurements providing high time resolution of CO₂ mixing ratios, and the flasks providing measurement of 50 species, many of which cannot be measured continuously or in situ. In addition, comparison between the two techniques provides a quality check of both instruments, and ensures that instruments at different sites are on the same scale. The in situ measurements utilize Picarro G2401 cavity ring-down spectrometers measuring CO₂, CO, CH₄ and H₂O, with a configuration to that described by Richardson et al. (2012), except that the air is also dried prior to measurement. The in situ and flask systems sample air from separate inlet lines located immediately adjacent to one another at the same tower height. Both flask and in situ data are corrected for the time offset of air flowing

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



from the inlet to the instrument, with different time offsets for the flask and in situ instruments. We examine all available measurements where flask samples were collected and the in situ instrument collected CO₂ data during the entire hour of integrated flask sampling. A total of 50 paired measurements were collected between April 2011 and January 2012 at two different tower sites, one urban and one rural.

First, we examine a randomly selected subset ($n = 12$, selected as it is too time-consuming to look at all results in this way) of samples in detail, comparing with the in situ Picarro CO₂ measurements made every ~ 2 s. For each of these samples, we calculate the predicted flask CO₂ mixing ratio from the in situ data, first explicitly using our weighting function, and second using a simple mean of all in situ measurements made during the sampling time. The weighted and simple means give similar results across a range observed of CO₂ mixing ratios and across a range of CO₂ variability during the sampling time (Fig. 6). The mean difference between predicted and measured flask CO₂ mixing ratio (calculated as predicted minus flask value) is 0.04 ± 0.26 ppm when the weighting function is used, and 0.02 ± 0.20 ppm when the simple mean is used. We conclude that the simple mean of in situ measurements is sufficient to approximate the weighting function.

We then examine all 50 paired flask/in situ measurements, using the simple mean of the in situ measurements. The comparison is made for all available flask samples when the side-by-side in situ system made measurements for the full hour during which the flask sample was collected. The mean difference is 0.04 ± 0.38 ppm (Fig. 7). The largest differences occur during times when the CO₂ variability within the hour is high. Of particular note are several samples collected during summer 2011 where the in situ – flask difference is greater than 1 ppm. These samples were collected at the rural site during periods of strong photosynthetic drawdown and with high variability in the measured in situ CO₂ mixing ratio during the hour. In many other instances of high variability, the in situ – flask agreement was excellent, and suggests that the large difference for these samples indicates a problem with either the flask or in situ measurement, which has not yet been resolved.

An integrated flask sample collection system

J. Turnbull et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

Acknowledgements. Flask measurements and field logistics were performed by NOAA/ESRL/CCGG. Funding for this project was provided by the National Institute of Standards and Technology Grant number 60NANB10D023.

References

- 5 Brand, W. A.: GAW Report No. 194, 15th WMO/IAEA meeting of experts on carbon dioxide, other greenhouse gases and related tracers measurement techniques, Jena, Germany 7–10 September 2009, Rep. 1553, 51 pp., World Meteorological Organization, 2009.
- Conway, T. J., Tans, P. P., Waterman, L., Thoning, K. W., Kitzis, D., Masarie, K. A., and Zhang, N.: Evidence for interannual variability of the carbon cycle from the National Oceanic and Atmospheric Administration/Climate Monitoring and Diagnostics Laboratory global air sampling network, *J. Geophys. Res.*, 99, 22831–22855, 1994.
- 10 Conway, T. J., Lang, P. M., and Masarie, K. A.: Atmospheric carbon dioxide dry air mole fractions from the NOAA/ESRL Carbon Cycle Global Cooperative Network, 1968–2010, version 2011-06-21, Path: available at: ftp://ftp.cmdl.noaa.gov/ccg/co2/flask/event/ (last access: December 2011), 2011.
- 15 Crosson, E. R.: A cavity ring-down analyzer for measuring atmospheric levels of methane, carbon dioxide, and water vapor, *Appl. Phys. B-Lasers O.*, 92, 403–408, 2008.
- Komhyr, W. D., Gammon, R., Harris, T., Waterman, L., Conway, T. J., Taylor, W., and Thoning, K. W.: Global atmospheric CO distribution and variations from 1968–1982 NOAA/GMCC CO flask sample data, *J. Geophys. Res.*, 90, 5567–5596, 1985.
- 20 Miller, J. B., Lehman, S. J., Montzka, S. A., Sweeney, C., Miller, B. R., Wolak, C., Dlugokencky, E. J., Southon, J. R., Turnbull, J. C., and Tans, P. P.: Linking emissions of fossil fuel CO₂ and other anthropogenic trace gases using atmospheric ¹⁴CO₂, *J. Geophys. Res.*, 117, D08302, doi:10.1029/2011JD017048, 2012.
- 25 Montzka, S. A., Myers, R. C., Butler, J. H., Elkins, J. W., and Cummings, S.: Global tropospheric distribution and calibration scale of HCFC-22, *Geophys. Res. Lett.*, 20, 703–706, 1993.
- Richardson, S., Miles, J. N. L., Davis, K. J., Crosson, E. R., Rella, C. W., and Andrews, A. E.: Field testing of cavity ring-down spectroscopic analyzers measuring carbon dioxide and water vapor, *J. Atmos. Ocean.-Tech.*, 29, 397–406, 2012.

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**An integrated flask
sample collection
system**

J. Turnbull et al.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[⏪](#)[⏩](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

- Sweeney, C., Karion, A., Wolter, S., Neff, D., Higgs, J. A., Heller, M., Guenther, D., Miller, B. R., Montzka, S. A., Miller, J. B., Conway, T. J., Dlugokencky, E. J., Novelli, P. C., Masarie, K. A., Oltmans, S., and Tans, P. P.: Three dimensional distribution of CO₂ across North America from the NOAA/ESRL greenhouse gas aircraft network, in preparation, 2012.
- 5 Turnbull, J. C., Miller, J. B., Lehman, S. J., Tans, P. P., Sparks, R. J., and Southon, J. R.: Comparison of ¹⁴CO₂, CO and SF₆ as tracers for determination of recently added fossil fuel CO₂ in the atmosphere and implications for biological CO₂ exchange, *Geophys. Res. Lett.*, 33, L01817, doi:10.1029/2005GL024213, 2006.
- 10 Zhao, C. and Tans, P. P.: Estimating uncertainty of the WMO mole fraction scale for carbon dioxide in air, *J. Geophys. Res.*, 111, D08S09, doi:10.1029/2005JD006003, 2006.

An integrated flask sample collection system

J. Turnbull et al.




Fig. 1. Integrating flask sampler schematic.

Title Page

Abstract Introduction

Conclusions References

Tables Figures

◀ ▶

◀ ▶

Back Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**An integrated flask
sample collection
system**

J. Turnbull et al.




Fig. 2. Measured CO₂ mixing ratio through time, starting with “ambient” 378.95 ppm CO₂ air inside integrating volume. Air entering the volume was switched to “zero” 0 ppm CO₂ air at time zero. Flow rate was set at 3800 SPLM.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

An integrated flask sample collection system

J. Turnbull et al.




Fig. 3. Weighting function for air inside the 15 l integrating volume after 60 min, under different flow scenarios. Black line: ideal linear weighting function where each timestep contributes an equal amount to the final sample (1.67% from each minute). Solid and dashed grey lines: constant flow at 1000 and 4000 SPLM, respectively. Red line: the weighting function we selected changing the flow rate every 12 min from 3800 SPLM to 1000 SPLM to 550 SPLM to 380 SPLM to 290 SPLM. The value at minute zero for each curve is the sum total of all air remaining in the flask from times prior to the last 60 min.

[Title Page](#)
[Abstract](#)
[Introduction](#)
[Conclusions](#)
[References](#)
[Tables](#)
[Figures](#)
[⏪](#)
[⏩](#)
[◀](#)
[▶](#)
[Back](#)
[Close](#)
[Full Screen / Esc](#)
[Printer-friendly Version](#)
[Interactive Discussion](#)





Fig. 4. Test of weighting function. Cylinder air at room pressure was flowed through the integrating system at the indicated flow rates (blue). Natural 363.5 ppm CO₂ air was flowed through the system for 20 min prior to the start of the experiment and during the experiment, except during the periods shaded in grey, when incoming air was switched to zero air. The measured CO₂ mole fraction exiting the system is indicated in black, and was 337.92 ppm at the end of the integrating time (60 min). Predicted value was 337.98 ppm.

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



An integrated flask sample collection system

J. Turnbull et al.




Fig. 5. Mean flask pair differences, typically averaged over 12 flask pairs. Measured mean pair differences for each species are shown in red, and the acceptable maximum pair difference is shown in green. Vertical axis uses a log scale for clarity, units vary by species. CO₂ is reported in ppm, CO, CH₄, N₂O and H₂ are in ppb, δ¹³CO₂ and δ¹³CH₄ are in ‰, all other species are in reported in ppt.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion






Fig. 6. Difference between predicted and measured flask CO₂ mole fraction, plotted against the CO₂ mole fraction. Predicted mole fractions were calculated from in situ (~2 s) CO₂ measurements. Red points use the full weighting function to calculate the predicted CO₂ mole fraction; blue points use a simple average of the in situ values during the flask sampling time. Error bars on the blue points indicate the standard deviation of the in situ measurements during the sampling time.




Fig. 7. Difference between in situ mean and flask CO₂ mole fraction for 50 flasks sampled at two sites near Indianapolis. Error bars are the standard deviation of the in situ values during the sampling time. Mean difference is 0.04 ppm.

An integrated flask sample collection system

J. Turnbull et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

