

Interactive comment on “Atmospheric column CO₂ measurement from a new automatic ground-based sun photometer in Beijing from 2010 to 2012” by Z. Q. Li et al.

Anonymous Referee #2

Received and published: 4 January 2013

This paper treats a topic of high relevance: the development of novel ground-based instrumentation for measuring atmospheric XCO₂. Simple, robust, and affordable instrumentation is required to achieve a significantly improved global coverage of ground-based XCO₂ records. The Total Carbon Column Observing Network (TCCON), which currently provides the reference XCO₂ dataset applies high-resolution FTIR spectrometers. The TCCON dataset is of utmost importance for satellite validation and inversion studies. However, the approach applied by the TCCON stations, as Li et al. correctly state, is not a promising candidate for a future global XCO₂ observation network with a considerably better sampling density. The spectrometers used by TCCON are expen-

C3479

sive, immobile, require regular qualified on-site maintenance and rely on a considerable logistic infrastructure.

Therefore, if XCO₂ measurements of useful quality would be achievable with the simple sun photometers investigated by Li et al., this would be a remarkable achievement. Unfortunately, the current version of the paper fails short to provide a reasonable proof for this claim. The authors do not make any serious attempt to derive a useful XCO₂ product from their raw spectral data, instead, they simply discuss a “difference absorption index” (DAI), without establishing a connection of this quantity with a meaningful physical result. In my opinion, the correction of the airmass dependency would be a minimum requirement to enter into a sensible discussion of diurnal or seasonal variations and a model comparison.

Error sources and cross sensitivities which are likely inherent to the method presented are all covered by the huge airmass-induced variation of DAI (e.g. H₂O column, ground-pressure, temperature profile, instrumental problems, e.g. spectral shift of filter response as fct of ambient temperature). However, to rank the sun photometer among the serious candidates for a network method, an XCO₂ precision in the order of < 0.25% needs to be demonstrated.

No attempt is made to relate the own investigations to relevant prior work. It is not clear to me whether the authors are aware of e.g. the following contributions:

Kobayashi et al.: "Remotely operable compact instruments for measuring atmospheric CO₂ and CH₄ column densities at surface monitoring sites", AMT, 2010

Petri et al.: "Remote sensing of CO₂ and CH₄ using solar absorption spectrometry with a low resolution spectrometer", AMT, 2012

Gisi et al.: "XCO₂-measurements with a tabletop FTS using solar absorption spectroscopy", AMT, 2012

In my opinion, in contrast to earlier studies, it becomes pretty clear that sections 3.2.2

C3480

and 4 of the paper do not meet the state of the art. Therefore, I cannot recommend publication of the current version of the paper in AMT.

Major revisions required:

- 1) As pointed out by anonymous referee #1, the use of English needs significant revision.
- 2) Insert a new section following 3.2.2: explain how to further process a DAI to achieve an estimate of the CO₂ column.
- 3) Insert a new section concerning error estimation and internal consistency before section 4. A few items: how consistent are CO₂ columns derived from different combination of channels (different DAIs)? Do the CO₂ columns behave as expected (the CO₂ column should correlate with ground pressure)? In the next step, calculate the XCO₂ (using ground pressure and H₂O column): is there an apparent unphysical correlation between XCO₂ and H₂O column or airmass?
- 4) In section 4, discuss the XCO₂ results indicated by your measurements. It is improper to interpret DAI values as if they were XCO₂ values.

Interactive comment on Atmos. Meas. Tech. Discuss., 5, 8313, 2012.