

Interactive comment on “Flask sample measurements for CO₂, CH₄ and CO using cavity ring-down spectrometry” by J.-L. Wang et al.

J.-L. Wang et al.

joechang@rcec.sinica.edu.tw

Received and published: 18 November 2013

We apologize that the water correction coefficients in the previous response to referee #1 need further corrections. The correct coefficients are as follows: a= -0.012000, b= -0.0002674 for CO₂ and a= -0.0098200, b= -0.0002393 for CH₄.

Interactive comment on Atmos. Meas. Tech. Discuss., 6, 7633, 2013.

C3343