

This discussion paper is/has been under review for the journal Atmospheric Measurement Techniques (AMT). Please refer to the corresponding final paper in AMT if available.

The AquaVIT-1 intercomparison of atmospheric water vapor measurement techniques

D. W. Fahey^{1,2}, R.-S. Gao¹, O. Möhler³, H. Saathoff³, C. Schiller^{4,†}, V. Ebert^{5,6,7}, M. Krämer⁴, T. Peter⁸, N. Amarouche⁹, L. M. Avallone^{10,*}, R. Bauer⁴, Z. Bozóki¹¹, L. E. Christensen¹², S. M. Davis^{1,2}, G. Durrý¹³, C. Dyroff¹⁴, R. L. Herman¹², S. Hunsmann⁷, S. M. Khaykin^{15,***}, P. Mackrodt⁵, J. Meyer⁴, J. B. Smith¹⁶, N. Spelten⁴, R. F. Troy¹², H. Vömel^{1,2,**}, S. Wagner^{4,6}, and F. G. Wienhold⁸

¹National Oceanic and Atmospheric Administration, Earth System Research Laboratory, Boulder, CO, USA

²Cooperative Institute for Research in Environmental Sciences, University of Colorado, Boulder, CO, USA

³Karlsruhe Institute of Technology, Institute for Meteorology and Climate Research, Atmospheric Aerosol Research (IMK-AAF), Karlsruhe, Germany

⁴Institute for Energy and Climate Research, Stratosphere (IEK-7), Forschungszentrum Jülich, Jülich, Germany

⁵University of Heidelberg, Physikalisch-Chemisches Institut (PCI), Germany

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



⁶Physikalisch-Technische Bundesanstalt (PTB, National Metrology Institute of Germany), Bundesallee 100, Braunschweig, Germany

⁷Technical University Darmstadt, Center of Smart Interfaces (CSI), Darmstadt, Germany

⁸Institute for Atmospheric and Climate Science, ETH Zurich, 8092 Zurich, Switzerland

⁹Division Technique de l'Institut National des Sciences de l'Univers, UPS 855 CNRS, Meudon, France

¹⁰Department of Atmospheric and Oceanic Sciences, University of Colorado, Boulder, CO USA

¹¹MTA-SZTE Research Group on Photoacoustic Spectroscopy, University of Szeged, Szeged, Hungary

¹²Jet Propulsion Laboratory, California Institute of Technology, Pasadena, CA, USA

¹³Groupe de Spectrométrie Moléculaire et Atmosphérique, UMR CNRS 7331, Université de Reims-Champagne-Ardenne, Reims, France

¹⁴Karlsruhe Institute of Technology (KIT), Institute for Meteorology and Climate Research, Atmospheric Trace Gases and Remote Sensing (IMK-ASF), Karlsruhe, Germany

¹⁵Central Aerological Observatory, Moscow, Russia

¹⁶School of Engineering and Applied Sciences, Harvard University, Cambridge, MA, USA

* now at: the National Science Foundation, Washington DC

** now at: Meteorologisches Observatorium Lindenberg, Lindenberg, Germany

*** now at: CNRS/INSU, LATMOS, IPSL, Université de Versailles St. Quentin,

Guyancourt, France

†deceased

Received: 5 January 2014 – Accepted: 6 February 2014 – Published: 1 April 2014

Correspondence to: D. W. Fahey (david.w.fahey@noaa.gov)

Published by Copernicus Publications on behalf of the European Geosciences Union.

Abstract

The AquaVIT-1 Intercomparison of Atmospheric Water Vapor Measurement Techniques was conducted at the aerosol and cloud simulation chamber AIDA at the Karlsruhe Institute of Technology, Germany, in October 2007. The overall objective was to intercompare state-of-the-art and prototype atmospheric hygrometers with each other and with independent humidity standards under controlled conditions. This activity was conducted as a blind intercomparison with coordination by selected referees. The effort was motivated by persistent discrepancies found in atmospheric measurements involving multiple instruments operating on research aircraft and balloon platforms, particularly in the upper troposphere and lower stratosphere where water vapor reaches its lowest atmospheric values (less than 10 ppm). With the AIDA chamber volume of 84 m³, multiple instruments analyzed air with a common water vapor mixing ratio, either by extracting air into instrument flow systems, locating instruments inside the chamber, or sampling the chamber volume optically. The intercomparison was successfully conducted over 10 days during which pressure, temperature, and mixing ratio were systematically varied (50 to 500 hPa, 185 to 243 K, and 0.3 to 152 ppm). In the absence of an accepted reference instrument, the reference value was taken to be the ensemble mean of a core subset of the measurements. For these core instruments, the agreement between 10 and 150 ppm of water vapor is considered good with variation about the reference value of about $\pm 10\%$ ($\pm 1\sigma$). In the region of most interest between 1 and 10 ppm, the core subset agreement is fair with variation about the reference value of $\pm 20\%$ ($\pm 1\sigma$). The upper limit of precision was also derived for each instrument from the reported data. These results indicate that the core instruments, in general, have intrinsic skill to determine unknown water vapor mixing ratios with an accuracy of at least $\pm 20\%$. The implication for atmospheric measurements is that the substantially larger differences observed during in-flight intercomparisons stem from other factors associated with the moving platforms or the non-laboratory environment. The success of AquaVIT-1 provides a template for future intercomparison efforts with water vapor

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



based on our understanding of the fundamental microphysics of ice formation. In contrast, in other aircraft and balloon measurements, liquid water supersaturation was not observed (Krämer et al., 2009).

The Intercomparison Campaign of Water Vapor Measurement Techniques (AquaVIT-1) was undertaken at the AIDA chamber at the Karlsruhe Institute of Technology (KIT) as an effort to clarify uncertainties in UT/LS water vapor measurements and help identify the cause(s) of the discrepancies such as instrument calibration and artifacts. The campaign did not include an evaluation of instrument sampling issues related to platform configuration, the ambient environment, or inlets, all of which can affect in-flight performance and inherently are more difficult to evaluate; nor an evaluation of instrument calibration procedures.

AquaVIT-1 was a controlled, refereed, and blind intercomparison of a large group of water vapor instruments using the AIDA chamber. The ranges of pressure, temperature and water vapor in the chamber included those found in the tropical UT/LS. The principal objective of AquaVIT-1 was to compare instruments in a controlled ground-based facility with the expectation that systematic measurement problems and perhaps their causes could be identified more readily and with less expense and effort than in airborne campaigns. In addition, the campaign included instruments that were relatively new to atmospheric studies or still under development in order to accelerate their progress in becoming reliable and accurate instruments for use in future field measurement campaigns.

The AquaVIT-1 experiments were initially planned as part of a SPARC workshop on Upper Tropospheric Humidity in June 2007 (Peter et al., 2008) and subsequently occurred in two one-week phases in October 2007 in Karlsruhe, Germany. The first phase was devoted to static intercomparisons with a separate experiment each day (15–19 October) at near-constant temperature conditions and several fixed pressure levels. The second phase was a week of dynamic intercomparisons, with several experiments each day (22–26 October) under varying pressure, temperature and humidity

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



conditions and with or without ice clouds present. In this paper, only the static experiments and their results are described.

2 AIDA chamber

The Aerosol Interaction and Dynamics in the Atmosphere (AIDA) chamber is located at KIT. The chamber is an aluminum vessel of volume 84 m^3 with facility to control pressure from one atmosphere to as low as 0.01 hPa and temperature from 313 K to as low as 183 K (Möhler et al., 2003). This range of conditions allows for simulating atmospheric aerosol and cloud formation processes under tropospheric and lower stratospheric conditions on short (minutes) to long (hours) time scales. Important features of the AIDA chamber for AquaVIT-1 were, first, the operation of the chamber allowed for conditions of near-constant pressure (± 1 hPa) and temperature (± 0.3 K). Under these conditions, water vapor amounts were constant or slowly changing for periods of many minutes, thereby allowing adequate time for all instruments to sample chamber air and make multiple determinations of water vapor content. Second, the water vapor mixing ratio and humidity could be altered by the addition of water vapor or dry air, or the partial removal of chamber air by pumping. Third, the large chamber volume has a small wall-to-volume ratio and allows multiple instruments to be located inside the chamber or to sample air from outside the chamber without significantly disturbing internal conditions. To this end, customized, extractive sampling probes were implemented for AquaVIT-1 to deliver chamber air to instruments located outside the chamber. The probes were made of stainless steel and heated to avoid water adsorption on the probe inner walls at low chamber temperatures.

3 Data protocol

All investigators signed the data protocol adopted for AquaVIT-1. The protocol encouraged rapid assessment and use of the results from the AquaVIT-1 tests while upholding

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



the rights of the individual scientists and treating all participants equitably. Key features of the protocol are:

- Quick-look data. Preliminary or quick-look data obtained during the AquaVIT-1 campaign were made available to the referees as soon as possible following each day's experiments (< 24 h). In the event of obvious difficulties, this allowed the referees to suggest corrections or amendments to data processing, instrument configuration, or instrument operation be made as soon as possible, thereby improving the overall outcome of the intercomparison. All water vapor mixing ratio values were archived as molar fractions in units of ppm (i.e., $\mu\text{mol mol}^{-1}$) unless otherwise noted.
- Blind intercomparison. A blind intercomparison was established so that preliminary data submitted during the campaign (typically within 48 h) and the short evaluation period immediately following the campaign were available only to the referees (O. Möhler, D. W. Fahey, and R. S. Gao) who were not affiliated with any participating instrument team.
- Final data. After the end of the short evaluation period (4 December 2007) the submitted datasets were released to all participants. Any further changes to a submitted dataset required documentation from an instrument's investigator and approval by the referees. All datasets were considered final on 10 January 2008.

A dedicated webpage with password protection enabled archiving and interchange of datasets among the participants and access to other AquaVIT-1 documents and information.

4 Instruments

AquaVIT-1 included more than 20 instruments utilizing either state-of-the-art or newly developed techniques and, thereby, represented a large fraction of the international

AMTD

7, 3159–3251, 2014

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



UT/LS water-vapor community. A subset of 15 instruments divided into core and non-core groups in Table 1 is the focus of this paper. This subset participated in the formal, blind, refereed intercomparison. The remaining instruments did not participate in the formal intercomparison or acquired insufficient science-quality data for analysis.

The core instrument subgroup includes APicT, CFH, FISH-1, FISH-2, FLASH-B1, FLASH-B2, HWV, JLH (see Appendix A). The APicT, as an AIDA facility instrument, has been involved in many AIDA chamber experiments. The other instruments have a long history of field measurements and intercomparisons on balloon and aircraft platforms operating in the upper troposphere and stratosphere. The mixing ratio discrepancies noted at low values in these regions derive from a number of datasets from these instruments. Establishing the accuracy of the core instruments under controlled laboratory conditions was one primary objective of AquaVIT-1. The reported uncertainty for each of the core instruments is listed in Table 2.

The non-core instrument subgroup includes MBW-373LX, SnowWhite, ISOWAT, OJSTER, PicoSDLA, WaSul-Hygro2 and CLH. This group includes mature instruments that have also been used in field measurements as well as instruments that were in the initial to later stages of development.

Following extensive discussion among the organizers and investigators, the decision was made that no AquaVIT-1 instrument would serve as the reference instrument. No participating instrument had played this role previously for a multi-instrument intercomparison in a chamber configuration and no sufficiently objective and analytical basis could be developed by the group prior to or following the campaign to determine which instrument(s) would qualify as a reference. See Sect. 6.5 for further discussion of reference choices.

The intercomparison results are expressed in units of water-vapor molar mixing ratio. The instruments suite measures different aspects of water vapor. For example, while Lyman- α instruments measure mixing ratio (HWV and FISH), the frostpoint hygrometer (CFH) measures partial pressure and TDL instruments measure column number density (JLH and APicT). Mixing ratio is preferred because it is easily derived from all

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



measurement techniques and is conserved in the atmosphere and within extractive instrument systems (in the absence of the condensed phase). As a consequence, mixing ratio is a customary unit found in scientific interpretation and modeling studies.

Brief descriptions of the intercompared instruments, their configuration in the AIDA chamber, their performance during the static experiments and lessons learned from AquaVIT-1 are included in Appendix A. Fundamental differences exist between the laboratory and flight environments (e.g., ambient pressure, temperature, flow rates, solar environment, and inlet sampling strategy) that generally represent some limit in the relevance of the laboratory results to instrument operation on moving platforms under ambient conditions. The nature of these differences and their impact on the measurements is different for each instrument. All flight instruments were modified or reconfigured to adapt to the laboratory setting, and in some cases, sources of error or uncertainty associated with the laboratory implementation degraded instrument performance. While the AIDA configuration is not ideal in this regard, the value of a systematic laboratory intercomparison was clear in advance of AquaVIT-1 given the large discrepancies found in field observations.

5 AIDA chamber instrument configuration

The overall configuration of instruments in the AIDA chamber facility is shown in Fig. 2. The instrument sampling techniques used in AquaVIT-1 can be classified into two distinct types. The first type is extractive sampling, which requires gas to be removed through a probe located inside the chamber that connects to a sample line that passes through the chamber and/or thermal enclosure walls. Most of the core instruments used extractive sampling with heated probes (Fig. 3). Three instruments were located outside the chamber but inside the chamber thermal enclosure.

The second type is internal or non-extractive sampling. Three core instruments and one non-core instrument used non-extractive sampling. APicT and JLH used open-path optical absorption spectroscopy inside the chamber (Fig. 3). FLASH-B1 used

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



5 a chamber window to measure Lyman- α fluorescence produced directly inside the chamber walls. Principal optical components of APicT and FLASH-B1 were mounted outside the chamber but inside the thermal enclosure. The JLH laser, detector, and open-path mirrors were mounted entirely inside the chamber (Fig. 3), as was the
10 SnowWhite instrument sensor. For these two instruments, the associated control and data recording electronics remained outside the thermal enclosure. The open-path white cell used by APicT was mounted completely inside the AIDA chamber. By folding its optical path between the inner chamber walls, APicT was the only instrument that provided a measurement of the water-vapor abundance averaged over the full diameter of the chamber.

6 AquaVIT-1 static experiments

6.1 AIDA chamber conditions

15 The chamber pressures and temperatures for the static experiments are shown in Fig. 4. Also shown are the nominal domains from approximately a decade of HWV observations in the tropics and midlatitudes (Weinstock et al., 2009) that show substantial overlap with the AquaVIT-1 experimental domains. In the experiments, temperature was reduced in daily steps from ~ 240 K on the first day to ~ 185 K on the last day. The transient temperature changes in Fig. 4a are the quasi-adiabatic responses to the pressure changes in Fig. 4b that occur while the chamber walls and enclosure remain at a constant temperature. The reported value was stable to ± 0.2 K (1σ) during the measurement segments. Transient temperature excursions in the time series of up to 10 K are the adiabatic responses to the occasional rapid addition or removal of air from the chamber. Chamber pressure was held constant to ± 1 hPa in 0.5–1 h intervals or segments (between 5 and 8) during each day. Water-vapor mixing ratio values varied
20 depending on the amount of water added directly to the chamber at the beginning of each day's experiment and on the subsequent changes in chamber conditions.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



at saturation values. Non-conservation was also observed to a lesser extent in the ice-free static experiments due to the adsorption and desorption of water on the walls and other chamber internal components. Examples of this are shown by the time series in Fig. 5 in which the water vapor mixing ratio increases when the chamber pressure is reduced in the second half of the day's experiment. If there are no other sources of water vapor then the mixing ratio should remain constant as air is pumped from the chamber. However, non-conservation of water vapor does not interfere fundamentally with the AquaVIT-1 results because the water vapor mixing ratios always changed slowly with time within an intercomparison segment and the stirring fan forced complete internal mixing within about 1 min (see discussion below).

6.2 Data processing

For each day of the static experiment series, the instrument teams submitted a data file reporting water vapor mixing ratios vs. UT time. The measurement interval for most instruments was 1 s. As an example, the 1 s datasets for 15 October are plotted in Fig. 5.

The data processing steps taken for the combined dataset were the following:

- Define segments. The time series were divided into constant pressure and near constant temperature segments for statistical analysis. Not all segments were used in the intercomparison analysis. The criteria for selecting a segment were near-constant or slowly and linearly varying water vapor mixing ratios within the segment and the availability of water vapor data for the segment from a majority of the core instruments. The first criterion ensured uniform mixing ratio conditions in the chamber and, hence, each sample line. Data gaps between segments generally correspond to rapidly changing chamber pressure (Fig. 4). Constant or linearly varying water vapor values in a segment were also required for the precision analysis described below. Segment lengths in the range 900–3600 s were chosen to provide good statistical confidence in the subsequent analysis steps.

The times and lengths of the segments used in the analysis along with average pressure, temperature, and water-vapor mixing ratios are provided in Table 3.

- b. Calculate linear fits for each instrument segment. Core-instrument linear fits to water vapor mixing ratios as a function of time were calculated from the respective time series data for each segment, thereby defining the mean values and their rates of change.
- c. Calculate reference water vapor mixing ratios. The reference water-vapor mixing ratio for each segment was obtained using only the core instrument data. The exception was that data from the FLASH instruments were excluded because of offsets discovered during the data evaluation (see Appendix). A two-step process was adopted to provide a consistent basis of comparison across and within segments. First, a single linear fit was performed on the complete set of core-instrument linear fits. This combined fit was chosen over a simple unweighted average of all core instrument data in order to give the same weight to each core instrument in deriving the reference value. This combined fit defines the reference function for the segment, which in turn defines the time evolution of values within the segment. Second, the reference water-vapor mixing ratio for the segment was defined to be the average of the reference function over the segment. These reference mixing ratios as listed in Table 3 are used throughout the analysis and plots presented here. Note that the number of instruments reporting data for each segment generally influences the reference value. Also note that it would be incorrect to use the reference values to infer the absolute accuracy of any instrument averages (see Sect. 6.5).
- d. Calculate 1 s probability distribution functions (PDFs) for each segment. For each instrument the differences between the instrument time series of 1 s mixing ratios and the reference function were used to form a PDF for each segment time series. A Gaussian function was fit to each PDF. The difference between the Gaussian mean and the reference value was assigned to be the average difference for the

**AquaVIT-1 water
vapor
intercomparison**

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



segment. These values appear for each instrument segment as separate symbols in Figs. 6 and 7.

- e. Calculate instrument precision. The standard deviation ($\pm 1\sigma$) of each Gaussian fit to differences from the reference function is defined to be the instrument precision for a segment. Values represent an upper limit since σ includes error in the reference function and non-linear variation of water vapor in the chamber over the segment period. Examples of instrument precision values are shown in Table 4 for core and non-core instruments.

6.3 Core instrument results

6.3.1 Reference value comparisons

A summary of the instrument intercomparison results for the 5 day static experiment series is shown in the plots in Figs. 6 and 7. The symbols represent the average difference within a segment from the reference water vapor value for that segment. Summary points for the core instruments are:

- 10–150 ppm H₂O: good agreement occurs in this range. Except for a few segments, all the segment values and the all-segment averages (circle/plus symbols in Fig. 6) for each instrument agree with the reference within $\pm 10\%$. The FLASH-B instrument shows the greatest segment-to-segment variability and the largest differences due to experimental setup issues described in Appendix A4. The instruments other than FLASH-B(1)/B(2) show a small segment-to-segment variability ($\sim 5\%$) indicating good instrument stability and systematic uncertainties that are constant throughout these experiments. There is some tendency of the largest differences to occur for the lowest pressure range (< 70 hPa).
- $1 \leq \text{H}_2\text{O} \leq 10$ ppm: fair agreement occurs in this range. All the segment values and the all-segment averages for each instrument agree with the reference within

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



about $\pm 20\%$. The segment-to-segment variability for each instrument is about 10% or greater indicating, in comparison to the results for 10–150 ppm H₂O, instrument stability issues and systematic uncertainties that are more important.

- $0.2 \leq \text{H}_2\text{O} \leq 1$ ppm: poor agreement occurs in this range. Fewer instruments reported data for these segments than for the other two mixing ratio ranges. In this range more so than in the two higher ranges, the reference value is influenced more strongly by the FISH results since the FISH-1 and FISH-2 data are essentially equal and fewer instruments contribute to the reference value calculation. All core instrument segment values agree with the reference and with each other within a range of about -100% to $+150\%$ with absolute differences less than 0.4 ppm. All segments have pressures > 150 hPa. Although mixing ratios in this range occur rarely in the UT/LS, these measurements help define the detection limits and performance limits of the instruments.

6.3.2 Uncertainties

The uncertainties of the core instruments as provided by the investigators are listed in Table 2 and further documented in Appendix A. Since no absolute reference value was established for AquaVIT-1, no independent conclusion can be drawn concerning whether the individual uncertainty ranges include the correct value for each segment. Since the core instrument uncertainties are derived with independent calibration procedures, there is an expectation of agreement within the combined uncertainties of the measurements. With the accuracy values given in Table 2 and data in Fig. 7, agreement can be expected within $\pm 20\%$ for the 1–150 ppm range and about $\pm 100\%$ for values < 1 ppm.

The precision upper limits calculated for the core instrument measurements for several example segments are shown in Table 4. PDFs of FISH-2 data for all segments are shown as an example in Fig. 8. Precision is not routinely reported for all instruments, in part, because precision, in contrast to accuracy, depends more on conditions of the

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



measurements and treatment of the raw data. Most of the segment PDFs show a good Gaussian fit indicating good stochastic behavior of the detection module in each instrument. The magnitude of the precision limits is typically in the range of 0.1 to 0.2 ppm, which suggests that the 1 s measurement precision is not a large component of the uncertainty in the 1–150 ppm range when averaging over segment lengths of 1800–3600 s. A comparison of precision derived from 1 s and 5 s time series measurements is also shown in Table 4. If the measurement variability is truly random, then the PDF of 5 s measurements should be a factor of $5^{0.5} = 2.2$ smaller than the 1 s precision values. Table 4 shows that this is not the case for most of the core and non-core instruments. The largest ratios of about 1.7 are found for the FISH instruments. This suggests that the measurement variability is not completely random on the 1 to 5 s timescale for most of the instruments and may contain significant contributions from instrument drift and varying conditions of instrument components. Rollins et al. (2014) further discuss precision and accuracy issues related to a suite of water vapor instruments that includes some AquaVIT-1 core instruments.

6.4 Non-core instrument results

Figure 6 includes results for the non-core instruments for the same segments. Non-core results do not influence the segment reference values as discussed above. Summary points for these results are:

- 10–150 ppm H_2O : the best overall agreement with the core reference values occurs in this range. Segment differences show a wider range than core instruments, varying from about –100 % to +200 % with most of the data falling within the –30 % to +50 % range. Instrument averages also show a wider range, varying from about –90 % to +40 %.
- $1 \leq \text{H}_2\text{O} \leq 10$ ppm: poorer agreement with the core reference values occurs in this range. Segment differences show a much wider range than core instruments,

varying from about -20% to $+1000\%$ with most instruments significantly higher than the reference value.

- $0.2 \leq \text{H}_2\text{O} \leq 1$ ppm: the poorest agreement with the core reference values occurs in this range as also found for the core instruments. Only two instruments submitted data for these low values. The results are $90\text{--}300\%$ higher than the reference value.

6.5 Absolute reference values

The ensemble average of the core instrument values was used as the reference value in the intercomparison in lieu of designating a single instrument as the reference. However, three AquaVIT-1 instruments have special merit in serving as absolute reference standards: the KIT MBW-373LX, APicT, and the German PTB water vapor permeation source (PTB-WVPS). The first and last have direct links to international, metrological primary standards. APicT evaluations show that its values are very consistent with ice saturation values expected in AIDA chamber ice-cloud measurements.

The MBW-373LX is a chilled-mirror frostpoint hygrometer from MBW Calibration Ltd. in Switzerland (<http://www.mbw.ch>) that is used regularly in the AIDA facility (Appendix A7). The unit has a frost-point accuracy of ± 0.1 K (corresponding to less than $\pm 2\%$ in water vapor mixing ratio) according to the manufacturer's calibration, which is linked to international standards. The KIT MBW-373LX unfortunately was not configured to operate with sample line pressures less than 500 hPa or sample flows below 0.2 L min^{-1} and thus could not be suitably intercompared in AquaVIT-1 for all conditions (Appendix A7). Near 1000 hPa, however, the KIT MBW hygrometer has intercompared well with the APicT in many previous AIDA experiments.

APicT values, derived from in-situ absorption profiles, are only based on H_2O spectral data (taken from the HITRAN2008 data base and from PCI/KIT reference measurements, Hunsmann et al., 2006) plus length, pressure and gas temperature measurements. Hence, APicT performance in AquaVIT-1 did not depend on a laboratory or

AMTD

7, 3159–3251, 2014

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Peter et al., 2006). In some cases, these discrepancies are large enough (50–100 %) to interfere with answering important scientific questions about water vapor in the UT/LS. An example of differences associated with core instruments is shown in Fig. 1. An important conclusion from AquaVIT-1 is that such differences as observed in the airborne field studies are significantly larger than those found for CFH and HWV in this laboratory study. However, the qualitative differences are similar, with CFH values less than HWV values. Further conclusions may follow from a systematic assessment of the field observations by instrument investigators that takes into account the AquaVIT-1 results.

Caution must be taken in using these results to infer instrument performance on moving platforms (e.g., balloons and aircraft) because AquaVIT-1 did not fully reproduce UT/LS instrument or sampling conditions for the diverse set of instruments involved, nor could it be expected to. Some instruments were modified from their flight configuration in order to adapt to the laboratory setting (see Appendix A). For those instruments mounted outside the AIDA chamber, environmental pressures and/or temperatures were generally significantly higher than typically encountered in UT/LS flights. Similarly, sample flows internal to the instruments were in some cases significantly lower than in flight and were often at higher temperatures than those encountered in the UT/LS. A closer simulation of external and internal pressures and temperatures occurred for JLH and SnowWhite because both were located inside the chamber. However, sample-volume flow rates that occur in flight for these two instruments (aircraft and balloon, respectively) were not well simulated in the chamber tests because of the limited stirring of chamber air. There are potentially other factors that influence in-flight performance in the UT/LS that did not influence the AquaVIT-1 experiments, such as rapid changes in mixing ratio, pressure and temperature. These effects will need to be carefully evaluated to make appropriate and optimal use of the results presented here.

8 Summary of the AquaVIT-1 experiments

The AquaVIT-1 results are summarized as follows:

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



- AquaVIT-1 successfully integrated a variety of instruments to measure water vapor in the AIDA chamber using extractive and non-extractive sampling methods. The scientific and technical participant group developed procedures and protocols to carry out the physical experiments and post-experiment data processing and analysis. For five days in October 2007, static experiments were conducted with chamber conditions covering a range of pressures (50–500 hPa), temperatures (185–243 K) and water vapor mixing ratios (< 1–150 ppm) in order to simulate conditions typically found in the UT/LS.
- The subset of the instruments examined here was divided into two categories: core and non-core instruments. Core instruments have been extensively used in field campaigns on moving platforms and some participated in campaigns, particularly in the tropics, that revealed large systematic discrepancies in the UT/LS.
- Time series of water vapor mixing ratios from the core and non-core instrument were divided into 31 segments of 900–3600 s duration for analysis. The intercomparison was based on reference values derived for each segment using linear fits to the core instrument data.
- The core instrument values are in fair to good agreement in the 1–150 ppm H₂O range with instrument averages over all segments agreeing within about $\pm 20\%$. For individual segments, agreement is generally found to be close or within the combined uncertainties for any instrument pair. Below 1 ppm H₂O, fewer instruments reported data and overall agreement is poorer.
- The non-core instrument group included mature instruments and newly developed prototypes. This group demonstrated generally poorer agreement with each other and with the core reference values than the core instruments.
- No statement can be made about the absolute accuracy of the instruments because no absolute reference instrument was available for the AquaVIT-1 experiments. However, each core instrument team independently and routinely

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



calibrates (or validates) their respective instrument and links it to one or more absolute reference sources. Most pairs of core instruments agree within the respective combined uncertainties over the water vapor range of 1–150 ppm. Thus, it is highly likely that the correct water vapor value for each segment is between the maximum and minimum of the segment averages of the core instruments. This corresponds to about $\pm 10\%$ of the reference value for the 1–150 ppm range. The APicT results from the dynamic experiments strengthen the conclusion that the correct water vapor values are bounded by the core instrument averages.

- The AquaVIT-1 results alone do not resolve the water vapor discrepancies observed in the atmosphere. Caution must be taken in using these results to infer instrument performance on moving platforms (e.g., balloons and aircraft) because AquaVIT-1 did not fully reproduce operating and sampling conditions in the UT/LS for the diverse set of instruments involved. All flight instruments were modified or reconfigured for chamber operation. In some cases, adaptation to laboratory environment introduced other sources of error or uncertainty. Furthermore, there are potentially other factors that influence in-flight performance that did not influence the AquaVIT-1 experiments. It is possible that these additional factors are responsible for the larger discrepancies observed in intercomparisons of some moving platform measurements.
- The success of AquaVIT-1 provides a template for future intercomparison efforts with water vapor or other species that are focused on improving the analytical quality of atmospheric measurements on moving platforms.

Appendix A

Instrument descriptions

A1 AIDA-PCI-in-cloud-TDL (APicT)

The APicT spectrometer is a near infrared, diode-laser based, fiber-coupled, high spectral resolution, absorption spectrometer with an open-path absorption cell mounted on the inner surface of the walls of the AIDA vessel. This open-path configuration avoids any gas sampling and allows a selective detection of the interstitial water vapor inside clouds and a continuous determination of the absolute water-vapor molar fraction inside the AIDA chamber. APicT was designed to investigate cloud formation within AIDA and thus can be used not only in particle-free conditions but also when clouds are formed within the APicT absorption path.

The main goal of the APicT design was – aside from high chemical selectivity and sensitivity, and sufficient temporal resolution – to achieve a highly accurate, absolute, water vapor measurement by complete avoidance of gas sampling and gas treatment. Special attention was given to providing absolute H₂O detection levels without the need for a regular calibration of the instrument response (i.e., by using water vapor generators). APicT's measurement volume is distributed equally over a central intersect through the AIDA chamber, which delivers representative, spatially averaged H₂O concentrations over the respective AIDA cross section with very little influence from wall effects or gas sampling problems.

The APicT instrument (Giesemann, 2003, Giesemann et al., 2003; Ebert, 2004, 2005, 2008) is based on the extensive tunable diode laser absorption spectroscopy (TDLAS) experience of the Ebert group previously located at the Physical Chemistry Institute (PCI) at the University of Heidelberg (now at PTB and Technical University Darmstadt). Extensive knowledge, particularly from very robust, laser-based, open-path species diagnostics in large-scale combustion processes (Schulz et al., 2007) as well as from a previously developed stratospheric, balloon-borne, open-multi-path

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



instrument (CHILD) for simultaneous H₂O and CH₄ detection (Gurlit et al., 2005; Ebert, 2006), has been combined for the development of the APicT spectrometer. APicT also takes advantage of knowledge of highly stable, open, White-type (White, 1942) multi-path cells at AIDA developed by KIT (Wagner et al., 2008).

Since 2003, APicT has been used in more and more cloud formation studies at AIDA in order to determine absolute water vapor and super saturation values dynamics during cloud formation (Murray, 2010; Wagner et al., 2008; Möhler et al., 2008a, b, 2005; Mangold, 2005).

APicT uses direct absorption spectroscopy (Schulz et al., 2007; Ebert and Wolfrum, 2001) (instead of the more common 2f wavelength modulation technique (Fernholz et al., 2002) to avoid calibration, while simultaneously permitting an efficient disturbance correction (e.g., for the strong, spectrally broad-band optical losses caused by the cloud particles). This technique has been derived from even more demanding measurements in extremely dust-laden combustion chambers (Dreier et al., 2011; Schulz et al., 2007; Awtry, 2006; Teichert et al., 2003; Ebert, 1998).

APicT detects H₂O vapor molecules via the (110–211) line (7299 cm⁻¹) in the $\nu_1 + \nu_3$ combination band (or for higher H₂O concentrations via the (211 322) line in the 2 ν_1 -overtone-band) both around at 1.37 μm (Gieseemann, 2003; Ebert et al., 2004, 2005, 2006; Hunsmann, 2006).

APicT consist of three basic parts (Fig. A1) (Gieseemann, 2003; Ebert et al., 2005):

- The outer part, placed within the user accessible space outside the AIDA insulated chamber, contains a fiber-coupled temperature-stabilized distributed-feedback (DFB) diode laser module in an N₂-purged container and all electronics (Laser/Peltier driver, signal generator, detector preamplifiers, data acquisition, and computer).
- The middle, opto-mechanical part is a cryogenic transfer optic (including two detectors for I and I_0) directly attached to the outside wall of the AIDA chamber. Using the fiber from the laser module and a lens collimator, a free space laser

**AquaVIT-1 water
vapor
intercomparison**

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



beam is formed, which is then focused through wedged CaF_2 windows into the absorption cell in the inner chamber. Through a second identical window the laser light behind the measurement zone is collected and directed onto an uncooled InGaAs detector (1 mm diam.) using a 2.54 cm spherical mirror. In the middle section, which is at atmospheric pressure but at cryogenic AIDA chamber temperatures, the total path length of the free space beam from the fiber end to the detector is 36 cm.

- c. The inner part is the cryogenic, open, White-type, multi-path absorption cell with heated mirrors. The cell is permanently mounted onto the walls inside the AIDA chamber (3.74 m mirror separation) and enables path lengths of up to 250 m. A path of 82 m was mostly used during AquaVIT-1.

For the measurements the laser is scanned across the target H_2O line with a repetition frequency of 139.8 Hz. The resulting absorption profiles are co-added up to 100 times to yield a 1–2 s time resolution. The averaged, in-situ raw signal is corrected for spectrally broad-band optical losses, electric-optical offsets, converted from time to frequency space and evaluated via a proprietary Labview-based fitting software taking into account (i) the ideal gas law and an extended Beer–Lambert law (Giesemann, 2003; Schulz et al., 2007; Ebert and Wolfrum, 2001; Teichert et al., 2003; Ebert et al., 2005), (ii) a Voigt line-shape-based spectral model that is fitted to the raw data using a non-linear Levenberg–Marquardt fitting algorithm, taking into account measured pressure, temperature and absorption path length inside AIDA, (iii) a highly precise characterization of the dynamic laser tuning (using a 10 cm air-spaced precision etalon), and (iv) H_2O spectral data like our measured air broadening coefficient ($g_{\text{air}} = 0.0981 \pm 0.0002 \text{ cm}^{-1} \text{ atm}^{-1}$ at 296 K) and temperature coefficient ($n = 0.74 \pm 0.01$) (Hunsmann et al., 2006) and the tabulated H_2O line strength from HITRAN2004/8, which was independently verified within a $\pm 3\%$ uncertainty (Hunsmann et al., 2006) during earlier AIDA measurements. For very low H_2O concentrations parasitic absorption effects in the middle section of the spectrometer were also corrected

for by using either a secondary detector channel on the cryo-transfer optics board, or the known ratio between the parasitic and the inner absorption path (0.36m/82m). No calibration or scaling parameters are used in order to derive absolute water mixing ratios.

5 With that procedure, APicT can report absolute H₂O molar fractions from above 10 000 to below 1 ppm at pressure/temperature ranges from 1000 to 1 hPa and 300 to 180 K and achieve precision (1 s) in the 20 ppb range (Ebert et al., 2004, 2005, 2006, 2008). The absolute accuracy estimated from an error budget is better than 5 % and dominated by the line strength uncertainty ($\pm 3\%$). Optical losses by cloud particles
10 of 99 % have been handled without additional systematic errors (even higher losses of 99.99 % have been compensated for in combustion applications using dedicated digital-signal-processing (DSP) based electronics (Awtry et al., 2006)).

The instrument, parts thereof, as well as the principles behind the APicT software and absolute evaluation procedure, was successfully validated in numerous cases in
15 view of achieving absolute accuracy without a regular calibration. Tuning accuracy has been checked via retrieved Doppler line widths at AIDA (yielding 0.15 % deviation). APicT water vapor values were compared at AIDA with the MBW-373LX and FISH instruments during cloud-particle free conditions yielding typically less than 3 % deviation (Hunsmann et al., 2006). A similar open-path TDLAS spectrometer for CH₄ was
20 compared at atmospheric background levels of CH₄ with a NOAA calibrated, extractive gas chromatograph and yielded < 1 % average deviation over 36 h (Lauer et al., 2007).

Recently an improved, extractive version of our TDLAS hygrometer series (using the same data evaluation principles) was investigated in a smaller water vapor intercom-
25 parison at FZ Jülich (Buchholz et al., 2013) and yielded an average offset of -3.9 % with respect to a frost point hygrometer and a Lyman- α absorption hygrometer. Particularly important is the direct validation of our extractive TDLAS spectrometer (named SEALDH) at the German primary national standard for humidity, which resulted in average relative absolute deviations of less than 1.5 % (Buchholz et al., 2014).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



During the AquaVIT-1 static experiments APicT performance however was hampered by a software problem that caused a sensitivity loss of nearly a factor of 10, which should not have impaired measurement accuracy. This precision loss was compensated by reducing the instrument temporal response. During the dynamic experiments (second week) this error was corrected.

A1.1 APicT ice-saturation experiments

The second phase of AquaVIT-1 was focused on dynamic experiments characterized by almost constant average temperature conditions each day (22–26 October) in the chamber superimposed with quick variations of pressure, temperature and humidity conditions. Dense ice clouds were generated during 10 segments (600 s to 2400 s duration) with almost constant pressure and temperature. During these segments, the average relative humidity inside the chamber was maintained at ice-saturation conditions at the respective gas temperature. At constant pressure and temperature conditions, the variability of the gas temperatures measured throughout the chamber volume is typically less than $\pm 0.2^\circ\text{C}$, which means that the variability of the water saturation pressure above the ice-crystal phase is less than about $\pm 3\%$. The average values of water vapor and temperature in the ice-saturated segments were in the range 0.01–40 Pa and 185–243 K, respectively. Figure A2 (top) shows the percentage differences between the APicT water vapor partial pressures and ice-saturation values (p_{wsat}) derived from measured temperatures for segments in the dynamic AquaVIT-1 experiments (Table 1). Figure A2 (bottom) shows the water vapor partial pressures from APicT measurements (p_w APicT) and ice-saturation calculations (p_w ice sat), and the chamber water vapor mixing ratio for each segment. During the segments, the water-vapor partial pressure measured in situ with the APicT instrument deviated by less than $\pm 3\%$ from the ice-saturation pressures calculated from laboratory vapor-pressure equation reported by Murphy and Koop (2005), which have an estimated uncertainty of $\pm 1\%$. APicT water retrievals over a large range are insensitive to the presence of ice clouds under the experiment conditions. The error bars in the figure are derived from

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



the widths of the Gaussian fits to the probability distribution functions of 1 s APiCT and ice-saturation water-vapor partial pressure data in each segment. Within its estimated uncertainty limits of about $\pm 5\%$, APiCT correctly measured the expected ice saturation conditions, with a slight tendency of a low bias at the higher water concentrations and higher temperatures, and a slight high bias at the lower water concentrations and lower temperatures (i.e., at high segment numbers).

A2 Cryogenic Frostpoint Hygrometer (CFH)

A2.1 Instrument description

The CFH is a small balloon borne instrument that measures water vapor continuously between the surface and the middle stratosphere (Vömel et al., 2007a) (See Fig. A3 and Table A1). It is based on the chilled mirror principle by measuring the temperature of a mirror carrying a thin dew or frost layer that is maintained in equilibrium with ambient water vapor. An optical phase-sensitive detector and light-emitting diode (LED) combination measures the bulk reflectivity of the mirror. A microprocessor feedback controller regulates the mirror temperature to maintain a constant bulk reflectivity and, hence, a constant condensate layer. Under this condition the condensate layer on the mirror is in thermal equilibrium with the water vapor in the ambient air passing over the mirror. The mirror temperature is then equal to the ambient dew point or frost point temperature and the water vapor mixing ratio and relative humidity can be calculated from this temperature using a variation of the Clausius–Clapeyron equation and a measure of ambient pressure.

The CFH is currently flown with an electrochemical concentration cell (ECC) ozone sonde, Global Positioning System (GPS), and a radiosonde, which is also used as data transmitter. Therefore, in addition to water vapor mixing ratio, every CFH payload provides ozone mixing ratio and radiosonde pressure, temperature and humidity data, which are used to determine altitude, potential temperature and mixing ratio.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



A2.2 CFH configuration during AquaVIT-1

Laboratory instrument configuration

During AquaVIT-1 a laboratory version of the balloon instrument was used because the latter cannot be operated at subambient pressures (i.e., is not vacuum sealed).

5 A stainless steel sensor housing was built that has largely the same geometry as the balloon borne instrument, but allows the connection to a vacuum system and a separate pressure measurement near the sample volume. The optical path of the laboratory instrument is nearly identical to the balloon instrument, with the only difference being that the detector lens was slightly recessed to allow the addition of a sapphire window,
10 which acted as vacuum seal. The detector and electronics were taken from a balloon instrument. The mirror holder is made of a thin-walled stainless steel tube, which was glued to the mirror disk and acted both as vacuum seal and thermal insulator to the sensor housing. Unlike the balloon instrument, the mirror of the laboratory instrument was fitted with two identical thermistors, each of which was independently calibrated.
15 The dimensions of the mirror, mirror stem and heater coil are identical to the balloon instrument as are the location of the thermistors. This ensures that the thermal properties of the assembly are the same as those of the balloon instrument. The cold finger, which is the part of the mirror assembly that is immersed in the cryogen, was longer than that of the balloon instrument, since accessibility to the sensor housing required
20 a larger separation between the cryogen container and sensor housing. The cooling rate of the instrument, however, is determined by the dimensions of the thin mirror stem, which remained unchanged. Thus the laboratory instrument differs from the balloon instrument largely in the use of different materials, which allowed the connection to a vacuum system. All key parameters of the laboratory version are identical to those
25 in the balloon instrument.

The pressure was measured using a Baratron pressure sensor, which was attached to a separate fitting about 20 cm downstream of the mirror. The flow tube diameter was

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



maintained at 2.54 cm through the CFH and, hence, no pressure gradient is expected for this configuration.

The new laboratory instrument allowed several additional features that are not implemented in the balloon instrument. The most important feature is the installation of a second optical detection axis. In the original configuration (see Fig. A3) the LED and photodetector are arranged such that mirror is lit with an incidence angle of about 85°. For the second optical detection axis small vacuum-sealed windows were installed such that the LED and photodiode observed the same mirror at an incidence angle of about 45°. Rotationally offset by 90° from this LED-detector axis, two white-light LEDs were placed to illuminate the mirror and frost layer. This allowed a visual observation of the mirror and frost coverage through the main sapphire window. Since the CFH is insensitive to ambient light, the addition of the white-light LEDs did not impact the detector operation. The relocation of the detector LED and photodiode from the specular axis resulted in some loss of signal, which is compensated for by referencing the frost coverage to the clear mirror signal. It did not impact the operation of the CFH or lead to a change in PID (proportional, integrating, differentiating) controller settings. Only one of the two detection axes at a time could drive the mirror temperature control.

A2.3 CFH operation during AquaVIT-1

Cryogen

The CFH during AquaVIT-1 was operated using liquid nitrogen. Since it is significantly colder than trifluoromethane (CHF₃), which is used in balloon soundings, a Teflon sleeve was placed over the cold finger to reduce the heat transfer into the cryogen. The Teflon sleeve nearly compensated for the lower temperature of liquid nitrogen since the mirror heater currents were comparable to those observed in flight. The cryogen container capacity was about twice that of the balloon instrument, liquid nitrogen was added more often (about every 2.5 h) because the heat capacity of liquid nitrogen is less than that of trifluoromethane.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Mirror cleaning and clearing

During the first week of operation the CFH remained connected to the AIDA chamber, was not opened, and the mirror was not cleaned during this set of experiments. A slight film deposit was observed on the fifth day of the first week, which prompted a mirror cleaning after the morning experiments. During the opening of the instrument, a vacuum fitting was damaged and had to be replaced. Therefore, no measurements could be performed during the afternoon of the fifth and last day of static experiments. Cleaning was done using cotton swabs and methanol, while monitoring the optics signal. This procedure is identical to the pre-launch instrument preparation for a balloon sounding. Cleaning was completed when the reflectivity signal was maximized.

In order to maintain a consistent frost layer, it was frequently cleared by heating the mirror to about 20 °C and evaporating the condensate layer completely. A similar process occurs during balloon flights. The frost clears within one second. After mirror clearing the instrument resumes normal operation with a newly formed frost layer. However, during AquaVIT-1 when the chamber water-vapor mixing ratio was low (frost-point temperatures typically less than 210 K), the frost layer could not be reformed readily and some measurement intervals were incomplete. When this occurred, water vapor was added to the system through a leak valve, which allowed the CFH to create a frost layer at a higher water vapor concentration than was available from the chamber. Establishing control at the higher water vapor concentrations using the leak valve took about one second. Closing the leak valve with continuous control took about one minute. Thus, it took about two to three minutes to clear the mirror to re-condense the frost layer and to bring the CFH into stable control at the water vapor concentration during the days of the lowest frostpoint temperatures. The times between mirror clears varied between several minutes to about two hours. The upper limit was set by the buffer overflow problem of the acquisition software. During a balloon sounding, which typically takes 90 min on ascent and 30 to 45 min on descent, two mirror clears

AMTD

7, 3159–3251, 2014

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



happen. Thus the time between mirror clears during AquaVIT-1 was comparable to that of a balloon sounding.

A2.4 CFH experiments in addition to the AquaVIT-1 program

Detector location and microscope observations

5 During the second and third day of AquaVIT-1 both detection axes were used to study if the reduced signal in the off-perpendicular detector configuration causes any negative impacts. No difference between these two configurations was observed. Starting on day four of AquaVIT-1 all experiments were run using the off-perpendicular detector configuration. This configuration allowed observing the mirror directly through the
10 perpendicular detection port using a microscope. For nearly all measurements in this configuration visual images of the frost layer exist. These visual images can be used as an independent quantification of a stable frost layer.

Controller settings

15 During the first week of operation the PID parameters used to run the CFH were those used by the balloon instrument, which depend only on frostpoint temperature. During the second week a number of experiments were performed with changed parameters to study whether the PID parameters could be optimized. All of these experiments remained inconclusive and aside from occasional controller instability, no changes in instrument performance were observed.

20 Frost layer morphology

The mirror clear capability was used to study the influence of different frost layer morphologies. These experiments showed that the frost layer morphology can have a significant influence on the performance of a frostpoint hygrometer. However, all data

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



submitted to the AquaVIT-1 archive used a frostlayer consistent with one that would have formed during a routine CFH sounding.

Liquid–ice transition

The AquaVIT-1 experiment focused on low water vapor concentrations, which translated to low frostpoint temperatures. However, in the atmosphere there exists a frostpoint temperature range in which the phase of the mirror condensate is not uniquely defined and in which the measurement of reflectivity does not provide sufficient data to distinguish the condensate phase. The transition between liquid phase and solid phase of the condensate on the mirror was studied after the completion of one experiment day. These experiments clearly show that the liquid to ice transition for the mirror condensate may extend to mirror temperatures colder than -30°C . Without special precaution, the condensate on the mirror may remain liquid down to these temperatures. Spontaneous frost formation will then create a frost layer whose morphology is undetermined. The poorly defined frost layer morphology may negatively impact the controller stability at much colder temperatures. The control routines used in the CFH force the condensate layer to freeze at a predetermined temperature (usually -15°C) and always lead to a well defined frost layer morphology. The same routine as used in flight was used in the AquaVIT-1 experiments, ensuring that the results of this experiment are transferable to the performance of the instrument in atmospheric soundings.

Flow rate dependency

The exhaust line of the lab CFH was connected to the vacuum system using a needle valve. Using this valve the flow through the instrument could be varied to study any possible flow rate dependency of the measurement. No such dependency could be found. In fact, flow rate variations led to strong pressure variations in the instrument as well as to strong frostpoint temperature variations. Measured frostpoint changes were as fast as pressure changes when the needle valve was either opened or closed.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



However, the changes in frostpoint perfectly canceled changes in pressure leading to constant mixing ratios throughout these experiments. This was taken as confirmation that the instrument responds as it should to changes in partial pressure, when measuring a constant mixing ratio.

5 **Comparison to PTB water vapor permeation source (PTB-WVPS)**

The PTB (Braunschweig, Germany) provided a primary referenced water-vapor permeation source, which was used as independent reference on day four of the second week of experiments. After 1 h of purging the combined flow systems, CFH water vapor values agreed with the permeation source to within 5 % and after 2.5 h to within 2 %.

10 **A3 Fast In situ Stratospheric Hygrometer (FISH-1 and FISH-2)**

A3.1 Instrument description

The FISH instruments, developed at the Forschungszentrum Jülich (Germany), are based on the Lyman- α photofragment fluorescence technique. FISH has been used in several balloon and aircraft campaigns and compared to many other hygrometers during atmospheric measurements.

The instrument is described in detail by Zöger et al. (1999). FISH consists of a closed, vacuum-tight fluorescence cell, a Lyman- α radiation source, a PMT in photon-counting mode, detectors to monitor the vacuum ultraviolet (VUV) radiation output of the Lyman- α lamp, and a mirror drive that controls the measuring cycle: the determination of the fluorescence N_g and background count rate N_u and of the lamp intensity I_0 . The water vapor mixing ratio is determined via $[\text{H}_2\text{O}] = c_k \cdot (N_g - f_u \cdot N_u) / I_0$. The coefficients c_k and f_u are determined in the calibration procedure.

The calibration of FISH is done in regular intervals between the measurement deployments using a calibration bench, consisting of a humidity mixing system and a frostpoint hygrometer (MBW DP30) as a reference instrument (Zöger et al., 1999; Meyer

et al., 2014). A calibration is done by varying the pressure (typically 80–500 hPa) and mixing ratio (1–500 ppm) to cover the relevant range of the atmospheric measurements.

A3.2 Performance during AquaVIT-1

5 During AquaVIT-1, two instruments were operated: FISH-1 designed for and used on Lear Jet missions and here operated from the upper AIDA platform, and FISH-2 designed for and used on Geophysica and Falcon missions and here operated from the lower AIDA platform. Both instruments are almost identical concerning the core components; however the lamp of FISH-1 couldn't be tuned in its optimum mode resulting
10 in a lower signal-to-noise ratio compared to FISH-2.

Both hygrometers worked nominally during all AquaVIT-1 experiments except for 22 October, as the high H₂O concentrations during that particular dynamic experiment caused optically thick conditions for FISH. Flow rates on the order of 10 standard liters per minute (SLM) could be achieved and are comparable to those during aircraft operations; only a few periods with flow rates below 5 SLM had to be removed from FISH-2
15 data when contamination could not be excluded.

Calibrations were carried out on 10, 15, 20, 21, 24 and 30 October. No significant drift of the calibration factors c_k and f_u could be detected over the AquaVIT-1 period. The reproducibility of c_k was better than 3% (FISH-2) and 4% (FISH-1) for individual
20 calibrations (Meyer, 2008).

The MBW-DP30 reference hygrometer was calibrated by the manufacturer with traceability to National Physical Laboratory (NPL, Middlesex, UK) standards a few months before AquaVIT-1. During AquaVIT-1, this DP30 was also compared to the secondary PTB standard (PTB-WVPS in Sect. 6.5). From this comparison, the MBW-DP30 accuracy was determined to be 2% in volume mixing ratio (Meyer, 2008; Meyer
25 et al., 2014).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



A3.3 Lessons learned during AquaVIT-1

At high mixing ratios and low pressure, an unexpected low bias of FISH data compared to the other AquaVIT-1 instruments occurred. In response, additional laboratory calibrations carried out over an extended pressure range in the home laboratory revealed a pressure dependence that is likely caused by O₂ absorption of the tails of the Lyman- α line (Meyer et al., 2014). This pressure dependence can be corrected by introducing a pressure dependence of c_k (not done for the data shown in this paper). The airborne data of FISH however, are not impacted by this effect as such conditions since mixing ratios of several ten to hundreds ppm at pressure below 100 hPa, do not occur in the atmosphere.

A4 FLuorescent Advanced Stratospheric Hygrometer for Balloon (FLASH-B1 and FLASH-B2)

A4.1 Instrument description

The FLASH-B instruments are a compact lightweight sonde developed at the Central Aerological Observatory (CAO) (Russia) for balloon-borne water vapor measurements in the upper troposphere and stratosphere (Yushkov et al., 1998, 2001). The instrument is based on the Lyman- α fluorescence technique. The source of Lyman- α radiation ($\lambda = 121.6$ nm) is a hydrogen discharge lamp while the detector of OH fluorescence at 308–316 nm is a Hamamatsu R647-P photomultiplier run in photon counting mode with an narrowband interference filter for selecting the fluorescence spectral region. The intensity of the fluorescent light sensed by the photomultiplier is directly proportional to the water vapor mixing ratio under stratospheric conditions (30–150 hPa) with small oxygen absorption (3% at 50 hPa). The background signal caused by the night sky emissions in the absence of fluorescence light is detected using lamp modulation with a 1-kHz square wave with 1/8 duty cycle and synchronous demodulation of the received signal. The H₂O measurement range is limited to below 400–500 hPa due to

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



strong Lyman- α absorption in the troposphere. FLASH-B uses the open layout (Khaplanov et al., 1992) with the optics looking directly into the outside air. This arrangement is suitable only for nighttime measurements.

The source of VUV radiation used in the FLASH-B instrument is a hydrogen glow-discharge lamp filled with a mixture of hydrogen and helium at the total pressure of 10 hPa. Unlike the more sophisticated hygrometers based on the fluorescence technique, FLASH-B does not have VUV photon flux control; however, this is compensated for by a very precise stabilization of the lamp's current, to which the emission intensity is directly proportional. The hydrogen glow-discharge lamps used in the FLASH-B instrument have been proven to have very stable intensity of the Lyman- α emission over both operation and storage time. The VUV light sources containing the mixture of hydrogen and helium are known to have stray helium line (318 nm) emission, which overrides the spectrum of hydroxyl fluorescence and thus may cause spurious signal from reflection (backscattering) of this emission. The FLASH-B lamp is equipped with a special MgF₂ window-filter for suppressing the 270–320 nm band emission.

The accuracy of the FLASH-B instrument is determined by the calibration error estimated as 4% in the 3–100 ppm range. The measurement precision is 5.5% calculated for 4 s integration time at stratospheric conditions. The total uncertainty of the measurement is less than 10% at the stratospheric mixing ratios greater than 3 ppm increasing to about 20% at mixing ratios less than 3 ppm. The total relative error of calibration amounts to 4%. The calibration fit function is linear in the pressure range of 30–150 hPa and water vapor mixing range of 1–300 ppm. At higher pressures the VUV absorption by oxygen and water vapor is taken into account. The lamp stray light backscattered from the chamber walls does not affect the calibration since the calibration coefficients are determined as the slope of a regression line.

FLASH-B has been successfully used in a number of balloon campaigns (e.g., LAUTLOS-WAVVAP, SCOUT-AMMA, TC4, LAPBIAT-II, TRO-pico) that included simultaneous measurements of stratospheric water vapor by different measurement techniques. In particular, point-by-point comparison with the frost point hygrometer from the

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



NOAA Climate Monitoring Diagnostics Laboratory (Boulder, CO, USA) showed a mean deviation of 2.4 % with 3.1 % standard deviation (1σ) (Vömel et al., 2007b) and comparison with CFH showed a mean deviation of 0.8 % with 4 % standard deviation (Khaykin et al., 2013). A number of unpublished field intercomparisons between FLASH-B and CFH flown simultaneously at various locations have provided very similar results with discrepancies below 5 % in the 1 to 10 ppm mixing ratio range.

A4.2 Performance during AquaVIT-1

During AquaVIT-1 two FLASH-B instruments were operating independently: FLASH-B(1) was mounted onto the AIDA vessel and made in situ measurements directly in the vessel, and FLASH-B(2) ran in an external warm chamber connected to AIDA via heated tube. An important point, which has to be thoroughly considered when assessing FLASH-B performance during the AquaVIT-1 experiments, is the effect of stray light backscattering in both AIDA vessel and FLASH-B2 external chamber. As mentioned above, the hydrogen lamp has a spurious emission line, overriding the fluorescence spectrum, and therefore the reflection of the lamp's emission from a surface would be treated by hygrometer as a water vapor signal. The latter is not an issue for the flight experiments, but does arise when the hygrometer is taking measurements inside a chamber.

During AquaVIT-1 the effect of stray light backscattering was observed for both FLASH-B1 and FLASH-B2 instruments. Attempts to estimate the contribution of stray light to the FLASH-B signal were performed for every static experiment by assuming a constant water vapor partial pressure. However, since the partial pressure was actually varying within every static experiment the estimates of stray light contribution (SLC) turned out incorrect. It was found that the SLC was slightly varying from one experiment to the other introducing an additive of 7 ± 2.5 ppm to the value of water vapor mixing ratios from the fluorescence signal. Errors in estimation of SLC introduced a systematic offset into the FLASH-B1 water-vapor data as described below.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



was a dramatic increase of noise in the data, known from field experiments to be the consequence of water contamination due to outgassing from balloon or instrument surfaces. Although it is somewhat doubtful that the chamber wall could be a source of such water contamination, it is absolutely clear that the low pressure effect is not an instrument-related issue.

During the AquaVIT-1 period FLASH-B2 was calibrated twice using the MBW-DP30 hygrometer and the FISH calibration bench. No change of calibration factor was detected. Both FLASH-B1 and 2 instruments were calibrated at CAO after the AquaVIT-1 campaign and did not show any change in calibration factors, which proves at best the stability of FLASH-B calibrations.

A4.3 Lessons learned during AquaVIT-1

The experimental set up of the FLASH-B instruments in the AquaVIT-1 campaign resulted in major issues that were not properly identified before and during the experiments. FLASH-B is an open-cell Lyman- α hygrometer; therefore when operated in a closed chamber it receives reflected stray light that is not produced during balloon flights. The latter introduces a constant additive offset to the water vapor signal. In order to avoid this, the stray light has to be accurately determined for every particular chamber experiment. The only way of accurate stray light signal determination is to measure a known water vapor concentration in the chamber. For this, a single measurement point is enough because the sensitivity of the sensor is governed by the calibration factor, which proved to remain constant during the entire campaign. The procedure of stray light detection as described above was not performed during AquaVIT-1. Other attempts to determine or estimate stray light contribution either failed or turned out highly inaccurate, resulting in a constant bias in the FLASH-B data.

The adaptation of FLASH-B1 instrument to AIDA chamber, in which FLASH-B analyzed volume is located very close to AIDA wall caused boundary wall effects, becoming apparent at low pressures as a sign of water contamination. For future chamber

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



experiments with FLASH-B the adaptation scheme and experimental set up have to be seriously reconsidered.

A5 Harvard Water Vapor (HWV)

A5.1 Instrument description

5 HWV is an aircraft-borne water vapor instrument that has flown on NASA's WB57F since 2001, successfully participating in the Clouds, Water Vapor and the Climate System mission (CWVCS), CRYSTAL FACE, the Middle latitude Cirrus Experiment (MidCiX), Pre-AVE (Aura Validation Experiment), AVE, AVE-WIIF (Water Isotope Intercomparison Flights), CRAVE
10 (Costa Rica AVE), and the Tropical Composition, Cloud, and Climate Coupling Experiment (TC4) missions prior to the AquaVIT-1 comparison (Weinstock et al., 2009). The instrument detection axis is a lightweight version of the one that has flown on the ER-2 aircraft since 1992 (Weinstock et al., 1994). The operating principle, optical configuration and laboratory calibrations of these instruments are nearly
15 equivalent, thus providing a self-consistent set of measurements of upper tropospheric and lower stratospheric water vapor from 1992 to 2007. The HWV instrument (as it flew until 2011) is shown on the left in Fig. A5 mounted in the WB57F left spearpod. On the right is a schematic representation of the flow through the instrument. The inner secondary duct picks up the laminar core of the ram-fed flow. Flow velocities to the detection axis are controlled by a throttle valve and are varied during flight from
20 ~ 40 to $\sim 80 \text{ ms}^{-1}$ to verify a "wall-less" flow system. This approach has been used for all our aircraft-borne water vapor measurements on the ER2 and WB57F since 1992.

The HWV instrument, and its companion, Harvard Total Water, HTW, have been previously described in detail by Weinstock et al. (1994, 2006a, b). Briefly, as illustrated
25 in the inset of Fig. A6, 121.6 nm (Lyman- α) radiation from a sealed radio-frequency discharge lamp photo-dissociates water vapor in a 5 cm \times 5 cm square duct. A fraction of the resulting OH fragments are formed in their first excited electronic state ($A^2\Sigma^+$), and

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**AquaVIT-1 water
vapor
intercomparison**

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



the OH fluorescence at ~ 315 nm is collected at right angles to both the Lyman- α beam and the air flow. The fluorescence passes through a filter assembly and is detected with a photomultiplier tube (not shown). Because the fluorescence is strongly quenched by collisions with O_2 and N_2 at a rate proportional to the air density, at altitudes of the upper troposphere and lower stratosphere the observed detector signal is effectively proportional to the water vapor mole-fraction. Solar and lamp scatter near 315 nm are measured by using a quartz window to periodically block the Lyman- α beam. This background measurement is typically executed for 2 s out of every 10 s. Changes in lamp intensity monitored with a vacuum ultraviolet (VUV) photodiode opposite the lamp are used to normalize the fluorescence signal. The lamp intensity at the fluorescence detection volume is calculated by correcting for absorption by water vapor, and the pressure-dependent absorption of 121.6 nm light by molecular oxygen. A rear-surface coated MgF_2 mirror adjacent to the VUV diode reflects some of the radiation back across the duct to a second VUV diode, permitting simultaneous measurements of water vapor by direct (Beer–Lambert) absorption at sufficiently high water vapor (mid to upper troposphere) concentrations.

Empirically determined proportionality constants define the sensitivity of the instrument to water vapor detection. These calibration constants describe the pressure-dependent relationship between the net normalized fluorescence signal and the water vapor concentration in the duct. Extensive laboratory calibrations are run over a range of water vapor mixing ratios (~ 0.5 to ~ 500 ppmv), pressures (~ 30 to 500 hPa), temperatures ($\geq -60^\circ\text{C}$), and flow rates, chosen to replicate the flight environment as closely as possible. Figure A6 shows a schematic of our laboratory calibration system. HWV is calibrated with two independent reference standards. A two-stage bubbler is used as both a water addition system and a primary standard. A small flow of air is brought to water vapor saturation by bubbling it through a glass frit immersed in distilled water in two stages; this flow is then combined with a main flow of dry air to produce air with known water vapor concentrations from 1–500 ppmv. The added water vapor concentration is known to within 5 % based on the uncertainty in the temperature

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



between our detection axes and the chamber to allow for laboratory calibrations without any instrument or ducting disassembly. Figure A7 shows calibration data for two days, 16 and 19 October, that were acquired immediately following AquaVIT-1 static experiments on those days. These calibration runs, in particular the one on 19 October, simulate the pressure and mixing ratio regimes of the chamber experiments (i.e., low water vapor mixing ratios and high pressures). Figure A7 (top) shows selected dilution steps from these two runs. The first calibration run explored mixing ratios from ~ 1 ppmv to 21 ppmv in steps of 1 to 5 ppmv, and had a background mixing ratio of ~ 1.1 ppmv. The second explored mixing ratios from ~ 0.5 to 3 ppmv in steps of ~ 0.5 ppmv, and had a stable background of ~ 0.6 ppmv when no water was added. Figure A7 (bottom) shows the inverse of the pressure-dependent calibration constant, C [counts/(# $_{\text{H}_2\text{O}}$ cm $^{-3}$)], derived from the dilution step data, plotted as a function of the number density in the duct, M [# cm $^{-3}$]. The solid black line represents the results of a linear fit to the complete set of calibration data acquired during the AquaVIT-1 campaign, and the dashed lines denote the $\pm 5\%$ range. The data from these two runs fall well within the $\pm 5\%$ uncertainty envelope. These results demonstrate the linearity and robustness of the calibration over an order of magnitude in pressure and from a few tenths of a ppmv to tens of ppmv.

The laboratory environment at AquaVIT-1 introduced sources of systematic error that were substantially different from those encountered in flight, and that degraded the performance of HWV. The HWV instrument was plagued by a systematic bias and large uncertainties associated with a combination of leaks into the sample duct and outgassing from the walls upstream of the detection axis. Note, neither outgassing nor leaks are expected to impact either the determination of the instrument sensitivity (i.e., calibration) or the airborne HWV measurements. In flight there is little pressure differential across the instrument duct, in fact pressures inside the duct are greater than ambient, often by ~ 10 hPa. Furthermore, to achieve fast time response and minimize the impact of outgassing in flight, HWV was specifically designed for high flow velocities and correspondingly high mass flow rates. Mass flows in the instrument duct during

fall just outside their combined uncertainty envelopes. However, these observed differences at the level of 10% cannot explain the larger systematic differences found in atmospheric observations as shown in Fig. 1.

A6 JPL Laser Hygrometer (JLH)

A6.1 Instrument description

The NASA Jet Propulsion Laboratory (JPL) JLH instrument is a single-channel, near-infrared, open-path, tunable diode laser spectrometer for in situ measurements of atmospheric water vapor from aircraft platforms in the troposphere and stratosphere (May, 1998). Three laser hygrometers were developed at the Jet Propulsion Laboratory, California Institute of Technology, for the NASA ER-2, WB-57 and DC-8 aircraft platforms and have participated in numerous NASA missions from 1997 to the present: POLARIS, CAMEX, ACCENT, SOLVE, CRYSTAL-FACE, Aura Validation Experiment (AVE), AVE-WIIF, MidCiX, PUMA-A, Costa Rica AVE, and TC⁴.

The light source for JLH is a near-infrared distributed feedback (DFB) tunable diode laser operated at 1370.0 nm wavelength. The laser beam passes through an aspheric lens, and is focused at the midpoint of a multipass cell in the Herriott optical configuration (Altmann et al., 1981; Herriott et al., 1964). Both laser and detector are temperature stabilized at 15 °C on a thermoelectrically cooled aluminum mount inside an evacuated aluminum housing (the blue sampling enclosure in Fig. 3). The laser beam enters the Herriott cell through a hole in the mirror closest to the laser. The mirrors are Au-coated spherical mirrors made of zerodur. The separation between the mirrors is maintained to within a tight tolerance by invar rods for optical stability over a wide range of temperatures (verified from 180 K to 300 K). The laser beam traverses 50 passes of the Herriott cell (corresponding to 1039.3 cm optical pathlength), and then returns through the same mirror hole to impinge on a detector. There is an additional 8.6 cm path through the air between the Herriott cell and the window to the evacuated laser/detector housing, yielding a total optical path of 1047.9 cm. The virtual sampling

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



volume is a 1.6 L hyperboloid between the two mirrors (plus the small path to the window).

JLH operates in two modes: harmonic wavelength modulation spectroscopy (May, 1998; May and Webster, 1993) and direct absorption spectroscopy. In harmonic spectroscopy, the laser is scanned at an 8 Hz repetition rate across the strong water absorption line at 1370.0 nm (7299.43 cm^{-1}), and modulated at $f = 64\text{ kHz}$ for sensitive harmonic detection at $2f$. Each ten scans are averaged to report data at 0.8 Hz (1.3 s). The modulated $2f$ peak-to-peak amplitude (“pp2f”), normalized by returned laser power, is used to calculate water mixing ratios as described in May (1998). The advantages of harmonic spectroscopy are fast response and high precision.

Every 24 s, JLH switches to its second operational mode, namely direct absorption spectroscopy. The advantages of direct absorption measurements are high accuracy, detection of background water, signal specific to water, and more highly constrained results (i.e., the laser linewidth is constrained by fitting the absorption line). The volume mixing ratio of water vapor is calculated from the Beer–Lambert law using the known pathlength and spectroscopic parameters from HITRAN 2004 database (Rothman et al., 2005) which did not change in the HITRAN 2006 update. One can either tie the direct absorption calculation of water to (1) the spectroscopic parameters in HITRAN and the measurements that were used to make the line list in the database, or (2) in situ calibration with a reference source, while varying the spectroscopic parameters accordingly.

For the AquaVIT-1 experiment, the JLH team assumed that the HITRAN 2004 parameters were correct, and calculated water mixing ratios on that basis. In addition to pathlength and spectroscopic parameters, the absolute wavelength scale must be known. For JLH, the wavelength scale was characterized by measuring the spacing of seven weak methane absorption lines in the 7299 cm^{-1} region. The linecenter positions of these methane lines have been accurately measured (Brown et al., 2005) and bracket the targeted water line at 7299.43 cm^{-1} . In the calibration experiments, the calibration chamber was back-filled with methane and trace water. The spectra of

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



water and methane were measured simultaneously. The water line position and wavelength scale of the laser scan were accurately determined by interpolation between the methane lines.

The largest source of error in this direct absorption measurement is the uncertainty in the spectroscopic parameters in the HITRAN line list. A secondary source of error arises from assumptions made in fitting the absorption line (e.g., Skrotzki et al., 2013). The absorption line is fitted with a Voigt lineshape, using occasional checks with a Galatry lineshape to ensure that the Voigt fit does not introduce errors greater than 3%. As an additional check on the validity of the spectroscopic parameters, the water vapor mixing ratios are compared with the known humidity output from a NIST-traceable Thunder Scientific 3900 Low-Humidity Generator. JLH is placed inside a stainless steel chamber with continuous gas flow from the humidity generator. The water vapor mixing ratio can be monitored at the input and/or output of the chamber by a mid-infrared direct absorption measurement at 5316 cm^{-1} (Troy et al., 2007). The final JLH data are the fast-response (1.3 s) harmonic measurements, scaled to match the 24 s direct absorption measurements for higher accuracy.

A6.2 Performance during AquaVIT-1

During the AquaVIT-1 static experiments (10–19 October 2007), the JLH laser, detector, and open-path optical cell were mounted inside the AIDA chamber. These components were mounted to a flat aluminum plate, offset by posts, and the plate was mounted 80 cm above the base of the AIDA vessel (Fig. 3). JLH experienced realistic UT/LS conditions of pressure, temperature, and water vapor mixing ratios in the open-path optical cell inside the chamber. These conditions are the same when JLH flies on research aircraft, except that the flow rates across the optical path are much faster on aircraft. In both AquaVIT-1 (outside the chamber) and aircraft field missions, the control electronics are maintained at $25\text{ }^{\circ}\text{C}$. For the data analysis, measured chamber pressure and temperature are used.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



JLH performed well on every day of the static experiments. The laser modulation was optimized for the 100 to 200 hPa pressure range, so this is where JLH had the greatest sensitivity and precision. In the range of 2 to 200 ppm, JLH agreed with APiCT to within 10 % on average. This accuracy is within the experimental uncertainty of JLH, but JLH is biased high relative to APiCT. Since JLH and APiCT use the same water absorption line (1370.0 nm), systematic differences can be narrowed down to four possibilities: misalignment of JLH optics at low temperature, the accuracy of measuring laser properties and tuning rate; assumptions about spectroscopic parameters, line broadening and line shape; and the method of fitting direct absorption lines.

JLH precision was not ideal during AquaVIT-1. JLH was installed with a 3 m long signal cable between the optics inside the chamber and the control electronics external to the chamber. The signal noise was an order of magnitude higher in the chamber than on aircraft. The long cable is suspected of picking up noise making the precision worse. As a result, the JLH limit of detection in AquaVIT-1 was 0.15 ppm, and no data less than 1 ppm was submitted to the archive. At higher mixing ratios, the precision did not adversely impact the comparisons.

A6.3 Lessons learned

Mounting the JLH laser, detector, and optical cell inside the chamber worked extremely well, and the JLH team recommends this approach for any open-path instrument. If this experiment is repeated, the precision could be improved by mounting the amplifier closer to the detector.

A7 MBW-373LX

The MBW-373LX is a chilled-mirror frost point hygrometer that is used as a reference instrument for AIDA water vapor measurements. Briefly, the measurement accuracy given by the manufacturer for the frost point is ± 0.1 K (i.e., ± 0.13 ppm at 10 ppm) with a precision of 0.05 K for operation at 2.5 bar with a sample flow between 0.5

clear. Besides possible problems in the permeation source or the dew point mirror instrument the flow determination seems to be the largest source of uncertainty.

A8 SnowWhite

The night version of the SnowWhite frost point hygrometer (Schmidlin, 1999; Fujiwara et al., 2003; Vömel et al., 2003; Vaughan et al., 2005; Miloshevich et al., 2006) manufactured by Meteolabor (Wetzikon, Switzerland) was installed inside the AIDA chamber. The entire radiosonde as set up to fly on meteorological balloons was mounted on a steel plate approximately one meter above the vessel floor. For the two-week campaign duration the instrument was not accessible. Therefore, the following modifications with respect to the balloon configuration were made: the batteries allowing operation for four hours flight time were replaced by an external supply located outside the chamber. Water bags that heat the instrument internally to temperatures close to 0 °C during freezing were removed and replaced by heaters with thermostats. The radio sonde transmitter was not enabled: instrument data were read out by a remote computer through a serial RS-232 interface, which is normally used during pre-flight setup. Water vapor partial pressure was calculated from the frost point temperature according to Murphy and Koop (2005). Meteolabor specifies 0.2 °C for both frost point temperature precision and accuracy leading to a total error of about $\pm 5\%$ (depending on temperature) in the water vapor mixing ratio.

Two major difficulties were encountered during the two weeks of operation. The heaters replacing the water bags caused periodic oscillation of the temperature inside the Styrofoam housing with strong transients as compared to the gradual temperature change during flight conditions. The resulting temperature gradients in the instrument electronics lead to erroneous reading of the frost point and ambient temperatures. Following advice from the manufacturer, this problem was solved during the second week by operating the heaters at a temperature below the thermostat switch-off level and manually adjusting their power such that the internal temperature remained in the range of 0 to 5 °C. Thus, conditions comparable to those during balloon flights were

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



obtained on four days during the second week. During the remaining time without the improved heating, especially for the static experiments carried out in the first week, data quality was much poorer.

The second problem was caused by the small ventilation of the sensor due to slow flow conditions inside the chamber. Under flight conditions the ascent of the balloons results in an air flow velocity 5 ms^{-1} . In contrast, at the instrument position near the chamber walls the corresponding downward air movement induced by a mixing fan circulating the air upward along the chamber center line was estimated to be at least a factor of ten less in magnitude. As a consequence, at low water vapor partial pressure corresponding to mixing ratios of 10 ppm or less, mass exchange at the frost point mirror was very limited leading in general to a response too slow for the feedback loop to stabilize safely on the frost point temperature. A further effect was that in the early phase of each experiment with mixing ratios less than 100 ppm it took very long for the frost layer to build up. In the second week however it was always established during the first dynamic cloud experiment. This second problem could not be solved during the campaign. Therefore, low mixing ratio data in general has to be analyzed with care. This experience shows that for future applications inside a climate chamber adequate thermal control and air flow are prerequisites to achieve performance comparable to flight conditions.

For the above reasons it could be misleading to judge SnowWhite's general performance based on an analysis restricted to the first week of static experiments as summarized in Fig. 6. In consequence, cloud free stationary segments defined by the referees for the second week of the intercomparison were consulted for additional data analysis. All segments in the second week in which SnowWhite delivered data (i.e., 17 out of 18 for the first four days) were subjected to the data treatment described in Sect. 6.2 with the APicT taken as reference (during the final day water vapor was below the SnowWhite detection limit). Figure A9 shows the relative deviation of the SnowWhite with respect to the APicT on its abscissa with the ordinate indicating the water vapor mixing ratio measured by the APicT during the respective segment. For

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



**AquaVIT-1 water
vapor
intercomparison**

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Wavelength modulation spectroscopy is used in combination with second-derivative detection in order to improve the signal-to-noise ratio as well as to minimize the contribution of absorption due to residual water inside the free optical paths.

In order to perform isotopic-ratio measurements it is necessary to probe absorption lines with relatively similar integrated line strengths. Due to the low abundance of the heavy isotopologues H_2^{18}O and H_2^{17}O it is therefore necessary to probe a relatively weak H_2^{16}O line. Although a good detection limit has been achieved in terms of optical density, the relatively low integrated line strength in turn leads to a somewhat reduced precision in the H_2^{16}O measurements.

For calibration purposes a calibration gas source was developed that produces a gas mixture of known isotopic composition by injecting small droplets of water ($\approx 30\ \mu\text{m}$ diameter) into a flow of dry N_2 . Since ISOWAT has been developed to measure water-isotopic ratios, all our efforts were focused to achieve stable isotopic ratios rather than water mixing ratios of ultimate stability. This calibration source has a calibration uncertainty of around ± 50 ppm. However, it is important to stress that this uncertainty does not lead to uncertainties in the isotopic ratios.

ISOWAT was connected to the chamber via heated stainless steel tubing to an extractive port. A flow of around 1.25 SLM through the MPC was established by a membrane pump downstream of the instrument. The pressure inside the MPC was controlled by a proportional valve located upstream of the cell.

During AquaVIT-1 the first prototype of ISOWAT was tested, which was finalized just prior to the campaign. During the campaign relatively large drift was observed in the mixing ratios (not in the isotopic ratios) that could later be linked to thermally induced changes in the optical alignment of the instrument. Since the campaign this issue has been largely eliminated. Furthermore, the sensitivity of ISOWAT was increased by careful investigation and elimination of noise sources. The current performance of ISOWAT can be found in Dyroff (2010). ISOWAT has been deployed since April 2010 aboard the CARIBIC A340-600 passenger aircraft (Brenninkmeijer et al., 2007), where it provides monthly measurements of $\delta^{18}\text{O}$ and δD in the upper troposphere.

A10 Open-path Jülich Stratospheric TDL Experiment (OJSTER)

A10.1 Instrument description

OJSTER is an open path cell diode laser experiment manufactured by MayComm Inc. (California, USA). The open-path Herriot multipath cell allows for absorption lengths of 4 m. The 1.37 μm diode laser can be tuned over a H_2O absorption line, and analysed both in direct absorption or the 2f mode. The instrument is similar to the ER-2 instrument described in May (1998), as well as the matrix method to compensate for pressure and temperature dependence. This calibration matrix is provided by the manufacturer and is derived from measurements made in a climate chamber against an Edgetech Dewprime 2000 Chilled Mirror system. This calibration has been re-evaluated in the Jülich laboratories against the MBW-DP30 frost point hygrometer, which also serves as reference hygrometer for FISH calibrations. First aircraft measurements and characterization of OJSTER in the laboratory are given in Schlicht (2006).

A10.2 Performance during AquaVIT-1

During AquaVIT-1, the open-path cell of OJSTER was mounted in a stainless-steel container with a volume of a few litres in the cold space outside the chamber walls (Fig. 2). Chamber air was extracted with a pump via stainless-steel tubes (Fig. 3). This experimental setup had two major disadvantages: first, the outgassing of the OJSTER cell components was very slow, even after several cycles of purging with dry warm air a background H_2O contamination of up to 20 ppm was detected. Secondly, the container leaked at the lowest temperatures; therefore, no measurements at all were obtained on AquaVIT-1 days with the lowest temperatures. OJSTER data during AquaVIT-1, therefore, cannot be used for a quantitative comparison with other hygrometers.

AMTD

7, 3159–3251, 2014

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



A11 PicoSDLA

PicoSDLA is a tunable diode laser spectrometer that measures in situ H₂O by absorption spectroscopy at 2.63 micron. The beam emitted by an antimonide laser diode is propagated in the open atmosphere over a 1 m distance and in situ spectra are recorded with an InAs detector at 1 s intervals. The sensor is designed to be operated from stratospheric balloons. At the time of the AquaVIT-1 experiment, PicoSDLA was still under development and it had just been tested in flight three months before the beginning of the campaign. For a detailed description of the laser hygrometer see Durry et al. (2008).

For the calculation of water vapor mixing ratios, a non-linear least-square fit was applied to the full molecular line shape using a molecular model based on the Beer–Lambert law in conjunction with pressure and temperature measurements yielded by onboard sensors. The detection technique is a standard direct-differential method. With all the sources of errors taken into account (baseline determination, spectroscopy, non-linearities in the laser spectral emission, achieved signal-to-noise ratio in the spectra, uncertainties in pressure and temperature determinations), the expected precision in the water-vapor mixing-ratio retrievals ranges between 5 and 10% for UT/LS conditions.

PicoSDLA was installed outside the AIDA chamber but inside the chamber thermal enclosure. The laser hygrometer was integrated into a 150 mm diameter cylinder with a length of 1.5 m. This cylinder was located at the bottom of the chamber enclosure (Fig. 2). The chamber inlet line was connected to the cylinder with a flexible connection (Tombac, DN40mm). The output of the cylinder was connected to a vacuum pump and to a flow rate regulator. Electric wires were passed through special vacuum connectors capable of operation at very low temperatures (down to -90°C). The flow rate was adjusted during the calibration campaign with a regulator in the AIDA control room. This regulator was set between 10 to 30 L min⁻¹ (at very low pressure, the real flow rate was less than 10 L min⁻¹). During AquaVIT-1, typical flow rates through the instrument

AMTD

7, 3159–3251, 2014

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



were of 1.0 SLM decreasing to about 0.3 SLM at a total chamber pressure of 100 hPa. At lower pressures, the small flow rates lead to a very slow effective time response of the instrument.

Once installed in the cylinder, PicoSDLA was operated continuously with no intervention during the full campaign. The behavior of the laser hygrometer installed in the cylinder was really satisfying in terms of laser stability and quality of the absorption spectra despite the low temperature encountered (down to -90°C) and the duration of the measurement (10 days, 8 h per day, to compare with the three hours of duration of a standard stratospheric flight).

PicoSDLA is not designed for operation in a closed cell like in the AquaVIT-1 configuration. It is typically operated in an open-path configuration suspended from a stratospheric balloon. This means that the air sample between the laser and detector is renewed every second based on a descent speed of the balloon of $2\text{--}3\text{ms}^{-1}$ and a cell length of 1 m. The equivalent flow rate is more than 30 SLM. The effect seen during AquaVIT-1 calibrations is that, for medium H_2O concentrations and with a sufficient flow rates, PicoSDLA gives a result in agreement with the reference data (see discussions in previous chapters). Nevertheless, one critical issue for PicoSDLA in the present configuration, is the contamination by water vapor outgassing from the instrument itself or from the cylinder; this effect could even be enhanced by the fact that flat heaters were used at different places in the instrument to maintain proper operation of the laser, detector and electronics. It could explain the disagreement with reference values observed at very low water-vapor concentrations. Retrospectively, after looking at the PicoSDLA data, it would have been perhaps more appropriate to locate the laser sensor in the main chamber with other core instruments. The relation between flow rate, heater operations and potential H_2O outgassing effect in the cylinder, is still unclear to us.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



A12 WaSul-Hygro2

During the campaign a prototype of the photoacoustic-spectroscopy based WaSul-Hygro (Bozóki et al., 2003) was operated fully automatically without being supervised by the Hungarian group. It is only after the campaign, during data evaluation it became clear that the WaSul-Hygro1 and WaSul-Hygro2 instruments were almost always operated under overload conditions caused by excessive acoustical noise inside the gas handling system, which ruined most of the measured data. Based on this experience the WaSul-Hygro instruments are now equipped with additional noise-cancelling acoustic buffer volumes, and there is also an overload detection algorithm implemented into its electronics.

A13 Closed-path Laser Hygrometer (CLH)

The University of Colorado Closed-path Laser Hygrometer (CLH) is a tunable-diode laser spectrometer designed to measure tropospheric water vapor and particulate water. This sensor is virtually identical, except for its internal absorption cell, to open-path instruments (JLH) flown by the JPL group (R. May and R. Herman) on the NASA ER-2, DC-8 and WB-57 during past missions. These water vapor spectrometers use near-infrared ($1.37\ \mu\text{m}$, $7306\ \text{cm}^{-1}$) tunable diode laser (TDL) sources that provide about 5 mW of single-mode output. In addition to their small size and dependability, these lasers are attractive because they operate near room temperature, eliminating the need for cryogenes. Instead, a thermoelectric element is used to stabilize the laser temperature to about $15\ ^\circ\text{C}$. The infrared detector – a standard InGaAs sensor – also does not require cryogenes, but is mounted to a thermostated baseplate to maintain a stable operating temperature.

The absorption path is defined by the length of the internal sample cell, a 6 mm I.D. stainless-steel tube, which is capped at both ends with anti-reflection-coated quartz windows. The path length is 27.62 cm, with a single pass between the laser and detector, which is sufficient for tropospheric water (the original intent), but is limited to

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



mixing ratios above about 10 ppm in the upper troposphere and lower stratosphere for 1 s integrations.

On aircraft, the CLH measures behind a forward-facing, heated inlet. The sample flow velocity within the instrument is fixed at less than 5 ms^{-1} using a mass flow controller, considerably slowing the flow compared to the free-stream velocity. This subisokinetic condition enhances sampling of particulates relative to the gas-phase. The particles are evaporated primarily by ram-heating, and the resulting water vapor is measured by absorption. The inlet has been designed such that the instrument is sensitive to particles larger than about $5 \mu\text{m}$ in diameter. Details of the CLH instrument and its performance on research aircraft can be found in Davis et al. (2007a, b).

At the AIDA facility, the CLH was operated in conditions quite different from the normal flight configuration. The instrument sat on a table at ambient pressure and temperature, and was connected to the chamber via a long, heated stainless steel line. The internal mass flow controller was bypassed for most of the experiments and air was pulled through the instrument by attaching to the general vacuum system via an external flow controller. When available, a low delta-P flow controller was used to allow for the highest flow rates possible.

A13.1 Performance during AquaVIT

The CLH experienced a number of problems during the AquaVIT-1 experiments that limited the ability to participate in several of the experiments, especially during the static experiments. Three separate issues affected the instrument performance and the amount of data collected: (1) a small ($\sim 1 \text{ standard cm}^3 \text{ min}^{-1}$ (sccm)) leak of ambient air into the instrument; (2) an instrument computer failure and (3) undermodulation of the laser frequency. The instrument computer was replaced, but a day of data collection was lost in the process. To address the leak, a set of tests was performed to assess the impact on the measured water vapor. As a result, only data with a total instrument flow rate greater than 1 SLM were used; under these conditions the leak contributed less water to the system than detectable within the precision of the instrument.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



calibration of the flow meter used to control the air flow (1 % relative uncertainty). Due to the nonlinearity of the mass flow controller calibration, the flow is corrected by a 3rd order polynomial. Using the full range of the flow meter and bath temperatures of 30 to 60 °C, the humidity range can be extended from 250 ppbv to 50 ppmv.

5

Dedication



This work is dedicated to our dear colleague and driving force in the AquaVIT-1 activity, the excellent and enthusiastic scientist Dr. Cornelius Schiller. He died much too soon at the age of 50 years on 3 March 2012. Starting in 1986 he built a group to measure atmospheric water vapor at the Forschungszentrum Jülich and became one of the world's leading scientists in this research area. He was chairman and member of several international scientific bodies. His research resulted in numerous highly relevant publications on the distribution and the changes in the global water vapor budget of the upper troposphere and stratosphere. With his passing we lose a colleague with the highest personal regard, a mentor for young researchers, and a paragon for all colleagues.

10

15

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Acknowledgements. In ordering the authors, the first eight include the referees (D. W. Fahey, R.-S. Gao, O. Möhler), the organizers (H. Saathoff, C. Schiller, V. Ebert) and those that helped conceive and initiate the intercomparison (MK, TP). The remaining authors from individual participating teams are in alphabetical order.

The success of the AquaVIT-1 campaign derived significantly from the excellent support of the staff scientists and technicians at the AIDA facility led by Ottmar Möhler and Harald Saathoff. Reimar Bauer of Forschungszentrum Jülich, Jülich, Germany, provided support for the AquaVIT-1 wikipage. Travel support for the referees (DWF and RSG) and some investigator groups was provided by the EU project EUROCHAMP, by SPARC – a core project of the world climate research program, and by the Institute for Meteorology and Climate Research of Karlsruhe, Germany. The NASA Upper Atmospheric Research Program also provided travel support for some USA participants. Elliot Weinstock is gratefully acknowledged for his valuable assistance with the HWV measurements. Ulrich Krieger and Martin Brabec are gratefully acknowledged for their help with SnowWhite. Zoltán Bozóki was partially supported by the OTKA Foundation from the Hungarian Research and Technology Innovation Fund (project number: NN109679).

This research and participation of JLH in AquaVIT-1 were supported by the NASA Atmospheric Composition Focus. The research described here was carried out by the Jet Propulsion Laboratory, California Institute of Technology, under a contact with NASA. (Copyright 2009 California Institute of Technology. Government sponsorship acknowledged.)

References

- Altmann, J., Baumgart, R., and Weitkamp, C.: Two-mirror multipass absorption cell, *Appl. Optics*, 20, 995–999, 1981.
- Awtry, A. R., Fisher, B. T., Moffatt, R. A., Ebert, V., and Fleming, J. W.: Simultaneous diode laser based in situ quantification of oxygen, carbon monoxide, water vapor, and liquid water in a dense water mist environment, *Proceedings of the Combustion Institute*, 31, 799–806, 2006.
- Bozóki, Z., Szakáll, M., Mohácsi, Á., Szabó, G., and Bor, Z.: Diode laser based photoacoustic humidity sensors, *Sensor. Actuat. B-Chem.*, 91, 219–226, 2003.

**AquaVIT-1 water
vapor
intercomparison**

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



- Brenninkmeijer, C. A. M., Crutzen, P., Boumard, F., Dauer, T., Dix, B., Ebinghaus, R., Filippi, D., Fischer, H., Franke, H., Frieß, U., Heintzenberg, J., Helleis, F., Hermann, M., Kock, H. H., Koepfel, C., Lelieveld, J., Leuenberger, M., Martinsson, B. G., Miemczyk, S., Moret, H. P., Nguyen, H. N., Nyfeler, P., Oram, D., O'Sullivan, D., Penkett, S., Platt, U., Pucek, M., Ramonet, M., Randa, B., Reichelt, M., Rhee, T. S., Rohwer, J., Rosenfeld, K., Scharffe, D., Schlager, H., Schumann, U., Slemr, F., Sprung, D., Stock, P., Thaler, R., Valentino, F., van Velthoven, P., Waibel, A., Wandel, A., Waschitschek, K., Wiedensohler, A., Xueref-Remy, I., Zahn, A., Zech, U., and Ziereis, H.: Civil Aircraft for the regular investigation of the atmosphere based on an instrumented container: The new CARIBIC system, *Atmos. Chem. Phys.*, 7, 4953–4976, doi:10.5194/acp-7-4953-2007, 2007.
- Brewer, P. J., Milton, M. J. T., Harris, P. M., Bell, S. A., Stevens, M., Scace, G., Abe, H., and Mackrodt, P.: EURAMET1002: international comparability in measurements of trace water vapour, *NPL REPORT AS*, 59, ISSN: 1754–2928, 2011.
- Brown, L. R.: Empirical line parameters of methane from 1.1 to 2.1 micrometers, *J. Quant. Spectrosc. Ra.*, 96, 251–270, 2005.
- Buchholz, B., Kühnreich, B., Smit, H. G. J., and Ebert, V.: Validation of an extractive, airborne, compact TDL spectrometer for atmospheric humidity sensing by blind intercomparison, *Appl. Phys. B.*, 110, 249–262, 2013.
- Buchholz, B., Böse, N., and Ebert, V.: Absolute validation of a diode laser hygrometer via intercomparison with the german national primary water vapor standard, *Appl. Phys. B.*, 1–17, doi:10.1007/s00340-014-5775-4, 2014.
- Cirisan, A., P. Luo, B., Engel, I., Wienhold, F. G., K. Krieger, U., Weers, U., Romanens, G., Levrat, G., Jeannet, P., Ruffieux, D., Philipona, R., Calpini, B., Spichtinger, P., and Peter, T.: Balloon-borne match measurements of mid-latitude cirrus clouds, *Atmos. Chem. Phys. Discuss.*, 13, 25417–25479, doi:10.5194/acpd-13-25417-2013, 2013.
- Davis, S. M., Hallar, A. G., Avallone, L. M., and Engblom, W.: Measurements of ice water content with a tunable diode laser hygrometer: calibration procedure and inlet analysis, *J. Atmos. Ocean. Tech.*, 24, 463, doi:10.1175/JTECH1975.1, 2007a.
- Davis, S. M., Avallone, L. M., Weinstock, E. M., Twohy, C. H., Smith, J. B., and Kok, G. L.: Comparisons of in situ measurements of cirrus cloud ice water content, *J. Geophys. Res.*, 112, D10212, doi:10.1029/2006JD008214, 2007b.
- Dessler, A. E.: Observations of climate feedbacks over 2000–10 and comparisons to climate models, *J. Climate*, 26, 333–342, doi:10.1175/JCLI-D-11-00640.1, 2013.

**AquaVIT-1 water
vapor
intercomparison**

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Dreier, T., Ebert, V., and Schulz, C.: Laser-based Combustion Diagnostics in Encyclopaedia of Analytical Chemistry, 2nd edn., edited by: Meyers, R. A., John Wiley, Chichester, doi:9780470027318/a0715.pub2, ISBN 978-0-470-97333-2, 2011.

Durry, G., Amarouche, N., Joly, L., Liu, X., Zéninari, V., and Parvitte, B.: Laser diode spectroscopy of H₂O at 2.63 micron for atmospheric applications, Appl. Phys. B., 90, 573–580, 2008.

Dyroff, C.: Tunable diode-laser absorption spectroscopy for trace-gas measurements with high sensitivity and low drift, Ph.D. thesis, University of Karlsruhe, ISBN: 978-3-86644-328-0, available at: <http://digbib.ubka.uni-karlsruhe.de/volltexte/1000010030> (last access: 20 March 2014), 2009.

Dyroff, C., Fütterer, D., and Zahn, A.: Compact diode-laser spectrometer ISOWAT for highly sensitive airborne measurements of water-isotope ratios, Appl. Phys. B., 98, 537–548, doi:10.1007/s00340-009-3775-6, 2010.

Ebert, V.: In situ Absorption Spectrometers using near-IR diode lasers and rugged multi-path optics for environmental field measurements, in: Laser Applications to Chemical and Environmental Analysis, OSA Technical Digest, paper WB1, Optical Society of America, Washington DC, 2006.

Ebert, V. and Wolfrum, J.: Absorption spectroscopy, in: Optical Measurements-Techniques and Applications, 2nd corr. edn, edited by: Mayinger, F. and Feldmann, O., Heat and Mass Transfer, Springer Verlag, Heidelberg, München, 227–265, 2001.

Ebert, V., Fitzer, J., Gerstenberg, I., Pleban, K.-U., Pitz, H., Wolfrum, J., Jochem, M., and Martin, J.: Simultaneous laser-based in-situ-detection of oxygen and water in a waste Incinerator for active combustion control purposes, Proceedings of the Combustion Institute, 27, 1301–1308, 1998.

Ebert, V., Teichert, H., Giesemann, C., Saathoff, H., and Schurath, U.: Fiber-coupled in situ-laser absorption spectrometer for the selective detection of water vapor traces down to the ppb-level, VDI Berichte 1863, VDI Verl., Düsseldorf, 197–208, 2004.

Ebert, V., Teichert, H., Giesemann, C., Saathoff, H., and Schurath, U.: Fiber-coupled in situ-laser absorption spectrometer for the selective detection of water vapor traces down to the ppb-level, Tech. Mess., 72, 23–30, 2005.

Ebert, V., Saathoff, H., Lauer, C., Hunsmann, S., and Wagner, S.: Simultaneous, absolute gas-phase and total water detection during cloud formation studies in the AIDA chamber using

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



a dual 1.37 μm TDL-Spectrometer, European Geosciences Union – General Assembly 2008, Vienna, Austria, 13–18 April, EGU2008-A-10066, 2008.

Fernholz, T., Teichert, H., and Ebert, V.: Digital, phase-sensitive detection for in situ diode laser spectroscopy under rapidly changing transmission conditions, *Appl. Phys. B.*, 75, 229–236, 2002.

Fujiwara, M., Shiotani, M., Hasebe, F., Vömel, H., Oltmans, S., Ruppert, P., and Hori-nouchi, T. T. T.: Performance of the meteolabor “Snow White” chilled-mirror hygrometer in the tropical troposphere: comparisons with the Vaisala RS80 A/H-humicap sensors, *J. Atmos. Ocean. Tech.*, 20, 1534–1542, 2003.

Giesemann, C.: Entwicklung und Einsatz eines Diodenlaserspektrometers zum quantitativen In-situ-Nachweis von Methan und Wasser in der Stratosphäre, Ph.D. thesis, Fakultät für Chemie und Geowissenschaften, Universität Heidelberg, 2003.

Giesemann, C., Teichert, H., Saathoff, H., Schurath, U., and Ebert, V.: Fiber-coupled near-infrared diode laser based in situ hygrometer for the detection of water traces in cryogenic aerosol clouds, 4th International Conference on Tunable Diode Laser Spectroscopy, Zermatt, CH, 14–18 July, 2003.

Gurlit, W., Burrows, J. P., Zimmermann, R., Platt, U., Giesemann, C., Wolfrum, J., and Ebert, V.: Light-weight diode laser spectrometer for balloon-borne measurements of water vapor and methane, *Appl. Optics*, 44, 91–102, 2005.

Herriott, D. R., Kogelnik, H., and Kompfner, R.: Off-axis paths in spherical mirror interferometers, *Appl. Optics*, 3, 523–526, 1964.

Hintsä, E. J., Weinstock, E. M., Anderson, J. G., and May, R. D.: On the accuracy of in situ water vapor measurements in the troposphere and lower stratosphere with the Harvard Lyman- α hygrometer, *J. Geophys. Res.*, 104, 8183–8189, 1999.

Hunsmann, S., Wagner, S., Saathoff, H., Möhler, O., Schurath, U., and Ebert, V.: Measurement of the temperature dependence of line strength and pressure broadening coefficients of water absorption lines in the 1.4 μm -Band, *VDI Berichte 1959*, VDI Verl., Düsseldorf, 149–164, 2006.

Jensen, E. J., Smith, J. B., Pfister, L., Pittman, J. V., Weinstock, E. M., Sayres, D. S., Herman, R. L., Troy, R. F., Rosenlof, K., Thompson, T. L., Fridlind, A. M., Hudson, P. K., Cziczo, D. J., Heymsfield, A. J., Schmitt, C., and Wilson, J. C.: Ice supersaturations exceeding 100% at the cold tropical tropopause: implications for cirrus formation and dehydration, *Atmos. Chem. Phys.*, 5, 851–862, doi:10.5194/acp-5-851-2005, 2005.

**AquaVIT-1 water
vapor
intercomparison**

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Jensen, E. J., Pfister, L., Bui, T. V., Lawson, P., Baker, B., Mo, Q., Baumgardner, D., Weinstock, E. M., Smith, J. B., Moyer, E. J., Hanisco, T. F., Sayres, D. S., Clair, J. M. St., Alexander, M. J., Toon, O. B., and Smith, J. A.: Formation of large ($\sim 100\ \mu\text{m}$) ice crystals near the tropical tropopause, *Atmos. Chem. Phys.*, 8, 1621–1633, doi:10.5194/acp-8-1621-2008, 2008.

Khaykin, S. M., Engel, I., Vömel, H., Formanyuk, I. M., Kivi, R., Korshunov, L. I., Krämer, M., Lykov, A. D., Meier, S., Naebert, T., Pitts, M. C., Santee, M. L., Spelten, N., Wienhold, F. G., Yushkov, V. A., and Peter, T.: Arctic stratospheric dehydration – Part 1: Unprecedented observation of vertical redistribution of water, *Atmos. Chem. Phys.*, 13, 11503–11517, doi:10.5194/acp-13-11503-2013, 2013.

Khaplanov, M., Astakhov, V., Lukjanov, A., Kretova, M., and Yushkov, V.: Fluorescent hygrometer for middle atmosphere measurements, *Proc. 19th Annual European Meeting on Atmospheric Studies by Optical Methods*, 540–545, 1992.

Kley, D.: $\text{Ly}(\alpha)$ absorption cross-section of H_2O and O_2 , *J. Atmos. Chem.*, 2, 203–210, 1984.

Krämer, M., Schiller, C., Afchine, A., Bauer, R., Gensch, I., Mangold, A., Schlicht, S., Spelten, N., Sitnikov, N., Borrmann, S., de Reus, M., and Spichtinger, P.: Ice supersaturations and cirrus cloud crystal numbers, *Atmos. Chem. Phys.*, 9, 3505–3522, doi:10.5194/acp-9-3505-2009, 2009.

Lauer, C., Weber, D., Wagner, S., and Ebert, V.: Calibration free measurement of atmospheric methane background via tunable diode laser absorption spectroscopy at $1.6\ \mu\text{m}$, *Laser Applications to Chemical, Security and Environmental Analysis (LACSEA)*, St. Petersburg, Florida, USA, 17–20 March 2008, Paper no. LMA2 and *Dresdner Beiträge zur Sensorik*, Hrsg. G. Gerlach, and P. Hauptmann, Band 29, ISBN 978-3-940046-45-1, 7–10, 2007.

Mackrodt, P.: A new attempt on a coulometric trace humidity generator, *Int. J. Thermophys.*, 33, 1520–1535, 2012.

Mangold, A., Wagner, R., Saathoff, H., Schurath, U., Giesemann, C., Ebert, V., Krämer, M., and Möhler, O.: Experimental investigation of ice nucleation by different types of aerosols in the aerosol chamber AIDA: implications to microphysics of cirrus clouds, *Meteorol. Z.*, 14, 485–497, 2005.

May, R. D.: Open-path, near-infrared tunable diode laser spectrometer for atmospheric measurements of H_2O , *J. Geophys. Res.*, 103, 19161–19172, 1998.

May, R. D. and Webster, C. R.: Data processing and calibration for tunable diode laser harmonic absorption spectrometers, *J. Quant. Spectrosc. Ra.*, 49, 335–347, 1993.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



- Meyer, J.: Vergleich und Charakterisierung stratosphärisch messender in-situ Hygrometer, Diploma thesis, University Aachen, 2008 (in German).
- Meyer, J., et al.: Calibration of a fluorescence hygrometer at low water mixing ratios, in preparation, Atmos. Meas. Tech. Discuss., in preparation, 2014.
- 5 Miloshevich, L., Vömel, H., Whiteman, D., Lesht, B., Schmidlin, F., and Russo, F.: Absolute accuracy of water vapor measurements from six operational radiosonde types launched during AWEX-G, and implications for AIRS validation, J. Geophys. Res., 111, doi:10.1029/2005JD006083, 2006.
- Möhler, O., Stetzer, O., Schaefers, S., Linke, C., Schnaiter, M., Tiede, R., Saathoff, H.,
10 Krämer, M., Mangold, A., Budz, P., Zink, P., Schreiner, J., Mauersberger, K., Haag, W., Kärcher, B., and Schurath, U.: Experimental investigation of homogeneous freezing of sulphuric acid particles in the aerosol chamber AIDA, Atmos. Chem. Phys., 3, 211–223, doi:10.5194/acp-3-211-2003, 2003.
- Möhler, O., Büttner, S., Linke, C., Schnaiter, M., Saathoff, H., Stetzer, O., Wagner, R.,
15 Krämer, M., Mangold, A., Ebert, V., and Schurath, U.: Effect of sulfuric acid coating on heterogeneous ice nucleation by soot aerosol particles, J. Geophys. Res., 110, D11210, doi:10.1029/2004JD005169, 2005.
- Möhler, O., Benz, S., Saathoff, H., Schnaiter, M., Wagner, R., Schneider, J., Walter, S., Ebert, V.,
20 and Wagner, S.: The effect of organic coating on the heterogeneous ice nucleation efficiency of mineral dust aerosol, Environ. Res. Lett., 3, 025007, doi:10.1088/1748-9326/3/2/025007, 2008a.
- Möhler, O., Georgakopoulos, D. G., Morris, C. E., Benz, S., Ebert, V., Hunsmann, S.,
25 Saathoff, H., Schnaiter, M., and Wagner, R.: Heterogeneous ice nucleation activity of bacteria: new laboratory experiments at simulated cloud conditions, Biogeosciences, 5, 1425–1435, doi:10.5194/bg-5-1425-2008, 2008b.
- Murphy, D. M. and Koop, T.: Review of the vapour pressures of ice and supercooled water for atmospheric applications, Q. J. Roy. Meteorol. Soc., 131, 1539–1565, 2005.
- Murray, B. J., Wilson, T. W., Dobbie, S., Cui, Z., Möhler, O., Schnaiter, M., Wagner, R., Benz, S.,
30 Niemand, M., Saathoff, H., Ebert, V., and Kärcher, B.: Glassy aerosol modify cirrus cloud properties, Nat. Geosci., 3, 233–237, 2010.
- Peter, T., Marcolli, C., Spichtinger, P., Corti, T., Baker, M. B., Koop, T.: When dry air is too humid, Science, 314, 1399–1402, doi:10.1126/science.1135199, 2006.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Peter, T., Krämer, M., and Möhler, O.: ETH, Upper Tropospheric Humidity: a Report on an International Workshop, Karlsruhe, Germany, 12–15 June 2007, SPARC Newsletter No. 30., available at: <http://www.sparc-climate.org/publications/newsletter/> (last access: 20 March 2014), 2008.

5 Rollins, A. W., Thornberry, T. D., Gao, R. S., Smith, J. B., Sayres, D. S., Sargent, M. R., Schiller, C., Krämer, M., Spelten, N., Hurst, D. F., Jordan, A. F., Hall, E. G., Vömel, H., Diskin, G. S., Podolske, J. R., Christensen, L. E., Rosenlof, K. H., Jensen, E. J., and Fahey, D. W.: Evaluation of UT/LS hygrometer accuracy by intercomparison during the NASA MACPEX mission, *J. Geophys. Res.*, doi:10.1002/2013JD020817, 2014.

10 Rothman, L. S., Jacquemart, D., Barbe, A., Chris Benner, D., Birk, M., Brown, L. R., Carleer, M. R., Chackerian, C., Chance, K., Coudert, L. H., Dana, V., Devi, V. M., Flaud, J.-M., Gamache, R. R., Goldman, A., Hartmann, J.-M., Jucks, K. W., Maki, A. G., Mandin, J.-Y., Massie, S. T., Orphal, J., Perrin, A., Rinsland, C. P., Smith, M. A. H., Tennyson, J., Tolchenov, R. N., Toth, R. A., Vander Auwera, J., Varanasi, P., and Wagner, G.: The HITRAN 2004 molecular spectroscopic database, *J. Quant. Spectrosc. Ra.*, 96, 139–204, 2005.

15 Sayres, D. S., Moyer, E. J., Hanisco, T. F., St. Clair, J. M., Keutsch, F. N., O'Brien, A., Allen, N. T., Lapson, L., Demusz, J. N., Rivero, M., Martin, T., Greenberg, M., Tuozzolo, C., Engel, G. S., Kroll, J. H., Paul, J. B., and Anderson, J. G.: A new cavity based absorption instrument for detection of water isotopologues in the upper troposphere and lower stratosphere, *Rev. Sci. Instrum.*, 80, 044102, doi:10.1063/1.3117349, 2009.

20 Schlicht, S.: Untersuchungen zur Wasserpartitionierung in Wolken, Ph.D. thesis, University of Aachen, Germany, 2006 (in German).

Schulz, C., Dreizler, A., Ebert, V., and Wolfrum, J.: Combustion diagnostics, in: *Springer Handbook of Experimental Fluid Dynamics*, edited by: Tropea, C., Foss, J., and Yarin, A., 1241–1316, 2007.

25 Schmidlin, F.: Relative humidity measurements from a chilled-mirror technique flown on low cost radiosonde, American Meteorological Society meeting, American Meteorological Society, Dallas, 1999.

Skrotzki, J., Habig, J. C., and Ebert, V.: Integrative fitting of absorption line profiles with high accuracy, robustness, and speed, *Appl. Phys. B*, doi:10.1007/s00340-013-5706-9, 2013.

30 Smith, J. B.: The sources and significance of stratospheric water vapor: mechanistic studies from equator to pole, Ph.D. thesis, Harvard University, Cambridge, MA, 2012.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Solomon, S., Rosenlof, K. H., Portmann, R. W., Daniel, J. S., Davis, S. M., Sanford, T. J., and Plattner, G.-K.: Contributions of stratospheric water vapor to decadal changes in the rate of global warming, *Science*, 327, 1219–1223, 2010.

SPARC Assessment of Upper Tropospheric and Stratospheric Water Vapour, edited by: Kley, D., Russell III, J. M., and Phillips, C., World Climate Research Programme, WCRP-113, available at: http://www.atmosp.physics.utoronto.ca/SPARC/WAVASFINAL_000206/WWW_wavas/Cover.html (last access: 20 March 2014), 2000.

Teichert, H., Fernholz, T., and Ebert, V.: In situ measurement of CO, H₂O and gas temperature in a lignite-fired power-plant, *Appl. Optics*, 42, 2043–2051, 2003.

Trenberth, K. E., Jones, P. D., Ambenje, P., Bojariu, R., Easterling, D., Klein Tank, A., Parker, D., Rahimzadeh, F., Renwick, J. A., Rusticucci, M., Soden, B., and Zhai, P.: Observations: Surface and Atmospheric Climate Change, in: *Climate Change 2007: The Physical Science Basis. Contribution of Working Group I to the Fourth Assessment Report of the Intergovernmental Panel on Climate Change*, edited by: Solomon, S., Qin, D., Manning, M., Chen, Z., Marquis, M., Averyt, K. B., Tignor, M., and Miller, H. L., Cambridge University Press, Cambridge, United Kingdom and New York, NY, USA, 2007.

Troy, R. F.: Field studies of ice supersaturations in the tropical tropopause layer, Ph.D. thesis, University of California at Los Angeles, Los Angeles, CA, 177 pp., 2007.

Vaughan, G., Cambridge, C., Dean, L., and Phillips, A. W.: Water vapour and ozone profiles in the midlatitude upper troposphere, *Atmos. Chem. Phys.*, 5, 963–971, doi:10.5194/acp-5-963-2005, 2005.

Vömel, H.: Report from the NDACC Meeting on Atmospheric Water Vapour Measurement, edited by: Braathen, G., University of Bern, Bern, Switzerland, available at: www.iapmw.unibe.ch/research/collaboration/ndsc-microwave/workshop/2006 (last access: 20 March 2014), 2006.

Vömel, H., Fujiwara, M., Shirokami, M., Hasebe, F., Oltmans, S., and Barnes, J.: The behavior of Snow White chilled-mirror hygrometer in extremely dry conditions, *J. Atmos. Ocean. Tech.*, 20, 1560–1567, 2003.

Vömel, H., David, D. E., and Smith, K.: Accuracy of tropospheric and stratospheric water vapor measurements by the cryogenic frost point hygrometer: instrumental details and observations, *J. Geophys. Res.*, 112, D08305, doi:10.1029/2006JD007224, 2007a.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Vömel, H., Yushkov, V., Khaykin, S., Korshunov, L., Kyrö, E., and Kivi, R.: Intercomparisons of stratospheric water vapour sensors: FLASH-B and NOAA/CMDL frost-point hygrometer, *J. Atmos. Ocean. Tech.*, 24, 941–952, 2007b.

Wagner, R., Benz, S., Bunz, H., Möhler, O., Saathoff, H., Schnaiter, M., Leisner, T., and Ebert, V.: Infrared optical constants of highly diluted sulfuric acid solution droplets at cirrus temperatures, *J. Phys. Chem. A*, 112, 11661–11676, doi:10.1021/jp8066102, 2008.

Webster, C. R., May, R. D., Trimble, C. A., Chave, R. G., and Kendall, J.: Aircraft (ER-2) laser infrared absorption spectrometer (ALIAS) for in-situ stratospheric measurements of HCl, N₂O, CH₄, NO₂, and HNO₃, *Appl. Optics*, 33, 454–472, 1994.

Weinstock, E. M., Hintsä, E. J., Dessler, A. E., Oliver, J. F., Hazen, N. L., Demusz, J. N., Allen, N. T., Lapson, L. B., and Anderson, J. G.: New fast response photofragment fluorescence hygrometer for use on the NASA ER-2 and the Perseus remotely piloted aircraft, *Rev. Sci. Instrum.*, 65, 3544–3554, 1994.

Weinstock, E. M., Smith, J. B., Sayres, D., Spackman, J. R., Pittman, J. V., Allen, N., Demusz, J., Greenberg, M., Rivero, M., Solomon, L., and Anderson, J. G.: Measurements of the total water content of cirrus clouds, Part I: Instrument details and calibration, *J. Atmos. Ocean. Tech.*, 23, 1397–1409, doi:10.1175/JTECH1928.1, 2006a.

Weinstock, E. M., Smith, J. B., Sayres, D., Pittman, J. V., Allen, N., and Anderson, J. G.: Measurements of the total water content of cirrus clouds, Part II: Instrument performance and validation, *J. Atmos. Ocean. Tech.*, 23, 1410–1421, doi:10.1175/JTECH1929.1, 2006b.

Weinstock, E. M., Smith, J. B., Sayres, D. S., Pittman, J. V., Spackman, J. R., Hintsä, E. J., Hanisco, T. F., Moyer, E. J., St. Clair, J. M., Sargent, M. R., and Anderson, J. G.: Validation of the Harvard Lyman- α in situ water vapor instrument: implications for the mechanisms that control stratospheric water vapor, *J. Geophys. Res.*, 114, D23301, doi:10.1029/2009JD012427, 2009.

White, J. U.: Long optical path of large aperture, *J. Opt. Soc. Am.*, 32, 285–288, 1942.

Yushkov, V., Merkulov, S., and Astakhov, V.: Optical balloon hygrometer for upper stratosphere and stratosphere water vapour measurements, in optical remote sensing of the atmosphere and clouds, edited by: Wang, J., Wu, B., Ogawa, T., Guans, Z.-H., *Proc. SPIE*, 3501, 439–445, 1998.

Yushkov, V., Sitnikov, N., Zaitcev, I., Pommereau, J.-P., and Garnier, A.: Stratospheric water vapor measurements in the winter arctic with optical fluorescence hygrometer on short and long duration balloons, in: *Proceedings of the 15th ESA Symposium on European Rocket and*

Balloon programmes and Related Research, Biarritz, France, ESA SP-471, 28–31, edited by: Warmbein, B., ESA, 263–268, 2001.

5 Zöger, M., Afchine, A., Eicke, N., Gerhards, M.-T., Klein, E., McKenna, D. S., Mörschel, U., Schmidt, U., Tan, V., Tuitjer, F., Woyke, T., Schiller, C.: Fast in situ stratospheric hygrometers: a new family of balloonborne and airborne Lyman- α photofragment fluorescence hygrometers, *J. Geophys. Res.*, 104, 1807–1816, 1999.

AMTD

7, 3159–3251, 2014

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water
vapor
intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Table 1. AquaVIT-1 instruments, participants, and institutes.

Instrument (Acronym) (technique) ^a	Type ^b	Participants	Institute
Core instruments			
AIDA-PCI-in-cloud-TDL (APicT) (Tunable Diode Laser technique (TDL))	NE	Volker Ebert ^{1,*} , Christian Lauer ¹ , Stefan Hunsmann ¹ , Harald Saathoff ² , Steven Wagner ^{1,*}	PCI, University of Heidelberg ¹ , Now PTB ² , Now Technical University Darmstadt ² , Karlsruhe Institute of Technology (KIT) ² , Germany.
Cryogenic Frostpoint Hygrometer (CFH) (frostpoint)	E	Holger Vömel	NOAA Earth System Research Laboratory and University of Colorado, Boulder, CO USA (now at Deutscher Wetterdienst, Lindenberg, Germany).
Fast In situ Stratospheric Hygrometer (FISH-1 and FISH-2) (Lyman- α)	E	Martina Krämer, Cornelius Schiller, Armin Afchine, Reimar Bauer, Jessica Meyer, Nicole Spelten, Andres Thiel, Miriam Kübbeler	Forschungszentrum Jülich, Jülich, Germany
FLUorescent Advanced Stratospheric Hygrometer for Balloon (FLASH-B1 and FLASH-B2) (Lyman- α)	NE	Sergey Khaykin, Leonid Korshunov	Central Aerological Observatory, Moscow, Russia
Harvard Water Vapor (HWV) (Lyman- α)	E	Jessica Smith, Elliot Weinstock	Harvard University, Cambridge, MA, USA
JPL Laser Hygrometer (JLH) (TDL)	NE	Robert Herman, Robert Troy, Lance Christensen	Jet Propulsion Laboratory, California Institute of Technology, Pasadena, CA, USA
Non-Core instruments			
MBW-373LX (frostpoint)	E	Harald Saathoff, Robert Wagner	KIT, Karlsruhe, Germany
SnowWhite (frostpoint)	NE	Frank Wienhold, Ulrich Krieger, Martin Brabec	Eidgenössische Technische Hochschule-ETH, Zürich, Zurich, Switzerland
ISOWAT (TDL)	E	Christoph Dyroff	KIT, Karlsruhe, Germany
Open-path Jülich Stratospheric Experiment (OJSTER) (TDL)	E	Martina Krämer, Cornelius Schiller, Armin Afchine, Reimar Bauer, Jessica Meyer, Nicole Spelten, Andres Thiel, Miriam Kübbeler	Forschungszentrum Jülich, Jülich, Germany
PicoSDLA (TDL)	NE	Georges Durry, Nadir Amarouche, Jacques Deleglise, Fabien Frerot	University of Reims, Champagne-Ardenne and Institut National des Sciences de l'Univers/Centre National de la Recherche Scientifique (INSU/CNRS), France
WaSul-Hygro2 (photoacoustic)	E	Zoltan Bozóki, Árpád Mohácsi	University of Szeged, Hilase Ltd., Szeged, Hungary
Closed-path Laser Hygrometer (CLH) (TDL)	E	Linnea Avallone, Sean Davis	University of Colorado, Boulder, CO, USA

^a Instrument descriptions in Appendix A.^b Instrument type based on standard use configuration in atmospheric or laboratory measurements: extractive sampling (E) and non-extractive sampling (NE).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Table 2. Core and non-core instrument uncertainty values.

Instrument name (technique) ^a	Uncertainty in final data ^b
Core instruments	
APicT (Tunable Diode Laser technique (TDL))	Accuracy: < 5 %; Precision: from $\approx 10\%$ at 0.25 ppm H ₂ O to 1% -0.1 % at > 20 ppm H ₂ O; for H ₂ O < 20 ppm: noise level approx. 0.025 ppm (1 σ at $\Delta t = 2$ s) at 80 m path
CFH (frostpoint)	10 % @ H ₂ O \leq 5 ppm 4 % @ H ₂ O > 5 ppm
FISH-1 and FISH-2 (Lyman- α)	H ₂ O \leq 20 ppm, $P > 80$ hPa: 6 % + 0.1 ppm (FISH-2) 6–8 % + 0.15 ppm (FISH-1)
FLASH-B1 and FLASH-B2 (Lyman- α)	$\pm(10\% + 0.1 \text{ ppm})$ @ 10 hPa < $P < 300$ hPa, H ₂ O > 3 ppm $\pm(20\% + 0.1 \text{ ppm})$ @ H ₂ O < 3 ppm
HWV ^b (Lyman- α)	$\pm 5\% + 0.53/-0.28$ ppm @ $P > 100$ hPa $\pm 10\% + 0.53/-0.28$ ppm @ $P \leq 100$ hPa
JLH (TDL)	10 % + 0.15 ppm (1 s) 10 % + 0.05 ppm (10 s)
Non-Core instruments	
MBW-373LX (frostpoint)	Accuracy $\pm 3\%$ and precision $\pm 1.5\%$ (± 0.1 °C frost point temperature) at pressures > 150 hPa and frostpoint temperatures > -70 °C. Unknown systematic errors at lower pressures and temperatures
SnowWhite (frostpoint)	5 % accuracy for mixing ratios > 10 ppm
ISOWAT (TDL)	4 % precision for H ₂ ¹⁸ O Accuracy suffered from instabilities of the optical alignment
OJSTER (TDL)	2 % accuracy for H ₂ O > 100 ppmv, 5 % accuracy for H ₂ O < 100 ppmv, 2 % precision for 1 s, 4 % for 100 Hz (Due to a contamination problem, the detection limit varied during the experiments)
PicoSDLA (TDL)	5 % to 10 % accuracy (measurement time of 800 ms)
WaSul-Hygro2 (photoacoustic)	Approaches 1 % above 100 ppm. Lower accuracy at lower mixing ratios
Closed-path Laser Hygrometer (CLH) (TDL)	15 % above 15 ppm

^aPrecision and accuracy values ($\pm 1\sigma$) provided by instrument investigators. See Table 4 for precision values derived from AquaVIT-1 datasets.

^b These values reflect instrument performance specific to the AIDA facility and may differ from values reported in the literature for atmospheric measurements.

AquaVIT-1 water
vapor
intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

**Table 3.** Details of AquaVIT-1 static segments used in the accuracy and precision evaluations.

Oct 2007	# ^a	Press. (hPa) ^b	Temp (K) ^b	Water vapor (ppm) ^c	Start time (UTs)	Stop time (UTs)	Length (s)
15th	1	100	243	12.22	28 500	30 300	1800
	2	200	243	5.16	32 400	34 200	1800
	3	500	243	1.91	36 000	39 600	3600
	4	200	243	3.58	43 500	45 300	1800
	5	100	242	6.05	47 100	48 900	1800
	6	50	242	10.41	50 700	52 500	1800
16th	7	200	225	33.87	30 900	32 700	1800
	8	500	224	12.8	35 700	38 400	2700
	9	<i>200</i>	<i>223</i>	<i>15.59</i>	<i>41 880</i>	<i>43 080</i>	<i>1200</i>
	10	100	223	21.11	44 880	46 680	1800
	11	50	223	31.09	48 600	50 400	1800
17th	12	100	214	88.99	26 880	28 680	1800
	13	<i>200</i>	<i>214</i>	<i>46.79</i>	<i>29 880</i>	<i>31 380</i>	<i>1500</i>
	14	300	214	31.96	34 200	37 800	3600
	15	200	213	44.44	41 280	43 080	1800
	16	<i>100</i>	<i>213</i>	<i>79.42</i>	<i>45 480</i>	<i>47 280</i>	<i>1800</i>
	17	<i>50</i>	<i>213</i>	<i>151.57</i>	<i>49 500</i>	<i>51 300</i>	<i>1800</i>
18th	18	<i>120</i>	<i>197</i>	<i>1.62</i>	<i>28 800</i>	<i>30 600</i>	<i>1800</i>
	19	<i>200</i>	<i>197</i>	<i>0.97</i>	<i>31 980</i>	<i>33 780</i>	<i>1800</i>
	20	<i>300</i>	<i>197</i>	<i>0.64</i>	<i>36 300</i>	<i>38 100</i>	<i>1800</i>
	21	<i>200</i>	<i>196</i>	<i>0.87</i>	<i>40 080</i>	<i>41 880</i>	<i>1800</i>
	22	80	196	1.77	45 000	48 600	3600
	23	50	196	2.59	50 400	52 200	1800
19th	24	80	186	1.64	23 400	25 200	1800
	25	120	186	1.03	27 480	29 280	1800
	26	<i>200</i>	<i>186</i>	<i>0.64</i>	<i>30 300</i>	<i>31 800</i>	<i>1500</i>
	27	<i>300</i>	<i>186</i>	<i>0.44</i>	<i>33 900</i>	<i>35 700</i>	<i>1800</i>
	28	<i>500</i>	<i>186</i>	<i>0.25</i>	<i>38 100</i>	<i>39 900</i>	<i>1800</i>
	29	<i>200</i>	<i>185</i>	<i>0.8</i>	<i>43 500</i>	<i>44 700</i>	<i>1200</i>
	30	<i>80</i>	<i>185</i>	<i>1.59</i>	<i>47 850</i>	<i>49 080</i>	<i>1500</i>
	31	<i>50</i>	<i>185</i>	<i>2.34</i>	<i>50 400</i>	<i>51 300</i>	<i>900</i>

^a Segment number. Segments not in *italics* were used in both accuracy and precision evaluations. Segments in *italics* were used only in the accuracy evaluation.

^b Average measured chamber conditions over the segment. See Fig. 5 for example segments.

^c Reference value derived from the linear fits to the core instrument time series (see Sect. 6.2c for details).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



Table 4. Experimental upper limits of instrument precision derived from the AquaVIT-1 inter-comparison data for selected segments during the static experiments^a.

Segment	5	22	24	25	Average
Reference H ₂ O (ppm)	6.1	1.8	1.6	1.0	–
Core instruments					
APicT	0.070 (1 s) 0.062 (5 s)	0.045 0.041	0.12 0.10	0.14 0.12	0.094 0.081
CFH	–	0.050 0.051	0.072 0.072	0.042 0.041	0.055 0.055
FISH-1	0.24 0.13	–	–	0.16 0.11	0.20 0.12
FISH-2	0.077 0.042	0.041 0.025	0.046 0.025	0.039 0.023	0.051 0.029
FLASH-B1	–	–	–	–	–
	0.11	0.099	–	0.29	0.17
HWV	0.083	–	–	–	0.083
	–	–	–	–	–
JLH	0.10 0.082	0.064 0.044	0.069 0.046	0.049 0.034	0.071 0.052
Non-core instruments					
MBW-373LX	0.022 (1 s) 0.020 (5 s)	–	–	–	0.022 0.020
SnowWhite	–	–	–	4.1 4.2	4.1 4.2
ISOWAT	0.15 0.13	–	–	–	0.17 0.13
OJSTER	0.75 0.67	–	–	–	0.75 0.67
PicoSDLA	0.40 0.39	0.087 0.081	–	0.38 0.38	0.29 0.28

^a Precision values are in ppm of water vapor. Segments were chosen to meet conditions of: (1 < water vapor mixing ratio < 10 ppm) and (70 < AIDA chamber pressure < 150 hPa) in order to represent typical UT/LS values. For each instrument, the top (bottom) row shows precision values for 1 s (5 s) measurement intervals. Segment details are shown in Table 3. Precision is defined as the standard deviation (1σ) of the Gaussian fit, P , to the differences from the reference values ($P = A \exp[-(x - \mu)^2 / 2\sigma^2]$ where A is a normalization factor, x is the measured value, μ is the reference value, and σ is the standard deviation).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Table A1. CFH specifications.

Vertical measurement range	0 to ~ 28 km
Dew point or frost point detection range	> +25°C to < -95°C
Mixing ratio detection range	> 25 000 ppm to < 0.8 ppm
Dew point or frost point uncertainty	0.5 °C
Mixing ratio uncertainty	~ 3.5% (surface), ~ 9% (tropopause), ~ 11% (28 km)
Vertical resolution	≲ 10 m (troposphere) to ≲ 100 m (stratosphere)
CFH instrument weight	400 g
CFH + ECC + radiosonde payload launch weight	1500 g

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion






Fig. 1. Profiles of water vapor mixing ratios from three core instruments included in the AquaVIT-1 intercomparison (CFH, HWV, and JLH), the Aircraft Laser Infrared Absorption Spectrometer (ALIAS) (Webster et al., 1994), and the Integrated Cavity Output Spectroscopy (ICOS) instrument (Sayres et al., 2009). All measurements were made in situ near San Jose, Costa Rica on 1 February 2006. The CFH measurements were made from a small balloon launched near the San Jose airport. The other measurements were made on board the NASA WB-57F high-altitude research aircraft on descent into the San Jose airport. Top left panel: profiles of temperature (bottom axis) and O_3 (top axis) vs. ambient pressure measured from the WB-57F (red) and balloon (black) showing good correspondence between air masses sampled from these different platforms. Top middle panel: water-vapor mixing ratios measured by the balloon and aircraft instruments. Saturation mixing ratios (blue dashed line) were calculated using balloon temperature and pressure. Top right panel: vertical profiles showing differences between each of the WB-57F instruments and the FP (CFH) balloon instrument. Bottom panel: deviations from a mean profile (mean profile = equally weighted HWV, ICOS, JLH ALIAS and CFH) for HWV (red), ICOS (green), JLH (violet), ALIAS (black) and CFH (gold)). From Rollins et al. (2013).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract Introduction

Conclusions References

Tables Figures

◀ ▶

◀ ▶

Back Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 Instrument Configuration




Fig. 2. AIDA chamber configuration of the AquaVIT-1 instruments in Table 1.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

AMTD

7, 3159–3251, 2014

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.



Fig. 3. Left: customized extractive sampling probes inside the chamber. This cluster of three probes was located on the lower right-hand chamber wall (lower right hand in Fig. 2), and provided chamber airflow to several instruments. These probes are either 10 mm or 16 mm inside diameter stainless-steel tubes surrounded by a sealed heating mantle and extending 35 cm into the chamber. Right: the JPL TDL open-path instrument mounted inside the chamber. The instrument electronic module remained outside the chamber.

[Title Page](#)[Abstract](#)[Introduction](#)[Conclusions](#)[References](#)[Tables](#)[Figures](#)[◀](#)[▶](#)[◀](#)[▶](#)[Back](#)[Close](#)[Full Screen / Esc](#)[Printer-friendly Version](#)[Interactive Discussion](#)

AquaVIT-1 water
vapor
intercomparison

D. W. Fahey et al.




Fig. 4. Time series of temperatures **(a)**, pressures **(b)** and the relationship of chamber pressure and temperature with water vapor mixing ratios **(c and d)** for all segments during the experiments (also see Table 3). The dashed lines indicate the nominal range of conditions observed by the HWV instrument in tropical and midlatitude flights during the last decade. During the two lowest-temperature experiments (18 and 19 October), humidity in the chamber was controlled by ice cover on the chamber walls.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. 5. Example time series of AIDA water vapor mixing ratios (1 s averages) as reported by the core instruments during the static experiments on 15 October (top) and 16 October (bottom). Data from the FLASH instruments are not included in the reference value (see text). The chamber pressures and temperatures are given in the top panel of each graph. The black lines are the reference functions for each segment and indicate the time intervals used in the intercomparison. The average of each reference function yields the reference value for a segment, which are given in Table 3.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. 6. Summary plot of the core and non-core instrument intercomparison results for the 5 day static experiment series. The instruments are identified on the left of each panel. Each panel is labeled by the range of reference water-vapor mixing ratios. The chamber pressure range is indicated by the symbol color. The symbols represent the average difference within a segment from the reference water vapor value for that segment. Data from the FLASH instruments are not included in the reference value (see text). The segments varied from 900 s to 3600 s in length with most being 1800 s (Table 3). The differences are plotted on two separate log scales for values more than 1 % above or below the reference value. Differences equal to or less than 1 % are plotted at a value of 1 %. The average of all segments for an instrument is shown with the circle/plus symbol. Data are not available for all segments for each instrument and some instruments have no data for the lower mixing ratio ranges. Instrument names in parentheses in a panel have no results in the associated mixing ratio range.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. 7. Summary plot of static experiment results for core instruments shown as the % difference between values from the listed instruments and the corresponding reference values for three ranges of the reference values. Data from the FLASH instruments are not included in the reference value (see text). A symbol represents the result for the segment number noted near the top axis. The use of a small symbol size for a segment indicates that the accuracy and precision cannot be defined based on the relationships in Table 3. Segment details are provided in Table 4. Colors represent the AIDA chamber average pressure during the segment. Differences less than or equal to 1 % are plotted as a 1 % value.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

AquaVIT-1 water
vapor
intercomparison

D. W. Fahey et al.




Fig. 8. Example summary plot of measurement precision in the static experiments. The results shown are for the FISH-2 instrument. Top: gaussian fits to the probability distribution functions (PDFs) of differences from the reference value function derived for each segment from the core instruments. The PDFs are derived from the 1 s time series data. Legend boxes indicate the reference water vapor value for each segment (see Table 3). The PDFs and fits are arbitrarily offset in the vertical for clarity. Bottom: plot of mean (symbols) and maximum and minimum differences from the reference values (thin vertical lines) and $1\text{-}\sigma$ precision (thick vertical lines) as defined in the footnote in Table 4. Color indicates the chamber pressure range in both top and bottom panels.




Fig. A1. Schematic of the general setup of the APiCT spectrometer: APiCT consists of three parts placed in different segments of the AIDA chamber facility. The parts are connected either via fiber optic cables or free space laser beams. The outer section, including most electronics, data acquisition components, and the diode laser, can be directly operated by the user. The middle section contains a cryogenic transfer optic, including detectors and fiber optics and is usually only serviced at sufficiently high temperatures (above about -30°C). The inner section is connected to the middle section via free space optics and mainly encompasses the open-path white cell, with opto-mechanic components directly attached to the inner walls of the AIDA chamber (Wagner et al., 2008; Giesemann et al., 2003).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page	
Abstract	Introduction
Conclusions	References
Tables	Figures
◀	▶
◀	▶
Back	Close
Full Screen / Esc	
Printer-friendly Version	
Interactive Discussion	



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. A2. (Top) Differences between APiCT water vapor partial pressures and those calculated from ice-saturation at measured temperatures (top axis) for AquaVIT-1 dynamic experiment segments. (Bottom) Water vapor partial pressures vs. dynamic segment number as calculated (p_w ice sat) and as measured by APiCT (p_w APiCT). Also shown are water-vapor mixing ratios in the chamber (right hand axis).

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion






Fig. A3. Schematic of the CFH sonde. The vertical dashed lines indicate the inlet and outlet tubes. The airflow inside the tubes is downward during balloon ascent and upward during descent. Only the lens and the mirror are exposed to the airflow inside the tube. The microprocessor controller regulates the mirror temperature such that the bulk reflectivity of the frost covered mirror remains constant (Vömel et al., 2007a).

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

⏪

⏩

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. A4. Summary plot of the static experiment results for FLASH-B1 data in the entire range of water vapor mixing ratios with the data from experiments 3 and 7 recalculated using correct SLC values.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. A5. (Left) The HWV instrument as it is mounted in the WB57F. (Right) Schematic illustrating the subsystems that control the flow of air through the instrument and acquire simultaneous measurements of temperature, pressure, and velocity in the duct.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures



Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. A6. A schematic of the HWV calibration system and heart of the Lyman- α detection axis with the components necessary for radial absorption labeled. The calibration system is routinely used in the laboratory and in the field. The bubbler serves as both a water vapor source and a primary standard. In the laboratory, axial absorption of 121.6 nm light down the 1 m length of the duct is a secondary reference standard. Radial absorption of 121.6 nm light provides independent verification of the fluorescence measurement accuracy in flight.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. A7. Linearity of Lyman- α detection method over a range of pressures and mixing ratios representative of chamber conditions on 16 and 19 October. The top panel shows selected dilutions from calibrations executed at low pressures, ~ 60 to 120 hPa (blue), and at low water vapor mixing ratios, 0 to 5 ppmv (red). The fluorescence data, evaluated using calibration constants from a composite of four calibration runs, are plotted in the colored dots; added water vapor, $WVMR_{add}$, utilizing the bubbler reference standard, is plotted in the black line; and the colored lines show the bubbler plus a constant offset representative of the background mixing ratio in the carrier flow for each run. The bottom panel shows the calibration figures (i.e., $1/C_{fir}$ vs. M) with the data from both calibration runs. The solid black line represents the best fit to the full set of calibrations performed at the AIDA facility, and the dashed lines show the $\pm 5\%$ uncertainty range.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion






Fig. A8. Comparison of HWV and APicT low water-vapor data for static experiments. Color indicates chamber pressure. ApicT data are multiplied by 1.09 to account for a calibration difference between these two instruments inferred from data acquired during the high water vapor experiments on 16 and 17 October. In addition, a constant offset of 0.3 ppmv has been subtracted from HWV. At mixing ratios lower than ~ 2 ppmv and chamber pressures less than ~ 200 hPa, systematic uncertainties associated with low mass flows in the HWV instrument plagued the measurement.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract Introduction

Conclusions References

Tables Figures

◀ ▶

◀ ▶

Back Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion






Fig. A9. Comparison of SnowWhite and APicT during the stationary segments during the second week's dynamic experiments. Data analysis is as described in Appendix A8. The black symbols and error bars indicate segment averages and standard deviations. The red bars denote the APicT precision derived from the Gaussian fit. Dashed lines indicate the $\pm 5\%$ error interval.

AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.

Title Page

Abstract

Introduction

Conclusions

References

Tables

Figures

◀

▶

◀

▶

Back

Close

Full Screen / Esc

Printer-friendly Version

Interactive Discussion



AquaVIT-1 water vapor intercomparison

D. W. Fahey et al.




Fig. A10. ISOWAT schematic. The beam of a tunable diode laser (LD) is focused into a multipass-absorption cell to achieve 76 m absorption pathlength. The sample detector (SD) records the multipass cell signal. A high-concentration reference cell in combination with the reference detector (RD) is used for laser wavelength stabilization.