

Interactive comment on “Field calibrations of a low-cost aerosol sensor at a regulatory monitoring site in California” by D. M. Holstius et al.

D. C. S. Beddows (Referee)

d.c.beddows@bham.ac.uk

Received and published: 29 January 2014

An excellent paper reporting the research and development associated with the, "Calibration of a low-cost aerosol sensor". By publishing this work, potentially new possibilities open up which were once limited by budgets and the expense of equipment. I now look forward to a similar device measuring particle number appearing in the literature.

Minor comments to consider which may enhance the paper:

1. Pg 614, Line 14. Not sure what you mean by, "absent any confounding, we should see no associations".

2. When introducing: (a) the PPD42NS and (b) the PANDA, simple diagrams illustrating how they work would help the reader's understanding.

AMTD

7, C10–C11, 2014

Interactive comment on *Atmos. Meas. Tech. Discuss.*, 7, 605, 2014.

Interactive
Comment

[Full Screen / Esc](#)

[Printer-friendly Version](#)

[Interactive Discussion](#)

[Discussion Paper](#)

