

Interactive comment on “The AquaVIT-1 intercomparison of atmospheric water vapor measurement techniques” by D. W. Fahey et al.

D. W. Fahey et al.

david.w.fahey@noaa.gov

Received and published: 16 August 2014

Anonymous Referee #1 Received and published: 18 July 2014 This reviewer would rate the paper excellent (1) for scientific significance, scientific quality, and presentation quality.

The laboratory intercomparison of water vapor instruments is relevant to addressing the differences observed in multiple airborne campaigns over many years. While not all flight conditions could be simulated, some limits could be placed on basic instrument calibration and performance characteristics. This paper serves as an important benchmark for future instrument intercomparisons. Substantial conclusions based on a robust dataset about the performance of both the core and non-core instruments over

C2229

3 water mixing ratio ranges are presented.

The paper is well written, well structured, with just the right amount of experimental detail and description for the reader to understand what was done, adequate supplemental material in the appendix on each instrument to understand the conclusions, appropriate references, and nothing extraneous included. The title reflects the content of the paper, and the abstract provides a clear and concise summary of the project.

The authors would like to thank the reviewer for his/her careful and thorough reading and supportive comments on the manuscript.

Only a few minor issues were noticed:

On page 3202 line 29, “AIDA TDL” is referenced, rather than the instrument name used elsewhere in the paper.

Done

Page 3209 line 22, suggest “Consequently,” at beginning of sentence.

Done

Page 3217 line 17 “VCSEL instruments” were not referenced earlier in the paper. Using the instrument names of Table 1 would be better.

Done

Page 3234 Figure 1 caption. Should be commas after JLH and ALIAS in second to last line.

Done

Interactive comment on Atmos. Meas. Tech. Discuss., 7, 3159, 2014.

C2230