

Interactive comment on “A semi-automated system for quantifying the oxidative potential of ambient particles in aqueous extracts using the dithiothreitol (DTT) assay: results from the Southeastern Center for Air Pollution and Epidemiology (SCAPE)” by T. Fang et al.

Anonymous Referee #2

Received and published: 29 October 2014

This paper describes an automated method for measuring aerosol oxidative load using an on-line DTT assay. The system is complicated but effective and provides the ability to generate kinetic curves. Previous concerns regarding the introduction and methods descriptions have been suitably addressed now.

Interactive comment on Atmos. Meas. Tech. Discuss., 7, 7245, 2014.

C3374