Atmos. Meas. Tech. Discuss., 8, 10755–10792, 2015 www.atmos-meas-tech-discuss.net/8/10755/2015/ doi:10.5194/amtd-8-10755-2015 © Author(s) 2015. CC Attribution 3.0 License.



This discussion paper is/has been under review for the journal Atmospheric Measurement Techniques (AMT). Please refer to the corresponding final paper in AMT if available.

Evaluation of two Vaisala RS92 radiosonde solar radiative dry bias correction algorithms

A. M. Dzambo¹, D. D. Turner², and E. J. Mlawer³

 ¹Cooperative Institute for Mesoscale Meteorological Studies, University of Oklahoma, Norman, OK, USA
 ²National Severe Storms Laboratory/NOAA, Norman, OK, USA
 ³Atmospheric and Environmental Research, Inc., Lexington, MA, USA

Received: 28 August 2015 - Accepted: 22 September 2015 - Published: 20 October 2015

Correspondence to: D. D. Turner (dave.turner@noaa.gov)

Published by Copernicus Publications on behalf of the European Geosciences Union.





Abstract

Solar heating of the relative humidity (RH) probe on Vaisala RS92 radiosondes results in a large dry bias in the upper troposphere. Two different algorithms (Miloshevich et al., 2009, MILO hereafter; and Wang et al., 2013, WANG hereafter) have been designed to

account for this solar radiative dry bias (SRDB). These corrections are markedly different with MILO adding up to 40 % more moisture to the original radiosonde profile than WANG; however, the impact of the two algorithms varies with height. The accuracy of these two algorithms is evaluated using three different approaches: a comparison of precipitable water vapor (PWV), downwelling radiative closure with a surface-based microwave radiometer at a high-altitude site (5.3 km MSL), and upwelling radiative closure with the space-based Atmospheric Infrared Sounder (AIRS).

The PWV computed from the uncorrected and corrected RH data is compared against PWV retrieved from ground-based microwave radiometers at tropical, midlatitude, and arctic sites. Although MILO generally adds more moisture to the original radiosonde profile in the upper troposphere compared to WANG, both corrections yield

similar changes to the PWV, and the corrected data agree well with the ground-based retrievals.

The two closure activities – done for clear-sky scenes – use the radiative transfer models MonoRTM and LBLRTM to compute radiance from the radiosonde profiles to compare against spectral observations. Both WANG- and MILO-corrected RH are statistically better than original RH in all cases except for the driest 30% of cases in the downwelling experiment, where both algorithms add too much water vapor to the original profile. In the upwelling experiment, the RH correction applied by the WANG vs. MILO algorithm is statistically different above 10 km for the driest 30% of cases and above 8 km for the moistest 30% of cases, suggesting that the MILO correction

²⁵ and above 8 km for the molecular 30% of cases, suggesting that the MILO correction performs better than the WANG in clear-sky scenes. The cause of this statistical significance is likely explained by the fact the WANG correction also accounts for cloud cover – a condition not accounted for in the radiance closure experiments.





1 Introduction

Water vapor (WV) is an important driver of weather and climate phenomena. Numerous studies have focused on modeling processes associated with water vapor and evaluating and improving water vapor observations (e.g. Ferrare et al., 1995, 2006; Revercomb

et al., 2003; Suortti et al., 2008; Krämer et al., 2009). Accurate measurements of water vapor are especially crucial in the upper troposphere; although very little water vapor is present in this part of the atmosphere (e.g., Ferrare et al., 2004), processes such as cirrus cloud formation and maintenance (Liou, 1986) and maintenance of stratospheric water vapor (e.g. Jensen et al., 1996a, b; Hartmann et al., 2001) require very accurate
 knowledge of the upper tropospheric water vapor budget. Our understanding of dynamic, thermodynamic and radiative processes, and even cloud water vapor budget, is impacted by the quality of water vapor measurements (Starr and Cox, 1985; Guichard et al., 1999; Wang et al., 2008).

Vaisala RS92 radiosondes have been launched by research and operational centers

- for over a decade and, compared to most ground and space based instruments, provide very high (~ 10 m) vertical resolution. The RS92 radiosonde utilizes two thin-film capacitive elements to measure water vapor, wherein the capacitance measured by the radiosonde is proportional to the number of water vapor molecules that are in contact with the sensor. The resulting relative humidity (RH) measurement is taken as a func-
- tion of this capacitance and the air temperature, which is measured by a separate thin capacitive wire sensor. While in flight, one of the RH sensors measures WV while the other RH sensor is artificially warmed to prevent ice buildup on the sensor; this process alternates between sensors. Unlike its predecessors (such as the RS80 radiosonde), the RH sensor is not shielded from solar radiation. If the RH sensor is warmer than
- the ambient environment due to solar heating, then the measured RH (as computed by Vaisala's DigiCORA[®] software) will be lower than its actual value. Many correction algorithms have been developed (e.g. Vömel et al., 2007b; Cady-Pereira et al., 2008; Yoneyama et al., 2008; Miloshevich et al., 2009; Wang et al., 2013) to correct for this





solar radiative dry bias (SRDB). Nearly all of the aforementioned algorithms correct RH as a function of pressure, solar elevation (zenith) angle, and/or RH itself.

Two of the most widely used correction algorithms come from the work of Wang et al. (2013) and Miloshevich et al. (2009); for brevity, these will be referred to as ⁵ WANG and MILO hereafter. WANG used Global Climate Observing System (GCOS) Reference Upper-Air Network (GRUAN) data (Seidel et al., 2009; Dirksen et al., 2014) to develop and test their RS92 correction algorithm. This physically based correction uses the following form:

$$\mathsf{RH}_{\mathsf{CORR}} = \mathsf{RH}\left(\frac{e_s(T + \mathsf{hf} \cdot \Delta T_{\mathsf{CORR}})}{e_s(T)}\right)$$

¹⁰ $\Delta T_{\text{CORR}} = \text{cf} \cdot \Delta T_{\text{CORR}_{\text{RSN}}},$

15

where *T* is the sonde-measured air temperature, hf is a heating factor (set to 13), cf is a correction factor (set to 0.4 below 500 hPa and 0.6 above 500 hPa) that accounts for both clear skies and cloud cover, and $\Delta T_{\text{CORR}_{RSN}}$ is a temperature correction given by Vaisala (http://www.vaisala.com/en/products/ soundingsystemsandradiosondes/soundingdatacontinuity/RS92DataContinuity/ Pages/revisedsolarradiationcorrectiontableRSN2010.aspx). Note that $\Delta T_{\text{CORR}_{RSN}}$ accounts for pressure and solar zenith angle.

The MILO correction was developed using cryogenic frost-point hygrometer (CFH), microwave radiometer (MWR) and reference humidity probes during the 2006 Water Vapor Validation Experiment-Satellite/Sondes (WAVES) campaign (Vömel et al., 2007).

²⁰ Vapor Validation Experiment-Satellite/Sondes (WAVES) campaign (Vomel et al., MILO consists of an empirically developed correction:

 $RH_{CORR} = G(P, RH) \times RH_{TLAG}$ SRE(α) = SRE(66°) × fraction(α),

where G(P,RH) is an empirically-derived function and given as a "look-up" table of coefficients in Miloshevich et al. (2009), and RH_{TLAG} is the original RH data that has



(1)

(2)

(3)

(4)



been corrected for time lag¹. The MILO correction also includes a correction based on solar zenith angle (Eq. 4), which is applied to Eq. (3): solar radiation error (SRE) is dependent on solar altitude angle (α) and expressed as a fraction of the SRE at 66°. which represents the mean solar zenith angle for the daytime CFH/RS92 soundings

during WAVES (Miloshevich et al., 2009). A comparison of these two correction algorithms in a typical atmospheric sounding is given in Fig. 1. In 2011, Vaisala upgraded its DigiCORA[®] software to version 3.64, which included their own SRDB correction algorithm; the details of this algorithm are not freely available to the public. This study focuses on RS92 radiosondes collected before this change to the DigiCORA software was made. 10

We evaluate the WANG and MILO SRDB corrections at sites maintained by the Department of Energy's (DOE) Atmospheric Radiation Measurement (ARM) program (Ackerman and Stokes, 2003; Mather and Voyles, 2013), at which numerous instruments are deployed that will aid in this evaluation. We use data from the ARM sites in

- the Southern Great Plains (SGP) in Lamont, OK, USA, North Slope Alaska (NSA) in 15 Barrow, AK, USA, and the Tropical Western Pacific (TWP) on Nauru Island, Republic of Nauru (Stokes and Schwartz, 1994). We also use ARM data collected during a 3 month experiment at a 5300 m MSL site at Cerro Toco (CJC) in northern Chile (Turner and Mlawer, 2010). Utilizing several distinct climate locations ensures a more accurate and
- in-depth analysis of the two correction algorithms. 20

Comparing the correction algorithms directly 2

The two correction algorithms were applied to RS92 data launched at the SGP. NSA. TWP, and CJC sites. These data spanned all months of the year. The mean change in

¹The time lag correction was developed for RS80 radiosondes. RS92 radiosondes do not requires a time lag correction due to the lack of a protective shield on the RH sensor. See Miloshevich et al. (2004) for more information





water vapor mixing ratio as a function of height (relative to the original radiosonde measurement) for each site is shown in Fig. 2. The largest difference between the two correction algorithms is in the middle-and-upper troposphere above 7 km, where the MILO algorithm moistens the original radiosonde much more than the WANG correction; the

difference between MILO and WANG approaches a factor of 1.8 by 14 km. Given the sensitivity of the outgoing longwave radiation to changes in upper tropospheric water vapor (e.g., Ferrare et al., 2004), understanding which of these corrections is more appropriate is very important. However, a close inspection of Fig. 2 also shows that the WANG correction moistens the radiosonde slightly more than the MILO correction in the lowest 2 km for the moister tropic and mid-latitude sites.

We compare the precipitable water vapor (PWV) values derived from integrating the moisture profiles from the original and corrected radiosonde profiles with those retrieved from the ARM 2-channel microwave radiometers (MWRs) using the so-called "MWRRET" algorithm (Turner et al., 2007). ARM has used the MWR-retrieved PWV as a "standard" for correcting for first-order radiosonde biases (Turner et al., 2003; Cady-

¹⁵ a "standard" for correcting for first-order radiosonde biases (Turner et al., 2003; Cady-Pereira et al., 2008), calibrating its Raman lidar (Turner and Goldsmith, 1999), and for evaluating infrared radiative transfer models (e.g., Turner et al., 2004).

The comparisons of the radiosonde PWV values with those from the MWR (Fig. 3) show that the original uncorrected radiosondes have a dry bias that increases as the

- PWV increases. Table 1 summarizes the median and standard deviations; in an effort to remove outliers, values that were below the 10 percentile and above the 90 percentile ile were removed before computing the PWV biases. For example, Fig. 3a1 shows that the mean PWV from the original radiosondes at SGP are approximately 0.35 cm drier than the MWR-retrieved value in the 4.25–4.75 cm bin; however, the Wang-corrected
- radiosonde, while moister than the original radiosonde, still has a slight dry bias of 0.10 cm relative to the MWR in this bin (Fig. 3a3). The magnitude of the PWV bias generally increases when more PWV is present in the atmosphere. Both the WANG and MILO corrections increase the sonde's derived PWV, and result in much better agreement with the MWR. This result is consistent with the findings in Yu et al. (2015),





where MWR retrievals of PWV and PWV derived from WANG-corrected RH data were found to be within the uncertainty of the MWR instrument (which is \sim 0.07 cm; Turner et al., 2007).

The PWV results (Fig. 3, Table 1), especially when we consider all three sites (SGP,
NSA, and TWP), demonstrate that both algorithms greatly improve the accuracy of the PWV relative to the MWR but do not distinguish which of the two corrections may be better. The WANG's drier correction (relative to MILO) in the upper troposphere is slightly offset by its wetter correction near the surface and thus yields similar PWV values. A close inspection of Table 1, however, suggests that the MILO correction seems
to add more PWV compared to MILO in the tropics, whereas WANG adds more PWV in drier climates such as SGP and NSA.

To evaluate the accuracy of the two SRDB corrections as a function of height, we first considered comparing the corrected radiosondes with water vapor measurements made by the ARM Raman lidars (Goldsmith et al., 1998; Ferrare et al., 2006) at the SGP and TWP/Darwin sites. Unfortunately, during the daytime the Raman lidar obser-

SGP and TWP/Darwin sites. Unfortunately, during the daytime the Raman lidar observations are limited to altitudes below 5 km, and thus unable to provide any insight into the accuracy of the two corrections in the upper troposphere.

Instead we use two radiance closure experiments to evaluate the two corrections in the upper troposphere: one downwelling experiment and one upwelling experiment.

- Radiance closure studies have been used in prior studies to validate sonde-derived brightness temperature (*T*_B) measurements (e.g. Turner et al., 2003; Soden et al., 2004; Mattioli et al., 2008; Kottayil et al., 2012) and offer another method for detecting systematic biases in radiosonde RH measurements. In each experiment, a radiative transfer model is used to transform the original RH data, along with the WANG and MILO cor-
- ²⁵ rected RH data, into simulated brightness temperatures. The model-derived $T_{\rm B}$ data is directly compared to an appropriate reference spectral radiance measurement, which will be described more thoroughly in the respective experiment sections. Statistical significance (for p = 0.05) is computed, where appropriate, to show the significance of the difference between WANG, MILO and the original data.



3 Downwelling experiment

The ARM program conducted the second phase of the Radiative Heating in Underexplored Bands Campaign (RHUBC-II) in Cerro Toco, Chile (CJC), in August through October 2009 (Turner and Mlawer, 2010). The CJC site is located approximately

5.3 kma.s.l. in the Atacama Desert; this site can be considered as a mid-tropospheric site due to its altitude and water vapor conditions. Also, during RHUBC-II, there was a high frequency occurrence of clear-sky and dry conditions, making it optimal for studying the accuracy of upper tropospheric water vapor measurements.

Our reference instrument is the G-band water-vapor radiometer profiler (GVRP). The

- ¹⁰ GVRP measures downwelling radiation in 15 channels at 170.0, 171.0, 172.0, ..., 182.0, 183.0, and 183.31 GHz. Cimini et al. (2009) showed that the GVRP (in that paper, referred to as "MP-183") agreed within uncertainty with two other collocated 183 GHz radiometers during RHUBC-I, which was held at the NSA site in February– March 2007. The lower frequency channels (e.g., below 178 GHz) are more sensitive
- to the total PWV, while the higher frequency channels are more sensitive to the water vapor profile shape (Cimini et al., 2009). The GVRP has an uncertainty of 1.5 K for $T_{\rm B}$ measurements (Cadeddu et al., 2013).

The corrected and uncorrected RH data from the 144 RS92 radiosondes launched during RHUBC-II were used as input into version 4.1 of the MonoRTM radiative trans-

- fer model (Clough et al., 2005) to compute monochromatic downwelling radiance at high spectral resolution (10 MHz) from 168–185 GHz. Since the Cerro Toco site almost always has clear skies, the model was run to compute clear sky radiances (methodology for identifying cases with environmental inhomogeneity or clouds is described in the next paragraph). These computed clear-sky monochromatic spectra were con-
- volved with the GVRP's instrument function to derive model calculations corresponding to each GVRP channel. These model-derived radiances, which were converted to $T_{\rm B}$, were directly compared to the $T_{\rm B}$ measurements made by the GVRP.

To reduce the complexity of the analysis, we restricted our comparisons to clear sky conditions only. To identify cloudy sky conditions as well as inhomogeneous environments (i.e., when there was a horizontal gradient in water vapor across the RHUBC-II site), the standard deviation of the GVRP T_B measurements at 174 GHz over a 30 min ⁵ window centered at the radiosonde launch time at both 30 and 150° was computed. If the standard deviation at either angle (where 90° corresponds to zenith) was more than 2.25 K, the sky conditions were not considered uniform and the sonde was removed from subsequent analysis. This additional screening also accounts for inhomogeneity created by localized mountain-scale circulations and a thermally driven circulation across the Cerro Toco site (Marín et al., 2013).

The comparison of the MonoRTM $T_{\rm B}$ calculations using the MILO vs. WANG corrected radiosondes as input demonstrated a different spectral character based upon the PWV in the profile. For the moistest 30% of the CJC radiosondes (i.e., where the PWV > 0.57 mm, where the maximum PWV observed at CJC was 1.20 mm), the

- ¹⁵ MILO-computed T_B was typically larger than the WANG-computed values at all GVRP frequencies (Fig. 4, green spectra), which implies that the MILO-corrected radiosondes are moister over the entire profile. However, for the driest 30 % of the CJC radiosondes (i.e., PWV < 0.37 mm), the T_B values computed using the WANG-corrected profiles are larger than the MILO-computed radiance for frequencies below 182 GHz (Fig. 4, or-
- ange spectra). This suggests that the WANG-corrected radiosondes are moister than the MILO-corrected data, especially in the lowest several kilometers of the atmosphere. Most importantly, this analysis suggests that the significant differences in how the two correction algorithms behave at different PWV amounts can be used with GVRP spectral observations to evaluate both algorithms.
- The median observed minus computed brightness temperature spectra for the WANG- and MILO-corrected radiosondes are shown in Fig. 5; these data are also divided into the 30% moistest and 30% driest profiles, each of which has 26 cases. Table 2 summarizes the median biases for the 30% moistest profiles and 30% driest profiles with standard deviations. For the median of the driest cases, the MonoRTM-

derived $T_{\rm B}$ calculations for both correction algorithms are approximately 1–4 K warmer than the GVRP observations for frequencies between 170 and 178 GHz, increasing to over 13 K warmer than the GVRP at the center of the water vapor absorption line at 183.3 GHz. This suggests that both correction algorithms actually worsen the MonoRTM-derived $T_{\rm B}$ measurements (compared to $T_{\rm B}$ measurements derived from

- the original RH data) in the most extreme of dry cases seen in the CJC dataset. Interestingly, the MonoRTM calculations that used the original uncorrected radiosondes provide a much better agreement with the GVRP observations for these very dry cases. Furthermore, the application of the two correction algorithms increases the scatter be-
- ¹⁰ tween the GVRP and MonoRTM-computed $T_{\rm B}$ at 183.0 and 183.31 GHz relative to the original uncorrected radiosonde (Table 2), suggesting that neither algorithm adds skill at the very low PWV amounts seen in this category of cases. We speculate that, given the extremely low RH values of ~ 10 % characteristic of the CJC site (Fig. 6), the precision of the RH measurement itself becomes a more significant source of er-
- ror. As a result, additional residual error is likely added to the otherwise bias-corrected MonoRTM-computed $T_{\rm B}$ values.

A much different story, however, is seen in the 30 % moistest profiles. The mean $T_{\rm B}$ bias between the GVRP observations and the MonoRTM calculations using both the WANG and MILO-corrected input data from this moist subset is much smaller than for the 200% drived profiles. The WANG (MILO Man a DTM extended)

- for the 30 % driest profiles. The WANG/MILO MonoRTM calculations also yield slightly moist-biased results compared to the original RH MonoRTM calculations, which are dry biased (Fig. 5). The good agreement between the observed and computed spectra for frequencies less than 177 GHz suggests that both algorithms have the PWV correct, as these channels have relatively constant weighting functions with height. At
- ²⁵ 183.0 and 183.31 GHz, the MonoRTM-derived $T_{\rm B}$ calculations for the WANG calculation is warm biased by 0.422 and 0.333 K respectively, whereas the $T_{\rm B}$ calculations using the MILO-corrected radiosondes are warm biased by 1.778 and 1.752 K. While these results seem to indicate that WANG-corrected radiosondes are in better agreement with the GVRP observations, this result is not statistically significant. Interestingly,

the scatter in the GVRP minus MonoRTM residuals at these two frequencies is very similar between the calculations that used the original RH profile and either of the two corrected RH profiles (Table 2). The moist 30 % cases in this analysis, when compared to other distinct climatological locations (Fig. 7), are considerably drier when compared 5 to a tropical location (e.g. the ARM TWP Nauru site).

As a consistency check for the $T_{\rm B}$ residuals (computed as observed minus computed) derived from original, WANG- and MILO-corrected RH data, a one-sided student *t* test is performed on the 30 % partitioned moist and dry cases for all 15 MonoRTM frequencies (results not shown here). For the moistest and driest 30 % of cases, WANG- and

- ¹⁰ MILO-corrected RH are statistically significant (at the p = 0.05 level) from the original RH data. A one-sided student *t* test between WANG and MILO for the moistest or driest 30 % of cases, however, reveals no statistical significance at any frequency. Despite the noted difference in biases from Fig. 5, we cannot reasonably conclude that one correction algorithm is better than the other. Hence, a second experiment is needed to further
- ¹⁵ deduce differences between the WANG and MILO corrections.

4 Upwelling experiment

The downwelling radiance closure experiment demonstrated that both WANG- and MILO-corrected RH data are improved over the original RH data only for the moister cases at CJC. However, while the CJC site is representative of a mid-tropospheric site in terms of altitude and pressure, its very dry climate resulted in water vapor amounts (as indicated by the IWV histograms in Fig. 7) that are significantly drier than those found at other ARM sites. Unfortunately, the CJC site was unique; downwelling radiance closure at the other sites would be modulated by the water vapor in the lower troposphere and there would be little sensitivity in the downwelling radiance from changes in the upper tropospheric humidity. The one-sided student *t* test results further suggest

²⁵ In the upper tropospheric humidity. The one-sided student *t* test results further suggest little variation between the correction algorithms despite the fact they correct differently in the upper troposphere.

However, upwelling spectral infrared radiance observations are very sensitive to the vertical distribution of water vapor. The SGP site experiences a wide range of weather phenomena throughout the year, which results in a wide range of upper-tropospheric IWV throughout the year (Fig. 7 – green line). During the cold season,
⁵ upper-tropospheric IWV at SGP is representative of that measured at the ARM's NSA (Barrow) site, whereas in the warm season at SGP the upper-tropospheric IWV is representative of a tropical location (e.g. the ARM's TWP sites). For this reason, radiosonde data from the SGP site is chosen for the upwelling radiance closure exercise.

We used the infrared radiance observations made by the Atmospheric Infrared Sounder (AIRS; Aumann and Pagano, 1994). Launched into a sun-synchronous polar orbit on 4 May 2002 aboard NASA's Aqua satellite (Parkinson, 2003), this instrument has provided extensive insight into a host of weather and climate related phenomena (e.g. Moustafa et al., 2006; Shu and Wu, 2009; Shimada and Minobe, 2011). The high spectral resolution of the AIRS, with 2378 channels, provides a wealth of infor-

- ¹⁵ mation for our study. Its data have been extensively compared with data from infrared spectrometers flown on aircraft (e.g., Tobin et al., 2006), demonstrating excellent calibration accuracy and stability. One caveat to using the AIRS, like any sun synchronous polar-orbiting satellite, is the temporal resolution of the data: although approximately 12.5 years of AIRS data are available, surface locations near the poles will have more
- ²⁰ measurements than surface locations in the mid-latitudes or near the equator. The ARM SGP site launches radiosondes around 18:00 UTC every day, which is about two to three hours before the AIRS overpass time (i.e. around 20:00 to 21:00 UTC). For this experiment, AIRS T_B and radiosonde data from a five-year period from January 2005 through December 2009 were used.
- ²⁵ Upwelling infrared radiation is highly sensitive to changes in water vapor, so we needed to ascertain if the PWV changed appreciably between the sonde launch and AIRS overpass. Upwelling infrared radiation measured is also sensitive to changes in cloud properties: the development or advection of clouds at the sonde launch site can obscure the atmosphere below the cloud top height. To minimize this impact, we in-

cluded data only where the AIRS overpass occurred within 135 min of the radiosonde launch. Also, we elected to analyze cloud-free scenes associated with time periods in which the PWV did not change appreciably. Thus, if the MWR PWV changed by more than 5% between the sonde launch and AIRS overpass, we excluded that data from ⁵ the study.

The 5% threshold was determined through a sensitivity study: for two standard atmospheres (summer and winter), we perturbed the column water vapor across a range of values for a fixed temperature profile typical for that season (results not shown here) and used the LBLRTM (Clough et al., 2005; described later in this section) to evaluate

- ¹⁰ changes in the peaks of the weighting function height computed for each of 467 total frequencies (subset from the 2378 AIRS channels) from each profile. The vertical resolution of the model for altitudes lower than 16 km was set to 100 m. For a change in PWV of 5 %, approximately 16 % (summer) and 14 % (winter) of the weighting function peak heights changed by more than 100 m. It should also be noted that 11 % of the
- total peaks (for each season) changed by less than 200 m (meaning than 5 % (3 %) of the summer (winter) weighting function peak heights changed by 200 m or more). Considering we use 1 km altitude bins in the main analysis, and the vertical resolution of the model is an order of magnitude smaller than this bin size, we feel this threshold is more than reasonable.

Additional screenings were implemented to account for the effects of cloud cover during this time threshold. The AIRS provides radiance measurements in a "footprint", which is a 3×3 set of pixels. Data were chosen such that the center pixel was the measurement closest to the SGP site. At 938 cm^{-1} the atmosphere is transparent to nearly all gases except for water vapor, thereby making this channel very sensitive to surface temperature in clear conditions. The standard deviation of the T_B values obtained from the 938 cm^{-1} channel radiances ($T_{B,938}$ hereafter) was computed for all nine pixels and thresholds were determined based on all available footprints (Table 3). To account for seasonal variability in the $T_{B,938}$ measurements, thresholds are deter-

To account for these deviations in surface temperature while keeping the error to within $\sim 6\%$ or ~ 3 K, we defined a clear-sky threshold equal to twice the 25 percentile of the

mined on a monthly basis: $T_{B,938}$ measurements in all pixels (for a clear sky scene)

For comparison sake, previous AIRS validation studies at this channel over the ocean

(e.g. Hagan and Minnett, 2003) demonstrated that the AIRS radiometric uncertainty is

demonstrated that the root mean square error of brightness temperature and water

vapor measurements over the ocean approached the theoretical expectations of clear sky conditions. Even in clear sky data, some variability in $T_{B 938}$ measurements occurs

as a result of local differences in surface temperature across the swath of the footprint.

⁵ approximately 1%, which is about 0.5 K at 300 K for 938 cm⁻¹. Tobin et al. (2006) later

result in a small standard deviation (generally less than 2 K).

10

 $T_{B,938}$ standard deviation for that month (Table 3). The factor of 2 ensures that enough cases make it into the analysis while staying under 3 K for any season, which accounts for the prescribed natural variability in $T_{B,938}$. High T_B standard deviations are primarily a signature of partly or mostly cloudy skies, since cloud tops are almost always colder than the surface.

Stratiform cloud decks are also accounted for: low $T_{B,938}$ standard deviations but lower than average $T_{B,938}$ values (relative to the mean for that month) signify a cloud deck and therefore are also screened from the data. Subvisible cirrus clouds, which ²⁰ affect the radiance budget but are too optically thin to be easily identified in the AIRS observations, were identified using the radiosonde RH data. Any original RH profile that has an RH_{ICE} measurement greater than 90% anywhere in the column is removed. Using all of the above criteria to account for cloud coverage and environmental homogeneity, 96 cases pass these screenings.

²⁵ The line-by-line radiative transfer model LBLRTM (Clough et al., 2005), which shares the physical basis as the MonoRTM used in the downwelling experiment, is used to compute upwelling infrared radiance from the original and corrected RH data. The LBLRTM computes very high-resolution radiance data; in order to match the 2378 AIRS channels, the monochromatic LBLRTM output is convolved with the AIRS instrument

spectral response function for each of the 2378 AIRS channels. The atmosphere is generally opaque in the spectral region between approximately 1300 and 2000 cm⁻¹ at the SGP site due to absorption by water vapor. Our analysis focused on the radiative closure in this spectral region, using only AIRS channels where the transmission of the atmosphere was zero. By restricting our analysis to this set of channels, uncertainties associated with the emission of the earth's surface were avoided.

For each radiosonde/AIRS overpass pair, the upwelling T_B was computed using the LBLRTM along the viewing angle of the AIRS instrument, and the observed minus computed T_B differences were assigned to different altitudes. We attributed the $T_B(\lambda)$ difference to the altitude where the weighting function for that wavelength (λ) had its maximum value. The weighting functions as a function of height W(z) were computed as:

 $W(z) = \beta(z) e^{-\tau(z)},$

10

where $\beta(z)$ is the gaseous absorption coefficient and $\tau(z)$ is the cumulative optical depth from the AIRS sensor to height *z* computed as

$$\tau(z) \equiv \int_{z}^{\infty} \beta(z') \,\mathrm{d}z', \tag{6}$$

and the wavelength dependence is inferred. In the 1300–2000 cm⁻¹ spectral region, water vapor is the primary gaseous absorber. Weighting functions "peak" at various heights depending on a channel's sensitivity to water vapor and the shape of the water

²⁰ vapor profile. For mid-latitude atmospheres, weighting functions for the different spectral channels generally peak between 5 to 12 km depending on the water vapor profile (which determines the optical depth profile) and the temperature profile. AIRS channels where the weighting function peaks above 2 km and below the tropopause are considered valid for this study. If a peak fell within a 1 km altitude range (e.g. 5–6, 6–7 km, ²⁵ etc.), the observed minus computed $T_{\rm B}$ residual for that channel was binned in this

(5)

CC D

height range. Similar to the downwelling experiment, mean residuals are computed according to the 30% moistest and 30% driest cases, which corresponded to IWV thresholds (for all radiosondes having valid measurements between 525 and 200 hPa) of above 0.96 mm and below 0.37 mm respectively.

- ⁵ Median brightness temperature biases between the AIRS and un/corrected RH data (Fig. 8) reveals an average correction for any given layer of approximately 0.2 to 0.4 K, depending on the correction. Below 5 km, $T_{\rm B}$ computations using WANG-corrected RH are less biased than $T_{\rm B}$ computations using MILO-corrected RH (a result consistent with Fig. 2). Above 5 km, MILO-corrected RH results in model-computed $T_{\rm B}$ that is less
- ¹⁰ biased than WANG, but both WANG- and MILO-corrected RH result in $T_{\rm B}$ computations that are statistically significant from $T_{\rm B}$ model computations using original RH as input (for all altitude levels). When comparing WANG- and MILO-corrected $T_{\rm B}$ residuals against one another, the corrections become statistically significant (at p = 0.05) from one another above the 5–6 km height bin. Also, MILO-corrected $T_{\rm B}$ residuals are
- ¹⁵ less biased than WANG-corrected $T_{\rm B}$ residuals except at the 12–13 km height bin. We reasonably conclude that MILO-corrected RH for all cases performs better than WANG-corrected RH, however, we feel it is necessary to partition the cases by uppertropospheric IWV in order to further deduce differences between the WANG and MILO RH correction algorithms.
- ²⁰ When evaluating the driest 30 % of data and moistest 30 % of data in Fig. 8, brightness temperature biases between the AIRS and un/corrected RH data (Fig. 9) are corrected, on average, by 0.2 to 0.5 K for the driest cases and 0.3 to 0.4 K for the moistest cases, depending on the correction algorithm that was used. Table 4 summarizes the median biases for the driest and moistest cases with standard deviations. Aside from
- the 12–13 km layer for WANG and the 6–7, 10–13 km height bins for MILO, the correction algorithms remain slightly dry biased. This result is consistent with the findings in Fig. 2: since MILO generally adds more WV in the middle and upper troposphere, it follows that MILO corrects more than WANG in these driest cases (though no more than about 0.2 K) and appears to be better. The moist cases, on the other hand, result in T_B

residuals closer to the observed AIRS T_B , with MILO-corrected T_B residuals being less biased than WANG-corrected T_B residuals at every height bin except the 12–13 km height bin. Again, these results are consistent with Fig. 2: MILO corrects more than WANG (as much as 0.10 to 0.15 K more), which is only possible in the presence of

- increased WV in the middle and upper troposphere. It should be noted that many more observations (i.e. usable channels resulting from the weighting function analysis) are available for the moist case category (especially above the 5–6 km height bin). In the drier profiles, the opacity of the atmosphere due to water vapor absorption decreases and thus more AIRS channels are eliminated from the analysis because the channel
- ¹⁰ is sensitive to surface emission, thereby making fewer measurements available. The number of measurements (i.e. number of brightness temperature measurements between 1300 and 2000 cm⁻¹ from the partitioned cases) per height bin for the driest 30 % and moistest 30 % of data is also given in Table 4.

For both WANG and MILO, Table 4 shows that both corrections have a slightly decreased standard deviations compared to the original measurements at nearly every height bin. MILO, in most cases, has a slightly lower standard deviations compared to WANG.

We also computed statistical significance among the $T_{\rm B}$ residuals for original, WANGand MILO-corrected $T_{\rm B}$ data (for the 30 % moistest and driest cases). Again, both the WANG- and MILO-corrected $T_{\rm B}$ are significantly different from the $T_{\rm B}$ derived from the

- ²⁰ WANG- and MILO-corrected T_B are significantly different from the T_B derived from the original RH data for all altitudes. When coupled with the fact T_B residuals among the correction algorithms are much less biased compared to T_B residuals using original RH data, we can conclude that WANG- or MILO-corrected RH is much improved over the original RH measurements. For the driest 30% of cases, the WANG and MILO
- ²⁵ corrections are statistically significant from each other (at the *p* = 0.05 level) at and above the 9–10 km bin. For the moistest 30 % of cases, WANG- and MILO-corrected *T*_B become statistically significant from one another at and above the 7–8 km bin. In both cases, MILO is less biased than WANG above the stated altitude bins (except

the 12–13 km bin); therefore we can also conclude that MILO-corrected RH is better representative of upper-tropospheric RH compared to WANG-corrected RH.

For both the upwelling and downwelling experiments, the dry thresholds are the same (0.37 mm), however, the T_B residuals computed for the upwelling experiment ⁵ from each correction algorithm reduced the bias, which was not the case for the driest 30% of results from the downwelling experiment. At this time, we cannot conclude why results for the respective subsets of data differ. The moist threshold is higher for the upwelling experiment compared to the downwelling experiment (0.96 vs. 0.57 mm) – likely because water vapor can more easily reach the upper troposphere due to phenomena such as deep convection at the SGP, while at CJC there are a range of pro-

- cesses at work keeping the troposphere relatively dry (Rutllant Costa, 1977). Figures 7 and 8 corroborate this idea as well considering the CJC observes lower RH and IWV (respectively) compared to the SGP site. With the exception of the 12–13 km bin, $T_{\rm B}$ residuals (Fig. 9) computed from MILO-corrected RH are less biased than $T_{\rm B}$ residuals
- ¹⁵ computed from WANG-corrected RH but remain slightly dry biased. Despite the limitations present in the upwelling experiment, but given the statistical significance between MILO- and WANG-corrected RH, the results from this experiment suggest that MILOcorrected RH is better representative of clear-sky RH compared to WANG-corrected RH in the upper troposphere, and both corrections represent improvements compared to uncorrected sondes.

5 Conclusions

25

Both the WANG and MILO corrections significantly improve the original Vaisala RS92 RH data, as demonstrated in an analysis of PWV at multiple sites, yielding approximately the same improvement in PWV relative to the MWR-retrieved value. However, the two algorithms differ in their corrections as a function of height due to their different methodologies.

Given this difference, radiative closure experiments were performed to determine if one of the two corrections was better than the other. Comparing radiative transfer calculations that use the WANG- and MILO-corrected radiosondes, an analysis of downwelling measurements at the 183.00 and 183.31 GHz channels of the CJC GVRP

- ⁵ indicated that the WANG median $T_{\rm B}$ calculation was not statistically different compared to the MILO median $T_{\rm B}$ calculation for the moist cases that are more typical of upper troposphere in mid-latitude atmospheres. Also, both corrections significantly improved the $T_{\rm B}$ bias for the moist cases: the original median $T_{\rm B}$ calculation was ~ 10 K too warm (implying the original sonde was too dry) at 183.00 and 183.31 K. However, radiosondes
- in the very dry category, corresponding to upper tropospheric conditions not typically found in mid-latitude or tropical locations, were made significantly too moist by both corrections, yielding much poorer agreement with the GVRP than the original uncorrected radiosonde profile. We find WANG- and MILO-corrected RH to be statistically better than the original RH for the moist cases, however, WANG- and MILO-corrected
 RH are not statistically different when tested against one-another.

The upwelling experiment using AIRS measurements revealed additional differences between WANG and MILO, likely owing to the fact the SGP site has a great seasonal dependence on upper-tropospheric IWV. The driest cases show that WANG is slightly less biased than MILO below 5 km and is likely due to the fact that WANG corrects more

- ²⁰ than MILO in the lower troposphere. Otherwise, MILO is less biased than WANG in nearly every other scenario, as indicated by the partitioning of radiances by height using weighting functions. Both the WANG and MILO corrections result in T_B computations that are statistically significant from T_B computations derived from original RH – a result consistent with the results found in the downwelling experiment. We find, however, that
- ²⁵ MILO is statistically different from WANG above 8 km in the moistest 30 % of cases and above 10 km in the driest 30 % of cases. We conclude that MILO offers a more realistic representation of upper-tropospheric RH compared to WANG because of the lower $T_{\rm B}$ bias at nearly all altitudes coupled with the statistical significance between MILO and WANG.

The outcome of the upwelling radiance closure experiment suggests that the correction factor "cf" used to scale the temperature correction in WANG may be too low. However, the intent of this correction factor is to account for both clear and cloudy conditions and despite the fact WANG offers a much better agreement than the original ⁵ RH measurements, our results indicate that WANG seemingly under-corrects for solar radiative dry bias. This also likely explains (from the upwelling experiment) why WANG is statistically different from MILO in the upper-troposphere. Given the ease-of-use of the WANG correction, we suggest that the "cf" be computed separately for clear and cloudy skies. Doing so would make WANG more robust and would be applicable to an increased number of applications. Regardless, our results demonstrate the utility of both correction algorithms across a wide range of climatic regimes, where MILO is

especially effective in the upper troposphere for clear-sky conditions.

Acknowledgement. The radiosonde, MWR, and GVRP data were obtained from the Atmospheric Radiation Measurement (ARM) Program sponsored by the US Department of Energy,
 Office of Science, Office of Biological and Environmental Research, Climate and Environmental Sciences Division. We would also like to thank the Dave Tobin for providing the AIRS footprint data needed to perform the upwelling experiment. This work was supported by the US Department of Energy's Atmospheric System Research (ASR) program with grant DE-SC0008830.

References

- Ackerman, T. P. and Stokes, G. M.: The atmospheric radiation measurement program, Phys. Today, 56, 38–44, 2003.
 - Aumann, H. H. and Pagano, R. J.: Atmospheric infrared sounder on the Earth Observing System, Opt. Eng., 33, 776–784, 1994.

Cadeddu, M. P.: G-Band Vapor Radiometer Profiler (GVRP) Handbook, DOE Office of Science

Atmospheric Radiation Measurement (ARM) Program, United States, No. DOE/SC-ARM/TR-091, Pacific Northwest National Laboratory (PNNL), Richland, WA, USA, 2010.

- 10775

Cadeddu, M. P., Turner, D. D., and Liljegren, J. C.: A neural network for real-time retrievals of PWV and LWP from arctic millimeter-wave ground-based observations, IEEE T. Geosci. Remote, 47, 1887–1900, 2009.

Cadeddu, M. P., Liljegren, J. C., and Turner, D. D.: The Atmospheric radiation measurement

- (ARM) program network of microwave radiometers: instrumentation, data, and retrievals, 5 Atmos. Meas. Tech., 6, 2359–2372, doi:10.5194/amt-6-2359-2013, 2013.
 - Cady-Pereira, K. E., Shephard, M. W., Turner, D. D., Mlawer, E. J., Clough, S. A., and Wagner, T. J.: Improved daytime column-integrated precipitable water vapor from vaisala radiosonde humidity sensors, J. Atmos. Ocean. Tech., 25, 873-883, 2008.
- ¹⁰ Chahine, M. T., Pagano, T. S., Aumann, H. H., Atlas, R., Barnet, C., Blaisdell, J., Chen, L., Divakarla, M., Fetzer, E. J., Goldberg, M., Gautier, C., Granger, S., Hannon, S., Irion, F. W., Kakar, R., Kalnay, E., Lambrigtsen, B. H., Lee, S., Le Marshall, J., McMillian, W. W., McMillin, L., Olsen, E. T., Revercomb, H., Rosenkranz, P., Smith, W. L., Staelin, D., Strow, L. L., Susskind, J., Tobin, D., Wolf, W., and Zhou, L.: AIRS: improving weather forecasting and providing new data on greenhouse gases, B. Am. Meteorol. Soc., 87, 911–926, 2006. 15
- Cimini, D., Nasir, F., Westwater, E. R., Payne, V. H., Turner, D. D., Mlawer, E. J., Exner, M. L., and Cadeddu, M. P.: Comparison of ground-based millimeter-wave observations and simulations in the Arctic winter, IEEE T. Geosci. Remote, 47, 3098-3106, doi:10.1109/TGRS.2009.2020743, 2009.
- Clough, S. A., Shephard, M. W., Mlawer, E. J., Delamere, J. S., Iacono, M. J., Cady-Pereira, K., 20 Boukabara, S., and Brown, P. D.: Atmospheric radiative transfer modeling: a summary of the AER codes, J. Quant. Spectrosc. Ra., 91, 233-244, 2005.
 - Dirksen, R. J., Sommer, M., Immler, F. J., Hurst, D. F., Kivi, R., and Vömel, H.: Reference quality upper-air measurements: GRUAN data processing for the Vaisala RS92 radiosonde, Atmos. Meas. Tech., 7, 4463-4490, doi:10.5194/amt-7-4463-2014, 2014.
 - Ferrare, R. A., Melfi, S. H., Whiteman, D. N., Evans, K. D., Schmidlin, F. J., and Starr, D. O'C.: A comparison of water vapor measurements made by Raman Lidar and radiosondes, J. Atmos. Ocean. Tech., 12, 1177-1195, 1995.

25

Ferrare, R. A., Browell, E. V., Ismail, S., Kooi, S. A., Brasseur, L. H., Brackett, V. G., Clayton,

M. B., Barrick, J. D. W., Diskin, G. S., Goldsmith, J. E. M., Lesht, B. M., Podolske, J. R., 30 Sachse, G. W., Schmidlin, F. J., Turner, D. D., Whiteman, D. N., Tobin, D., Miloshevich, L. M., Revercomb, H. E., Demoz, B. B., and Di Girolamo, P.: Characterization of upper-troposphere

mation and persistence of subvisible cirrus clouds near the tropical tropopause, J. Geophys. Res., 101, 21361–21375, 1996b.

- Kottayil, A., Buehler, S. A., John, V. O., Miloshevich, L. M., Milz, M., and Holl, G.: On the importance of Vaisala RS92 radiosonde humidity corrections for a better agreement between measured and modeled satellite radiances, J. Atmos. Ocean. Tech., 29, 248–259, 2012.
- ³⁰ Krämer, M. and coauthors, 2009: Ice supersaturations and cirrus cloud crystal numbers. Atmos. Chem. Phys., 9, 3505–3522.

water vapor measurements during AFWEX using LAS E., J. Atmos. Ocean. Tech., 21.12, 1790–1808, 2004.

- Ferrare, R. A. Turner, D., Clayton, M., Schmid, B., Redemann, J., Covert, D., Elleman, R., Ogren, J., Andrews, E., Goldsmith, J. E. M., and Jonsson, H.: Evaluation of daytime measure-
- ments of aerosols and water vapor made by an operational Raman lidar over the Southern Great Plains, J. Geophys. Res., 111, D05S08, doi:10.1029/2005JD005836, 2006.
 - Goldsmith, J. E. M., Blair, F. H., Bisson, S. E., and Turner, D. D.: Turn-key Raman Lidar for profiling atmospheric water vapor, clouds, and aerosols, Appl. Optics, 37, 4979–4990, 1998.Guichard, F., Parsons, D., and Miller, E.: Thermodynamic and radiative impact of the correction
- of sounding humidity bias in the tropics, J. Climate, 13, 3611–3624, 2000.
 - Hagan, D. E. and Minnett, P. J.: AIRS radiance validation over ocean from sea surface temperature measurements, IEEE T. Geosci. Remote, 41, 432–441, 2003.
 - Hartmann, D. L., Holton, J. R., and Fu, Q.: The heat balance of the tropical tropopause, cirrus, and stratospheric dehydration, Geophys. Res. Lett., 28, 1969–1972, doi:10.1029/2000GL012833, 2001.
 - Heymsfield, A. J. and Miloshevich, L. M.: Relative humidity and temperature influences on cirrus formation and evolution: observations from wave clouds and fire II., J. Atmos. Sci., 52, 4302–4326, 1995.

Ingram, W. J.: Water vapor feedback in a small ensemble of GCMs: two approaches, J. Geo-

²⁰ phys. Res., 117, D12114, doi:10.1029/2011JD017221, 2012.

15

25

Jensen, E. J., Toon, O. B., Pfister, L., and Selkirk, H. B.: Dehydration of the upper troposphere and lower stratosphere by subvisible cirrus clouds near the tropical tropopause, Geophys. Res. Lett., 23, 825–828, 1996a.

Jensen, E. J., Toon, O. B., Selkirk, H. B., Spinhirne, J. D., and Schoeberl, M. R.: On the for-

8, 10755–10792, 2015 Evaluation of two

Discussion

Paper

Discussion

Paper

Discussion Paper

Discussion Paper

vaisala RS92 radiosonde solar radiative dry bias correction algorithms

AMTD

A. M. Dzambo et al.

10777

- Liljegren, J. C., Clothiaux, E. E., Mace, G. G., Kato, S., and Dong, X.: A new retrieval for cloud liquid water path using a ground-based microwave radiometer and measurements of cloud temperature, J. Geophys. Res.-Atmos., 106, 14485–14500, 2001.
- Liou, K. N.: Influence of cirrus clouds on weather and climate processes: a global perspective, Mon. Weather Rev., 114, 1167–1200, doi:10.1175/1520-0493(1986)114<1167:IOCCOW>2.0.CO;2, 1986.
 - Marín, J. C., Pozo, D., Mlawer, E., Turner, D. D., and Curé, M.: Dynamics of local circulations in mountainous terrain during the RHUBC-II project, Mon. Weather Rev., 141, 3641–3656, 2013.
- ¹⁰ Mather, J. H. and J. W. Voyles: The ARM Climate Research facility: a review of structure and capabilities, B. Am. Meteorol. Soc., 94, 377–392, 2013.
 - Mattioli, V., Westwater, E. R., Cimini, D., Gasiewski, A. J., Klein, M., and Leuski, V. Y.: Microwave and millimeter-wave radiometric and radiosonde observations in an Arctic environment, J. Atmos. Ocean. Tech., 25, 1768–1777, doi:10.1175/2008JTECHA1078.1, 2008.
- Miloshevich, L. M., Paukkunen, A., Vömel, H., and Oltmans, S. J.: Development and validation of a time-lag correction for Vaisala radiosonde humidity measurements, J. Atmos. Ocean. Tech., 21, 1305–1327, 2004.
 - Miloshevich, L. M., Vömel, H., Whiteman, D. N., and Leblanc, T.: Accuracy assessment and correction of Vaisala RS92 radiosonde water vapor measurements, J. Geophys. Res.-Atmos,
- ²⁰ 114, D11305, doi:10.1029/2008JD011565, 2009.
- Parkinson, C. L.: Aqua: an earth-observing satellite mission to examine water and other climate variables, IEEE T. Geosci. Remote, 41, 173–183, doi:10.1109/TGRS.2002.808319, 2003.
- Revercomb, H. E., Turner, D. D., Tobin, D. C., Knuteson, R. O., Feltz, W. F., Barnard, J., Bösenberg, J., Clough, S., Cook, D., Ferrare, R., Goldsmith, J., Gutman, S., Halthore, R., Lesht,
- B., Liljegren, J., Linné, H., Michalsky, J., Morris, V., Porch, W., Richardson, S., Schmid, B., Splitt, M., Van Hove, T., Westwater, E., and Whiteman, D.: The ARM program's water vapor intensive observation periods: overview, initial accomplishments, and future challenges, B. Am. Meteorol. Soc., 84.2, 217–236, 2003.
- Rutllant Costa, J.: On the extreme aridity of costal and Atacama deserts in northern Chile, PhD thesis, University of Wisconsin, Madison, 175 pp, 1977.
- Seidel, D., Turner, D. D., Tobin, D. C., Knuteson, R. O., Feltz, W. F., Barnard, J., Bösenberg, J., Clough, S., Cook, D., Ferrare, R., Goldsmith, J., G utman, S., Halthore, R., Lesht, B., Liljegren, J., Linné, H., Michalsky, J., Morris, V., Porch, W., Richardson, S., Schmid, B.,

Splitt, M., Van Hove, T., Westwater, E., and Whiteman, D.: Reference upper-air observations for climate: rationale, progress, and plans, B. Am. Meteorol. Soc., 90, 361–369, doi:10.1175/2008BAMS2540.1, 2009.

Shimada, T. and Minobe, S.: Global analysis of the pressure adjustment mechanism over sea surface temperature fronts using AIRS/Agua data, Geophys. Res. Lett., 38, L06704,

doi:10.1029/2010GL046625, 2011.

Shu, S. and L. Wu: Analysis of the influence of Saharan air layer on tropical cyclone intensity using AIRS/Aqua data, Geophys. Res. Lett., 36, L09809, doi:10.1029/2009GL037634, 2009.

Soden, B. J., Turner, D. D., Lesht, B. M., and Miloshevich, L. M.: An analysis of

¹⁰ satellite, radiosonde, and lidar observations of upper tropospheric water vapor from the Atmospheric Radiation Measurement Program, J. Geophys. Res., 109, D04105, doi:10.1029/2003JD003828, 2004.

Starr, D. and Cox, S. K.: Cirrus clouds. Part II: Numerical experiments on the formation and maintenance of cirrus, J. Atmos. Sci., 42.23, 2682–2694, 1985.

Stokes, G. M. and Schwartz, S. E.: The atmospheric radiation measurement program: programmatic background and design of the cloud and radiation test bed, B. Am. Meteorol. Soc., 75, 1201–1221, 1994.

Suortti, T. M., Kivi, R., Kats, A., Yushkov, V., Kämpfer, N., Leiterer, U., and Vömel, H.: Tropospheric comparisons of Vaisala radiosondes and balloon-borne frost-point and Lyman-*a*

²⁰ hygrometers during the LAUTLOS-WAVVAP experiment, J. Atmos. Ocean. Tech., 25.2, 149– 166, 2008.

Tobin, D. C., Revercomb, H. E., Knuteson, R. O., Best, F. A., Smith, W. L., Ciganovich, N. N., Dedecker, R. G., Dutcher, S., Ellington, S. D., Garcia, R. K., Howell, H. B., LaPorte, D. D., Mango, S. A., Pagano, T. S., Taylor, J. K., van Delst, P., Vinson, K. H., and Werner, M.

W.: Radiometric and spectral validation of atmospheric infrared sounder observations with the aircraft-based scanning high-resolution interferometer sounder, J. Geophys. Res., 111, D09S02, doi:10.1029/2005JD006094, 2006.

Turner, D. D. and Goldsmith, J. E. M.: Twenty-four-hour Raman Lidar water vapor measurements during the Atmospheric Radiation Measurement program's 1996 and 1997 water va-

 por intensive observation periods, J. Atmos. Ocean. Tech., 16, 1062–1076, 1999.
 Turner, D. D. and Mlawer, E. J.: Radiative heating in underexplored bands campaign (RHUBC), B. Am. Meteorol. Soc., 91, 911–923, doi:10.1175/2010BAMS2904.1, 2010.

Turner, D. D., Lesht, B. M., Clough, S. A., Liljegren, J. C., Revercomb, H. E., and Tobin, D. C.: Dry bias and variability in Vaisala RS80-H radiosondes: the ARM experience, J. Atmos. Ocean. Tech., 20, 117–132, 2003.

Turner, D. D., Tobin, D. C., Clough, S. A., Brown, P. D., Ellingson, R. G., Mlawer, E. J., Knuteson,

- R. O., Revercomb, H. E., Shippert, T. R., Smith, W. L., and Shephard, M. W.: The QME AERI LBLRTM: a closure experiment for downwelling high spectral resolution infrared radiance, J. Atmos. Sci., 61, 2657–2675, 2004.
 - Turner, D. D., Clough, S. A., Liljegren, J. C., Clothiaux, E. E., Cady-Pereira, K., and Gaustad, K. L.: Retrieving liquid water path and precipitable water vapor from the atmospheric
- radiation measurement (ARM) microwave radiometers, IEEE T. Geosci. Remote, 45, 3680– 3690, 2007.
 - Vömel, H., Barnes, J. E., Forno, R. N., Fujiwara, M., Hasebe, F., Iwasaki, S., Kivi, R., Komala, N., Kyrö, E., Leblanc, T., Morel, B., Ogino, S.-Y., Read, W. G., Ryan, S. C., Saraspriya, S., Selkirk, H., Shiotani, M., Valverde-Canossa, J., and Whiteman, D. N.: Validation of Aura Microwave
- Limb Sounder water vapor by balloon-borne Cryogenic Frost point Hygrometer measurements, J. Geophys. Res.-Atmos., 112, D24S37, doi:10.1029/2007JD008698, 2007a.
 - Vömel, H., Selkirk, H., Miloshevich, L., Valverde-Canossa, J., Valdés, J., Kyrö, E., Kivi, R., Stolz, W., Peng, G., and Diaz, J. A.: Radiation dry bias of the Vaisala RS92 humidity sensor, J. Atmos. Ocean. Tech., 24, 953–963, 2007b.
- ²⁰ Wang, J., Carlson, D. J., Parsons, D. B., Hock, T. F., Lauritsen, D., Cole, H. L., and Chamberlain, E.: Performance of operational radiosonde humidity sensors in direct comparison with a chilled mirror dew-point hygrometer and its climate implication, Geophys. Res. Lett., 30, 1860, doi:10.1029/2003GL016985, 2003.

Wang, J., Zhang, L., Dai, A., Immler, F., Sommer, M., and Vömel, H.: Radiation dry bias correc-

- tion of Vaisala RS92 humidity data and its impacts on historical radiosonde data, J. Atmos. Ocean. Tech., 30, 197–214, 2013.
 - Yoneyama, K., Fujita, M., Sato, N., Fujiwara, M., Inai, Y., and Hasebe, F.: Correction for radiation dry bias found in RS92 radiosonde data during the MISMO field experiment, SOLA, 4, 13–16, 2008.
- Yu, H., Ciesielski, P. E., Wang, J., Kuo, H-C., Vömel, H., and Dirksen, R.: Evaluation of humidity correction methods for Vaisala RS92 tropical sounding data, J. Atmos. Ocean. Tech., 32, 397–411, 2015.

Discussion

Paper

Discussion

Paper

Discussion Paper

Discussion Paper

Printer-friendly Version Interactive Discussion

Full Screen / Esc

Close

Back

scussion Pa	AM 8, 10755–1	AMTD 8, 10755–10792, 2015					
aper Discussion Pape	Evaluation of two vaisala RS92 radiosonde solar radiative dry bias correction algorithms A. M. Dzambo et al.						
<u> </u>	Title Page						
	Abstract	Introduction					
scus	Conclusions	References					
sion P	Tables	Figures					
aper	14	►I					
_		•					
	Back	Close					
scussio	Full Scre	Full Screen / Esc					
on Pa	Printer-frier	Printer-friendly Version					
aper	Interactive Discussion						

Table 1. A summary of the microwave radiometer and radiosonde un/corrected PWV biases with ± 1 sigma uncertainty from the ARM's SGP, NSA and TWP (Darwin) sites.

	SGP Site, Lamont, OK, USA	NSA Site, Barrow, AK, USA	TWP Site, Darwin, Australia
(ORIG-MWR) PWV Bias N = 1745 points	$-0.07 \pm 0.08 \text{cm}$	$-0.02 \pm 0.03 \text{cm}$	$-0.20 \pm 0.13 \text{cm}$
(WANG-MWR) PWV Bias N = 371 points	$-0.02 \pm 0.06 \text{cm}$	$0.00 \pm 0.02 \text{cm}$	$-0.09 \pm 0.09 \mathrm{cm}$
(MILO-MWR) PWV Bias N = 1009 points	$-0.02 \pm 0.06 \text{cm}$	$-0.00 \pm 0.02 \text{cm}$	$-0.06 \pm 0.08 \text{cm}$

Table 2. A summary of the median T_b biases between the GVRP T_b and MonoRTM-derived T_b measurements using original radiosonde RH data and WANG/MILO corrected radiosonde RH data. Data is represented as a median bias with \pm one standard deviation.

GHz	170.0	172.0	174.0	176.0	178.0	179.0	180.0	181.0	182.0	183.0	183.31
Moist 30 % ORIG	0.70 ± 0.97 K	0.96 ±1.15 K	1.65 ±1.25 K	2.20 ±1.37 K	2.22 ±2.17 K	3.03 ±2.79 K	4.94 ±4.20 K	3.98 ±5.47 K	8.27 ±5.67 K	10.41 ±4.11 K	10.38 ±3.89 K
Moist 30 % WANG	–0.27 ±0.97 K	−0.22 ±1.12 K	0.14 ±1.16K	0.08 ±1.31 K	−1.09 ±2.07 K	−1.27 ±2.64 K	−0.65 ±4.08 K	–3.75 ±5.22 K	–2.46 ±5.67 K	−0.42 ±4.40 K	−0.33 ±4.09 K
Moist 30 % MILO	−0.34 ±0.96 K	–0.32 ±1.11 K	−0.01 ±1.15 K	–0.17 ±1.36 K	−1.63 ±2.08 K	–2.19 ±2.57 K	−2.14 ±4.08 K	–5.795 ±5.32 K	–3.737 ±5.65 K	–1.778 ±4.37 K	−1.75 ±4.06 K
Dry 30 % ORIG	-1.19 ±0.61 K	−1.56 ±0.62 K	−1.67 ±0.80 K	−1.58 ±0.65 K	−2.47 ±1.24 K	–2.71 ±1.55 K	-2.29 ±2.048 K	-4.906 ±2.90 K	−6.618 ±4.81 K	-3.208 ±5.89 K	−1.73 ±5.99 K
Dry 30 % WANG	−1.53 ±0.62 K	–1.98 ±0.66 K	-2.22 ±0.80 K	–2.37 ±0.68 K	–3.78 ±1.29 K	–4.54 ±1.65 K	–5.17 ±2.17 K	–9.977 ±3.08 K	–15.82 ±5.24 K	–14.75 ±6.67 K	−13.58 ±6.80 K
Dry 30 % MILO	−1.47 ±0.60 K	−1.90 ±0.65 K	–2.12 ±0.78 K	−2.22 ±0.67 K	−3.55 ±1.32 K	-4.21 ±1.66 K	-4.51 ±2.20 K	-8.349 ±3.08 K	–14.09 ±5.17 K	−15.67 ±6.62 K	–14.47 ±6.71 K

Table 3. A summary of the monthly brightness temperature thresholds used to screen cloudy sky scenes from the AIRS data.

Month	$T_{\rm B}$ Threshold (K)
Jan	0.99 K
Feb	1.64 K
Mar	1.82 K
Apr	2.29 K
May	2.13 K
Jun	2.57 K
Jul	2.58 K
Aug	2.66 K
Sep	2.35 K
Oct	1.88 K
Nov	1.28 K
Dec	0.90 K

Table 4. A summary of the brightness temperature biases between the AIRS and the LBLRTM derived data over the SGP site using un/corrected RH data as input as a function of height, where the height for each spectral residual was determined as the height where the weighting function for that profile peaks. The driest 30% and moistest 30% of the data correspond to upper-tropospheric IWV thresholds of less than 0.37 mm and greater than 0.96 mm respectively.

	Dry 30 % ORIG	Dry 30 % WANG	Dry 30 % MILO		Moist 30 % ORIG	Moist 30 % WANG	Moist 30 % MILO
3-4 km N=334	$-0.94 \pm 0.95 \mathrm{K}$	-0.63 ± 0.88 K	$-0.74 \pm 0.86 \mathrm{K}$	3–4 km N = 0			
4–5 km N = 767	$-0.72 \pm 0.84 \mathrm{K}$	$-0.37 \pm 0.76 \mathrm{K}$	$-0.43 \pm 0.76 \mathrm{K}$	4 - 5 km N = 0			
5–6 km N = 1076	$-0.48 \pm 0.69 \mathrm{K}$	$-0.22 \pm 0.66 \mathrm{K}$	$-0.21 \pm 0.66 \text{K}$	5–6 km N = 558	$-0.55 \pm 1.34 \text{K}$	$-0.20 \pm 1.26 \text{K}$	$-0.13 \pm 1.23 \text{K}$
6–7 km N = 681	$-0.92 \pm 0.45 \mathrm{K}$	$-0.60 \pm 0.42 \text{K}$	$-0.59 \pm 0.41 \text{K}$	6–7 km N = 2061	$-0.38 \pm 0.78 \text{K}$	$-0.08 \pm 0.77 \text{K}$	$-0.04 \pm 0.77 \text{K}$
7–8 km N = 952	$-0.91 \pm 0.75 \text{K}$	$-0.60 \pm 0.69 \mathrm{K}$	$-0.57 \pm 0.70 \text{K}$	7–8 km N = 1277	$-0.41 \pm 0.70 \text{K}$	-0.12 ± 0.71 K	–0.01 ± 0.72 K
8–9 km N = 498	$-0.85 \pm 0.61 \text{K}$	-0.54 ± 0.54 K	$-0.46 \pm 0.51 \text{K}$	8–9 km N = 1307	$-0.59 \pm 0.49 \text{K}$	$-0.25 \pm 0.49 \mathrm{K}$	$-0.10 \pm 0.52 \text{K}$
9–10 km N = 532	$-0.77 \pm 0.51 \mathrm{K}$	$-0.45 \pm 0.41 \text{ K}$	$-0.38 \pm 0.41 \text{K}$	9–10 km N = 658	$-0.49 \pm 0.42 \text{K}$	$-0.20 \pm 0.42 \mathrm{K}$	$-0.06 \pm 0.43 \text{K}$
10–11 km N = 255	$-0.78 \pm 0.43 \mathrm{K}$	$-0.44 \pm 0.37 \text{K}$	$-0.36 \pm 0.37 \text{K}$	10–11 km N = 1247	$-0.40 \pm 0.44 \text{K}$	$-0.13 \pm 0.42 \text{K}$	0.01 ± 0.42 K
11–12 km N = 191	$-0.66 \pm 0.62 \mathrm{K}$	$-0.32 \pm 0.55 \mathrm{K}$	$-0.18 \pm 0.49 \text{K}$	11–12 km N = 421	$-0.47 \pm 0.33 \text{K}$	$-0.14 \pm 0.27 \text{K}$	$0.04 \pm 0.25 \text{K}$
12-13 km N = 0				12–13 km N = 82	$-0.23 \pm 0.37 \text{K}$	$0.07\pm0.33\mathrm{K}$	$0.21 \pm 0.31 \text{K}$

Discussion Pa	AM 8, 10755–1	AMTD 8, 10755–10792, 2015					
aper Discussion Pa	Evaluatio vaisala radiosor radiative correction A. M. Dza	Evaluation of two vaisala RS92 radiosonde solar radiative dry bias correction algorithms A. M. Dzambo et al.					
per	Title	Title Page					
	Abstract	Introduction					
scus	Conclusions	References					
sion	Tables	Figures					
Paper	I	►I					
—	•	•					
	Back	Close					
scussic	Full Screen / Esc						
on Pa	Printer-frier	Printer-friendly Version					
aper	Interactive Discussion						

Figure 1. A comparison of the WANG and MILO corrected RH profiles (red and green, respectively) compared to the original RH profile (black). The light blue line represents the saturation RH with respect to ice. This example is the 18Z sounding for the SGP site from 15 June 2006.

Discussion Paper

Figure 2. The mean relative increase in the water vapor mixing ratio caused by the two correction algorithms for RS92 radiosondes launched at the SGP, NSA, TWP, and CJC sites (left), and the standard deviation (right) as a function of height. The MILO (WANG) corrected data are shown with dotted (solid) lines. The number of comparisons for each site is shown in the figure. NSA results are only shown up to the mean tropopause height (10 km). The inset plot on the main figure is the mean relative increase in the water vapor mixing ratio caused by the two correction algorithms, but only from 0 to 4 km.

Figure 3. A comparison between the PWV derived from the original radiosonde data (top), WANG-corrected (middle), and MILO-corrected (bottom) radiosonde data with the PWV derived from the collocated MWR at the SGP site (panels **a**1, **a**2, and **a**3), NSA site (panels **b**1, **b**2, and **b**3) and TWP Darwin site (panels **c**1, **c**2, and **c**3). The solid black line superimposed on the data denotes the mean values for each PWV bin, and the vertical lines represent the standard deviations.

Figure 4. Downwelling brightness temperature differences between MonoRTM calculations using the WANG and MILO corrected RH profile as input. Data is sorted by the moistest 30 % and driest 30 % of all profiles in the CJC data set (green and orange, respectively). The thick black lines are the mean spectral residual for the two subsets of data.

Figure 5. Median GVRP minus MonoRTM-computed spectral residuals, where the MonoRTM was driven by WANG- and MILO-corrected radiosondes (red/green and orange/blue, respectively) and uncorrected radiosondes (gray lines). These median residuals were computed for the moistest and driest 30 % of the CJC radiosondes, as shown in Fig. 4.

Discussion Paper

Discussion Paper

Figure 6. Median RH profiles for four arm sites. RH is grouped in 25 hPa bins (starting at 1000 hPa), and the median is computed from that bin. There are 142 soundings for the CJC site, 2500 soundings across the annual cycle for the SGP and TWP (Nauru) sites, and 1712 soundings for the NSA site.

Figure 7. Distributions of upper-tropospheric integrated water vapor (IWV) from 530 to 200 hPa for four ARM sites. The mean surface pressure at the CJC site is 530 hPa, while 200 hPa is the approximate height of the tropopause.

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Figure 8. The median AIRS minus LBLRTM brightness temperature difference (residual) as a function of height (for all data), where the residual in a spectral channel was assigned to a particular height (in 1 km intervals) based upon where the weighting function for that channel peaks with altitude. The height each channel was determined is based on the original RH profile, and the number of brightness temperature residuals averaged at each height bin are denoted to the right. Error bars represent the 25th/75th percentile of brightness temperature residuals.

Figure 9. Same as in Fig. 8, but where the residuals are for the moistest 30% and driest 30% of the water vapor profiles. The median values shown in this plot, along with the standard deviations, are given in greater detail in Table 3.

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Close

Back

Discussion Paper