

## ***Interactive comment on “The SPectrometer for Ice Nuclei (SPIN): An instrument to investigate ice nucleation” by S. Garimella et al.***

### **Anonymous Referee #2**

Received and published: 9 March 2016

One more point of importance with regard to using the homogeneous freezing line of Koop. Typically a nucleation rate must be specified when plotting this curve. For example, the value  $J = 5 \times 10^8 \text{ cm}^{-3}\text{s}^{-1}$  from Hoose and Möhler (2012) is often used.

### **References**

Hoose, C. and Möhler, O.: Heterogeneous ice nucleation on atmospheric aerosols: a review of results from laboratory experiments, Atmos. Chem. Phys., 12, 9817–9854, 10.5194/acp-12-

Printer-friendly version

Discussion paper



9817-2012, 2012.

---

Interactive comment on Atmos. Meas. Tech. Discuss., doi:10.5194/amt-2015-400, 2016.

**AMTD**

---

Interactive  
comment

Printer-friendly version

Discussion paper

