

Comparison of cavity enhanced optical–feedback laser spectroscopy and gas chromatography for ground-based and airborne measurements of atmospheric CO concentration

Irène Ventrillard¹, Irène Xueref-Remy², Martina Schmidt^{2,3}, Camille Yver Kwok², Xavier Faïn⁴, and Daniele Romanini¹

¹Univ. Grenoble Alpes, CNRS-UMR5588, Laboratoire Interdisciplinaire de Physique (LIPhy), Grenoble, France

²Laboratoire des Sciences du Climat et de l'Environnement (LSCE), UMR CEA-CNRS 1572, Gif-sur-Yvette, France

³Now at Institut für Umweltphysik (IUP) University Heidelberg, Germany

⁴Univ. Grenoble Alpes, Institut des Géosciences de l'Environnement (IGE), Grenoble, France

Correspondence to: Irène Ventrillard (irene.ventrillard@univ-grenoble-alpes.fr)

Abstract.

We present the first comparison of carbon monoxide (CO) measurements performed with a portable laser spectrometer that exploits the Optical–Feedback Cavity–Enhanced Absorption Spectroscopy (OF-CEAS) technique, against a high performance automated gas chromatograph (GC) with mercuric oxide reduction gas detector. First, measurements of atmospheric CO mole fraction were continuously collected in Paris (France) suburb over one week. Both instruments showed an excellent agreement within typically 2 ppb (part per billion in volume) fulfilling the World Meteorological Organization (WMO) recommendation for CO inter-laboratory comparison. The compact size and robustness of the OF-CEAS instrument allowed its operation aboard a small aircraft employed for routine tropospheric air analysis over the French Orléans forest area. Direct OF-CEAS real-time CO measurements in tropospheric air were then compared with later analysis of flask samples by the gas chromatograph. Again, a very good agreement was observed. This work establishes that the OF-CEAS laser spectrometer can run unattended at a very high level of sensitivity (<1 ppb) and stability without any periodic calibration.

1 Introduction

Carbon monoxide (CO) is a reactive trace gas that plays a significant role in global atmospheric chemistry by being a major sink of tropospheric hydroxyl radicals (OH). Hydroxyl radical is the main tropospheric oxidant, thus its abundance affects the lifetimes of radiatively important gases such as methane. Oxidation of CO by OH also provides a source or a sink, respectively in high or low NO_x conditions, for tropospheric ozone (Logan et al., 1981). CO concentration in the atmosphere have thus crucial implications for both climate and air quality issues, and accurate CO measurements in the troposphere are important when modeling climate-chemistry interactions with global coupled models (Voulgarakis et al., 2013).

Consequently, monitoring of tropospheric CO has been conducted over the last decades on the global scale (Novelli et al., 1998). Recently satellite-based observations have become an important contribution to regional monitoring of atmospheric CO (Worden et al., 2013). There is still a need, however, to strengthen direct CO observations from both surface stations and

aircraft to assess the large spatio-temporal variability of CO, especially within the boundary layer at the regional scale for a better understanding of atmospheric chemistry and transport, and towards improving forecast modeling of air quality (Sahu et al., 2013; Warner et al., 2013; Té et al., 2016).

Historical methods, such as gas chromatography, have been used for many years for surface monitoring of CO (Derwent et al., 2001; Langenfelds et al., 2002; Yver et al., 2009; Schmidt et al., 2014). A gas chromatograph (GC) equipped with mercuric oxide reduction gas detector allows for very sensitive laboratory measurements but requires hourly calibration procedures with calibration gases and an expert operator to achieve uniform high-quality results. In addition, the mercuric oxide reduction detectors are known for their non-linear response function, which needs to be quantified on a regular basis several times per year (Yver et al., 2009). However, recent developments in optical spectroscopy methods have brought new alternatives for in-situ CO monitoring (Zellweger et al., 2012; Chen et al., 2013; Yver Kwok et al., 2015). The most sensitive optical techniques allow detection limit at the ppb level and below. Among them, Optical-Feedback Cavity Enhanced Absorption Spectroscopy (OF-CEAS) (Morville et al., 2005) exploits a high finesse optical cavity in which a laser source is coupled to enhance the interaction of photons with gas molecules present inside the cavity (Morville et al., 2014). OF-CEAS offers many advantages for quantitative and selective trace gas analysis: it allows real-time absolute measurements with a smallest detectable absorption coefficient in the range of a few $10^{-10}/\text{cm}$ for 1 s acquisition time (Landsberg et al., 2014), it does not require periodic calibrations with certified gas mixtures, its sampling volume is small (20 cm^3), its response time can be faster than 1 s, and it enables the development of compact instruments to be operated by non-specialists.

Another advantage that follows from the high sensitivity of the OF-CEAS technique is the ability to work in the near infrared region (NIR) where widely used optics are commercially available together with room temperature lasers and detectors. Traditional near infrared (NIR) OF-CEAS instruments reach limit of detection (LOD) at the sub-ppb level for CO (Faïn et al., 2014) that is comparable to other instruments exploiting the mid infrared (MIR) spectral region where absorption coefficient are typically two orders of magnitude higher. Indeed, commercial MIR laser spectrometers based on different laser spectroscopy techniques offer CO sub ppb LOD, such as Picarro instruments by CRDS with a resonant cavity or instruments exploiting a multipass cell like Aerodyne and Los Gatos products. The performance of the OF-CEAS technique in the NIR led a private company (AP2E, Aix-en-Provence, France) to exploit the patent for commercially available analyzers (namely ProCEAS). Exploiting the MIR with OF-CEAS instruments allows to reach sub-ppb levels for several species of interest in trace detection and ppm levels for isotopic ratio measurements (Maisons et al., 2010; Gorrotxategi-Carbajo et al., 2013; Manfred et al., 2016; Richard et al., 2016).

OF-CEAS based measurements of CO have been conducted before around various applications, for example for in-situ trace measurements on geothermal gases (Kassi et al., 2006), for continuous and high resolution measurement of air extracted from ice cores drilled out of polar glaciers (Faïn et al., 2014) or for breath analysis in different medical settings (Ventrillard-Courtillot et al., 2009; Maignan et al., 2014). ProCEAS analyzers are now commercialized in the domains of industrial and air quality monitoring, with some very stringent applications such as air quality control onboard nuclear submarines. In order to further establish for different user communities that OF-CEAS can become a work-horse in many CO applications, which demand robust and compact instrumentation with ppb sensitivity and a fast response time, this paper reports on the comparison of

CO measurements performed by OF-CEAS against those obtained by the well established gas chromatography technique. GC measurements were done with a high performance gas chromatograph equipped with a mercuric oxide reduction gas detector (Yver et al., 2009). First, the atmospheric CO concentration in Gif-sur-Yvette, France, was continuously analyzed at ground level over one week. Then, the OF-CEAS instrument was set aboard a small aircraft employed for periodic tropospheric air measurements over the French Orléans forest area. Airborne in-situ CO measurements by OF-CEAS were then compared with flask samples later analyzed with the GC at LSCE.

All values reported in this paper are dry air mole fractions (expressed in ppm or ppb) but are called concentration as commonly done by the community.

10 2 Materials and Methods

We briefly describe the GC set-up and outline the OF-CEAS technique, highlighting the characteristics most relevant for the measurements reported here such as instruments calibrations. In particular two steps of post data processing were needed to come to an excellent agreement between the optical and chromatographic measurements performed during autumn 2006. Firstly, the non-linearity of the GC reduction gas detector was corrected following a procedure established in 2010. Secondly, the two instruments had to be calibrated on the same standard scale (WMO CO X2004). This was performed with a recent re-evaluation (in 2014) on this scale of the gas standards initially used for the OF-CEAS spectrometer calibration.

2.1 Gas chromatograph

The LSCE laboratory at Gif-sur-Yvette is equipped with two coupled gas chromatographs (HP-6890, Agilent and PP1, Peak laboratories) which run fully automated, alternating between calibration gas and ambient air, in order to analyse CO, H₂, CO₂, CH₄, N₂O and SF₆ concentration in atmospheric measurements, flask samples or high pressure cylinders. Detailed descriptions of the GC system for CO analysis is given by Yver et al. (2009). CO is analyzed with the PP1 chromatograph equipped with a reduction gas detector (RGD) after reduction of mercuric oxide and detection of mercury vapor by UV absorption. Each analysis takes less than 6 min allowing between two and six injections of ambient air alternating with calibration gases and flask samples. The air is dried before injection in two steps. First, it passes through a glass trap which is hosted in a commercial refrigerator kept at 5° C in order to remove a large fraction of water vapor and in a second step, air is further dried by passing through a second glass trap cooled in an ethanol bath at -55° C using a cryogenic cooler (designed as the "cooling trap" in the following). An operator is only required to change the cooling trap 2-3 times per week and to restart the acquisition.

Calibration of the gas chromatograph: Correction of the reduction gas detector nonlinearity

The GC is calibrated for CO with cylinders certified by NOAA/GMD on the WMO CO X2004 scale (Novelli et al., 1994). CO concentrations are calculated using regular measurements of one calibration cylinder with a typical atmospheric concentration value (here 168.0 ± 0.8 ppb) and a non linear correction function of the detector response as described in Yver et al. (2009) and Yver (2010). The correction function is determined on an annual frequency using a set of 5 cylinders with a CO

concentration ranges from 57 ± 1.0 ppb to $523 \text{ ppb} \pm 10.9$ ppb, and applied as a post-run correction (Yver et al., 2009). This non linear correction was validated using flask measurement comparisons between LSCE and NOAA, with a mean difference of 4.5 ± 2.2 ppb for the period of July 2006 to July 2009. For the one week comparison campaign with the OF-CEAS instrument in November 2006, the correction function applied to CO in-situ measurements by the GC (CO_{meas}) to obtain the calibrated CO concentrations reported in the following is given by :

$$\Delta CO_{corr} = 11.4 + 0.077 \times CO_{meas} - 7.1 \cdot 10^{-4} \times (CO_{meas})^2 + 1.03 \cdot 10^{-8} \times (CO_{meas})^3 \quad (1)$$

It applies a correction for the non-linear behavior of the analyser in the range of - 15 to + 15 ppb for measured CO concentration up to 500 ppb.

The calibration cylinder is analyzed every 30-40 minutes as well as a quality control gas, a so called target gas, with a CO concentration of 68 ppb, that is treated as unknown. Over the entire comparison period, the repeatability defined as 1-sigma standard deviation of the target gas is 0.4 ppb.

The flask samples filled during the airborne campaign are measured in a similar way as the ambient air concentration with two injections per flask.

2.2 Optical-Feedback Cavity Enhanced Absorption Spectrometer

The laser spectroscopy technique going under the name of OF-CEAS was introduced by Morville et al. (2005) and has been further detailed in different publications (Kerstel et al., 2006; Kassi et al., 2006; Ventrillard-Courtillot et al., 2009; Faïn et al., 2014; Maignan et al., 2014; Morville et al., 2014). In particular the OF-CEAS instrument used in this study has been described in Kassi et al. (2006). It provides in situ CO measurements with a detection limit of 0.2 ppb in 20 s (Faïn et al., 2014) with no calibration and running unattended. Here we will just recall the basic principle of OF-CEAS.

Spectroscopic measurements of trace gas concentrations require a long light absorption path. Like other spectroscopy techniques, OF-CEAS is based upon the use of a sample cell made with an optical cavity in order to enhance light interaction with the gas sample. Specifically in the spectrometer used here for CO monitoring, the resonant optical cavity composed of high-reflective mirrors (mirror reflectivity; $R \simeq 99.995\%$) allows a 20 km effective absorption length with a compact set-up: the cavity length is only 1 m long, folded to 0.5 m external size. The complexity of using a resonant cavity with high reflectivity mirrors (thus very small transmissivity) resides in the coupling of a sufficient amount of light in the cavity by injecting laser light through one of the cavity mirrors. The originality of OF-CEAS is that the optical cavity is made of three mirrors placed in a "V-shaped" configuration. In this way, a fraction of the light trapped inside the optical cavity, and therefore frequency-selected by a resonant mode of the cavity, can be returned to the laser. The non-linear response of the laser then forces it to lase on the exact frequency of the excited cavity "mode". This optical feedback (OF) effect is also responsible for a narrowing of the laser emission line width and an increase of the cavity transmission to a level that is orders of magnitude larger than in competing techniques (Morville et al., 2005). OF-CEAS absorption spectra are acquired on a small spectral region, as shown in Fig. 1 for the present case, by scanning the laser frequency at a fast acquisition rate (6 Hz here). However the measurement response time for $1/e$ change in a concentration value is not limited by this rate but by the gas exchange rate inside the sample volume.




Figure 1. Single OF-CEAS spectra in absolute absorption units, recorded in 150 ms with a gas sample at a pressure of 200 mbar and a temperature of 295 K. CO and CH₄ concentrations are deduced from the real time fit. The standard deviation of the residuals is $9 \times 10^{-10} \text{ cm}^{-1}$ (in absorption units). The base line has been subtracted for comparison with HITRAN simulated absorption spectra (Rothman et al., 2013). HITRAN spectra and the residuals have been offset for clarity.

Therefore, the cell is designed to allow minimal dead spaces and a small internal (sample) volume, which does not exceed 18 cm³. The gas is continuously flowing and the cell pressure is stabilized with a downstream pressure regulator, to 200 mbar
 5 in this study. The flow is adjusted manually with a needle valve at the inlet of the sampling cell, to 250 sccm (standard cm cubes per minute) for ground measurements and around 50 sccm for airborne measurements. The corresponding gas exchange times are then 0.9 s and 4.3 s respectively. If needed, shorter response time can be obtained by using a lower sample pressure or a higher pumping rate. The design of the spectrometer is robust and compact: the optical assembly and all the electronics, for real time control and data acquisition, fit inside a 19" chassis where the V-shaped cavity is placed in the diagonal as shown
 10 in figure 1 of Kassi et al. (2006). The device is temperature stabilized around 22°C using heating adhesive ribbons.

Importantly, OF-CEAS provides quantitative absorption measurements in real-time without the need for a periodic calibration with certified gas mixtures. A normalization procedure of the absorbance scale is realized continuously based on cavity optical loss measurements performed by Cavity Ring Down Spectroscopy (CRDS) (Kerstel et al., 2006; Morville et al., 2014). Pressure stabilization inside the cavity allows for a real-time numerical fit of the measured absorption spectra with a reduced number of parameters (Gorrotxategi-Carbajo et al., 2013). This enables the selective determination of the concentrations of all compounds that possess absorption lines in the selected spectral window (CO and CH₄ here). This is a key point in trace gas monitoring, atmospheric air being a highly complex gas mixture. To optimize CO detection limit, the laser emission is chosen

in the near infrared region, in the [2.3- 2.4] μm range (Fig. 1), an interesting region that includes an atmospheric window where water vapor absorption lines are sparse and weak, while several light species such as CO display relatively strong absorption bands. The smallest detectable absorption coefficient is typically in the range of several $10^{-10}/\text{cm}$. As a result, an OF-CEAS instrument optimized for CO monitoring has a LOD of 0.2 ppb of CO for an acquisition time of 20 s, value derived from the Allan variance in Faïn et al. (2014).

Fast response time and low LOD allow OF-CEAS instruments to perform fast trace gas monitoring. This has already been exploited in airborne atmospheric measurements -of methane in Romanini et al. (2006), water isotopes in Iannone et al. (2009a) and Kerstel et al. (2006), and CO in this work- as well as in other fields with laboratory prototypes and commercial instruments (ProCEAS) as mentioned in the introduction section.

Calibration of the OF-CEAS Spectrometer: Conversion of absolute molecular absorption to CO concentration

The ring-down calibration included in the OF-CEAS technique allows for direct absolute molecular absorption measurements (in cm^{-1} unit, Fig. 1). Then line intensity is directly converted to CO concentration with a conversion factor specific to the fitted absorption line for the temperature and pressure operation conditions (298.5 ± 1 K and 200 ± 1.8 mbar in this work). This allows to account for temperature and pressure effects on the line intensity parameters. It is important to stress that this factor is a constant that does not depend on the cavity finesse (continuously measured by ring-down) nor on the gas sample composition (the multiline fit allows independent fit of each species) as far as the foreign pressure broadening effect on CO absorption lines from water can be neglected. This is justified for atmospheric measurements where water concentration remains small (it varies from 0.8% to 1.6% for ground based measurements reported in this work). As a consequence, the conversion factor of each molecule needs to be determined for the OF-CEAS spectrometer working conditions only once and then during operation the instrument delivers absolute concentrations in real-time without the need of any calibration with certified mixture.

The conversion factor can be derived from spectral database or by direct calibration using certified mixture. Even if the line intensities for CO in this spectral region are well defined -to better than 1% in the HITRAN database (Li et al., 2015)-, in practice calibration with gas standards is found to be more accurate because it cancels sensors pressure and temperature absolute accuracy and allows to minimize line profile effects by considering a specific model in the fit procedure -a Rautian model is used as in Gorrotxategi-Carbajo et al. (2013).

For the comparison campaign in 2006, the OF-CEAS spectrometer was calibrated with 2 high pressure cylinders containing air whose CO concentration had been certified in 1995 by the Commonwealth Scientific and Industrial Research Organisation (CSIRO). However, the GC was calibrated on the WMO X2004 scale provided by the NOAA. Differences between CSIRO and NOAA CO scales in the order of 6 ppb have been reported by Masarie et al. (2001). Therefore we reevaluated CO concentrations in the CSIRO cylinders against the WMO CO X2004 scale. This was done in 2014 using another OF-CEAS instrument, designed for ice core analysis (Faïn et al., 2014). This instrument was calibrated with three standards certified in 2011 by the NOAA GMD (Global Monitoring Division) Carbon Cycle Group on the WMO CO X2004 scale (33.2 ± 0.5 ppb, 51.8 ± 0.1 ppb and 102.1 ± 0.1 ppb of CO). The two working standards used to calibrate the instrument for the 2006 campaign, were recalibrated.

brated to 35.0 ± 1.5 ppb and 104.0 ± 1.5 ppb while the CSIRO values certified in 1995 were 32.6 ± 0.7 ppb and 98.7 ± 1 ppb, respectively.

5 These WMO-scaled CO standard gases values were then used to calibrate the entire 2006 dataset using a linear relationship (i.e., without offset adjustment), which is consistent with the fact that the zero of spectral measurements is intrinsically accurate. Furthermore, the high linearity of OF-CEAS was previously reported for concentrations ranging over more than three decades (data published for water measurements in Iannone et al. (2009b)). The accuracy of this calibration is estimated to be of 2 % limited by the accuracy of the NOAA standards. But the obtained conversion factor corresponds to a 10% overestimation of
10 the line intensity specified in HITRAN with a 1% accuracy. CSIRO specifications of the two standards being offset by 5% and 7% as compare to NOAA standards are neither compatible with HITRAN database. Nonetheless, the good agreement of OF-CEAS and GC measurements reported in the following shows that this calibration on the same reference scale is a crucial point for the inter-comparison.

The LOD of the OF-CEAS spectrometer is much smaller than the accuracy of the NOAA standards, at the level of 0.2 ppb
15 for an acquisition time of 20 s. At longer acquisition times, small drifts prevent for better averaging. It is partly to be attributed to drifts in the sensors that are used to control sample pressure and temperature, thus selection of more stable sensors can decrease the drift. Other causes of drift are changes in parasitic optical etalon effects (Morville et al., 2014). However the drifts associated to these optical effects can be made quite small and cannot increase arbitrarily and remain bounded at all times as shown by the Allan variance of CO measurements in Faïn et al. (2014).

20 3 Comparison: Results and discussion

3.1 In-situ ground measurements

Direct comparison of atmospheric CO concentration measurements by GC and OF-CEAS over one week (8/11/2006-14/11/2006), was performed at LSCE in Gif-sur-Yvette, 20 km southwest of Paris ($48^{\circ}43'N$ / $02^{\circ}09'E$ / 120 m above sea level). The LSCE GC setup routinely monitors atmospheric concentration with a sampling inlet located on the roof of the building, 7 m above
25 ground level. The OF-CEAS instrument from LIPhy was set to run in the same building but with an independent sampling line. Sampling lines measured about 20 m and were made of 3/8" diameter Dekabon tubes. The estimated sample propagation delay along the tube from the roof to the OF-CEAS instrument is about 6 min (with a gas flow of 250 sccm). A larger delay is observed on the GC data due to the use of the cold trap. The volume of this trap corresponds to the sample volume collected over about 15 min by the GC, inducing a smoothing of the signal of the semicontinuous injections. To eliminate the time delay
30 between both instruments, the time shift was fixed to 14 min (Fig. 2).

Reported GC CO values are dry air mole fractions. For the comparison, OF-CEAS CO mole fraction ($x_{CO,air}$) is converted into dry air mole fraction ($x_{CO,dry}$) according to:

$$x_{CO,dry} = \frac{x_{CO,air}}{1 - x_{H_2O}} = \frac{x_{CO,air}}{1 - [RH \times e(T)]/P}. \quad (2)$$

Where the water mole fraction (x_{H_2O}) is computed from the humidity rate (RH), the atmospheric pressure (P) and the water saturation vapor pressure ($e(T)$) given by a polynomial function of the atmospheric temperature (T) (Lowe, 1976). The meteorological data used (RH , P and T) are routinely monitored at the Saclay tower located about 1.5 km North-Northwest
5 from the sampling point (data provided by the SPR group from Saclay CEA).

OF-CEAS and GC raw data were post-treated as detailed in section 2. In Fig. 2 are shown typical variations of atmospheric CO dry concentration measured by both the GC and the OF-CEAS analysers during a week day (2.a), a Sunday (2.b) and a Wednesday night (2.c). During night-time and most of the day on the week-ends, CO concentration slowly varies within typically 100-300 ppb. But during week days emissions from nearby traffic usually induce two rush hour peaks in the morning
10 at around 8 am and in the evening after 5 pm. The persistence and higher intensity of the second CO peak could be related to air mass change or to the Friday evening traffic jams all around Paris suburbs (Fig. 2.a). During daytime and in the evening, the fast response time of the OF-CEAS instrument (1 s averaged here to 2 s) allows to record many short but very strong peaks (sometime rising up to more than 1 ppm during only 1 or 2 minutes). These are due to local pollution of vehicles passing by the laboratory. The 14/11/2006 night (Fig. 2.c) when CO concentration remained around 100 ppb and very small concentration
15 fluctuations are measured (less than 20 ppb over 7 hours), allows a comparison over several hours with nearly no effect from the slower GC response time.

OF-CEAS and GC measurements show an excellent agreement. When CO concentration varies slowly, such as during night-time and Sunday measurements shown in Fig. 2, the agreement is within about 2 ppb rms over several hours for concentration values ranging from 100 ppb to 300 ppb for the whole comparison period of one week. This difference is fully compatible with
20 the calibration accuracy of the two instruments reported before. When CO concentration is subject to fast changes such as in Fig. 2.a, the strong difference in the GC and OF-CEAS measurements is explained by the slower response time of the GC instrument due to the buffering effect of the cooling trap. The air sample being continuously flushed in the cooling trap, the effect on CO concentration measurements by the GC is not equivalent to a simple time average. A more complex weighted moving averaging could be performed on the faster OF-CEAS measurement to try to mimic the GC measurement, but a study
25 concerning this averaging issue appears to be beyond the scope of this paper.

The OF-CEAS instrument measures CO and CH₄ at the same time (Fig. 1). Contrary to CO, CH₄ concentration is not sensitive to traffic pollution. Daily variability is usually less than 10% with a background value of about 1900 ppb. The rms noise of the OF-CEAS measurements for CH₄ is 4 ppb for an averaging time of 20 s. A good agreement between OF-CEAS and GC is also found with maximum deviations of ± 20 ppb, corresponding to about 1% in relative unit. Such performance
30 has been previously reported in Romanini et al. (2006) with a similar OF-CEAS instrument compared to the same GC.

3.2 Airborne measurements

In the framework of the French RAMCES observation network for greenhouse gases monitoring, regular weekly flights have been carried out by LSCE since 1996 above the Orléans forest, located about 100 km south of Gif-sur-Yvette. This flight program aims at improving our understanding of transport processes into the atmospheric boundary layer, and to better assess the relative role of local, regional and continental anthropogenic and biospheric fluxes on the observed trace gas concentrations.

Especially, vertical profiles of trace gases are very useful for assessing atmospheric transport model performances. During the flights, air samples are collected in flasks, as described in Chevalier et al. (2009), and later analyzed at LSCE by GC to measure the concentration of CO₂, CH₄, CO, N₂O and SF₆ (Xueref-Remy et al., 2011; Haszpra et al., 2012). Glass flasks are filled at 5 10 different altitudes (between 100 m and 3000 m above ground level). Those used for the present comparison were analyzed one week after collection. On the 15/11/2006, the OF-CEAS instrument was installed inside the aircraft to measure in situ CO atmospheric concentration during the entire flight. The OF-CEAS instrument was mounted in a 19'' rack fixed in place of a seat.

Tropospheric air was sampled upwind of the aircraft engines exhaust: a 2 m Dekabon inlet line carries outside air to the 10 set-up entrance, passing through a customized window of the aircraft. The same inlet was used for the OF-CEAS instrument. For the GC, the sampling unit consists of a diaphragm pump which draws air through the chemical drying cartridge filled with Mg(ClO₄)₂. Air is collected in 1 L glass flasks sealed with PTFE O-rings. Flasks are collected in pairs and pressurized to 2 bar absolute pressure. The filling step takes between 30 s and 1 min during which the plane covers a typical distance of 5 km.

The entire set of measurements is shown in Fig. 3: starting from the airport of Toussus-le-Noble (48°45'N/ 2°08'E/ 164 m 15 asl), during the flight to the Orléans forest area (47°50'N/ 2°30'E/ 135 m asl) where the plane starts a routine flight that consists of legs at the 10 pre-defined altitudes for the flask samples collection, and during flying back to the airport. During the flight above the Orléans forest, CO levels remain around 90 ppb (± 10 ppb) above 1000 m, while an increase at lower altitude is clearly measured up to 150 ppb at 100 m due to surface CO sources like traffic and heating.

It should be highlighted that the OF-CEAS instrument is robust enough to operate in the harsh environment of a small 20 aircraft including during take-off and landing phases. On the tarmac, CO rises up to 16 ppm due to airplane exhaust gases. This illustrates the wide sensitivity range of the measurements, of about 4 orders of magnitude. During the whole flight, the instrument ran un-attended. The only not automatized action by the operator required during the flight consisted in adjusting the flow in the sampling cell with a needle valve. Given that this valve was placed at instrument inlet, the flow changed linearly with pressure, thus decreased with altitude. However for GC comparison, measurements were taken during constant altitude 25 sections, allowing averaging on time scales largely exceeding the sample exchange time for flow around 50 sccm. Later, flow regulation has been automatized using a numerically controlled flow regulator. During the flight, the flow was slowly varying between 40 sccm and 70 sccm. OF-CEAS data are averaged for 5.5 s to be consistent with the largest value of the response time. It corresponds to a space resolution of 300 m according to the aircraft velocity. Due to the harsh environment in the plane, the standard deviation of the measurements is increased to typically 2 ppb (zoom in Fig. 3) while for ground-based measurements 30 it is 0.6 ppb for 2 s averaging time (Fig. 2.c).

As explained in the previous sections, for the comparison of GC and OF-CEAS measurements post data processing was performed to correct for the RGD non linearity and to bring both instruments on the same calibration scale. Additionally, OF-CEAS concentrations have to be expressed in dry air. The water mole fraction was not monitored during the flight but was derived from a model allowing to compute the specific humidity q from meteorological data (analysis from the European

Center for Medium-Range Weather Forecasts -ECMWF). Water mole fraction is then given by:

$$x_{H_2O} = \frac{q \times M_{dry}}{M_{H_2O} + q \times (M_{dry} - M_{H_2O})}. \quad (3)$$

Where M_{H_2O} and M_{dry} are respectively the molar mass of water and dry air. During the flight, the values obtained for the water mole fraction vary typically from 0.2% to 1% at respectively high and low altitude inducing a correction on CO values between +0.2 to +1.6 ppb. The model was compared to the meteorological data (RH , P and T) provided by the closest station that is the radiosounding of Trappes (48°46'N/ 2°1'E/ 168 m asl). These data are recorded by Meteo-France and are available on the SIRTA website (sir). Humidity rates derived from ECWMF model and Trappes data are in agreement within 20%, which means that corrections obtained from one or the other model will be closer than the measurement error.

In the bottom of Fig. 3 is plotted the difference between OF-CEAS and GC CO concentrations. To be consistent with the typical filling duration of the flasks, OF-CEAS values are averaged during 1 min around the flask filling times. A good agreement is obtained: for the set of ten measurements recorded at different altitudes, the difference has a mean value of -2.2 ppb with a standard deviation of 1.7 ppb. This small systematic difference could not be explained even when different effects like residual effect of the non-linearity of the RGD or humidity correction of the OF-CEAS measurements were examined. The agreement between the OF-CEAS spectrometer and the GC measurements is very close to the 2015 World Meteorological Organization (WMO) compatibility goal at 1σ for CO that is ± 2 ppb (see WMO/GMA report edited by Tans and Zellweger (2016)).

4 Conclusions

The OF-CEAS technique allows for the development of sensitive, compact, robust and reliable instruments to perform in-situ trace gas analysis. After a single calibration with a reference standard, an OF-CEAS instrument delivers in real time absolute CO concentrations that are in excellent agreement over one week with state of the art gas chromatograph referenced to the same calibration scale. Similar performance is expected on other trace molecules for which sufficiently strong absorption lines are available. To reach the best accuracy, the GC is periodically calibrated with a standard gas every 30 min and is corrected from the RGD non linearity with data post processing. The agreement between the OF-CEAS spectrometer and the GC for CO concentrations is typically better than 2 ppb that meets the 2015 WMO recommendation for CO inter-laboratory comparison (Tans and Zellweger, 2016). This agreement shows that OF-CEAS instrumental drift on the long term remains acceptable, at the level of accuracy required for atmospheric CO monitoring. Periodic calibrations with a standard gas could become necessary to attain a higher degree of accuracy, since these calibrations could be used to correct the effect of these small drifts.

OF-CEAS instruments offer other advantages that are rarely associated with high sensitivity and selectivity in gas analysis. The sample volume inside the cavity is below 20 cm^3 (standard temperature and pressure conditions) and the pressure can be lowered down to a few mbar, opening field applications in trace detection where small volume samples are available such as bubbles of gas trapped in ice cores for climate studies (Faïn et al., 2014; Rhodes et al., 2016; Grilli et al., 2014).

Additionally, some OF-CEAS analyzers can reach a high sensitivity (below 1 ppb for CO) associated with a short response time (of typically 1 s) that can be exploited in different applications such as in breath analysis to distinguish the respiratory phases (Ventrillard-Courtillot et al., 2009; Maignan et al., 2014) or during tropospheric and stratospheric airborne campaigns to deliver high spatial resolution data for atmospheric models (Romanini et al., 2006; Iannone et al., 2009a). OF-CEAS gas analyzers are now commercialized by AP2E (ProCEAS) that offers presently to measure the concentration of 15 molecular species at high sensitivity with high selectivity.

In order to further enhance the development of the OF-CEAS technique in trace detection and isotopic ratio measurements, the spectral regions that can be exploited have been enlarged to allow new specific molecular absorption signatures. It has been demonstrated that this technique is compatible with different kinds of semiconductor lasers. Indeed, while OF-CEAS was previously developed in the NIR with distributed feed-back telecom diode lasers (Morville et al., 2005; Kassi et al., 2006), it has been demonstrated that it is compatible with extended cavity diode lasers that operates in the visible (Courtillot et al., 2006; Horstjann et al., 2014) and with quantum cascade lasers (QCL) (Maisons et al., 2010; Gorrotxategi-Carbajo et al., 2013) and more recently with interband cascade lasers (ICL) (Manfred et al., 2015; Richard et al., 2016) in the mid infrared region.

15 *Acknowledgements.* The authors are grateful to the whole RAMCES team for its participation to the flights. This work was partly funded by the CarboEurope-IP EU project and supported by the LabexOSUG@2020 program.

References

- http://sirta.ipsl.polytechnique.fr/sirta.old/data_policy.html, page:<http://sirta.ipsl.polytechnique.fr/sirta.old/data{ }policy.html>.
- 20 Chen, H., Karion, a., Rella, C. W., Winderlich, J., Gerbig, C., Filges, a., Newberger, T., Sweeney, C., and Tans, P. P.: Accurate measurements of carbon monoxide in humid air using the cavity ring-down spectroscopy (CRDS) technique, *Atmospheric Measurement Techniques*, 6, 1031–1040, doi:10.5194/amt-6-1031-2013, 2013.
- Chevalier, F., Engelen, R. J., Carouge, C., Conway, T. J., Peylin, P., Pickett-heaps, C., Ramonet, M., and Rayner, P. J.: AIRS-based versus flask-based estimation of carbon surface fluxes, *Journal of Geophysical Research*, 114, D20 303, doi:10.1029/2009JD012311, 2009.
- 25 Courtillot, I., Morville, J., Motto-Ros, V., and Romanini, D.: Sub-ppb NO₂ detection by optical feedback cavity-enhanced absorption spectroscopy with a blue diode laser, *Applied Physics B*, 85, 407–412, doi:10.1007/s00340-006-2354-3, <http://www.springerlink.com/index/10.1007/s00340-006-2354-3>, 2006.
- Derwent, R. G., Ryall, D. B., Jennings, S. G., Spain, T. G., and Simmonds, P. G.: Black carbon aerosol and carbon monoxide in European regionally polluted air masses at Mace Head, Ireland during 1995-1998, *Atmospheric Environment*, 35, 6371–6378, doi:10.1016/S1352-2310(01)00394-6, 2001.
- 30 Faïn, X., Chappellaz, J., Rhodes, R. H., Stowasser, C., Blunier, T., McConnell, J. R., Brook, E. J., Preunkert, S., Legrand, M., Debois, T., and Romanini, D.: High resolution measurements of carbon monoxide along a late Holocene Greenland ice core: evidence for in situ production, *Climate of the Past*, 10, 987–1000, doi:10.5194/cp-10-987-2014, <http://www.clim-past.net/10/987/2014/>, 2014.
- Gorrotxategi-Carbajo, P., Fasci, E., Ventrillard, I., Carras, M., Maisons, G., and Romanini, D.: Optical-feedback cavity-enhanced absorption spectroscopy with a quantum-cascade laser yields the lowest formaldehyde detection limit, *Applied Physics B*, 110, 309–314, doi:10.1007/s00340-013-5340-6, <http://link.springer.com/10.1007/s00340-013-5340-6>, 2013.
- 35 Grilli, R., Marrocco, N., Desbois, T., Guillerme, C., Triest, J., Kerstel, E., and Romanini, D.: Invited Article : SUBGLACIOR : An optical analyzer embedded in an Antarctic ice probe for exploring the past climate, *Review of Scientific Instruments*, 85, 1–7, doi:10.1063/1.4901018, 2014.
- Haszpra, L., Ramonet, M., Schmidt, M., Barcza, Z., Pátkai, Z., Tarczay, K., Yver, C., Tarniewicz, J., and Ciais, P.: Variation of CO₂ mole fraction in the lower free troposphere, in the boundary layer and at the surface, *Atmospheric Chemistry and Physics*, 12, 8865–8875, doi:10.5194/acp-12-8865-2012, <http://www.atmos-chem-phys.net/12/8865/2012/>, 2012.
- 5 Horstjann, M., Hernández, M. D. A., Nenakhov, V., Chrobry, A., and Burrows, J. P.: Peroxy radical detection for airborne atmospheric measurements using absorption spectroscopy of NO₂, *Atmospheric Measurement Techniques*, 7, 1245–1257, doi:10.5194/amt-7-1245-2014, 2014.
- Iannone, R. Q., Kassi, S., Jost, H.-J., Chenevier, M., Romanini, D., Meijer, H. a. J., Dhaniyala, S., Snels, M., and Kerstel, E. R. T.: Development and airborne operation of a compact water isotope ratio infrared spectrometer., *Isotopes in environmental and health studies*, 45, 303–20, doi:10.1080/10256010903172715, <http://www.ncbi.nlm.nih.gov/pubmed/19670069>, 2009a.
- 10 Iannone, R. Q., Romanini, D., Kassi, S., Meijer, H. a. J., and Kerstel, E. R. T.: A Microdrop Generator for the Calibration of a Water Vapor Isotope Ratio Spectrometer, *Journal of Atmospheric and Oceanic Technology*, 26, 1275–1288, doi:10.1175/2008JTECHA1218.1, <http://journals.ametsoc.org/doi/abs/10.1175/2008JTECHA1218.1>, 2009b.
- Kassi, S., Chenevier, M., Gianfrani, L., Salhi, A., Rouillard, Y., Ouvrard, A., and Romanini, D.: Looking into the volcano with a Mid-IR DFB diode laser and Cavity Enhanced Absorption Spectroscopy., *Optics express*, 14, 11 442–52, <http://www.ncbi.nlm.nih.gov/pubmed/19529562>, 2006.
- 15

- Kerstel, E., Iannone, R., Chenevier, M., Kassi, S., Jost, H.-J., and Romanini, D.: A water isotope (^2H , ^{17}O , and ^{18}O) spectrometer based on optical feedback cavity-enhanced absorption for in situ airborne applications, *Applied Physics B*, 85, 397–406, doi:10.1007/s00340-006-2356-1, <http://link.springer.com/10.1007/s00340-006-2356-1>, 2006.
- 20 Landsberg, J., Romanini, D., and Kerstel, E.: Very high finesse optical-feedback cavity-enhanced absorption spectrometer for low concentration water vapor isotope analyses., *Optics Letters*, 39, 1795–8, doi:10.1364/OL39.001795, <http://www.ncbi.nlm.nih.gov/pubmed/24686607>, 2014.
- Langenfelds, R. L., Francey, R. J., Pak, B. C., Steele, L. P., Lloyd, J., Trudinger, C. M., and Allison, C. E.: Interannual growth rate variations of atmospheric CO_2 and its $\delta^{13}\text{C}$, H_2 , CH_4 , and CO between 1992 and 1999 linked to biomass burning, *Global Biogeochemical Cycles*, 25 16, 21–1–21–22, doi:10.1029/2001GB001466, <http://doi.wiley.com/10.1029/2001GB001466>, 2002.
- Li, G., Gordon, I. E., Rothman, L. S., Tan, Y., Hu, S.-M., Kassi, S., Campargue, A., and Medvedev, E. S.: Rovibrational Line Lists for Nine Isotopologues of the CO Molecule in the $X\ 1\ \Sigma^+$ Ground Electronic State, *The Astrophysical Journal Supplement Series*, 216, 15, doi:10.1088/0067-0049/216/1/15, <http://stacks.iop.org/0067-0049/216/i=1/a=15?key=crossref.ed63c46ee3fac474bebf91c04ae7d6>, 2015.
- 30 Logan, J. a., Prather, M. J., Wofsy, S. C., and McElroy, M. B.: Tropospheric chemistry - A global perspective, 86, 7210–7254, 1981.
- Lowe, P.: An Approximating Polynomial for the Computation of Saturation Vapor Pressure, *Journal of Applied Meteorology*, 16, 100–103, doi:10.1175/1520-0450(1978)017<0413:COAPFT>2.0.CO;2, 1976.
- Maignan, M., Briot, R., Romanini, D., Gennai, S., Hazane-Puch, F., Brouta, A., Debaty, G., and Ventrillard, I.: Real-time measurements of endogenous carbon monoxide production in isolated pig lungs., *Journal of biomedical optics*, 19, 047001, doi:10.1117/1.JBO.19.4.047001, 35 <http://www.ncbi.nlm.nih.gov/pubmed/24699633>, 2014.
- Maisons, G., Gorrotxategi-Carbajo, P., Carras, M., and Romanini, D.: Optical-feedback cavity-enhanced absorption spectroscopy with a quantum cascade laser, *Optics letters*, 35, 3607–3609, 2010.
- Manfred, K. M., Ritchie, G. A. D., Lang, N., Röppcke, J., and van Helden, J. H.: Optical feedback cavity-enhanced absorption spectroscopy with a $3.24\ \mu\text{m}$ interband cascade laser, *Applied Physics Letters*, 106, 221106, doi:10.1063/1.4922149, <http://scitation.aip.org/content/aip/journal/apl/106/22/10.1063/1.4922149>, 2015.
- Manfred, K. M., Hunter, K. M., Ciaffoni, L., and Ritchie, G. A. D.: ICL based OF-CEAS: a sensitive tool for analytical chemistry, *Analytical Chemistry*, 89, 902–909, doi:10.1021/acs.analchem.6b04030, <http://pubs.acs.org/doi/abs/10.1021/acs.analchem.6b04030>, 2016.
- 5 Masarie, K. A., Langenfelds, R. L., Allison, C. E., Conway, T. J., Dlugokencky, E. J., Francey, R. J., Steele, L. P., and Vaughn, B.: NOAA / CSIRO Flask Air Intercomparison Experiment : A strategy for directly assessing consistency among atmospheric measurements made by independent laboratories, *Journal of Geophysical Research*, 106, 20,445–20,464, 2001.
- Morville, J., Kassi, S., Chenevier, M., and Romanini, D.: Fast, low-noise, mode-by-mode, cavity-enhanced absorption spectroscopy by diode-laser self-locking, *Applied Physics B*, 80, 1027–1038, doi:10.1007/s00340-005-1828-z, 10 <http://link.springer.com/10.1007/s00340-005-1828-z>, 2005.
- Morville, J., Romanini, D., and Kerstel, E.: Cavity enhanced Absorption Spectroscopy with Optical Feedback, in: *Cavity-Enhanced Spectroscopy and Sensing*, edited by Gagliardi, G. and Looock, H.-P., pp. 163–207, Springer-Verlag, Berlin Heidelberg, doi:10.1007/978-3-642-40003-2, 2014.
- Novelli, P. C., Collins, J. E., Myers, R. C., Sachse, G. W., and Scheep, H. E.: Reevaluation of the NOAA / CMDL carbon monoxide reference scale and comparisons with CO reference gases at NASA-Langley and the Fraunhofer Institut, *Journal of Geophysical Research: Atmospheres*, 99, 12 833–12 839, 1994.
- 15

- Novelli, P. C., Masarie, K. a., and Lang, P. M.: Distributions and recent changes of carbon monoxide in the lower troposphere, *Journal of Geophysical Research*, 103, 19 015, doi:10.1029/98JD01366, 1998.
- Rhodes, R. H., Faïn, X., Brook, E. J., McConnell, J. R., Maselli, O. J., Sigl, M., Edwards, J., Buizert, C., Blunier, T., Chappellaz, J., and
20 Freitag, J.: Local artifacts in ice core methane records caused by layered bubble trapping and in situ production: a multi-site investigation, *Climate of the Past*, 12, 1061–1077, doi:10.5194/cp-12-1061-2016, <http://www.clim-past.net/12/1061/2016/>, 2016.
- Richard, L., Ventrillard, I., Chau, G., Jaulin, K., Kerstel, E., and Romanini, D.: Optical-feedback cavity- enhanced absorption spectroscopy with an interband cascade laser: application to SO₂ trace analysis, *Applied Physics B*, p. 247, doi:10.1007/s00340-016-6502-0, 2016.
- Romanini, D., Chenevier, M., Kassi, S., Schmidt, M., Valant, C., Ramonet, M., Lopez, J., and Jost, H.-J.: Optical-feedback cavity-
25 enhanced absorption: a compact spectrometer for real-time measurement of atmospheric methane, *Applied Physics B*, 83, 659–667, doi:10.1007/s00340-006-2177-2, <http://link.springer.com/10.1007/s00340-006-2177-2>, 2006.
- Rothman, L., Gordon, I., Babikov, Y., Barbe, A., Chris Benner, D., Bernath, P., Birk, M., Bizzocchi, L., Boudon, V., Brown, L., Campargue, A., Chance, K., Cohen, E., Coudert, L., Devi, V., Drouin, B., Fayt, A., Flaud, J.-M., Gamache, R., Harrison, J., Hartmann, J.-M., Hill, C., Hodges, J., Jacquemart, D., Jolly, A., Lamouroux, J., Le Roy, R., Li, G., Long, D., Lyulin, O., Mackie, C.,
30 Massie, S., Mikhailenko, S., Müller, H., Naumenko, O., Nikitin, A., Orphal, J., Perevalov, V., Perrin, A., Polovtseva, E., Richard, C., Smith, M., Starikova, E., Sung, K., Tashkun, S., Tennyson, J., Toon, G., Tyuterev, V., and Wagner, G.: The HITRAN2012 molecular spectroscopic database, *Journal of Quantitative Spectroscopy and Radiative Transfer*, 130, 4–50, doi:10.1016/j.jqsrt.2013.07.002, <http://linkinghub.elsevier.com/retrieve/pii/S0022407313002859>, 2013.
- Sahu, L. K., Sheel, V., Kajino, M., and Nedelec, P.: Variability in tropospheric carbon monoxide over an urban site in Southeast Asia,
35 *Atmospheric Environment*, 68, 243–255, doi:10.1016/j.atmosenv.2012.11.057, <http://dx.doi.org/10.1016/j.atmosenv.2012.11.057>, 2013.
- Schmidt, M., Lopez, M., Yver Kwok, C., Messenger, C., Ramonet, M., Wastine, B., Vuillemin, C., Truong, F., Gal, B., Parmentier, E., Cloué, O., and Ciais, P.: Six years of high-precision quasi-continuous atmospheric greenhouse gas measurements at Trainou Tower (Orléans Forest, France), *Atmospheric Measurement Techniques Discussions*, 7, 569–604, doi:10.5194/amtd-7-569-2014, <http://www.atmos-meas-tech-discuss.net/7/569/2014/>, 2014.
- Tans, P. and Zellweger, C.: 18th WMO/IAEA Meeting on Carbon Dioxide, Other Greenhouse Gases and Related Tracers Measurement Techniques (GGMT-2015), 229, <http://www.wmo.int/pages/prog/arep/gaw/documents/FINAL{ }GAW{ }REPORT{ }229.pdf>, 2016.
- 5 Té, Y., Jeseck, P., Franco, B., Mahieu, E., Jones, N., Paton-Walsh, C., Griffith, D. W. T., Buchholz, R. R., Hadji-Lazaro, J., Hurtmans, D., and Janssen, C.: Seasonal variability of surface and column carbon monoxide over megacity Paris, high altitude Jungfraujoch and Southern Hemispheric Wollongong stations, *Atmospheric Chemistry and Physics Discussions*, pp. 1–26, doi:10.5194/acp-2015-884, <http://www.atmos-chem-phys-discuss.net/acp-2015-884/>, 2016.
- Ventrillard-Courtilot, I., Gonthiez, T., Clerici, C., and Romanini, D.: Multispecies breath analysis faster than a single respiratory cycle by optical-feedback cavity-enhanced absorption spectroscopy., *Journal of biomedical optics*, 14, 064 026, doi:10.1117/1.3269677, <http://www.ncbi.nlm.nih.gov/pubmed/20059264>, 2009.
- Voulgarakis, a., Naik, V., Lamarque, J. F., Shindell, D. T., Young, P. J., Prather, M. J., Wild, O., Field, R. D., Bergmann, D., Cameron-Smith, P., Cionni, I., Collins, W. J., Dalsøren, S. B., Doherty, R. M., Eyring, V., Faluvegi, G., Folberth, G. a., Horowitz, L. W., Josse, B., MacKenzie, I. a., Nagashima, T., Plummer, D. a., Righi, M., Rumbold, S. T., Stevenson, D. S., Strode, S. a., Sudo, K., Szopa, S., and
15 Zeng, G.: Analysis of present day and future OH and methane lifetime in the ACCMIP simulations, *Atmospheric Chemistry and Physics*, 13, 2563–2587, doi:10.5194/acp-13-2563-2013, 2013.

- Warner, J., Carminati, F., Wei, Z., Lahoze, W., and Attié, J. L.: Tropospheric carbon monoxide variability from AIRS under clear and cloudy conditions, *Atmospheric Chemistry and Physics*, 13, 12 469–12 479, doi:10.5194/acp-13-12469-2013, 2013.
- Worden, H. M., Deeter, M. N., Frankenberg, C., George, M., Nichitiu, F., Worden, J., Aben, I., Bowman, K. W., Clerbaux, C., Coheur, P. F.,
20 De Laat, a. T. J., Detweiler, R., Drummond, J. R., Edwards, D. P., Gille, J. C., Hurtmans, D., Luo, M., Martínez-Alonso, S., Massie,
430 S., Pfister, G., and Warner, J. X.: Decadal record of satellite carbon monoxide observations, *Atmospheric Chemistry and Physics*, 13,
837–850, doi:10.5194/acp-13-837-2013, 2013.
- Xueref-Remy, I., Messenger, C., Filippi, D., Pastel, M., Nedelec, P., Ramonet, M., Paris, J. D., and Ciais, P.: Variability and budget of CO₂ in Europe: Analysis of the CAATER airborne campaigns-Part 1: Observed variability, *Atmospheric Chemistry and Physics*, 11, 5655–5672, doi:10.5194/acp-11-5655-2011, 2011.
- 435 Yver, C.: Estimation des sources et puits du dihydrogène troposphérique : développements instrumentaux , mesures atmosphériques et assimilation variationnelle, Ph.D. thesis, Université de Versailles, 2010.
- Yver, C., Schmidt, M., Bousquet, P., Zahorowski, W., and Ramonet, M.: Estimation of the molecular hydrogen soil uptake and traffic emissions at a suburban site near Paris through hydrogen, carbon monoxide, and radon-222 semicontinuous measurements, *Journal of Geophysical Research*, 114, D18 304, doi:10.1029/2009JD012122, <http://doi.wiley.com/10.1029/2009JD012122>, 2009.
- 440 Yver Kwok, C., Laurent, O., Guemri, A., Philippon, C., Wastine, B., Rella, C. W., Vuillemin, C., Truong, F., Delmotte, M., Kazan, V., Darding, M., Lebègue, B., Kaiser, C., Xueref-Rémy, I., and Ramonet, M.: Comprehensive laboratory and field testing of cavity ring-down spectroscopy analyzers measuring H₂O, CO₂, CH₄ and CO, *Atmospheric Measurement Techniques*, 8, 3867–3892, doi:10.5194/amt-8-3867-2015, <http://www.atmos-meas-tech.net/8/3867/2015/>, 2015.
- Zellweger, C., Steinbacher, M., and Buchmann, B.: Evaluation of new laser spectrometer techniques for in-situ carbon monoxide measure-
445 ments, *Atmospheric Measurement Techniques*, 5, 2555–2567, doi:10.5194/amt-5-2555-2012, 2012.




Figure 2. Two days and one night monitoring of atmospheric concentration in Gif sur Yvette selected for their different ranges. CO values are given in dry air mole fraction. The GC data are time-shifted by 14 min in order to eliminate time delay between the two instruments. Upper graphs: OF-CEAS measurements are averaged for 2 s while GC measurements are performed twice an hour. Lower graphs: Difference of the measurements after averaging OF-CEAS data for 1 minute around the GC time measurement. Mean values of the difference and standard deviations are written in green.




Figure 3. Top : CO airborne measurements by OF-CEAS in real time and by GC with latter flask analysis at the LSCE. Altitudes are indicated during collection sample for GC measurements. OF-CEAS data are averaged for 5.5 s. Bottom: Difference between OF-CEAS and GC measurements, where OF-CEAS values are computed in dry air and averaged for 1 min around the flask filling times. Error bars in this graph indicate the standard deviation of OF-CEAS measurements during 1 min around the comparison time. The difference of OF-CEAS and GC measurements has a mean value of -2.2 ppb with a standard deviation of 1.7 ppb.