

Interactive comment on “A new instrument for stable isotope measurements of ^{13}C and ^{18}O in CO_2 – Instrument performance and ecological application of the Delta Ray IRIS analyzer” by Jelka Braden-Behrens et al.

Anonymous Referee #3

Received and published: 13 July 2017

Interesting paper but needs some major revisions. Please find below some listed points that should be changed or at least answered.

1a) Page 2 lines 13ff: text passage about IRMS: Pls cite Schnyder et al. there (citation below) 1b) in the same text passage: I think “sample preparation effort and cost” might be a minus for IRMS techniques. But here the main disadvantages should be mentioned like (storage) problems with vials (see Gemery et al., 1996 and Knohl et al., 2004) and the advantage of quasi-continuous measurements relative to the “dis-continuous” measurement by IRMS. 2) Page 2 lines 22ff. text passage about different

spectrometer types: should be shortened as this manuscript is not a review on optical methods for measurement of isotope ratios 3) Page 3 lines 25ff: “to characterize the Delta Ray IRIS and its performance under field conditions”: I think measurement of the “internal cell turnover” and “Allen deviation” is not sufficient to fulfill this topic here. The reference gas box from the Delta Ray is said to offer possibilities to adjust CO₂ conc of the “reference” gas to the measured [CO₂] to cancel out a possible concentration effects on the measured d-values. The authors need to go more in detail here by showing data (!) from multiple CO₂-in air-standards with different [CO₂] and different d¹³C- and d¹⁸O values measured with IRMS (preferred) in comparison to measurement with Delta Ray or a comparison with different optical measurement devices (more problematic). I suppose you have measured the data, so show them here please. 4) Please give more info (citation if available) on the kind of measurements performed at the MPI in Jena (isotopes and concentration). 5) The link to VPDP was done with the gas tank measured in Jena? Please extend the info on how this is done. Fig. 3 describes your quality control standard? Is there a way to compare measured values (+ stdev.) with a target value (+stdev.)? 6) Page 3 line 26 “b)” please add one or two sentences why R¹³Ceco and R¹⁸Oeco is interesting. 7) Page 11 line 21 “lighter” here means only ¹³C-depleted or also ¹⁸O-depleted ? Please specify (also in whole manuscript) 8) Page 13 line 26: more “enriched” in what? Please check that also in whole manuscript, depleted in ¹³C, enriched in ¹⁸O ... (page 14 line 21 ...) 9) I’m not totally happy to read a manuscript with 2 hypotheses where one hypothesis can be discarded but the 2nd one cannot be proven. The authors should find a way around this, at least the additional measurements for finally testing should be mentioned and discussed here 10) the unit “‰¹ is not conform to the SI unit system, what about using “mUr”? It might be more a editorial decision ...

Gemery et al. (1996): Oxygen isotope exchange between carbon dioxide and water following atmospheric sampling using glass flasks. J Geophys Res 101, D9, 14514-14420. Knohl et al. (2004): Kel-FTM discs improve storage time of canopy air samples in 10-mL vials for CO₂-d¹³C analysis. Rapid Comm Mass Spectrom. 18, 1663-1665.

Schnyder et al. (2004): Mobile, outdoor continuous-flow isotope-ratio mass spectrometer system for automated high-frequency ^{13}C - and ^{18}O - CO_2 analysis for Keeling plot applications. *Rapid Comm Mass Spectrom.* 18, 3068-3074.

Interactive comment on *Atmos. Meas. Tech. Discuss.*, doi:10.5194/amt-2017-120, 2017.

[Printer-friendly version](#)

[Discussion paper](#)

