

List of changes on "Sensitivity study of the instrumental temperature corrections on Brewer total ozone column measurements".

Alberto Berjón et al.

We would like to thank the editor for her time spent on reviewing our manuscript and her comments helping us improving the article.

The English structure and grammar of the manuscript has been thoroughly reviewed. The title of the manuscript has been changed to "Sensitivity study of the instrumental temperature corrections on Brewer total ozone column measurements" .

We include below our point-by-point response to the editor comments and a version of the paper with all changes highlighted.

Associate Editor Decision: Publish subject to technical corrections (16 May 2018) by Irina Petropavlovskikh

Comments to the Author:

Dear Authors,

Thank you for submitting responses to comments from three reviewers. The discussion of results was significantly improved. Based on modifications to the manuscript, I made a decision to accept the paper for publication.

The manuscript text still has many grammatical errors.

For example in lines 8 and 9 of the abstract, the following sentence was edited, but now the meaning of what you are trying to say is lost.

"The results clearly show that the traditional methodology based on the internal halogen lamps is not affected by possible temperature affecting the light source". Is it better to say "The results clearly show that the traditional methodology based on the internal halogen lamps is not sensitive to the temperature-caused changes in the spectrum of the light source"? Is it external or internal light source? Please add to the sentence

The sentence is replaced by:

"The results clearly show that the traditional methodology based on the internal halogen lamps is not sensitive to the temperature-caused changes in the spectrum of the internal light source."

P.1 ,line 17, it is installed "IN" the wide range...

This is replaced.

P. 2, line 3, suggestion to change to "Preventing mistakes in the wavelength selection are of importance...."

The sentence is replaced by:

"Preventing errors in the wavelength selection is of highest importance for the ozone determination."

p. 2 line 6 "(?)" - was it a placeholder for the reference to a paper?

p. 2 line 13, (?) - please remove.

There are several places in the text when the sentence ends with "(?)". Please remove or add reference if it is missing.

Issues with references have been corrected.

p. 6, line 19 Not sure what you mean by "verify Equation (5)" Do you mean "confirm to the condition described by Equation (5)"?

The sentence is replaced by:

“satisfy the conditions defined by Equation”

p. 7 line 9 suggestion to change to "...for the experiments, sensitivity study was performed."

The sentence is replaced by:

“In order to determine the most suitable temperature range for the experiments, a statistical analysis of the Brewer operating temperatures using the EUBREWNET database was performed.”

p. 10, line 19 "relative" -> "relative coefficient "? Not clear what you are saying here. Please clarify.

The sentence is replaced by:

“We can clearly see the improvement when using the analysis of the relative coefficients. In most of the cases we can assume a linear relationship between relative Brewer measurements and the temperatures.”

p. 10 , line 22 - "are" - > "is" , i.e. "the summary is shown"

The sentence is replaced by:

“A summary of the results obtained by both linear regressions methods is shown in Tables 1 and 2.”

p. 11, line 17, "show" -> "found"

This is replaced.

p. 11, lines 22-23. This sentence is unfinished.

The sentence is replaced by:

“This result might be useful when analyzing different operating issues, such as wrong temperature coefficients or incorrect values of ETC, that may introduce diurnal cycles in the final ozone values.”

p. 13, line 7-8, another use of "verify Equation (5)". May be "satisfies the conditions defined by Equation (5)"? Please re-write this and next sentence.

This is replaced.

p. 13, line 9 -should be "is done"

This is replaced.

Please read the manuscript carefully after finalizing the edits and correct text with proper English. You might need to find the native English speaker to proof read your manuscript before the re-submission of the manuscript.

The English structure and grammar of the manuscript has been thoroughly reviewed.

The paper should be resubmitted after the technical corrections are done.

Thank you for the valuable contribution to the special issue.

Regards,

Irina Petropavlovskikh

~~Study on the effect~~ Sensitivity study of the instrumental temperature corrections on Brewer total ozone column measurements

Alberto Berjón^{1,2}, Alberto Redondas^{3,2}, Meelis-Mait Sildoja⁴, Saulius Nevas⁴, Keith Wilson⁵, Sergio F. León-Luis^{3,2}, Omar el Gawhary⁶, and Ilias Fountoulakis⁷

¹University of La Laguna, Department of Industrial Engineering, S.C. de Tenerife, Spain

²Regional Brewer Calibration Center for Europe, Izaña Atmospheric Research Center, Tenerife, Spain

³Agencia Estatal de Meteorología, Izaña Atmospheric Research Center, Spain

⁴Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany

⁵Kipp & Zonen BV, Delft, Netherlands

⁶Dutch Metrology Institute, Delft, Netherlands

⁷Aristotle University of Thessaloniki, Laboratory of Atmospheric Physics, Thessaloniki, Greece

Correspondence to: Alberto Redondas (aredondasm@aemet.es)

Abstract. The instrumental temperature ~~correction~~ corrections to be applied to the ozone measurements by the Brewer spectrophotometers ~~is derived from measurements of the irradiance from an internal halogen lamp in the instrument. These characterisations are derived from the irradiance measurements of internal halogen lamps in the instruments. These characterizations~~ of the Brewer ~~instruments~~ spectrophotometers can be carried out within a thermal chamber, varying the temperature from -5°C to $+45^{\circ}\text{C}$, or during field measurements, making use of the natural change in ambient temperature. However, the internal light source used to determine the thermal sensitivity of the instrument could be affected in both methods by the temperature variations as well, which may affect the determination of the temperature coefficients. In order to validate the standard procedures for determining Brewer's temperature coefficients, two independent experiments using both external light sources and the internal halogen lamps have been performed within the ATMOZ Project. The results clearly show that the traditional methodology based on the internal halogen lamps is not ~~affected by possible temperature affecting the~~ sensitive to the temperature-caused changes in the spectrum of the internal light source. The three methodologies yielded equivalent results, with differences in total ozone column below 0.08% for a mean diurnal temperature variation of 10°C .

1 Introduction

The Brewer spectrophotometer has been used for decades as a reference instrument to retrieve total ozone column (TOC) and for the validation of satellite-based measurements. TOC is retrieved from direct sun measurements at four wavelengths in the ultraviolet (UV) spectral range, from 310.1 nm to 320.1 nm. To be able to make ~~this such~~ measurements, this ~~equipment instrument~~ should be suitable for outdoor use. Moreover, ~~this instrument it~~ should operate at any temperature, since it is installed ~~on in~~ a wide range of environments, from subtropical deserts to polar zones. Changes in the ambient temperature may affect the Brewer instrument in two different ways: ~~changing its sensitivity and causing a spectrum change its spectral responsivity and cause a wavelength~~ shift.

~~Prevent changes in wavelength selection are of special relevance~~ Preventing errors in the wavelength selection is of highest importance for the ozone determination ~~and different strategies are already applied~~. Therefore, different strategies had been already implemented to minimize this problem. ~~Therefore, and~~ the impact of temperature on the ~~spectrum spectral~~ shift is expected to be minimal. The spectrophotometer ~~consists of~~ contains a temperature-compensated monochromator that allows accurate measurements in the UV range (McElroy, 2014). Materials used in the monochromator are selected to minimize the effect of the internal temperature changes on the spectrum position relative to the exit slits. Nevertheless, mechanical tolerances in the manufacturing may cause imperfections in this temperature compensation. Thus, the Brewer operational procedure recommends to perform an internal Hg-lamp test (HG test) when the internal temperature, which is registered for each ozone measurement by a sensor located near the PMT, varies more than 3°C (Environment Canada, 2008). The HG test uses a mercury discharge lamp (line 302.15 nm or 296.73 nm) to check the stability of the wavelength calibration during Brewer instrument operations. During the test, the diffraction grating is positioned such that the operating wavelengths are dispersed onto the appropriate exit slits (Kipp & Zonen, 2008). Finally, ~~the wavelength wavelengths~~ used to determine TOC ~~are the TOC were~~ selected at stationary points in the solar spectrum in order to minimize the effect of any residual spectrum shift (Brewer, 1973).

15 The Brewer spectrophotometer quality assurance protocol includes the thermal characterization of the instrument. This characterization is initially carried out inside a thermal chamber at Kipp & Zonen, manufacturer of the Brewer spectrophotometer, by measuring the output of an internal halogen lamp while varying the temperature of the chamber from -5°C to $+45^{\circ}\text{C}$ during a period of 72 hours. Temperature coefficients are determined during this characterization ~~for at~~ all Brewer operating wavelengths using a linear approximation.

20 If an appreciable temperature dependence in the retrieved ozone is detected during the calibration campaigns, the temperature coefficients are ~~corrected adjusted~~ using the in-field data of the internal halogen lamp measurements along the diurnal temperature variation (Redondas and Rodríguez-Franco, 2015). A drawback of this method is the narrower temperature range that is available for the field measurements compared with the thermal chamber method. However, the natural ambient temperature variation applies to the instrument during its normal operation mode, which generally yields acceptable temperature coefficients.

25 The determination of the thermal sensitivity of the instrument by means of the internal halogen lamp, used in both described procedures, implies that the internal lamp itself and the power supply circuit are also subjected to temperature changes and ~~that~~ they can also exhibit some temperature sensitivity. This can potentially modify the lamp irradiance or its alignment, hampering the determination of the temperature coefficients. In addition, there are different elements involved in the direct sun measurements but not in the measurement of the internal halogen lamp, such as the quartz window and the neutral density filters, which can make the results of the characterization to be ineffective when applied to the operational measurements.

30 While the temperature effect ~~over on~~ the global UV ~~measurements of the Brewer spectrophotometer irradiance measurements by the Brewer spectrometer~~ has been studied by different authors (Weatherhead et al., 2001; Siani et al., 2003; Fountoulakis et al., 2017), so far no validation of the temperature sensitivity of the TOC retrieved from the Brewer measurements has been reported. On this basis, ~~the a~~ validation of the ~~procedures standard procedures used~~ for the retrieval of the temperature

coefficients was included as one of the objectives ~~of the~~ within the European Metrology Research Programme (EMRP) project "Traceability for atmospheric total column ozone" (ATMOZ) ~~project of the European Metrology Research Programme (EMRP).~~

In this work, we report on a comparative study of the temperature coefficients retrieved ~~using by~~ the standard procedures (using ~~the internal halogen lamp~~ internal halogen lamps in a thermal chamber and during field measurements) and ~~by using~~ two
5 alternatives setups employing external lamps for thermal chamber measurements. For this purpose, we have made measurements with #185 and #233 MKIII Brewer spectrophotometers ~~respectively~~ at PTB (Physikalisch-Technische Bundesanstalt) in Braunschweig, Germany, and Kipp&Zonen in Delft, Netherlands. ~~Field,~~ respectively. For comparison purpose, also field data from the EUBREWNET (COST Action ES1207) database is ~~also~~ used (Rimmer et al., 2018).

2 Principle of measurement of the Brewer spectrophotometers

10 In order to understand how the temperature correction is applied to the ~~measurements~~ measurement data by the Brewer spectrophotometers we review in this section the basic measurement principles that are used to obtain the TOC from this instrument. A comprehensive description of the Brewer instrument and the TOC calculation from the results of the measurements can be found in Kerr et al. (1985).

TOC is the main product derived from ~~solar direct~~ direct solar irradiance measurements by the Brewer spectrophotometers.
15 The direct irradiance measurement is performed by pointing the direct entrance port normally to the sun based on an azimuth tracker and a rotating quartz prism which follows the sun's elevation. These measurements are made through a quartz window covering the direct irradiance port. To select the different wavelengths used in the ~~calculation~~ TOC derivation, the Brewer spectrophotometer maintains a fixed position of the diffraction grating and uses a rotating slit mask to select successively each wavelength. The rapid movement of the slit mask assures that all wavelengths are measured almost simultaneously. TOC, in
20 Dobson units (DU) or milli-atm-cm, is obtained from Equation 1 (following manufacturer's nomenclature):

$$TOC = \frac{ETC - R_6 - B}{A\mu} \quad (1)$$

where R_6 is usually defined on the basis of double ratios ~~between intensity measurements~~ of measured intensities, $I_c(\lambda_i)$, at certain wavelengths (Kipp & Zonen, 2008), but it can be also written as the linear combination of the common logarithm of the intensity:

$$25 \quad R_6 = \sum_{i=1}^4 \omega_i F_c(\lambda_i) \quad (2)$$

$$F_c(\lambda_i) = 10^4 \log(I_c(\lambda_i)) \quad (3)$$

The coefficients, ω_i , take values of -1.0, 0.5, 2.2 and -1.7 for the wavelength of 310.1 nm, 313.5 nm, 316.8 nm and 320.1 nm, respectively. These wavelengths correspond to slits 2 to 5 in the rotating slit mask of the monochromator. $I_c(\lambda_i)$ are obtained

from the Brewer raw ~~intensity-signal counts~~ after dark count, dead time, temperature and filter transmittance corrections. The wavelengths, λ_i , used in Equations 2 and 3 have been ~~especially-specially~~ selected to minimize ~~effects of~~ any small shift in ~~the~~ wavelength (Fioletov et al., 2005). Moreover, coefficients ω_i have been determined to suppress any influence of the aerosol and the SO_2 ~~in-on~~ the ozone retrieval (Dobson, 1957; Kerr et al., 1981). In general, any linear effects with wavelength ~~is-are~~ also suppressed, as λ_i and ω_i ~~verify-Equations-satisfy the conditions defined by Equation~~ 4 and 5.

$$\sum_{i=1}^4 \omega_i = 0 \quad (4)$$

$$\sum_{i=1}^4 \omega_i \lambda_i \approx 0 \quad (5)$$

ETC is a linear combination (Equation 6) of the extraterrestrial ~~constant-constants~~ of the instrument, $F_{ext}(\lambda_i)$, which can be obtained from a comparison with a calibrated instrument or from the Langley plot method (Redondas, 2007).

$$10 \quad ETC = \sum_{i=1}^4 \omega_i F_{ext}(\lambda_i) \quad (6)$$

B is a linear combination of the Rayleigh ~~transmittance-transmittances~~ of the air, $\beta(\lambda_i)$, corrected by the Rayleigh air mass, ν , and the ratio ~~between-of~~ the pressure at the observation position, p , and the standard pressure at sea level, p_0 .

$$B = \nu \frac{p}{p_0} \sum_{i=1}^4 \omega_i \beta(\lambda_i) \quad (7)$$

A is a linear combination of the ozone absorption coefficients, $\alpha(\lambda_i)$.

$$15 \quad A = \sum_{i=1}^4 \omega_i \alpha(\lambda_i) \quad (8)$$

Both the Rayleigh air mass, ν , and the ozone air mass, μ , are calculated assuming an effective altitude of 5 km and 22 km, respectively (Bernhard et al., 2005).

R_6 is derived from the intensity ratios ~~,or~~ or equivalently from Equation 2 ~~,andtherefore-and, therefore,~~ it has no units, just like ETC and B .

3 Temperature correction

Most ~~photo-detectors~~ photodetectors have some sensitivity to temperature. If the sensitivity can be linearly approximated, the intensity I [c/s] ~~measured of a stable light source measured by the detector~~ at different temperatures T [$^{\circ}C$], ~~while the detector is illuminated with a stable light source,~~ can be expressed as:

$$5 \quad I = I_c - \tau_0(T - T_0) \quad (9)$$

where T_0 is the reference temperature, I_c is the intensity of the source measured at the reference temperature, and τ_0 is the variation rate of the intensity with the temperature [$c/s^{\circ}C$]. We can rewrite this expression as:

$$I_c = \frac{I}{1 - \tau(T - T_0)} \quad (10)$$

10 where $\tau = \tau_0/I_c$ is the temperature coefficient having units of $1/^{\circ}C$. This last expression has an advantage that, while τ_0 depends on the intensity of the light source, τ is independent and we can use it to determine the intensity of the source at the reference temperature. This process is generally referred to as temperature correction.

The temperature coefficient is usually determined in a laboratory measuring a stable light source while the detector temperature is varied. τ is calculated from the linear regression between measured intensity and the temperature. From the previous
15 equation and applying the natural logarithm, we can write:

$$\ln(I_c) = \ln(I) - \ln(1 - \tau(T - T_0)) \quad (11)$$

For Brewer ~~Spectrophotometers~~ spectrophotometers $\tau \approx 10^{-3} \text{ }^{\circ}C^{-1}$, then $\tau(T - T_0) \ll 1$ and we can approximate the natural logarithm $\ln(1 + x)$ to the first order term of its Taylor expansion:

$$\ln(I_c) = \ln(I) + \tau(T - T_0) \quad (12)$$

20 In the Brewer data processing, this expression is multiplied by 10^4 , the natural logarithm is replaced by common logarithm and T_0 is set to $0^{\circ}C$.

$$10^4 \log(I_c) = 10^4 \log(I) + \tau_b T \quad (13)$$

Where $\tau_b = 10^4 \log(e) \tau$ is the Brewer temperature coefficient. Using Equation 3 we can rewrite this ~~expressed expression~~
as:

$$25 \quad F_c = F + \tau_b T \quad (14)$$

We can define a Brewer relative temperature coefficient, $\tau'_b(\lambda_i)$, by subtracting the reference coefficient from coefficients derived for other spectral channels. Usually the reference coefficient is the one corresponding to the wavelength $\lambda_0 = 303.2nm$. In terms of the Brewer temperature coefficient we can express the relative coefficient as:

$$\tau'_b(\lambda_i) = \tau_b(\lambda_i) - \tau_b(\lambda_0) \quad (15)$$

5 Since the weights used in the common Brewer algorithm to calculate ozone are chosen to ~~verify~~ satisfy the Equation 4, we can ~~write~~ rewrite the Equation 2 as:

$$R_6 = \sum_{i=1}^4 \omega_i F_c(\lambda_i) = \sum_{i=1}^4 \omega_i F(\lambda_i) + T \sum_{i=1}^4 \omega_i \tau_b(\lambda_i) = \sum_{i=1}^4 \omega_i F(\lambda_i) + T \sum_{i=1}^4 \omega_i \tau'_b(\lambda_i) \quad (16)$$

Equation 16 shows that we can use interchangeably $\tau_b(\lambda_i)$ or $\tau'_b(\lambda_i)$ to calculate R_6 and ~~therefor,~~ therefore, TOC.

The relative temperature coefficients, $\tau'_b(\lambda_i)$, can be calculated from Equation 15, but they can also be experimentally
10 retrieved by the linear regression between the ratio of intensities, expressed as $F(\lambda_i) - F(\lambda_0)$, and the temperature:

$$\begin{aligned} F(\lambda_i) &= F_c(\lambda_i) - \tau_b(\lambda_i)T \\ F(\lambda_i) - F(\lambda_0) &= F_c(\lambda_i) - F_c(\lambda_0) - (\tau_b(\lambda_i) - \tau_b(\lambda_0))T \\ F(\lambda_i) - F(\lambda_0) &= F_c(\lambda_i) - F_c(\lambda_0) - \tau'_b(\lambda_i)T \end{aligned} \quad (17)$$

The relative coefficients have the advantage that they can be reliably derived even if the illumination condition is not stable. The only requirement for the determination of the relative coefficients is that the change of the light source is proportional at all wavelengths.

15 Furthermore, from the previous Equations we can see that the temperature effect over R_6 can be reduced to the ~~lineal~~ linear combination of the temperature coefficients, τ_{R6} .

$$\tau_{R6} = \sum_{i=1}^4 \omega_i \tau_b(\lambda_i) \quad (18)$$

Where $\tau_b(\lambda_i)$ can refer indistinctly to the Brewer temperature coefficients or to the relative temperature coefficients. τ_{R6} is expected to be more robust to changes in the light source than the relative coefficients $\tau'_b(\lambda_i)$. Since w_i ~~verify~~ satisfy the
20 Equation 5, τ_{R6} can be correctly determined not only if the change of the light source is proportional at all wavelengths, but also if the change is linear with the wavelength.

4 ~~Operating temperature~~ EUBREWNET data on operating temperatures and thermal sensitivity in EUBREWNET sensitivities

In order to determine the most suitable temperature range for the experiments, ~~we have made a study on the operating temperature~~ a statistical analysis of the Brewer operating temperatures using the EUBREWNET database was performed.

5 The standard operating ambient temperature range provided by the manufacturer of the Brewer spectrophotometer is from 0°C to $+40^{\circ}\text{C}$. ~~This limitation comes from~~ These range limits are related to the operating temperature range of the photomultiplier tube (PMT), which is from 0°C to $+50^{\circ}\text{C}$. As the heat dissipated by the internal electronics increases the temperature in the instrument by about 5°C , the operating temperature range of the Brewer has a safety margin of 5°C .

In the case that the ambient temperature drops below 0°C , the equipment should be operated using an internal heater that
10 allows to extend the operating ambient temperature range down to -20°C ~~to $+40^{\circ}\text{C}$~~ . In practice, the internal heater is activated when instrument temperature drops below 10°C or 20°C depending of the instrument configuration. Additionally, a cold-weather cover is furnished by the manufacturer for extreme weather conditions, which allows operating the instrument at ambient temperatures as low as -50°C .

From the EUBREWNET database (<http://rbcce.aemet.es/eubrewnet>) we have studied the Brewer spectrophotometer internal
15 temperatures at which the ozone measurements have been made in 32 measurement stations worldwide from 1996 to 2017. For this analysis, 4.2 million recorded temperatures were used. The stations involved in the study are mainly in Europe, but there are also data from Greenland, Australia, Uruguay and Algeria. They are, therefore, a very representative sample of the different environmental conditions under which the Brewer spectrophotometers are measuring throughout the world. Figure 1 shows a boxplot for all the stations. The median temperature values for all the stations are between 16°C and 32°C , while the 1st and
20 3rd quartiles are always above 11°C and below 39°C , respectively. The mean diurnal temperature variation is 12°C . Figure 2 shows a histogram with data from all stations together. Considering all the data, the mean temperature value is 23.0°C and a median value is 22°C , with a standard deviation of 16.6°C . The 1 and 99 percentiles are estimated to correspond to 5°C and 44°C . Only a small number of measurements (0.04%) are outside the safety limits for the PMT that we discussed earlier (0°C , 50°C).

25 We have also analyzed the thermal sensitivity, τ_{R6} , of 44 Brewer spectrophotometers at included in the EUBREWNET database, Figure 3. These values range from $-0.9^{\circ}\text{C}^{-1}$ to 4.0°C^{-1} . Two different ~~distribution clearly appears~~ distributions appear to be clearly related to the different Brewer models. MKIII model has a mean τ_{R6} value of $0.20^{\circ}\text{C}^{-1}$ and a standard deviation of $0.52^{\circ}\text{C}^{-1}$. ~~But in~~ In the case of MKII and MKIV, the mean value rises to $1.54^{\circ}\text{C}^{-1}$ and the standard deviation is $0.70^{\circ}\text{C}^{-1}$. While Brewer MKIII has a double monochromator to assure a low stray light influence in UV, MKII and MKIV
30 have a single monochromator, so that they use a $\text{NiSO}_4/\text{UG11}$ filter in front of the PMT to eliminate the effect of visible light on the measurements, i.e., to reduce the stray light in the UV range. The higher temperature dependence observed in the single-monochromator Brewers is commonly attributed to this $\text{NiSO}_4/\text{UG11}$ filter (Fountoulakis et al., 2017; Cappellani and Kochler, 2000).

5 Experimental setups

To study the temperature sensitivity of the Brewer spectrophotometers by using external lamps, two different experimental setups have been used: a first one at PTB in Braunschweig, Germany and a second one at Kipp & Zonen in Delft, Netherlands (both shown in Figure 4). For these studies, MKIII Brewers #185 and #233 were chosen. These instruments are the traveling
5 master instrument of the Regional Brewer Calibration Center for Europe (RBCC-E) triad at the Izaña observatory of the Spanish National Meteorological Agency (AEMET) and a research instrument of Kipp & Zonen, respectively.

During the experiments the internal heater was turned off, but the air circulation fan was left on to evenly distribute the air inside the Brewer instrument, allowing a uniform heating up and cooling down of the internal components.

For the Brewer #185 characterization, a dedicated climate chamber at PTB was used to provide the necessary conditions.
10 The schematic of the measurement system is presented in Figure 5 (left). The temperature and humidity of the chamber was monitored using the built-in sensors of the chamber and two extra sensors, one PT-100 thermometer and one Almemo humidity and temperature sensor. A Hamamatsu model LC8 UV source with a built-in *Xe* lamp and equipped with a quartz fiber bundle as a light guide was used to illuminate simultaneously both ~~global-and-the~~ global and the direct input ports of the Brewer spectrophotometer. The light guide was terminated with light-shaping-diffuser (LSD) to provide a uniform illumination. To
15 monitor the output stability of the UV source, a set of monitor detectors were placed close to the Brewer input ports. Those included two *SiC* photodiodes and a ~~calibrated-spectrometer~~ entrance optics of a calibrated spectroradiometer, which was outside of the chamber. One of the *SiC* photodiodes was located inside the chamber and the other one outside. To direct the UV-radiation onto the external *SiC* diode a similar light guide was used as for the *Xe* lamp system. For optimal irradiation conditions the internal *SiC* photodiode and the entrance optics of the ~~spectrometer~~ spectroradiometer included special quartz-
20 based Primusil diffusers. For the external *SiC* photodiode, no diffuser was used to compare the readings with the diffuser-covered detectors and register any possible change of the diffuser transmittance due to the change of ~~temperature-or-the~~ temperature or the relative humidity in the chamber during the experiment. The Brewer observations consisted of alternately measuring the internal and the external lamps. The external *Xe* lamp was continuously on during the whole cycle of the ~~characterisations~~ characterizations while the internal halogen lamp was turned on and off for each measurement. The drift of
25 the *Xe* source irradiance at the Brewer entrance port was corrected by using the calculated mean of the normalized integrated spectral data from the monitor spectroradiometer and the temperature-corrected *SiC* detector readings.

The experiment was done twice at the PTB facilities. On the first occasion in January 2016, ~~referred-referred~~ in the results as PTB1, the quartz fiber bundle was used to illuminate simultaneously both ~~global-and-the~~ global and the direct input ports of the Brewer instrument. The temperature of the climate chamber was varied between $-5^{\circ}C$ to $+40^{\circ}C$ over a 70 hours period. Sepa-
30 rate cycles were used above and below $0^{\circ}C$ to achieve a better control over ~~temperature-and-the~~ temperature and the humidity. Due to some inconsistencies in the retrieved results of the first experiment, the temperature ~~characterisations~~ characterizations were repeated at the PTB facilities in February 2017. On this second occasion, ~~referred-referred~~ to as PTB2, the light guide of the external lamp was aligned ~~with-to~~ illuminate only the direct entrance port. The global port was not used for the measurements. In addition, the internal halogen lamp was replaced since anomalous behavior was observed during its operation at the

time of the first measurements in January 2016 and later also at the RBCC-E. Two different temperatures cycles were used at different temperature change ~~rates~~. First, the temperature of the climate chamber was varied between 8°C and 45°C over a 64 hours period with a temperature change rate of $1.2^{\circ}\text{C}/\text{h}$. A second cycle was done varying the temperature from -8°C to 30°C over 50 hours with a temperature change rate of $2.8^{\circ}\text{C}/\text{h}$.

5 The experimental ~~set-up~~ setup for characterizing Brewer #233 is shown in Figure 5 (right). The temperature in the chamber was varied from -5 to +45 during a period of 72 hours. A Laser Driven Light Source (LDLS) Energetiq EQ-99, from Dutch Metrology Institute (VSL), was used as an external lamp. By means of an optical fibre bundle, the light was guided into the chamber, collimated by a 25 mm lens and then illuminated the Brewer's quartz window at normal incidence. The collimated beam illuminated the rotating prism, which was aligned in accordance with the incoming light. During the external lamp
10 measurements, no lamp monitor was used, as one of the main characteristics of the LDLS is its high stability (Islam et al., 2013). The other components in the beam delivery part were assumed to be stable and independent of temperature. Two ~~separated~~ separate experiments were performed using the internal halogen lamp in the first case and the external LDLS via the quartz window in the second case. The ~~respectively~~ respective lamps were continuously turned on during each experiments.

In all cases, each measurement cycle included an HG test, repositioning of the micrometer of the diffraction grating to locate
15 the 302.15 nm line of the mercury discharge lamp.

Since different instruments have been used in each experiment (#185 at PTB and #233 at K&Z) the differences in results may be due not only to the differences of the experiment setup, but also due to the different Brewer ~~instrument~~ instruments.

In this work we also include an analysis of the measurements based on the internal halogen lamp during field measurements at dates close to the characterizations in the temperature chambers.

20 6 Results

Results from the experiments carried out at ~~the~~ PTB and at Kipp & Zonen, as well as ~~the analysis of field measurements,~~ analysis of the field measurements are presented in this section.

In Figure 6 we show results of the Brewer measurements of both the external sources and the internal halogen ~~lamp~~ lamps at different temperatures. For the sake of brevity, only the data for 310.1 nm wavelength are shown. The measurements at the
25 other wavelengths show a very similar behavior. The first evident thing apparent in the figures is the difficulty of assuming a linear behavior in these measurements. Despite of using different experimental setups and instruments, the results are not as expected. Only the relation between the measurement results of the internal halogen lamp and the temperature in the K&Z experiment can be considered linear. Nevertheless, some behaviors are repeated in the two experiments at ~~the~~ PTB, which makes us assume that they are probably due to real changes in the behavior of lamps, detectors or the different mechanical
30 elements during the experiments. One of these clearly observed behaviors is the presence of hysteresis cycles. This is possibly related to an inhomogeneous temperature distribution within the instrument. In any case, it is difficult to extract information from the data presented in this way.

The total variation shown in the plots in Figure 6 is only about 1% for a temperature change of more than 50°C over ~~near~~ nearly three days. ~~In~~ Under these conditions it is difficult to ensure sufficient stability of the light sources and the required precision in the alignment. Any uncontrolled effect during the measurements may negatively impact the determination of the absolute temperature coefficients. These results reflect the difficulty of determining the absolute temperature coefficients for
5 the Brewer spectrophotometer.

As stated in section 3, the relative coefficients are intended to be more robust against variations of the experimental conditions. Figures 7, 8 and 9 show results of the relative analysis of the measurements presented versus temperature and the relative temperature coefficients derived from both the external and the internal lamps in the three experiments. The analysis of the internal halogen lamp data from field measurements is also presented.

10 We can clearly see the improvement when using the ~~relative analysis~~. ~~We~~ analysis of the relative coefficients. ~~In most of the cases we can assume a linear relationship between relative Brewer measurements and the temperatures~~. Figures 7, 8 and 9 show two different linear regressions for each data sets. In ~~the first one a first approximation~~ we use the individual ~~measurements and in the second, the linear regression has been~~ relative Brewer measurements. In a second step, a linear regression is performed with respect to the average values at each temperature. This is done to avoid overrepresentation of any of the temperatures
15 since the number of measurements ~~for at~~ certain temperatures is much higher than ~~for others~~. ~~The result summary for at the others~~. A summary of the results obtained by both linear regressions methods ~~are~~ is shown in Tables 1 and 2.

The analysis of the internal halogen lamp measurements in PTB1 shows a very marked nonlinear behavior when using ~~intensities 5 and 6 relative to intensity 2~~. data from slits 4 and 5 relative to measurements with slit 1. Due to this, only data below 30°C have been used to make the linear regression ~~used~~ to obtain the relative temperature ~~coefficient~~ coefficients. This
20 behavior ~~is was~~ not repeated in PTB2. This change in the ~~behaviour could be~~ behavior must have been due to the replacement of the internal halogen lamp between both experiments. One week before the PTB1 experiment, the internal halogen lamp was burned out and it had to be replaced. However, this replaced lamp used in the PTB1 experiment did not show a stable ~~behaviour~~ behavior also during later field measurements. Therefore, it was replaced again in March 2016. This is also the reason why we have only a few number of measurement points to make the field data analysis shown in Figure 7. Moreover, differences
25 between external lamp measurements in PTB1 and PTB2 ~~are~~ may be due to the alignment changes ~~of to couple~~ the external lamp ~~and radiation into~~ the direct entrance port. All these problems makes us to consider Brewer #185 not stable during the PTB1 experiment and, therefore, its results will not be included in the final analysis.

~~On the other hand~~ Fortunately, the PTB2 experiment (Figure 8) ~~shows~~ showed more consistent results. ~~In this case, the relations~~ Here, the relationship between the relative ~~measurements~~ measurement results and the temperature are mainly linear, although some ~~non-linearity is~~ non-linearities are observed for both the internal and the external lamp measurements between 40°C and 50°C . This non-linearity for high temperatures may be explained by the photomultiplier tube ~~behaviour~~ behavior ~~at this~~ behavior at these temperatures. Figure 10 shows the dark signals for Brewers #185 and #233 measured during PTB2 and K&Z experiments. They demonstrate different ~~behaviour~~ behaviors in the operating temperature range. ~~However, for~~ Moreover, at temperatures between 40°C and 50°C both show an increase in the dark signals. Additionally, dark signals of
35 #185 show two different ~~behaviours~~ trends between 40°C and 50°C along the whole experiment time. From the data sets of

the EUBREWNET data base we can see that the most usual behavior corresponds to the one shown by #233. The K&Z experiment gives also a linear relation between the relative data and the temperature for both the internal and the external lamp measurements, Figure 9.

From Tables 1 and 2 we can see that the determined relative temperature coefficients present important differences depending on the used data set (thermal chamber measurements with internal halogen lamp and external lamp or field measurements with internal lamp) in both PTB2 and K&Z experiments. However, when calculating τ_{R6} the differences are strongly reduced and very similar values are obtained, as shown in Figure 11.

Differences between the τ_{R6} values retrieved from the internal and the external ~~lamps-lamp measurements~~ are less than 0.06 for both PTB2 and K&Z experiments, independently if the linear regression is done using individual values or mean values ~~for-at~~ each temperature. Including the field data retrieval, differences are within about 0.08 in both PTB2 and K&Z experiments when the linear regression is done using mean values ~~for-at~~ each temperature. For the regression using individual values, difference rise to 0.14 and 0.07 for the PTB2 and K&Z data sets, respectively. The higher coefficient ~~show-found~~ in the case of the PTB2 field data set when the linear regression is done using individual values, indicates that the greatest number of measurements taken at lower temperatures has a negative impact on the estimated coefficient.

The typical value of the numerator in Equation 1, $ETC - R_6 - B$, is about 1000 for an optical airmass of 1, thus differences in τ_{R6} of 0.08 represent a 0.08% of TOC for a temperature variation of $10^\circ C$. As the mean diurnal variation is close to $10^\circ C$, that value can be considered the diurnal uncertainty due to the temperature correction. This result ~~is important when we try to distinguish between~~ might be useful when analyzing different operating issues ~~that may generate diurnal cycles,~~ such as wrong temperature coefficients or incorrect values of ETC , that may introduce diurnal cycles in the final ozone values.

Note that the uncertainty associated to the different coefficients in Tables 1 and 2 correspond to the standard uncertainty of the slopes from the linear regressions in Figures 7, 8, 9 and 11. τ_{R6} coefficients can also be calculated as a linear combination of the relative coefficients or directly from the linear regressions in Figure 11. ~~But~~ However, to derive the associated uncertainty of τ_{R6} from the uncertainties of the relative coefficients we should assume they are not independent variables. Therefore, the calculation of τ_{R6} uncertainty is more direct from the linear regressions.

It is worth to note that temperature correction is usually applied to measurement data using a reference temperature close to the most ~~frequent~~ frequently faced operation temperature. A reference temperature close to the mean operational temperature means that the applied temperature correction is most of the time small and ~~thus a low accurate,~~ thus, a low accuracy estimation of the temperature sensitivity will not have a high effect ~~over-on~~ the TOC retrieval. However, this is not the case with the Brewer spectrophotometer, which use a reference temperature of $0^\circ C$ while the mean operation temperature is $23^\circ C$ and a median value of $22^\circ C$. Considering ~~negligible the~~ the negligible uncertainty of the temperature ~~measurement~~ measurements, we can write the uncertainty associated with the temperature correction as $\delta\tau_{R6}(T - T_0)$, where $\delta\tau_{R6}$ is the uncertainty associated to the coefficient τ_{R6} . Figure 12 shows ~~a-an~~ estimation of the uncertainties of the temperature ~~correction-for~~ corrections applied to the measurements in EUBREWNET database using the distribution of ~~temperature~~ temperatures previously shown in Figure 2 and the mean value of $\delta\tau_{R6} = 0.08^\circ C^{-1}$ that we have obtained from the differences between the methods. As we can see, the estimated uncertainties are close to zero for most of the cases when $T_0 = 22^\circ C$ and the maximum uncertainty would

only be half the maximum uncertainty in the case of $T_0 = 0^\circ C$. While a change of temperature reference from $T_0 = 0^\circ C$ to $T_0 = 22^\circ C$ may increase the uncertainty associated with some measurements made at low temperatures, in general it would result in a reduction of the uncertainty.

7 Conclusions

5 Two experiments were conducted at ~~the PTB (January, PTB (in January 2016 and February, in February 2017)~~ and the Kipp & Zonen (~~October, in October 2016~~) facilities to validate the standard methods for the determination of the temperature dependence of measurements by the Brewer MKIII ~~measurements instruments~~ used to retrieve atmospheric TOC. We have prioritized the MKIII model in this study since it is the most extended model and generally used as reference, as in the case of the RBCC-E.

10 The first experiment ~~, performed at PTB in January 2016 with a non-stable spectrophotometer, led to unsuccessful an unstable internal lamp in the Brewer spectrophotometer led to unusable~~ results but it ~~shows~~ showed that it is necessary to guarantee a good performance of the Brewer instrument before carrying out the temperature sensitivity analysis. This highlights the importance of the method based on the internal halogen lamp measurement data in the field since it presents the best way to ensure the correct functioning of the spectrophotometer throughout its operation. In order to apply this method it is advisable
15 to schedule SL tests throughout the day to record as wide a range of temperature variation as possible.

The absolute measurement temperature coefficients obtained through the different methods present important inconsistencies that prevent ~~its use to derive absolute temperature coefficients~~ their use. These problems are probably due to the difficulties to control the whole system with the required precision. The total variation of the ~~measurements~~ measurement results is only about 1% with a temperature change of more than $50^\circ C$. Uncontrolled ~~effect~~ effects during the measurements may negatively
20 impact the determination of the absolute temperature coefficients.

The relative ~~measurements seems~~ results seem to be more robust against uncontrolled ~~effects, systematic effects~~ and they present an approximately linear behavior with the temperature. However, the derived relative temperature coefficients show important differences depending on the data used for their calculation.

The calculated τ_{R6} values are very much stable. The combination of the four wavelengths clearly increase the stability in all
25 the experiments. This is probably because the linear combination removes any linear effect with ~~the~~ wavelength, as it ~~verify~~ satisfies the conditions defined by Equation 5. Instead, the relative coefficients does not ~~verify~~ satisfy this property.

Better results are found when the linear regression ~~are~~ is done using the mean value ~~for~~ at each temperature. The TOC differences due to the method used to calculate the temperature coefficients τ_{R6} are below 0.08% for a mean diurnal temperature variation of $10^\circ C$.

30 These experiments confirm that the ~~characterisations~~ characterizations performed in a thermal chamber using either the internal halogen lamp of the Brewer instrument or an external lamp as well as those carried out with the internal halogen lamp during field measurements lead to small differences in the retrieved τ_{R6} . This is ~~so~~ the case even though the values of the relative coefficients are obtained using different types of lamps.

The analysis of the EUBREWNET data shows some temperature sensitivity differences between Brewer MKIII model and the MKII and MKIV models, which may be related ~~with~~to $NiSO_4$ /UG11 filter. Therefore, the conclusions of this work may not be directly applicable to the MKII and MKIV models. Further studies are necessary in order to analyze these specific models.

- 5 Finally, it is advisable to consider a change ~~in the temperature reference~~of the reference temperature from the actual $0^\circ C$ to $22^\circ C$. While this could increase the uncertainty associated with some measurements made at low temperatures, in general it would result in a reduction ~~in~~of the uncertainty associated with the temperature correction.

Competing interests. The authors declare that they have no conflict of interests.

- Acknowledgements.* This work has been supported by the European Metrology Research Programme (EMRP) within the joint research
10 project ENV59 "Traceability for atmospheric total column ozone" (ATMOZ). The EMRP is jointly funded by the EMRP participating countries within EURAMET and the European Union. We thank the European Brewer Network (<http://rbcce.aemet.es/eubrewnet/>) for providing access to the data and the PI investigators and their staff for establishing and maintaining the 32 sites used in this investigation. We further acknowledge the support of the Fundación General de la Universidad de La Laguna.

References

- Bernhard, G., Evans, R. D., Labow, G. J., and Oltmans, S. J.: Bias in Dobson total ozone measurements at high latitudes due to approximations in calculations of ozone absorption coefficients and air mass, *Journal of Geophysical Research: Atmospheres*, 110, n/a–n/a, <https://doi.org/10.1029/2004JD005559>, <http://dx.doi.org/10.1029/2004JD005559>, d10305, 2005.
- 5 Brewer, A. W.: A replacement for the Dobson spectrophotometer?, *Pure and Applied Geophysics*, 106, 919–927, <https://doi.org/10.1007/BF00881042>, 1973.
- Cappellani, F. and Kochler, C.: Temperature effects correction in a Brewer MKIV spectrophotometer for solar UV measurements, *Journal of Geophysical Research: Atmospheres*, 105, 4829–4831, <https://doi.org/10.1029/1999JD900254>, <http://dx.doi.org/10.1029/1999JD900254>, 2000.
- 10 Dobson, G. M. B.: Observers' handbook for the ozone spectrophotometer, in: *Annals of the International Geophysical Year*, V, pp. 46–89, Pergamon Press, 1957.
- Environment Canada: Standard Operating Procedures Manual for the Brewer Spectrophotometer, Tech. rep., Environment and Climate Change Canada, <http://woudc.org/resources/sop.php>, 2008.
- Fioletov, V. E., Kerr, J. B., McElroy, C. T., Wardle, D. I., Savastiouk, V., and Grajnar, T. S.: The Brewer reference triad, *Geophysical Research Letters*, 32, <https://doi.org/10.1029/2005GL024244>, l20805, 2005.
- 15 Fountoulakis, I., Redondas, A., Lakkala, K., Berjon, A., Bais, A. F., Doppler, L., Feister, U., Heikkila, A., Karppinen, T., Karhu, J. M., Koskela, T., Garane, K., Fragkos, K., and Savastiouk, V.: Temperature dependence of the Brewer global UV measurements, *Atmospheric Measurement Techniques*, 10, 4491–4505, <https://doi.org/10.5194/amt-10-4491-2017>, <https://www.atmos-meas-tech.net/10/4491/2017/>, 2017.
- 20 Islam, M., Ciaffoni, L., Hancock, G., and Ritchie, G. A. D.: Demonstration of a novel laser-driven light source for broadband spectroscopy between 170 nm and 2.1 μm , *Analyst*, 138, 4741–4745, <https://doi.org/10.1039/C3AN01020A>, 2013.
- Kerr, J. B., McElroy, C. T., and Olafson, R. A.: Measurements of ozone with the Brewer ozone spectrophotometer, in: *Proceedings of the Quadrennial Ozone Symposium held in Boulder, Colorado, August 1980*, pp. 74–79, J. London, ed., 1981.
- Kerr, J. B., McElroy, C. T., Wardle, D. I., Olafson, R. A., and Evans, W. F. J.: The Automated Brewer Spectrophotometer, in: *Atmospheric Ozone: Proceedings of the Quadrennial Ozone Symposium held in Halkidiki, Greece 3–7 September 1984*, edited by Zerefos, C. S. and Ghazi, A., pp. 396–401, Springer Netherlands, Dordrecht, https://doi.org/10.1007/978-94-009-5313-0_80, 1985.
- 25 Kipp & Zonen: Brewer MKIII Spectrophotometer Operators Manual., Tech. rep., Kipp & Zonen Inc., <http://www.kippzonen.com/Download/207/Brewer-MkIII-Operator-s-Manual>, 2008.
- McElroy, C. T.: Brewer Ozone Spectrophotometer Mechanical Design Principles, Izaña Atmospheric Research Center, Tenerife, Spain, http://kippzonen-brewer.com/wp-content/uploads/2014/10/BrewMechDesign1_McElroy.pdf, 14th WMO-GAW Brewer Users Group Meeting / COST Action ES1207 EUBREWNET Open Congress, 2014.
- 30 Redondas, A.: Ozone absolute Langley calibration, *The Tenth Biennial WMO Consultation on Brewer Ozone and UV Spectrophotometer Operation, Calibration and Data Reporting*, Edited by C. T. McElroy and E. W. Hare, Gaw Report. No. 176, WMO TD No. 1420, 12–14, 2007.
- 35 Redondas, A. and Rodríguez-Franco, J.: Ninth Intercomparison Campaign of the Regional Brewer Calibration Center Europe (RBCC-E), no. 224 in *WMO/GAW Reports*, World Meteorological Organization, http://www.wmo.int/pages/prog/arep/gaw/documents/Final_GAW_224.pdf, 2015.

Rimmer, J. S., Redondas, A., and Karppinen, T.: EuBrewNet – A European Brewer network (COST Action ES1207), an overview, *Atmospheric Chemistry and Physics Discussions*, 2018, 1–14, <https://doi.org/10.5194/acp-2017-1207>, <https://www.atmos-chem-phys-discuss.net/acp-2017-1207/>, 2018.

5 Siani, A. M., Benevento, G., and Casale, G. R.: Temperature dependence of Brewer UV measurements at Rome station, in: *Proceedings Volume 5156, Ultraviolet Ground- and Space-based Measurements, Models, and Effects III*, San Diego, United States, <https://doi.org/10.1117/12.505389>, optical Science and Technology, SPIE's 48th Annual Meeting, 2003.

Weatherhead, E., Theisen, D., Stevermer, A., Enagonio, J., Rabinovitch, B., Disterhoft, P., Lantz, K., Meltzer, R., Sabburg, J., DeLuisi, J., Rives, J., and Shreffler, J.: Temperature dependence of the Brewer ultraviolet data, *Journal of Geophysical Research: Atmospheres*, 106, 34 121–34 129, <https://doi.org/10.1029/2001JD000625>, 2001.




Figure 1. Statistics of each of the 32 stations used for the analysis of the operational instrumental temperatures. Median temperatures are represented by lines within boxes determined by the 1st and 3rd quartiles. Whiskers represent Tukey's limits. The 1st and 3rd quartiles are always above 11°C and below 39°C respectively.




Figure 2. Frequency distribution of instrument temperatures from EUBREWNET database. 4.2 million temperature data points have been used. Only a 0.04% of measurements are outside the recommended limits ($0^{\circ}C$ and $50^{\circ}C$).




Figure 3. Temperature sensitivity of the different Brewer models.




Figure 4. Pictures showing Brewer instruments in the controlled environment chambers for the measurements at the PTB (left) and at the Kipp & Zonen facilities (right).




Figure 5. Measurement setups for used for the Brewer characterizations at PTB (left) and at Kipp & Zonen (right).




Figure 6. Scatter plot of Brewer measurement data (Slit 2) versus temperature in the temperature chambers from PTB1, PTB2 and K&Z experiments when illuminating with the internal (upper plots) and the external lamps (lower plots).




Figure 7. PTB1 experiment. Linear regressions between relative Brewer data (Slit 2, 3, 4 and 5 relative to Slit 1) and temperature coefficients. Coefficients are determined from internal and external lamp measurements in the temperature chamber and from field data. Linear regression using individual measurements (grey points) is represented by dashed lines, while solid lines represent the linear regression when using average values at each temperature.




Figure 8. PTB2 experiment. Linear regression between relative Brewer data (Slit 2, 3, 4 and 5 relative to Slit 1) and temperature coefficients. Coefficients are determined from internal and external lamp measurements in the temperature chamber and from field data. Linear regression using individual measurements (grey points) is represented by dashed lines, while solid lines represent the linear regression when using average values at each temperature.




Figure 9. K&Z experiment. Linear regression between relative Brewer data (Slit 2, 3, 4 and 5 relative to Slit 1) and temperature coefficients. Coefficients are determined from internal and external lamp measurements in the temperature chamber and from field data. Linear regression using individual measurements (grey points) is represented by dashed lines, while solid lines represent the linear regression when using average values at each temperature.





Figure 10. Dark signal values during the PTB2 and the K&Z experiments.




Figure 11. Linear regression regressions between τ_{R6} and temperature for PTB1, PTB2 and K&Z experiments using internal and external lamp in the temperature test chamber and from field data. Linear regression using individual measurements (grey points) is represented by dashed lines, while solid lines represent the linear regression when using average values at each temperature.




Figure 12. Distribution of the temperature correction uncertainties using $\delta\tau_{R6} = 0.08^\circ\text{C}^{-1}$ for the reference temperature $T_0 = 0^\circ\text{C}$ and $T_0 = 22^\circ\text{C}$.

Table 1. Temperature coefficients obtained from the linear regression of the individual measurements, for slit 2, 3, 4 and 5 retrieved from Brewer measurements relative to slit 1 (τ'_b) and temperature coefficient for R_6 (τ_{R6}). Rows are grouped in three blocks representing the tree experiments (PTB1, PTB2 and K&Z). For each experiment, results from external (top row) and internal (middle row) lamp measurements in the temperature chamber, and derived from field data (bottom row) are shown.

Experiment	$\tau'_b(310nm)$ [$^{\circ}C^{-1}$]	$\tau'_b(313nm)$ [$^{\circ}C^{-1}$]	$\tau'_b(316nm)$ [$^{\circ}C^{-1}$]	$\tau'_b(320nm)$ [$^{\circ}C^{-1}$]	τ_{R6} [$^{\circ}C^{-1}$]
PTB1 Ext.Lamp	-0.33 ± 0.02	-0.40 ± 0.03	-0.19 ± 0.04	-0.41 ± 0.05	0.40 ± 0.04
PTB1 Int.Lamp	-0.17 ± 0.03	-0.17 ± 0.03	-0.03 ± 0.03	-0.14 ± 0.03	0.24 ± 0.05
PTB1 Field	0.42 ± 0.11	1.06 ± 0.12	1.97 ± 0.15	2.18 ± 0.18	0.74 ± 0.18
PTB2 Ext.Lamp	-0.67 ± 0.02	-1.21 ± 0.02	-1.64 ± 0.02	-2.16 ± 0.03	0.16 ± 0.03
PTB2 Int.Lamp	-0.20 ± 0.01	-0.21 ± 0.01	-0.13 ± 0.01	-0.22 ± 0.01	0.18 ± 0.01
PTB2 Field	-0.21 ± 0.03	-0.28 ± 0.03	-0.09 ± 0.03	-0.25 ± 0.03	0.30 ± 0.05
K&Z Ext.Lamp	0.13 ± 0.02	0.27 ± 0.03	0.45 ± 0.03	0.67 ± 0.03	-0.15 ± 0.05
K&Z Int.Lamp	-0.05 ± 0.01	-0.02 ± 0.01	0.15 ± 0.01	0.33 ± 0.01	-0.21 ± 0.02
K&Z Field	0.13 ± 0.01	0.37 ± 0.01	0.64 ± 0.01	0.95 ± 0.01	-0.14 ± 0.01

Table 2. Temperature coefficients obtained from the linear regression of the mean values for each temperature, for slit 2, 3, 4 and 5 retrieved from Brewer measurements relative to slit 1 (τ'_b) and temperature coefficient for R_6 (τ_{R6}). Rows are grouped in three blocks representing the tree experiments (PTB1, PTB2 and K&Z). For each experiment, results from external (top row) and internal (middle row) lamp measurements in the temperature chamber, and derived from field data (bottom row) are shown.

Experiment	$\tau'_b(310nm)$ [$^{\circ}C^{-1}$]	$\tau'_b(313nm)$ [$^{\circ}C^{-1}$]	$\tau'_b(316nm)$ [$^{\circ}C^{-1}$]	$\tau'_b(320nm)$ [$^{\circ}C^{-1}$]	τ_{R6} [$^{\circ}C^{-1}$]
PTB1 Ext.Lamp	-0.31 ± 0.03	-0.37 ± 0.05	-0.15 ± 0.07	-0.35 ± 0.09	0.39 ± 0.04
PTB1 Int.Lamp	-0.16 ± 0.03	-0.16 ± 0.02	-0.03 ± 0.04	-0.15 ± 0.04	0.28 ± 0.05
PTB1 Field	0.22 ± 0.24	0.81 ± 0.29	1.59 ± 0.45	1.77 ± 0.49	0.69 ± 0.07
PTB2 Ext.Lamp	-0.61 ± 0.02	-1.11 ± 0.05	-1.51 ± 0.06	-2.00 ± 0.08	0.15 ± 0.03
PTB2 Int.Lamp	-0.20 ± 0.01	-0.21 ± 0.01	-0.12 ± 0.01	-0.22 ± 0.01	0.20 ± 0.01
PTB2 Field	-0.28 ± 0.07	-0.41 ± 0.06	-0.23 ± 0.07	-0.39 ± 0.09	0.22 ± 0.08
K&Z Ext.Lamp	0.14 ± 0.02	0.32 ± 0.03	0.44 ± 0.03	0.71 ± 0.03	-0.21 ± 0.04
K&Z Int.Lamp	-0.04 ± 0.01	-0.00 ± 0.01	0.17 ± 0.01	0.37 ± 0.02	-0.21 ± 0.02
K&Z Field	0.13 ± 0.01	0.37 ± 0.01	0.64 ± 0.01	0.94 ± 0.01	-0.13 ± 0.01