

Wavelength calibration of Brewer spectrophotometer using a tunable pulsed laser and implications to the Brewer ozone retrieval

Alberto Redondas^{1,2}, Saulius Nevas³, Alberto Berjón^{4,2}, Meelis-Mait Sildoja³, Sergio F. León-Luis^{1,2}, Virgilio Carreño^{1,2}, and Daniel Santana^{4,2}

¹Agencia Estatal de Meteorología, Izaña Atmospheric Research Center, Spain

²Regional Brewer Calibration Center for Europe, Izaña Atmospheric Research Center, Tenerife, Spain

³Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany

⁴University of La Laguna, Department of Industrial Engineering, S.C. de Tenerife, Spain

Correspondence to: Alberto Redondas (aredondasm@aemet.es)

We thank the referee for their detailed comments on our manuscript. In the attached document we address their questions and suggestions.

Comments by Reviewer #1

J. Gröbner (Referee) julian.groebner@pmodwrc.ch The paper describes a very detailed and unique characterisation experiment of a Brewer spectrophotometer to determine its spectral characteristics (wavelength scale and spectral resolution) which is necessary to calculate the ozone absorption coefficient required for the total column ozone determination from the solar irradiance measurements. The experiment was performed using a tunable laser source to compare and validate the standard procedure used by the Brewer community. The results show that the two procedures provide consistent results to within 0.1% which is very satisfying and confirms that the current standard procedure is valid. Apart from minor grammatical errors the manuscript is well written and certainly interesting to the scientific community. I have a few comments which I would the authors to answer, pending those I support the publication of the manuscript.

page 6, line 10: It is a cubic polynomial fit, not cubic spline.

corrected

- page 7, lines 13-17: It is true that the hg test of the Brewer is repeated when the discrepancy between the actual and determined position is larger than 1.5 steps. However the hg routine sets the position of the micrometer according to the calculation, and repeating the hg routine serves mainly for confirmation. Therefore the hg routine is accurate to ± 0.5 steps, since this is the resolution of the system, not ± 1.5 steps as written in the manuscript. This considerably improves the estimated wavelength uncertainty.

Yes this reduces considerably the error, 1.5 steps are an estimation of the maximum error in the EUBREWNET processing, in this network a mercury test is considered failed when is repeated, the ozone observations are discarded between the last test and the failed one, how ever most of the test gives a correction of less than step.

In section 3.1, I would suggest to add some information on the wavelength uncertainty of the tunable laser setup, which will affect the Brewer wavelength dispersion. I expect in fact the Brewer wavelength dispersion to have less uncertainties when using spectral discharge lamps with published emission line wavelengths ($\approx 1\text{pm}$), than the wavelength obtained by the tunable laser system ($\approx 10\text{ pm}$).

- 5 The text in Section 3.1 was expanded to explain the wavelength calibrations of the instruments used to monitor the laser wavelengths. As to the uncertainty of the Brewer calibration using emission lines, one has to consider also the effect of varying ambient conditions (pressure, temperature, humidity, CO₂ content) on the refractive index of air, which is also in the range of several picometers (e.g. ambient temperature change from 20 C to 25 C has an effect of 2 pm, atmospheric pressure change from 101kPa to 98 kPa would cause a 3pm change). So that the uncertainties in both cases should be indeed quite comparable.

- In section 3.1.2 the authors compare the ozone absorption coefficient calculated with the parametrized and the actual slit functions and show that the difference is of the order of 0.9% (Table 2). The parametrized slits however are trapezoidal, with a plateau at 0.87 (13% from the top). However as shown in Figure 6, this is not representing the true slits, and therefore the parametrization might be closer to reality when using the full triangle as parametrization. This might show that the method using the parametrization with a full triangle will have less differences to the tunable laser results using the actual measured slit functions. (I have made some tests and the full triangle parametrization resolves about half of the 0.9% discrepancy). I would suggest that the authors add a third column in table 2 showing this information.

10

Thanks for the suggestion, a new column were added and the calculation repeated, we found the discrepancy is more near to 0.85% +/- 0.1 for the high resolution cross sections rather than 0.9% and the use of the triangular parametrization in fact reduce the difference to the half 0.45%. The results slightly differ from cross section to cross section with the exception of t Bass&Paur probably due his low spectral resolution (Table 1).

Table 1. Ozone absorption coefficient in atm cm^{-1} calculated using four absorption cross sections . And the percentage difference to the measured to the paramerized with a trapezoid and triangular

Trapezoid	Triangular	Measured	% Trapezoid	% Triangular
0.3381	0.3395	0.3406	-0.73	-0.33
0.3331	0.3344	0.3359	-0.85	-0.44
0.3483	0.3498	0.3514	-0.86	-0.45
0.3393	0.3407	0.3422	-0.84	-0.43

- In the conclusion, page 13, last sentence, I do not understand the statement saying that both methods agree to 0.1% if the parametrized or measured slits are used. To my understanding, the standard method using a scanning grating is not able to use the measured slits, since the method relies on interpolating the slit functions to the ozone position, which therefore requires a parametrized slit.

- In my opinion the abstract should also mention the positive result that the tunable laser and the scanning grating method give the same ozone absorption coefficients (to within 0.3% or so)?

The comment were added

1 Minor comments

5 :
 - The different wavelength scales (nanometer, angstroem) used in the manuscript and the figures is confusing, and I would recommend to use a single one (nanometer)?

Following your suggestion, we have modified the Figures and, now, its scale is in nanometers. In the text we have tried to express all the results in nanometers. However, some results were expressed as Angstroem or picomentros, because it is a very small value. (Also, the other referee has indicated that some results may be better expressed in picometers)

-page 4, line 8 : i would explicitly state that the method is an ozone calibration (not to be confused with a radiometric irradiance calibration for example).

revised.

page 5, point 3: The FWHM also depends on wavelength, which therefore requires some sort of parametrization of the slit function when using the standard scanning grating method.

revised.

- Figure 2: The units on the left axis seem too small (maximum of 7 counts/second)?

5

The exponential were missing, now is corrected.

- page 6, line 3: I would remove the value in parenthesis (0.0080 nm), or replace the picometer values.

The value has been removed in the text.

-Table 3, I did not find the acronym for SGW. Could it be added in the caption of the Table, for clarity?

10 The SGW makes reference to the article Weber et al. (2016). In this paper, the authors studied the ozone absorption cross section and its uncertainty.

2 initial review

Page 1, Background: Some references on the brewer spectrophotometer should be included.

References to the brewer instrumet are added.

I have never seen such an equation to determine the ozone absorption coefficient within the Beer-Lambert law. I even think that it is a circular argument, because the solar spectrum would be included twice, one time directly in the Beer-Lambert Law, the second time for the convolution. Therefore I think this equation is in fact wrong, and the correct one is the one mentioned later (e.g. equation 8. I would suggest that either the authors can provide a reference for this usage of Equation 7, or delete this part of the discussion, which is not used anyway.

Similar equation were used in a tutorial by David Wardle during the Brewer workshop,

Wardle, D.: Physical Principles II: Optical Characteristics of the Brewer, in The Tenth Biennial WMO Consultation on Brewer Ozone and UV Spectrophotometer Operation, Calibration and Data Reporting, World Meteorological Organization.

- 5 [online] Available from: "ftp://ftp.wmo.int/Documents/PublicWeb/arep/gaw/gaw176_10thbrewer.pdf" 2008. Unfortunately the presentation is not longer available but a similar equation can be derived A similar equation is also presented by Volodya Savastouk during the Brewer Workshop in Beijing with different notation (https://www.wmo.int/pages/prog/arep/gaw/documents/13th_Brewer_d2_BrewerXsec.pdf) and reproduced here (Fig1)

$$\tilde{\alpha}(X, \mu) = \sum w_i \frac{\int \sigma(\lambda) * S(\lambda, \lambda') * E(\lambda, \lambda', X, \mu) d\lambda'}{\int S(\lambda, \lambda') * E_o(\lambda, \lambda', \mu) d\lambda'} \quad (1)$$

- 10 The equation shows the advice of Wardle “The best choice for alpha values is weighting by the solar spectrum absorbed by a typical slant column ozone.” This derivation doesn’t take into account that is the spectra and not its logarithm who is filtered by the Slit function (Aliwell et al., 2002). This last approach is used by Diemoz et al. for the Brewer NO_2 calculations

So the equation is removed and replaced for:

$$\alpha(\lambda) = \frac{\log \left(\frac{\int E_o(\lambda) S(\lambda, \lambda_i) 10^{-\alpha(\lambda) X \mu - \beta(\lambda) \frac{P}{P_o} \nu} d\lambda}{\int E_o(\lambda) S(\lambda, \lambda_i) 10^{-\beta(\lambda) \frac{P}{P_o} \nu} d\lambda} \right)}{X \mu} \quad (2)$$

- 15 This references were added to the article:

Page 6 line 10, Figure number is missing

corrected




Figure 1. Slide from Volodya Savastiouk 2011

Page 7, lines 1-8: I think there is a mistake in the discussion in the precision of the hg test: The HG test allows the wavelength setting to be set to within 0.5 micrometer steps, not 1.5. Then, since one step equals (approximately), 0.0075 nm, 0.5 steps would correspond to 0.00375 nm, or 0.0375Å.

Yes there was a mistake, it should be 0.866 steps not Å. The Hg test allows indeed the wavelength setting to be set within 0.5 micrometer step. However, the correction is only applied if the difference is greater than 2 steps. According to the Brewer SOP : “Corrections to the micrometer position are made, and if the adjustment required is greater than 2 steps (.012nm) then the scan is repeated.” So that we assume that the maximum error is +1.5 steps with a rectangular probability distribution function. Hence, the standard uncertainty in such a case is $1.5/\sqrt{3} = 0.866$ steps, which is approx 0.065Å (0.0075 nm per step)

Figure 5. Why use of log scale? In this manuscript, the interest lies in the slit function shape, and therefore in the linear scale, not in the wings.

The reason to use log scale is to visualize the dark count issue, in linear scale is difficult to see.

page 10, last line. I would add "as already noted in Gröbner et al., 1998", at the end of the sentence finishing with (310–320 nm).

added.

I am confused by the conclusion, and would suggest that the authors rephrase it be less ambiguous: Point 1 of the conclusion states that the tunable laser results give an ozone absorption coefficient 0.8% higher than the ones obtained with the standard approach using the parametrized slits. The last point states however that the use of the cubic spline agrees with the tunable laser results to within 0.1% and therefore confirms that the use of the tunable laser (fixed grating position), with the standard procedure (rotating grating) is equivalent. While I understand that cubic and quadratic dispersion fittings might give rise to differences, I do not understand how this affects the use of parametrized slits and actually measured ones using a tunable laser.

We rephrase the conclusions, to reflect from one side the use of parametrized slit function gives a significant difference of 0.8% if we use the measured slit, but the use of the laser confirm the operational rotating grating method (using the parametrized slits with the laser)

References

- Aliwell, S. R., Van Roozendaal, M., Johnston, P. V., Richter, A., Wagner, T., Arlander, D. W., Burrows, J. P., Fish, D. J., Jones, R. L., Tørnkvist, K. K., Lambert, J.-C., Pfeilsticker, K., and Pundt, I.: Analysis for BrO in zenith-sky spectra: An intercomparison exercise for analysis improvement, *J.-Geophys.-Res.*, 107, ACH 10–1, <https://doi.org/10.1029/2001JD000329>, <http://onlinelibrary.wiley.com/doi/10.1029/2001JD000329/abstract>, 00113, 2002.
- 5 Diemoz, H., Siani, A. M., Redondas, A., Savastouk, V., McElroy, C. T., Navarro-Comas, M., and Hase, F.: Improved retrieval of nitrogen dioxide (NO₂) column densities by means of MKIV Brewer spectrophotometers, *Atmospheric Measurement Techniques*, 7, 4009–4022, 2014.
- Weber, M., Gorshelev, V., and Serdyuchenko, A.: Uncertainty budgets of major ozone absorption cross sections used in UV remote sensing ap-
10 plications, *Atmospheric Measurement Techniques*, 9, 4459–4470, <https://doi.org/10.5194/amt-9-4459-2016>, <https://www.atmos-meas-tech.net/9/4459/2016/>, 2016.

Wavelength calibration of Brewer spectrophotometer using a tunable pulsed laser and implications to the Brewer ozone retrieval

Alberto Redondas^{1,2}, Saulius Nevas³, Alberto Berjón^{4,2}, Meelis-Mait Sildoja³, Sergio F. León-Luis^{1,2}, Virgilio Carreño^{1,2}, and Daniel Santana^{4,2}

¹Agencia Estatal de Meteorología, Izaña Atmospheric Research Center, Spain

²Regional Brewer Calibration Center for Europe, Izaña Atmospheric Research Center, Tenerife, Spain

³Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany

⁴University of La Laguna, Department of Industrial Engineering, S.C. de Tenerife, Spain

Correspondence to: Alberto Redondas (aredondasm@aemet.es)

We thank the referee for their detailed comments on our manuscript. In the attached document we address their questions and suggestions.

Comments by Reviewer #2

1 General comment

- 5 The paper describes an experiment of a spectral characterization of a meteorological instrument: The Brewer (principally used for UV radiation and total ozone column monitoring). This experimentation has been done in a laboratory of a metrological institution (the German PTB in Braunschweig) with the help of one of the most modern and precise material: A tunable laser. The originality of this experiment (part of the EMRP/ATMOZ project), is that scientific teams of the meteorological community worked together with metrology specialists, what is a unique gain for the quality management of the meteorological
- 10 measurements. The spectral characterization that has been done is necessary for the determination of the total ozone column from the UV irradiance measured with the instrument Brewer. The results of the experiment allow a high quality validation of the standard procedures of spectral characterization used in the Brewer community, and allow defining quantitatively the limits of these standard procedures and of their assumptions. The results are well presented in this article, and the discussion is of high scientific quality and will be useful for the Brewer community.
- 15 Apart from minor (but many) typology and grammatical errors, that should be corrected (I have listed a part of them at the end of this report), the manuscript is well written pleasant to read and well understandable. I have a few (5) comments/questions which I would the authors to answer, pending those I support the publication of the manuscript. Technical comments ("x-y = page x, Line y):

5 Comments/questions about the content:

1) In the beginning of the background paragraph, when you mention the slits of the "slit mask" (2-1), it would be welcome to have a brief description of the brewer's slits, and to what they are useful. Maybe you should introduce a table of them and refer to the table in the text.

A table with the slits (Number and its wavelength) and a description has been added in the text.

2) (in 1. Background), lines 2-10 to 2-15, you mention some assumptions about the slit functions (2-14). It would be nice to explain before or in a table, what are the most important assumptions that are done.

The assumptions are described later, we add a reference to the section on the text.

3) In the description of the experiment (Background from 2-19 to 2-28):

3a) I guess the use of the tunable laser is useful for point 2, not point 3. So the mention "using the tunable laser" (2-24) should go in the title of point 2 (2-20). yes 3b) In point 3, you mention that the Brewer scans $\pm 2\text{nm}$ around a fixed wavelength of the laser. Further in Point 3, you mention that the Brewer can scan with 5nm increment between 290 nm and 365 nm. You need to precise how you can scan with a lower step than the increment.

Yes is an error, the experiment were performed with the laser emitting from 290 to 355 every 5 nm whereas the brewer is scanning every 5 steps.

4) In the alpha-formula (4-24) you write the ozone cross-section with $\alpha(\lambda)$ and then in the text, you mention σ for it.

Corrected

5) In the discussion, it would be welcome to explain how the different Brewer users can use these results to optimize the TOC retrieval with their own Brewer. Can these results be generalized to all Brewers? To all Mk-III Brewers? Or should every Brewer go to a laboratory with a tunable laser to characterize its slits? Do you have assumption concerning the age stability of this slit characterization? Is it expected to change with the time? If yes why?

The work validate the method currently in use with the brewer network, the limitations of the quadratic dispersion is an issue on MK-III and MK-IV brewer and the new operating software recently introduce the cubic dispersion. We don't think necessary the characterization with the laser of network brewer.

We introduce this on the discussion.

5 **Grammar and typos**

You use sometimes "tuneable", sometimes "tunable", please choose one terminology, and I guess "tunable" is the correct one.

-> So please correct in title, in 1-5 (Abstract), in 2-9, 2-13 and 2-16 (Chapter 1), in 7-22 and 8-1 (Chapter 3) **Done**

Abstract

1-9: "a underestimation" -> an underestimation **Done**

10 **1. Background**

1-13: "total column ozone (TOC)" -> total ozone column (TOC) 2-9 "the use of the use of" -> the use of 2-9 "allow us" -> allows us **Done**

2-10 "coeficients" -> coefficients **Done**

2-14 "the need for the assumptions" -> the need of assumption **Done**

15 **2. Calibration of the Brewer spectrophotometer** 3-7 "weighting coefficients w" -> "weighting coefficients w_i " **Done**

5-11 "referered" -> referred **Done**

5-12 "double Brewer ," -> double Brewer, (without space before the ",") **Done**

6-1 "one of the six exit slit" -> one of the six exit slits **Done**

6-7 "0.7 A" -> 0.7 Å (Å with "o" on the top for Ångström) **Done**

20 **7-1 "The cubic approximation method. . . use knowledge" -> ". . . uses . . ." **Done****

3. Pulsed laser-based measurements

8-10 and 8-11 "counts/seconds" (3 times) -> counts/second **Done**

9-7 "parametrised" -> parametrized **Done**

9-8 "(Orphal et al. (2016)." -> ", Orphal et al. (2016)." or (Orphal et al. (2016)) **Done**

Wavelength calibration of Brewer spectrophotometer using a [..*]tunable pulsed laser and implications to the Brewer ozone retrieval

Alberto Redondas^{1,2}, Saulius Nevas³, Alberto Berjón^{4,2}, Meelis-Mait Sildoja³, Sergio Fabian León-Luis^{1,2}, Virgilio Carreño^{1,2}, and Daniel Santana^{4,2}

¹Agencia Estatal de Meteorología, Izaña Atmospheric Research Center, Spain

²Regional Brewer Calibration Center for Europe, Izaña Atmospheric Research Center, Tenerife, Spain

³Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany

⁴University of La Laguna, Department of Industrial Engineering, S.C. de Tenerife, Spain

Correspondence to: Alberto Redondas (aredondasm@aemet.es)

Abstract.

In this contribution we present the wavelength calibration of the traveling reference Brewer spectrometer of the Regional Brewer Calibration Center for Europe (RBCC-E) at PTB in Braunschweig, Germany. The wavelength calibration is needed for the calculation of the ozone absorption coefficients used by the Brewer ozone algorithm. In order to validate the standard
5 procedure for determining Brewer's wavelength scale, a calibration has been performed by using a [..²]]tunable laser source at PTB in the framework of the EMRP project ENV59 ATMOZ "Traceability for the total column ozone". Here we compare these results to those of the standard procedure for the wavelength calibration of the Brewer instrument. Such a comparison allows validating the standard methodology used for measuring the ozone absorption coefficient with respect to several assumptions. The results of the laser-based calibrations reproduces those obtained by the standard operational methodology and shows that
10 there is [..³]]an underestimation of 0.8% [..⁴]]of the ozone absorption coefficients due to the use of the parametrized slit functions.

1 Background

[..⁵]

Nowadays the primary ground-based instruments used to report total ozone column (TOC) are Dobson and Brewer
15 spectrophotometers. Based on the irradiances measured by these instruments in the ultraviolet (UV) spectral range and on well-defined retrieval procedures, TOC values are derived. The Brewer spectrometer (Brewer, 1973; Kerr et al., 1981; Kerr, 2010) was introduced in the 1980s as an automatic device measuring direct solar UV radiation and global UV irradiance. Both the Brewer and the Dobson instruments were considered by the World Meteorological Organization (WMO)

*removed: tuneable

²removed: tuneable

³removed: a

⁴removed: due

⁵removed: The

in the framework of the Global Atmosphere Watch program (GAW) as the standard instruments for TOC monitoring. The wavelength calibration is needed for calculation of the ozone absorption coefficient used by the Brewer ozone retrieval algorithm. The Brewer spectrophotometer has two operating modes. In the ozone mode, used for the total ozone column measurements, [..⁶]and in the aerosol mode the diffraction grating stays at a fixed position while the six operational wavelengths are selected by a rotating slit mask (Table 1). The scanning mode is used to perform spectral irradiance measurements in the [..⁷]UV spectral range. In this mode, the slits are fixed and the spectral scan is carried out by turning the diffraction grating. The usual wavelength calibration procedure is performed in the scanning mode by analyzing recorded emission lines of the spectral discharge lamps, which are usually mercury (Hg), cadmium (Cd) , and zinc (Zn). [..⁸]The use of the spectral lines provided in Table 2 allow us to determine the central wavelengths and the corresponding [..⁹]full width at half maximum (FWHM) of the slit functions as well as the relation between the positions of the grating and the corresponding instrument wavelengths (dispersion relation) required to determine the operational wavelengths used for the ozone determination. To obtain the ozone absorption coefficients, the instrumental slit functions are convolved [..¹⁰]with the Bass & Paur ozone absorption cross-section data. The use of [..¹¹]a tunable laser source allows us to:

-
- Calculate the ozone absorption [..¹²]coefficients from the calibration directly in the ozone mode. The normal determination of the ozone absorption coefficients involves scanning of the spectral lines in the scanning mode of the instrument so that the dispersion relation is required to convert the grating positions in micrometer steps to the respective wavelengths. Here we can determine the instrumental slit functions directly in the ozone mode by scanning them with the [..¹³]tunable laser and weight them with the ozone absorption cross-sections without the need [..¹⁴]of assumption about the slit functions and the dispersion relations used in the normal calibration procedure [..¹⁵](See section 2).
- Calculate the dispersion relation based on regularly spaced reference spectral lines provided by the [..¹⁶]tunable laser instead of the irregularly distributed emission lines of the Hg, Cd and Zn spectral lamps.

During the experiment we performed three measurements:

1. The standard method for the dispersion measurements using spectral lamps described in section 2.
2. Direct dispersion measurements (laser scanning).

⁶removed: the

⁷removed: ultraviolet (UV)

⁸removed: Using

⁹removed: FWHM (

¹⁰removed: whit

¹¹removed: the of the tuneable laser source allow

¹²removed: coefficients

¹³removed: tuneable

¹⁴removed: for the assumptions

¹⁵removed: .

¹⁶removed: tuneable

Table 1. Slits and wavelengths used in the Brewer operative algorithms. The table provides the mean and the standard deviation of the central wavelengths and the Full Width Half Maximum (FWHM) in nanometers of the slits of the “average” Brewer spectrophotometer determined during RBCC-E campaigns (Redondas et al., 2014)

slit	Wavelength	std	FWHM	std
0	306.301	0.014	0.548	0.016
2	310.051	0.014	0.539	0.015
3	313.501	0.015	0.555	0.012
4	316.801	0.017	0.545	0.012
5	320.002	0.019	0.538	0.012

[..¹⁷]While the Brewer spectrophotometer is measuring in the ozone mode and in the aerosol mode, the laser wavelength is swept ±2nm with a step of 0.04nm around the six Brewer operational wavelengths selected by the rotating slit mask for different grating positions (Figure 1)

3. Dispersion measurements using the tunable laser (Brewer scanning). While the laser is emitting at a fixed wavelength
- 5 in the range from 290 nm to 365 nm with an increment of 5nm, the Brewer instrument scans ±2nm around [..¹⁸]these wavelengths in the scanning mode by moving the grating and using the 6 slits. [..¹⁹][..²⁰]The results allow us to estimate the dispersion approximation error due to the lack of spectral lines available from the discharge lamps at the end of spectral range of the Brewer spectrophotometer and due to the fact that the emission lines of the used lamps are not equally spaced.

10 **2 Calibration of the Brewer spectrophotometer**

The Brewer instrument measures the irradiance of direct sunlight at six nominal wavelengths (λ) in the UV range (303.2, 306.3, 310.1, 313.5, 316.8, and 320.1) nm[..²¹], each spectral band covering a bandwidth of 0.5 nm ([..²²]resolving power λ/δλ of around 600). The spectral resolution is achieved by a holographic grating in combination with a slit mask that selects the channel to be analyzed by a photomultiplier tube (PMT). The longest four wavelengths are used for the ozone calculation.

- 15 Based on the Lambert-Beer law, the total ozone column in the Brewer algorithm can be expressed as:

$$X = \frac{F - ETC}{\alpha \mu}$$
(1)

¹⁷removed: Direct dispersion measurements (laser scanning).
¹⁸removed: this wavelength
¹⁹removed: Such Brewer scans are carried out at wavelengths ranging from 290
²⁰removed: to 365 nm with an increment of 5nm.
²¹removed:)
²²removed: resolution

Table 2. Emission lines of the discharge lamps used for Brewer calibration

Lamp	Line (nm)	Slits
Mercury (Hg)	289.36	0–1
Hg	296.728	0–3
Zinc (Zn)	301.836	0–5
Zn	303.578	0–5
Cd (multiplet)	308.082	0–5
Cd	313.3167	0–5
Cd	326.1055	0–5
Zn	328.233	0–5
Hg	334.148	0–5
Cd	340.3652	0–5
Cd	349.995	4–5
Cd (multiplet)	361.163	5

where F are the measured double ratios corrected for Rayleigh effects, α is the ozone absorption coefficient, μ is the ozone air mass factor, and ETC is the extra-terrestrial constant. The F , α and ETC parameters are weighted functions at the operational wavelengths with weighting coefficients $[.^{23}]w_i$:

$$F = \sum_i^4 w_i F_i - \frac{p}{p_0} \beta_i \mu \quad (2)$$

$$5 \quad \alpha = \sum_i^4 w_i \alpha_i \quad (3)$$

$$ETC = \sum_i^4 w_i F_{0i} \quad (4)$$

where, β_i are the Rayleigh coefficients, p is the climatological pressure at the measurement site, p_0 is the pressure at sea level, and F_0 are the individual extra-terrestrial constants at each wavelength. The weights $w = [1, -0.5, -2.2, 1.7]$ have been chosen so as to minimize the influence of SO_2 and verify:

$$10 \quad \sum_i^4 w_i = 0 \quad (5)$$

²³removed: w

$$\sum_i^4 w_i \lambda_i = 0 \quad (6)$$

This widely eliminates absorption features which depend, in local approximation, linearly on the wavelength, like for example the contribution from aerosols.

We can divide the calibration in three steps including instrumental calibration, wavelength calibration, and ETC transfer:

- 5 1. The instrumental calibration includes all the parameters that affect the [\[.24\]](#) double ratios (F') obtained from the signals measured with the different slits in front of the PMT, in particular, photomultiplier dead time correction, temperature coefficients and filter attenuation.
2. Wavelength calibration is needed to determine the ozone absorption coefficient. The so-called "dispersion test" is used to obtain the particular wavelengths for the instrument and the slits, or instrumental functions, of each spectrophotometer.
- 10 Note that the precise wavelengths of every Brewer spectrophotometer are slightly different from instrument to instrument.
3. Finally, the ETC transfer is performed by comparison with the reference Brewer instrument or, in the case of the reference instruments, by the Langley method.

The Brewer wavelength calibration follows the operative procedure (Gröbner et al., 1998; Kerr, 2002) used by the Regional Brewer Calibration Center-Europe (RBCC-E) at the calibration campaigns. In summary, the individual wavelengths (bands) in the Brewer instrument are selected through the use of a stainless steel mask of seven slits located at the focal plane of the spectrometer. The particular wavelength is determined by analyzing the measurements of a series of discharge lamps during so-called dispersion test, which determines the central wavelength and FWHM of every slit. Then the wavelength setting is optimized to minimize the effect of wavelength shift during the operation of the instrument by performing the so-called sun-scan test (Lamb, K and Asbridge, A.I. , 1996). Finally, the ozone absorption coefficient is determined for every slit.

20 The ozone absorption coefficient [\[.25\]](#) were defined as (Vanier and Wardle, 1969; Bernhard et al., 2005):

$$\alpha_i([.26]\lambda) = [.27] \frac{[.28] \log \left(\frac{\int E_o(\lambda) S(\lambda, \lambda_i) 10^{[.29] - \sigma(\lambda) X_\mu - \beta(\lambda) \frac{P}{P_o} \nu} d\lambda}{\int E_o(\lambda) S(\lambda, \lambda_i) 10^{-\beta(\lambda) \frac{P}{P_o} \nu} d\lambda} \right)}{X_\mu} \quad (7)$$

Where S is the instrumental slit function for the corresponding wavelength, [\[.30\]](#) E is the sun spectrum at this wavelength that depends mostly on the ozone concentration and airmass, and σ is the ozone cross-section at the temperature of -46.3 °C for Dobson and at -45 °C for Brewer instruments.

25 The Brewer operative method uses the following assumptions:

²⁴ removed: measured signal counts

²⁵ removed: is defined as

³⁰ removed: F'




Figure 1. Results of Brewer measurements in the ozone mode obtained while the laser wavelength [..³⁵] was changed every [..³⁶] 0.04 nm. The central wavelength and the FWHM calculated are displayed in red using the same methodology of the dispersion test.

1. Use “ideal” slits; the slit functions are parametrized as trapezoids, i.e., isosceles triangles truncated at 0.87 height.
2. Stray light is not considered, i.e., zero slit function values are assumed outside the triangle.
3. The FWHM of the triangle is [..³¹] considered different for every slit. It is calculated from the dispersion test [..³²], determined in micrometer steps and then converted to wavelengths using the dispersion relation (Fig. 2).
- 5 4. The ozone cross sections are expressed by the Bass & Paur absorption coefficient data set.
5. Solar spectrum is [..³³] considered constant on the slit.

Under these assumptions, the ozone effective absorption is essentially obtained the same way as in the approximation method of Bernhard et al. (2005) used with Dobson spectrophotometers. (see Eq. 8).

$$\alpha_i = [..^{34}] \frac{\int \sigma(\lambda) S_i(\lambda, \lambda') d\lambda'}{\int S_i(\lambda, \lambda') d\lambda'} \quad (8)$$

³¹removed: dependent of the slit and it is derived

³²removed: .

³³removed: not considered, ($F=1$)

2.1 Dispersion Test

The Brewer spectrophotometer is constructed based either on a single or a double ^[..³⁷]monochromator of modified Ebert–Fastie type, generally ^[..³⁸]referred to as single or double Brewer, respectively. The first monochromator disperses the incoming radiation onto six exit slits. In the case of the ^[..³⁹]double-monochromator Brewer the six exit slits (intermediate slits) of the first monochromator are the entrance slits to a second monochromator that is used in subtractive mode. The wavelength is selected by choosing one of the six exit ^[..⁴⁰]slits (ozone mode) or rotating the grating (scanning mode). The rotation of the grating is managed by a drive mechanism consisting of a motor-driven micrometer linked to an arm that rotates the grating. The smallest wavelength increment corresponding to one stepper motor step varies steadily from approximately 8.0 pm to 7.0 pm ^[..⁴¹]^[..⁴²](Gröbner et al., 1998).

The dispersion relation, which provides the relation between the micrometer steps and the monochromator set wavelengths, is determined by scanning the emission lines as described in section 1. The line scans are carried out with ^[..⁴³]an increment of 10 motor steps (^[..⁴⁴]0.7^[..⁴⁵]Å). From the results, the central position and the FWHM of the slit function are calculated in motor steps assuming an isosceles triangle. The both sides of the peak are fitted to a straight line taking only the function values above 20% and below 80% of the normalized peak. The central point is calculated by the intersection point and the FWHM is the width of the triangle (Figure ^[..⁴⁶]2). Finally, the dispersion relation is calculated using a quadratic polynomial or the ^[..⁴⁷]cubic polynomial approximation. This relation is used to transform the previously determined central positions and FWHMs of the slit functions in micrometer steps to a wavelength scale. ^[..⁴⁸]

The cubic approximation method of Gröbner et al. ^[..⁴⁹]uses knowledge of the optical design of the ^[..⁵⁰]Brewer spectrometer to transfer results of the spectral line measurements from one slit to the other slits^[..⁵¹], which reduces the number of free parameters to ^[..⁵²]be adjusted compared with the quadratic method. However, there is a systematic difference ^[..⁵³]

³⁷removed: monochromator

³⁸removed: referered

³⁹removed: double Brewer ,

⁴⁰removed: slit

⁴¹removed: (0.0080

⁴²removed:) Gröbner et al. (1998)

⁴³removed: with

⁴⁴removed:

⁴⁵removed: A

⁴⁶removed: ??

⁴⁷removed: spline cubic

⁴⁸removed: Measured slit function of the 320 nm (slit #3): the method uses the normalised values only between 0.2 and 0.8 (point with cross). The calculated center step and the FWHM are shown for the up (red) and down scan (blue). The resulting parametrized slit are represented on the right axis. Uncorrected counts/second (blue diamonds) and non-linearly-corrected values (black circles). The central wavelength determination is not affected by the nonlinearity but the FWHM is bigger if the correction is not applied. The hysteresis is evident from the asymmetry of the function at the low values region of the plot.

⁴⁹removed: use

⁵⁰removed: brewer to transfer line measures

⁵¹removed: reducing

⁵²removed: adjust

⁵³removed: on the calculation of the ozone absorption coefficient between




Figure 2. Slit function of the slit #3 (320 nm) measured in the scanning mode: the standard method uses the normalized values only between 0.2 and 0.8 (points with crosses). The figure shows the signals recorded during the scans (orange diamonds) and the non-linearity-corrected values (blue circles) in counts/second. The hysteresis is evident from the asymmetry of the uncorrected signals in the low signal region of the plot. The center steps and the FWHMs calculated from the up-scan are shown in red while the values derived from the down-scanned slit function are presented in blue text. The central wavelength determination is not affected by the nonlinearity but the apparent FWHMs would be larger if the non-linearity correction were not applied. The resulting parametrized slits are represented on the right axis.

between the ozone absorption coefficients calculated by the two methods. Both methods generally agree [..⁵⁴] only in the ozone spectral range [..⁵⁵] of the Brewer instrument (Redondas and Rodriguez-Franco, 2012).

The stability of the wavelength calibration during Brewer operations is checked by measuring the internal Hg lamp. In most of the Brewers, the 302 nm double line (302.150 nm and 302.347 nm) is used due to its proximity to the Brewer operational wavelengths. However, for Brewer #185 and for an increasing number of other Brewers the test is performed using the more powerfull 296.7 nm line. The wavelength test includes 12 measurements of the line from the mercury lamp on slit 0, with an increment of ± 10 steps of the micrometer motor. The obtained curve for the line peak is compared to a stored reference one. The comparison is done by shifting the two [..⁵⁶]scanned curves against each other and calculating the correlation coefficient between the two after each shift. The interpolated step number yielding the maximum of the correlation coefficient provides

⁵⁴removed: if only the brewer

⁵⁵removed: is used

⁵⁶removed: scanned




Figure 3. Pulsed optical parametric oscillator (OPO)-based setup at PTB that was used for measuring the slit functions of the Brewer spectrophotometer.

the reference micrometer position (Savastiouk, 2005). If the required adjustment of the micrometer position is more than one and a half motor steps, the test is repeated. [..⁵⁷] Hence, the accuracy of the wavelength setting of the Brewer instrument achieved by such an approach is [..⁵⁸] defined by a rectangular probability distribution function on the interval of ± 0.5 steps. Thus, the respective standard uncertainty is $0.5/\sqrt{3} = 0.289$ steps. During the brewer operation at RBCC-E calibration campaigns the ozone observations are discarded if the subsequent hg test is repeated, the bigger discrepancy for the accepted measurements will be 1.5 steps. This 1.5 affects the ozone [..⁵⁹] absorption coefficient by approximately of [..⁶⁰] 0.1 atmcm^{-1} and the ozone concentration by 0.3%.

3 Pulsed laser-based measurements

3.1 Instrumental setup

⁵⁷removed: The

⁵⁸removed: limited by one and half motor stepsand cannot be better than about 0.86A. This

⁵⁹removed: abortion

⁶⁰removed: 0.10

[..⁶¹] For the characterisation of the bandpass functions of the Brewer instrument, an upgraded PLACOS setup (Nevas et al., 2009) featuring a [..⁶²] tunable pulsed laser system based on an Optical Parametric Oscillator (OPO) was used at Physikalisch-Technische Bundesanstalt (PTB) in Braunschweig (Figure 3). The new OPO system generates 3-6 ns pulses at 1 kHz repetition rate in the spectral range from 210 nm to 2600 nm. The laser [..⁶³] beam was guided into the direct port of the Brewer spectrophotometer by using a liquid light guide. A fraction of the beam was directed to a monitor photodiode in order to account for the output power changes of the laser beam. The photocurrent of the silicon photodiode was measured by [..⁶⁴] an electrometer. During the measurements, the laser wavelengths were monitored by a wavemeter (laser spectrum analyzer) and a high-resolution spectrometer with a FWHM of 0.1 nm. The wavelength accuracy of the wavemeter according to the manufacturer's specifications is between 6 pm and 3 pm in the spectral range from 290 nm to 360 nm, respectively. The calibration of the instrument is accomplished by means of an autocalibration procedure based on an internally built-in neon lamp. The wavelength scale of the high-resolution spectrometer was calibrated by using the spectral emission lines of a Hg lamp. The wavelength uncertainty of the calibrated spectrometer in the above-mentioned spectral range was estimated to be 10 pm. A side-by-side comparison of the laser wavelength values measured by both instrument types showed an agreement well within 10 pm.

15 In contrast with the standard calibration procedure, where the Brewer instrument scans the lines of the spectral lamps, in this experiment the Brewer measures in ozone mode. Here, the Brewer grating is fixed at the ozone position while the coupled laser beam is measured using the seven slits (slit #1 is used to obtain the dark signal values). During these measurements the wavelength of the OPO system is scanned with 0.04 nm step. The experiment is complemented by the measurements in the Brewer scanning mode, where the [..⁶⁵] tunable laser is used as a source of spectral lines covering the range from 290 nm to 20 360 nm on a regular grid with 5 nm step.

3.1.1 Non-linearity of the PMT

The Brewer detector system, which is based on a PMT, responds non-linearly to pulsed sources. For the measurements of pulsed sources, the PMT manual advises to change the electronics configuration. As the main objective was to validate the operational wavelength calibration of the Brewer, we decided to keep the instrument configuration equivalent to that during the field operations. The non-linearity problem was solved by determining the respective correction function. For this purpose, the power of the laser beam was varied while simultaneously measuring signals of the PMT and the linear monitor photodiode.

The ratios of the measured Brewer counts to the recorded monitor photodiode signals are shown in figure 4. The non-linearity is evident [..⁶⁶] together with hysteresis region near 10^4 Brewer counts/[..⁶⁷] second. The correction is not reliable around 10^3

⁶¹removed: For scanning the

⁶²removed: tuneable

⁶³removed: wavelengths were monitored during the measurements by a wavemeter (laser spectrum analyzer) and a high-resolution spectrometer with an uncertainty of 0.01 nm. The laser

⁶⁴removed: a charge meter.

⁶⁵removed: tuneable

⁶⁶removed: in the figure

⁶⁷removed: seconds




Figure 4. Log-log plot of the normalized ratio of measured Brewer counts to the monitor photodiode signal, which is proportional to the laser power, plotted as a function of the Brewer counts/s. The black points shows the measured points while the red curve is the fit used to correct the Brewer signals for the nonlinearity.

counts/[..⁶⁸]second and lower than 100 counts/[..⁶⁹]second. As we can control the power of the laser beam, it is possible to work on the “flat regions” of the non-linearity characteristics and apply the determined correction. This correction does not affect the calculated central wavelength of the slit functions, though, it does affect the determined FWHM values (Figure 2) if the correction is not applied.

- 5 [..⁷²]We observed that the recorder dark signal values (measurements performed with the blocked slit #1) were highest immediately after exposing the PMT to the laser light. The dark signal of the PMT was then gradually fast decreasing with time after the excitation, which may cause the signal values obtained for slit #1 (measured immediately after slit #0) be higher than for the other slits measured afterwards.

3.1.2 Slit parametrization

- 10 The Brewer algorithms assume trapezoidal slits functions cut at 0.87 of the height (Figure 6) with the center wavelength and the FWHM calculated for every slit from the dispersion relation. The laser measurements allow us to estimate the effect on the

⁶⁸removed: seconds

⁶⁹removed: seconds

⁷²removed: From the Figure 5 we can see that correction is not reliable for signals lower than 100 counts/seconds. Also we




Figure 5. Brewer measurements in the ozone mode while the laser wavelength is changed every 0.04 nm. The orange curve corresponds to the dark counts obtained from the measurements of slit 1.

ozone calculation if we use the directly measured slit functions instead of the parametrized ones. For this purpose we calculate the ozone absorption coefficients for the four ozone cross sections evaluated in the "ACSO" committee ("Absorption Cross Sections of Ozone") (Orphal et al. (2016)).

Among the available data sets there are versions of Bass and Paur (1985) cross-sections denoted as Brewer operational (Brw), IGACO quadratic coefficient (B&P), the cross-sections of Daumont Brion Malicet (DBM) (Daumont et al. (1992), Brion et al. (1993), and Malicet et al. (1995)), and the newly recommended data set for ozone ground-based calculation by Serdyuchenko, Groshelev, Weber (SGW) (Serdyuchenko et al., 2012; Gorshchev et al., 2012; Weber et al., 2016).

⁷³removed: parametrised




Figure 6. Plot of the parametrized (thick lines, left axis) and the measured slit functions (dots, left axis) as well as the different ozone [\[.74\]](#) cross-sections in cm^{-1} (right axis) used for the Brewer effective ozone absorption coefficient calculation.

Table 3. Ozone absorption [\[.77\]](#) coefficients in atm cm^{-1} calculated using four ozone absorption [\[.78\]](#) cross-sections together with the trapezoidal and the triangular parametrization of the slit functions as well as the directly measured slit functions. Also percentage difference is given between the absorption coefficients yielded by the parameterised and the directly measured slit functions.

	[.79] Trapezoid	[.80] Triangular	Measured	% Trapezoid	% Triangular
BRW	0.3381	[.81] 0.3395	0.3406	-0.73	-0.33
B&P	[.82] 0.3331	[.83] 0.3344	0.3359	-0.85	-0.44
DMB	0.3483	0.3498	0.3514	-0.86	-0.45
[.84] SGW	[.85] 0.3393	0.3407	0.3422	-0.84	-0.43

Using the measured slit functions, the calculated effective ozone cross sections are [\[.75\]](#) 0.85% higher compared to those obtained by using the parametrized Brewer slits in the standard procedure (Table 3) [\[.76\]](#), this difference is reduced approximately to the half when a triangular parametrization is used 0.44%, the results are very similar for all the cross section with the exception of the Bass & Paur used by the brewer due his low resolution.

⁷⁵removed: 0.9%

⁷⁶removed: independently of the cross sections used

4 Discussion

The experiment allows us to validate the Brewer standard methodology used to perform the wavelength calibration. For this purpose, we compare laser-based wavelength calibration results to those yielded by the standard operative method based on scanning the spectral lamps in case of both the quadratic and the cubic fit to the dispersion relation.

5 Figure [..⁸⁶]7 shows discrepancies between the central wavelengths calculated by the quadratic and the cubic fits to be bigger than 0.1Å for wavelength above 320 nm and much bigger near 350 nm. This is also indicated by a systematic shape of the residuals of the quadratic fit that are much larger than for the cubic fit ([..⁸⁷]Figure 8). This indicates that the quadratic fit is only valid in the ozone range ([..⁸⁸]310 nm - 320 nm) as already noted in Gröbner et al. (1998). The comparison of the calculated FWHMs (Figure [..⁸⁹]7) shows a different pattern with a difference of 0.1Å between the direct and the scanning
10 methods in the ozone range and with a smaller difference between the quadratic and the cubic fits.

[..⁹⁰][..⁹¹]

[..⁹²]

[..⁹³]The differences between the ozone absorption coefficients calculated from the scanning and the direct measurements of the slit functions are summarized for the six measurements in [..⁹⁴]Table 4, taking as a reference the direct measurements. The quadratic fits result in bigger differences of around 1% whereas in the case of the cubic fits the differences decrease
15 to 0.3% and 0.1% when the laser or the discharge lamps are used, respectively.

5 Conclusions

1. Using the measured slit functions instead of the [..¹¹⁴]parametrized ones increases the ozone absorption coefficients and consequently the calculated ozone values by 0.8%.
- 20 2. The quadratic dispersion relation fit used in the standard Brewer algorithm is not suitable outside the ozone spectral range 310 nm - 320 nm. The residuals show a systematic pattern, which is particularly important [..¹¹⁵]at the upper end of the spectral range.

⁸⁶removed: ??

⁸⁷removed: see figures 9 and 10

⁸⁸removed: 310-320 nm)

⁸⁹removed: ??

⁹⁰removed: Differences between the central wavelengths determined directly and by the scanning methods: with the laser wavelengths in equally spaced grid every 5

⁹¹removed: and lines of the discharge lamps; in both cases quadratic and cubic fitting are used.

⁹²removed: Differences between the FWHMs determined directly and by the scanning methods: with the laser wavelengths in equally spaced grid every 5nm and lines of the discharge lamps; in both cases quadratic and cubic fitting are used.

⁹³removed: The differences to the directly calculated

⁹⁴removed: table

¹¹⁴removed: parametrized

¹¹⁵removed: in the higher




Figure 7. ^[.95] Differences between the central wavelengths (upper panel) and FWHMs (lower panel) determined directly and by the scanning methods: with the laser wavelengths in ^[.96] equally spaced grid every 5 nm and lines of the discharge lamps; in both cases quadratic and cubic fits are used.

^[.97] ^[.98] ^[.99] ^[.100] ^[.101] ^[.102] ^[.103] ^[.104] ^[.105] ^[.106] ^[.107] ^[.108] ^[.109] ^[.110] ^[.111] ^[.112] ^[.113]

3. The comparison of the results of the three experiments shows a maximum difference of 0.3% if the cubic fit is used to approximate the dispersion relation of the Brewer instrument. The respective ^[.116] difference between the ozone absorption coefficient ^[.117] that is obtained from the direct measurements of the tunable laser in the ozone mode and from the operative discharge lamp method is only of 0.1%, if both use the parametrized or measured slit. This confirms the standard procedure used for the RBCC-E calibrations.

4. This work validates the current wavelength calibration method of the Brewer network and shows that in general there is no need for the characterization of the network Brewer instruments with a tuneable source. However, the limitations of the quadratic dispersion approximation used over extended spectral range of Brewers MK-III and MK-IV are evident. Thus, we suggest to update the operating Brewer software for this model instruments with the recent version, which already includes the cubic dispersion.

¹¹⁶ removed: error in

¹¹⁷ removed: obtained using direct measurement in




Figure 8. Residuals of the quadratic (filled circles, solid lines) and cubic (dashed lines) fits. The color indicates the six Brewer slits with the laser wavelengths in equally spaced grid every 5nm

Table 4. Ozone absorption coefficient in atm cm^{-1} calculated using four absorption cross sections

	<i>brw_scan</i> ¹	<i>opo_quad</i> ²	<i>opo_cubic</i> ³	<i>lamp_quad</i> ⁴	<i>lamp_cubic</i> ⁵
SGW	0.3409	0.3442	0.342	0.3446	0.3412
ratio	1	0.9881	1.0033	1.0108	1.001

1. Laser wavelength scanned at fixed Brewer grating position
2. Brewer grating position. changed at fixed laser wavelength, dispersion approx. by a quadratic function
3. Brewer grating pos. changed at fixed laser wavelength, dispersion approx. by a cubic function
4. Brewer grating pos. changed at fixed lamp emission wavelength, dispersion approx. by a quadratic function
5. Brewer grating pos. changed at fixed lamp emission wavelength, dispersion approx. by a cubic function

Acknowledgements. This work has been supported by the European Metrology Research Programme (EMRP) within the joint research project ENV59 “Traceability for atmospheric total column ozone” (ATMOZ). The EMRP is jointly funded by the EMRP participating countries within EURAMET and the European Union.

References

- Bernhard, G., Evans, R. D., Labow, G. J., and Oltmans, S. J.: Bias in Dobson total ozone measurements at high latitudes due to approximations in calculations of ozone absorption coefficients and air mass, *Journal of Geophysical Research: Atmospheres*, 110, n/a–n/a, <https://doi.org/10.1029/2004JD005559>, <http://dx.doi.org/10.1029/2004JD005559>, d10305, 2005.
- 5 Brewer, A. W.: A replacement for the Dobson spectrophotometer?, *Pure and Applied Geophysics*, 106, 919–927, <https://doi.org/10.1007/BF00881042>, 1973.
- Brion, J., Chakir, A., Daumont, D., Malicet, J., and Parisse, C.: High-resolution laboratory absorption cross section of O₃. Temperature effect, *Chem. Phys. Lett.*, 213, 610–612, [https://doi.org/10.1016/0009-2614\(93\)89169-I](https://doi.org/10.1016/0009-2614(93)89169-I), <http://www.sciencedirect.com/science/article/pii/000926149389169I>, 1993.
- 10 Daumont, Brion, J., Charbonnier, J., and Malicet, J.: Ozone UV spectroscopy I: Absorption cross-sections at room temperature, *Journal of Atmospheric Chemistry*, 15, 145–155, <https://doi.org/10.1007/BF00053756>, 1992.
- Gorshchev, V., Serdyuchenko, A., Weber, M., and Burrows, J. P.: High spectral resolution ozone absorption cross-sections: Part I. Measurements, data analysis and comparison around 293K., *J. Quant. Spectrosc. Radiat. Transfer*, http://www.iup.uni-bremen.de/UVSAT_material/manuscripts/serdyuchenko_o3x_part1_v1.1.pdf, 2012.
- 15 Gröbner, J., Wardle, D. I., McElroy, C. T., and Kerr, J. B.: Investigation of the wavelength accuracy of Brewer spectrophotometers, *Appl. Opt.*, 37, 8352–8360, <https://doi.org/10.1364/AO.37.008352>, 1998.
- Kerr, J.: New methodology for deriving total ozone and other atmospheric variables from Brewer spectrophotometer direct sun spectra, *Journal of Geophysical Research: Atmospheres*, 107, 2002.
- Kerr, J. B.: The Brewer Spectrophotometer, in: *UV Radiation in Global Climate Change*, edited by Gao, W., Slusser, J. R., and Schmoldt, D. L., pp. 160–191, Springer Berlin Heidelberg, http://link.springer.com/chapter/10.1007/978-3-642-03313-1_6, 2010.
- 20 Kerr, J. B., McElroy, C. T., and Olafson, R. A.: Measurements of ozone with the Brewer ozone spectrophotometer, in: *Proceedings of the Quadrennial Ozone Symposium held in Boulder, Colorado, August 1980*, pp. 74–79, J. London, ed., 1981.
- Lamb, K and Asbridge, A.I. : Brewer Technical Note No.96.01 Sun scans and relationship to the wavelength calibration step number, Tech. rep., International Ozone Services, 1996.
- 25 Malicet, Daumont, D., Charbonnier, J., Parisse, C., Chakir, A., and Brion, J.: Ozone UV spectroscopy. II. Absorption cross-sections and temperature dependence, *J. Atmos. Chem*, 21, 263–273, <https://doi.org/10.1007/BF00696758>, 1995.
- Nevas, S., Lindemann, M., Sperling, A., Teuber, A., and Maass, R.: Colorimetry of LEDs with array spectroradiometers, *Mapan*, 24, 153–162, 2009.
- Orphal, J., Staehelin, J., Tamminen, J., Braathen, G., De Backer, M.-R., Bais, A., Balis, D., Barbe, A., Bhartia, P. K., Birk, M., et al.: Absorption cross-sections of ozone in the ultraviolet and visible spectral regions: Status report 2015, *Journal of Molecular Spectroscopy*, 327, 105–121, 2016.
- 30 Redondas, A. and Rodriguez-Franco, J. J.: Regional Brewer Calibration Center Europe CEOS campaigns: Instrumental Findings, in: *Quadrennial International Ozone Symposium*, <http://hdl.handle.net/20.500.11765/7850>, ID5967, 2012.
- Redondas, A., Evans, R., Stuebi, R., Köhler, U., and Weber, M.: Evaluation of the use of five laboratory-determined ozone absorption cross sections in Brewer and Dobson retrieval algorithms, *Atmospheric Chemistry and Physics*, 14, 1635–1648, 2014.
- 35 Savastiouk, V.: Improvements to the direct-sun ozone observations taken with the Brewer spectrophotometer, 2005.

- Serdyuchenko, A., Gorshelev, V., Weber, M., Chehade, W., and Burrows, J. P.: High spectral resolution ozone absorption cross-sections: Part II. Temperature dependence, *J. Quant. Spectrosc. Radiat. Transfer*, 2012.
- Vanier, J. and Wardle, D.: The effects of spectral resolution on total ozone measurements, *Quarterly Journal of the Royal Meteorological Society*, 95, 399, 395, <http://adsabs.harvard.edu/abs/1969QJRMS..95..395V>, 1969.
- 5 Weber, M., Gorshelev, V., and Serdyuchenko, A.: Uncertainty budgets of major ozone absorption cross sections used in UV remote sensing applications, *Atmospheric Measurement Techniques*, 9, 4459–4470, <https://doi.org/10.5194/amt-9-4459-2016>, <https://www.atmos-meas-tech.net/9/4459/2016/>, 2016.