

Interactive comment on “Effects of Temperature, Pressure, and Carrier Gases on the Performance of an Aerosol Particle Mass Analyser” by Ta-Chih Hsiao et al.

Ta-Chih Hsiao et al.

tchhsiao@gmail.com

Received and published: 21 June 2018

The authors would like to thank the positive feedbacks to our work and appreciate the reviewers' valuable comments for helping to significantly improve the manuscript. We agreed with most of the comments and have revised the manuscript accordingly. Following are our point-by-point response to each of the comments made by the reviewers.

Reviewer 1 Comments The authors reported evaluations of temperature, pressure, and carrier gas on the performance of a commercially available aerosol particle mass analyser (APM). The effects of the first two parameters (temperature and pressure) were

mainly evaluated through theoretical calculation of the transfer function of APM, while effects of carrier gases (air, O₂ and CO₂) were experimentally evaluated with DMA pre-selected 50-nm and 100-nm PSL spheres. Results suggested that the mass detection limit of particles can be as low as 10⁻² fg, and can be further extended to low values with other carrier gases such as hydrogen or with a lower operation pressure down to 80 kPa. The treatment of the theoretical calculation of the transfer function is rigorous, and experiments were well designed and performed. The writing of the manuscript is clear, and the study is within the scope of AMT. However, I have concerns, some of which appeared in the initial review, that the authors are over-interpreting their results and ignoring a few other studies of the same sort. These concerns are detailed below in Major Comments, while a few editorial suggestions are listed as Minor Comments. I suggest publication of this manuscript in AMT after the authors address them in the revised manuscript.

We appreciate the positive feedbacks to our work and would revise the manuscript according to reviewer's suggestions. The abstract was rewrite to explicitly express that the effects of pressure and temperature were theoretically analysed, while the effect of carrier gas was evaluated experimentally. Using gases other than air, such as CO₂ and O₂, as carrier gas for APM are just trying to test whether or not the APM performance would change significantly under various conditions, which may be the case for ambient monitoring or characterizing atmospheric aerosols. For example, argon would be required as the carrier gas if the APM was used as an aerosol particle classifier coupled with inductively coupled plasma mass spectrometry (ICP-MS; in a similar manner to the DMA-ICP-MS system). Tandem APM-ICP-MS could be employed for realtime characterization of trace elements in atmospheric aerosols. In addition, the theoretical calculation and numerical simulation were conduct for explaining the experimental results.

Major Comments: 1. While first stating the importance of the DMA-APM system in measuring atmospheric particle mass and effective density, the authors might be a bit

Printer-friendly version

Discussion paper



over-emphasizing the "low" mass detection limits using hydrogen as the carrier gas and operating under 80-kPa condition. We do not have either of these often in the atmosphere (at least in the lower portion). It is good to characterize the DMA-APM system under those conditions, but it is a bit misleading (giving an impression that 10×10^{-4} fg is easily achieved in ambient measurements) to state those numbers in the abstract. And some of them were from theoretical calculation instead of direct measurement (see below in Major Comment #2).

Thanks for your comments. We rewrote the abstract and try to avoid over-emphasizing the detection limits. P1.L15-23: "...In this study, the effects of temperature and pressure were analysed through theoretical calculation, and the influence of varying carrier gas was experimentally evaluated. The transfer function and APM operational region were further calculated and discussed to examine their applicability. Based on the theoretical analysis of the APM's operational region, the mass detection limits are changed with the properties of carrier gases under a chosen λ value. Moreover, the detection limit can be lowered when the pressure is reduced, which implies that performance may be affected during field study. In experimental evaluation, air, oxygen, and carbon dioxide were selected to atomize aerosols in the laboratory with the aim of evaluating the effect of gas viscosity on the APM's performance. Using monodisperse polystyrene latex (PSL) spheres with nominal diameters of 50 and 100 nm, the classification performance of the APM were slightly varied with carrier gases, while the classification accuracy were consistently within 10%."

2. While the authors stated throughout the manuscript (and the title) that temperature, pressure, and carrier gases were evaluated for the APM performance, most of them (at least for temperature and pressure) were from theoretical calculation. I suggest the authors stating this caveat explicitly in their abstract/conclusion, in order not to mislead readers that temperature and pressure were also evaluated experimentally.

Thanks for your comment. As you see in the revised manuscript, to avoid the potential misleading, the theoretical calculation and experimental evaluation were clearly defined



in abstract and main text. P1.L15-16: "...In this study, the effects of temperature and pressure were analysed through theoretical calculation, and the influence of varying carrier gas was experimentally evaluated..."

3. Kuwata et al. (ES&T, 2012; AS&T, 2015) used DMA-APM to study particle (specifically SOA) density (ES&T, 2012) and developed equations to characterize the DMA-APM system (AS&T, 2015). It is highly suggested that this manuscript is put in the context of those previous studies with comparison and discussion.

As suggested by reviewer, the related references were included in the context. P2.L22-24: "...Throughout the past decades, this scheme has also been adopted extensively to determine the Df of aerosol aggregates (Lall et al., 2008; McMurry et al., 2002; Park et al., 2003; Park et al., 2004a; Park et al., 2004b; Scheckman et al., 2009) and atmospheric aerosols (Kuwata and Kondo, 2009; Kuwata et al., 2011)."

P6.L5-9: "The operation of DMA-APM is identical to Kuwata and Kondo (2009) and Kuwata et al. (2011), in which the DMA selects particles with +1 charge and predetermined mobility diameters and then subjects them to the APM. Following, the APM was set to scan across a range of voltage (V) while the number concentration (CN) of the passing particles was measured by a CPC. The peak of the CN-V distribution was subsequently inspected to determine the particle mass (m)."

P7.L5-7: "...It should be noted that, according to the work done by Kuwata (2015), even when the resolution of the DMA-APM system appears to be controlled by the APM, the particle classification by the DMA-APM at a certain operating condition still could be regulated by both DMA and APM."

P7.L26-32: "According to Kuwata's theoretical analysis of transfer function and resolution of the DMA-APM system, the common operation of constant \bar{m} and varying V could not maintain the transfer function because of the range of d_p, m passing the DMA (Kuwata, 2015). In such case, the transfer function may not be symmetric, and the transfer function is narrower for larger m because the dependence of \bar{m} on m . It was

[Printer-friendly version](#)

[Discussion paper](#)



then concluded that the operation of constant V and varying \dot{V} , on the other hand, could better maintain the DMA-APM resolution because \dot{V}^2 can be constant under constant V. However, this ideal operation protocol is less employed for the DMA-APM system, mainly due to the practical impediment of quickly and accurately scanning \dot{V} over a range. Therefore, the common constant \dot{V} operation is investigated here.”

4. The authors mentioned that viscosity and density of the carrier gases might affect the classification capability of APM (page 7, line 6). While viscosity was included in their theoretical treatment, is it possible to include different densities of those carrier gases tested? These three gases (air, O₂, and CO₂) have quite different molecular weights too. Is that going to affect the classification capability of APM as well?

The effect of densities of carrier gas may affect the classification performance of APM through influencing the flow field. In this study, our analytical treatment does not explore the velocity term. However, the effect of gas density was included in the numerical simulation of flow field using air, O₂, and CO₂ as carrier gases. The setting values are listed as below:

Density (kg/m³) Viscosity (Pa S) air 1.205 0.000018203 CO₂ 1.842 0.000014673 O₂ 1.331 0.000020229

Base on ideal gas law, the molecular weight is proportional to the density of gas under constant pressure and temperature. Therefore, we can regard the molecular weights effect as the density effect.

5. As the authors stated quite frequently the usefulness of using the DMA-APM system to measure effective density, what is the measured density of PSL spheres when compared to reported values?

The measured density for 50-nm and 100-nm PSL are 1142-1148 Kg/m³ and 1142-1156 Kg/m³, respectively. The PSL density reported by manufacturer is 1050 Kg/m³.

Minor Comments:



1. Page 2, line 15: "(Ehara et al., 1996)" to "Ehara et al. (1996)".

Corrected accordingly.

2. Page 2, line 22: "the past decade" to "the past decades". You cited papers in as early as 2002.

Revised as suggested.

3. Page 4, line 6: "defined by(Ehara et al., 1996)" should be "defined by Ehara et al. (1996).

Corrected accordingly.

4. Page 4, line 25: "low molecular weight" to "low-molecular-weight".

Revised as suggested.

5. Page 6, line 2: "APM C_N-V spectra". Is "C_N" defined?

The related description is added in P6.L7-8: "...the APM was set to scan across a range of voltage (V) while the number concentration (CN) of the passing particles was measured by a CPC. The peak of the CN-V distribution was subsequently inspected to determine the particle mass (m)."

6. Page 6, line 27: "peak voltage is indicate" to "peak voltage is indicated".

Revised as suggested.

7. Page 7, line 31: "by an approximate order of" to "to approximately"?

Revised as suggested.

8. Page 13, Figure 5: better to separate this figure into two panels (say for 50-nm and 100-nm PSL), and join the symbol with lines. It is very difficult to tell which one is which in the current form. Also I am not sure if this Penetration-Voltage plot can be called "experimental transfer function".

[Printer-friendly version](#)

[Discussion paper](#)



Interactive
comment

Figure revised as suggested. The Penetration-Voltage plot is revised as APM spectrum, which was used in Tajima et al. (2013).

9. Page 14, Figure 6: although symbols and lines were defined in the text, it would be better to put the legends here in the figure for readers to follow easily.

Revised as suggested.

10. Page 15, Figure 7: this figure has different appearance in tick labels and legends (too small to see) compared to other figures. Suggest to change those labels to a larger font size.

Revised as suggested.

Please also note the supplement to this comment:

<https://www.atmos-meas-tech-discuss.net/amt-2017-480/amt-2017-480-AC1-supplement.pdf>

Interactive comment on *Atmos. Meas. Tech. Discuss.*, doi:10.5194/amt-2017-480, 2018.

[Printer-friendly version](#)

[Discussion paper](#)

