

Reply to the editor review of the manuscript “On the parametrization of optical particle counter response including instrument-induced broadening of size spectra and a self-consistent evaluation of calibration measurements” by A. Walser et al.

We wish to thank the editor for carefully reading our manuscript and for providing valuable feedback which helped us to improve the manuscript. In the following, the questions and comments raised by the editor are marked in blue. Our answers are written in black and include a description of changes done to the manuscript in separate paragraphs.

Reviewer #1 raised questions about assessing the value of this new approach for practical determinations of particle size distribution in ambient conditions with unknown mixtures of aerosol species with differing index of refraction as well as shape. Clearly, without additional information about the actual particle populations, the new approach cannot realistically reduce uncertainty about the intrinsic uncertainties in the ambient particles. Your response made this clear, yet I am concerned that the reader of the revised manuscript may still suffer from some misunderstanding. Reviewer #1's question appeared to include the suggestion that your approach would provide an *optimized* return of index of refraction. This is not the case - it will only return a set of possible solutions without any “relative value” associated with each solution. Please add short discussion making this explicitly clear to the (new) reader, perhaps with some guidance about whether you recommend that this approach be performed for all data collected, for special sampling opportunities, or for "spot assessments" of uncertainties in sub-sets of larger data sets.

We never intended to cause any misunderstanding about the capabilities of the proposed size distribution retrieval method and we agree that it is important to make its limitations clear to the reader. The method allows for a thorough propagation of all initial parameter PDFs into a final size distribution PDF, but it cannot rate the value of individual parameter picks, i.e. it cannot refine the initial parameter PDFs/uncertainties.

Changes to manuscript

Section 2.5.3:

We complemented the discussion about the benefits of the proposed size distribution retrieval method by a sentence clarifying the above-said. We further added a recommendation for when to use the proposed approach.

The use of “spectral” (eg at lines 18 - 22 of page 6, and line 10 of page 10, etc.), could be confusing to some readers (who may think of wavelength-dependencies). Perhaps “signal” broadening? or “size broadening” would be clearer? Or simply “light intensity variations”?

We appreciate this hint and the proposed alternatives.

Changes to manuscript

The (misleading) term “spectral broadening” is replaced by “signal broadening”.

In section 2.5.1, the discussion about light intensity variability focused on in Section 2.3 is rehashed; later you mention (page 9) changes in scattering geometry again. Please consider if these “partial” references should simply be redirects to a complete discussion of all the sources in Sec 2.3. Perhaps this section should also include a wider range of possible broadening source. Ones that occur to me include: the time response of the detectors/sensing system to particles passing through the lit region at different speeds, the temperature dependent variation in detector sensitivity, and the altitude dependent variations in background-noise that are possible with some systems...others were referenced at other points in the paper.

We concur that the rehashed discussion about light intensity variability in Sect. 2.5.1 is redundant. Yet, the subject of the discussion on page 9 isn’t instantaneous broadening of size spectra (“signal broadening”) but tries to give possible explanations for remaining (temporal) deviations between measured and modeled OPC response, e.g. induced by light source intensity fluctuations (on larger time scales). This discussion is not really belonging to the signal broadening topic treated in Sect. 2.3. Therefore, we think it is better kept separated from the latter.

Changes to manuscript

Section 2.3 and 2.5.1:

The detailed discussion about the non-uniformity of light intensity in the OPC sampling volume and its implications for the width of recorded size spectra is now included in Sect. 2.3. The respective part in Sect. 2.5.1 is replaced by a reference.

Section 2.5.2 (Page 9):

The discussion of potential reasons for remaining (time-dependent) model deviations is kept in place, but the differences to the signal broadening topic (Sect. 2.3) are now made explicitly clear. Further, some of the proposed additional sources for time-dependent OPC response fluctuations are included.

On the parametrization of optical particle counter response including instrument-induced broadening of size spectra and a self-consistent evaluation of calibration measurements

Adrian Walser^{1,2}, Daniel Sauer², Antonio Spanu³, Josef Gasteiger³, and Bernadett Weinzierl³

¹Ludwig Maximilians University (LMU), Meteorological Institute, Munich, Germany

²German Aerospace Center (DLR), Institute of Atmospheric Physics, Oberpfaffenhofen, Germany

³University of Vienna (UNIVIE), Aerosol Physics and Environmental Physics, Vienna, Austria

Correspondence to: B. Weinzierl (bernadett.weinzierl@univie.ac.at)

Abstract. Optical particle counters (OPCs) are common tools for the in situ measurement of aerosol particle number size distributions. As the actual quantity measured by OPCs is the intensity of light scattered by individual particles, it is necessary to translate the distribution of detected scattering signals into the desired information, i.e. the distribution of particle sizes. A crucial part in this challenge is the modeling of OPC response and the calibration of the instrument, i.e. in other
5 words, to establish the relation between instrument-specific particle scattering cross-section and measured signal amplitude. To date, existing methods lack a comprehensive parametrization of this-instrument-OPC response, particularly regarding the instrument-induced broadening of size-distribution-widths signal amplitude distributions. This deficiency can lead to significant
size distribution biases. We introduce an approach-overcoming-the-present-shortcomings-by-implementing-advanced-OPC
response model including a simple parametrization of the broadening effect and a self-consistent way to evaluate calibration
10 measurements using a Markov chain Monte Carlo (MCMC) method. We further outline how to obtain-consistently-derive
particle number size distributions with realistic uncertainty estimates for-OPC-size-distributions within this new framework.
Measurements-Based-on-measurements of particle standards for two OPCs, the Grimm model 1.129 (SkyOPC) and the DMT
Passive Cavity Aerosol Spectrometer Probe (PCASP), show-substantial-reduction-in-we-demonstrate-that residuals between
measured and modeled response compared-to-can-be-substantially-reduced-when-using-the-new-approach-instead-of existing
15 methods. For-the-presented-More-importantly-for-the-investigated set of measurements only the new approach yields results that
are conform with the true size distributions within the range of model uncertainty. The offered-approach-will-help-to-improve
presented innovations will help improving the accuracy of derived-OPC-derived size distributions and the assessment of their
precision for-OPC-measurements-in-general.

1 Introduction

20 The size distribution of aerosol particles is a key property to understand the impact of aerosols on human health and Earth's climate. To measure aerosol size distributions, optical particle counters (OPCs) are widely used in air quality programs and atmospheric studies. However, several studies directly comparing size distributions from different OPC instruments (e.g. Belosi

et al., 2013; Renard et al., 2016) and OPCs with other sizing methods (e.g. Reid et al., 2003; Müller et al., 2012) find significant disagreements and in some cases OPCs show systematic mis-sizing and artificial broadening of size spectra. This highlights that, although OPCs allow for a fast assessment of qualitative size information, the task to gain proper particle number size distributions can be challenging. One reason for this is the measurement principle itself, as particle size is only indirectly in-
5 ferred from scattered light intensity. This intensity, in general, is a non-monotonic function of particle size and depends also on particle intrinsic properties, such as complex refractive index and shape (Szymanski and Liu, 1986; Szymanski et al., 2009). Especially for particle sizes that are comparable or larger than the wavelength of the incident light, the size-dependence of scattered intensity tends to be flat and occasionally ambiguous, so that uncertainties in the particle intrinsic properties can introduce large sizing uncertainties (Reid et al., 2003; Formenti et al., 2011). Another reason lies in the existing methods for
10 OPC calibration and response parametrization. The available approaches (e.g. Cerni, 1983; Bemer et al., 1990; Rosenberg et al., 2012; Cai et al., 2013) are not consistent with each other. Further, they do not allow for a comprehensive description of instrument response and a satisfactory quantification of corresponding uncertainties. Figure 1 summarizes the major sources of uncertainty adjunct to OPC measurements. They can be divided into the last-mentioned uncertainty in the instrument re-
15 sponse and calibration, the uncertainty in the particle intrinsic properties and the uncertainty in the measured concentrations themselves, e.g. arising from counting statistics, eventual particle losses etc. In order to allow for inter-comparability between different OPC instruments and the comparison with other measurement techniques it is necessary to correct for systematic errors and to quantify all uncertainties as good as possible, i.e. to improve OPC data accuracy and assess its precision (Formenti et al., 2011; Mahowald et al., 2014).

In the following manuscript we focus on the central aspect of OPC response modeling and calibration and present a new
20 approach that

- allows for a more accurate description of OPC instrument response and
- yields realistic associated uncertainty estimates.

We discuss the advantages of the new approach against the background of the prevailing concepts and present its superiority by means of measurement results for two optical particle counters that were used during the Saharan Aerosol Long-range
25 TRansport and Aerosol-Cloud-Interaction Experiment (SALTRACE) (Weinzierl et al., 2017). Moreover, we outline a possible way to obtain adequate uncertainties for OPC size distributions within the new framework.

2 Methods

2.1 OPC Measurement Principle

The basic principle behind OPC measurements is that particles passing through a sampling volume illuminated by a light source
30 — usually a monochromatic laser — scatter light into a photosensitive detector. The amplitudes of the detected scattering signal pulses are a function of particle size. Counting the pulses and sorting them into discrete bins according to their amplitudes the

measurement products of a typical OPC are scattering signal amplitude histograms. The mathematical problem of retrieving number size distributions from recorded scattering signal amplitude histograms is of inverse nature and is described by a set of so called Fredholm integral equations of the first kind

$$N_i = \int_0^{\infty} \kappa_i(D) F(D) dD (+\Delta N_i) \quad (1)$$

- 5 with the number of particles N_i counted in bin i , a term ΔN_i accounting for potential counting errors, the corresponding kernel function $\kappa_i(D)$ giving the probability for each particle diameter D to be sorted into bin i and the number size distribution $F(D)$ (Kandlikar and Ramachandran, 1999; Fiebig et al., 2005).

Connecting the OPC output, i.e. the particle count histograms, and the desired information, i.e. particle number size distribution, the kernel functions are the key aspect of every OPC measurement. Deriving the kernel functions requires knowledge on the scattering signal amplitude threshold values defining the bin limits, the instrument-specific relationship between scattering signal amplitude and particle scattering cross-section and the theoretical relationship between scattering cross-section and particle size. The latter is subject to intrinsic particle properties such as complex refractive index and shape. For given ~~size-and~~ intrinsic properties the size-dependent particle scattering cross-section $C_{scat}(D)$ with respect to the incident light and OPC scattering geometry, i.e. the solid-angle range covered by the detector, can be calculated. In case of a homogeneous sphere Mie–
15 Lorenz theory (Mie, 1908) provides an analytical solution. For more complex particle shapes complementary frameworks like the T-matrix method (Waterman, 1965) or the discrete dipole approximation (Purcell and Pennypacker, 1973) can be applied.

Bridging the gap between theoretical calculations and the instrument output, i.e. finding the instrument-specific parameters linking ~~theoretical-particle-scattering-cross-section-and~~ $C_{scat}(D)$ with the measured scattering signal amplitude is the purpose of an OPC calibration. The set of instrument-specific parameters resulting from the calibration in combination with scattering
20 theory allows us to predict the OPC output, i.e. to determine the kernel functions, for any other material with given optical properties.

2.2 Existing Concepts for Size Assignment and Calibration Evaluation

Though ~~scattering-cross-section~~ $C_{scat}(D)$ and, hence, the scattering signal amplitude generally ~~is-a-are~~ non-monotonic ~~function~~ functions of particle size (see Fig. 2), the most popular approach of OPC bin size assignment is to assume or establish ~~monotony~~ monotonicity in order to simplify Eq. (1) by allowing for a one-to-one mapping between particle diameter and bin threshold values. One way to achieve ~~monotony~~ monotonicity is to replace the correct ~~theoretical-size-dependence-of-the-scattering-cross-section~~ $C_{scat}(D)$ by a smoothed monotonic approximation (Cerni, 1983; Osborne et al., 2008). Another option is to simply merge ~~the-bins-in-the-affected-size-regions~~ bins in size regions affected by ambiguities in $C_{scat}(D)$ accepting a reduction in resolution (Pinnick et al., 1981). Following these concepts OPC manufacturers usually provide their instruments with a table
30 of predefined (polystyrene latex equivalent) diameter bin threshold values. Mathematically, this means expressing the kernel

functions as sharp, adjacent step functions in diameter space

$$\begin{aligned}\kappa_i(D) &= \begin{cases} 1 & \text{for } D \in [D_i, D_{i+1}) \\ 0 & \text{otherwise} \end{cases} \\ &= \int_0^D \delta(\tilde{D} - D_i) - \delta(\tilde{D} - D_{i+1}) d\tilde{D}\end{aligned}\quad (2)$$

with delta functions at D_i and D_{i+1} , i.e. the lower and upper diameter threshold values of bin i . In doing so, Eq. (1) simplifies

5 to

$$N_i = \int_{D_i}^{D_{i+1}} F(D) dD$$

and the size distribution can be directly represented by the measured counts in a discrete way as

$$F(D) = \frac{N_i}{D_{i+1} - D_i} \text{ for } D \in [D_i, D_{i+1})$$

This simplification, however, involves fundamental shortcomings:

- 10 – Even if ~~quasi-monotony in the theoretical scattering cross-section size dependence~~ quasi-monotonicity between particle size and (discretized) scattering signal amplitude can be established for particles of certain intrinsic properties (e.g. polystyrene latex spheres) by a smart choice of OPC collecting optics (Barnard and Harrison, 1988) and/or bin threshold values, this does not automatically hold for particles of different intrinsic properties (e.g. different refractive index or shape) (Szymanski et al., 2009).
- 15 – Due to the involved approximations (e.g. a smoothing of ~~the theoretical scattering cross-section relationship~~ $C_{scat}(D)$) nominal manufacturer values can significantly deviate from reality for certain parts of the instrument size range. Such deviations are regularly reported (Szymanski and Liu, 1986; Rosenberg et al., 2012; Ryder et al., 2013).
- The instrument response can change over time, e.g. due to degradation of OPC light source intensity, pollution or misalignment of optical elements. Such changes usually do not induce a uniform shift in the apparent size distributions, but
- 20 rather cause a complicated deformation.
- No uncertainty estimates are provided for the nominal diameter threshold values. This lack entails an underestimation of size distribution uncertainties.

Some studies stick to the simplified concept of ~~bin diameter threshold values~~ adjacent bins in diameter space but try to ~~correct for possible diameter deviations. Lance et al. (2010); Cai et al. (2013)~~ reduce possible sizing deviations. Lance et al. (2010)

- 25 and Cai et al. (2013) use an empirical diameter offset to uniformly shift the manufacturer threshold values in order to yield best agreement between measured histogram modes and nominal diameter values of reference particles. A more universal

calibration approach commonly used is to find the parameters for the linear relationship between ~~the measured (mean or mode)~~ measured mean scattering signal amplitudes and theoretical ~~(mean) mean~~ scattering cross-sections for reference particles (Cerni, 1983; Bemer et al., 1990). Still assuming a monotonic ~~relation between scattering cross-section and particle diameter~~ $C_{scat}(D)$ they use the resulting linear fit parameters ~~(, i.e. slope m and intercept c)~~ to derive the size dependence of the to
 5 derive a size-dependent scattering signal amplitude $U(D) = m \cdot C_{scat}(D) + c$ and calculate the bin diameter threshold values D_i from their predefined scattering signal amplitude counterparts U_i .

Rosenberg et al. (2012) presented another way of size assignment avoiding workarounds for the non-monotonic behavior of ~~particle scattering cross-section~~ $C_{scat}(D)$. Their main new concept is to use the ~~linear fit parameters from the calibration and the unmodified theoretical relationship~~ same calibration parameters m and c and the unmodified/unsmoothed $C_{scat}(D)$ to
 10 define the kernel functions as diameter projections of the scattering signal amplitude bins (see Fig. 2a)

$$\begin{aligned} \kappa_i(D) &= \begin{cases} 1 & \text{for } C_{scat}(D) \in [C_{scat,i}, C_{scat,i+1}) \\ 0 & \text{otherwise} \end{cases} \\ &= \begin{cases} 1 & \text{for } C_{scat}(D) \in \left[\frac{U_i - c}{m}, \frac{U_{i+1} - c}{m} \right) \\ 0 & \text{otherwise} \end{cases} \end{aligned}$$

$C_{scat,i}$ and $C_{scat,i+1}$ denote the lower and upper scattering cross-section threshold values of bin i that are a linear function of the actual thresholds given by the scattering signal amplitude values U_i and U_{i+1} . This means, that particle diameters ~~with an~~
 15 ~~instrument-specific scattering cross-section of $C_{scat}(D)$~~ will be sorted into bin i if $C_{scat}(D)$ falls within the limits defined by U_i and U_{i+1} scaled with the linear coefficients m and c . This can be further expressed as

$$\kappa_i(D) = \int_0^D \delta \left(C_{scat}(\tilde{D}) - \frac{U_i - c}{m} \right) - \delta \left(C_{scat}(\tilde{D}) - \frac{U_{i+1} - c}{m} \right) d\tilde{D} \quad (3)$$

To simplify the inverse problem of Eq. (1) and, again, directly gain size distribution information from OPC histogram data they use the kernel functions to calculate so-called ~~perfect~~ “perfect” (mean) diameters $D_{p,i}$ and widths W_i to characterize all bins

$$\begin{aligned} 20 \quad D_{p,i} &= \int_0^\infty D \cdot \kappa_i(D) dD \\ W_i &= \int_0^\infty \kappa_i(D) dD \end{aligned}$$

With these values a discrete representation for the size distribution is given by

$$F(D_{p,i}) = \frac{N_i}{W_i}$$

The uncertainties in the calibration parameters m and c are used to derive instrument-related uncertainties for $D_{p,i}$, W_i and,
 25 therewith, the resulting size distribution values $F(D_{p,i})$. Although this approach supersedes workarounds for the ambiguities in ~~the size dependence of the scattering cross-section~~ $C_{scat}(D)$, it still has shortcomings. One conceptual inadequacy is that

representing the bins by their perfect diameter $D_{p,i}$ is ultimately not appropriate, as it will only match with the real mean diameter of particles sorted into bin i in the unrealistic case of a flat size distribution. If, for instance, the size distribution is (strongly) dropping towards larger particles, the occurrence of smaller particle diameters is more likely, meaning that the real mean diameter of particles falling into bin i would be (much) smaller than $D_{p,i}$. As a result, this causes a sizing bias between the real and calculated size distribution.

Spiegel et al. (2012) offer yet another approach to directly estimate size distributions from measured histograms. They translate the range of possible scattering geometries seen by individual particles¹ into a range of possible ~~particle diameter to scattering signal amplitude relationships~~ $U(D) \propto C_{scat}(D)$, a so-called “Mie band”. From this “Mie band” they calculate the a priori probabilities for discrete (equidistant) particle diameter intervals to contribute to the count rate of each bin. This approach potentially allows particle diameters to be sorted into more than one bin, i.e. overlapping OPC bins in diameter space. According to the derived contribution probabilities they then distribute the measured bin counts to the discrete diameter intervals. The major shortcoming of this method is similar to the one discussed above. Even if the a priori probabilities for two diameter intervals (of equal width) to contribute to a certain bin’s count rate are the same, the true particle abundance in these intervals will not be the same for the realistic case of a non-flat size distribution. Therefore, the inverse direction, i.e. to equally distribute the counted particles to the two intervals, is generally incorrect.

In summary, none of the existing concepts for OPC calibration and bin size assignment ~~proofs~~ proves completely satisfactory. The simplifications to the inverse problem of Eq. (1) and the attempts to directly gain size information from OPC histogram data are always accompanied by systematic errors. Further, some approaches, especially the use of the manufacturer provided set of bin diameter threshold values, do not offer instrument-related (sizing) uncertainty estimates.

2.3 Instrumental Broadening of Size Spectra

A shortcoming common to all available methods is that they do not consider the artificial broadening of size distributions in the basic parametrization of OPC response. ~~The inhomogeneity~~ One primary cause for the increase in apparent size distribution width is the non-uniformity of light intensity inside the ~~sampling volume (Wendisch et al., 1996) contributes significantly to this spectral broadening, i. e. the increase in apparent sizedistribution width~~ OPC sampling volume (e.g. Wendisch et al., 1996). An inhomogeneity of incident light intensity leads to differences in scattering signal amplitudes for particles passing the sampling volume at slightly varying locations. This means that owing to the intrinsic nature of real OPCs even spherical and homogeneous identical particles appear to vary in size. So far this signal broadening has, if at all, been treated separately from the basic instrument calibration, although it is an instrument-specific property, meaning it is always present in OPC measurements. Depending on the degree of instrument-induced ~~spectral-signal~~ broadening in relation to the actual size distribution width, this effect may lead to significant measurement biases being most pronounced for narrow size distributions (or size distribution modes). In atmospheric research, such narrow size spectra are for example met during ice residual measurements in (contrail) cirrus (Voigt et al., 2017) or aerosol chamber experiments (e.g. Schnaiter et al., 2012).

¹ In contrast to the instruments presented in this study, this is an important aspect for open path OPCs where particle positions with respect to the optics vary considerably.

~~Spectral~~Signal broadening can be further enhanced by other effects such as varying orientation of aspherical particles with respect to the direction of the incident light (Reid et al., 2003) and coincident count events (Baumgardner et al., 1985; Cooper, 1988). Latter becomes relevant for very high particle concentrations when average inter-particle distances are not larger than the size of the sampling volume any more. In such a case, the probability to erroneously interpret the sum of several
5 scattering signals from multiple particles as a single particle's signal increases. In addition to an artificial deformation of the size distribution towards larger sizes this entails an underestimation of total particle number concentration.

To correct for artificial ~~spectral broadening~~broadening of size spectra the common procedure is to define a matrix that contains the probabilities (associated with the broadening effect) to find a particle of a certain size class in adjacent size classes in its elements (Cooper, 1988; Baumgardner and Spowart, 1990; Wendisch et al., 1996; Brenguier et al., 1998). The resulting
10 inverse matrix equation is then solved for the true size distribution. One disadvantage of such methods based on empirical matrices is that their elements might not be universally valid, as for instance the magnitude of broadening that is related to varying particle orientations depends on the degree of particle asphericity. Moreover, the number of uncertainty-afflicted parameters becomes quite large. Assuming an OPC with K bins, the number of parameters required to describe ~~spectral~~signal broadening is K^2 .

15 Therefore, for the inversion of OPC histogram data it is advantageous to treat ~~spectral~~signal broadening in a more universal way. In Sect. 2.5.1 we present a new approach that includes the instrument-specific part of ~~spectral~~signal broadening within the basic parametrization of OPC response. In addition, ~~spectral~~signal broadening resulting from different orientations of aspherical particles can be included in the inversion process via a set of possible size to scattering cross-section relations as outlined in Sect. 2.5.3.

20 2.4 Uncertainty in Particle Properties

So far, we discussed the inverse nature of the OPC measurement principle, challenges and shortcomings in the parametrization of basic OPC response and the artificial broadening of size spectra. An aspect that further complicates OPC measurements is that, in most situations, the (~~size-dependent~~size-dependent) optical properties of the aerosol particles are a priori unknown or at least subject to a considerable degree of uncertainty. Externally or internally mixed individual particles can be combinations of
25 different non-homogeneously distributed materials (e.g. Kandler et al., 2011) making it difficult to find representative complex refractive indices for the bulk aerosol. In any case, the quality of OPC derived size distributions depends on the quality of information on the optical particle properties.

In order to derive a size distribution uncertainty estimate from uncertainties in the particle properties, yet, most studies follow the pragmatic approach and report the maximum impact on the size distribution as a conservative estimate (e.g. Osborne et al.,
30 2008; Weinzierl et al., 2009; Brock et al., 2011; Ryder et al., 2013; Hermann et al., 2016). However, the size distribution uncertainty induced by the uncertainties in the particle properties can be substantially ~~size-dependent~~size-dependent. To yield improved size distribution uncertainty estimates one needs realistic estimates for the set of possible size to scattering cross-section relations and a proper way to propagate these estimates (as, for example, outlined in Sect. 2.5.3).

2.5 New Approach

In this section, we introduce an approach to the parametrization of OPC response that involves instrument-specific ~~spectral~~ signal broadening and overcomes the shortcomings of existing methods. We further propose a way to evaluate calibration measurements and to obtain aerosol particle number size distributions with realistic uncertainty estimates from OPC data.

5 2.5.1 Parametrization of the Instrument Response

Let a particle of intrinsic properties ϑ (complex refractive index etc.) and diameter D have the scattering cross-section $C_{scat,\vartheta}$ with respect to the incident light and OPC scattering geometry. For an ideal instrument the scattering signal amplitude U_ϑ in the detector corresponding to this scattering cross-section would be given by the (assumed) linear relationship between scattering cross-section and scattering signal amplitude only, described by the coefficients ~~for slope~~ m and ~~intercept~~ c . Yet, ~~due to the nature of real OPCs even spherical and homogeneous identical particles do not necessarily generate the same scattering signal amplitudes. One primary cause for this is the non-uniformity of light intensity in the sampling volume. Particles passing at slightly different locations scatter light of different incident intensity which linearly confers on the intensity of the scattered light. As a consequence, the relative spread in scattering signal amplitudes can be assumed to be independent from their absolute value. Therefore, as explained in Sect. 2.3, instrument-induced signal broadening causes~~ the single signal amplitude U_ϑ ~~is to~~ be replaced by a probability density function (PDF) for a set of possible $\{U_\vartheta\}$. Under the assumption that the non-uniformity of light intensity is a primary reason for the broadening, the relative variance of this PDF is independent of the absolute value of U_ϑ . In a simplified approach, ~~this PDF can be assumed to follow the U_ϑ PDF is, hence, approximated by~~ a Gaussian distribution with a constant relative standard deviation b . ~~Equivalently, one could describe~~ This is equivalent to a “blurring” of the initially sharp OPC scattering cross-section bin threshold values (resulting from the predefined scattering signal amplitude bin thresholds) ~~as being “blurred”~~ with the same relative standard deviation b . Replacing the delta functions in Eq. (3) (i.e. ~~Gauss-Gaussian~~ functions of vanishing standard deviation) by ~~Gauss-Gaussian~~ functions of constant relative standard deviation yields the new kernel functions

$$\kappa_i(D | b, m, c) = \frac{1}{\sqrt{2\pi}b} \int_0^D \frac{1}{U_i} \exp \left(-\frac{\left(C_{scat,\vartheta}(\tilde{D}) - \frac{U_i - c}{m} \right)^2}{2b^2 U_i^2} \right) - \frac{1}{U_{i+1}} \exp \left(-\frac{\left(C_{scat,\vartheta}(\tilde{D}) - \frac{U_{i+1} - c}{m} \right)^2}{2b^2 U_{i+1}^2} \right) d\tilde{D} \quad (4)$$

with the new instrument-specific parameter triplet (b, m, c) and the scattering signal amplitude threshold values U_i and U_{i+1} defining bin i . Figure 2 illustrates the difference in OPC kernel functions between an ideal instrument that follows Eq. (3) and an instrument with a finite relative Gaussian signal broadening. While the assumption of a Gaussian broadening is an adequate approximation for the OPCs used in this study, it is possible to customize the ~~spectral~~ shape of the broadening effect for other OPC types (e.g. open-path geometries), if necessary.

2.5.2 Calibration Evaluation

~~Including spectral~~ Taking account of signal broadening, the new parametrization allows for an extension of the classical OPC calibration evaluation approach that is restricted to the determination of the linear coefficients m and c .

Given a set of particle standards with known intrinsic properties and size distributions the forward solution of Eq. (1) using Eq. (4) for the kernel functions yields the model count histograms, i.e. the parametrized theoretical instrument response

$$M_{ij} = \int_0^{\infty} \kappa_i(D | b, m, c) F_j(D) dD \quad (5)$$

with the model counts M_{ij} for OPC bin i and particle standard j and the corresponding number size distribution $F_j(D)$. With the real measured particle counts N_{ij} the task of a calibration within the new framework is now to inversely find the values for the parameters that bring M_{ij} and N_{ij} into best agreement.

For stable measurement conditions, i.e. constant OPC volumetric sample flow etc., the uncertainties ~~on~~ of the measured particle counts follow the Poisson counting statistics. With increasing number of counts, the relative uncertainty hence decreases with $N_{ij}^{-1/2}$. Naturally, the simplified model will not be able to reproduce the calibration measurements perfectly, because there will be additional deviations that are not parametrized. Provided sufficiently high numbers of counts in the course of the sampling, the relative bin count uncertainties due to Poisson counting statistics will become negligible compared to these additional deviations. In consequence, bringing model and measurement into agreement corresponds to maximizing the probability for the model counts M_{ij} afflicted with a priori unknown uncertainties σ_{ij} , that cover the additional model deviations, to occur given the measured N_{ij} . To ensure that the modeled instrument response later agrees with reality within its margin of uncertainty, it is necessary to find a good representation of the unknown uncertainties σ_{ij} and quantify them in the course of the calibration, too.

~~Particle standard measurements show that the~~ Comparing measured particle standard histograms with the corresponding model results for a suitable instrument parameter tuple (b, m, c) reveals that remaining deviations between ~~modeled and measured OPC response~~ the two mainly appear as (non-uniform) small shifts ~~between the corresponding histograms~~, meaning that compared to the model histograms some measured histograms are shifted to smaller while others are shifted to larger scattering signal amplitudes. It thus seems natural to treat these deviations as a remaining uncertainty of the modeled scattering signal amplitudes. Apart from the experimental finding there are also theoretical explanations for the observed shifts. ~~Small differences between the actual OPC scattering geometry and the one used in theoretical calculations as studied by Rosenberg et al. (2012) can to first-order approximation be assumed to lead to a collective relative shift in particle scattering cross-sections. OPC~~ OPC light source intensity fluctuations around a temporal average ~~can further cause histogram induce~~ time-dependent scattering signal amplitude variations, causing measured histograms to slightly move up and down with time. In contrast to the instantaneous signal broadening discussed in Sect. 2.3, these fluctuations act on greater time scales leading to histograms shifts that are differently pronounced for samples recorded with certain time lags. ~~These~~ Other possible sources for such time-dependent ~~intensity shifts are changes in detector sensitivity or signal background noise. Due to the linear relationship, time-dependent signal amplitude~~ fluctuations can equivalently be thought of as relative fluctuations in the scatter-

ing cross-sections assuming a fixed light source intensity. Therefore, we express the model count uncertainties σ_{ij} in terms of a relative uncertainty of the theoretical particle scattering cross-sections $C_{scat,\vartheta}(D)$.

For a given instrument parameter tuple (b, m, c) the set of model bin counts M_{ij} results from Eq. (5) and (4) with the (best estimate) theoretical function $C_{scat,\vartheta}(D)$. A relative shift in the theoretical scattering cross-sections corresponding to a multiplication of $C_{scat,\vartheta}(D)$ by a factor $\varepsilon \neq 1$ leads to a different set of model bin counts $M_{ij,\varepsilon}$. Assuming the PDF of the possible relative shifts ε to be a Gaussian function centered at 1 and having a standard deviation of σ_ε one can derive the respective PDFs for the model bin counts $M_{ij,\varepsilon}$. For the sake of convenience and simplicity the resulting model bin count PDFs can themselves be approximated by Gaussian PDFs, which is usually an adequate approximation. This leads to the following expression for the unknown model bin count uncertainties

$$\sigma_{ij}^2 = \frac{1}{\sqrt{2\pi\sigma_\varepsilon^2}} \int_0^\infty (M_{ij,\varepsilon} - M_{ij})^2 \exp\left(-\frac{\varepsilon^2}{2\sigma_\varepsilon^2}\right) d\varepsilon \quad (6)$$

with $M_{ij,\varepsilon}$ defined by Eq. (5) and (4) replacing $C_{scat,\vartheta}(D)$ with $C_{scat,\vartheta,\varepsilon}(D) = \varepsilon \cdot C_{scat,\vartheta}(D)$. In summary, the new calibration evaluation should yield the set of model parameters $(b, m, c, \sigma_\varepsilon)$ composed of the instrument-specific parameter tuple (b, m, c) and, according to the above considerations, the remaining relative uncertainty of the theoretical scattering cross-sections σ_ε .

A way to meet the challenge of model parameter probability maximization under initially unknown model uncertainties is to make use of Bayesian statistics and Markov chain Monte Carlo (MCMC) methods (e.g. Goodman and Weare, 2010). Following Bayes' theorem (Bayes and Price, 1763) the (posterior) probability P for a set of model bin counts $\{M_{ij}\}$ to occur under a set of measured bin counts $\{N_{ij}\}$ can be expressed as

$$P(\{M_{ij}\} | \{N_{ij}\}) \propto P(\{N_{ij}\} | \{M_{ij}\}) \cdot P(b, m, c, \sigma_{ij}) \quad (7)$$

, i.e. the product of the likelihood function determining the probability of the $\{N_{ij}\}$ to occur given the $\{M_{ij}\}$ and the so-called prior probability $P(b, m, c, \sigma_{ij})$, including all prior knowledge on the model parameters for instance from physical constraints or invariance considerations (e.g. Jaynes, 1968). The proportionality factor equating both sides of Eq. (7) can be thought of as a normalization constant. Upon the assumption of Gaussian model bin count PDFs the likelihood function can be expressed as

$$P(\{N_{ij}\} | \{M_{ij}\}) = \prod_{ij} \frac{1}{\sqrt{2\pi\sigma_{ij}^2}} \exp\left(-\frac{(N_{ij} - M_{ij})^2}{2\sigma_{ij}^2}\right) \quad (8)$$

with M_{ij} and σ_{ij} defined by Eq. (5) and (6) respectively. MCMC methods allow us to efficiently sample the model parameter space utilizing the forward solution to the problem to find the region of maximum probability according to Eq. (7). This way, the PDFs for the instrument parameters (b, m, c) and the relative uncertainty of the theoretical particle scattering cross-sections σ_ε are obtained together with all correlations between the individual parameters. In this study we utilize the Python-based sampler tool *emcee* (Foreman-Mackey et al., 2013).

2.5.3 Retrieval of Size Distributions within the New Framework

The new instrument parametrization, including instrument-specific ~~spectral-signal~~ broadening and the parameter PDFs resulting from the MCMC-based calibration evaluation, now permits to derive size distributions from OPC measurements in a self-consistent way. Propagating the parameter uncertainties yields improved estimates for the corresponding size distribution uncertainties. Figure 3 illustrates a possible workflow within the proposed framework to go from measured OPC count histogram data to PDFs in size distribution solutions. Similar to what has been proposed by Fiebig et al. (2005) the basic idea is to start with random Monte Carlo samples drawn from the model parameter PDFs and a set of possible theoretical particle diameter to scattering cross-section relationships $\{C_{scat,\vartheta}(D)\}$, e.g. given by the likely range of aerosol particle complex refractive indices, their shape and orientation². In addition, a random relative shift from the chosen $C_{scat,\vartheta}(D)$ is picked according to its potential (time-dependent) systematic deviation, e.g. induced by light source intensity fluctuations. This relative shift is drawn from the Gaussian PDF $\mathcal{N}(1, \sigma_\varepsilon^2)$ parametrized by σ_ε , which is derived as part of the calibration evaluation. With the resulting (shifted) $C_{scat,\vartheta,\varepsilon}(D)$ and the instrument parameter tuple (b, m, c) the set of OPC kernel functions $\{\kappa_i(D)\}$ can be calculated following Eq. (4). Given this set of bin kernel functions the aerosol size distribution $F(D)$ is adjusted such that the (uncertainty-weighted) deviations between modeled and measured bin counts are minimized. The result is then either one best solution for $F(D)$ or an ensemble of possible solutions for each iteration, depending on the respective inversion algorithm (see e.g. Kandlikar and Ramachandran, 1999; Fiebig et al., 2005, and the references herein). By repeating this procedure multiple times one finally acquires a collective solution ensemble, representing members of the size distribution solution PDF, considering all uncertainties in instrument specific parameters and the theoretical diameter to scattering cross-section relationships. In this work (see Sect. 4) we use a parametrized size distribution and, again, a MCMC method for the inversion. The corresponding size distribution parameter solutions obtained for each Monte Carlo iteration are merged into a final size distribution parameter solution ensemble.

Apart from a thorough and transparent derivation of size distribution uncertainties, the proposed retrieval method has further advantages. For one, the Monte Carlo sampling enables a one-to-one mapping between each size distribution solution (ensemble member) and the corresponding initial parameter picks, thereby for instance facilitating parameter sensitivity studies. ~~Moreover, the resulting~~ This can help to identify dominating initial parameter uncertainties and even allow to confine the initial parameter estimates by comparing the traceable solutions to size distribution results obtained by independent measurements. It should be clarified, however, that the proposed retrieval method alone simply propagates the initial parameter PDFs and cannot provide information on their adequacy, i.e. per se it cannot judge the value of individual parameter picks. A big advantage of the method is that the solution ensemble itself allows for a simple while appropriate further propagation of size distribution uncertainties, ~~e.g. for the calculating of quantities like the~~. This is explicitly useful when calculating quantities depending on the size distribution. For instance, PDFs for the effective particle diameter or the aerosol extinction coefficient ~~-can easily be (numerically) derived by collecting the individual results obtained for each size distribution solution (ensemble member). In~~

²Other uncertainty-afflicted instrument properties as, for instance, sample flow rate or size-dependent aspiration efficiency can be randomly sampled in a comparable manner.

view of all benefits, we recommend using the proposed retrieval method (or equivalent approaches) for every occasion, despite the additional effort involved.

3 Experimental Setup

3.1 Involved OPCs

5 The two central OPCs examined in this study are the Grimm model 1.129 (SkyOPC) and the DMT Airborne Passive Cavity Aerosol Spectrometer Probe (PCASP-100X with an upgraded signal processing package SPP-200, abbreviated PCASP hereafter). Both aerosol spectrometers were part of the airborne in situ instrumentation used in the SALTRACE campaign. The SkyOPCs were operated inside the cabin of the German Aerospace Center’s Falcon research aircraft behind an isokinetic aerosol inlet, the PCASP was mounted in one of the under-wing stations. Detailed descriptions of the instruments can be found
10 in Bundke et al. (2015) for the SkyOPC and in Liu et al. (1992) and Strapp et al. (1992) for the PCASP. Both are closed path spectrometers in which the aerosol particle beam is confined to the inner area of light source focus. The particles scatter light coming from a monochromatic laser of visible red wavelength (633nm Helium-Neon laser for the PCASP and 655nm diode laser for the SkyOPC) which is then detected in a sideways direction. The applied wide-angle collection of scattered light minimizes ambiguities in ~~the particle size to scattering cross-section relationship~~ $C_{scat}(D)$ (see Heim et al. (2008) and Rosenberg
15 et al. (2012) for details on the scattering geometry).

During SALTRACE and the lab measurements presented here the SkyOPC was operated in the fast mode for smaller sizes covering a nominal diameter range of 0.25 to about $3\mu\text{m}$. The corresponding scattering signal amplitude range is separated into 16 preset bins defined by set of digital threshold values. In standard configuration, the PCASP sorts scattering pulses into 30 bins over a nominal diameter range of 0.1 to $3\mu\text{m}$. It further allows for a custom selection of the digital bin threshold values. In
20 this study, we only consider the PCASP low gain stage³. In order to better study differences between the approaches discussed in Sect. 2, we present results for a custom high resolution binning of this gain stage in addition to the default binning. For the custom binning the gain stage’s signal amplitude range is divided into bins with constant width.

The DMT Ultra-High Sensitivity Aerosol Spectrometer (UHSAS) (Cai et al., 2008) lab version covering a size range of about 0.06 to $1\mu\text{m}$ in high resolution (99 bins) was utilized as a reference for total particle concentration during the SkyOPC
25 calibration measurements and further served for a qualitative assessment of sizing to support the particle mobility filtering described in Sect. 3.2.

3.2 Measurements

The calibration measurements were performed using monodisperse aerosols of polystyrene latex (PSL) spheres. For the Sky-OPC, the data set is complemented by di(2-ethylhexyl) sebacate (DEHS) aerosol samples. The complex refractive indices for

³As the intensity of scattered light intensity over the PCASP size range covers more than six orders of magnitude the PCASP optical detection system is divided into three amplification stages, called the high gain, mid gain, and low gain stage.

PSL and DEHS are approximately $1.585 + i0$ (Sultanova et al., 2009) and $1.45 + i0$ (manufacturer data sheet) in the wavelength range of the SkyOPC and PCASP. PSL spheres dispersed in distilled water were mobilized via nebulization with the DMT portable aerosol generator running with aerosol-free carrier air. The resulting aerosol was subsequently dehumidified by an arrangement of silica gel dryers and diluted to avoid OPC measurement issues related to coincident count events. The DEHS aerosol was produced using a TSI Model 3475 condensation aerosol generator based on the Sinclair-LaMer principle (Altmann and Peters, 1992): with nitrogen as the carrier gas an aqueous sodium chloride solution is nebulized and dried to yield a high-concentration condensation nuclei aerosol. Passing through a heated vessel filled with liquid DEHS, a reheater unit and a condensation chimney the precursor particles allow for heterogeneous condensation of the supersaturated DEHS vapor. The mean particle size of the resulting (quasi-)monodisperse aerosol is a function of the ratio between vapor concentration and condensation nuclei number concentration. Again, the DEHS aerosol was diluted prior to the measurement in the OPC to circumvent counting coincidences. For all measurements we chose sampling interval times long enough to minimize relative uncertainties from counting statistics.

For mean particle diameters up to 800nm the aerosol was additionally filtered with the aid of a differential mobility analyzer (Grimm Vienna type L-DMA, abbreviated DMA hereafter, Reischl et al. (1997)). DMA filtering substantially reduced the widths of the DEHS aerosol size distributions and allows us to obtain quantitative information on their mean diameters and widths. Using the UHSAS, it was taken care that the DMA transfer function was centered to the middle of the initial DEHS generator aerosol size distribution to guarantee that the resulting size distribution is in good approximation represented by the Gaussian DMA transfer function itself. The relative standard deviations of the DMA transfer functions are calculated by means of the formulae given in Reischl et al. (1997) and Stolzenburg (1988). The initial PSL particle size distributions, that are traceable via the United States National Institute of Standards and Technology (NIST), are of Gaussian shape with known mean and standard deviation. Although they are narrower than the width of the DMA transfer functions, additional DMA filtering helped to effectively remove the interfering background at smaller particle diameters that is caused by the nebulization (cf. Hermann et al. (2016)). For mean particle diameters larger than 800nm the presented counting histograms are empirically separated from this background.

25 4 Results and Discussion

In this section, we present the results for the evaluation of the PSL calibration measurements following the new method proposed in Sect. 2.5. We compare these results with the theoretical instrument response for nominal manufacturer diameter bin thresholds and results obtained for the approach of Rosenberg et al. (2012), representing a state-of-the-art conventional method. Hereafter, we abbreviate these approaches as MFR and R12, respectively. We further demonstrate the impact of method choice on size distribution inversion results for measurements of DEHS samples.

The measurements of PSL particle standards, carried out as described in Sect. 3.2, are utilized to calibrate the OPCs following both the new and the R12 approach (introduced in Sect. 2.2). Figure 4 and 5 contrast the resulting modeled relative bin count histograms and the measured relative histograms for the SkyOPC and the PCASP (low gain stage) respectively. The model

histograms are calculated by means of Eq. (5) with the well-defined Gaussian PSL size distributions and the kernel functions given by Eq. (3) for the R12 (shown in red brown colors) and Eq. (4) for the new approach (shown in blue colors). The best estimate model histograms, i.e. the model histograms for the maximum probability model parameter tuple — $(m, c)_{best}$ for the R12 and $(b, m, c, \sigma_\varepsilon)_{best}$ for the new approach — are represented by the color-framed white histogram bars. For the SkyOPC, additionally the model histograms for the MFR approach following Eq. (2) and using the manufacturer-supplied set of nominal values are displayed in golden colors. The underlying measured histograms are depicted by the gray bars. For the new and the R12 approach the parameter PDFs resulting from the evaluation of the calibration measurements (see Fig. A1 and A2) are sampled using a Monte Carlo method to yield the corresponding PDFs of the model histogram bin counts that are visualized by error bars spanning the range between the 16 and 84th percentiles. In each panel of Fig. 4 and 5 the mean diameter and standard deviation of the Gaussian PSL size distribution is displayed in the left upper corner. Figure 6 supplements the SkyOPC histogram comparisons with scatter plots showing all modeled and measured relative bin counts for the different approaches. Finally, Fig. 7 quantitatively compares the total sum of residuals

$$\sum_{ij} R_{ij,best} = \sum_{ij} \sqrt{(N_{ij} - M_{ij,best})^2}$$

between the measured N and best estimate model relative bin counts M for the two instruments and the different approaches. The subscripts i and j represent the different OPC bins and used particle standards respectively.

The model histograms for the MFR approach (e.g. Fig. 4, golden colors) exhibit significant deviations from the underlying measured histograms. They offer much smaller widths than their measured counterparts. In addition, absolute offsets between the histogram modes are apparent for both SkyOPC and PCASP (not shown). Deviations are largest for the SkyOPC, because it was operating under dusty conditions during SALTRACE over a longer period previous to the presented measurements, presumably causing a pollution of optical elements. In consequence, the scatter plots for the MFR approach in the upper row of Fig. 6 show the largest discrepancy between model and measurements. This becomes also obvious for both instruments when looking at the total sums of residuals in Fig. 7. The residuals for the MFR approach are substantially enhanced compared to the others.

The R12 approach allows for the correction of the absolute shifts of the histogram modes. Nevertheless, instrument-specific ~~spectral-signal~~ broadening is still ignored. The modeled histograms, thus, continue to underestimate the widths of the actually measured histograms, which is visible in the histogram plots in Fig. 4 and 5. Here, and especially in Fig. 6 it is also apparent that the R12 approach remains unable to reproduce the measurements within the margins of model uncertainty for most of the relative bin counts. Particularly for the smaller relative count values the absence of a parametrization of ~~spectral-signal~~ broadening leads to large model deviations. However, in comparison to the MFR approach total residuals for the model best estimates are reduced by 25 and 35% for the SkyOPC and PCASP (low gain stage, default binning) respectively. Beyond that, an estimate for the model uncertainty is established.

By introducing a simple parametrization of instrument-specific ~~spectral-signal~~ broadening and a self-consistent way of evaluating OPC calibration measurements, the new method succeeds in modeling the measured histogram widths correctly (see Fig. 4 and 5 rightmost columns). As a result, the total residuals between measured and modeled relative bin counts for the

model best estimates decrease by 82 and 77% compared to the MFR approach for the SkyOPC and PCASP (low gain stage, default binning) respectively. With respect to the R12 approach total residuals for the SkyOPC, the PCASP default binning and the finer PCASP custom binning are lowered by 77, 64 and 76%. Further, Fig. 6 shows that the new approach ~~proofs~~proves capable to correctly reproduce the measured histograms within the margins of model uncertainty over the complete range of relative bin counts.

Figure 8 shows the SkyOPC counting efficiency curve obtained by parallel measurements with the UHSAS as a reference counter during the PSL calibration measurements. These measurements offer another perspective on the comparison between the two approaches. The mean total concentration fractions measured by the SkyOPC

$$f_{msm,j}(D_j) = \frac{\sum_i N_{ij, \text{SkyOPC}}}{\sum_k N_{kj, \text{UHSAS}}} \cdot \frac{\Phi_{\text{UHSAS}}}{\Phi_{\text{SkyOPC}}}$$

- are calculated from the respective total number of counts and the volumetric instrument flow rates Φ for each particle standard j and are depicted by the red diamond markers. The associated 68% confidence intervals (approximately corresponding to \pm one standard deviation) that result from error propagation involving count rate scatter and instrument sample flow uncertainties are represented by the red error bars. The modeled concentration fractions are derived from the bin kernel functions κ_i as

$$f_{mdl}(D) = \sum_i \kappa_i(D)$$

- and are visualized by the solid lines for the model best estimates, again in red brown for the R12 and in blue for the new approach. The shaded areas show the range between the 16 and 84th percentiles derived from the model parameter PDFs. The R12 approach predicts a sharp drop-off to smaller particle diameters in contrast to the measurements. The new approach is able to correctly model both shape and absolute values of the observed sigmoidal behavior of the counting efficiency curve.

- Measurements of DEHS samples, as outlined in Sect. 3.2, allow us to test the possible implication of the choice of method for size distribution inversion results using an independent material. As proposed in Sect. 2.5.3 and illustrated in Fig. 3, the inversion of measured OPC histogram data is based on the parametrization of instrument response, the respective parameter PDFs derived from the calibration and ~~the particle diameter to scattering cross-section relationship~~ $C_{scat,\vartheta}(D)$ for the new material. The use of DEHS spherical droplets guarantees that this latter relationship is well-defined for the given scattering geometry as complex refractive index and shape of the aerosol particles are known, thus adding no further complexity to the retrieval. Moreover, the size distribution of the filtered DEHS samples approximately follows a Gaussian distribution simplifying the inversion in this case to the determination of the size distribution parameters, mean diameter μ_{sd} and standard deviation σ_{sd} . The inversion algorithm used here to solve Eq. (1) for the parametrized size distribution is based on a MCMC method (Goodman and Weare, 2010; Foreman-Mackey et al., 2013). To obtain adequate size distribution parameter PDFs 10000 Monte Carlo samples are drawn from the corresponding instrument parameter PDFs. Figure 9 shows the inversion results for two DEHS samples and the two methods, i.e. the R12 approach (red brown) and the new one (blue). The theoretical (true) size distributions and the corresponding values for the means and standard deviations are depicted by the red lines and markers. For both the new and R12 approach the retrieved size distribution means agree with the theoretical values within their range of uncertainty, meaning that both methods allow for a correct (mean) sizing. This finding additionally ~~proofs~~proves

the validity of the used ~~particle-diameter-to-scattering-cross-section-relationships~~ $C_{scat,\vartheta}(D)$ (for DEHS and the calibration material PSL). Upon closer inspection the retrieved means tend to slightly underestimate the true values, which could imply minor deviations between the true and the used OPC scattering geometry and/or refractive index values for PSL and DEHS. The parameter PDFs for the size distribution means μ_{sd} are almost identical for the two methods concerning both PDF median values and widths, i.e. uncertainty ranges. This agreement disappears for the size distribution standard deviations σ_{sd} . The new method again agrees with the theoretical values within the range of parameter uncertainty and, hence, successfully predicts the full shape of the size distribution. The R12 approach attributes the ~~spectral~~-width of a measured histogram completely to the width of the size distribution, thus overestimating σ_{sd} significantly. For the examples shown here, the R12 approach overestimates the true values for σ_{sd} by 714 and 302% with respect to the medians of the retrieved parameter PDFs. The widths of the σ_{sd} parameter PDFs, i.e. the estimated range of uncertainty in this parameter, also differs for the two methods. With respect to the distance between the 16 and 84th percentiles the R12 approach yields 285 and 224% higher PDF widths than the new method leading to greater overall uncertainties in the retrieved size distributions, which are nonetheless unable to encompass the true ones.

It should be noted, though, that the standard deviations of the DEHS size distributions used here are quite small. When size distributions become broader the impact of instrument-specific ~~spectral~~-signal broadening on the width of the recorded histograms decreases and, hence, differences between the methods will become less pronounced. Besides, uncertainties in aerosol properties like complex refractive index and shape might be the dominant source of size distribution uncertainty in many situations. However, this example demonstrates that the new method is able to retrieve even narrow size distributions correctly and, hence, to provide access to realistic uncertainty estimates for all situations. The results also imply that even for the same data and OPC instrument, calibrated with the same set of measurements, retrieved size distributions can be contradictory solely due to different instrument response parametrizations and calibration evaluation approaches.

5 Conclusions

Retrieving aerosol particle number size distributions and associated uncertainties from OPC histogram data is a challenging task. Scattered light intensity (the measurand) generally is a non-monotonic function of particle size (the quantity of interest) and depends also on particle intrinsic properties such as complex refractive index. Besides, due to the non-ideal behavior of real OPCs, measured intensity distributions are artificially broadened. To realistically model OPC response, i.e. to find suitable OPC bin kernel functions defining the probabilities for particle diameters to be sorted into the instrument's discrete scattering signal amplitude bins, hence, is a crucial requirement.

We have introduced a new approach to model OPC response and, within this framework, a self-consistent way for the evaluation of calibration measurements. Two OPCs involved in the SALTRACE campaign, the SkyOPC and the PCASP, and measurements of PSL particles have been utilized to compare the new approach with existing concepts. The results lead to the following conclusions:

The manufacturer provided set of (PSL equivalent) nominal diameter threshold values for the OPC bin borders should be treated with caution and the resultant size distributions should be considered as rather qualitative measures. Not only the concept of adjacent continuous bins in diameter space can be problematic given the non-monotonic relation between particle size and scattering signal amplitude, but the values are also material-dependent and drifts in size-assignment, e.g. due to pollution of OPC optics or light source intensity drifts, can occur over time. We have shown that the corresponding size distributions can significantly deviate from reality, even for the reference material. Furthermore, no uncertainty estimates are provided for the nominal diameter values that could be used to infer instrument-related size distribution uncertainties.

Calibrating the instrument can remove absolute sizing offsets. The results for a state-of-the-art OPC calibration and response parametrization approach (Rosenberg et al., 2012) exhibit clear improvements in sizing and, therewith, a reduction in total residuals between modeled and measured bin histograms. The introduction of instrument parameter uncertainties that goes along with the calibration evaluation allows to derive related size distribution uncertainty estimates. However, these estimates fail to explain remaining differences between modeled and measured instrument response for the presented data. The main reason for this is the absence of a parametrization of instrument-induced ~~spectral~~-signal broadening. This artificial increase in apparent size distribution width, which is stronger the narrower the actual size distribution is compared to the degree of broadening, may involve significant systematic OPC measurement biases (in atmospheric research) when disregarded.

By introducing a simple (one parameter) approach to describe this ever-present broadening of size spectra the new method leads to substantial improvements. Residuals between modeled and measured OPC response are considerably reduced compared to the other methods. The new method further correctly predicts the size-dependence of OPC counting efficiency. Most importantly, the measurements are successfully reproduced within the range of model uncertainty.

In the context of the new method we have also outlined a self-consistent way to thoroughly propagate parameter uncertainties and gain realistic size distribution PDFs without avoiding to address the actual inverse problem underlying OPC measurements. Besides the advanced uncertainty assessment, a benefit of the proposed Monte Carlo retrieval procedure is the facilitation of subsequent uncertainty propagation for quantities calculated from the size distribution (e.g. the effective diameter). When this procedure is combined with the new OPC response model, exemplary results for measurements of DEHS samples demonstrate that even narrow size distributions are retrieved correctly. For the conventional method the same retrieval procedure, propagating the corresponding parameter uncertainties, yields larger size distribution uncertainties and significantly overestimated size distribution widths.

In summary, the new method has the following major advantages over existing concepts for OPC bin size assignment:

- The inevitable instrument-specific broadening of measured size spectra is parametrized for the first time leading to a more accurate modeling of OPC response.
- The model parameter PDFs resulting from the evaluation of calibration measurements allow for realistic uncertainty estimates for this response and, in consequence, provide a basis for proper size distribution uncertainties.

Appendix A

Figure A1 and A2 show the calibration results for the PCASP low gain stage with the custom high resolution binning following the new and the R12 approach, respectively. As explained in detail in Sect. 2.5 the parameter solution ensemble for the new approach is obtained with the aid of the Python-based MCMC sampler tool *emcee* (Foreman-Mackey et al., 2013) using Eq. (8) for the likelihood function.

For the R12 approach the theoretical particle scattering cross-section means $\overline{C_{scatj}}$ and standard deviations $(\sigma_{C_{scat}})_j$ for each particle standard j are calculated from the known PSL size distributions and the theoretical ~~size-dependence-of-the scattering-cross-section~~ $C_{scat}(D)$ following Eq. (4) and (5) in Rosenberg et al. (2012). Accordingly, for the scattering signal amplitudes the measured histogram mode bin centers \overline{U}_j and half widths $(\sigma_U)_j$ are calculated. Further adapting their procedure, for the linear fit between the two properties the scattering cross-section and signal amplitude PDFs for each particle standard are approximated by uncorrelated Gaussian distributions with the afore-mentioned values defining the respective means and standard deviations. In order to yield a parameter solution ensemble similar to the new approach, in this study the fitting is conducted by means of the same MCMC tool using the following logarithmic likelihood function for a linear relation between two properties with uncorrelated Gaussian uncertainties:

$$\ln P = K - \sum_j \frac{(m \cdot \overline{C_{scatj}} + c - \overline{U}_j)^2}{2(m^2 \cdot (\sigma_{C_{scat}})_j^2 + (\sigma_U)_j^2)}$$

m and c represent the fit slope and intercept, K is a (neglectable) constant offset that has no influence on the maximization process and the resulting parameter solution PDFs.

The two approaches' results for m and c are consistent with respect to the corresponding uncertainty ranges (see Fig. A1 and A2) with the new method yielding smaller uncertainty ranges for both m and c . The median values for the new method's additional parameters b and σ_ϵ , describing the relative standard deviation ("blurring") of the scattering cross-section bin threshold values and the remaining relative uncertainty of ~~the-theoretical-scattering-cross-section-relation~~ $C_{scat}(D)$, approximately amount to 22 and 10%, respectively. The slight cutting of larger values in the PDF for c resulting for the R12 approach is caused by the physical constraint that the scattering cross-section values for the bin thresholds may not be negative.

Acknowledgements. The research leading to these results has received funding from the Helmholtz Association under grant number VH-NG-606 (Helmholtz-Hochschul-Nachwuchsforschergruppe AerCARE) and from the European Research Council under the European Community's Horizon 2020 research and innovation framework program/ERC grant agreement number 640458 (A-LIFE). The SALTRACE campaign was mainly funded by the Helmholtz Association, DLR, TROPOS, and LMU. We further acknowledge funding from the LMU Munich's Institutional Strategy LMUexcellent within the framework of the German Excellence Initiative, and from the European Union through the European Seventh Framework Programme (FP7 2007-2013) under grant agreement number 607905 (Marie Curie Initial Training Network VERTIGO). We also would like to thank M. Richter (GRIMM Aerosol Technik GmbH & Co. KG, Ainring, Germany) for fruitful discussions and information about their instruments. We acknowledge the use of the software module providing the corner plots by Foreman-Mackey et al. (2016).

References

- Altmann, J. and Peters, C.: Proceedings of the 1992 European Aerosol Conference The adjustment of the particle size at a Sinclair-La Mer-Type Aerosol Generator, *Journal of Aerosol Science*, 23, 277–280, doi:10.1016/0021-8502(92)90403-I, 1992.
- Barnard, J. C. and Harrison, L. C.: Monotonic responses from monochromatic optical particle counters, *Applied Optics*, 27, 584–592, doi:10.1364/AO.27.000584, 1988.
- Baumgardner, D. and Spowart, M.: Evaluation of the Forward Scattering Spectrometer Probe. Part III: Time Response and Laser Inhomogeneity Limitations, *Journal of Atmospheric and Oceanic Technology*, 7, 666–672, doi:10.1175/1520-0426(1990)007<0666:EOTFSS>2.0.CO;2, 1990.
- Baumgardner, D., Strapp, W., and Dye, J. E.: Evaluation of the Forward Scattering Spectrometer Probe. Part II: Corrections for Coincidence and Dead-Time Losses, *Journal of Atmospheric and Oceanic Technology*, 2, 626–632, doi:10.1175/1520-0426(1985)002<0626:EOTFSS>2.0.CO;2, 1985.
- Bayes, M. and Price, M.: An Essay towards Solving a Problem in the Doctrine of Chances. By the Late Rev. Mr. Bayes, F. R. S. Communicated by Mr. Price, in a Letter to John Canton, A. M. F. R. S., Royal Society of London, 1763.
- Belosi, F., Santachiara, G., and Prodi, F.: Performance Evaluation of Four Commercial Optical Particle Counters, *Atmospheric and Climate Sciences*, 03, 41–46, doi:10.4236/acs.2013.31006, 2013.
- Bemer, D., Fabries, J. F., and Renoux, A.: Calculation of the theoretical response of an optical particle counter and its practical usefulness, *Journal of Aerosol Science*, 21, 689–700, doi:10.1016/0021-8502(90)90123-F, 1990.
- Brenguier, J.-L., Bourrianne, T., Coelho, A. A., Isbert, J., Peytavi, R., Trevarin, D., and Weschler, P.: Improvements of Droplet Size Distribution Measurements with the Fast-FSSP (Forward Scattering Spectrometer Probe), *Journal of Atmospheric and Oceanic Technology*, 15, 1077–1090, doi:10.1175/1520-0426(1998)015<1077:IODSDM>2.0.CO;2, 1998.
- Brock, C. A., Cozic, J., Bahreini, R., Froyd, K. D., Middlebrook, A. M., McComiskey, A., Brioude, J., Cooper, O. R., Stohl, A., Aikin, K. C., de Gouw, J. A., Fahey, D. W., Ferrare, R. A., Gao, R.-S., Gore, W., Holloway, J. S., Hübler, G., Jefferson, A., Lack, D. A., Lance, S., Moore, R. H., Murphy, D. M., Nenes, A., Novelli, P. C., Nowak, J. B., Ogren, J. A., Peischl, J., Pierce, R. B., Pilewskie, P., Quinn, P. K., Ryerson, T. B., Schmidt, K. S., Schwarz, J. P., Sodemann, H., Spackman, J. R., Stark, H., Thomson, D. S., Thornberry, T., Veres, P., Watts, L. A., Warneke, C., and Wollny, A. G.: Characteristics, sources, and transport of aerosols measured in spring 2008 during the aerosol, radiation, and cloud processes affecting Arctic Climate (ARCPAC) Project, *Atmos. Chem. Phys.*, 11, 2423–2453, doi:10.5194/acp-11-2423-2011, 2011.
- Bundke, U., Berg, M., Houben, N., Ibrahim, A., Fiebig, M., Tettich, F., Klaus, C., Franke, H., and Petzold, A.: The IAGOS-CORE aerosol package: instrument design, operation and performance for continuous measurement aboard in-service aircraft, *Tellus B*, 67, doi:10.3402/tellusb.v67.28339, 2015.
- Cai, Y., Montague, D. C., Mooiweer-Bryan, W., and Deshler, T.: Performance characteristics of the ultra high sensitivity aerosol spectrometer for particles between 55 and 800 nm: Laboratory and field studies, *Journal of Aerosol Science*, 39, 759–769, doi:10.1016/j.jaerosci.2008.04.007, 2008.
- Cai, Y., Snider, J. R., and Wechsler, P.: Calibration of the passive cavity aerosol spectrometer probe for airborne determination of the size distribution, *Atmos. Meas. Tech.*, 6, 2349–2358, doi:10.5194/amt-6-2349-2013, 2013.
- Cerni, T. A.: Determination of the Size and Concentration of Cloud Drops with an FSSP, *Journal of Climate and Applied Meteorology*, 22, 1346–1355, doi:10.1175/1520-0450(1983)022<1346:DOTSAC>2.0.CO;2, 1983.

- Cooper, W. A.: Effects of Coincidence on Measurements with a Forward Scattering Spectrometer Probe, *Journal of Atmospheric and Oceanic Technology*, 5, 823–832, doi:10.1175/1520-0426(1988)005<0823:EOCOMW>2.0.CO;2, 1988.
- Fiebig, M., Stein, C., Schröder, F., Feldpausch, P., and Petzold, A.: Inversion of data containing information on the aerosol particle size distribution using multiple instruments, *Journal of Aerosol Science*, 36, 1353–1372, doi:10.1016/j.jaerosci.2005.01.004, 2005.
- 5 Foreman-Mackey, D., Hogg, D. W., Lang, D., and Goodman, J.: emcee: The MCMC Hammer, *Publications of the Astronomical Society of the Pacific*, 125, 306–312, doi:10.1086/670067, arXiv: 1202.3665, 2013.
- Foreman-Mackey, D., Vousden, W., Price-Whelan, A., Pitkin, M., Zabalza, V., Ryan, G., Emily, Smith, M., Ashton, G., Cruz, K., Kerzendorf, W., Caswell, T. A., Hoyer, S., Barbary, K., Czekala, I., Hogg, D. W., and Brewer, B. J.: corner.py: corner.py v1.0.2, doi:10.5281/zenodo.45906, 2016.
- 10 Formenti, P., Schütz, L., Balkanski, Y., Desboeufs, K., Ebert, M., Kandler, K., Petzold, A., Scheuven, D., Weinbruch, S., and Zhang, D.: Recent progress in understanding physical and chemical properties of African and Asian mineral dust, *Atmos. Chem. Phys.*, 11, 8231–8256, doi:10.5194/acp-11-8231-2011, 2011.
- Goodman, J. and Weare, J.: Ensemble samplers with affine invariance, *Communications in Applied Mathematics and Computational Science*, 5, 65–80, doi:10.2140/camcos.2010.5.65, 2010.
- 15 Heim, M., Mullins, B. J., Umhauer, H., and Kasper, G.: Performance evaluation of three optical particle counters with an efficient “multi-modal” calibration method, *Journal of Aerosol Science*, 39, 1019–1031, doi:10.1016/j.jaerosci.2008.07.006, 2008.
- Hermann, M., Weigelt, A., Assmann, D., Pfeifer, S., Müller, T., Conrath, T., Voigtländer, J., Heintzenberg, J., Wiedensohler, A., Martinsson, B. G., Deshler, T., Brenninkmeijer, C. A. M., and Zahn, A.: An optical particle size spectrometer for aircraft-borne measurements in IAGOS-CARIBIC, *Atmos. Meas. Tech.*, 9, 2179–2194, doi:10.5194/amt-9-2179-2016, 2016.
- 20 Jaynes, E. T.: Prior Probabilities, *IEEE Transactions on Systems Science and Cybernetics*, 4, 227–241, doi:10.1109/TSSC.1968.300117, 1968.
- Kandler, K., Lieke, K., Benker, N., Emmel, C., Küpper, M., Müller-Ebert, D., Ebert, M., Scheuven, D., Schladitz, A., Schütz, L., and Weinbruch, S.: Electron microscopy of particles collected at Praia, Cape Verde, during the Saharan Mineral Dust Experiment: particle chemistry, shape, mixing state and complex refractive index, *Tellus B*, 63, 475–496, doi:10.1111/j.1600-0889.2011.00550.x, 2011.
- 25 Kandlikar, M. and Ramachandran, G.: INVERSE METHODS FOR ANALYSING AEROSOL SPECTROMETER MEASUREMENTS: A CRITICAL REVIEW, *Journal of Aerosol Science*, 30, 413–437, doi:10.1016/S0021-8502(98)00066-4, 1999.
- Lance, S., Brock, C. A., Rogers, D., and Gordon, J. A.: Water droplet calibration of the Cloud Droplet Probe (CDP) and in-flight performance in liquid, ice and mixed-phase clouds during ARCPAC, *Atmos. Meas. Tech.*, 3, 1683–1706, doi:10.5194/amt-3-1683-2010, 2010.
- Liu, P. S. K., Leaitch, W. R., Strapp, J. W., and Wasey, M. A.: Response of Particle Measuring Systems Airborne ASASP and PCASP to NaCl and Latex Particles, *Aerosol Science and Technology*, 16, 83–95, doi:10.1080/02786829208959539, 1992.
- 30 Mahowald, N., Albani, S., Kok, J. F., Engelstaeder, S., Scanza, R., Ward, D. S., and Flanner, M. G.: The size distribution of desert dust aerosols and its impact on the Earth system, *Aeolian Research*, 15, 53–71, doi:10.1016/j.aeolia.2013.09.002, 2014.
- Mie, G.: Beiträge zur Optik trüber Medien, speziell kolloidaler Metallösungen, *Annalen der Physik*, 330, 377–445, doi:10.1002/andp.19083300302, 1908.
- 35 Müller, D., Lee, K.-H., Gasteiger, J., Tesche, M., Weinzierl, B., Kandler, K., Müller, T., Toledano, C., Otto, S., Althausen, D., and Ansmann, A.: Comparison of optical and microphysical properties of pure Saharan mineral dust observed with AERONET Sun photometer, Raman lidar, and in situ instruments during SAMUM 2006, *Journal of Geophysical Research: Atmospheres*, 117, D07 211, doi:10.1029/2011JD016825, 2012.

- Osborne, S. R., Johnson, B. T., Haywood, J. M., Baran, A. J., Harrison, M. a. J., and McConnell, C. L.: Physical and optical properties of mineral dust aerosol during the Dust and Biomass-burning Experiment, *Journal of Geophysical Research: Atmospheres*, 113, D00C03, doi:10.1029/2007JD009551, 2008.
- Pinnick, R. G., Garvey, D. M., and Duncan, L. D.: Calibration of Knollenberg FSSP Light-Scattering Counters for Measurement of Cloud Droplets, *Journal of Applied Meteorology*, 20, 1049–1057, doi:10.1175/1520-0450(1981)020<1049:COKFLS>2.0.CO;2, 1981.
- Purcell, E. M. and Pennypacker, C. R.: Scattering and Absorption of Light by Nonspherical Dielectric Grains, *The Astrophysical Journal*, 186, 705, doi:10.1086/152538, 1973.
- Reid, J. S., Jonsson, H. H., Maring, H. B., Smirnov, A., Savoie, D. L., Cliff, S. S., Reid, E. A., Livingston, J. M., Meier, M. M., Dubovik, O., and Tsay, S.-C.: Comparison of size and morphological measurements of coarse mode dust particles from Africa, *Journal of Geophysical Research: Atmospheres*, 108, 8593, doi:10.1029/2002JD002485, 2003.
- Reischl, G. P., Mäkelä, J. M., and Nécid, J.: Performance of Vienna Type Differential Mobility Analyzer at 1.2–20 Nanometer, *Aerosol Science and Technology*, 27, 651–672, doi:10.1080/02786829708965503, 1997.
- Renard, J.-B., Dulac, F., Berthet, G., Lurton, T., Vignelles, D., Jégou, F., Tonnelier, T., Jeannot, M., Couté, B., Akiki, R., Verdier, N., Mallet, M., Gensdarmes, F., Charpentier, P., Mesmin, S., Duverger, V., Dupont, J.-C., Elias, T., Crenn, V., Sciare, J., Zieger, P., Salter, M., Roberts, T., Giacomoni, J., Gobbi, M., Hamonou, E., Olafsson, H., Dagsson-Waldhauserova, P., Camy-Peyret, C., Mazel, C., Décamps, T., Piringer, M., Surcin, J., and Daugeron, D.: LOAC: a small aerosol optical counter/sizer for ground-based and balloon measurements of the size distribution and nature of atmospheric particles – Part 1: Principle of measurements and instrument evaluation, *Atmos. Meas. Tech.*, 9, 1721–1742, doi:10.5194/amt-9-1721-2016, 2016.
- Rosenberg, P. D., Dean, A. R., Williams, P. I., Dorsey, J. R., Minikin, A., Pickering, M. A., and Petzold, A.: Particle sizing calibration with refractive index correction for light scattering optical particle counters and impacts upon PCASP and CDP data collected during the Fennec campaign, *Atmos. Meas. Tech.*, 5, 1147–1163, doi:10.5194/amt-5-1147-2012, 2012.
- Ryder, C. L., Highwood, E. J., Rosenberg, P. D., Trembath, J., Brooke, J. K., Bart, M., Dean, A., Crosier, J., Dorsey, J., Brindley, H., Banks, J., Marsham, J. H., McQuaid, J. B., Sodemann, H., and Washington, R.: Optical properties of Saharan dust aerosol and contribution from the coarse mode as measured during the Fennec 2011 aircraft campaign, *Atmos. Chem. Phys.*, 13, 303–325, doi:10.5194/acp-13-303-2013, 2013.
- Schnaiter, M., Büttner, S., Möhler, O., Skrotzki, J., Vragel, M., and Wagner, R.: Influence of particle size and shape on the backscattering linear depolarisation ratio of small ice crystals – cloud chamber measurements in the context of contrail and cirrus microphysics, *Atmos. Chem. Phys.*, 12, 10465–10484, doi:10.5194/acp-12-10465-2012, 2012.
- Spiegel, J. K., Zieger, P., Bukowiecki, N., Hammer, E., Weingartner, E., and Eugster, W.: Evaluating the capabilities and uncertainties of droplet measurements for the fog droplet spectrometer (FM-100), *Atmos. Meas. Tech.*, 5, 2237–2260, doi:10.5194/amt-5-2237-2012, 2012.
- Stolzenburg, M. R.: An Ultrafine Aerosol Size Distribution System, Ph.D. thesis, University of Minnesota, 1988.
- Strapp, J. W., Leaitch, W. R., and Liu, P. S. K.: Hydrated and Dried Aerosol-Size-Distribution Measurements from the Particle Measuring Systems FSSP-300 Probe and the Deiced PCASP-100X Probe, *Journal of Atmospheric and Oceanic Technology*, 9, 548–555, doi:10.1175/1520-0426(1992)009<0548:HADASD>2.0.CO;2, 1992.
- Sultanova, N., Kasarova, S., and Nikolov, I.: Dispersion Properties of Optical Polymers, *Acta Physica Polonica A*, 116, 585–587, doi:10.12693/APhysPolA.116.585, 2009.

- Szymanski, W. W. and Liu, B. Y. H.: On the Sizing Accuracy of Laser Optical Particle Counters, *Particle & Particle Systems Characterization*, 3, 1–7, doi:10.1002/ppsc.19860030102, 1986.
- Szymanski, W. W., Nagy, A., and Czitrovsky, A.: Optical particle spectrometry—Problems and prospects, *Journal of Quantitative Spectroscopy and Radiative Transfer*, 110, 918–929, doi:10.1016/j.jqsrt.2009.02.024, 2009.
- 5 Voigt, C., Schumann, U., Minikin, A., Abdelmonem, A., Afchine, A., Borrmann, S., Boettcher, M., Buchholz, B., Bugliaro, L., Costa, A., Curtius, J., Dollner, M., Dörnbrack, A., Dreiling, V., Ebert, V., Ehrlich, A., Fix, A., Forster, L., Frank, F., Fütterer, D., Giez, A., Graf, K., Groß, J.-U., Groß, S., Heimerl, K., Heinold, B., Hüneke, T., Järvinen, E., Jurkat, T., Kaufmann, S., Kenntner, M., Klingebiel, M., Klimach, T., Kohl, R., Krämer, M., Candra Krisna, T., Luebke, A., Mayer, B., Mertes, S., Molleker, S., Petzold, A., Pfeilsticker, K., Port, M., Rapp, M., Reutter, P., Rolf, C., Rose, D., Sauer, D., Schäfler, A., Schlage, R., Schnaiter, M., Schneider, J., Spelten, N., Spichtinger, P., Stock, P., Walser, A., Weigel, R., Weinzierl, B., Wendisch, M., Werner, F., Wernli, H., Wirth, M., Zahn, A., Ziereis, H., and Zöger, M.:
 10 ML-CIRRUS - The airborne experiment on natural cirrus and contrail cirrus with the high-altitude long-range research aircraft HALO, *Bulletin of the American Meteorological Society*, doi:10.1175/BAMS-D-15-00213.1, 2017.
- Waterman, P.: Matrix formulation of electromagnetic scattering, *Proceedings of the IEEE*, 53, 805–812, doi:10.1109/PROC.1965.4058, 1965.
- Weinzierl, B., Petzold, A., Esselborn, M., Wirth, M., Rasp, K., Kandler, K., Schütz, L., Koepke, P., and Fiebig, M.: Airborne measurements of dust layer properties, particle size distribution and mixing state of Saharan dust during SAMUM 2006, *Tellus B*, 61, 96–117,
 15 doi:10.1111/j.1600-0889.2008.00392.x, 2009.
- Weinzierl, B., Ansmann, A., Prospero, J. M., Althausen, D., Benker, N., Chouza, F., Dollner, M., Farrell, D., Fomba, W. K., Freudenthaler, V., Gasteiger, J., Groß, S., Haarig, M., Heinold, B., Kandler, K., Kristensen, T. B., Mayol-Bracero, O. L., Müller, T., Reitebuch, O., Sauer, D., Schäfler, A., Schepanski, K., Spanu, A., Tegen, I., Toledano, C., and Walser, A.: The Saharan Aerosol Long-Range Transport
 20 and Aerosol–Cloud–Interaction Experiment: Overview and Selected Highlights, *Bulletin of the American Meteorological Society*, 98, 1427–1451, doi:10.1175/BAMS-D-15-00142.1, 2017.
- Wendisch, M., Keil, A., and Korolev, A. V.: FSSP Characterization with Monodisperse Water Droplets, *Journal of Atmospheric and Oceanic Technology*, 13, 1152–1165, doi:10.1175/1520-0426(1996)013<1152:FCWMWD>2.0.CO;2, 1996.




Figure 1. Sources of uncertainty for size distributions derived from OPC measurements. The main purpose of this study is to introduce an advanced description of OPC response and to offer improved estimates for the corresponding uncertainties (red-rimmed box).




Figure 2. An example subset of kernel functions for the SkyOPC describing the probabilities for particle diameters and corresponding scattering cross-sections to be sorted into the predefined OPC scattering signal amplitude histogram bins, visualized by the different colors. The theoretical relationship between particle diameter and scattering cross-section for PSL is represented by the black curve. The upper graph (a) shows an ideal case without instrumental broadening of size spectra, whereas the lower graph (b) shows a more realistic case where the effect of spectral-signal broadening is considered. The broadening is parametrized by a constant relative Gaussian uncertainty on the scattering cross-section bin threshold values.




Figure 3. Flow chart demonstrating a possible pathway for the retrieval of size distribution information from OPC histogram data within the new framework.




Figure 4. Comparison of modeled relative histograms (colored) and measured counterparts (gray, hatched) for the SkyOPC and different PSL particle standards (rows) for different approaches of OPC kernel function parametrization (columns). The colored histogram bars represent each model's best estimate, the error bars the range between the 16 and 84th percentiles of the corresponding PDFs. The leftmost column shows the theoretical instrument response according to the manufacturer provided set of nominal diameter threshold values in gold (MFR), the middle column the results following the calibration and instrument parametrization approach by Rosenberg et al. (2012) in red brown (R12) and the rightmost column the results of the new approach in blue.




Figure 5. Comparison of modeled relative histograms (colored) and measured counterparts (gray, dotted) for the PCASP and different PSL particle standards (rows) for different approaches of OPC kernel function parametrization (columns), by analogy with Fig. 4. Instead of the PCASP (low gain stage) default binning a custom high resolution linear partitioning is applied here to better highlight the differences between the approaches.




Figure 6. All modeled relative SkyOPC bin counts for the PSL standards scatter-plotted versus their measured counterparts for the three different approaches (rows). The comparisons are shown on linear and logarithmic scales on the left and right hand side respectively. The markers represent the model best estimates, the error bars the range between the 16 and 84th percentiles of the corresponding PDFs. The black lines follow the one-to-one relationship. Significant model underestimations, i.e. vanishingly small model values where non-vanishing bin counts are measured, occur in the two upper rows. The number fraction of significantly underestimated values is noted in the upper left corner of the logarithmic scale plots and the corresponding values are shown with triangular markers in the linear scale plots.




Figure 7. Total sum of residuals between measured relative bin counts and the corresponding model best estimates including all PSL calibration measurements for the SkyOPC and PCASP. In addition to the absolute residual values (solid bars), the arrows and percentage numbers demonstrate the relative reduction by changing the approach.




Figure 8. Comparison between modeled and measured SkyOPC (bin 1-15) counting efficiency. The measured mean counting efficiency values are plotted with red diamond markers and their associated 68% confidence intervals with red error bars. The solid lines represent the model best estimates for the different approaches. The shaded areas correspond to the range between the 16 and 84th percentiles.




Figure 9. Parametrized size distribution retrieval results for two DEHS samples with mean diameters of 0.4 (upper row, graphs (a1) to (a3)) and 0.5 μm (lower row, graphs (b1) to (b3)). The (normalized) Gaussian size distributions are shown in graphs (a1) and (b1). For both the R12 and the new approach the size distribution retrieval PDFs are represented by their diameter-wise medians (solid lines) and 2nd, 16, 84 and 98th percentiles (shaded areas). The theoretical (true) size distributions are indicated by the red lines. The corner plots display the solution PDFs for the size distribution parameters, i.e. mean and standard deviation for the R12 approach in graphs (a2) and (b2) and the new approach in graphs (a3) and (b3). The dashed lines in the 1D histograms represent the parameter PDF medians, 16 and 84th percentiles (in μm). The median values and their distances to the percentiles are noted on top of each histogram. The 2D correlation plots show the solution scatter (black points) superposed with color-coded 2D histograms and smoothed Gaussian contours at 0.5, 1., 1.5 and 2 sigma. The true parameter values are again indicated by the red lines and markers.




Figure A1. PCASP (low gain stage) calibration results following the new method. The model parameter solution ensemble is visualized in the same way as in Fig. 9.




Figure A2. PCASP (low gain stage) calibration results for the approach of Rosenberg et al. (2012). The black markers in the left hand plot represent the centers of the scattering signal amplitude histogram modes measured for the PSL standards and the corresponding calculated mean scattering cross-sections. The error bars represent the histogram modes' half widths and scattering cross-section standard deviations respectively. The red brown solid line shows the best fit, the shaded area the range between the 16 and 84th percentiles of the fit function PDFs. To allow for direct comparison with the results of the new method the right hand plot shows the parameter solution ensemble in the same way as in Fig. A1.