

Response to Reviewer #1

We thank Reviewer #1 for her/his very detailed and helpful comments. Please find below the reviewer's comments (black) and our responses (blue) which also indicate the changes made in the manuscript.

1. General comments

The paper is well written and gives a good overview of the re-processing of GOME data together with its in-flight performance during the entire mission. The new level 1 product and new results on the instrument monitoring are definitely worth a long publication.

I find it a very good and practical idea to include a description of the new data format in the appendix, this will give the users a good starting point.

Although the overall quality of the paper content is already high, there are a number of points where the structure and phrasing can be changed to improve the overall readability.

It is not clear, which of the new in-sights and degradation corrections are included in the re-processed data set. An overview (table or graph) of the main level 1 processing steps indicating the changed steps might be an easy remedy for this. Are all the long term monitoring drift and degradation corrections included in the (ir-)radiance data now?

New users of GOME data would be helped by a few brief explanations on the mentioned GOME specifics. Especially if this new data-set is to become a reference, the paper should be as stand-alone as possible.

For the different radiometric steps a few sentences at the beginning of Section 3 would be helpful. There you could explain that the main cause for degradation is the scan mirror, the biggest correction is done based on irradiance monitoring, then the differences between radiance and irradiance degradation are corrected for in the reflectance. Explaining the approach first and then going into the details would improve the readability of this part a lot.

Section 2 of this review lists specific comments and questions about the content and understanding of the paper. These also include the issues summarized in the general comments. Once these points are clarified, I would strongly recommend the paper for publication.

Section 3 contains suggestions where to edit the text to allow for a smoother read, minor spelling and grammar errors and formatting issues.

This review is based on the version amt-2018-118.pdf retrieved on the 4th of June from <https://doi.org/10.5194/amt-2018-118>. The earlier version (amt-2018-118-manuscript-version1.pdf retrieved on the 1st of June) was not considered.

| # | Page | Line | Section | Comment |
|-----|------|------|---------|---|
| SC1 | 1 | 21 | 1 | You could also add S5P, S5, S4. <i>Done.</i> |
| SC2 | 3 | 12 | 2.1 | Is the integration time for the forward scan also 1.5 s? This is not clear from the text. <i>Yes. We made this more clear in the text.</i> |
| SC3 | 4 | 12 | 2.1 | What kind of channel separator? In line 14 a di-chroic is mentioned, is a di-chroic used here too? <i>It is a channel separator prism. We added an explanation.</i> |
| SC4 | 5 | 5 | 2.2 | Please add which parameters are calibrated in-flight and which are based on on-ground data only. <i>Done. In the list we indicate the in-flight calibration parameters with an asterisk.</i> |
| SC5 | 5 | 7-17 | 2.2 | It would be very useful to have a flowchart showing the actual order of |

| | | | | |
|------|---|--------|-------|--|
| | | | | <p>the steps, also indicating which steps have been changed/improved with the new version. Maybe even show which step is based on in-flight data. The word “steps” in line 7 suggests that the list shows the order of the steps. Consider writing “the basic calibration algorithms are:” instead.</p> <p>We added two more figures (Figs. 2 and 3) indicating the processing flow for calculating the in-flight calibration parameters and the science data, respectively.</p> <p>The description of the algorithms using the on-ground calibration has been moved into a separate section (new Sec. 3).</p> |
| SC6 | 6 | 8/9/10 | 2.2 | <p>Are these calibration constants then fed back immediately into the L1 processor?</p> <p>Yes, they are fed back immediately. We added this information to the text.</p> |
| SC7 | 6 | 9 | 2.2 | <p>The use of the word “comprise” is confusing here. It suggests that the list is complete, that only data from dark measurements, the PtNeCr lamp and the LED is used to directly derive calibration constants during L1 processing. Is this indeed correct? I would also expect that for example transient filtering is performed for all data during L1 processing.</p> <p>Yes, this list is complete. Transient filtering is not performed.</p> <p>No changes were made in the text.</p> |
| SC8 | 6 | 30 | 2.3.1 | <p>What are the other sources? What is “slow” noise? “Slow” with respect to what?</p> <p>“Additional” is not needed here. In this case “slow” means “the noise varies slowly with time and not from readout to readout”. We reformulated the sentence.</p> |
| SC9 | 6 | 32 | 2.3.1 | <p>“from one typical orbit”: Isn’t this section about on-ground calibration data? Does this scaling factor ever change?</p> <p>This scaling factor has been obtained from one orbit during the commissioning phase. It has not changed since then.</p> |
| SC10 | 6 | 33 | 2.3.1 | <p>So only band 1a earthshine needs this correction as all other bands and modes have a shorter integration time?</p> <p>Yes.</p> |
| SC11 | 7 | 2 | 2.3.1 | <p>Are all Peltier signals involved for all detectors? Or only the Peltier belonging to band 1 a? Please clarify.</p> <p>Only the Peltier output belonging to channel 1 is used. We added the information in the text.</p> |
| SC12 | 7 | 6/7 | 2.3.1 | <p>So the additional residual correction is not part of the L1 processor discussed here? Was it not feasible to include this?</p> <p>The correction is implemented in the L1 processor. We clarified this in the text.</p> |
| SC13 | 7 | 12 | 2.3.2 | <p>“In-flight calibration exercise” : do you actually include in-flight calibration into the correction? The rest of the section does not suggest this. Or do you mean “in-flight correction” ? Please clarify. The use of the word calibration suggests calibration measurements to me, and I cannot see how that could be done in-flight.</p> <p>We replaced “in-flight calibration exercise” with “L1 processing”.</p> |
| SC14 | 7 | 25 | 2.3.2 | <p>“currently” , do you expect more ghosts to become significant?</p> <p>We deleted “currently”.</p> |
| SC15 | 7 | 28 | 2.3.2 | <p>Accuracy: Is this for the combined straylight or only the uniform part? I</p> |

| | | | | |
|------|----|-------|-------|--|
| | | | | <p>guess there must have been a discussion at some stage whether the ghost correction is useful or detrimental. You could consider adding a reference here, if there ever was any research done on this. (This is more a note of personal interest than a comment on your manuscript.)</p> <p>It is for the uniform part. Unfortunately, there is no reference available. It has been discussed pre-flight (1994), but we do not have more detailed information.</p> |
| SC16 | 7 | 10-22 | 2.3.2 | <p>To make this section a bit clearer, I would change the order a bit: Lines 10-12, then lines 20 to mid 22, then lines 12 to 19, then from line 22 on.</p> <p>Done as suggested.</p> |
| SC17 | 8 | 14 | 2.3.3 | <p>Do I understand this correctly: There are two BSDF steps with a different parametrization? Why are they not combined? Is that because the BSDF depends too much on the degradation? Please clarify.</p> <p>Yes, there are two BSDF steps. This has historical reasons. In the previous GDP version the application of a separate extraction software was required in which the second step was performed. The second BSDF step contains an improved azimuth dependence. We made this more clear in the text.</p> |
| SC18 | 8 | 3-21 | 2.3.3 | <p>This part might be a better read if you state what is performed for the solar, the earthshine and the moon measurements. Or is the BSDF correction also applied for earthshine and moon? Also here a figure with the processing steps and paths might help.</p> <p>This section (now Sec. 3.3) has been rewritten to make this more clear.</p> |
| SC19 | 8 | 25-26 | 2.3.4 | <p>I would guess that the polarization sensitivity is from on-ground calibration and that only the characterization has two main parts. This is not clear from the sentence.</p> <p>We split the sentence in order to make this more clear.</p> |
| SC20 | 8 | 31 | 2.3.4 | <p>I assume the interpolation should be followed by a multiplication with the sensitivity? Or where does the on-ground data come back in?</p> <p>We added the information in line 26.</p> |
| SC21 | 9 | 9 | 2.3.4 | <p>Why were the iterations made if they are not needed? Do you mean "for practical reasons these iterations.."? </p> <p>The iterations were made in the course of several (unplanned) reprocessings. We added this in the text.</p> |
| SC22 | 9 | 18 | 2.4 | <p>Can you explain which calibration parameters are included? Alternatively you could add it to the appendix.</p> <p>In the appendix we added the reference to the Product User Manual (Aberle, 2018) that contains the complete list of all parameters.</p> |
| SC23 | 10 | 10 | 3.0 | <p>For the different radiometric steps a few introductory sentences at the beginning of Section 3 would be helpful. Here the approach can be outlined: that the main cause for degradation is the scan mirror, the biggest correction is done based on irradiance monitoring, then the differences between radiance and irradiance degradation are corrected in the reflectance. I found myself wondering why it was done in such a roundabout way and finding the explanation pages later. To explain the approach at the beginning of the section removes this confusion.</p> <p>We added a brief introduction to this section (now Sec. 4).</p> |

| | | | | |
|------|----|--------|-----|--|
| SC24 | 10 | 14 | 3.1 | <p>1) I would make very clear here, that the degradation has been shown to be mainly/only the scan mirror and not the diffuser, otherwise the degradation correction doesn't seem logical.</p> <p>2) Furthermore it would be important to mention that it is a first order correction and differences between radiance and irradiance are corrected in the reflectance.</p> <p>1) We added a sentence and a reference here to make this more clear.</p> <p>2) We mention this in Section 'Reflectance Degradation' (former Sec. 3.3)</p> |
| SC25 | 10 | 26 | 3.1 | <p>Somewhat more explanation is needed here, why does the loss of the gyroscope functionality only affect one channel? What does this functionality do?</p> <p>Our text was a bit misleading in this case. The loss of the gyroscope functionality did affect all channels. We reordered the sentences to make this more clear.</p> |
| SC26 | 11 | Fig. 2 | 3.1 | <p>From 2004 on the entire wavelength range from 450nm seems to be above 1, this is not explained in the text. Where does it come from?</p> <p>Values above 1 might be due deposits on the coatings which can lead to changes in interference patterns and to an increase in intensity (Snel, 2001). We added the explanation in the text.</p> |
| SC27 | 11 | Tab. 1 | 3.1 | <p>Could you also add the values for end-of-life?</p> <p>End-of-life values have been added (and also values for 290nm, see comments Reviewer #2).</p> |
| SC28 | 11 | 4/5 | 3.1 | <p>The same degradation is applied to both irradiance and radiance? Wouldn't that only work when all degradation occurs within the common path and none in the diffuser? Has this been verified? Then it should really be mentioned here. (OK, I now see it's mentioned later in the text. I have added SC 23 and SC24.)</p> <p>Please see responses to SC23 and SC24.</p> |
| SC29 | 12 | 13 | 3.2 | <p>For new users of GOME data it is not clear why a platform pointing problem would only affect one channel. Please add a brief explanation.</p> <p>See response to SC25. We reordered the sentences here, too.</p> |
| SC30 | 13 | 12/13 | 3.2 | <p>So PMD 1 does not decay as bad as channel 2. Is it known why? Is it maybe related to the wavelength dependent difference between s-and p- reflectance of the scan mirror?</p> <p>We assume that this might be related to long-term changes in the mean wavelength of the PMD which is in the order of ~10nm for PMD1.</p> |
| SC31 | 15 | 8 | 3.3 | <p>So there is a contribution of the diffuser after all? Or is the only difference the angle on the scan mirror?</p> <p>The only difference is the angle on the scan mirror.</p> |
| SC32 | 16 | 19-22 | 3.3 | <p>Is there an explanation for the degradation getting better and worse? Have there been studies for other wavelengths too? If yes, do they show the same behaviour?</p> <p>We did not analyze other wavelengths. Degradation getting better and worse might be due to changes in interference patterns (Snel, 2001). We added a sentence and the reference.</p> |
| SC33 | 17 | Fig 6 | 3.3 | <p>It is striking that the two wavelengths appear to behave the same until 2004 and then they start deviating, is there a reason for that?</p> |

| | | | | |
|------|-------|--------|--------|--|
| | | | | We assume that this is related to the scan angle dependence that increased over time (see Snel, 2001). |
| SC34 | 17 | | 4.2 | This section is very well written. |
| SC35 | 19 | 2 | 4.2 | Which are the thermally sensitive optical elements? Does the degradation of thermally sensitive optical elements also cause the changes in the reflectance? We removed this sentence. |
| SC36 | 21 | Fig. 9 | 4.3.1 | The caption and plots' y-axes are not consistent with the unit and what is shown. I think you mean leakage signal or dark signal and not current. To make the plot a bit clearer, you could add the co-addition times in the plot or the caption. Corrected. |
| SC37 | 21 | Fig. 9 | 4.3.1 | For channel 2 in the normal scanning mode: the spread is much larger than for the other channels and modes. Is that explained? Unfortunately, we do not have an explanation for this behavior. |
| SC38 | 21 | 17 | 4.3.1 | Shouldn't it be leakage signal? Yes. Corrected. |
| SC39 | 23 | 3 | 4.3.1 | Is this plot representative for other channels and modes? Yes, a three-fold broadening of the distribution was also found for other bands and modes. |
| SC40 | 25 | 4 | 4.3.2 | I assume the tape recorder failure changed the power conditioning? Or how can it have the shown effect? The noise level changed because of the significantly reduced number of measurements and in particular because measurements from the South Atlantic Anomaly region are missing since the tape recorder failure. |
| SC41 | 26 | 5 | 4.4 | When using a monochromatic LED, the pixel response and quantum efficiency is monitored for the LED's wavelength but not necessarily for the wavelength the pixel is normally detecting. Have there ever been other measurements, for example on-ground with a white light source, to verify the results from the LEDs? Unfortunately, this has not been verified. |
| SC42 | 28/29 | | 5 | It's not entirely clear from the summary (or elsewhere) which insights from the long term monitoring of irradiance, degradation, spectral calibration have been included in the L1 processor. If they are, are they part of the calibration data or are corrections already included in the (ir) radiance? More results of the study were included in the summary in order to make this more clear. |
| SC43 | 29/30 | | App. A | Great idea to include the file format. Thanks. |

3. Technical corrections

3.1. Definitions

Is there a reason to explicitly name the detector brand Reticon? No other brands are named as far as I could see.

No, there is no explicit reason, it is rather a 'leftover'. We replaced it with “array detector”.

Figure 11: “PDF” is not explained.
We added the explanation.

3.2. Formatting of plots

Figure 9: The y-axis says “DC”, which normally is the dark current, but the dark signal is shown.
Corrected.

3.3. Typos

For the following words, the spelling/capitalization is not consistent throughout the article:

- Sun
- Polarization Measurement Device
- The word ‘data’ is used both in the singular and the plural, please pick one of the two
- Please reconsider you capitalization, either capitalize all new abbreviations or none. For example “Focal Plane Assembly (FPA)” but “pixel-to-pixel variations (PPG) in quantum efficiency” on page 5

Spelling/capitalization should be consistent now.

| # | Page | Line | Section | Comment |
|------|------|-------|----------|--|
| TC1 | 1 | 2 | Abstract | Shouldn't it be “ozone and other trace gases”? Changed. |
| TC2 | 1 | 29 | Abstract | Shouldn't it be “polarization correction, and dark current correction” ? Changed. |
| TC3 | 2 | 18 | 1 | Similar changes ...[], whereas they are Corrected. |
| TC4 | 2 3 | 27 | 2.1 | Full stop missing. Inserted. |
| TC5 | 4 | 1 | 2.1 | It's “GOME Users” not “User's” Corrected. |
| TC6 | 4 | 16 | 2.1 | To clarify: “that consists for each channel of ... “ Added. |
| TC7 | 4 | 22 | 2.1 | “... and it had a repeat cycle ...” Added. |
| TC8 | 4 | 22/23 | 2.1 | The sentence should also be in the past tense. Changed. |
| TC9 | 4 | 26 | 2.1 | “... additional ground stations had been ..” Changed. |
| TC10 | 9 | 12 | 2.4 | Was thus, not thus was. Corrected. |
| TC11 | 9 | 15 | 2.4 | It has turned out ... Corrected. |
| TC12 | 9 | 17 | 2.4 | Contains Corrected. |
| TC13 | 10 | 8 | 3.1 | Word order: the latter serve themselves... Changed. |
| TC14 | 10 | 10 | 3.1 | 3rd of July |

| | | | | |
|------|----|----|-------|--|
| | | | | Corrected. |
| TC15 | 11 | 7 | 3.1 | 3rd of July Corrected. |
| TC16 | 13 | 10 | 3.1 | 3rd of July Corrected. |
| TC17 | 18 | 5 | 4.2 | ...lines ... have Corrected. |
| TC18 | 19 | 11 | 4.2 | The use of “however” is a bit confusing here, I would first state that they didn’t find a dependence on longitude and then “However they found the maxima...” Reformulated. |
| TC19 | 19 | 13 | 4.2 | Temperature rise not raise Corrected. |
| TC20 | 22 | 12 | 4.3.1 | The second “which” is not needed. Changed. |
| TC21 | 22 | 17 | 4.3.1 | The second comma is not needed. Changed. |
| TC22 | 23 | 4 | 4.3.1 | ..pixel the standard ... Changed. |
| TC23 | 23 | 12 | 4.3.1 | Rephrase to “The most significant decrease in the number of available measurements is for the LED dark signal calibration measurements.” Changed. |
| TC24 | 25 | 3 | 4.3.2 | Don’t you mean the following section ? Corrected. |
| TC25 | 27 | 8 | 4.4 | The date format is different than before. Changed. |

3.4. References

| Page | Line | Section | Comment |
|------|-------|--------------|--|
| 33 | 10/11 | Bibliography | Link seems to be faulty. Corrected. |
| 33 | 12/13 | Bibliography | Link seems to be faulty. We double-checked the link and it should be valid. |

3.5. Author contributions

The authors’ contributions are not listed separately, is this intentional?

[This section is optional and we would like to leave it out.](#)

Response to Reviewer #2

We thank Reviewer #2 for her/his very detailed and helpful comments. Please find below the reviewer's comments (black) and our responses (blue) which also indicate the changes made in the manuscript.

1) Scientific Significance: The manuscript provides a good description of the methods and results of a study to evaluate and improve the stability of the GOME Level 1 record.

2) Scientific Quality: The results are well-structured and well-referenced and use good statistical analysis methods. There are good references to detailed reports for interested readers.

3) Presentation Quality: The paper is well-written and the figures and tables are good in both content and structure.

Editorial Comments and Suggestions:

Make Figure 2 larger. It should at least be full page width.

Done as suggested.

In Section 2.3.2, rewrite and clarify the last line. Is this 10% the accuracy of the stray light estimates relative to the true stray light? That is, is if the stray light error is 20 units, then the correction will be between 18 and 22 units and the final result will have an error of ± 2 units?

We added: "... not more accurate than $\sim 10\%$, i.e. processing errors of 10% of true straylight."

Page 19, Line 14, "raise" should be "rise".

Corrected.

The value of 1100 for the SNR for Channel 1 in the Table 1 seems high even for the 305 nm wavelength. What is the corresponding integration time and the size of the FOV? I believe there was a change in the Channel 1A/1B wavelength boundary during the mission. Is this before or after that change? Also, provide an SNR value for a shorter wavelength in the table, say 290 nm.

We added to Table 1:

- the information on the integration time for channel 1 (6s);
- values for 290nm;
- values for middle (2001) and end (2010) of the mission.

The change in the Channel 1A/1B wavelength boundary was implemented in June 1998.

While the views of the Moon are complicated by scan mirror differences with angle and the phases of the Moon, more accurate lunar models are now available. For example, Eumetsat's GSICS Implementation of the ROLO model (GIRO) and the GSICS Lunar Observation Dataset (GLOD) introduced at https://www.eumetsat.int/website/home/News/DAT_3460357.html?lang=EN&pState=1 could be explored to allow the lunar measurements to be used to monitor instrument changes.

Thank you very much for the reference to these data sets! For the GOME instrument one problem is that the Moon measurements do not fill the entire slit so that the calibration key data cannot be applied just like that. Moreover, the Moon is always observed at a scan angle in the western Limb, whereas scan angle dependency observations corresponding to East or Nadir pixels would be needed to significantly improve on the degradation correction using the Sun. And since the Moon merely reflects Sun light, it cannot be used as calibration source independent of solar activity.

Questions on Science:

Section 3.1

Page 11 line 8 et seq. While arguments can be made for estimating degradation by avoiding lines with high solar activity, this will not work well for Channel 1. See

V. Marchenko, Sergey & Deland, Matthew & Lean, Judith. (2016). Solar Spectral Irradiance Variability in Cycle 24: Observations and Models. Journal of Space Weather and Space Climate. 6. 10.1051/swsc/2016036.

for estimates of solar variability for 270 nm to 500 nm over a solar cycle. After estimating the changes in the instrument throughput, the final time-dependent solar provided in Level 1 should be constructed with realistic solar activity variations.

We agree with the reviewer that in principle realistic solar variations should be taken into account. Thank you for pointing us to this reference. Nonetheless, providing an optimum solar irradiance product is not the main focus of this study.

Also, how large are the Etalon Effects in Figure 2? What errors would they be expected to produce in the radiance/irradiance ratios? Why wasn't a correction applied? It appears that the authors have access to estimates of these corrections from other analysis:

https://wdc.dlr.de/sensors/gome/degradation_files/degradation.php

The etalon amplitudes may be estimated from Figure 4 (former Fig. 2) as the amplitude of the semi-regular wiggles (~10 wiggles in channel 1 to ~5 in channel 4). Since GOME does not have a flat field mode (e.g. using a white light source) etalon cannot be directly derived. Correction using the Sun would be possible but only relative to a certain reference date, not in an absolute sense. The main focus of the Level 1 product has been to function as input for Level 2 retrievals. For Level 2, etalon is irrelevant as long as the structure is identical for solar and for earth-shine measurements. There are in fact some indications that this may not completely be the case, depending on the solar azimuth, but attempts to characterize solar-azimuth dependent etalon-like structures of the diffuser BSDF have not been deemed reliable enough to be applied in the GOME calibration. The errors are shown in Fig.8 of reference (Slijkhuis 2004) on

https://wdc.dlr.de/sensors/gome/degradation_files/degradation.php

And there are the earlier results in

Weber, Mark & Burrows, John & Cebula, R. (1998). GOME Solar UV/VIS Irradiance Measurements between 1995 and 1997 – First Results on Proxy Solar Activity Studies. Solar Physics. 177. 63-77. 10.1023/A:1005030909779.

Section 3.3

From Section 2.3.2, the angle for the mirror for Solar measurements is 41° and those for the Earth measurements range from 49° ±15°. What are the results for Figure 6 for the ground pixels at this matching angle? If they are not equal to 1.0 what are the likely instrument changes that produce time-dependent differences in the radiance/irradiance ratios?

In Fig. 8 (former Fig. 6) the results for the west pixels (incidence angle 44°-34°), cyan curves, match the angle for the solar measurements. In general, west pixels show the minimal degradation in reflectance compared to the other ground pixels types. Explicit characterization and indication of the instrument changes that produce the reflectance degradation is difficult. As mentioned on page 10: "The main degradation as a consequence of extensive exposure to the space environment can be attributed to deposits on the scan mirror (which is coated with a MgF 2 layer) thereby changing its reflective properties". This change in mirror coating also changes the scan-angle dependent polarization properties of the instrument (Snel, 2001) .

Is it correct that the analysis in the section is just an evaluation of errors in the Level 1 product and that no

corrections based on the PICS results have been applied?

Yes, this is correct. We added this to the summary.

If so, degradation is only shown for 325 nm and 335 nm measurements and the changes are over 20% and differ by over 5%. This does not suggest that the shorter channel are well characterized for absolute radiance / irradiance calibration. All algorithms are sensitive to the reflectance if they need parameters associated with cloud cover. What are the effects of a +10% error in the UV cloud fraction on GODFIT ozone retrievals?

A 10% error in cloud fraction is expected to have an impact of <1% on total column ozone retrieval (Christophe Lerot, personal communication, July 2018).

Do the authors recommend that Channel 1 data in this product be used for ozone profile retrievals? What about the use of data from 300-310 for tropospheric retrievals requiring radiance / irradiance calibration?

We agree with the reviewer that the retrieval of ozone profiles and tropospheric columns from GOME requires a very careful handling of the measured spectra and additional corrections to account for degradation. However, several studies successfully demonstrated the feasibility (Liu et al., 2005, Cai et al., 2012, Miles et al., 2015). Moreover, Keppens et al. (2018) have shown that decadal drift values for GOME level-2 ozone profiles are overall insignificant.

Section 4.2

How large are the variations in the wavelength scales along an orbit from measurement-based estimates? Do they match with the variations predicted from the effects of the measured pre-disperser prism temperature changes combined with the laboratory sensitivity characterization or are there other complicating factors?

On average 9 different wavelength scales are used along one orbit. The variation in the wavelengths depends on the spectral region. In general, the variation is <0.002 nm, except for the beginning of channel 3 and the end of channel 4, where the variation is 0.004-0.005 nm along one orbit. This analysis is based on ~2000 randomly selected orbits.

The use of spectral calibration as function of pre-disperser temperature was a recommendation based on on-ground measurements of instrument performance. However, we cannot retrieve the original data.

DOAS- based retrievals often generate internal estimates of the wavelength scale shifts as part of the fitting process. Have any of these been compare to this bottom-up analysis based on the prism temperatures?

No, unfortunately, these comparisons have not been performed.

References:

Cai, Z., Y. Liu, X. Liu, K. Chance, C. R. Nowlan, R. Lang, R. Munro, and R. Suleiman (2012), Characterization and correction of Global Ozone Monitoring Experiment 2 ultraviolet measurements and application to ozone profile retrievals, *J. Geophys. Res.*, 117, D07305, doi: 10.1029/2011JD017096.

Keppens, A., Lambert, J.-C., Granville, J., Hubert, D., Verhoelst, T., Compernelle, S., Latter, B., Kerridge, B., Siddans, R., Boynard, A., Hadji-Lazaro, J., Clerbaux, C., Wespes, C., Hurtmans, D. R., Coheur, P.-F., van Peet, J. C. A., van der A, R. J., Garane, K., Koukouli, M. E., Balis, D. S., Delcloo, A., Kivi, R., Stübi, R., Godin-Beekmann, S., Van Roozendaal, M., and Zehner, C.: Quality assessment of the Ozone_cci Climate Research Data Package (release 2017) – Part 2: Ground-based validation of nadir ozone profile data products, *Atmos. Meas. Tech.*, 11, 3769-3800, <https://doi.org/10.5194/amt-11-3769-2018>, 2018.

Liu, X., K. Chance, C. E. Sioris, R. J. D. Spurr, T. P. Kurosu, R. V. Martin, and M. J. Newchurch (2005),

Ozone profile and tropospheric ozone retrievals from the Global Ozone Monitoring Experiment: Algorithm description and validation, *J. Geophys. Res.*, 110, D20307, doi: 10.1029/2005JD006240.

Miles, G. M., Siddans, R., Kerridge, B. J., Latter, B. G., and Richards, N. A. D.: Tropospheric ozone and ozone profiles retrieved from GOME-2 and their validation, *Atmos. Meas. Tech.*, 8, 385-398, <https://doi.org/10.5194/amt-8-385-2015>, 2015.

The Global Ozone Monitoring Experiment: Review of in-flight performance and new reprocessed 1995–2011 level 1 product

Melanie Coldewey-Egbers¹, Sander Slijkhuis¹, Bernd Aberle¹, Diego Loyola¹, and Angelika Dehn²

¹German Aerospace Center (DLR), Remote Sensing Technology Institute, Oberpfaffenhofen, Germany

²European Space Agency (ESA), ESRIN, Frascati, Italy

Correspondence: Melanie Coldewey-Egbers (Melanie.Coldewey-Egbers@dlr.de)

Abstract. The Global Ozone Monitoring Experiment (GOME) on-board the second European Remote Sensing satellite provided ~~global~~ measurements of atmospheric constituents such as ozone ~~or~~ and other trace gases for the 16 year period from 1995 to 2011. In this paper we present a detailed analysis of the long-term performance of the sensor and introduce the new homogenized and fully calibrated level 1 product which has been generated using the recently developed GOME Data Processor level-0-to-1b (GDP-L1) Version ~~5.05.1~~ 5.05.1. By means of the various in-flight calibration parameters we monitor the behavior and stability of the instrument during the entire mission. Severe degradation of the optical components has led to a significant decrease in intensity in particular in channels 1 and 2 covering the spectral ranges of 240–316 nm and 311–405 nm, respectively. Thus, a soft correction based on using the ~~sun~~ Sun as a stable calibration source is applied. Revision and optimization of other calibration algorithms such as the wavelength assignment, polarization correction, ~~or~~ and dark current correction resulted in an improved and homogeneous level 1 product that can be regarded as the European satellite reference data for successor atmospheric composition sensors and that provides an excellent prerequisite for further exploitation of GOME measurements.

1 Introduction

The Global Ozone Monitoring Experiment (GOME) was launched on 21 April 1995 by the European Space Agency (ESA) on-board the second European Remote Sensing satellite (ERS-2). It was the first European UV-VIS-NIR (ultraviolet–visible–near-infrared) spectrometer in space dedicated to observe atmospheric trace constituents such as ozone, ~~or~~ nitrogen dioxide, sulfur dioxide, formaldehyde, bromine, water vapor as well as cloud and aerosol parameters on a global scale (Burrows et al., 1999). The sensor operated for more than 16 years, which is a world record for this kind of instruments, until the retirement of the ERS-2 platform in early July 2011. GOME is the predecessor of a series of similar follow-up instruments like SCIAMACHY (Scanning Imaging Absorption Spectrometer for Atmospheric Chartography, 2002–2012, Bovensmann et al., 1999) on-board Envisat, OMI (Ozone Monitoring Instrument, launched in 2004 on-board Aura, Levelt et al., 2006), ~~or~~ GOME-2 (Munro et al., 2016) on-board the MetOp satellite series, and or the Copernicus Sentinel missions TROPOMI on-board the Sentinel-5 Precursor, Sentinel-4, and Sentinel-5, and marks the beginning of European operational, global, long-term monitoring of climate-relevant atmospheric parameters.

The existing atmospheric data archive of GOME is of very high value and may be considered (in conjunction with SCIAMACHY) as the basis for a future reference data set for successor sensors. The status of scientific results of the miscellaneous GOME level 2 data products is presented in numerous publications (e.g., Balis et al., 2007; de Smedt et al., 2008; Loyola et al., 2010; van Roozendael et al., 2012; Lerot et al., 2014). Furthermore, GOME data ~~form~~forms a substantial part of recently developed long-term climate data records, for example the GOME-type Total Ozone Essential Climate Variable and the harmonized tropical tropospheric ozone data records generated within the framework of ESA's outstanding Climate Change Initiative (Coldewey-Egbers et al., 2015; Heue et al., 2016).

In 2012 at ESA's Atmospheric Science Conference, as a result of the discussion rounds, the scientific user community formulated a set of recommendations (ATMOS, 2012) also addressing the preservation and further exploitation of the 16 years of GOME measurements. These recommendations have led to ESA's GOME-Evolution project that started in April 2014. Among other topics, the objective of this activity is to provide the Earth Observation (EO) user community with improved and consolidated GOME level 1 products, in an easily accessible common data format, based on updated GOME calibration algorithms and improved in-flight calibration characterization for the complete mission. Homogenization of the current GOME level 1 products has become necessary because so far they were generated using different processor versions and, thus, were not fully consistent during the complete mission.

Furthermore, a detailed investigation of the long-term performance of the GOME instrument for the entire mission period was carried out in the framework of GOME-Evolution. The results will be presented in this paper. This part of the study is an extension to the work by Coldewey-Egbers et al. (2008) who introduced a first overview of the long-term behavior for the 11 year time span from 1995 to 2006. Special emphasis is put on the analysis of the Sun mean reference spectra in order to monitor and correct for the gradual degradation of the instrument's optical properties. Coldewey-Egbers et al. (2008) have shown that the degradation is severe in particular in the UV channels 1 (~70-90%) and 2 (~35-65%) covering the spectral range of 240–316 nm and 311–405 nm, respectively. Similar changes are observed for SCIAMACHY and GOME-2 (Noël et al., 2007; Bramstedt et al., 2009; Munro et al., 2016), whereas ~~it is~~they are considerably smaller for OMI (Schenkeveld et al., 2017).

The various in-flight calibration parameters are a good means to monitor the long-term stability of the sensor and its measurements. Instrument stability is one of the most important prerequisites to meet the challenge of measuring very small changes in atmospheric parameters associated with long-term climate change from space. For example, satellite sensors are required to detect ozone trends in the order of 1% decade⁻¹ (GCOS, 2011). Amongst other things, in our study particular attention was paid to the analysis of the long-term performance and stability of the spectral calibration since errors in wavelength assignment may have a significant impact on the earth albedo and trace gas retrievals (Voors et al., 2006; van Geffen et al., 2015; Pan et al., 2017).

In addition to the new GOME level 1 product and the revised in-flight calibration data set, a "Climate" total column water vapor product ~~has been developed within ESA GOME-Evolution~~(Beirle et al., 2018) has been developed within the ESA GOME-Evolution project. It is based on homogenized GOME, SCIAMACHY, and GOME-2 observations and provides a consistent time series that is dedicated to study the temporal evolution of water vapor over the past two decades on a global

scale. Another part of the project was the creation of a web gallery (GOME Web Gallery) featuring the GOME/ERS-2 mission and related scientific achievements.

The paper is organized as follows: In Section 2 we provide an overview of the GOME instrument design (~~Sect.2.1~~), ~~a brief description and brief descriptions~~ of the level-0-to-1 ~~algorithms and processing chain~~ (~~Sect. processing chain and the new level 1 product. Section 2.2~~), ~~3 contains~~ a summary of the calibration algorithms using results of the on-ground instrument characterization (~~Sect. ??~~), ~~and a short description of the new level 1 product generated in the framework of GOME-Evolution~~ (~~Sect.2.3~~). An analysis of the Sun mean reference spectra and the description of the degradation correction algorithm is presented in Sect. 4.1, followed by the investigation of the Polarization Measurement Device data (Sect. 4.2). In Section 4.3 we show results of the reflectance degradation analysis. Section 5 contains the detailed results of the long-term analysis of the most important GOME in-flight calibration parameters needed for, e.g., the spectral calibration or the dark current correction. Summary and concluding remarks are finally given in Section 6.

2 GOME/ERS-2

2.1 Instrument and platform characteristics

GOME is a nadir-viewing, across-track scanning spectrometer that covers the ultraviolet, visible and near-infrared wavelength range from 240 to 790 nm with moderate spectral resolution of 0.2 to 0.4 nm. It measures the solar radiation reflected and scattered by the Earth's atmosphere and surface as well as the solar irradiance. Its primary objective is the determination of the amounts and distributions of atmospheric trace constituents, such as ozone, nitrogen dioxide, sulfur dioxide, formaldehyde, or bromine oxide as well as cloud and aerosol parameters (Burrows et al., 1999). In normal viewing mode, there are three forward scans (footprint size of $320 \times 40 \text{ km}^2$ each – across-track \times along-track) followed by a backscan with 1.5 s integration time each. The maximum swath width is 960 km, global coverage is achieved at the equator within three days.

GOME is a double monochromator, which has as dispersing elements a predisperser prism, combined with a holographic grating in each of the four optical channels. The earthshine radiance and solar irradiance spectra are recorded with four linear Si-diode arrays with 1024 spectral elements each. These detectors are cooled to 235 K by means of Peltier coolers to reduce dark current and to improve the signal-to-noise ratio. The four channels cover the wavelength regions of 240–316 nm (channel 1), 311–405 nm (channel 2), 405–611 nm (channel 3), and 595–793 nm (channel 4). Channels 1 and 2 are further electronically divided into two bands ('a' and 'b') covering the short-wavelength and long-wavelength parts of the channels, respectively. In addition there are four stray light bands: two short-wave of band 1a, one long-wave of band 1b, and one short-wave of band 2a. Part of the light is branched out at the predisperser prism and recorded with three fast broadband silicon photo-diodes, the ~~polarization measurement devices~~ Polarization Measurement Devices (PMDs), whose spectral ranges cover approximately the optical channels 2 (300–400 nm), 3 (400–580 nm), and 4 (580–750 nm), respectively. They measure the amount of light polarized parallel to the instrument slit, which is perpendicular to the plane of incidence of the scan mirror. The PMDs are non-integrating detectors which are continuously sampled, albeit over an RC-circuit which has an averaging effect over the

sampling time. The corresponding sampling time of the PMD measurements is 93.75 ms, i.e. 16 PMD measurements are available for one detector channel measurement at the default integration time of 1.5 s.

The various pointing geometries of the GOME scan mirror permit in addition to solar and earth nadir viewing, polar viewing (viewing angle of 45°), and lunar observations (viewing angle of about 80°) at selected times during a year. A calibration unit adjacent to the spectrometer part consists of the ~~sun~~-Sun view port and a compartment housing a platinum-neon-chromium (Pt/Ne/Cr) hollow cathode discharge lamp. The solar radiation is attenuated by a 20% transmission mesh and directed via a diffuser plate (wet-sanded aluminium plate with chromium/aluminium coating) onto the entrance slit of the spectrometer. The calibration unit becomes optically coupled to the spectrometer by appropriate positioning of the scan mirror.

A detailed overview of the GOME instrument, its operation, and scientific methods can be found in the GOME ~~User's~~-Users Manual (GOME Users Manual) and in Burrows et al. (1999). For understanding the algorithm principles described in the following sections a simple functional diagram of the GOME instrument is shown in Fig. 1. The most important instrument components relevant to the level-0-to-1 calibration are

- the scan mirror whose position is linked to the observation mode, e.g., nadir or pole scanning, static or moon view, and the calibration mode; the latter comprises solar measurements, dark signal measurements, and spectral lamp measurements;
- the calibration unit that hosts the spectral calibration lamp and the Sun diffuser;
- the slit that limits the instantaneous field-of-view to $2.9^\circ \times 0.142^\circ$ or $40 \times 2 \text{ km}^2$ on the ground; the slit function, i.e., the instrument spectral response to monochromatic input, is a convolution of projected slit width, pixel response, and optical aberrations;
- the quartz predisperser prism where part of the light is branched out and directed towards the PMD unit (see below);
- the channel separator ~~that separates~~ prism in the intermediate focus that acts as a spatial filter to separate the wavelengths for channel 1, for channel 2, and for channels 3 and 4, respectively; this separation serves to reduce stray light on the UV detectors, i.e. channels 1 and 2;
- the dichroic filter that separates the wavelengths of channel 3 from those of channel 4; in the spectral range from 590 nm to 610 nm the filter changes from reflection to transmission;
- the channel optics that consists for each channel of 4 quartz lenses mounted in one barrel;
- four red LEDs which illuminate the detectors directly and which are used to characterize the pixel-to-pixel sensitivity;
- the Focal Plane Assembly (FPA) which holds the ~~Reticon~~-array detector and the pre-amplifier electronics;
- the PMD unit that contains three broadband ~~polarization measurement devices~~ Polarization Measurement Devices whose spectral bandwidths correspond roughly to the detector array channels 2, 3, and 4.




Figure 1. Functional diagram of the GOME instrument (see text for more explanations).

ERS-2 orbited the Earth at an altitude of about 790 km in a Sun-synchronous near polar orbit; the descending node local equator crossing time was about 10:30 UTC, and it had a repeat cycle of 35 days. Each orbit takes-took ~100 min and the spacecraft completes-completed ~14 orbits per day. Operational GOME observations are available since July 1995, although global coverage was lost in June 2003 due to a permanent failure of the ERS-2 on-board tape recorder. Since then availability of GOME data coverage is limited to the region where ERS-2 was in direct contact with ground stations in the European-Atlantic sector. Over the years additional ground stations have-had been brought online to incrementally increase the data gathering abilities of the satellite. The ERS-2 active mission was completed on 4 July 2011 at orbit no. 84719. The ESA mission operations overview (GOME Mission Operations Overview) provides a detailed review of the most important events over the entire mission lifetime which may have had an impact on the GOME data quality. Anomalies such as cooler or instrument switch-offs, spectral lamp failures, or data gaps are reported on a yearly basis.

2.2 GOME Data Processor

The GOME Data Processor (GDP) is the operational off-line ground segment for the GOME instrument, incorporating, among other things, a level 0-to-1 processing chain (GDP-L1) and the complete GOME data archive (Loyola et al., 1997). During the level 0-to-1 processing, GOME data is converted into calibrated physical quantities by applying a series of calibration algorithms. Some of the calibration data were-was obtained during the pre-flight on-ground calibration. Other parameters which can be directly derived from measurements using on-board calibration sources are derived during the level 0-to-1 processing; they are marked with an asterisk (*) in the following list. The basic calibration steps-needed-algorithms are:

- signal correction, i.e. correction for dark signal, FPA crosstalk, pixel-to-pixel variations (*), Focal Plane Assembly (FPA) crosstalk, Pixel-to-Pixel Gain (PPG) (*) in quantum efficiency, and stray light;

- wavelength calibration^(*), i.e. assigning to each detector pixel its associated wavelength;
- radiance calibration, i.e. conversion of the corrected detector signals to radiance units by application of the radiance response function; this step also includes the polarization correction;
- irradiance calibration, i.e. conversion of the corrected detector signals to irradiance units, including the correction for BSDF (Bi-directional Scattering Distribution Function) of the diffuser plate;
- geolocation, i.e. determination of the geographical position for each detector readout using ESA's ERS-2 orbit propagator;
- quality assessment, i.e. identification of dead pixels, hot pixels, saturation, and sun-glint.

The GOME on-ground calibration was performed during the pre-flight calibration phase by TPD/TNO (Netherlands Organization for Applied Scientific Research). The output was a data set containing the so-called "Calibration Key Data" such as stray light correction, BSDF coefficients, radiance response function, ~~or~~ and polarization correction. In the course of switching from on-ground to the in-flight situation various adjustments in the key data had to be applied which were mostly due to air-vacuum wavelength shifts and outgassing of optical coatings. Over the years further updates of the key data have been implemented which were related to the radiance response and to the diffuser BSDF (e.g., Aben et al., 2000; Slijkhuis et al., 2006). An overview of these algorithms using the on-ground calibration data is given in Sect. ~~2.2~~ 3.

Calibration constants which can be directly deduced from measurements using on-board calibration sources (in-flight calibration parameters) are derived during the level 0-to-1 processing. They are fed back immediately to the processor. This comprises the dark signal measurements on the night side of each orbit, the internal LED measurements, and at regular intervals wavelength calibration using the spectral lamp measurements. Figure 2 depicts the processing flow for calculating the respective in-flight calibration parameters including the solar reference measurements. The calibration parameters as well as the Sun Mean Reference (SMR) spectrum are stored in the calibration data base. Monitoring these calibration parameters provides an excellent insight into the long-term stability of the instrument. A detailed description of the corresponding algorithms and the results of the long-term analysis is presented in Sect. 5.

~~A third~~

Figure 3 is a flowchart indicating the order of the steps for processing and calculating the level 1 science data after the calculation of the calibration data. The individual algorithms are applied to the pre-processed solar data, moon and earthshine measurements. 'Normalize' means the normalization of the signal to 1 second exposure time. Detailed descriptions of the individual algorithms are presented in Sec. 3 for the on-ground calibration, and in Sec. 5 for the in-flight calibration. Another step in the entire calibration procedure is the correction of degradation (see Sect. 4.1). Due to degradation in optical components the calibration ~~constants~~ parameters for radiance and irradiance change in time. However, this degradation cannot be derived from on-board calibration sources and the correction has to be obtained offline and externally from the data processor. For GOME this has been done by scientific analysis of the solar observations.




Figure 2. Processing flow for calculating the in-flight calibration parameters from the dark signal measurements (DARK), the internal LED measurements (LED), the spectral lamp measurements (LAMP), and the solar measurements (SUN). The calibration parameters as well as the Sun Mean Reference (SMR) spectrum are stored in the calibration data base.




Figure 3. Processing flow indicating the order of steps for calculating the calibrated level 1 science data, i.e. irradiance (left column), moon radiance (middle column), and earthshine radiance (right columns). See text for detailed explanations.

The last major GDP-L1 processor update has been developed in 2006 in order to provide a first complete reprocessing of the data set available at that time. The main driver for this activity has been the gaps in the solar calibration spectra over long

periods caused by pointing issues on the ERS-2 platform. Furthermore, other algorithmic developments have been included and a detailed analysis of the long-term performance of GOME in terms of numerous diagnostic in-flight calibration parameters has been performed for the first 11 year period (Coldewey-Egbers et al., 2008).

5 In the framework of ESA's GOME-Evolution project, the GDP-L1 Version ~~5.0~~5.1 has been developed in order to generate a completely homogenized, fully calibrated level 1 product for the entire 16 year mission period. Algorithm improvements comprise a new polarization correction (Sect. ~~??~~3.4) and an updated degradation correction (Sect. 4.1), an improved usage of dark signal measurements (Sect. 5.3), as well as revised and improved spectral calibration (Sect. 5.2).

2.3 New GOME level 1 product

10 The previous GOME level 1 data products from the predecessor GDP-L1 Version 4.x and lower contained geolocation, uncalibrated measurements, plus all necessary calibration data (and was thus in modern terminology more like a level 1a product). In addition an external post-processing software "extractor" tool was needed to convert this data to calibrated radiances, or to calibrated solar irradiance, respectively. The advantages were a small product size and the flexibility for the scientific user to perform sensitivity studies on the impact of different calibration steps. However, in the course of time it has turned out that both arguments are no longer valid.

15 In lieu thereof, the new GOME level 1 product generated with GDP-L1 Version 5.1 contains fully calibrated (ir)radiances, corresponding geolocation information, and selected calibration parameters in NetCDF-4 format. Running a separate extraction tool is not necessary anymore; several former extraction software options are now integrated in GDP-L1, others are no longer used (Slijkhuis and Aberle, 2016). The product format and structure are designed to be similar to currently developed or planned EO products, in particular to the Sentinel-5 Precursor mission launched in October 2017. This should enable the
20 application of common reading software to the different atmospheric composition sensors with little or no adaptations required for the various products. In addition to radiance and irradiance data, cloud parameters retrieved with the OCRA (Optical Cloud Recognition Algorithm) and ROCINN (Retrieval of Cloud Information using Neural Networks) algorithms (Lutz et al., 2016; Loyola et al., 2018) have been integrated in the new level 1 product which required reprocessing of the data record in several iterations. Following the request from the users, another addition compared to the old product is geolocation information for
25 each single PMD measurement. A more detailed description of the content and structure of the new level 1 product can be found in Appendix A and in the GOME/ERS-2 Product User Manual (Aberle, 2018).

3 GOME on-ground calibration data and correction algorithms

In this section we provide an overview of the GOME on-ground calibration data and the basic principles of the corresponding correction algorithms. For more details we refer to Slijkhuis and Aberle (2016).

3.0.1 ~~Correction for FPA noise and band 1a residual offset~~

3.1 Correction for FPA noise and band 1a residual offset

~~An additional source of temporary slow noise on the Reticon detector signal is crosstalk~~ Crosstalk correlated to the voltage controlling the Peltier coolers on the ~~FPA~~ Focal Plane Assembly leads to noise on the array detector signal that varies slowly with time. It can be approximated by multiplying the Peltier cooler control signal by a scaling factor which has been obtained during the commissioning phase (April 1995 to July 1996) from one typical orbit and is stored in the calibration key data file. The noise is correlated to the integration time and correction is only necessary for integration times of 6 s or longer typical for band 1a measurements. Furthermore, the correction is only applied to Earthshine measurements. The correction algorithm comprises four steps: (i) apply a high-pass filter to all Peltier output signals belonging to channel 1 from one orbit, (ii) calculate an average value of the filtered Peltier output, (iii) multiply the mean Peltier output by the scaling factor specified for the actual integration time, and (iv) subtract the noise from the signals of the entire band to be corrected.

It appears that after the removal of the Peltier noise as described above, a residual offset remains which has to be corrected since it is too large for, e.g. ozone profile retrieval. This additional correction has been ~~implemented~~ developed in the framework of the CHEOPS-GOME study (Slijkhuis, 2006; Slijkhuis et al., 2006) and is implemented in the L1 processor. It uses the signal of the stray light band 1a (just before the beginning of the nominal band 1a).

3.1.1 ~~Stray light correction~~

3.2 Stray light correction

After the first calibration and characterization measurements of GOME at TPD/TNO, it became obvious that stray light, i.e. light from wavelengths other than the nominal wavelength of a specified detector pixel, is a major issue and needs to be corrected during the ~~in-flight calibration exercise~~. L1 processing. In order to reduce the impact of stray light several improvements were applied before launch such as tilt changes to the gratings, the use of anti-reflection coatings, change of the channel separation between channel 1 and 2, and improvement of internal baffling. Despite these improvements a correction algorithm is still required. Specifically in channels 1 and 2 the signal readouts are spoiled by a non-negligible amount of stray light whose main sources are:

- a uniform or very slowly changing quantity of stray light over the detector pixels induced by diffuse reflections within the FPA;
- ghost stray light signals induced by reflections from the surfaces of the detector arrays and the lenses of the channel telescope; symmetrical ghosts (signals mirrored at the middle of the detector) and asymmetrical ghosts (signals mirrored at some arbitrary detector pixel) were detected;
- out-of-band stray light on the PMDs induced by radiation outside the wavelength range of the detector arrays.

In order to reduce the impact of stray light several improvements were applied before launch such as tilt changes to the gratings, the use of anti-reflection coatings, change of the channel separation between channel 1 and 2, and improvement of internal baffling. Despite these improvements a correction algorithm is still required. Therefore, summed Summed contributions from uniform and ghost stray light are subtracted from the measured signal. The relative uniform stray light levels obtained during the pre-flight calibration are 0.2% for channels 1 and 2, and 0.1% for channels 3 and 4, respectively. These levels are multiplied with the averaged signal fluxes per detector array to get the uniform stray light contribution. For GOME, ~~currently~~ there is only one significant ghost. Its efficiencies (0.05% for channels 1, 2, and 4 and 0.1% for channel 3) were determined during the pre-flight characterization and are multiplied with the mirrored (around the pixel center of the ghost) signal flux to get the ghost stray light contribution. However, the calibration key data for stray light ~~are is~~ probably not more accurate than $\sim 10\%$.
10 i.e. processing errors of 10% of true stray light

3.2.1 Radiometric calibration

3.3 Radiometric calibration

The objective of the radiometric calibration is to transform the 16-bit ~~binary units~~ Binary Units (BU) of the detector pixel readouts into calibrated radiances (photons $\text{s}^{-1} \text{cm}^{-2} \text{nm}^{-1} \text{sr}^{-1}$) or, for the ~~sun~~ Sun, into calibrated irradiance (photons $\text{s}^{-1} \text{cm}^{-2} \text{nm}^{-1}$). ~~For earthshine measurements the intensity calibration also includes the application of a polarization correction (see Sect. ??), whereas for moon measurements a 'basic' intensity calibration is applied.~~

In GDP-L1 the radiometric calibration is divided into several steps. ~~A basic calibration is carried out using the key data from the on-ground calibration. This step comprises the application of the~~ (see also Fig. 3).

The radiance response function, which depends on wavelength, scan angle, and temperature ~~as well as the application of a basic BSDF which is dependent on the wavelength, the azimuth angle, and the elevation of the sunlight on the diffuser. The radiance response function,~~ is applied to the solar, moon, and earthshine measurements. It is a compound function in which the scan angle dependent part and the temperature dependent part are given per channel, for 9 scan angles and for 5 temperatures, respectively. These key data are then interpolated to the actual values of the respective measurement. ~~The BSDF is expressed as parametrization using polynomials. Subsequently, refinements are made to correct~~ Then, solar and earthshine spectra are
25 corrected for instrument degradation (see Sect. 4.1) ~~and to correct the solar irradiance using~~.

The BSDF correction is applied to the solar measurements and comprises two parts. The basic BSDF from the on-ground calibration depends on wavelength, azimuth angle, and the elevation of the sunlight on the diffuser. It is expressed as parametrization using polynomials. The second step uses an improved azimuth dependence of the diffuser BSDF (Slijkhuis et al., 2006). ~~For the latter the~~ The azimuth dependence is fitted using a third-order polynomial in wavelength for all channels. The polynomial
30 coefficients are stored in a look-up-table for a number of azimuth angles which are then linearly interpolated to the actual azimuth angle.

~~Furthermore, the~~ The earthshine radiance is additionally corrected for the so-called 'radiance jump' effect that is caused by the serial readout of the detector, i.e. the last pixel of the array is read out 93.75 ms later than the first pixel. In case of

inhomogeneous ground scenes this effect may be visible as a jump in radiance between two neighboring detectors. The last pixels of one detector record the same wavelengths as the first pixels of the next channel, but at an integration time shifted by 93.75 ms. A linear correction in wavelength is applied which re-normalizes all intensities to the same integration time thereby using information from the PMDs (which are read out every 93.75 ms synchronized with the first detector pixel). Although the correction adjusts the continuum level, it cannot account for any difference in spectral features that may arise from viewing a slightly different ground pixel. For earthshine measurements the intensity calibration also includes the application of a polarization correction (see Sect. 3.4).

3.3.1 Polarization-correction

3.4 Polarization correction

- 10 GOME is a polarization sensitive instrument. The radiance response function described in Sect. ~~??~~3.3 calibrates the instrument assuming unpolarized light. Therefore a correction factor must be applied which describes the ratio

$$c_{pol} = \frac{\text{throughput_for_actual_input_polarization}}{\text{throughput_for_unpolarized_light}} \cdot$$

- ~~The polarization-correction-algorithm~~ of the throughput for actual input polarization to the throughput for unpolarized light. The Polarization Correction Algorithm (PCA) needs the polarization sensitivity of the instrument as well as a characterization of the atmospheric polarization, ~~and it.~~ It is divided into two main parts which both use on-ground key data. The first step is to derive the atmospheric polarization from theory and from measurement for a few wavelengths. Three of these polarization points come from the comparison of channel array signals with broadband PMD signals; the corresponding wavelengths are approximately 360 nm, 500 nm, and 700 nm. A fourth point is obtained from theoretical assumptions and comes from a Rayleigh single-scatter model simulation of polarization in the UV. The second step of the PCA is to interpolate the polarization points to wavelength and to apply the correction to the whole spectrum. Akima interpolation is used for the better part of the spectrum longward of 325 nm. Below 300 nm polarization is taken as a constant which is calculated based on a Rayleigh single-scattering model. In the UV region 300-325 nm the Generalized Distribution Function (GDF, Schutgens and Stammes, 2002) is used. The impact of the polarization correction on the spectra is in the order of $-1 \pm 5\%$. The largest change ($-5 \pm 12\%$) arises in band 1b (283–316 nm) and the minimum impact is found in channel 4 ($-0.5 \pm 0.5\%$).
- 25 Within GOME-Evolution one important improvement for the GDF parametrization has been implemented, that is to use GOME's own retrieved total ozone columns instead of climatological ones. To this end the level 2 ozone values are inserted into the level 1 calibration database. This was hardly possible during the operational phase of the instrument, but for reprocessing there was no limitation, especially because the ozone retrieval is not critically dependent on the polarization curve itself, i.e. within the accuracy needed for this parametrization. Thus, in principle no iterations between level 1 processing and level 2 processing are necessary. Nevertheless, in practice these iterations were made in the course of several intermediate reprocessings so that ozone columns used for the final version are fully compatible with the level 1 polarization.

4 Solar irradiance, PMD measurements, and reflectance

In this section we present the long-term evolution of the solar irradiance, PMD, and reflectance measurements. Monitoring the irradiance (Sec. 4.1) was used to develop a first-order degradation correction algorithm that is routinely applied in GDP-L1 to irradiance and radiance data. Differences between the irradiance and radiance degradation due to different light paths and a strong scan angle dependence are analyzed later in Sec.4.3.

5 ~~Solar irradiance, PMD measurements, and reflectance~~

4.1 Sun mean reference spectrum and degradation correction algorithm

Once per day GOME ~~observes-recorded~~ a short series of Sun spectra via the solar port and a diffuser plate. Thereby, the incidence angle on the diffuser is (i) constant in azimuth (which varies only with season) and (ii) changes in elevation as the Sun moves through the field-of-view. The incidence angle of the scan mirror is 41° (compared to $49^\circ \pm 15^\circ$ for the nadir measurements). All measurements within an elevation angle of $\pm 1.5^\circ$ with respect to the center are averaged and corrected for the azimuth dependence of the diffuser BSDF (see Sect. ~~??3.3~~). This yields the so-called daily Sun ~~mean-reference~~ Mean Reference (SMR) spectrum which is stored in the calibration database and used for the calculation of the earthshine reflectivity spectra. The latter ~~themselves-serve~~ serve themselves as input for almost all retrieval algorithms for atmospheric constituents as well as cloud and aerosol properties.

The relative intensity of the GOME SMR spectra with respect to a reference spectrum from 3rd of July 1995 is depicted in Fig. 4 (January 1996 to January 2011, one spectrum per year) to demonstrate the severe impact of degradation of the optical properties. This comparison shows that the pre-flight radiance parameters were no longer applicable to the in-flight situation (Aben et al., 2000; Hegels et al., 2001). The main degradation as a consequence of extensive exposure to the space environment can be attributed to deposits on the scan mirror (which is coated with a MgF_2 layer) thereby changing its reflective properties. Degradation due to changes related to the diffuser were declared to be negligible (Snel, 2001) . The loss in throughput is especially severe in channel 1. Below 300 nm intensity decreased by 80-95% which implies a significant deterioration of the signal-to-noise ratio. ~~This-Table 1 indicates the approximate signal-to-noise ratios for typical radiance values for channel 1 at 290 and 305 nm, channel 2 above 325 nm, and channels 3 and 4, respectively, at the beginning, in the middle, and at the end of the GOME mission. For channels 2, 3, and 4, the signal-to-noise ratio is well above 1000. Towards shorter wavelengths the ratio significantly decreases due to strong ozone absorption and a weaker solar irradiance. These values are comparable to those obtained for SCIAMACHY (Bovensmann et al., 1999, their Fig. 4) . In channel 1 (and also in the other channels) the signal-to-noise ratio for GOME is expected to decrease linearly with the degradation of the signal (for signal levels below $\sim 15,000$ BU) since the detector noise is exceeding the shot noise. Above signal levels of $\sim 15,000$ BU shot noise becomes dominant. Thus, the strong degradation observed in channel 1 may have a severe impact on the retrieval of atmospheric parameters using this specific spectral region, e.g. ozone profiles (van Peet et al., 2014). The decrease in channel 2 is 40-80%. In channel 3 the decrease-decline (10-40%) started in 2001-when-the-2001. Throughput changes in channel 4 are relatively~~




Figure 4. Relative intensity of GOME Sun mean reference spectra (January 1996 to 2011, one spectrum per year) with respect to a reference spectrum from 3rd of July 1995. Corresponding smooth solid lines denote results of the polynomial fit performed during the degradation correction. Wavelength regions around 470 nm and 600 nm are affected by changes (outgassing) in the dichroic filter (see text for more details).

Table 1. Approximate signal-to-noise ratios at the beginning, mid and end of the GOME mission for five spectral regions. (*) Note that the values for channel 1 in 1995 are for an integration time of 6 s whereas the other values are for an integration time of 1.5 s.

| Spectral region | <u>Channel 1</u> <u>at 290 nm</u> | <u>Channel 1</u> <u>at 305 nm</u> | <u>Channel 2</u> <u>>325 nm</u> | <u>Channel 3</u> | <u>Channel 4</u> |
|--|--------------------------------------|--------------------------------------|---------------------------------------|------------------|------------------|
| <u>Radiance [photons s⁻¹ cm⁻² nm⁻¹ sr⁻¹]</u> | <u>~3.0e10</u> | <u>~4.0e11</u> | <u>~2.0e13</u> | <u>~3.0e13</u> | <u>~3.0e13</u> |
| <u>Signal-to-noise: 1995</u> | <u>~140 (*)</u> | <u>~1100 (*)</u> | <u>~3500</u> | <u>~4000</u> | <u>~2500</u> |
| <u>2001</u> | <u>~20</u> | <u>~180</u> | <u>~2700</u> | <u>~4000</u> | <u>~2500</u> |
| <u>2010</u> | <u>~6</u> | <u>~55</u> | <u>~1600</u> | <u>~3500</u> | <u>~2500</u> |

small. Values above 1 might be due deposits on the coatings which can lead to changes in interference patterns and an increase in intensity (Snel, 2001). Since mid-2001 the measurements were additionally affected by an ERS-2 pointing problem as a consequence of the loss of the ~~gyroscope functionality. Throughput changes in channel 4 are relatively small. gyroscopes~~ (which govern the platform steering) functionality. SCIAMACHY as well as GOME-2 suffer from degradation in pretty much the same way (Noël et al., 2007; Bramstedt et al., 2009; Munro et al., 2016), whereas OMI irradiances degraded by only 3-8% from 2005 to 2015 (Schenkeveld et al., 2017).

In Fig. 4 at wavelength regions around 470 nm and 600 nm changes in intensity are affected by changes (outgassing of coatings) in the dichroic filter which separates the wavelengths of channels 3 and 4. Unpredictable polarization-sensitive changes were observed and the radiometric calibration in these regions might be doubtful. Furthermore the outgassing is assumed to be responsible for the slight transmission increase in channel 3 in the early part of the mission. The low-frequency oscillating structure appearing in all channels is the result of the etalon effect which is caused by a changing thickness of ice deposits on the detectors and which leads to spectral interference patterns (Mount et al., 1992). At present no attempts are being made to correct for this effect.

In order to remediate the observed GOME science channel degradation a correction algorithm was developed in the framework of the ESA project GDAQI (GOME Data Quality Improvement, Aben et al., 2000). This degradation correction is applied to irradiance and radiance spectra as additional part of the radiometric calibration. The degradation correction approach that was chosen is the comparison of all available solar data from the entire mission period with the corresponding solar data of a reference day in the early GOME lifetime (3rd of July 1995). The Sun is a reliably stable input source to monitor the instrument throughput despite of small changes in the solar spectrum due to changes in solar activity. This study was done for both the GOME science channels and the PMDs. The temporal changes have been determined by building ratios of all solar spectra with the solar spectrum of the reference day t_0 which may be written as

$$\frac{I_{Sun}(\lambda, t)}{I_{Sun}(\lambda, t_0)} = P_{Deg}(\lambda, t) \cdot C_{SED}(t) \cdot Residual(\lambda, t), \quad (1)$$

where $P_{Deg}(\lambda, t)$ is the used degradation function dependent on wavelength λ and time t . $C_{SED}(t)$ is the intensity correction due to the seasonal variation in Sun-Earth distance, and $Residual(\lambda, t)$ is the remaining structure. Note that the impact of the etalon effect and the changes in the dichroic filter are not accounted for. For the determination of the degradation correction function $P_{Deg}(\lambda, t)$ a two-step approach was developed: (i) each irradiance ratio (per channel) is approximated by a polynomial function in wavelength and (ii) each coefficient of this polynomial in wavelength is subsequently described by a time-dependent expression. Thus, for the degradation function $P_{Deg}(\lambda, t)$ per channel the following expression has been obtained:

$$P_{Deg}(\lambda, t) = \sum_{k=0}^n a_k(t) \cdot (\lambda - \lambda_0)^k. \quad (2)$$

λ_0 is the center wavelength in each channel. Each coefficient $a_k(t)$ of the polynomial in wavelength is taken from a look-up table (LUT). For channels 1 and 2 third order polynomials ($n = 3$) are used, whereas in channels 3 and 4 quadratic ($n = 2$) and linear ($n = 1$) polynomials are used, respectively (see smooth curves in Fig. 4). The LUT is generated by smoothing the time series of each polynomial wavelength coefficient using a Savitzky-Golay filter (Press et al., 1992) with a filter width of 250 days. Figure 5 shows, for each channel, the first polynomial coefficient (a_0) and the corresponding smoothed curve as a function of time. In channel 1 degradation started almost immediately after launch. Until 2000 the intensity decreased to $\sim 50\%$ of the early-mission values. In 2001/02 measurements were additionally affected by the ERS-2 pointing problem. In channel 2 significant decrease in intensity is observed especially during this period. Channels 2000-2002. Since mid-2001 all measurements were additionally affected by the ERS-2 pointing problem. Furthermore, channels 3 and 4 are additionally affected by changes in the dichroic filter (i.e. outgassing of coatings).




Figure 5. First polynomial coefficient a_0 from wavelength fit as a function of time (small dots, blue: channel 1, red: channel 2, green: channel 3, and magenta: channel 4). The light blue, orange, light green, and light violet curves denote the corresponding smoothed curves using a Savitzky-Golay smoothing filter with a filter width of 250 days.

In GDP-L1 the degradation correction is then applied according to:

$$I_{SunCorr}(\lambda, t) = I_{Sun}(\lambda, t) / P_{Deg}(\lambda, t). \quad (3)$$

In addition to the Sun the moon provides an independent irradiance source and in principle GOME lunar measurements can be used to characterize and monitor instrument performance and degradation (Dobber, 1997; Dobber et al., 1998). The moon is viewed on the eclipse side of the orbit over the nadir-scan mirror at an incidence angle of $5^\circ - 15^\circ$. The amount of light is of the same order of magnitude as for the earthshine observations, though calibration measurements are complicated by several factors such as moon availability and phase, non-uniformity of the moon surface, polarization, and partial slit filling. Orbit requirements were so strict that measurements are only possible for a very limited number of orbits per year, with the moon phase always being ~ 0.6 between full moon and the last quarter. After 2003 no more moon measurements are available. Thus, GDP-L1 does not attempt to generate calibrated radiances for the moon and a long-term analysis of moon observations has not been performed. Furthermore, for an accurate monitoring of instrument degradation using lunar measurements a more precise characterization of the reflective and scattering properties of the moon would be necessary. Early investigations by Dobber (1997) using the first 18 months of GOME's lifetime confirmed the assumption that the scan mirror (instead of the diffuser) is primarily subject to degradation.

4.2 PMD measurements and Q_factors

The relative change of the solar PMD measurements as a function of time with respect to a reference measurement from 3rd of July 1995 are shown in Fig. 6 for all three PMDs. Note that the measurements were normalized to 1 astronomical unit (AU) in order to eliminate seasonality. As for the SMR spectra the degradation for the PMDs is strongest for PMD 1 which corresponds




Figure 6. Relative change of PMD signals as a function of time from 1995 to 2011 for PMDs 1, 2, and 3 (from top to bottom). Reference measurement is from 3rd of July 1995.

to channel 2. The signal decreases to about 40% of the original value. The temporal evolution for PMDs 2 and 3 is similar to the behavior of the signals in channels 3 and 4, respectively.

PMD Q_factors are self-calibration constants which ensure that the calculated fractional polarization p of the Sun is unpolarized with $p = 0.5$. They are defined as the relative difference between the measured solar signal of PMD $_i$, with $i = 1, 2, 3$, and the expected PMD signal calculated from the key data and the corresponding channel signals, when unpolarized input is assumed:

$$Q_factor_i = (PMD_i - \sum_j X_j \times channel_j) / PMD_i, \quad (4)$$

where $channel_j$ is the channel signal of pixel j and X_j is the ratio of the PMD signal to the channel signal for a monochromatic input signal as obtained from on-ground calibration measurements. Q_factors thus involve the differential degradation between

PMD signals and the channel signals since the time of on-ground calibration.

Figure 7 shows the Q_factors for PMD 1, 2, and 3 (from top to bottom) as a function of time from 1995 to 2011. In principle, the behavior of the Q_factors as detected in the previous study (Coldewey-Egbers et al., 2008) continued. For the first Q_factor a decrease until 2001 is observed. From 2002 to 2011 Q_factor 1 steadily increased. That means that in the first period the degradation of the PMD signal was stronger than the degradation of the signal in channel 2, whereas in the second period the channel signal decreased faster than the PMD signal. Q_factor 2 increased slowly from the beginning of the measurements – indicating that the PMD signal degraded less than the average signal in channel 3 – and reached nearly the same value as Q_factor 1 at the end of the mission. For Q_factor 3 note that it is nonzero already at the beginning of the measurements. This is related to stray light (wavelength > 790 nm) which affected in particular PMD 3, whereas PMD 1 had a negligible stray light




Figure 7. Q_factors 1, 2, and 3 (from top to bottom) as a function of time from 1995 to 2011.

effect. Q_factor 3 remained more or less stable until 1999 followed by a slow increase until 2011. Outliers are due to GOME operation anomalies such as cooler switch-offs, instrument or satellite switch-offs, on-board anomalies, or special operations (see also: GOME Mission Operations Overview).

4.3 Reflectance degradation

- 5 In Sect. 4.1 the approach used to correct for instrument degradation was described. This 'soft' correction is applied a first order correction as it is applied in GDP-L1 to both irradiance and radiance spectra ~~in GDP-L1~~ thereby assuming that both spectra degrade in the same way. The ratio

$$R = \frac{\pi I}{\mu_0 E}, \quad (5)$$

- where I is the Top-Of-Atmosphere (TOA) radiance reflected and scattered by the Earth's atmosphere, E is the solar irradiance, and μ_0 is the cosine of the solar zenith angle, defines the reflectance R , which is used by many algorithms to retrieve the amount of atmospheric constituents. From Eq. 5 it is clear that the reflectance remains unchanged in the level 0-to-1 processing since the applied degradation correction cancels out under the assumption that the BSDF does not degrade. However, the light paths for radiance and irradiance measurements are different and the degradation of the scan mirror indicates a strong dependence on the incidence angle (Snel, 2001). This leads to a substantial differential degradation of radiance and irradiance spectra (Tanzi et al., 2001; van der A et al., 2002) and, thus, to degradation in the reflectance, which may affect for example ozone profile retrievals (van der A et al., 2002; Liu et al., 2007) or the determination of total ozone columns using a direct fitting approach (Lerot et al., 2014), because these algorithms are sensitive to absolutely calibrated reflectances. Correction approaches for the reflectance degradation have been developed in the past which rely on, e.g. the comparison of experimental and simulated data (van der A et al., 2002; Cai et al., 2012), the comparison of satellite reflectance spectra with ground-based reference

spectra (Lerot et al., 2014), or the comparison of global average reflectance with respect to global average reflectance from the beginning of the mission (Liu et al., 2007; Tilstra et al., 2012). For the latter approach the underlying assumption is that the global average reflectance does not change in time. For irradiance degradation correction (see Sect. 4.1) this assumption can be regarded fulfilled but the earthshine radiance and, thus, the reflectance on the other hand depend strongly on highly variable atmospheric conditions such as clouds, trace gases, aerosols, or surface albedo and on the viewing angle. Therefore, retrievals using this correction may be inadequate for trend studies (Liu et al., 2007). On the other hand, Garane et al. (2018) have shown that using the latest version of the direct fitting approach GODFIT for ozone retrieval (GODFIT version 4) GOME (as well as OMI) ~~perform~~ performs in an extremely stable way and ~~do~~ does not require any spectral soft calibration procedure.

In the framework of GOME-Evolution we analyzed the long-term behavior of the GOME reflectance using measurements over so-called Pseudo Invariant Calibration Sites (PICSs) which have been identified and characterized by the Committee on Earth Observation Satellites (CEOS) to be suitable to detect the radiometric stability of satellite sensors (CEOS Pseudo Invariant Calibration Sites). The advantages of these sites are the spatial uniformity and homogeneity, their stable spectral characteristics over time, and generally high reflectance to enhance the signal-to-noise ratio. At the moment there are six CEOS reference PICSs all located in the Saharan desert: Libya-1, and -4, Mauritania-1 and -2, and Algeria-3 and -5, respectively. They are usually made up of sand dunes with climatologically low aerosol loading, little rainfall, and practically no vegetation or human impact. More details on the PICSs can be found in Helder et al. (2010). In the past these sites have been widely used in post launch calibration and validation of satellite sensors (e.g., Smith and Cox, 2013; Mishra et al., 2014; Sun et al., 2014; Uprety and Cao, 2015).

For our study we selected four reference sites: Libya-1 (24.42°N, 13.35°E) and -4 (28.55°N, 23.39°E) as well as Algeria-3 (30.32°N, 7.66°E) and -5 (31.02°N, 2.23°E). The geolocation in the parentheses denotes the center latitudes and longitudes. Fortunately, in this area the impact of the ERS-2 tape recorder failure in June 2003 is quite small, so that the time series are almost complete with only a short gap in 2003. We limit our analysis to two single wavelengths in the UV part of the spectrum (325 and 335 nm, respectively) which mark the lower and upper limit of the fitting window for total ozone retrieval (Loyola et al., 2011; Lerot et al., 2014). All GOME ground pixels with cloud fraction less than 0.2 were extracted that fall into a square area of $\pm 1.5^\circ$ in latitude and longitude around the center geolocation of the reference site. About 3000 pixels were found for each reference site which fulfill these criteria. In general, the top-of-atmosphere reflectance of a scene measured by the satellite sensor depends on the viewing geometry because of the anisotropy of the surface reflectance. The reflectance is higher for west pixels, when the Sun and the satellite are on the same side of the scene (backward scattering viewing geometry), than for east pixels in forward scattering viewing geometry (Zoogman et al., 2016; Lorente et al., 2018). The anisotropy depends on wavelength and on the surface properties. In case of the Saharan PICSs scenes and for wavelengths of 325 and 335 nm the difference in reflectance between west and east pixels is about 25%.

Figure 8 shows the reflectance normalized to the reflectance at the beginning of the mission, segregated by pixel type (east, nadir, and west which mark the three forward scans) for both wavelengths (solid curves denote 325 nm and dashed curves denote 335 nm) as a function of time for the entire mission for PICS Libya-4. The data gap from mid 2003 to early 2004 is due to the tape recorder failure. Fluctuations in reflectance related to the seasonal variation and to short-term variation of




Figure 8. Relative reflectivity as a function of time for cloud-free GOME measurements at 325 nm (solid) and at 335 nm (dashed) for Libya-4 reference site. Colors denote the different GOME pixel types: east (green), nadir (blue), and west (cyan) ground pixels. The data gap between mid 2003 and early 2004 is due to the ERS-2 on-board tape recorder failure in June 2003.

atmospheric conditions were smoothed out. Until late 1999 the curves show only little degradation. Afterward they start to increase, reach a maximum in 2001, and then decrease again to values below 1 in early 2003. From 2004 to 2011 the curves steadily increase except for the reflectance degradation for west pixels at 325 nm (solid cyan curve) which show a slight decrease at the very end of the mission. ~~The degradation~~ These ups and downs in the degradation might be related to changes in interference patterns which can lead to an increase or decrease in reflectance. Furthermore, the degradation depends strongly on the pixel type, i.e. the line-of-sight. Until 2003 west pixels are much less affected than nadir and east pixels, and also after 2003 the behavior of the reflectance from west pixels is slightly different compared to nadir and east pixels. For the other three PICSs Libya-1, Algeria-3, and Algeria-5 we found very similar results (without figure) with negligible differences compared to Libya-4. In addition, we analyzed the reflectance for cloud free pixels over the Mediterranean Sea between Greece and Egypt, which shows the same temporal evolution as the reflectance over the Saharan desert (without figure) although the surface albedo is much lower there. As mentioned earlier, in principle this analysis could be used for correcting the reflectance degradation. However, the underlying requirement that the reflectance remains stable over a long time period might not be fulfilled in every case.

5 GOME in-flight calibration parameters

- 15 In this section we present the analysis of the GOME calibration parameters obtained from measurements using on-board calibration sources and applied during the level 0-to-1 processing as described in Sect.2.2. For a detailed description of the individual calibration algorithms related to the parameters we refer to the GOME Algorithm Theoretical Basis Document (Slijkhuis and Aberle, 2016). Monitoring the individual parameters was performed with special emphasis on the analysis of the long-term stability.

5.1 Overview

In the framework of the GOME-Evolution project the complete set of in-flight calibration data has been revisited and re-analyzed in order to draw conclusions on the long-term stability of the GOME sensor and to optimize the GDP-L1 usage of the in-flight calibration for the entire mission. The database contains spectral lamp measurements for the wavelength calibration (see Sect. 5.2), dark current measurements for all integration time patterns (see Sect. 5.3), LED measurements for the pixel-to-pixel gain correction (see Sect. 5.4), as well as the Sun mean reference spectra, and moon and PMD measurements. After the ERS-2 tape recorder failure in June 2003 the number of available calibration data is significantly reduced since only data within accessibility of an ERS-2 receiving station were transmitted to ground. In particular, no more moon measurements are available after 2003.

5.2 Spectral calibration

The objective of the spectral calibration is to assign a certain wavelength to each individual GOME detector pixel. Therefore, the instrument houses a platinum-chromium-neon hollow cathode emission lamp (Murray, 1994). This lamp provides a sufficient number of atomic emission lines of these three elements with well-known spectral positions which allow the wavelength allocation. At first, spectral calibration parameters are calculated by the determination of the pixel number center of the spectral lines and the subsequent fitting of a polynomial through these pixel-wavelength pairs. The second step is the application of the calibration parameters from the previous step to the measurements.

Several lamp spectra were measured (i) over the orbit approximately once per month, during the calibration timeline which was run for five orbits, and (ii) every day just before and after the Sun calibration. The latter measurements are available until April 1998. Since September 2001 the calibration lamp was used only during the five orbits of the monthly calibration due to numerous lamp failures since the voltage has not reached its nominal value (see Fig.?? and also GOME Mission Operations Overview).

For the spectral calibration a total of 68 candidate emission lines within GOME's spectral range from 240 to 790 nm ~~has~~ have been selected from the reference lamp atlas (Murray, 1994) and is stored in the calibration key data base. The lamp measurements of the individual lines can be regarded as statistical distributions from which the moments can be calculated.

They contain characteristic information about the spectral lines that are needed to select those lines suitable for an accurate calibration. The aforementioned moments are the mean value, i.e. the pixel number center of the maximum intensity, as well as the variance, the standard deviation σ , and the skewness *skew*. The full width half maximum (FWHM) is computed from the standard deviation. To be selected, the moments of a spectral line must meet the following statistical criteria: (i) the signal of the center pixel shall not be below a certain minimum, i.e. well above the noise level, (ii) the FWHM shall not be below a certain value in order to fulfill the Nyquist criteria for the digital recording of analogue signals, and (iii) the skewness shall not be larger than a certain value, i.e. the line must be roughly symmetric. Reasonable thresholds for the criteria have been determined during the pre-flight measurements and the commissioning phase. Current values are 50 BU/s for channel 1 and 300 BU/s for channels 2–4 for the first criterion, $\sigma \geq 0.6$, $\text{FWHM} \geq 1.5 \text{ pixel}$, and $\text{skew} \leq 0.6$ for the second and third criterion, respectively.




Figure 9. Change in wavelength (with respect to the beginning of the mission) as a function of time from 1995 to 2011 for two selected lamp lines per channel. From top to bottom: channel 1, channel 2, channel 3, and channel 4.

As mentioned before, the calibration parameters are obtained by fitting a polynomial through the pixel-wavelength pairs. In channels 1 and 2 third-order polynomials are used, whereas in channels 3 and 4 fourth-order polynomials are used, respectively. At least seven spectral lines per channel are needed for the fit which is performed using the singular value decomposition algorithm (Press et al., 1992).

- 5 The statistical parameters of each individual emission line were analyzed in terms of both long- and short-term stability. Regarding short-term variability few lines were found whose moments show jumps between two values leading to jumps in the fitted polynomial coefficients. Other lines fulfill the aforementioned criteria only in very few cases which also results in jumps in the fitted coefficients. This analysis has lead to a revised spectral line list (by excluding the identified unstable lines) that improved the stability of the spectral calibration for the complete mission. Figure 9 shows the wavelength changes of selected
- 10 lamp lines (two per channel) as a function of time. Depicted is the difference (in nm) with respect to the wavelength at the beginning of the mission. The stability of the wavelengths is excellent until 2004. Toward the end of the mission the variability increases slightly, in particular in channels 1 and 3. The standard deviation of the wavelength changes is 0.0015 nm in channel 1, 0.0025 nm in channel 3, and less than 0.001 nm in channels 2 and 4, respectively. These values are comparable to the analysis by van Geffen (2004) who ~~use~~ used a different wavelength calibration approach (van Geffen and van Oss, 2003). They found
- 15 temporal variations of the wavelength calibration from 0.0015 to 0.0034 nm for nine narrow spectral bands.

One of the key elements in the optical system of GOME is a quartz predisperser prism. The wavelength calibration is sensitive to the dispersion of this prism, whose refractive index varies with temperature. Thus, the calibration parameters from




Figure 10. Temperature measured at the pre-disperser prism as a function of time from 1995 to 2011 (blue) and linear fit (red). The temperature increase is about $2.5 \text{ K decade}^{-1}$. Magenta dots denote the increase of the pre-disperser temperature along one orbit in Kelvin hour⁻¹ (right y-axis) for the years 1995 to 2003. Analysis of later years is not possible due to the tape recorder failure (incomplete orbits).

the lamp measurements are stored in the database as a function of this temperature. In the operational processing the most recent calibration parameters are then selected from the database according to the predisperser temperatures encountered in the actual orbit. Each individual GOME spectrum is, thus, implicitly corrected for temperature variations that are caused by seasonal variations, the position in the orbit, and by the rate of degradation of thermally sensitive optical elements.

- 5 Figure 10 shows the time series of the predisperser temperature from 1995 to 2011 (blue dots). An increase of $\sim 4 \text{ K}$ within the instrument's lifetime is found which is due to degradation of the thermal system. Furthermore, the curve exhibits a seasonal cycle with maximum values in December/January when the Sun-Earth distance is at a minimum. Outliers are caused by instrument and cooler switch-offs. Magenta dots denote the increase of the predisperser temperature along an orbit in Kelvin hour⁻¹ for the years 1995 to 2003. The increase along an orbit is due to warming of the satellite by the Sun and because
- 10 light passes through the instrument. This analysis relies on the average of 60-70 days per year; for each day the temperature measured along the first orbit, which is always located between 120-160°E, was investigated. The errorbars are a measure of the intra-annual variability. We did not analyse the dependence of the predisperser temperature itself on longitude as in ~~van Geffen (2004, their Fig. 3)~~ van Geffen (2004, his Fig. 3). They found a maximum of the predisperser temperature over the Atlantic and a minimum over the Pacific. ~~However, they also~~ Furthermore, they stated that the temperature increase
- 15 along the orbits does not show a dependence on longitude. During the first eight years of the mission the temperature ~~raise~~ rise along the illuminated part of one orbit increased from about 0.7 K/h to about 0.9 K/h. Unfortunately, analysis of later years is not possible due to the ERS-2 tape recorder failure and incomplete orbits.

5.3 Dark signal correction

- 20 The detectors integrated in GOME are random access linear photo-diode arrays. One of the characteristics of these devices is a certain amount of dark current due to thermal leakage. It is expected that this current will depend on the orbital position

of the satellite and also the time into the mission. Therefore it is necessary to continuously monitor the dark current and the associated noise which is done by means of periodically taken dark-side measurements (Sect. 5.3.1). In this case the scan mirror points toward the GOME interior. The PMD detectors are non-integrating devices and, therefore, do not have a leakage current. Nevertheless, those detectors must be corrected for their zero offsets and the noise must be monitored (see Sect. 5.3.2).

5 5.3.1 Dark current and dark current noise

The complete dark signal comprises two parts: (i) a constant value of $\sim 140\text{--}150$ ~~BU~~ Binary Units (BU) which is called the fixed pattern readout noise (FPRN) and (ii) the time-dependent leakage current itself which is about ~ 2 BU/s. This value is quite small because of the low temperature (-38°C) of the detector arrays. The dark signal measurements have to be taken with the same integration time patterns as those used for scanning and other calibration measurements since it was found that a certain amount of cross talk is present which depends on the integration time. However, the detector temperature is not taken into account for GOME as it is the case for the dark signal correction of the GOME-2 instrument (Munro et al., 2016).

The dark signal correction is the subtraction of a mean dark signal spectrum from the measured signal $S_i^{meas,k}$:

$$S_i = S_i^{meas,k} - \overline{S_i^{dark,k}}, \quad (6)$$

where $i = 1, \dots, 1024$ detector pixels. The integration time pattern k describes the number of clock pulses, where one pulse takes 93.75 ms, e.g., a time pattern of 640 is equivalent to 60 s.

The mean dark signal for $n = 10$ consecutive measurements is defined as

$$\overline{S_i^{dark,k}} = \frac{1}{n} \sum_{j=1}^n \left(S_i^{dark,k} \right)_j. \quad (7)$$

Figure 11 shows the dark signal as a function of time for the three most representative integration time patterns: (i) the normal scanning orbits with 12 s integration time for band 1a and 1.5 s for the other bands (with co-adding applied), (ii) the LED measurements for the pixel-to-pixel gain correction (see next Section) with 30 s integration time for all bands, and (iii) the polar view mode with 60 s integration time for band 1a and 6 s for the other bands. Figure 11 shows the dark signal for bands 1a, 2b, 3, and 4 (from top to bottom) and for time patterns (i) to (iii) from left to right. All panels denote a significant increase ~~in the leakage current~~ over time.

Note that the dark signal in bands 2b, 3, and 4 for the normal scanning orbits (Fig. 11, left panel) is much higher due to co-adding of four measurement sequences. At present there is no explanation for the behavior of the signal from 2005 to 2007. It is most obvious in channel 4 for the normal scanning mode (bottom left panel) and for the polar view mode (bottom right panel). The signal decreased significantly in 2005 (by 40 BU for the normal scanning mode), reached a minimum in the beginning of 2006 and increased again, in which the entire development of this anomaly is quite smooth. The jumps in the time series (e.g., seen in channels 2 and 3 for the normal scanning mode and in channel 3 for the polar view mode) are due to instrument or cooler switch-offs or instrument anomalies.

The noted increase in the dark signal is an increase in the ~~second part of the~~ leakage current, i.e. the time-dependent part. Figure 12 shows the ~~leakage current~~ dark signal as a function of the integration time for four different years 1997, 2002, 2007,




Figure 11. ~~Leakage-current~~ Dark signal in Binary Units (BU) as a function of time from 1995 to 2011 for three integration time patterns (see text for more explanation): normal scanning mode (12 s integration time for band 1a and 1.5 s for the other bands, co-adding applied for bands 2a, 2b, 3, and 4, left panels), LED measurement mode (30 s integration time for each band, middle panels), and polar view mode (60 s integration time for band 1a and 6 s for the other bands, right panels). From top to bottom: band 1a, band 2b, band 3, and band 4.

and 2011. Different symbols and linestyles denote the individual bands 1a, 2b, 3, and 4. The y-intercept represents the FPRN which is about 140-150 BU and remains constant over the entire time period. The slope denotes the time-dependent leakage current which is quite similar for all channels (~ 2 BU/s) and which increases over time. The increase is also almost identical for all channels and amounts to about 4 BU/s per decade (~ 6.5 BU/s from 1995 to 2011). This is comparable to earlier work by Dehn (2003) and our predecessor study (Coldewey-Egbers et al., 2008) as well as to Munro et al. (2016) who analyzed the dark signal for the GOME-2 instrument on-board the MetOp series of satellites using the same type of detectors. For OMI, which is a nadir viewing UV-VIS imaging spectrograph using two-dimensional charge-coupled device (CCD) detectors (Levelt et al., 2006; Dobber et al., 2006), a 7-fold dark current increase was found from 2005 to 2015 (Schenkeveld et al., 2017), and for GOMOS/ENVISAT (Global Ozone Monitoring by Occultation of Stars) – using the same CCD detectors as OMI an even higher increase was found (Bertaux et al., 2010). Although the increase in the dark current seems to be significant, there is not necessarily a negative impact on the quality of the level 1 data products as long as appropriate dark current measurements are available and applied during the level 0-to-1 processing.

We found that it is not only the leakage current itself which changed over time, but also its distribution ~~which~~ considerably widened. Figure 13 shows histograms of the dark signal for spectral band 1a (240–283 nm) for an integration time of 12 s (nominal scanning mode) for every two years from 1997 to 2011. The data ~~correspond~~ corresponds to the upper left panel of Fig. 11. Colored numbers stuck to the individual histograms denote the median values and the FWHM of the distribution. The latter is additionally indicated by the filled rectangles. As seen in Fig. 11 the dark signal significantly increased over time from ~ 176 BU to ~ 242 BU. Furthermore a noticeable, almost three-fold ~~γ~~-broadening of the distribution was found. FWHM




Figure 12. Dark ~~signale~~signal in Binary Units (BU) as a function of integration time for January 1997 (red), 2002 (black), 2007 (green), and 2011 (blue), respectively. Different symbols and linestyles denote channels 1a (solid), 2b (dotted), 3 (dashed), and 4 (dash-dotted).

increased from 2.7 BU to 7.4 BU. A widening of the dark current distribution was also noticed for OMI (Schenkeveld et al., 2017) and GOMOS (Bertaux et al., 2010).

The noise on the signals of the detector pixel readouts is expected to be constant over all individual pixels. For each detector pixel ~~are the~~standard deviation from all leakage measurements from one orbit with the same integration time is computed.

- 5 The noise is then the average of all standard deviations. The annual mean noise level is shown in Figure 14 (blue curves, left y-axis) as a function of time for three different integration time patterns (scanning, moon and LED). The error bars denote the standard deviation for the annual mean. The lowest noise level (~ 2 BU) is found for LED dark signal calibration measurements which have the longest integration time (30 s), whereas the noise level for scanning and moon integration time pattern are quite similar and about 4 BU. The values remain more or less constant until June 2003. Afterward the noise level for LED dark
- 10 signal calibration measurements slightly increased (dot-dashed blue line), whereas a decrease is found for moon dark signal calibration measurements (dashed blue curve). Red curves (right y-axis) denote the number of available dark signal calibration measurements. ~~In case of LED dark signal calibration measurements the~~The most significant decrease in the number of available measurements is ~~found for the LED dark signal calibration measurements~~found for the LED dark signal calibration measurements (dot-dashed red curve).

5.3.2 PMD offset and noise

- 15 The signals of the PMD detectors as non-integrating devices must be corrected for their zero offsets and the associated noise must be monitored. Figure 15 shows PMD offsets for each PMD as a function of time for the entire mission period. The offset




Figure 13. Histograms-Probability Density Function (PDF) of the dark signal in spectral band 1a for an integration time of 12 s for every two years from 1997 to 2011. Colored numbers denote the median value in Binary Units (BU) and the FWHM (in parenthesis) of the distribution. The latter is additionally indicated by the colored rectangles.




Figure 14. Annual mean dark current noise (blue curves, left y-axis) in Binary Units (BU) for three integration time patterns: scanning (solid curve), moon (dashed curve), and LED (dot-dashed curve). Red curves (right y-axis) denote the number of available calibration measurements for the three individual time patterns.

of PMD 1 is about 1320 BU, whereas it is about 510 BU for PMDs 2 and 3. All offsets indicate a very small increase of 0.8% in 16 years of the mission. The increase is nearly linear for PMDs 2 and 3, whereas for PMD 1 the increase started in 1999; the PMD 1 offset reached a maximum at the end of 2004, decreased in 2005 and increased again thereafter. For all PMDs




Figure 15. PMD offset in Binary Units (BU) as a function of time for PMD 1, 2, and 3 (from top to bottom).

the offsets seem to have two states, and jumps between the two states are due to cooler and instrument switch-offs as well as instrument anomalies.

Figure 16 shows the annual mean PMD noise as a function of time. The PMD noise is defined as the mean value of the standard deviations which are calculated for each PMD over all 16 individual PMD measurements. It is about 0.5-1.5 BU. The previous study (Coldewey-Egbers et al., 2008, their Fig. 9d) has shown the impact of the South Atlantic Anomaly on the noise level which increases significantly when measurements from this area are taken into account. In the new GDP-L1 version these calibration measurements are discarded (see previous-following section). In general the noise level remains stable over the entire period although – as a consequence of the tape recorder failure in June 2003 – a slight change in the noise level was found, in particular for the moon dark signal calibration measurements.

10 5.3.3 Impact of South Atlantic Anomaly

The South Atlantic Anomaly (SAA) is an area of enhanced flux of energetic particles due to a dip in the Earth's inner Van Allen radiation belt. In this region low Earth orbit spacecrafts are exposed to higher-than-normal radiation levels and may suffer from damage (Heirtzler, 2002; Casadio and Arino, 2011). High energy protons impact the detectors of GOME, i.e. the background signal is higher than the normal dark signal, the noise is enhanced, and the measured spectra are also prone to intensity spikes caused by cosmic particles.

For this reason all calibration measurements in the SAA are discarded. The algorithm to identify the SAA uses the signal from PMD 1 since it has been found that the noise level on PMD 1 is a reliable indicator of the enhanced particle bombardment in the SAA region. Figure 17 shows a map of the GOME long-term mean PMD 1 noise derived from the first five years of the mission. The impact of the SAA clearly appears in terms of significantly enhanced PMD 1 noise in an oval-shaped region




Figure 16. Annual mean PMD noise (blue curves, left y-axis) in Binary Units (BU) for three integration time patterns: scanning (solid curve), moon (dashed curve), and LED (dot-dashed curve). Red curves (right y-axis) denote the number of available calibration measurements for the three individual time patterns.

centered at the east coast of Brazil. The SAA spans from 50°S–0° in latitude and from 90°W–30°E in longitude. During the level 0-to-1 processing PMD measurements are grouped and for each group a noise value w.r.t. the median value is calculated. If the noise value exceeds a certain threshold all calibration measurements from the group are discarded. This includes also the lamp measurements for the spectral calibration and the LED measurements for the pixel-to-pixel gain correction (see next Section). The new algorithm defines an 'inside SAA' and an 'outside SAA' region for dark signal values in the calibration database.

5.4 Pixel-to-pixel gain correction

The pixel-to-pixel Pixel-to-Pixel Gain (PPG) variability in quantum efficiency of each diode detector array is characterized and corrected using internal LEDs. Each channel has a monochromatic red LED located between the channel optics and the detector window (see Fig. 1), i.e. the detectors are illuminated directly without any dispersing element in between that may suffer from degradation effects. The monitored detector signal corresponds to a superposition of a smoothly varying signal caused by the LED characteristics and a small-scale structure due to the slightly different sensitivity of each pixel. The determination of the correction spectra for each of the four channels is based on a mean value of several consecutive LED measurements and a smoothed curve through this average using a triangle filtering window:

$$c_i = \frac{S_i^{smooth}}{S_i^{LED}}, \quad (8)$$




Figure 17. Map of the GOME long-term mean PMD 1 noise in A/D[-converter] Units (= Binary Units) derived from the first five years of the mission. Enhanced noise levels indicate measurements affected by the South Atlantic Anomaly.

where c_i is the correction factor of detector pixel i , $\overline{S_i^{LED}}$ is the mean value of several consecutive LED measurements, and S_i^{smooth} is the smoothed curve through this averaged measurements. The latter is calculated by means of

$$S_i^{smooth} = \frac{\sum_{k=-n}^n \frac{n-|k|}{n} \times \overline{S_{i+k}^{LED}}}{\sum_{k=-n}^n \frac{n-|k|}{n}} \quad (9)$$

using a triangle filtering window of width $n = 5$. The application of the PPG correction is then simply

$$S_i^{corr} = S_i c_i, \quad (10)$$

where S_i is the measured signal value of detector pixel i , and S_i^{corr} is the corrected value.

Typically, the LED spectra were obtained in monthly intervals until 2003. From 2003 onward LED measurements are limited to two or three sequences per year. The absolute radiance correction due to the pixel-to-pixel variability is very small ($\sim 0.02\%$). However, it may not be negligible in wavelength regions used for the retrieval of weak absorbers such as bromine oxide. Figure 18 shows the relative intensity of the LED spectra as a function of time and wavelength with respect to a reference spectrum from the beginning of the measurements (27th of June 1995). The nearly linear decrease which was already detected in the previous study (Coldewey-Egbers et al., 2008) continued until the end of the mission in 2011 and is due to the degradation of the LEDs' brightnesses themselves. The output decreased to $\sim 60\%$. It is almost homogeneous over the complete wavelength range of each channel. The steepest decrease is found in channel 1.

In addition we analyzed the distribution of the PPG spectra-correction factors as a function of time. Figure 19 shows box-whisker plots of the distribution for each channel and as a function of time. We show one distribution per year. In channel 1 the amplitude of the PPG correction spectrum is slightly larger than for the other channels. Nevertheless, the distribution of the correction spectrum remains roughly stable over the entire period, whereas for channel 2 a significant broadening of the




Figure 18. Relative intensity of LED spectra as a function of time (x-axis) and wavelength (y-axis) with respect to a reference spectrum from ~~27~~^{27th} of June 1995.

distribution is found. The standard deviation increased by a factor of ~ 2.5 in this channel, which indicates that the variability in sensitivity between the individual pixels increased significantly. For channels 3 and 4 a broadening of the distribution of $\sim 40\%$ was found. For all channels we noticed that the number of outliers did not increase over the years (not shown in this plot) which indicates that the detector as a whole is affected and that the increase is not just due to a few strongly battered

5 pixels.

6 Summary and conclusions

The Global Ozone Monitoring Experiment, launched in April 1995 on-board the second European Remote Sensing satellite, provided measurements of atmospheric trace constituents such as O_3 , NO_2 , SO_2 , $HCHO$, BrO , H_2O as well as aerosol and cloud parameters on a global scale for more than 16 years, before it was decommissioned in July 2011. The existing data
10 archive of GOME can be considered as the European reference for follow-up atmospheric composition sensors like SCIA-MACHY, OMI, GOME-2, and the Copernicus Sentinel missions S5P/S4/S5, and preservation as well as further improvement and exploitation of this unique data set are highly recommended.

Within the framework of the ESA's GOME-Evolution project a homogenized level 1 data product for the complete mission was generated for the first time, based on the new GDP-L1 Version ~~5.05.1~~^{5.05.1}, that contains fully calibrated radiances, irradiances,
15 geolocation information, and selected calibration parameters. In addition, cloud parameters retrieved with the well-established




Figure 19. Box-Whisker plots of the distribution of the PPG correction pattern as a function of time (one selected distribution per year) for channels 1, 2, 3, and 4 (from top to bottom). Note the slightly different y-axis range for channel 1. Red horizontal lines denote the median, the blue boxes denote the lower (25%) and the upper (75%) quartile, and the green caps denote the minimum and maximum values (except the outliers).

OCRA and ROCINN algorithms have been integrated in the new product. The format and structure of the GOME L1 NetCDF-4 files is similar to other state-of-the-art EO products like S5P. Furthermore, a detailed investigation of the long-term performance of the GOME instrument in terms of monitoring the various in-flight calibration parameters was carried out. This should ensure the high quality of the GOME (ir)radiance measurements that is needed to retrieve atmospheric geophysical products with

5 highest accuracy.

The polarization correction algorithm was improved in the new GDP-L1 5.1. Instead of climatological values the ozone columns derived from the GOME measurements themselves are used for the parametrization of the Generalized Distribution Function. By means of the daily solar irradiance measurements the degradation was monitored and corrected. Degradation can be explained in terms of deposits on the GOME scan mirror. Below 300 nm intensity decreased by 80-95% which im-

plies a significant deterioration of the signal-to-noise ratio and which may have a severe impact on the challenging retrieval of atmospheric parameters such as ozone profiles. The decrease in channel 2 is 40-80%. In channel 3 the decrease (10-40%) started in 2001 ~~when~~, whereas throughput changes in channel 4 are relatively small. Since 2001 the measurements were additionally affected by an ERS-2 pointing problem. ~~Throughput changes in channel 4 are relatively small.~~ A correction algorithm

5 has been developed and further improved which relies on the intensity measured in the early part of the mission and which comprises a wavelength- and a time-dependent part. In GDP-L1 this correction is routinely applied to irradiance and radiance measurements. The degradation in reflectance, i.e. the differential degradation between solar irradiance and earth radiance measurements has been monitored for two wavelengths, 325 and 335 nm (lower and upper limit of the ozone fitting window), using cloud-free pixels over the Saharan desert. Changes are in the order of -10 to 30% and depend on wavelength and the

10 viewing angle. Since changes in reflectance may result from both changes in instrument performance or changes in atmospheric conditions, no routine corrections are applied in GDP-L1.

For the spectral calibration special attention was paid to the identification of lamp lines that remain stable (with respect to the statistical moments) over the whole mission. This has resulted in an updated spectral line list that improved the temporal stability of the wavelength assignment. For the leakage current an increase of 4 BU/s per decade ~~was found. In general, this~~

15 ~~has not necessarily a negative impact on measurement quality as long as appropriate and a widening of the distribution were~~ found. Typically, in GDP-L1 dark signal measurements from the same orbit are available and applied or a very close-by orbit are applied so that these changes do not have a negative impact on the measurement quality. The existing dark signal correction has been further improved by differentiating between measurements from outside and inside the SAA. Thereby, the enhanced background signal and noise level, which are typical for measurements from inside the SAA, are better accounted for. The

20 output of the LEDs that are used to monitor the pixel-to-pixel sensitivity decreased to about 60% of the early-mission values. For channel 2 a significant broadening of the PPG distribution was observed.

Data availability. The new GOME L1 products can be accessed via https://earth.esa.int/web/guest/news/-/asset_publisher/G2mU/content/new-ers-2-gome-level-1-v5-1-dataset-available-online.

Appendix A: NetCDF structure of new level 1 files

25 The GOME level 1 product filename is constructed as follows:

<MMM>_<CCCC>_<TTTTTTTTTT>_<instance ID>.nc,

where <MMM> is the mission ID, <CCCC> is the file class, and <TTTTTTTTTT> (= <FFFF><DDDDDD>) is a mission specific file type. <FFFF> is the file category and <DDDDDD> is a product semantic descriptor. <instance ID> consists of start time, end time, orbit number, packet version, processor version, and processing time. The packet version is a version number

30 which is specific for the combination of processor version, input data (for example calibration data) version and configuration version. For GOME Level 1 products, the mission ID is ER2 for ERS-2. The file class can be TEST, ~~REP1 or REP2~~ or RPRO

for test data or reprocessing~~1 or 2~~. The file type field contains an instrument identifier (GOM) as file category and the processing level (L1B_ or L2__ or L0__) as semantic descriptor. The packet version is 2 and the processor version is currently 5.1. We encode these versions into "02SUBSCRIPTNB051000" as the versions part of the "instance ID". The file extension is ".nc" that is typically used for netCDF files. All time strings in the filename and product are formatted in ISO 6801 format. Following this scheme, the result would be for example: ER2_TEST_GOM_L1B____20010811T032404_20010811T050712_32981_0~~12~~_0~~50000~~51000_20150311T151024.ncThe packet version is 1 and the processor version is currently 5.0. We encode these versions into "01SUBSCRIPTNB050000" as the versions part of the "instance ID". The file extension is ".nc" that is typically used for netCDF files. All time strings in the file name and product are formatted in ISO 6801 format.

Product size may vary between 60 and 75MB. Products which are measured after the ~~ERS-ERS-2~~ tape recorder problem in June 2003 are typically smaller because they don't ~~cover a complete~~ comprise measurements for the entire orbit.

The different dimensions in the GOME level 1 netCDF file are time (=1), which corresponds to ~~a~~one time per orbit, scanline (≈ 500), which corresponds to one complete scan comprising three forward and one backward scan, ground_pixel (~~=3~~ +or 1), which corresponds to the number of across-track scans, detector channel (=4), which corresponds to the number of detectors, band (=6), which corresponds to the number of spectral windows, and spectral_channel, which corresponds to the total number of detector pixels.

Figure A1 provides an overview of the netCDF structure of the level 1 file. In addition to metadata and instrument related parameters, calibration data and irradiance measurements are available. The radiance measurements themselves are organized in groups for different modes~~in which mode is one of the following types~~: nadir, static_view, narrow_swath, north- and southpo-
lar_view, sun, or moon. Forward and backward scans are separated in different groups. Inside these groups there are subgroups for bands and PMDs. A band is a part of a channel which can have its own integration time and co-adding factor. Integration times may change during one orbit. All subgroups contain several variables and attributes. For a detailed description and the complete list of all variables we refer to Aberle (2018).

Competing interests. The authors declare that they have no conflicts of interest

Acknowledgements. This work was performed in the framework of the ESA GOME-Evolution project. Particular thanks to Wolfgang Lengert, ERS-2 Mission Manager, for making this work possible. Thanks to Kai-Uwe Eichmann (IUP-B), Andreas Richter (IUP-B), Mark Weber (IUP-B), Steffen Beirle (MPIC), and Gabriele Brizzi (SERC0) for helpful feedback on the prototype version of the new GOME L1 product.




Figure A1. NetCDF structure of new GOME level 1 file [version 5.1](#). Measurements are organized in groups for different modes and bands (see text for more explanations).

References

- Aben, I., Eisinger, M., Hegels, E., and Tanzi, C.: GDAQI Final Report, Tech. rep., ESA/ESRIN, Frascati, Italy, http://wdc.dlr.de/sensors/gome/degradation_files/degradation.php, [accessed:26-July-2018], TN-GDAQI-003 SR/2000, 2000.
- Aberle, B.: GOME ERS-2 Level 1 Product User Manual, Tech. rep., German Aerospace Centre (DLR), Oberpfaffenhofen, Germany, https://earth.esa.int/documents/700255/3497594/GOME-DLR-L1-PUM_6D.pdf/f33fee60-20a8-4065-859f-e79a5b1b740b, [accessed: 26-July-2018], ER-PS-DLR-GO-0016, Issue 6/D, 2018.
- ATMOS, 2012: ATMOS 2012 Conference – Advances in Atmospheric Science and Applications: Session Summaries and Recommendations, EOEP-DTEX-EOPS-RD-12-0002, available at: http://old.esaconferencebureau.com/docs/default-source/atmos_docs2/recommendations-summaries.pdf [accessed: 26-July-2018], 2012.
- Balis, D., Lambert, J.-C., van Roozendaal, M., Spurr, R., Loyola, D., Livschitz, Y., Valks, P., Amiridis, V., Gerard, P., Granville, J., and Zehner, C.: Ten years of GOME/ERS2 total ozone data — The new GOME data processor (GDP) version 4: 2. Ground-based validation and comparisons with TOMS V7/V8, J. Geophys. Res., 112, <https://doi.org/doi:10.1029/2005JD006376>, d07307, 2007.

- Beirle, S., Lampel, J., Wang, Y., Mies, K., Dörner, S., Grossi, M., Loyola, D., Dehn, A., Danielczok, A., Schröder, M., and Wagner, T.: The ESA GOME-Evolution “Climate” water vapor product: a homogenized time series of H₂O columns from GOME, SCIAMACHY, and GOME-2, *Earth Syst. Sci. Data*, 10, 449–468, <https://doi.org/10.5194/essd-10-449-2018>, 2018.
- Bertaux, J. L., Kyrölä, E., Fussen, D., Hauchecorne, A., Dalaudier, F., Sofieva, V., Tamminen, J., Vanhellemont, F., Fanton d’Andon, O., Barrot, G., Mangin, A., Blanot, L., Lebrun, J. C., Pérot, K., Fehr, T., Saavedra, L., Leppelmeier, G. W., and Fraisse, R.: Global ozone monitoring by occultation of stars: an overview of GOMOS measurements on ENVISAT, *Atmos. Chem. Phys.*, 10, 12091–12148, <https://doi.org/10.5194/acp-10-12091-2010>, 2010.
- Bovensmann, H., Burrows, J. P., Buchwitz, M., Frerick, J., Noël, S., Rozanov, V. V., Chance, K. V., and Goede, A. P. H.: SCIAMACHY: Mission Objectives and Measurement Modes, *J. Atmos. Sci.*, 56, 127–150, [https://doi.org/10.1175/1520-0469\(1999\)056<0127:SMOAMM>2.0.CO;2](https://doi.org/10.1175/1520-0469(1999)056<0127:SMOAMM>2.0.CO;2), 1999.
- Bramstedt, K., Noël, S., Bovensmann, H., Burrows, J. P., Lerot, C., Tilstra, L., Lichtenberg, G., Dehn, A., and Fehr, T.: SCIAMACHY Monitoring Factors: Observation and End-to-End Correction of Instrument Performance Degradation, in: *Proceedings of the ESA Atmospheric Science Conference, Barcelona, Spain, 7-11 Sep 2009*, edited by SP-676, E., 2009.
- Burrows, J. P., Weber, M., Buchwitz, M., Rozanov, V. V., Ladstädter-Weissenmayer, A., Richter, A., de Beek, R., Hoogen, R., Bramstedt, K., Eichmann, K.-U., Eisinger, M., and Perner, D.: The Global Ozone Monitoring Experiment (GOME): Mission Concept and First Scientific Results, *J. Atmos. Sci.*, 56, 151–175, [https://doi.org/10.1175/1520-0469\(1999\)056<0151:TGOMEG>2.0.CO;2](https://doi.org/10.1175/1520-0469(1999)056<0151:TGOMEG>2.0.CO;2), 1999.
- Cai, Z., Liu, Y., Liu, X., Chance, K., Nowlan, C., Lang, R., Munro, R., and Suleiman, R.: Characterization and correction of Global Ozone Monitoring Experiment 2 ultraviolet measurements and application to ozone profile retrievals, *J. Geophys. Res.*, 117, <https://doi.org/10.1029/2011JD017096>, D07305, 2012.
- Casadio, S. and Arino, O.: Monitoring the South Atlantic Anomaly using ATSR instrument series, *Adv. Space Res.*, 48, 1056–1066, <https://doi.org/https://doi.org/10.1016/j.asr.2011.05.014>, 2011.
- CEOS Pseudo Invariant Calibration Sites: https://calval.cr.usgs.gov/rst-resources/sites_catalog/ceos-reference-sites/, [accessed: 26-July-2018].
- Coldewey-Egbers, M., Slijkhuis, S., Aberle, B., and Loyola, D.: Long-term analysis of GOME in-flight calibration parameters and instrument degradation, *Appl. Opt.*, 47, 4749–4761, <https://doi.org/10.1364/AO.47.004749>, 2008.
- Coldewey-Egbers, M., Loyola, D. G., Koukouli, M., Balis, D., Lambert, J.-C., Verhoelst, T., Granville, J., van Roozendael, M., Lerot, C., Spurr, R., Frith, S. M., and Zehner, C.: The GOME-type Total Ozone Essential Climate Variable (GTO-ECV) data record from the ESA Climate Change Initiative, *Atmos. Meas. Tech.*, 8, 3923–3940, <https://doi.org/10.5194/amt-8-3923-2015>, 2015.
- de Smedt, I., Müller, J.-F., Stavrakou, T., van der A, R., Eskes, H., and van Roozendael, M.: Twelve years of global observations of formaldehyde in the troposphere using GOME and SCIAMACHY sensors, *Atmos. Chem. Phys.*, 8, 4947–4963, <https://doi.org/10.5194/acp-8-4947-2008>, 2008.
- Dehn, A.: Long Term Monitoring of GOME Diffuser Reflectivity and Dark Signal Analysis, Tech. rep., ESA-ESRIN, ERSE-SPPA-EOPG-TN-03-0010, Issue 1.0, available at: <https://earth.esa.int/documents/10174/1596664/Gome28.pdf> [accessed: 26-July-2018], 2003.
- Dobber, M. R.: GOME moon measurements, including instrument characterisation and moon albedo, in: *3rd ERS Scientific Symposium*, Florence, Italy, ESA SP-414 Vol.II, p. 743, 1997.
- Dobber, M. R., Goede, A. P. H., and Burrows, J. P.: Observations of the moon by the global ozone monitoring experiment: radiometric calibration and lunar albedo, *Appl. Opt.*, 37, 7832–7841, <https://doi.org/10.1364/AO.37.007832>, 1998.

- Dobber, M. R., Dirksen, R. J., Levelt, P. F., van den Oord, G. H. J., Voors, R. H. M., Kleipool, Q., Jaross, G., Kowalewski, M., Hilsenrath, E., Leppelmeier, G. W., de Vries, J., Dierssen, W., and Rozemeijer, N. C.: Ozone Monitoring Instrument Calibration, *IEEE Trans. Geosc. Rem. Sens.*, 44, 1209–1238, <https://doi.org/10.1109/TGRS.2006.869987>, 2006.
- Garane, K., Lerot, C., Coldewey-Egbers, M., Verhoelst, T., Koukouli, M. E., Zyrichidou, I., Balis, D. S., Danckaert, T., Goutail, F., Granville, J., Hubert, D., Keppens, A., Lambert, J.-C., Loyola, D., Pommereau, J.-P., Van Roozendaal, M., and Zehner, C.: Quality assessment of the Ozone_cci Climate Research Data Package (release 2017) – Part 1: Ground-based validation of total ozone column data products, *Atmos. Meas. Tech.*, 11, 1385–1402, <https://doi.org/10.5194/amt-11-1385-2018>, 2018.
- GCOS: Systematic Observation Requirements for Satellite-Based Data Products for Climate - 2011 Update, <http://www.wmo.int/pages/prog/gcos/Publications/gcos-154.pdf>, [accessed: 26-July-2018], 2011.
- 10 GOME Mission Operations Overview: <https://earth.esa.int/web/sppa/mission-performance/esa-missions/ers-2/gome/mission-operations-overview>, [accessed: 26-July-2018].
GOME Users Manual: Tech. rep., ESA/ESTEC, Noordwijk, The Netherlands, ESA SP-1182, available at: <https://earth.esa.int/documents/10174/1596664/GOME05.pdf> [accessed: 26-July-2018], 1995.
GOME Web Gallery: <https://earth.esa.int/web/sppa/mission-performance/mission-web-galleries/uvn-galleries>, [accessed: 26-July-2018].
- 15 Hegels, E., Aben, I., Slijkhuis, S., and Tanzi, C. P.: GOME Degradation Correction at Solar Activity Maximum, in: *Proceedings of the ERS-ENVISAT Symposium*, edited by SP-461, E. S. P., CD-ROM, ISBN 92-9092-685-6, 2001.
Heirtzler, J. R.: The future of the South Atlantic anomaly and implications for radiation damage in space, *Journal of Atmospheric and Solar-Terrestrial Physics*, 64, 1701–1708, [https://doi.org/https://doi.org/10.1016/S1364-6826\(02\)00120-7](https://doi.org/https://doi.org/10.1016/S1364-6826(02)00120-7), 2002.
- Helder, D. L., Basnet, B., and Morstad, D. L.: Optimized identification of worldwide radiometric pseudo-invariant calibration sites, *Canadian Journal of Remote Sensing*, 36, 527–539, <https://doi.org/10.5589/m10-085>, 2010.
- 20 Heue, K.-P., Coldewey-Egbers, M., Delcloo, A., Lerot, C., Loyola, D., Valks, P., and van Roozendaal, M.: Trends of tropical tropospheric ozone from 20 years of European satellite measurements and perspectives for the Sentinel-5 Precursor, *Atmos. Meas. Tech.*, 9, 5037–5051, <https://doi.org/10.5194/amt-9-5037-2016>, 2016.
- Lerot, C., van Roozendaal, M., Spurr, R., Loyola, D., Coldewey-Egbers, M., Kochenova, S., van Gent, J., Koukouli, M., Balis, D., Lambert, J.-C., Granville, J., and Zehner, C.: Homogenized total ozone data records from the European sensors GOME/ERS-2, SCIAMACHY/Envisat, and GOME-2/MetOp-A, *J. Geophys. Res.*, 119, 1639–1662, <https://doi.org/10.1002/2013JD020831>, 2014.
- 25 Levelt, P., van den Oord, G., Dobber, M., Malkki, A., Visser, H., de Vries, J., Stammes, P., Lundell, J., and Saari, H.: The Ozone Monitoring Instrument, *IEEE Transactions on Geoscience and Remote Sensing*, 44, 1093–1101, <https://doi.org/10.1109/TGRS.2006.872333>, 2006.
- Liu, X., Chance, K., and Kurosu, T. P.: Improved ozone profile retrievals from GOME data with degradation correction in reflectance, *Atmos. Chem. Phys.*, 7, 1575–1583, <https://doi.org/10.5194/acp-7-1575-2007>, 2007.
- 30 Lorente, A., Boersma, K. F., Stammes, P., Tilstra, L. G., Richter, A., Yu, H., Kharbouche, S., and Muller, J.-P.: The importance of surface reflectance anisotropy for cloud and NO₂ retrievals from GOME-2 and OMI, *Atmos. Meas. Tech.*, 11, 4509–4529, <https://doi.org/10.5194/amt-11-4509-2018>, 2018.
- Loyola, D. G., Balzer, W., Aberle, B., Bittner, M., Kretschel, K., Muehle, H., Ruppert, T., Schmid, C., Slijkhuis, S., Spurr, R., Thomas, W., Wieland, T., and Wolfmueller, M.: Ground Segment for ERS-2/GOME Sensor at the German D-PAF, in: *3rd ERS Scientific Symposium*, Florence, Italy, ESA SP-414 Vol.II, pp. 591–596, 1997.
- 35 Loyola, D. G., Thomas, W., Spurr, R., and Mayer, B.: Global patterns in daytime cloud properties derived from GOME backscatter UV-VIS measurements, *Int. J. Remote Sensing*, 31, 4295–4318, <https://doi.org/10.1080/01431160903246741>, 2010.

- Loyola, D. G., Koukouli, M. E., Valks, P., Balis, D. S., Hao, N., van Roozendaal, M., Spurr, R. J. D., Zimmer, W., Kiemle, S., Lerot, C., , and Lambert, J.-C.: The GOME-2 total column ozone product: Retrieval algorithm and ground-based validation, *J. Geophys. Res.*, 116, <https://doi.org/10.1029/2010JD014675>, 2011.
- Loyola, D. G., Gimeno García, S., Lutz, R., Argyrouli, A., Romahn, F., Spurr, R. J. D., Pedergrana, M., Doicu, A., Molina García, V.,
5 and Schüssler, O.: The operational cloud retrieval algorithms from TROPOMI on board Sentinel-5 Precursor, *Atmos. Meas. Tech.*, 11, 409–427, <https://doi.org/10.5194/amt-11-409-2018>, 2018.
- Lutz, R., Loyola, D., Gimeno García, S., and Romahn, F.: OCRA radiometric cloud fractions for GOME-2 on MetOp-A/B, *Atmos. Meas. Tech.*, 9, 2357–2379, <https://doi.org/10.5194/amt-9-2357-2016>, 2016.
- Mishra, N., Helder, D., Angal, A., Choi, J., and Xiong, X.: Absolute Calibration of Optical Satellite Sensors Using Libya 4 Pseudo Invariant
10 Calibration Site, *Remote Sensing*, 6, 1327–1346, <https://doi.org/10.3390/rs6021327>, 2014.
- Mount, G. H., Sanders, R. W., and Brault, J. W.: Interference effects in reticon photodiode array detectors, *Appl. Opt.*, 31, 851–858, <https://doi.org/10.1364/AO.31.000851>, 1992.
- Munro, R., Lang, R., Klaes, D., Poli, G., Retscher, C., Lindstrot, R., Huckle, R., Lacan, A., Grzegorski, M., Holdak, A., Kokhanovsky, A.,
Livschitz, J., and Eisinger, M.: The GOME-2 instrument on the Metop series of satellites: instrument design, calibration, and level 1 data
15 processing – an overview, *Atmos. Meas. Tech.*, 9, 1279–1301, <https://doi.org/10.5194/amt-9-1279-2016>, 2016.
- Murray, J. E.: Atlas of the spectrum of a platinum/chromium/neon hollow-cathode reference lamp in the region 240–790 nm, Final Report to
ESA, John Wheaton Associates, 1994.
- Noël, S., Bramstedt, K., Bovensmann, H., Burrows, J., Gottwald, M., and Krieg, E.: SCIAMACHY Degradation Monitoring Results, in:
Proceedings of the ENVISAT Symposium, 23-27 April 2007, Montreux, Switzerland, edited by SP-636, E. S. P., 2007.
- Pan, C., Weng, F., Beck, T., Flynn, L., and Ding, S.: Recent Improvements to Suomi NPP Ozone Mapper Profiler Suite Nadir Mapper Sensor
20 Data Records, *IEEE Transactions on Geoscience and Remote Sensing*, 55, 5770–5776, <https://doi.org/10.1109/TGRS.2017.2714103>, 2017.
- Press, W. H., Teukolsky, S. A., and Vetterling, W. T.: Numerical Recipes in Fortran 77, Cambridge University Press, 1992.
- Schenkeveld, V. M. E., Jaross, G., Marchenko, S., Haffner, D., Kleipool, Q. L., Rozemeijer, N. C., Veefkind, J. P., and Levelt, P. F.: In-flight
25 performance of the Ozone Monitoring Instrument, *Atmos. Meas. Tech.*, 10, 1957–1986, <https://doi.org/10.5194/amt-10-1957-2017>, 2017.
- Schutgens, N. A. J. and Stammes, P.: Parametrisation of Earth’s polarisation spectrum in the ultra-violet, *J. Quant. Spectrosc. Rad. Transfer*,
75, 239–255, [https://doi.org/10.1016/S0022-4073\(01\)00248-5](https://doi.org/10.1016/S0022-4073(01)00248-5), 2002.
- Slijkhuis, S.: CHEOPS-GOME– Algorithm Theoretical Basis Document – Level 0 to 1 Processing Update, Tech.
rep., German Aerospace Centre (DLR), Oberpfaffenhofen, Germany, CH-TN-DLR-GO-0003, Issue 1/B, available at:
30 https://wdc.dlr.de/sensors/gome/degradation_files/literature/atbd_cheops.pdf [accessed: 26-July-2018], 2006.
- Slijkhuis, S. and Aberle, B.: GOME/ERS-2 Level 0 to 1b ATBD, Tech. rep., German Aerospace Centre (DLR), Oberpfaffenhofen, Ger-
many, https://earth.esa.int/documents/700255/3497594/GOME-DLR-L1-ATBD_V7A.pdf/c8c4a130-3fd9-4f98-b6d9-5859689f52ed,
[accessed:26-July-2018], GOME-DLR-L1-ATBD, Issue 7.0, 2016.
- Slijkhuis, S., Aberle, B., and Loyola, D.: Improvements of GDP Level 0-1 processing system in the framework of
35 CHEOPS-GOME, in: Proceedings of the Atmospheric Science Conference, edited by ESA Spec. Publ. SP-628, available at:
http://earth.esa.int/workshops/atmos2006/participants/516/paper__paper_ASW_cheops_final.pdf [accessed: 26-July-2018], 2006.
- Smith, D. L. and Cox, C. V.: (A)ATSR Solar Channel On-Orbit Radiometric Calibration, *IEEE Transactions on Geoscience and Remote
Sensing*, 51, 1370–1382, <https://doi.org/10.1109/TGRS.2012.2230333>, 2013.

- Snel, R.: In-orbit optical path degradation: GOME experience and SCIAMACHY prediction, in: Proceedings of the ERS-ENVISAT Symposium, edited by ESA Spec. Publ. SP-461, CD-ROM, ISBN 92-9092-685-6, 2001.
- Sun, J., Xiong, X., Angal, A., Chen, H., Wu, A., and Geng, X.: Time-Dependent Response Versus Scan Angle for MODIS Reflective Solar Bands, *IEEE Transactions on Geoscience and Remote Sensing*, 52, 3159–3174, <https://doi.org/10.1109/TGRS.2013.2271448>, 2014.
- 5 Tanzi, C., Snel, R., Hasekamp, O., and Aben, I.: Degradation of UV Earth Albedo Observations by GOME, in: Proceedings of the ERS-ENVISAT Symposium, edited by SP-461, E. S. P., CD-ROM, ISBN 92-9092-685-6, 2001.
- Tilstra, L. G., de Graaf, M., Aben, I., and Stammes, P.: In-flight degradation correction of SCIAMACHY UV reflectances and Absorbing Aerosol Index, *J. Geophys. Res.*, 117, <https://doi.org/10.1029/2011JD016957>, D06209, 2012.
- Upreti, S. and Cao, C.: Suomi NPP VIIRS reflective solar band on-orbit radiometric stability and accuracy assessment using desert and
10 Antarctica Dome C sites, *Remote Sensing of Environment*, 166, 106–115, <https://doi.org/10.1016/j.rse.2015.05.021>, 2015.
- van der A, R. J., van Oss, R. F., Peters, A. J. M., Fortuin, J. P. F., Meijer, Y. J., and Kelder, H. M.: Ozone profile retrieval from recalibrated Global Ozone Monitoring Experiment data, *J. Geophys. Res.*, 107, ACH 2–1–ACH 2–10, <https://doi.org/10.1029/2001JD000696>, 2002.
- van Geffen, J. H. G. M.: Wavelength calibration of spectra measured by the Global Ozone Monitoring Experiment: variations along orbits and in time, *Appl. Opt.*, 43, 695–705, <https://doi.org/10.1364/AO.43.000695>, 2004.
- 15 van Geffen, J. H. G. M. and van Oss, R. F.: Wavelength calibration of spectra measured by the Global Ozone Monitoring Experiment by use of a high-resolution reference spectrum, *Appl. Opt.*, 42, 2739–2753, <https://doi.org/10.1364/AO.42.002739>, 2003.
- van Geffen, J. H. G. M., Boersma, K. F., van Roozendaal, M., Hendrick, F., Mahieu, E., de Smedt, I., Sneep, M., and Veeffkind, J. P.: Improved spectral fitting of nitrogen dioxide from OMI in the 405–465 nm window, *Atmos. Meas. Tech.*, 8, 1685–1699, <https://doi.org/10.5194/amt-8-1685-2015>, 2015.
- 20 van Peet, J. C. A., van der A, R. J., Tuinder, O. N. E., Wolfram, E., Salvador, J., Levelt, P. F., and Kelder, H. M.: Ozone Profile Retrieval Algorithm (OPERA) for nadir-looking satellite instruments in the UV–VIS, *Atmos. Meas. Tech.*, 7, 859–876, <https://doi.org/10.5194/amt-7-859-2014>, 2014.
- van Roozendaal, M., Spurr, R., Loyola, D., Lerot, C., Balis, D., Lambert, J.-C., Zimmer, W., van Gent, J., van Geffen, J., Koukouli, M., Granville, J., Doicu, A., Fayt, C., and Zehner, C.: Sixteen years of GOME/ERS-2 total ozone data: The new direct-fitting GOME Data
25 Processor (GDP) version 5 — Algorithm description, *J. Geophys. Res.*, 117, <https://doi.org/10.1029/2011JD016471>, D03305, 2012.
- Voors, R., Dobber, M., Dirksen, R., and Levelt, P.: Method of calibration to correct for cloud-induced wavelength shifts in the Aura satellite’s Ozone Monitoring Instrument, *Appl. Opt.*, 45, 3652–3658, <https://doi.org/10.1364/AO.45.003652>, 2006.
- Zoogman, P., Liu, X., Chance, K., Sun, Q., Schaaf, C., Mahr, T., and Wagner, T.: A climatology of visible surface reflectance spectra, *J. Quant. Spectrosc. Rad. Transfer*, 180, 39–46, <https://doi.org/10.1016/j.jqrst.2016.04.003>, 2016.