



Development of an incoherent broadband cavity-enhanced absorption spectrometer for measurements of ambient glyoxal and NO₂ in a polluted urban environment

Shuaixi Liang¹², Min Qin¹, Pinhua Xie¹²³, Jun Duan¹, Wu Fang¹, Yabai He¹, Jin Xu¹, Ke Tang¹², Fanhao Meng¹², Kaidi Ye¹²,
5 Jianguo Liu¹²³, and Wenqing Liu¹²³

¹ Key Laboratory of Environmental Optics and Technology, Anhui Institute of Optics and Fine Mechanics, Chinese Academy of Sciences, Hefei 230031, China

² University of Science and Technology of China, Hefei 230026, China

³ CAS Center for Excellence in Regional Atmospheric Environment, Institute of Urban Environment, Chinese Academy of Sciences, Xiamen, 361021, China

Correspondence to: Min Qin (mqin@aiofm.ac.cn); Pinhua Xie (phxie@aiofm.ac.cn)

Abstract: We report the development of an instrument for simultaneous fast measurements of glyoxal (CHOCHO) and NO₂ based on incoherent broadband cavity-enhanced absorption spectroscopy (IBBCEAS) in the 438 – 465 nm wavelength region. The highly-reflective cavity mirrors were protected from contamination by N₂ purge gas. The reduction of the effective cavity
15 length was calibrated by measuring collision-induced oxygen absorption at ~477 nm of pure oxygen gas input with and without the N₂ mirror purge gas. The detection limits of the developed system were evaluated to be 23 parts per trillion by volume (pptv, 2σ) for CHOCHO and 29 pptv (2σ) for NO₂ with a 30-s acquisition time, respectively. A potential cross-interference of NO₂ absorption on accurate CHOCHO measurements has been investigated in this study, as the absorption of NO₂ in the atmosphere could often be several hundred-fold higher than that of glyoxal, especially in contaminated areas. Due to nonlinear
20 spectrometer dispersion, simulation spectra of NO₂ based on traditional convolution-simulation did not match the measurement spectra well enough. In this work, we applied actual NO₂ spectral profile measured by the same spectrometer as a reference spectral profile in subsequent atmospheric spectral analysis and retrieval of NO₂ and CHOCHO concentrations. This effectively reduced the spectral fitting residuals. The instrument was successfully deployed for 24 days of continuous measurements of CHOCHO and NO₂ in atmosphere in a comprehensive field campaign in Beijing in June 2017.

1. Introduction

Glyoxal (CHOCHO) is a typical intermediate for most volatile organic compounds (VOC) oxidations in the atmosphere. It plays an important role in quantifying VOC emissions, understanding VOC oxidation mechanisms, and further understanding the formation of O₃ and secondary organic aerosol (SOA). On a global scale, simulations show that biogenic isoprene is the
30 largest source of glyoxal (47% of total contributions); anthropogenic acetylene also contributes significantly to glyoxal (20% of contributions) (Fu et al., 2008). The loss of glyoxal is mainly due to photolysis, OH and NO₃ oxidation reactions, wet and dry deposition, and irreversible absorption of water-soluble aerosols and clouds (Fu et al., 2008; Min et al., 2016). The ratio of glyoxal to formaldehyde, R_{GF} , is often used as an indicator of hydrocarbon precursor speciation in contaminated areas; observations in the field can give divergent conclusions (Vrekoussis et al., 2010; Kaiser et al., 2015; DiGangi et al., 2012).
35 Glyoxal readily undergoes heterogeneous reactions to form SOA, but the contribution to SOA has a high uncertainty (Li et al., 2016; Washenfelder et al., 2011; Volkamer et al., 2007). Therefore, accurate quantification of glyoxal is a prerequisite for studies of the source, sink, and atmospheric chemistry of glyoxal.

Several technologies are currently used for measurements of glyoxal in the atmosphere, including chemical and spectroscopic methods. The common wet chemistry method is based on a derivatization reagent such as agent o-(2,3,4,5,6-
40 pentafluorobenzyl) hydroxylamine (PFBHA), 2,4-dinitrophenylhydrazine (DNPH), or pentafluorophenyl hydrazine (PFPH),



with subsequent analysis using liquid chromatography or mass spectrometry techniques (Temime et al., 2007; Ho et al., 2004; Munger et al., 1995; Pang et al., 2014). Some successful spectroscopic techniques for glyoxal include differential optical absorption spectroscopy (DOAS), laser-induced fluorescence (LIF), and incoherent broadband cavity enhanced absorption spectroscopy (IBBCEAS). Long-path DOAS (LP-DOAS) was used to measure the glyoxal concentration for the first time at
5 a total atmospheric light-path of 4420 m with a detection limit of 0.1 parts per billion by volume (ppbv, 2σ) in Mexico City (Volkamer et al., 2005a). In 2008, LP-DOAS was used to measure glyoxal above the rainforest and then compared with multi-axis DOAS (Max-DOAS), suggesting that local CHOCHO was confined to the first 500 m of the boundary layer (MacDonald et al., 2012). LIF can quantify both glyoxal and methylglyoxal with a detection limit of 2.9 pptv (2σ) in 5 min for glyoxal (Henry et al., 2012). IBBCEAS is an excellent method for measuring atmospheric trace gases. It features high sensitivity, small
10 chemical interference, and simultaneous measurement of multiple components. IBBCEAS has been rapidly developed since Fiedler et al. first described it in 2003 (Fiedler et al. 2003). More recently, the technology has been successfully applied to measure a variety of trace gases (Min et al., 2016; Wang et al., 2017; Yi et al., 2016; Volkamer et al., 2015), weakly absorbed cross-sections of different trace gases (Chen et al., 2011; Kahan et al., 2012) and aerosol extinction (Washenfelder et al., 2013). Using a xenon arc lamp as a light source, Washenfelder et al. reported first measurement of glyoxal using the IBBCEAS
15 technique in the laboratory with a detection limit of 58 pptv (2σ) within 1 min (Washenfelder et al., 2008). Later, Thalman et al. coupled CEAS hardware with a DOAS retrieval algorithm to measure glyoxal in open cavity mode with a detection limit of 19 pptv (2σ , 1 min) (Thalman et al., 2010). Coburn et al. subsequently measured the eddy covariance flux of glyoxal with LED-CE-DOAS for the first time and found that the nocturnal oxidation reaction on an ocean surface organic microlayer was a source of the oxygenated VOCs (Coburn et al., 2014). With significant improvements, Min et al. developed an aircraft
20 IBBCEAS instrument and used it to measure tropospheric glyoxal with a detection limit of 34 pptv (2σ) within 5 s (Min et al., 2016). Table 1 compares different measurement techniques for glyoxal. Based on these technologies, Thalman et al. conducted a comprehensive instrument inter-comparison campaign for glyoxal (Thalman et al., 2015).

Spectral measurement techniques using broadband light sources, such as DOAS and IBBCEAS, can simultaneously observe a wide range of spectral bands during a single measurement. Thus, many contaminants can be measured concurrently.
25 The overlap of the NO_2 and glyoxal absorption bands at 438–465 nm allows us to simultaneously measure their concentrations. However, NO_2 can interfere with the measurement of glyoxal, especially for high concentration of NO_2 (Thalman et al., 2015). This is a key factor that needs to be considered to improve data retrieval of glyoxal in China's highly polluted environment.

Here, we describe the development of an incoherent broadband cavity-enhanced absorption spectrometer for sensitive detection of CHOCHO and NO_2 in the atmosphere. The effective length of the optical cavity with purge-gas protected mirrors
30 was accurately calibrated based on the collision-induced oxygen (O_4) absorption at 477 nm. The instrument detection limit was estimated using the Allan variance analysis. The effects of NO_2 on glyoxal were evaluated via spectral simulation and measurements. The results show that using the measured NO_2 reference spectrum can overcome NO_2 interference to glyoxal due to conventional convolution methods from the uneven dispersion of the grating spectrometer. We then applied the measured reference spectrum to the retrieval of glyoxal in the same wavelength band and obtained the glyoxal concentration
35 in heavily polluted air in China. The IBBCEAS instrument was successfully deployed during the APHH-China (Air Pollution and Human Health in a Chinese Megacity) project, and we obtained the profiles of glyoxal and NO_2 concentrations in Beijing's summer atmosphere during the APHH-China campaign (2–26 June 2017).

Table 1. Comparison of different techniques for measuring glyoxal



Analytical technique	Research Unit	Time resolution	Detection Limit (2σ)	Field Applications	Purge flows	Ref.
Microfluidic	University of York	30 min	53 pptv	No	/	Pang et al. (2014)
Mad LIP	University of Wisconsin-Madison	1 min	12 pptv	USA	/	Huisman et al., (2008)
LIP	University of Wisconsin-Madison	5 min	2.9 pptv	No	/	Henry et al., (2012)
LP-DOAS	Massachusetts Institute of Technology	2-15 min	0.1 ppbv	Mexico City	/	Volkamer et al., (2005)
IBBCEAS	University of Colorado	1 min	58 pptv	Laboratory	no	Washenfelder et al. (2008)
LED-CE-DOAS	University of Colorado	1 min	19 pptv	Laboratory	Yes	Thalman et al., (2010)
ACES	University of Colorado & NOAA	5 s	34 pptv	USA & China	no	Min et al., (2016)
IBBCEAS	Anhui Institute of Optics and Fine Mechanics, CAS	30 s	23 pptv	China	Yes	This work

2. System and principle

2.1 Description of the IBBCEAS setup

The IBBCEAS technology is an absorption spectroscopy technique. It improves the effective path length via multiple light reflections in an optical cavity. This leads to a significant improvement of the detection sensitivity. Our design of the IBBCEAS setup consists of a light-emitting diode (LED) light source, a pair of off-axis parabolic mirrors, a pair of high-reflectivity cavity mirrors, Teflon perfluoroalkoxy polymer resin (PFA) optical cavity, optical band-pass filter, an optical fiber-coupled grating spectrometer and some other components. A schematic diagram of the instrument is shown in Fig. 1.

The light from a high-power blue LED (LedEngin) with a peak wavelength of ~ 448 nm was coupled to the optical cavity via a 90° off-axis parabolic mirror (Edmund Optics). The temperature of the LED was measured by a temperature sensor (PT1000) and controlled by a thermoelectric cooler (TEC) at $20^\circ\text{C} \pm 0.1^\circ\text{C}$ to reduce the impact of temperature fluctuations on the LED. Multiple reflections of light between two high-reflectivity cavity mirrors (Advanced Thin Films) increased the effective absorption path length. The light exiting the cavity passed through an optical band-pass filter (Thorlabs FB450-40) to eliminate stray light. It was then focused onto a 1-m optical fiber (600 μm in diameter with a numerical aperture of 0.22) by a second off-axis parabolic mirror. Finally, the other end of the fiber cable was coupled to a compact Czerny-Turner spectrometer (Ocean Optics, QE65000) with a spectral resolution of ~ 0.57 nm around 450 nm. A 2- μm teflon polytetrafluoroethylene (PTFE) membrane filter (Tisch Scientific) was used in the front of the inlet to remove aerosols—this reduced scattering losses by particulate matter and its impacts on the effective path length (Thalman et al., 2010). The cavity mirrors were purged with a constant flow of dry nitrogen at a rate of 0.2 sL min^{-1} (standard liters per minute) to block their contact with air samples inside the cavity. This ensured cleanness of the cavity mirror throughout the experiment. A mass flow controller maintained a constant combined sample and purge gas flow rate of 1.2 sL min^{-1} , which resulted in a gas residence time of about 16 s in the optical cavity.




Fig. 1. Schematic of the incoherent broadband cavity enhanced absorption spectrometer.

2.2 Theory of IBCEAS

The total extinction in the optical cavity includes the absorption by trace gases, Rayleigh scattering by gas molecules, and Mie scattering by particles. The use of a filter in air sampling pipeline removes the particles. The Rayleigh scattering extinction of pure N_2 is about $2.5 \times 10^{-7} \text{ cm}^{-1}$ at 455 nm, which is comparable to the cavity loss $\sim 8.1 \times 10^{-7} \text{ cm}^{-1}$ based on mirror reflectivity and cavity length (i.e. $(1 - R)/d$). Thus, the general description of the total optical extinction α_{abs} within the optical cavity is (Washenfelter et al., 2008):

$$\alpha_{abs}(\lambda) = \left(\frac{1 - R(\lambda)}{d_{eff}} + \alpha_{Ray}(\lambda) \right) \left(\frac{I_0(\lambda) - I(\lambda)}{I(\lambda)} \right), \quad (1)$$

where $R(\lambda)$ is the wavelength-dependent reflectivity of the cavity mirrors, $\alpha_{Ray}(\lambda)$ is the extinction for Rayleigh scattering, $I_0(\lambda)$ and $I(\lambda)$ are the light intensities transmitted through the optical cavity without and with the absorbing species, respectively, and d_{eff} is the effective cavity length. The mirror reflectivity $R(\lambda)$ is determined from the Rayleigh scattering of N_2 and He via the following equation (Washenfelter et al., 2008):

$$R(\lambda) = 1 - \frac{\frac{I_{N_2}(\lambda)}{I_{He}(\lambda)} \cdot \alpha_{Ray}^{N_2}(\lambda) d_0 - \alpha_{Ray}^{He}(\lambda) d_0}{1 - \frac{I_{N_2}(\lambda)}{I_{He}(\lambda)}}, \quad (2)$$

Here, $I_{N_2}(\lambda)$ and $I_{He}(\lambda)$ are the light intensities measured when the cavity is filled with N_2 and He, respectively. Term $\alpha_{Ray}^{N_2}(\lambda)$ and $\alpha_{Ray}^{He}(\lambda)$ are the extinction caused by Rayleigh scatterings of N_2 and He, respectively. Term d_0 is the distance between the two cavity mirrors. Terms d_0 and d_{eff} are not equal due to cavity mirror purging. Determination of the d_{eff} will be described in the Section 3.2. After obtaining the mirror reflectivity $R(\lambda)$, the absorption coefficient α_{abs} can be calculated according to Eq. (1). If the chamber contains a variety of gas absorbers (including NO_2 and $CHOCHO$), then the absorption coefficient α_{abs} will be the sum of their individual contributions and can be written via the following equation:

$$\alpha_{abs} = \sum_i^n \alpha_i(\lambda) = \sum_i^n \sigma_i N_i = \sigma_{NO_2}(\lambda) [NO_2] + \sigma_{CHOCHO}(\lambda) [CHOCHO] + \dots, \quad (3)$$



Here, $\sigma_i(\lambda)$ and N_i are the absorption cross-section and number density for the i -th trace absorber, and n is the total number of absorbers. Finally, the absorber concentrations can be retrieved from the measured broadband spectrum via a nonlinear least square fitting.

3. Results and Analysis

5 3.1 Determination of the cavity mirror reflectivity

The cavity mirror reflectivity needs to be accurately determined for subsequent measurements of the concentrations of trace gases inside the cavity. Using the difference of Rayleigh scattering cross-sections between N_2 and He, we calculated the mirror reflectivity $R(\lambda)$ according to Eq. (2). The values of $\alpha_{Ray}^{N_2}(\lambda)$ and $\alpha_{Ray}^{He}(\lambda)$ were taken from published references (Shardanand et al., 1977; Snee et al., 2005). The black and red curves in Fig. 2 were the spectrometer's signal intensity when the cavity was filled with high purity N_2 (99.999%) and He (99.999%), respectively. The difference in light intensity due to Rayleigh scattering by N_2 versus He is well visible. The shaded spectral region (438 nm – 465 nm) indicated in the figure contains the main absorption peak of glyoxal and is of primary interest for its spectral retrieval. The mirror reflectivity at the maximum absorption position of glyoxal (455 nm) is greater than 0.99994. The cross-sections were obtained by convolving the high-resolution literature cross-sections of CHOCHO (Volkamer et al., 2005b), NO_2 (Voigt et al., 2002) and H_2O with the nominal spectrometers's instrument function of 0.57 nm full width at half maximum (FWHM). The H_2O absorption cross section was calculated with the SpectraPlot program based on the HITRAN2012 database.




Fig. 2. (a) Calibration of the mirror reflectivity. The black and red curves represent spectrometer's CCD traces of nitrogen and helium, respectively, with a spectral acquisition time of 30 s. The blue line is the resulting mirror reflectivity curve. (b) The green, magenta and black lines are convolution-based literature absorption cross-sections of NO_2 , glyoxal and H_2O vapor, respectively.

20 3.2 Calibration of the effective cavity length

Considering the intended application's environmental conditions of high-load particulate matter and high-concentration polluting gases, we used an aerosol filter to reduce particles entering the optical cavity and purged the immediate space in front of the cavity mirrors with pure N_2 gas to keep the cavity mirrors clean (see Fig. 1). This purging made it difficult to accurately measure the effective cavity length. However, the effective cavity length is required for retrieving trace gas concentrations. Here, we utilized the collision-induced oxygen absorption (referred as O_2 - O_2 or O_4 absorption) (Thalman et al., 2013) at ~477



nm within our operation wavelength region to quantify the effective cavity length. Pure O₂ gas was introduced into the optical cavity and the O₂-O₂ 477-nm absorptions with and without the N₂ mirror purges were then measured. The O₂ flow rate was 1 sL min⁻¹ and the total N₂ purge flow rate was 0.2 sL min⁻¹. Figures 3a and 3b show an example of O₂-O₂ measurement spectrum, its model fitting, and the fit residuals. Figure 3c shows the time series of equivalent O₂ concentrations when N₂ mirror purge gas was alternated between On and Off. A coarse estimation for the cavity length reduction factor was calculated to be 0.87 at

$$d_{eff_O_4_based} = d_0 \times \frac{[\sqrt{O_4 \text{ Signal}}]_{Purge\ on}}{[\sqrt{O_4 \text{ Signal}}]_{Purge\ off}}, \quad (4)$$

Here, the O₄ signals were based on the absorption measurements with and without the N₂ purge flows, respectively. Furthermore, we modelled the reduction factor of the effective cavity length due to purge gas to include the effect of the dilution of sample gases by purge gases inside the cavity and the fact that the measured O₄ spectra were proportional to the product of [O₂]×[O₂] concentrations. According to the simulation results, if N₂ purge gases distributed evenly to both ends of the cavity and 50% of the total purge N₂ was involved in the dilution of O₂, the reduction factor for linear absorption process was 0.841, which was 3.3% less than the coarse estimation value of 0.87. An uncertainty of the purge N₂ participating in the O₂ mixture at 40% or 60% could cause a ~2% uncertainty in the cavity length reduction factor. In this experiment, d₀ = 70.0 cm and the calculated d_{eff} was 58.9 cm.




Fig. 3. (a) Example of retrieved and fitted absorption spectrum of O₄. The blue line is the measured spectrum and the red line is the fitted spectrum of O₄; (b) Fit residuals; (c) The time series of equivalent O₂ concentrations when N₂ purge gas was alternated between On and Off.

20 3.3 Instrument stability and detection limit

The stability of the system affects its detection sensitivity. An ideal stable system can theoretically achieve an extremely high sensitivity by averaging measurements over a long period of time. However, there are practical considerations that limit this to a certain time range (Werle et al., 1993). For an IBBCEAS system, its stability is mainly affected by the mechanical drifts of the system and the change in the intensity and central wavelength emission of the light source due to temperature variations.

25 We used two methods to describe the performance of the system: distribution analysis and Allan variance analysis of a large number of measurements. For more than 8 hours, 10,000 spectra were continuously acquired with the optical cavity filled with dry nitrogen. As the cavity was free of any NO₂ and CHOCHO, these measurements reveal the fluctuations around zero concentration. The acquisition time of each spectrum was 3 s (which combined 10 spectrometer's CCD traces with an exposure time of 300 ms each). The concentrations of NO₂ and CHOCHO time series (Fig. 4a and 4b) were obtained by retrieving the



spectral measurements. The histograms (Figs. 4c and 4d) were constructed from this data (Fig. 4a and 4b) respectively. The standard deviation ($\sigma_{Gaussian}$) and mean value (μ) were calculated from the Gaussian distributions of the histograms for each gas. The mean value was an offset from the expected zero and was considered to be a residual “background”. The limit of detection (LOD) can be defined as Eq. (5) from analytical chemistry and this method was also commonly used in cavity-enhanced systems to evaluate instrument performance (Thalman et al., 2015; Fang et al., 2017).

$$LOD_{exp} = 2 \times \sigma_{Gaussian} + |background|, \quad (5)$$

According to Eq. (5), the detection limits (with a 3 s acquisition time) for NO₂ and CHOCHO were calculated to be about 0.094 ppb (2σ) and 0.058 ppb (2σ), respectively.

Allan variance analysis has been also a convenient way to describe the stability and detection limit of a system as a function of averaging time. We used Allan variance analysis to characterize the overall stability of our system and to determine the optimum averaging time and predict the detection limit of the system. The above mentioned 10,000 spectral concentration values were divided into M groups—each containing N values ($N=1, 2, \dots, 2000$; $M = 10000/N = 10000/1, 10000/2, \dots, 10000/2000$). The average of N values is denoted as y_i ($i = 1, 2, \dots, M$), and the corresponding averaging time is $t_{avg} = N \times 3$ s. Since each spectrum was measured in the optical cavity filled with dry nitrogen, the y_i values contain only measurement noise as a function of averaging times (Langridge et al., 2008). The Allan variance and standard deviation of NO₂ and CHOCHO concentrations are calculated according to Eqs. (6) and (7), respectively, as shown in Figs. 4e and 4f. The Allan deviation initially decreases with a gradient -0.5 as averaging time increases, before it starts gradually to increase towards a longer averaging time. The optimum integration time (210 s for CHOCHO) of the instrument is around the minimum of Allan deviation. Further increase of the integration times yield no more decrease in the Allan deviation due to system drift. For a total acquisition time of 3 s, the detection limits (standard deviation) of NO₂ and CHOCHO are 0.083 ppbv and 0.052 ppbv (2σ), respectively. This result is consistent with LOD_{exp} (0.094 ppbv and 0.058 ppbv). By increasing the spectral averaging time to 30 s (which combined 100 spectrometer’s CCD traces with an exposure time of 300 ms each), the NO₂ and CHOCHO detection limits (standard deviation) were reduced to 29 pptv (2σ) and 23 pptv (2σ), respectively. To capture the rapid variation of CHOCHO in the field, time resolution of the IBBCEAS instrument was typically set to 30 s. During field measurements, system drift was managed by frequently measuring the I₀ spectrum and stabilizing the temperature of the system.

$$\sigma_A^2(t_{avg}) = \frac{1}{2(M-1)} \sum_{i=1}^{M-1} [y_{i+1}(t_{avg}) - y_i(t_{avg})]^2, \quad (6)$$

$$\sigma_s^2(t_{avg}) = \frac{1}{M-1} \sum_{i=1}^M [y_i(t_{avg}) - \mu]^2, \quad (7)$$

In the above formulas, $y_i(t_{avg})$ is the averaging concentration of the i -th group. Term μ is the average concentration over the entire measurement period.




Fig. 4. Evaluation of the instrument performance. Panels (a) and (b) are the time series of NO₂ and CHOCHO with 3 s acquisition time. Panels (c) and (d) show the histogram analyses of the measurements of NO₂ and CHOCHO, respectively. Panels (e) and (f) are Allan deviation plots for measurements of NO₂ and CHOCHO, respectively.

5 3.4 Sampling loss of glyoxal and Measurement of glyoxal sample gas

In order to obtain a stable concentration of glyoxal, we used a mass flow controller to allow the quantitative high purity nitrogen through the trap containing solid glyoxal at atmospheric pressure and at -72 °C. The sample stream out of the glyoxal trap was further diluted with dry high purity nitrogen in a sampling bag (PFA) before entering the inlet of the IBBCEAS.

3.4.1 Sampling tube loss of glyoxal

10 We measured the glyoxal sample gas in the sampling bag alternately using 3 m and 10 m sampling tubes (PFA) at a flow rate of 1 L/min to study the loss of CHOCHO in the sampling tube. The experimental results showed that sampling tube length has no obvious impact on glyoxal loss (Fig. 5). This is consistent with previous findings (Min et al., 2016).





Fig. 5. Measurement of CHOCHO loss in the sampling tube. Blue dots correspond to the measured CHOCHO with the extra 3 m PFA inlet tube; Red dots correspond to the measured CHOCHO with the extra 10 m PFA inlet tube.

3.4.2 Measurements of CHOCHO standard additions

- 5 The high concentration of glyoxal was diluted several times in proportion to obtain the concentration time series as shown in the Fig. 6a. The last five low concentration gradients in the Fig. 6a are diluted proportionally by the first maximum concentration gradient. Figure 6b shows the correlation between the average of these concentration gradients and the normalized mixing ratio with a correlation coefficient of $R^2 = 0.9996$.



10 Fig. 6. Measurement of glyoxal sample gas. (a) Different concentrations of CHOCHO measured by IBBCEAS. (b) shows the correlation between the average of these concentration gradients and the normalized mixing ratio.

3.5 Interference from NO₂ and spectral fitting

- Both the glyoxal and NO₂ have absorption bands in the same wavelength region as shown in Fig. 2. Therefore, it is important to select suitable absorption features for their retrieval to reduce cross interferences. Various factors, such as the performance of the instrument (e.g. the intensity wavelength range of the LED light source, the mirror reflectivity, and the spectrometer resolution), the absorption strength of the gas, the concentration level in the actual atmosphere, and the correlation between the absorption features of different gas species in the same wavelength region should be considered to obtain the best-fitting wavelength interval. Figure 7 shows the correlations matrix of absorption cross-sections of CHOCHO and NO₂ for a range of fitting intervals starting between 429 and 448 nm and ending between 457 and 475 nm. We hope to find an optimal fit interval with minimal correlation (Pinardi et al., 2013). In this paper, the retrieval band of glyoxal and NO₂ is finally 438 – 465 nm.

When the concentration of NO₂ exceeds ~12 ppbv in the actual atmosphere, the absorption due to NO₂ is more than 100-fold higher than that due to a typical 0.1 ppbv glyoxal in the atmosphere. The concentration of NO₂ in the atmosphere of



polluted urban areas in China could reach tens or even hundreds of ppbv (Qin et al. 2009). Concurrently, the accumulation of NO₂ at night further challenges accurate glyoxal measurements. Therefore, accurate data analysis of the NO₂ absorption contributions became critical to reduce its impact on the determination of the glyoxal absorption and concentration. For modelling of measurement spectra, one common approach was to first determine a nominal spectrometer's resolution profile as instrumental function and then the literature reference spectrum was convoluted with the instrumental function of the spectrometer. However, we noticed that the grating spectrometer had nonuniform dispersions. We measured the wavelength dependence of the grating spectrometer's resolution by using narrow atomic emission lines of low-pressure Hg, Kr and Zn lamps. These results were summarized in the Table 2. The nonuniform dispersions make spectral modelling less accurate. Subsequently inaccurate modelling makes it difficult to overcome cross-interference of strongly-absorbed interference gases with weakly-absorbed gases of interest within the same wavelength region. A more reliable approach we used to obtain NO₂ reference spectra was to make a direct measurement of known concentrations of NO₂ standard gases with the spectrometer and further calibrate with the convoluted literature reference spectrum. We verified the measured NO₂ reference spectrum and the convoluted literature NO₂ reference spectrum by retrieval of the same NO₂ spectra. The difference was about 1.4%.




Fig. 7. Correlations matrix of absorption cross-sections of CHOCHO and NO₂ for different wavelength intervals in the 429 – 475 nm wavelength range.

Table 2. Wavelength dependence of the grating spectrometer's resolution

Wavelength (nm)	407.78	431.96	437.61	450.24	468.01	472.21	481.05
FWHM (nm) [*]	0.55	0.56	0.56	0.57	0.60	0.61	0.62

^{*} The full width at half maximum (FWHM) values were determined from the emission linewidth measurements of low-pressure Hg, Kr and Zn lamps.

3.5.1 Residual structure from NO₂ fitting

For our application, inaccurate NO₂ fitting produces a large residual structure especially in the case of high concentrations of NO₂ (see Fig. 8). Figure 8 shows the variation characteristics of fit residuals from fitting different concentrations of standard NO₂ when using the convolution-based NO₂ reference spectrum. As is clear from Fig. 8, there is a similar residual structure in the fit residual and it increases with increasing NO₂ concentration. Such a residual structure will have a disastrous effect on the retrieval of glyoxal in the atmosphere. Figure 9 shows that the standard deviation of these fit residuals has a good dependence with the NO₂ concentration.




Fig. 8. Fitting residuals of different concentrations of standard NO_2 when using the convolution-based NO_2 reference spectrum




Fig. 9. The standard deviation of the fit residual from fig.7 as a function of NO_2 concentrations.

5 3.5.2 Spectral simulation of NO_2 interference with glyoxal

The influence of residual structures in the absorption spectra has been evaluated by Stutz et al. (Stutz et al., 1996). Due to the nonuniform dispersion of the spectrometer, a stable residual structure was produced when a NO_2 reference spectrum based on a simple convolution calculation (by using a nominal function for the instrument line profile) was used for experimental spectral profile fitting (Fig.8). We evaluated this nonuniform dispersion effect of the co-existing NO_2 absorption on glyoxal spectral analysis.

The simulation spectra we used to test the accuracy of the spectral extraction comprised three components: NO_2 reference spectra based on measurements according to Eq. (1), as well as the Rayleigh spectrum of N_2 at 1atm, and convolution-simulated spectrum of 0.1 ppbv CHOCHO. We obtained simulation spectra containing different concentrations of NO_2 (0 – 200 ppbv) and 0.1 ppbv glyoxal according to Eq. (3) as a summation. The spectral retrieval was conducted by using a nonlinear least-squares fitting routine. We tested the retrieval accuracy of CHOCHO by applying either a convolution-based NO_2 reference



spectrum or a measurement-based NO₂ reference spectrum in the nonlinear least-squares fitting routine for the modelling of the NO₂ spectral contribution.

Figure 10a shows the deviation of the retrieved CHOCHO concentration from its nominal 0.1 ppbv value as a function of NO₂ concentration. The blue line in Fig. 10a is the retrieval result when using convolution-based NO₂ reference spectrum as its model function (The gray area indicates the range of fitting uncertainties). The deviation of the extracted glyoxal concentration and estimated uncertainty increase linearly as the concentration of NO₂ increases. The deviation of glyoxal reaches 0.58 ppbv when the concentration of NO₂ is 198 ppbv. In other words, for our instrument, the large bias is characterized as 2.9 pptv glyoxal/ppbv NO₂. Thalman et al. showed in their experiment that the CE-DOAS and BBCEAS systematic bias is 1 pptv glyoxal/ppbv NO₂ at higher NO₂ (Thalman et al., 2015). The difference between our findings and his findings may be due to differences in instruments—especially the spectrometers. When the concentration of NO₂ is less than ~8 ppbv, its effect on the deviation of glyoxal is less than the detection limit of the instrument (23 pptv, 2σ). The NO₂ likely has only a minor effect on glyoxal measurements in this low concentration case. When the retrieval is performed using the measurement-based NO₂ reference spectrum, the deviation of the extracted glyoxal concentration value (Fig. 10a, red line) remains close to zero. The uncertainty of the fitting error (gray area) is also small, indicating that the effect of NO₂ on glyoxal is negligible. Figure 10b compares the standard deviations of the fit residual as a function of NO₂ concentration. The standard deviation is reduced from 5.1×10^{-11} cm⁻¹/ppbv NO₂ when using the convolution-based NO₂ reference spectrum in the least-squares fitting, which reduced to 1.7×10^{-12} cm⁻¹/ppbv NO₂ when using the measurement-based NO₂ reference spectrum. This is an improvement of over 30 times.





Fig. 10. The simulation results of the effect of NO₂ on glyoxal. (a) The deviation of the retrieved glyoxal concentration from its nominal value of 0.1 ppbv as a function of NO₂ concentrations. The blue line is the retrieval result using the convolution-simulated NO₂ reference spectra (gray area is the range of fitting uncertainty). The red line is the result of retrieval via the measured NO₂ reference spectrum (gray area is the range of fitting uncertainty). (b) The corresponding standard deviations of the spectral fit residual.

3.5.3 Spectral fitting of field measurement spectra

We compared the effects of using the convolution-based and the measurement-based NO₂ reference spectra to fit real atmospheric spectral measurements. As the simulation analysis in the previous Section 3.5.2 indicated already, using the measurement-based NO₂ reference spectrum for data analysis of the real atmospheric measurements achieved more precise NO₂ fitting, as both the NO₂ reference spectrum and the real atmospheric measurements share the same instrument (i.e. the grating spectrometer) function. The results of the real experimental spectral fitting are shown in Fig. 11. A comparison of the spectral retrievals using both the convolution-based NO₂ reference spectrum and the measurement-based NO₂ reference spectrum are displayed in the left and right columns of Fig. 11, respectively. Corresponding fit residuals are shown in the bottom panels, and the standard deviations are 1.31×10^{-9} cm⁻¹ and 8.78×10^{-10} cm⁻¹, respectively. The standard deviation of



the fitting residuals by using measurement-based NO₂ reference spectrum is 33% smaller than those with convolution-based reference spectrum. Moreover, the fitting residual using measurement-based NO₂ reference spectrum showed no obvious structure. The fitting of glyoxal is more precise, and the fitting error is reduced by 31.7% (Figs. 11g and 11h) when using the measured NO₂ reference spectrum. For NO₂, the fitting exhibits almost no difference. The result demonstrates that it is critical to use the measured NO₂ reference spectrum. Any tiny distortion in the NO₂ reference spectral profile could have severe effect on the CHOCHO extraction, because NO₂ absorption is about two orders of magnitude stronger than that of the CHOCHO in the local atmosphere. Figure 12 shows the standard deviation of the fitting residual of the absorption coefficient as a function of NO₂ concentration for measurements conducted during the APHH-China project (June 2017). The standard deviation is reduced from $5.1 \times 10^{-11} \text{ cm}^{-1}/\text{ppbv NO}_2$ to $2.2 \times 10^{-11} \text{ cm}^{-1}/\text{ppbv NO}_2$ by using the measurement-based NO₂ reference spectrum, which is 2.3 times smaller. The uncertainties in absorption cross-sections are 4% for NO₂ (Voigt et al., 2002), and 5% for CHOCHO (Volkamer et al., 2005). The difference in NO₂ between the literature reference spectrum and the measured reference spectrum is 1.5%. Our experimental uncertainties in cavity mirror reflectivity and effective cavity length are 5% and 2%, respectively. The combined uncertainty (summed in quadrature) are estimated to be 6.7% or 6.9% for NO₂ using convolution- or measurement-based reference spectrum, and 7.3% for CHOCHO.



15

Fig. 11. A comparison of the experimental atmospheric spectral retrievals using both the convolution-based NO₂ reference spectrum (left column) and the measurement-based NO₂ reference spectrum (right column). (a) and (b) show the same atmospheric spectrum (recorded on 9 June 2017 at 12:28 local time). The retrieved NO₂, H₂O, and CHOCHO concentrations are shown in (c) and (d), (e) and (f), (g) and (h), respectively. Two overall fit residuals are shown in the bottom panels (i) and (j), with the standard deviations of $1.31 \times 10^{-9} \text{ cm}^{-1}$ and $8.78 \times 10^{-10} \text{ cm}^{-1}$, respectively.

20




Fig. 12. The standard deviation of the fit residual of the absorption coefficients as a function of NO₂ concentrations for spectral data analysis during the APHH field measurements in Beijing, June 2017. The red dots were the standard deviations of the fit residuals by using a convolution-based NO₂ spectral profile and the black line is the linear fit of the data. The blue dots were the standard deviations of the fit residuals by using a measurement-based NO₂ reference spectral profile and the green line is the linear fit of the corresponding data.

5

3.6 Field measurements

The field campaign was conducted in the city Beijing at the Iron Tower Department of the Institute of Atmospheric Physics, Chinese Academy of Sciences during the APHH-China project (June 2–26, 2017). The IBBCEAS system was deployed to measure both CHOCHO and NO₂, supplemented by many other atmospheric measurement instruments. The sampling height of the IBBCEAS system was about 4 m above the ground. A cavity-attenuated phase shift (CAPS) spectroscopy system (University of York) for NO₂ data comparison was located in another container about 30 meters away from the IBBCEAS system. Figure 13 shows the 24-day continuous measurements of CHOCHO and NO₂ in the atmosphere by our IBBCEAS instrument. Each measurement data point was derived from each absorption spectrum acquired over 30 s (which averaged 100 spectrometer's CCD traces with an exposure time of 300 ms each). The concentration of glyoxal in the city reached 0.572 ppbv at the maximum; the average was 0.091 ppbv. Time series data for NO₂ measured by IBBCEAS was compared with the data from the CAPS spectroscopy system (Fig. 13b). Overall both sets of measurements were in very good agreement. The average concentration of NO₂ was ~20.0 ppbv and the maximum value was ~80 ppbv. A correlation plot comparing the IBBCEAS and CAPS NO₂ concentration data is shown in Fig. 14, with the data averaged to 1 h. The linear regression exhibits [NO₂] CAPS = 1.03 × [NO₂] IBBCEAS with a correlation coefficient of $R^2 = 0.99$. Discrepancies of ~3% between the two data sets may be partly due to the different air sampling locations of these two instruments and the uncertainty of the effective cavity length calibration of the IBBCEAS. Overall this 3% deviation was within the expected 7.3% uncertainty mentioned in Section 3.5.3.

15
20




Fig. 13. Results of 24-day continuous measurements of CHOCHO and NO₂ in atmosphere.




Fig. 14. Correlation plot of NO₂ concentration values between IBBCEAS and CAPS measurements. Each NO₂ data point was represented by averaged values to 1 h. The slope of the straight line fit (red line) is 1.03 with R² = 0.99

5

3 Conclusions

This paper describes the development of an IBBCEAS system and its field application to high-sensitivity measurements of atmospheric glyoxal and NO₂. The mirror reflectivity of the optical cavity was calibrated using the difference in Rayleigh scattering cross-sections between N₂ and He gases. The mirror reflectivity R is greater than 0.99994 at 455 nm, and the corresponding effective absorption pathlength is about 11.7 km (cavity dimension 0.7 m, in the absence of Rayleigh scattering). To accurately obtain a reduction factor for the cavity length when the cavity mirrors were protected by N₂ pure gases, the O₄ absorption in pure oxygen (at the 477 nm band) was used to calibrate the effective cavity length. The reduction factor of the cavity length was 0.841 at an inlet flow rate of 1 sL min⁻¹ and a total purge flow rate of 0.2 sL min⁻¹. Here, the cavity length d₀ = 70 cm, and the calculated d_{eff} = 58.9 cm. We used Allan variance analysis to identify the system's detection limits for NO₂ and CHOCHO. They were 0.083 ppbv (2σ) and 0.052 ppbv (2σ) at a 3-s time resolution in the laboratory, respectively. Further increases in acquisition time to 30 s improve the detection limits of CHOCHO and NO₂ to 23 pptv (2σ) and 29 pptv (2σ), respectively. The overall uncertainties of the instrument are 6.7% or 6.9% for NO₂ using convolution-based or measurement-based reference spectrum and 7.3% for CHOCHO. The effect of NO₂ on glyoxal was evaluated via spectral simulations and measurements. When using a convolution-based NO₂ reference spectral profile, the high concentration of NO₂ had a large effect on glyoxal and the bias was characterized as 2.9 pptv glyoxal/ppbv NO₂. The effect of NO₂ on glyoxal became negligible when retrieval was performed using the measurement-based NO₂ reference spectral profile. The measured NO₂ reference spectrum was applied to the retrieval of the actual atmospheric spectrum effectively reducing the impact of NO₂ on the retrieval

20



of CHOCHO during the APHH-China field measurement project (June 2–26, 2017). The standard deviation of the fitting residual was reduced from $5.1 \times 10^{-11} \text{ cm}^{-1}/\text{ppbv NO}_2$ to $2.2 \times 10^{-11} \text{ cm}^{-1}/\text{ppbv NO}_2$ by using the measured NO_2 reference spectrum, which is 2.3 times smaller. The concentrations of CHOCHO and NO_2 in the Beijing summer atmosphere were obtained during the APHH-China project. There was good agreement in NO_2 concentrations acquired by the IBBCEAS and another independent instrument using a different measurement technique—CAPS. The maximum concentrations of glyoxal and NO_2 in Beijing in summer reached 0.572 ppbv and ~80 ppbv respectively. This has demonstrated that our IBBCEAS instrument is capable of making accurate continuous measurements in atmospheric environments of high-load particulate matters and high-concentration polluting gases.

Acknowledgements

This work was supported by the National Natural Science Foundation of China (Grant No. 91544104, 41705015 and 41571130023) and the Science and Technology Major Special Project of Anhui Province, China (16030801120). The co-author Y. He is also associated with Macquarie University, Australia. The authors would like to thank Lee James from the University of York for providing NO_2 data. We gratefully acknowledge the discussions with Bin Ouyang from the Lancaster Environment Centre, Lancaster University, UK.

References

- Chen, J. and Venables, D. S.: A broadband optical cavity spectrometer for measuring weak near-ultraviolet absorption spectra of gases, *Atmos. Meas. Tech.* 4, 425–436, doi:10.5194/amt-4-425-2011, 2011.
- Coburn, S., Ortega, I., Thalman, R., Blomquist, B., Fairall, C. W., and Volkamer, R.: Measurements of diurnal variations and eddy covariance (EC) fluxes of glyoxal in the tropical marine boundary layer: description of the Fast LED-CE-DOAS instrument, *Atmos. Meas. Tech.* 7, 3579–3595, doi:10.5194/amt-7-3579-2014, 2014.
- DiGangi, J. P., Henry, S. B., Kammrath, A., Boyle, E. S., Kaser, L., Schnitzhofer, R., Graus, M., Turnipseed, A., Park, J. H., Weber, R. J., Hornbrook, R. S., Cantrell, C. A., Maudlin Iii, R. L., Kim, S., Nakashima, Y., Wolfe, G. M., Kajii, Y., Apel, E. C., Goldstein, A. H., Guenther, A., Karl, T., Hansel, A., and Keutsch, F. N.: Observations of glyoxal and formaldehyde as metrics for the anthropogenic impact on rural photochemistry, *Atmos. Chem. Phys.*, 12, 9529–9543, doi:10.5194/acp-12-9529-2012, 2012.
- Fu, T.-M., Jacob, D. J., Wittrock, F., Burrows, J. P., Vrekoussis, M., and Henze, D. K.: Global budgets of atmospheric glyoxal and methylglyoxal, and implications for formation of secondary organic aerosols, *J. Geophys. Res.*, 113, D15303, doi:10.1029/2007JD009505, 2008.
- Fiedler, S. E., Hoheisel, G., Ruth, A. A., and Hese, A.: Incoherent broad-band cavity-enhanced absorption spectroscopy of azulene in a supersonic jet, *Chem. Phys. Lett.*, 382, 447–453, doi:10.1016/j.cplett.2003.10.075, 2003.
- Fang, B., Zhao, W., Xu, X., Zhou, J., Ma, X., Wang, S., Zhang, W., Venables, D. S., and Chen, W.: Portable broadband cavity-enhanced spectrometer utilizing Kalman filtering: application to real-time, in situ monitoring of glyoxal and nitrogen dioxide, *Opt. Express*, 25, 26910–26922, doi:10.1364/OE.25.026910, 2017.
- Ho, S. S., Yu, J. Z.: Determination of airborne carbonyls: comparison of a thermal desorption/GC method with the standard DNPH/HPLC method, *Environ. Sci. Technol.*, 38, 862–870, doi:10.1021/es034795w, 2004.
- Henry, S. B., Kammrath, A., and Keutsch, F. N.: Quantification of gas-phase glyoxal and methylglyoxal via the Laser-Induced Phosphorescence of (methyl)GlyOxal Spectrometry (LIPGLOS) Method, *Atmos. Meas. Tech.*, 5, 181–192, doi:10.5194/amt-5-181-2012, 2012.
- Huisman, A. J., Hottle, J. R., Coens, K. L., DiGangi, J. P., Galloway, M. M., Kammrath, A., and Keutsch, F. N.: Laser-induced phosphorescence for the in situ detection of glyoxal at part per trillion mixing ratios, *Anal. Chem.* 80, 5884–5891, 2008.
- Kaiser, J., Wolfe, G. M., Min, K. E., Brown, S. S., Miller, C. C., Jacob, D. J., deGouw, J. A., Graus, M., Hanisco, T. F., Holloway, J., Peischl, J., Pollack, I. B., Ryerson, T. B., Warneke, C., Washenfelder, R. A., and Keutsch, F. N.: Reassessing the ratio of glyoxal to formaldehyde as an indicator of hydrocarbon precursor speciation, *Atmos. Chem. Phys.* 15, 7571–7583, doi:10.5194/acp-15-7571-2015, 2015.
- Kahan, T. F., Washenfelder, R. A., Vaida, V., and Brown, S. S.: Cavity-enhanced measurements of hydrogen peroxide absorption cross sections from 353 to 410 nm, *J. Phys. Chem. A*, 116, 5941–5947, 2012.
- Li, J., Mao, J., Min, K. E., Washenfelder, R. A., Brown, S. S., Kaiser, J., Keutsch, F. N., Volkamer, R., Wolfe, G. M., Hanisco, T. F., Pollack, I. B., Ryerson, T. B., Graus, M., Gilman, J. B., Lerner, B. M., Warneke, C., de Gouw, J. A., Middlebrook, A. M., Liao, J., Welti, A., Henderson, B. H., McNeill, V. F.,



- Hall, S. R., Ullmann, K., Donner, L. J., Paulot, F., and Horowitz, L. W.: Observational constraints on glyoxal production from isoprene oxidation and its contribution to organic aerosol over the Southeast United States, *J. Geophys. Res. Atmos.* 121, 9849-9861, doi:10.1002/2016JD025331, 2016
- Langridge, J. M., Ball, S. M., Shillings, A. J. L., and Jones, R. L.: A broadband absorption spectrometer using light emitting diodes for ultrasensitive, in situ trace gas detection, *Rev. Sci. Instrum.*, 79, 123110, doi: 10.1063/1.3046282, 2008.
- 5 Min, K. E., Washenfelder, R. A., Dubé W. P., Langford, A. O., Edwards, P. M., Zarzana, K. J., Stutz, J., Lu, K., Rohrer, F., Zhang, Y., and Brown, S. S.: A broadband cavity enhanced absorption spectrometer for aircraft measurements of glyoxal, methylglyoxal, nitrous acid, nitrogen dioxide, and water vapor, *Atmos. Meas. Tech.*, 9, 423-440, doi:10.5194/amt-9-423-2016, 2016.
- Munger, J. W., Jacob, D. J., Daube, B. C., and Horowitz, L. W.: Formaldehyde, glyoxal, and methylglyoxal in air and cloudwater at a rural mountain site in central Virginia, *J. Geophys. Res. Atmos.*, 100, 9325-9333, 1995.
- 10 MacDonald, S. M., Oetjen, H., Mahajan, A. S., Whalley, L. K., Edwards, P. M., Heard, D. E., Jones, C. E., and Plane, J. M. C.: DOAS measurements of formaldehyde and glyoxal above a south-east Asian tropical rainforest, *Atmos. Chem. Phys.*, 12, 5949-5962, doi:10.5194/acp-12-5949-2012, 2012.
- Pang, X., Lewis, A. C., Rickard, A. R., Baeza-Romero, M. T., Adams, T. J., Ball, S. M., Daniels, M. J. S., Goodall, I. C. A., Monks, P. S., Peppe, S., Ródenas Garc ía, M., Sánchez, P., and Muñoz, A.: A smog chamber comparison of a microfluidic derivatisation measurement of gas-phase glyoxal and methylglyoxal with other analytical techniques, *Atmos. Meas. Tech.*, 7, 373-389, doi:10.5194/amt-7-373-2014, 2014.
- 15 Pinardi, G., Van Roozendaal, M., Abuhassan, N., Adams, C., Cede, A., Clémer, K., Fayt, C., Frieß U., Gil, M., Herman, J., Hermans, C., Hendrick, F., Irie, H., Merlaud, A., Navarro Comas, M., Peters, E., Piter, A. J. M., Puentedura, O., Richter, A., Schönhardt, A., Shaiganfar, R., Spinei, E., Strong, K., Takashima, H., Vrekoussis, M., Wagner, T., Wittrock, F., and Yilmaz, S.: MAX-DOAS formaldehyde slant column measurements during CINDI: intercomparison and analysis improvement, *Atmos. Meas. Tech.*, 6, 167-185, doi:10.5194/amt-6-167-2013, 2013.
- Qin, M., Xie, P., Su, H., Gu, J., Peng, F., Li, S., Zeng, L., Liu, J., Liu, W., and Zhang, Y.: An observational study of the HONO-NO₂ coupling at an urban site in Guangzhou City, South China, *Atmos. Environ.*, 43, 5731-5742, doi:10.1016/j.atmosenv.2009.08.017, 2009.
- 20 Shardanand, and Rao, A. D. P.: Absolute Rayleigh scattering cross sections of gases and freons of stratospheric interest in the visible and ultraviolet regions, NASA Technical Note, 1977.
- Sneep, M., and Ubachs, W.: Direct measurement of the Rayleigh scattering cross section in various gases, *J. Quant. Spectrosc. Radiat. Transf.*, 92, 293-310, 2005.
- 25 Stutz, J., and Platt, U.: Numerical analysis and estimation of the statistical error of differential optical absorption spectroscopy measurements with least-squares methods, *Appl. Optics*, 35, 6041-6053, 1996.
- Temime, B., Healy, R. M., and Wenger, J.: A denuder-filter sampling technique for the detection of gas and particle phase carbonyl compounds, *Environ. Sci. Technol.*, 41, 6514-6520, doi:10.1021/es070802v, 2007.
- Thalman, R., and Volkamer, R.: Inherent calibration of a blue LED-CE-DOAS instrument to measure iodine oxide, glyoxal, methyl glyoxal, nitrogen dioxide, water vapour and aerosol extinction in open cavity mode, *Atmos. Meas. Tech.*, 3, 1797-1814, doi:10.5194/amt-3-1797-2010, 2010.
- 30 Thalman, R., Baeza-Romero, M. T., Ball, S. M., Borrás, E., Daniels, M. J. S., Goodall, I. C. A., Henry, S. B., Karl, T., Keutsch, F. N., Kim, S., Mak, J., Monks, P. S., Muñoz, A., Orlando, J., Peppe, S., Rickard, A. R., Ródenas, M., Sánchez, P., Seco, R., Su, L., Tyndall, G., Vázquez, M., Vera, T., Waxman, E., and Volkamer, R.: Instrument intercomparison of glyoxal, methyl glyoxal and NO₂ under simulated atmospheric conditions, *Atmos. Meas. Tech.*, 8, 1835-1862, doi:10.5194/amt-8-1835-2015, 2015.
- 35 Thalman, R., and Volkamer, R.: Temperature dependent absorption cross-sections of O₂-O₂ collision pairs between 340 and 630 nm and at atmospherically relevant pressure, *Phys. Chem. Chem. Phys.*, 15, 15371-15381, doi: 10.1039/c3cp50968k, 2013.
- Vrekoussis, M., Wittrock, F., Richter, A., and Burrows, J. P.: GOME-2 observations of oxygenated VOCs: what can we learn from the ratio glyoxal to formaldehyde on a global scale?, *Atmos. Chem. Phys.*, 10, 10145-10160, doi:10.5194/acp-10-10145-2010, 2010.
- Volkamer, R., Martini, F.S., Molina, L. T., Salcedo, D., Jimenez, J. L., and Molina, M. J.: A missing sink for gas-phase glyoxal in Mexico City: Formation of secondary organic aerosol, *Geophys. Res. Lett.* 34, L19807, doi:10.1029/2007GL030752, 2007.
- 40 Volkamer, R., Molina, L. T., Molina, M. J., Shirley, T., and Brune, W. H.: DOAS measurement of glyoxal as an indicator for fast VOC chemistry in urban air, *Geophys. Res. Lett.*, 32, L08806, doi:10.1029/2005GL022616, 2005a.
- Volkamer, R., Baidar, S., Campos, T. L., Coburn, S., DiGangi, J. P., Dix, B., Eloranta, E. W., Koenig, T. K., Morley, B., Ortega, I., Pierce, B. R., Reeves, M., Sinreich, R., Wang, S., Zondlo, M. A., and Romashkin, P. A.: Aircraft measurements of BrO, IO, glyoxal, NO₂, H₂O, O₂-O₂ and aerosol extinction profiles in the tropics: comparison with aircraft-/ship-based in situ and lidar measurements, *Atmos. Meas. Tech.*, 8, 2121-2148, doi:10.5194/amt-8-2121-2015, 2015.
- 45 Volkamer, R., Spietz, P., Burrows, J., and Platt, U.: High-resolution absorption cross-section of glyoxal in the UV-vis and IR spectral ranges, *J. Photochem. Photobiol. Chem.* 172, 35-46, doi:10.1016/j.jphotochem.2004.11.011, 2005b.
- Voigt, S., Orphal, J., and Burrows, J. P.: The temperature and pressure dependence of the absorption cross-sections of NO₂ in the 250-800 nm region measured by Fourier-transform spectroscopy, *J. Photochem. Photobiol. Chem.*, 149, 1-7, 2002.
- 50



- Washenfelder, R. A., Young, C. J., Brown, S. S., Angevine, W. M., Atlas, E. L., Blake, D. R., Bon, D. M., Cubison, M. J., de Gouw, J. A., Dusanter, S., Flynn, J., Gilman, J. B., Graus, M., Griffith, S., Grossberg, N., Hayes, P. L., Jimenez, J. L., Kuster, W. C., Lefer, B. L., Pollack, I. B., Ryerson, T. B., Stark, H., Stevens, P. S., and Trainer, M. K.: The glyoxal budget and its contribution to organic aerosol for Los Angeles, California, during CalNex 2010, *J. Geophys. Res. Atmos.*, 116, D00V02, doi:10.1029/2011JD016314, 2011.
- 5 Wang, H., Chen, J., and Lu, K.: Development of a portable cavity-enhanced absorption spectrometer for the measurement of ambient NO₃ and N₂O₅: experimental setup, lab characterizations, and field applications in a polluted urban environment, *Atmos. Meas. Tech.*, doi:10.5194/amt-10-1465-2017, 2017.
- Washenfelder, R. A., Flores, J. M., Brock, C. A., Brown, S. S., and Rudich, Y.: Broadband measurements of aerosol extinction in the ultraviolet spectral region, *Atmos. Meas. Tech.*, 6, 861-877, doi:10.5194/amt-6-861-2013, 2013.
- 10 Washenfelder, R. A., Langford, A.O., Fuchs, H., and Brown, S. S.: Measurement of glyoxal using an incoherent broadband cavity enhanced absorption spectrometer, *Atmos. Chem. Phys.*, 8, 7779-7793, www.atmos-chem-phys.net/8/7779/2008/, 2008.
- Werle, P., Mücke, R., and Slemr, F.: The limits of signal averaging in atmospheric trace-gas monitoring by tunable diode-laser absorption spectroscopy (TDLAS), *Appl. Phys. B.*, 57, 131-139, 1993.
- 15 Yi, H., Wu, T., Wang, G., Zhao, W., Fertein, E., Coeur, C., Gao, X., Zhang, W., and Chen, W.: Sensing atmospheric reactive species using light emitting diode by incoherent broadband cavity enhanced absorption spectroscopy, *Opt. Express*, 24, A781-790, doi:10.1364/OE.24.00A781, 2016.