- 1 Aerosol Optical Depth comparison between GAW-PFR and AERONET-
- 2 Cimel radiometers from long-term (2005-2015) 1-minute synchronous
- 3 measurements
- 4 Emilio Cuevas¹, Pedro Miguel Romero-Campos¹, Natalia Kouremeti², Stelios Kazadzis²,
- 5 Petri Räisänen³, Rosa Delia García^{4,1}, Africa Barreto^{5,1,4}, Carmen Guirado-Fuentes^{4,1}, Ramón Ramos¹,
- 6 Carlos Toledano⁴, Fernando Almansa^{5,1,4}, and, Julian Gröbner²
- ⁷ ¹Izaña Atmospheric Research Center (IARC), State Meteorological Agency (AEMET), Spain
- 8 ²Physikalisch-Meteorologisches Observatorium Davos, World Radiation Center (PMOD/WRC), Davos,
- 9 Switzerland
- 10 ³Finnish Meteorological Institute, Helsinki, Finland
- ⁴Atmospheric Optics Group, Valladolid University, Valladolid, Spain
- 12 ⁵Cimel Electronique, Paris, France
- 13 Correspondence: Emilio Cuevas
- 14 (ecuevasa@aemet.es)
- 15 Abstract

16 A comprehensive comparison of more than 70000 synchronous 1-minute aerosol optical depth (AOD) data 17 from three Global Atmosphere Watch-Precision Filter Radiometers (GAW-PFR), traceable to the World 18 AOD reference, and 15 Aerosol Robotic Network-Cimel (AERONET-Cimel) radiometers, calibrated 19 individually with the Langley plot technique, was performed for four common or near wavelengths 380 20 nm, 440 nm, 500 nm and 870 nm in the period 2005-2015. The goal of this study is to assess whether, 21 despite the marked technical differences between both networks (AERONET, GAW-PFR) and the number 22 of instruments used, their long-term AOD data are comparable and consistent. The percentage of data 23 meeting the WMO traceability requirements (95% of the AOD differences of an instrument compared to 24 the WMO standards lie within specific limits) is > 92 % at 380 nm, > 95 % at 440 nm and 500 nm, and 98 25 % at 870 nm, with the results being quite similar for both AERONET V2 and V3 versions. For the data 26 outside these limits the contribution of calibration and differences in the calculation of the optical depth 27 contribution due to Rayleigh scattering, and O₃ and NO₂ absorption have a negligible impact. For AOD > 28 0.1, a small but non-negligible percentage (~ 1.9 %) of the AOD data outside the WMO limits at 380 nm 29 can be partly assigned to the impact of dust aerosol forward scattering on the AOD calculation due to the 30 different field of view of the instruments. Due to this effect the GAW-PFR provides AOD values which are 31 ~3% lower at 380 nm, and ~2% lower at 500 nm, compared with AERONET-Cimel. The comparison of 32 the Angström exponent shows that under non-pristine conditions (AOD > 0.03 and AE < 1) the AE 33 differences remain < 0.1. This long-term comparison shows an excellent traceability of AERONET-Cimel 34 AOD with the World AOD reference at 440 nm, 500 nm and 870 nm channels and a fairly good agreement 35 at 380 nm, although AOD should be improved in the UV range.

36 *Copyright statement*. TEXT

37 **1. Introduction**

38 In recent decades there has been a growing interest in the role played by atmospheric aerosols in the

- 39 radiation budget and the Earth's hydrological cycle, mainly through their physical and optical properties
- 40 (IPCC, 2013). The most comprehensive and important parameter that accounts for the optical activity of
- 41 aerosols in the atmospheric column is the aerosol optical depth (AOD) (WMO, 2003, 2005). This is also a
- 42 key parameter used in atmospheric column aerosol modelling (e.g. Basart et al., 2012; Benedetti et al., 2018;
- 43 Cuevas et al., 2015; Huneeus et al., 2016) and in satellite observations (e.g. Sayer et al., 2012, 2013; Kahn

1 and Gaitley, 2015; Amiridis et al., 2015). The second aerosol optical parameter in importance is the 2 Angström exponent (AE) (Angstrom, 1929) that accounts for the spectral dependency of the AOD. Since 3 the AE is inversely related to the average size of the aerosol particles, it is a qualitative indicator of the 4 atmospheric aerosol particle size and therefore a useful parameter to assess the aerosol type (WMO, 2003). 5 At present, two global ground-based radiometer networks provide aerosol optical properties of the 6 atmospheric column using centralized data processing procedures based on their respective standard criteria 7 and also centralized protocols for calibration and quality control, linking all network instruments. These 8 are GAW-PFR (Global Atmosphere Watch - Precision Filter Radiometer; http://www.pmodwrc.ch/worcc/; 9 last access: 05 September 2018) and AERONET-Cimel (AErosol RObotic NETwork - Cimel Electronique 10 radiometer; https://aeronet.gsfc.nasa.gov; last access: 01 September 2018) networks. AERONET is, in fact, 11 a federation of ground-based remote sensing aerosol networks established by NASA (National Aeronautics 12 and Space Administration) and PHOTONS (PHOtométrie pour le Traitement Opérationnel de 13 Normalisation Satellitaire; University of Lille- Service d'Observation de l'INSU, France; Goloub et al., 2007), being complemented by other sub-networks, such as, AEROCAN (Canadian sunphotometry 14 15 network; Bokoye et al., 2001), AeroSibnet (Siberian system for Aerosol monitoring; Sakerin et al., 2005), 16 AeroSpan (Aerosol characterisation via Sun photometry: Australia Network; Mitchell et al., 17 2017), CARSNET (China Aerosol Remote Sensing NETwork; Che et al., 2015), and RIMA (The Iberian 18 network for aerosol measurements; Toledano et al., 2011). There are other radiometer networks that in 19 recent years have incorporated centralized protocols for data evaluation and databases, and performed 20 regular intercomparisons with GAW-PFR and AERONET-Cimel. These include, for example, SKYNET 21 (SKYradiometer NETwork), and its seven associated sub-networks, that uses the Prede-POM sky radiometer to investigate aerosol-cloud-solar radiation interactions (e.g. Campanelli et al., 2004; Nakajima 22 23 et al., 2007; Takamura et al., 2004).

24 The World Optical Depth Research Calibration Center (WORCC) was established in 1996 at the 25 Physikalisch Meteorologisches Observatorium Davos / World Radiation Center (PMOD / WRC). The 26 GAW-PFR network (Wehrli, 2005) was initiated within PMOD/WRC for global and long-term atmospheric 27 aerosol monitoring and accurate detection of trends. Aerosol data series measured at 12 core sites away 28 from local and regional pollution sources, representative of atmospheric background conditions in different 29 climates and environments of the planet, in addition to another 20 associated stations are included in this 30 global network (Kazadzis et al., 2018a). For this reason, GAW-PFR uses the PFR, an accurate and reliable 31 instrument regarding its absolute response stability over time that was designed for long-term AOD 32 measurements (Wehrli, 2008a). The GAW-PFR was specifically designed by WORCC for this goal 33 following the technical specifications defined by WMO (2003; 2016). In 2006, the Commission for 34 Instruments and Methods of Observation (CIMO) of WMO (WMO, 2007) recommended that the WORCC 35 at the PMOD/WRC should be designated as the primary WMO Reference Centre for AOD measurements 36 (WMO, 2005).

The AERONET-Cimel network (Holben et al., 1998) was, in principal, designed to validate satellite products and to characterize the spatial-temporal distribution of atmospheric aerosols based on their optical

1 properties. It is the largest surface-based global aerosol network with more than 84 sites with measurement 2 series longer than 10 years and more than 242 sites having data sets > 5 years. Cimel radiometer data, part 3 of AERONET, are processed centrally and freely delivered in near real time by the NASA Goddard Space 4 Flight Center. Both networks, although designed to meet different objectives, are now global benchmarks 5 for the study and characterization of aerosol optical properties worldwide, and for the evaluation of aerosol observations on board satellites and simulations with models. Multiple studies have proliferated in recent 6 7 years to obtain aerosol climatology and to determine AOD trends in different parts of the world (e.g. Nyeki 8 et al., 2012; Klingmüller et al., 2016; Chedin et al., 2018). However, these networks use radiometers with 9 significant technical differences. Moreover, calibration methodologies, AOD calculation algorithms and 10 data evaluation methods are also relatively different between the two networks. Consequently, the objective 11 of this study is to assess whether, despite the marked differences between both networks, including the 12 different day-to-day maintenance and operation procedures of the respective instruments during the study 13 period, the long-term AOD provided by the two networks is comparable and consistent.

14 The WMO has defined the GAW-PFR Triad (three Master PFR instruments) as the world-wide 15 reference for AOD measurements (WMO, 2005). Based on this concept, an instrument provides traceable 16 measurements of AOD to this WMO reference, when this instrument can demonstrate an unbroken chain 17 of calibrations between itself and the GAW-PFR Triad with AOD measurements within specified limits of 18 the GAW-PFR reference. This can either be achieved by a direct comparison to the GAW-PFR Triad 19 (Kazadzis et al., 2018a), or by using a portable transfer standard radiometer as presented in this study. 20 Several comparisons between AERONET-Cimel, GAW-PFR and other radiometers have been carried out 21 in different places (Barreto et al., 2016; Kazadzis et al., 2014, 2018b; Kim et al., 2008; McArthur et al., 22 2003; Mitchell and Forgan, 2003; Nyeki et al., 2015; Schmid et al., 1999; Toledano et al., 2012). However, 23 these comparisons have been performed during field intercomparison campaigns or during relatively short 24 periods of time, so they are not representative of a large variety of atmospheric conditions. In addition, the 25 type of instrument maintenance and the number and qualifications of staff serving them during campaigns 26 is generally of a higher quality compared to that of the instrument daily operation in unattended mode. This 27 might cause an improvement of the instrument performance during intensive campaigns compared to the 28 operational mode.

The growing interest in the analysis of long-term AOD and AE data series for climatological purposes requires an assessment of their quality assurance and long-term intercomparability. This is the first study to analyse the long-term traceability of AERONET-Cimel with respect to GAW-PFR, and therefore to assess the validity of the long AOD and AE AERONET-Cimel data series for climatological and climate change studies under specific quality control requirements.

GAW-PFR has a comprehensive calibration system (Kazadzis et al., 2018a; Schmid and Wehrli, 1995)
that is transferred by a worldwide suite of reference instruments. AERONET-Cimel does not have a CIMOWMO linked reference and, as described by Holben et al. (1998), Eck et al. (1999), and Toledano et al.
(2018), is based on maintaining Reference AERONET radiometers based on the Langley calibration
technique at Izaña, Spain and Mauna Loa, USA. Calibration of all other instruments based on raw voltage

ratios comparisons with Reference instruments at dedicated sites (Carpentras-France, Washington DC-USA, Valladolid-Spain). There are few places in the world where synchronous observations of these two networks are available for long time periods and variable AOD conditions. The Izaña Observatory (IZO; Tenerife, Canary Islands) is one of them. The GAW-PFR measurements started at Izaña Observatory in 2001 (Wehrli, 2005) while AERONET-Cimel started in 2003 (Goloub et al., 2007). Since 2005, synchronous measurements (1-minute values), that have been evaluated following the calibration procedures of each of the networks, are available.

8 In addition, the Izaña Observatory is one of the two places in the world (the other is Mauna Loa -9 Hawaii, USA) where sun-calibrations are performed using the Langlev plot technique for both AERONET-10 Cimel and GAW-PFR reference instruments (Toledano et al., 2018) because of stable (and very low) AOD 11 conditions during many days per year. Consequently, the instruments compared at the Izaña Observatory 12 have been calibrated under the same environmental conditions, and therefore AOD differences can be 13 directly linked with calibration principles, AOD post-processing and other instrumental differences. In this 14 work, we analyse and evaluate the comparison of 11 years (2005-2015) of 1-minute synchronous 15 observations of AOD with AERONET-Cimel and GAW-PFR in four common or near wavelengths, 16 assessing the results and explaining the possible causes of these differences. Some preliminary technical 17 details on the traceability between GAW-PFR and AERONET-Cimel were reported in a technical report 18 by Romero-Campos et al. (2017).

19 In Section 2 the facility in which this long-term comparison has been carried out is described. The technical 20 characteristics of the AERONET-Cimel and GAW-PFR instruments are shown in Section 3, with special 21 emphasis on the technical and methodological differences of both networks. Section 4 describes the 22 methodology followed in this intercomparison based on the concept of WMO-GAW traceability. Results 23 are given in Section 5. A summary and conclusions are provided in Section 6. The Supplement contains 24 case analyses of inaccurate calibration and cloud contamination, results of the comparison between PFR and Cimel with AERONET V3, additional information of the very short natural AOD variability, and the 25 26 simulations performed with the Monte Carlo model to evaluate the impact of dust forward scattering 27 radiation on AOD determination."

28 **2. Site Description**

29 Izaña Observatory (28.3° N, 16.5° W; 2373 m a.s.l.) is located in Tenerife (Canary Islands, Spain) and is managed by the Izaña Atmospheric Research Center (IARC), which is part of the State Meteorological 30 31 Agency of Spain (AEMET). It is a suitable place for long-term studies of aerosol optical properties under contrasting atmospheric and meteorological conditions. This is because IZO is located in the free 32 33 troposphere (FT) above the temperature inversion caused by the trade wind regime in lower levels and 34 general subsidence associated with the branch of the decay of Hadley's cell aloft (Carrillo et al., 2016). 35 This meteorological feature favours, during most of the year, the presence of pristine skies and clean air 36 representative of atmospheric background conditions (Cuevas et al., 2013; Rodríguez et al., 2009). On the 37 other hand, its proximity to the African continent makes it a privileged site for observing and characterizing

1 the Saharan Air Laver (SAL) that normally presents a high burden of desert mineral dust, especially during 2 the summer months (Basart et al., 2009; Rodríguez et al., 2011; Cuevas et al., 2015). At this time of the year, the SAL impacts the subtropical free troposphere over the North Atlantic with large interannual 3 (Rodríguez et al., 2015) and sharp intraseasonal (Cuevas et al., 2017a) variability. The contrasting 4 5 atmospheric conditions that occur at IZO allow the comparison of the two networks, which can be 6 performed under a wide range of AOD values; mostly for pristine conditions (AOD ≤ 0.03) but also for 7 relatively high turbidity (AOD > 0.6) linked with dust aerosol related intrusions. In addition, the location 8 offers the possibility of observing rapid changes in AOD, going from pristine conditions to dusty skies, and 9 vice versa, in a matter of a few hours, especially in the summer period. The periodical presence of a dust 10 laden SAL allows us to evaluate the impact that the dust forward scattering into the field of view has on 11 AOD retrieval. All this defines IZO as an excellent atmospheric aerosol natural laboratory to compare the 12 performance of different radiometers measuring AOD. One of the first international AOD intercomparison 13 campaigns was carried out at IZO in April 1984 (WMO, 1986) promoted and coordinated by PMOD / WRC. 14

15 The privileged conditions of pristine skies that characterize IZO during many days a year have allowed this 16 observatory to become a calibration site for the GAW-PFR and AERONET-Cimel networks since 2001 17 and 2003, respectively, where the extraterrestrial constants are determined with direct sun observations 18 using the Langley plot technique (Toledano et al., 2018). Note that the extraterrestrial constant (calibration 19 constant) is the signal the instrument would read outside the atmosphere at a normalized earth-sun distance. 20 In addition, since July 2014, IZO has also been designated by the WMO as a CIMO (WMO, 2014) testbed 21 for aerosols and water vapour remote sensing instruments. IZO is a station of the Baseline Surface Radiation 22 Network (BSRN) (Driemel et al., 2018; García et al., 2019). Details of IZO facilities, measurement 23 programmes and main research activities can be found in Cuevas et al. (2017b).

24 **3. GAW-PFR and AERONET-Cimel radiometers**

25 The two types of radiometers intercompared in this study are Cimel CE318-N (Holben et al., 1998), 26 hereinafter referred to as Cimel, the standard instrument of AERONET until the recent appearance of 27 CE318-T (Barreto et al., 2016), and the PFR (Wehrli, 2005) standard instrument of the GAW-PFR network. 28 The main features of these two radiometers are described in Table 1. The Cimel (Holben et al., 1994, 1998) 29 is a radiometer equipped with a 2-axis robot that performs two types of basic radiation measurements: direct 30 solar irradiance and sky (radiance) observations, thanks to an automatic pointing robot that executes the 31 observation sequences that have been scheduled. The robot performs automatic pointing to the sun by 32 stepping azimuth and zenith motors using ephemeris based on time, latitude and longitude. Additionally, a 33 four-quadrant detector is used to improve the sun tracking before each scheduled measurement sequence. 34 This sensor guides the robot to the point where the intensity of the signal channel is maximum. Diffuse-sky 35 measurements are also performed by Cimel to infer aerosol optical and microphysical properties. Two 36 different routines are executed: almucantar (varying the azimuth angle keeping constant the zenith angle) 37 and principal plane (varying the zenith angle keeping constant the azimuth angle). The ability of Cimel to

1 perform both direct and diffuse-sky measurements makes it necessary to use a specific robot rather than a simple sun tracker. The field of view angle (FOV) of the instrument is 1.29° (~1.3° from now on) (Torres 2 3 et al., 2013). The wavelengths in which the measurements are sequentially made by a single detector depend on the interference filters that each version of the radiometer has installed in the filter wheel, which is 4 5 located inside the sensor head and which is moved by a stepper motor. The Cimel versions used in this study have at least eight interference filters centred at 340 nm, 380 nm, 440 nm, 500 nm, 675 nm, 870 nm, 6 7 940 nm, and 1020 nm and 10 nm full width at half maximum (FWHM) bandwidth, except for 340 nm and 8 380 nm which have 2 nm and 4 nm FWHM, respectively. Solar irradiance is measured with a Silicon 9 detector in these channels. The possible deterioration of the interference filters is reduced since they are 10 only sun-exposed during three consecutive 1-second direct-sun measurements per channel, this cycle being 11 scheduled every ~15 minutes. The rest of the time the Cimel is taking sky radiance measurements, or at rest 12 position, looking downwards.

13 The PFR (Wehrli, 2000, 2005, 2008a, b) is designed for continuous and automated operation under a 14 broad range of weather conditions. It accurately measures direct solar radiation transmitted in four 15 independent narrow wavelength channels centred at 368 nm, 412 nm, 500 nm and 862 nm, with 5 nm 16 FWHM bandwidth. The FOV of the instrument is 2.5° and the slope angle is 0.7°. Dielectric interference filters manufactured by the ion-assisted-deposition technique are used to assure significantly larger stability 17 18 in comparison to the one manufactured by classic soft-coatings. The PFR was designed for long-term stable 19 measurements, therefore the instrument is hermetically sealed with an internal atmosphere slightly 20 pressurized (2000 hPa) with dry nitrogen, and is stabilized in temperature with a Peltier-type thermostatic 21 system maintaining the temperature of the detector head at 20° C $\pm 0.5^{\circ}$ C. This system makes corrections 22 of the sensitivity for temperature unnecessary, and also prevents accelerated ageing of filters, ensuring the 23 high stability of the PFR. The PFR is mounted on a sun tracker, pointing always at the sun without any 24 active optimization of the sun-pointing. The detectors are only exposed for short time periods, since an 25 automated shutter opens every minute for 10s for sun measurements, minimizing degradation related to the 26 filters exposure.

The expected uncertainty of AOD in the four channels of the PFR radiometer is from 0.004 (862 nm) up to 0.01 (368 nm) (Wehrli, 2000). For the Cimel radiometer, the expected uncertainty of level 2-AOD product is between 0.002 and 0.009, for reference instruments, larger for shorter wavelengths, and between 0.01 and 0.02 for field instruments, larger in the UV, under conditions of clear skies (Eck et al., 1999; Barreto et al., 2016). It should be taken into account that, in general, in the ultraviolet range the AOD uncertainty is higher and depends on the optical air mass (Carlund et al., 2017).

In relation to the calibration of both networks, GAW and AERONET, they use measurements at high mountain stations with very stable and low AOD over a day in which consecutive measurements can be performed over a wide range of optical air mass (approximately between 2 and 5) in the shortest possible time, in order to calibrate Reference instruments using the Langley plot technique. In case of AERONET-Cimel these calibrations are subsequently transferred to the field instruments of the network in other sites through regular intercomparison campaigns. In case of the GAW-PFR, the calibration system is more

1 complex in order to ensure traceability with the WORCC world reference. The maintenance of the AOD standard by the WORCC Calibration Central laboratory is described in Kazadzis et al. (2018). It consists 2 3 of a triad of instruments that measure continuously, and three additional portable transfer standard 4 radiometers located at Mauna Loa (one instrument) and Izaña (two instruments) observatories. Every six 5 months, one of the portable transfer standard radiometers visits the reference triad based at PMOD/WRC 6 (Davos) and compares the calibration constants defined by the 6-month Langley calibrations in the two 7 high mountain stations (Table 1 of Kazadzis et al. 2018) with the one defined by the triad. The comparison 8 is based on the signals (voltages) and not on AOD values. The differences between the Izaña GAW-PFR 9 radiometers and the reference triad have been always lower than 0.5%, being within the uncertainty of the 10 Langley method plus the small possible instrument degradations that can be detected in a 6-12 month 11 period. Such degradations are quite small and are accounted for in the calibration analysis since 12 extraterrestrial constants are linearly interpolated between two triad visits or every 6-month periods. 13 Additionally, the Izaña GAW-PFR "field" radiometers are calibrated on a routine basis using the Langley-14 plot technique for double checking quality assurance. Therefore, these radiometers cannot be considered as simple "field" instruments, but as regularly calibrated radiometers with assured traceability with the 15 16 WORCC triad reference.

17 IZO is one of the two sites of Langley-plot calibration of both networks, which represents an advantage when comparing the two instruments, eliminating, to a large extent, errors caused by the calibration transfer. 18 19 However, there are differences between the calibration methodologies used by both networks. AERONET 20 obtains the calibration by means of the average of a few extraterrestrial constants (V_0), obtained from 21 Langleys, performed in a relatively short time (the time needed to collect data from at least 10 morning Langley plots). However, PFR related Langleys are calculated by temporal linear fit to a larger number of 22 23 extra-terrestrial constants V₀ obtained from Langley plots performed during 6 months (Wehrli, 2000; 24 Kazadzis et al., 2018a). Details on requirements for performing Langley calibrations of reference 25 instruments by GAW-PFR and AERONET, and their uncertainties, are analysed in detail by Toledano et al. (2018). 26

27 **4. Data and methodology used in this study**

The AOD at each wavelength is obtained from the Beer-Bouguer-Lambert law (Thomason et al. 1982;
WMO, 2003) for radiometers collecting spectral direct sun measurements.

30
$$I(\lambda) = I_0(\lambda) \exp(-\tau m)$$

(1)

31 where $I(\lambda)$ is the direct sun signal at ground level at wavelength λ , $I_0(\lambda)$ is the extraterrestrial signal of the 32 instrument corrected by the Earth-Sun distance, and *m* is the optical air mass in the measurement path 33 (Kasten and Young, 1989). A detailed description of how AOD is obtained and the determination of 34 extraterrestrial constants by GAW-PFR and AERONET-Cimel are provided by Holben et al. (1994, 1998; 35 2001), Toledano et al. (2018), and Wehrli (2000; 2008b).

36 4.1. GAW-PFR and AERONET-Cimel data

1 GAW-PFR provides AOD values every 1 minute as an average of 10 sequential measurements of total 2 duration less than 1 second (20 ms for each channel), then dark current is measured, going to the sleep 3 mode until the next minute. AERONET-Cimel takes a sequence of three separate measurements (1-second per filter) in one minute interval (each one every 30 seconds). This sequence of measurements is called 4 5 "triplet" and it is performed every ~15 minutes for air masses lower than 2, and with higher frequency for lower solar elevations. Therefore, AERONET-Cimel provides AOD values for each triplet, at least, every 6 7 ~15 minutes. Note that AERONET-Cimel performs AOD measurements interspersed with sky radiance 8 measurements, whose duration varies throughout the day, and therefore the AOD measurements are not 9 necessarily provided at full minutes. We consider the 1-minute data as synchronous when GAW-PFR and 10 AERONET-Cimel AOD data were obtained with a difference of ~30 s.

11 GAW-PFR and AERONET-Cimel instruments use the same time reference. The synchronization 12 between PC and GAW-PFR data-logger was performed every 12 hours since 2005, and improved to 6 hours after 2013 using NTP servers via Internet. From 2005 to 2012 the time of the AERONET-Cimel reference 13 14 instruments was checked manually once per day using a handheld GPS. From 2012 onwards, the time was 15 adjusted automatically three times per day using the ASTWIN Cimel software. In turn, the PC time is 16 adjusted through the AEMET internal time server every 15 minutes. The AOD comparison has been performed using 1-minute synchronous data from the four closest channels of both instruments in the period 17 2005-2015 (more than 70000 data-pairs in each channel). Thus, in the case of GAW-PFR, the four available 18 19 channels of 368 nm, 412 nm, 500 nm and 862 nm were analysed, while in the case of AERONET-Cimel, 20 only the 380 nm, 440 nm, 500 nm and 870 nm channels were considered (Table 1). For the 500 nm channel, 21 the nominal wavelengths of the two networks differ by a maximum of 1.8 nm. However, the nominal 22 wavelengths in the rest of the compared channels present higher differences. Therefore, the AOD values of 23 the original GAW-PFR 368 nm, 412 nm and 862 nm channels have been interpolated or extrapolated to the 24 corresponding AERONET-Cimel channels (380 nm, 440 nm and 870 nm) using the Angström power law, 25 and the GAW-PFR AE calculated from the four PFR AOD measurements.

Synchronous AE data provided by both instruments have also been compared (see section 5.5). GAW-PFR determines AE using all four PFR wavelengths (Nyeki et al., 2015), while AERONET-Cimel uses different wavelength ranges (340-440 nm, 380-500 nm, 440-675 nm, 440-870 nm, 500-870 nm) (Eck et al., 1999). As a consequence, we have calculated a new AE for the Cimel radiometer using the four channels equivalent to those of the PFR.

In this study we have used the two versions of the AERONET database. Version 2 (V2;) has been used so far in many scientific publications in high impact journals, and Version 3 (V3) has been released just recently (Giles et al., 2019). In section 5.1., a comparison of V2 and V3 is presented. A total of three GAW-PFR and 15 AERONET-Cimel instruments have participated in this intercomparison study covering the period 2005-2015. Their corresponding reference numbers are shown in Table 2.

- 36
- 37
- 38

Table 1. Main features of the GAW-PFR (PFR (Wehrli, 2000, 2005, 2008a, b) and AERONET-Cimel (Holben et al., 1994, 1998; Torres et al., 2013) radiometers used in this study.

	GAW-PFR	AERONET-Cimel
Type of instrument	Standard version	Standard version
		Reference instrument
Type of observation	Automatic continuous direct sun irradiance	Automatic sun-sky tracking
Available standard channels	368, 412, 500, 862 nm	340, 380, 440, 500, 675, 870, 1020 nm
FWHM	5 nm	2 nm (340 nm), 4 nm (380 nm), 10 nm(VIS- NIR)
AOD uncertainty	± 0.01	0.002-0.009 spectrally dependent with the higher errors in the UV (Reference instruments) (Eck et al., 1999)
FOV (FWHM)	2.5°	1.3° (slope angle unknown)
	(1.2° plateau, 0.7° slope angle)	
Sun tracker	Any sun tracker with a resolution of at least 0.08°	Robot specifically designed by Cimel and controlled in conjunction with the radiometer
Temperature control/correction	Temperature controlled 20°C ± 0.5°C	Temperature correction to 1020 nm is applied in V2. Corrections from filter-specific temperature characterization in V3 for VIS and NIR spectral bands (Giles et al., 2019)
Power	Grid	Solar panels/grid
Data transmission	Local PC /FTP	Local PC / FTP Satellite transmission
Calibration	Comparison with reference triad. Additional in situ long- term Langleys	At least 10 good morning Langleys plots

6 *4.2. Cloud filtering*

The data matching in our comparison analysis was performed with synchronous 1-minute AOD values of both networks labelled with quality control (QC) flags that guarantee proven quality data not affected by the presence of clouds. In the case of the AERONET-Cimel network, the selected AOD data are Level 2 data from both V2 and V3 AERONET databases, which have been cloud filtered by the Smirnov algorithm (Smirnov et al., 2000), based on the triplet method, with a second-order temporal derivative constraint (McArthur et al., 2003), and visually screened in V2. The cloud screening in AERONET V3 has been completely automated, and notably improved, especially by refining the triplet variability and cirrus cloud detection and removal (Giles et al., 2019). These two cloud screening methods are able to detect rapid
changes in the atmosphere and remove those measurements in which AOD variability within the triplet is
higher than the following criteria:

- AOD triplet variability > MAX{0.02 or 0.03*AOD} at all wavelengths (V2)
- AOD triplet variability > MAX{0.01 or 0.015*AOD} at 675 nm, 870 nm, and 1020 nm
 simultaneously (V3)

The selection of these thresholds ensures the triplet average does not exceed 0.02 (V2) or 0.01 (V3)
within 1 min in case of low AOD conditions.

GAW-PFR cloud screening algorithms also use the Smirnov triplet measurement, and the second-order derivative check, but add a test for optically thick clouds with AOD_{500nm} > 2 (Kazadzis et al., 2018b). In the case of the GAW-PFR network (Wehrli, 2008a) the flags take the value 0 (cloudless conditions, no wavelength crossings and sun pointing within certain limits, more details in Kazadzis et al. (2018a)) for all those selected records.

14

4

Table 2. GAW-PFR and AERONET-Cimel instrument numbers used in this study in the period 2005-2015.
Data from Reference Cimel #398 was not upgraded to Level 2 in V3 during the period 12 July 2008 - 15
September 2008.

18

Instruments used in this study	Period 2005-2009	Period 2010-2015
GAW-PFR	2 instruments: #6,#25	2 instruments: #6,#21
AERONET-Cimel	13 instruments: #25,#44,#45,#79,#117,#140 #244,#245,#380,#382,#383,#398,#421	5 instruments: #244, #347, #380 #421, #548

19

20 *4.3. WMO traceability criteria*

The criterion for traceability used in this study follows the recommendation of the WMO (WMO, 2005) which states that 95% of the AOD measurements fall within the specified acceptance limits, taking the PFR as a reference:

24
$$U_{95} = \pm (0.005 + 0.010/m)$$
 (2)

where *m* is the optical air mass. Note that the U95 range is larger for smaller optical air mass. The acceptance limits proposed by WMO take into account, on the one hand, the uncertainty inherent in the calculations of the AOD, and on the other hand, the uncertainty associated with the calibration of the instrument. The latter, for the case of instruments with finite field of view direct transmissions, such as the PFR and the Cimel, is dominated by the influence of the top-of-the-atmosphere signal determined by Langley plot measurements, divided by the optical air mass. The first term of Eq. 2 (0.005) represents the maximum tolerance for the uncertainty due to the atmospheric parameters used for the AOD calculation (additional atmospheric trace gas corrections, and Rayleigh scattering). The second term describes the calibration related relative
 uncertainties. The WMO recommends an upper limit for the calibration uncertainty of 1%.

4.4. Modelling the impact of near-forward scattering on the AOD measured by the PFR and Cimel
radiometers

5 In order to study the impact of near-forward scattering on the irradiance measured by the PFR and Cimel instruments, a forward Monte Carlo model (Barker 1992, Barker 1996, Räisänen et al. 2003) was 6 7 employed. For the present work, the model was updated to account for the finite width of the solar disk. 8 The starting point of each photon was selected randomly within the solar disk, assuming a disk half-width 9 of 0.267° and the impact of limb darkening on the intensity distribution was included following Böhm-10 Vitense (1989). Some diagnostics were also added to keep track of the distribution of downwelling photons at the surface with respect to the angular distance from the centre of the sun. Gaseous absorption was 11 12 accounted for following Freidenreich and Ramaswamy (1999), while the Rayleigh scattering optical depth was computed using Bodhaine et al. (1999). 13

14

15 **5. Results**

16 5.1. Comparison of long-term AERONET V2 and V3 datasets at Izaña site

17 Since V3 has been released recently (Giles et al., 2019), we present a comparison between V2 and V3 18 for the Cimel channels 380 nm, 440 nm, 500 nm and 870 nm for the period 2005-2015. The results indicate 19 that for the Izaña site the agreement and consistency between the two AERONET versions is very high for 20 the four channels (R^2 >0.999) in full agreement with the results of the V2-V3 comparison reported by Giles 21 et al. (2019). It follows that the results of the comparison between GAW-PFR and the two versions of 22 AERONET are very similar as shown throughout this work. A detailed description of AERONET V3 and its improvements with respect to V2 is given in Giles et al. (2019). As such improvements depend on aerosol 23 24 type, according to the changes introduced in V3, for the high mountain site such as Izaña characterized by low background AOD values or, alternatively, by the presence of dust (no pollution or biomass burning 25 aerosols), the AOD differences between V2 and V3 are expected to be minimum as is confirmed in this 26 27 study (Figure 1).

However, it should be noted that AERONET V3 does not restrict the calculation of AOD to optical air masses less than 5.0 (Giles et al., 2019), as V2 does. This results in an increase in the number of solar measurements occurring in the early morning and the late evening. Consequently, the GAW-PFR comparisons with AERONET V3 consisted of ~ 9000 more data pairs than the GAW-PFR comparison with V2 (see Supplement S.1.).

33

34



Figure 1. AERONET Version 3 (V3) vs Version 2 (V2) AOD 1-minute data scatterplot at Izaña
Observatory for the period 2005-2015: a) 380 nm; b) 440 nm; c) 500 nm and d) 870 nm. The
corresponding equations of the linear fits, the coefficients of determination (R2), Mean Bias (MB), Root
Mean Square Error (RMSE) and the number of data pairs (N) used are included in each legend.

6 5.2. AERONET-Cimel AOD comparison with GAW-PFR data

7 The comparison with GAW-PFR AOD shows that the AOD from AERONET-Cimel radiometers meet 8 the WMO traceability criteria ("traceable AOD data" from now on) at 440 nm, 500 nm and 870 nm 9 channels. The lowest agreement is found in the UV channel (380 nm) with 92.7 % of the data, and the highest in the infrared channel (870 nm) with 98.0 % for V2 (Figure 2; Table 3) because this channel is less 10 11 affected by trace gases absorption. Almost identical results are obtained for V3 (Supplement S1 and 12 S2.1). However, in the first half of the comparison period (2005-2009) there were some mechanical problems in the solar tracker where the GAW-PFR was mounted on, which caused sporadic problems of 13 sun pointing. This finding was confirmed with data from the four-quadrant silicon detector (Wehrli, 2008a) 14 that showed diurnal variation of the PFR sensors position up to 0.3°. From 2010 onwards, the PFR was 15 16 mounted on an upgraded solar tracker of higher performance and precision. This reduced problems in sun 17 pointing, that were the main cause of most of the AOD discrepancies between PFR and Cimel, and therefore 18 not attributable to the instruments themselves.

1 In addition, since 2010, Cimel #244 has been in continuous operation for most of the time at the Izaña 2 Observatory, greatly simplifying calibration procedures and the corresponding data evaluation, and 3 minimizing errors of calibration uncertainties introduced by the use of a high number of radiometers in the intercomparison. During the 2010-2015 period, the fraction of traceable AOD measurements of the total 4 5 between the AERONET-Cimel radiometer and the GAW-PFR improves to 93.46 % in the 380 nm channel, and this percentage rises to 99.07 % for the 870 nm channel. Despite the technical differences between both 6 radiometers, described above, and the different calibration protocols, cloud screening and data processing 7 8 algorithms, the data series of both instruments, can be considered as equivalent, except for 380 nm, 9 according to the WMO traceability criteria defined previously (Eq. 2). This explains the excellent 10 agreement in the long-term AOD climatology shown for GAW-PFR and AERONET-Cimel in Toledano et 11 al. (2018).

12 We have compared the percentages of AERONET-Cimel AOD V2 data meeting the WMO criteria for 13 the four interpolated GAW-PFR channels with those of AERONET V3 (Table 3). A more detailed statistical 14 evaluation for different scenarios of aerosol loading (three ranges of AOD) and aerosol size (three ranges 15 of AE) for each compared channel has been performed (see Table 4). We observe that the poorest agreement 16 is obtained at the shorter wavelength channels (440 nm, and especially 380 nm). Kazadzis et al. (2018b) 17 also found a decrease in the percentage of AOD meeting the WMO criteria for 368 nm and 412 nm spectral bands during the Fourth WMO Filter Radiometer Comparison for aerosol optical depth measurements. As 18 19 these authors pointed out, the shorter the wavelength, the poorer the agreement because of several reasons: 20 AOD in the UV suffers from out-of-band or at least different blocking of the filters, small differences in 21 central wavelength or FWHM have a larger impact, the Rayleigh correction is more critical, and NO₂ absorptions are treated differently. Regarding the effect of the aerosol load and particle size on the AOD 22 23 differences, our results confirm the decrease of agreement between the two instruments for very large 24 particles coincident with almost pure dust (AE \leq 0.3), and high turbidity conditions (AOD > 0.1). However, 25 it should be noted that the percentage of data pairs in these situations is relatively low (e.g., 6% for AOD > 0.1, and 3.2% for AE > 0.25 at 380nm) with respect to the total data (Table 4). A similar result was reported 26 27 by Kim et al. (2008), who attributed these discrepancies to the possible spatial and temporal variability of 28 aerosols under larger optical depths in addition to the effect of the different FOV of both radiometers. In 29 our case, and according to previous studies on AOD climatology at IZO (Barreto et al., 2014), the presence 30 of high mineral dust burden when the station is within the SAL, does not necessarily imply lower 31 atmospheric stability conditions resulting in daily AOD means with greater standard deviation. For these 32 reasons, we assumed that the different FOV of these instruments is the main cause of part of the AOD 1-33 minute differences outside the U95 limits, under high AOD conditions. This issue is specifically addressed 34 in Section 5.3.



Figure 2. One-minute AOD data differences between AERONET-Cimel (V2) and GAW-PFR for (a) 380
nm (70838 data-pairs), (b) 440 nm (71645 data-pairs), (c) 500 nm (70833 data-pairs) and (d) 870 nm
(71660 data-pairs) for the period 2005-2015. Black dots correspond to the U95 limits. A small number
of outliers are out of the ±0.06 AOD differences range. Black arrows indicate a change of Reference
AERONET-Cimel radiometer and red arrows indicate a change of the GAW-PFR instrument.

7 Table 3. Percentage of AERONET-Cimel (V2 and V3) 1-minute AOD data meeting the WMO criteria
8 for the four interpolated GAW-PFR channels for the period 2005-2015.

ChannelV2 (%)V3 (%)380 nm92.792.3440 nm95.795.2500 nm95.895.7870 nm98.097.8

- 2 Table 4. Percentage of AERONET-Cimel 1-minute AOD data (V2) meeting the WMO criteria for the four
- 3 compared channels, and different AOD and AE scenarios for the period 2005-2015, number of data pairs
- 4 are shown in brackets. The last row corresponds to the total percentages for the sub-period 2010-2015.
 - AOD and AE traceability > 95% are marked in bold. Number of data pairs are in brackets.

% of data within WMO limits	380 nm	440 nm	500 nm	870 nm
AOD ≤ 0.05	94.4 (57008)	96.8 (59130)	97.0 (58572)	98.5 (60191)
$0.05 < \text{AOD} \le 0.10$	91.0 (4723)	93.1 (4850)	92.8 (4817)	94.2 (4908)
AOD > 0.10	75.0 (3938)	86.5 (4615)	85.1 (4466)	95.9 (5118)
$AE \le 0.25$	73.1 (2145)	82.3 (2417)	80.1 (2351)	96.2 (2824)
$0.25 < \mathrm{AE} \leq 0.6$	91.2 (5407)	96.2 (5810)	96.0 (5691)	97.9 (5911)
AE > 0.6	94.6 (55114)	96.9 (57089)	97.0 (56504)	98.7 (58146)
Total 2005-2015	92.7 (65669)	95.7 (68595)	95.8 (67855)	98.0 (70217)
Total 2010-2015	93.5 (41977)	97.4 (43745)	97.2 (43627)	99.1 (44498)

1

5

7 In general, the agreement obtained with the 1-minute AOD data is slightly lower than that obtained 8 during short campaigns, such as those reported by Barreto et al. (2016) at IZO (5566 data-pairs), with 9 agreements > 99 % for AOD_{870nm} and AOD_{500nm}. However, our results for AOD_{500nm} (> 95 % of 70833 data-10 pairs) are significantly better that that observed by Kazadzis et al. (2014) (~ 48 % of 468 data-pairs) 11 covering a relatively narrow range of AOD. In addition, short-term campaigns usually cover a small range 12 of AOD, and instruments are carefully and frequently supervised. On the contrary, during our 13 intercomparison over a period of 11 years, the operation of the instruments can be considered as the normal 14 operation of such a system. An additional interesting aspect of this study is that it is not a simple 15 intercomparison exercise between two instruments but a comparison of a number of instruments that acted 16 as reference instruments for the AERONET/Europe Network.

- 17
- Table 5. Basic skill-scores from the AOD intercomparison between GAW-PFR and AERONET-Cimel
 V2 for the period 2005-2015. The skill scores definitions are found in Huijnen and Eskes (2012).

Period	2005-2015				
Wavelengths (nm)	380	440	500	870	
Mean Bias (MB)	-0.0026	-0.0018	-0.0021	-0.0001	
Modified Normalized Mean Bias (MNMB)	-0.1301	-0.1046	-0.1474	0.0129	
Fractional Gross Error (FGE)	0.1727	0.1546	0.1918	0.1837	
Root Mean Square Error (RMSE)	0.0081	0.0070	0.0064	0.0049	
Pearson's correlation coefficient (r)	0.9910	0.9925	0.9939	0.9949	

Number of data-pairs	70838	71645	70833	71660
----------------------	-------	-------	-------	-------

2 In the first period (2005-2009), a total of 13 Cimel radiometers were used, while in the second period 3 (2010-2015), five Cimel radiometers have participated, and for much of this period, the Cimel #244 was operating as the permanent AERONET reference instrument at IZO. Once the most important causes of 4 non-traceability in the first period, which were associated with a poor pointing of GAW-PFR due to 5 6 problems in the sun-tracker, were discounted, we can conclude that there are no significant differences in 7 the percentages of traceable data between the two periods. This means that the continuous change of 8 Reference Cimel instruments used in the 2005-2009 period did not have a significant impact on AOD data 9 comparison differences. This provides proof of the consistency and homogeneity of the long AERONET-10 Cimel AOD data series, and their comparability with the GAW-PFR AOD data series, regardless of the 11 number of instruments used to generate these data series. In our study, with a number of comparison data-12 pairs one or two orders of magnitude higher than those used in short campaigns, the results shown in Table 13 4 can be considered as fairly good.

14 In addition to the traceability scores, we have introduced some basic skill scores corresponding to the AOD 15 intercomparison between GAW-PFR and AERONET-Cimel for the period 2005-2015 (Table 5) to be in line with previous studies that have performed short-term comparisons between these two instruments. The 16 17 definitions of the used skill scores can be found in Huijnen and Eskes (2012). The Pearson's correlation 18 coefficient (r) values of the PFR-Cimel 1-minute AOD data-pairs are higher than 0.99 in all channels. Concerning Mean Bias (MB) and Root Mean Square Error (RMSE) associated with AOD differences, our 19 20 results show quite similar skill scores to those found at Mauna Loa, USA for AOD_{500nm} (Kim et al., 2008), 21 although the number of data pairs used at IZO (\sim 71000) is much higher, and the AOD range of our study 22 is much larger than that of the comparison performed in Mauna Loa. Kim et al. (2008) summarize results 23 of previous short-term intensive studies (McArthur et al., 2003; Mitchell and Forgan, 2003; Kim et al., 24 2005; Schmid et al., 1999) carried out in stations where the radiometers were calibrated by intercomparison 25 with Reference instruments. These results show MB values to be within 0.01 bias, one order of magnitude 26 lower than in Mauna Loa and Izaña Observatories, highlighting the importance of having well calibrated 27 instruments to carry out these type of comparisons. For the period 2010-2015 (not shown here), as expected, 28 the RMSE and the Pearson's correlation improve slightly compared with the whole period 2005-2015.

In relation to the comparison between GAW-PFR and AERONET-Cimel V3, we have calculated the percentage of AERONET-Cimel 1-minute AOD data (V3) meeting the WMO criteria for optical air mass > 5.0 for the period 2005-2015 (Supplement S2.2). The results are somewhat poorer than for optical air mass < 5.0, since the solar elevation is very low. Only for the 870 nm channel 95% of the data meet with the WMO criteria, although the percentages of data in the 440 nm and 500 nm channels are close to this value. This would be the main reason to find slightly poorer traceability results with all V3 data compared to those found with V2 for which the AOD data are limited to optical air mass < 5.0.

36 5.3. Non-traceability assessment

1

As presented in Table 3, data outside the WMO traceability criteria vary from 2% for 870 nm up to 7.3% for 380 nm. In this section, the different possible causes of non-traceability in AOD are evaluated and, if possible, quantitatively estimated. In order to assess the relevance and quantitative impact of these causes, and estimate errors derived from a non-perfect AOD data synchronization, we first made an analysis on the natural variability of AOD in a very short time period (1 minute) shown below.

6 7

8

9 5.3.1. Short-time AOD variability

10 In order to determine the variability of AOD within one minute, we have performed two independent 11 analyses with AOD data from the PFR and Cimel for the 368/380 nm and 501/500nm channels during one 12 year (2013). On the one hand, and taking into account that GAW-PFR provides AOD every minute, we 13 have calculated all the AOD differences for each channel in the successive minutes. So we have the 14 variation of AOD from one minute to the next one during a whole year. On the other hand, for AERONET-15 Cimel, we have taken advantage of the triplets, since each triplet consists of three successive measurements 16 made in one minute. In this case, the strategy has been to calculate the standard deviation of the triplet AOD 17 measurements during a whole year.

18

19 Table 6. Percentage of AOD data with variability within 1 minute less than 0.01 and 0.005, respectively,

20 using AOD data from GAW-PFR (at 368 and 501nm) and AERONET-Cimel (at 380 and 500 nm) for

21 2013. A total of ~32000 data-pairs per channel have been used from GAW-PFR, and 20117 triplets

22 (60351 individual AOD measurements) from the Cimel#244 to calculate the AOD variability.

GAW- PFR							
Percentage of c	Percentage of data with 1-minute AOD variability (%)						
	368 nm	501 nm					
< 0.01	99.88	99.91					
< 0.005	99.21	99.35					
AERONET-Ci	mel						
Percentage of c	lata with 1-minute A	OD variability (%)					
	380 nm	500 nm					
< 0.01	99.87	99.99					
< 0.005	99.82	99.42					

23

The results obtained on the AOD variability in 1-minute from PFR data are very similar and consistent to those obtained with Cimel. Less than ~ 0.8% of the AOD data show variability higher than 0.005 in all wavelength ranges. It should be noted that the possible instrumental noise is included in this variability, so
that the actual natural AOD variability would be, in any case, lower than that expressed in Table 6. The
percentage of data with 1-minute AOD variability for all four GAW-PFR channels are given in Supplement
\$3.1.

5 We have also determined the percentage of 1-minute AOD data from the Cimel triplets (year 2013) 6 whose diurnal range of variation AOD_{max} - $AOD_{min} > 0.015$, for several AOD intervals. Note that this value 7 is half of the WMO traceability interval when m = 1 (maximum possible interval) (see Eq.2). The results 8 shown in S3.2 indicate that the 1-minute AOD variability is responsible for only 0.11% (0.01%) of 1-minute 9 Cimel AOD values outside the WMO limits in the [0-0.03] AOD range (pristine conditions) for 380 nm 10 (500 nm). The AOD variability maximizes in the 0.1-1 AOD range causing 2.31% and 1.69% of the AOD data outside WMO limits for 380 and 500 nm, respectively. This last scenario corresponds, as expected, to 11 12 changes of air masses, such as transitions from pristine to dusty conditions, and vice versa, or to sharp onset 13 and disappearance of very sporadic biomass burning plumes. In any case, the AOD data with 1-minute 14 variability exceeding 0.02 (V2) or 0.01 (V3) are filtered by AERONET (see Section 4.2) and therefore are 15 not included in the GAW-PFR and AERONET-Cimel comparison.

16 These results indicate that the natural AOD variability is very low thus the non-ideal measurement 17 synchronization cannot explain the percentages of non-traceable AOD cases shown in Tables 3 and 4.

18 19

20 5.3.2. Uncertainties of GAW-PFR channel interpolation to AERONET-Cimel channels

The interpolation of the CIMEL AODs to the PFR AOD wavelengths can be one of the sources of uncertainty in this comparison assessment. The greatest uncertainty arises in the extrapolation of the AOD_{412 nm} of the PFR to the Cimel wavelength 440 nm. Using the Angström formula we have calculated that for an uncertainty of ± 0.5 in the AE the introduced uncertainty in the AOD extrapolation from 412 nm to 440 nm is ~5% (i.e., 0.005 for AOD_{412nm}=0.1). The introduced uncertainty in AOD extrapolation is reduced to ~2% for an uncertainty of ± 0.3 in AE. For all other AOD interpolations the errors are smaller.

27 5.3.3 Calibration related errors

28 As described in Section 3, the calibration procedures of the AERONET-Cimel and GAW-PFR 29 radiometers are different. While in the case of GAW-PFR, frequent calibrations are established throughout 30 the year and the calibration value is linearly interpolated in time, in AERONET- Cimel a constant 31 calibration value is assumed in the intermediate period between two consecutive calibrations carried out on 32 an annual basis. The typical calibration uncertainty for a single Langley plot is 0.7-0.9 % (at the 95 % 33 confidence level), and it is reduced to 0.4 % in the case of IZO when averaging at least 10 Langley-derived 34 extraterrestrial constants (which is the normal procedure) (Toledano et al., 2018). Regarding the GAW-35 PFR radiometers operated at IZO, a direct yearly comparison of the Langley based V_o's with the reference 36 triad at PMOD/WRC showed differences lower than 1 % for all channels for the 2005-2015 period.

1 A not sufficiently accurate determination of the calibration constant results in a fictitious AOD diurnal evolution presenting a concave or convex characteristic curve due to the calibration error dependence on 2 solar air mass. The largest error occurs in the central part of the day (lower air masses), mainly, in clean 3 4 days with very low aerosol load (< 0.02 in 500 nm) as reported by Romero and Cuevas (2002) and Cachorro 5 et al. (2004), and as can be derived from Equation 2. According to Cachorro et al. (2004, 2008) fictitious differences of up to 0.06 between the minimum and the maximum AOD can be recorded in a day with 6 7 constant AOD as a result of a non-accurate calibration or non-cleaned instruments. However, these fictitious 8 differences in AOD depend on the related calibration magnitude errors.

We have represented the AOD differences between GAW-PFR and AERONET-Cimel versus optical air mass for the four channels for pristine conditions (AOD_{500nm} \leq 0.03) for both V2 and V3 (See Supplement S4). It should be noted that although the few outliers are evenly distributed throughout the whole airmass range, they are not equally distributed with respect to the zero of the AOD difference, but there is a bias with positive large outliers (higher Cimel AOD), already reported by Nyeki et al. (2013), and small negative outliers for optical air mass lower than 2.

The total percentage of AOD traceable data pairs under pristine conditions (AOD_{500nm} ≤ 0.03) is very 15 16 high for all wavelengths (>97.7 %) falling within the U95 limits (Table 7), except for 380 nm. There is no dependence on 1-minute AOD differences with optical air mass for 440, 500 and 870 nm, and a slight 17 dependence for 380 nm (Table 7) with higher percentage of AOD differences outside the U95 limits at 18 19 lower optical air masses. For the extended range of optical air mass > 5 in V3, the AOD differences do not 20 increase with optical air mass (Supplement S5). The lower traceability at 380 nm for low air masses is 21 especially clear in V3 with 92.9% of traceable data (See Supplement S5). This result is consistent with the 22 fact that the highest uncertainty in the determination of the calibration constants is observed in the UV 23 range, and the lowest uncertainty in the near-infrared channel (Eck et al., 1999; Jarosyawski et al. 2003; 24 Toledano et al., 2018). This is attributable to an imperfect calibration, or to very small changes in the filters' 25 transmittance, that can only be detectable in extreme conditions: UV range, very low optical air mass, and 26 pristine conditions. According to Toledano et al. (2018), the greatest variance in the extraterrestrial constant 27 in the UV channel could be due to a number of factors: 1) higher AOD variability at the shorter wavelengths; 28 2) filter blocking issues; and 3) temperature effects affecting AERONET-Cimel instruments that have not 29 been accounted for in the UV range.

- 30Table 7. Percentage of 1-minute AOD data (V2) meeting the WMO criteria for each wavelength for31different optical air mass intervals under pristine conditions ($AOD_{500nm} \le 0.03$) in the period 2005-322015. See Supplement S5 for equivalent results with V3.
- 33

Percentage of AOD differences within the U95 limits	Total	$1 \le m < 2$	$2 \le m < 3$	$3 \le m < 4$	$4 \le m < 5$
AOD500 <i>nm</i> ≤ 0.03	(%)	(%)	(%)	(%)	(%)
380 nm	95.8	94.5	96.0	97.4	97.2
440 nm	97.9	97.9	97.7	98.2	97.7

500 nm	98.3	98.4	98.1	98.6	98.4
870 nm	99.2	99.4	99.3	99.2	98.6

2 The correct cause attribution of each outlier would require manual inspection and additional specific 3 information on instrumental checking and maintenance information that is not always available. We have 4 investigated in more detail the origin of the outliers and whether one of the two instruments predominantly caused them. Thus, we have calculated for the non-traceable AOD data the diurnal range of AOD variation 5 6 (maximum value minus minimum value of AOD in one day) at 380 nm for each instrument under pristine conditions (Figure 3) using Cimel AOD_{500nm} daily mean < 0.03 to select the pristine days. According to this 7 8 approach, the instrument that shows the highest daytime AOD range is the one that is responsible for the 9 outlier. As the wavelength increases both the number of outliers and the magnitude thereof decreases 10 significantly (Supplement S6). Then, we identified those outliers with a diurnal AOD range higher than 11 25% of the mean daily AOD value and investigated their possible causes. A total of 51 cases for GAW-12 PFR and 81 cases for AERONET Cimel V3 were obtained and analysed in detail, using auxiliary 13 information, such as 1-minute in-situ meteorological data, 5-minute all-sky images, 1-minute BSRN data, 14 and satellite imagery (not shown here). We obtained the percentage of AOD outliers of GAW-PFR and 15 AERONET Cimel (V3) for which a certain cause has been identified, such as calibration uncertainties, 16 cloud screening algorithm failures, mixture of the two previous causes, poor sun pointing, or not well-17 defined causes (electronic problems, humidity inside the lenses, filter dirtiness, obstruction of the lenses 18 collimators, insects on the optics outside, etc.) (see Supplement S7).

19 From the analysis of these cases, under the conditions described above, it should be noted that $\sim 44\%$ 20 of the cases with fictitious AOD diurnal cycles were due to small uncertainties in the calibration of 21 AERONET-Cimel (V3), while for this same cause ~ 8% of cases were identified in GAW-PFR. Some 22 examples of AOD non-traceability for both AERONET-Cimel and GAW-PFR in the ~380 nm channel are 23 shown in Supplement S8. The fictitious diurnal AOD cycle is mainly visible in the UV channels as shown 24 in the examples reported in Supplement S9, where the convex or concave diurnal AOD curvature 25 symmetrical at noon provides a hint of calibration inaccuracies. Note that the fictitious diurnal AOD can 26 be more easily identified under very low AOD conditions. We should emphasize that the rare finding of 27 small calibration inaccuracies in a high mountain site with pristine skies and stable atmosphere does not 28 detract from the quality of any instrument as they often measure near or below the detection limit. Simply, 29 these small inaccuracies are the result of limitations in the photometric measurement technique.



1

Figure 3. AOD diurnal range variation (maximum value minus minimum value of AOD in one day) at
380nm corresponding to AOD outliers (non-traceable AOD) under pristine conditions (AOD_{Cimel-500nm}≤
0.03) in the period 2005-2015 for AERONET V2 (a) and V3 (b).

- 5
- 6

7 5.3.4. Differences in cloud-screening and sun tracking

8 We have examined the effect that the presence of clouds might have on AOD differences and the 9 percentage of cases outside the U95 limits. The impact of clouds on AOD differences only occurs when 10 both GAW-PFR and AERONET-Cimel cloud screening algorithms fail to identify clouds in the direct sun 11 path. In order to assess the impact that cloud conditions might cause on AOD traceability, we have used 12 the concept of daily fractions of clear sky (FCS) that has been applied before to solar radiation data at IZO (García et al., 2014). FCS represents the percentage of observed sunshine hours in a day with respect to the maximum possible sunshine hours in that day. The higher the daily FCS, the higher the clear sky percentage we have on that day. The percentages of traceable and non-traceable AOD data versus FCS values grouped into five intervals are shown in Table 8. It should be emphasized that the number of cases linked with FCS between 0% and 60% are less than 2% of the total cases. As the fraction of clear sky increases, the percentage of traceable AOD data significantly exceeds the number of non-traceable AOD data. The percentage of traceable data is especially large (> 90 %) when FCS > 80 % (almost clear skies).

8 This is the FCS range in which a significant percentage of days with cases presenting scattered clouds 9 are recorded, which qualitatively confirms that V3 has introduced more efficient cloud screening than V2. 10 However, the real impact of clouds on AOD traceability at IZO is very low due to its special characteristics 11 of a high mountain station with very little cloudiness. Therefore, in practice, the possible impact of clouds 12 on the non-traceability of AOD data-pairs is insignificant at IZO. GAW-PFR and AERONET-Cimel cloud 13 screening algorithms provide successful identification on clear direct-sun conditions during cloudy skies 14 (FCS < 40 %) for 99.75 % of the cases, excluding those with very thin clouds.

In the particular case of Izaña there are some very specific cloud scenarios in which cloud screening algorithms could fail resulting in non-AOD traceability: 1) Cirrus clouds (see Supplement S10); and 2) low clouds (stratocumulus) that sometimes exceed the observatory height level (see Supplement S11). As can be deduced from the analysis of these cloud cases, the impact of the different types of clouds on AOD retrieval is very complex and further specific investigations are required in order to understand the reasons behind failures in the GAW-PFR and AERONET-Cimel cloud screening algorithms.

- 21
- 22
- 23
- 24
- 25
- 26
- 27
- 28

Table 8. Percentage of AOD data within the U95 limits for each channel and 5 daily fractions of clear
sky (FCS) intervals. In brackets, relative frequency of each FCS interval for AERONET V2 and V3,
respectively. In bold, the percentages of V3 that are greater than those of V2.

32

	380	nm	440	nm	500	nm	870	nm
	V2	V3	V2	V3	V2	V3	V2	V3
0%≤FCS<20%	47.6	44.4	43.5	44.4	47.6	44.4	87.0	92.6
(0.03%) (0.04%)								
20%≤FCS<40%	69.3	76.6	73.3	82.2	73.6	80.8	86.3	94.1
(0.22%) (0.22%)								

			1					
40%≤FCS<60%	79.1	77.5	87.8	84.8	88.8	87.2	91.9	92.0
(1.08%) (1.09%)								
60%≤FCS<80%	88.4	89.6	93.9	93.9	93.4	94.4	97.8	97.6
(7.10%) (7.17%)								
FCS≥80%	93.3	92.8	96.2	95.6	96.2	96.1	98.3	98.1
(91.6%) (91.5%)								

1 5.3.5. Rayleigh scattering, absorption by O₃ and NO₂ corrections

2 In this section, we evaluate the possible impact on the 1-minute AOD data outside the U95 limits due 3 to the different processing of each network regarding the correction by Rayleigh scattering and by the light 4 absorption of column O₃ and NO₂. Although GAW-PFR and AERONET-Cimel use spectral channels with 5 weak absorption by atmospheric gases, AOD can only be determined if optical depth contributions from those gases are well estimated and subtracted from the total optical depth (τ). GAW-PFR and AERONET-6 7 Cimel separate the contributions of the molecules (Rayleigh scattering, τ_R), aerosols (τ_a ; in this study 8 referred to as AOD) and absorbing gases: total column ozone (τ_{O3}) and nitrogen dioxide (τ_{NO2}) due to their 9 different optical air masses at low solar elevation:

$$I(\lambda) = IO(\lambda)\exp(\tau_R m_R + AODm_a + \tau_{O3}m_{O3} + \tau_{NO2}m_{NO2})$$
(3)

So, AOD can be derived from: $I_{L_{2}}(\lambda)$

10
$$AOD = \frac{1}{m_a} \left(ln \frac{I_0(\lambda)}{I(\lambda)} - \tau_R m_R - \tau_{O3} m_{O3} - \tau_{NO2} m_{NO2} \right)$$
(4)

- 11 5.3.5.1 Rayleigh scattering
- 12 13

14

The Rayleigh scattering contribution to total optical depth would be:

$$\tau_R = \delta_R \frac{m_R}{m_a} \tag{5}$$

15 where, m_R is calculated, according to Kasten and Young (1989):

1

$$m_R = \frac{1}{\sin\theta + 0.50572(\theta + 6.07995)^{(-1.6364)}} \tag{6}$$

17 and m_a , according to Kasten (1966), has the following expression:

18
$$m_a = \frac{1}{\sin\theta + 0.0548(\theta + 2.65)^{(-1.452)}}$$
(7)

- 19 where θ is the sun elevation, and δ_R can be expressed as (Bodhaine et al., 1999): 20 $\delta_R(\lambda) = 0.00864\lambda^{-(3.916+0.074\lambda+\frac{0.050}{\lambda})}\frac{P}{P_o}$ (8)
- 21 where $P_o = 1013.25$ hPa, λ is the wavelength in microns (μ) and P is the pressure in hPa at the

22 measurement site. The depolarization factor recommended by (Young, 1980) is already included in Eq.

23 8. From Eq. 8, we can derive the differences in τ_R contribution $(4\tau_R)$:

24
$$\Delta \tau_R = (0.00864\lambda^{-(3.916+0.074\lambda+\frac{0.050}{\lambda})}) \frac{1}{1013.25} \frac{m_R}{m_a})(P_{PFR} - P_{Cimel})$$
(9)

1 Accordingly, the main $\Delta \tau_R$ from GAW-PFR and AERONET-Cimel can arise from the different way 2 the two instruments obtain the atmospheric pressure (P_{PFR} and P_{Cimel} , respectively). While AERONET-3 Cimel obtains the site station pressure from the National Centers for Environmental Prediction (NCEP) and 4 the National Center for Atmospheric Research (NCAR) reanalysis at standard levels, GAW-PFR has a 5 solid-state pressure transducer in the control box to read barometric pressure simultaneously with each PFR measurement. As Giles et al. (2018) have stated, the expected error in the station pressure P_{Cimel} is generally 6 7 < 2 hPa provided the elevation of the station is well-known and the weather conditions are stable. In order 8 to assess this possible difference, we have compared the 1-minute synchronous pressure data of both 9 instruments, and the corresponding 1-minute $\Delta \tau_R$ from Eq. 9. Note that, in practice, this comparison is 10 performed at 6-hour intervals since the NCEP/NCAR reanalysis data are available routinely with this 11 temporal resolution (Kalnay et al., 1996). The results are depicted in Figure 4.

12 The results indicate that most of the 1-minute pressure differences are within \pm 5 hPa (Figure 4a), 13 resulting in 1-minute $\Delta \tau_R$ data within ± 0.001 . However, when pressure differences are significantly higher, such as those registered at the end of 2014 (> 30 hPa) (Figure 4a), $\Delta \tau_R$ increases significantly (~ 0.01) 14 (Figure 4b). However, it should be noted that only 99 AOD data pairs have been registered for which the 15 16 pressure difference between PFR and Cimel is greater than 20 hPa at 870nm and 440nm, and one AOD 17 data pair at 500nm and 380nm channels. Taking into account that the accuracy of the new barometers built into new radiometers is ~3 hPa only dramatic barometer malfunctioning could cause $\Delta \tau_R > 0.01$. As stated 18 19 by Kazadzis et al. (2018b), the use of erroneous pressure values can lead to wavelength-dependent AOD 20 errors and to large errors in AE. However, these flagrant barometer malfunctions are quickly detected and 21 easily corrected if there are other pressure measurements at the station, as is the case in Izaña.



1

Figure 4. (a) 1-minute pressure data (hPa) from GAW-PFR and 6-hour pressure data at Izaña Observatory altitude from the National Centers for Environmental Prediction (NCEP) and the National Center for Atmospheric Research (NCAR) reanalysis for the case of AERONET-Cimel, and (b) corresponding 1-minute $\Delta \tau_R$ caused by pressure differences in the period 2005-2015.

8 5.3.5.2 Differences in O₃ absorption

9

10

6

7

The O₃ related optical depth is determined with the following expression:

11
$$\tau_{O3}(\lambda) = \sigma_{O3}(\lambda) \frac{O_3}{1000} \frac{m_{O3}}{m_a}$$
(10)

Where O₃ is expressed in Dobson units (DU), and the absorption coefficients ($\sigma_{O3}(\lambda)$) take the following values (Gueymard, 1995): 0.0026 cm⁻¹ (440 nm), 0.03150 cm⁻¹ (500 nm), and 0.00133 cm⁻¹ (870 nm). The ozone absorption is maximum in the 500 nm channel and practically zero in the 380 nm channel. GAW-PFR uses the following expression for m_{O3} (Komhyr, 1980):

16
$$m_{O3} = \frac{R+h}{\sqrt{(R+h)^2 - (R+r)^2(\cos\theta)^2}}$$
(11)

1 where R = 6370 km is the mean radius of the Earth, r = 2.370 km is the altitude of the station, h = 22 km is 2 the estimated height of the ozone layer, and θ is the solar elevation. However, AERONET-Cimel uses an 3 updated expression (Komhyr et al., 1989) in which h is not fixed and takes a value as a function of the latitude, and the absorption coefficients are obtained for each particular filter using the spectral response 4 5 provided by the manufacturer. For most of the period covered in this study, measured total ozone values from IZO (Brewer spectrometer) were used to calculate τ_{O3} (Wehrli, 2008a). If no Brewer data is available, 6 7 data are retrieved from the Total Ozone Mapping Spectrometer (TOMS) or more recently from the Ozone 8 Monitoring Instrument (OMI) (McPeters et al., 2015) for daily operations (Kazadzis et al., 2018b). In the 9 case of Izaña, if the OMI overpass fails, GAW-PFR uses the Brewer O₃ climatology. Concerning 10 AERONET-Cimel Version 2, a NASA TOMS 1° x 1.25° resolution O₃ climatology is used. From Eq. 10, 11 the differences in O₃ optical depth $\Delta \tau_{O3}$ can be derived:

$$\Delta \tau_{O3} = \sigma_{O3}(\lambda) \frac{1}{1000} \frac{m_{O3}(O_{3PFR} - O_{3Cimel})}{m_a}$$
(12)

12

13 The largest influence of total ozone data uncertainty in τ_{O3} occurs at 500 nm (Figure 5). According to 14 Wehrli (2008b) and Kazadzis et al. (2018b), total ozone needs to be determined to \pm 30 DU or 10 % of 15 typical values, to ensure an uncertainty of \pm 0.001 in τ_{03} at 500 nm. In the case of the GAW-PFR / AERONET-Cimel comparison, and due to the very different method in which both networks obtained O₃ 16 values for their corresponding corrections, the ozone differences found on some days (1761 out of 71965 17 18 days; 2.4 %) are very large (> 40 DU), exceeding a difference in the ozone optical depth of 0.001. Even so, 19 the potential contribution to AOD differences outside the U95 limits between the two networks is 20 negligible. Total O₃ over IZO shows a relatively small amplitude throughout the year, but both surface 21 ozone concentrations and column ozone amount could sharply increase under the influence of cut-off lows 22 injecting air from the high-mid troposphere into the lower subtropical troposphere, which is not uncommon 23 in spring and the first half of summer (Cuevas et al., 2015; Kentarchos et al., 2000). In addition through 24 exchange processes in the Upper Troposphere Lower Stratosphere (UTLS) due to the presence of the 25 subtropical jet (mainly from February to April) (Rodríguez-Franco and Cuevas, 2013). However, if we 26 wanted to repeat this traceability study of 1-minute AOD data in mid or high latitude stations where sharp 27 O₃ variations (several tens of DU) could be registered in a few hours, the correction of 1-minute AOD 28 measurements by τ_{O3} might be a challenging issue.



Figure 5. (a) Total O_3 used by GAW-PFR (measured Brewer O_3 values from IZO, OMI O_3 overpass or Brewer O_3 climatology) and AERONET-Cimel (TOMS O_3 climatology), and (b) $\Delta \tau_{O3}$ (λ) caused by differences in daily total O_3 between the two instruments in the period 2005-2015.

7 5.3.5.3 Differences in NO₂ absorption

8 AERONET-Cimel applies a correction by absorption of NO₂, but GAW-PFR does not include this 9 correction. AERONET-Cimel V2 obtains daily total NO₂ data from a 0.25° x 0.25° resolution NO₂ monthly 10 climatology obtained from the ESA Scanning Imaging Absorption SpectroMeter for Atmospheric 11 CHartographY (SCIAMACHY) (Eskes and Boersma, 2003). AERONET-Cimel V3 uses a geographic and 12 temporally dependent multiyear monthly climatology from the Ozone Monitoring Instrument (OMI) NO2 13 concentration (Giles et al., 2019). In order to assess the contribution to AERONET-Cimel 1-minute AOD 14 data non-traceability by NO₂ absorption we have to estimate the NO₂ optical depth ($\tau_{NO2}(\lambda)$) of AERONET-15 Cimel since GAW-PFR does not perform this correction. Analogously to $\Delta \tau_{O3}$, the differences in nitrogen 16 dioxide optical depth $\Delta \tau NO_2$ can be obtained from:

17
$$\Delta \tau_{NO2} = \sigma_{NO2}(\lambda) \frac{1}{1000} \frac{m_{NO2}}{m_a} (-NO_{2Cimel})$$
(13)

1 Where m_a is given by Eq. 7, NO_{2Cimel} (DU) is the daily total NO₂ used by AERONET-Cimel, $\sigma_{NO2}(\lambda)$ is 2 the NO₂ absorption (Gueymard, 1995) weighted by the specific filter response: 15.6 cm⁻¹ (380 nm), 12.3 3 cm⁻¹ (440 nm), and 4.62 cm⁻¹ (500 nm). Finally m_{NO2} has the following expression (Gueymard, 1995):

$$m_{NO2} = \frac{1}{\sin\theta + 602.30(90 - \theta)^{0.5}(27.96 + \theta)^{-3.4536}}$$
(14)

5

4





7 Figure 6. (a) NO_2 monthly climatology obtained from the ESA SCanning Imaging Absorption 8 SpectroMeter for Atmospheric CHartographY (SCIAMACHY), used by AERONET-Cimel at IZO, and 9 (b) $\Delta \tau_{NO2}(\lambda)$ caused by differences in daily total NO_2 between GAW-PFR and AERONET-Cimel in the 10 period 2005-2015. Note that GAW-PFR does not take into account the correction for the NO_2 11 absorption.

12

6

- 13
- 14
- 14
- 15

2 3

Table 9. Percentage (%) of additional traceable AERONET-Cimel AOD 1-minute data (V2 and V3) after correcting by pressure, and total column O_3 and NO_2 for the period 2005-2015.

5

4

6

Channel	Increment (%) of traceable AOD data after P, O ₃ and NO ₂ corrections				
	V2	V3			
380 nm	1.3	1.7			
440 nm	0.2	0.3			
500 nm	0.3	0.1			
870 nm	~0.0	~0.1			

7

8 In Figure 6a the total NO₂ used by AERONET-Cimel to evaluate $\tau_{NO2}(\lambda)$ is depicted. Figure 6b shows 9 the $\Delta \tau_{NO2}(\lambda)$ caused by differences in daily total NO₂ between GAW-PFR and AERONET-Cimel. $\Delta \tau_{NO2}$ is 10 of the order of 10^{-3} for 380 and 440 nm channels, while, for 500 nm channel, it is of the order of 10^{-4} . 11 However, it should be noted that an impact on AOD calculation is expected when replicating similar 12 analysis in highly NO₂ polluted regions. Such cases include large industrial cities from East Asia and 13 Central and Eastern Europe (e.g., Chubarova et al., 2016).

Taking into account the corrections for Rayleigh scattering and for the absorptions by O_3 and NO_2 , we have calculated the additional traceable AOD data (Table 9). This percentage is maximum at 380 nm with 1.3% (V2) and 1.7% (V3) of the whole dataset. The 870 nm channel is only affected by the Rayleigh correction component and therefore the increment of traceable data after the mentioned corrections is minimal.

5.4. GAW PFR and AERONET-Cimel comparison under high AOD conditions: the impact of dust forward scattering for different FOVs

21 When we present the AOD differences between AERONET-Cimel and GAW-PFR versus AOD 22 (GAW-PFR) for AOD > 0.1 (dusty conditions), we note that AERONET-Cimel shows slightly higher AOD 23 values than GAW-PFR (Figure 7). Similar results for V3 are shown in S12. In fact, the percentage of data 24 outside the U95 limits increases as AOD increases (Table 10), so for dust-related aerosol conditions 25 $(AOD_{500nm} > 0.3)$ the percentage of AOD data outside the U95 limits is > 50 % for 380 nm and 440 nm 26 (Table 10, percentages in brackets). Similar results are found when using AERONET V3 (see Supplement 27 S13). Taking into account the number of data compared with the total cases, these results show a small but 28 non-negligible percentage of AOD differences outside the U95 limits for AOD > 0.1, ranging from ~ 0.3 29 % at 870 nm to ~ 1.9 % at 380 nm (Table 10).





Figure 7. Actual AOD differences between AERONET-Cimel V2 and GAW-PFR vs AOD_{PFR} at (a) 380 nm (b) and 500 nm for the period 2005-2015. The fitting line has been calculated with those data points with AOD > 0.1 and Cimel-PFR AOD difference > 0. The number of data used in the plots are indicated in the legend. The percentage of non-traceable AOD data with these conditions is ~24% for 380 nm, and ~8% for 500 nm. Note that some traceable (black) points show larger AOD differences than nontraceable (red) points because of the air mass dependence of the WMO traceability criterion.

9

10

- 1 Table 10. Percentage of AERONET V2 AOD data outside the U95 limits at 380, 440, 500 and 870 nm
- 2 channels and for three AOD_{500nm} thresholds respect to all data and respect to all data for each AOD
 - 3 *interval (in brackets).*

	Percentage of AOD data outside the U95 limits (%)						
	AOD _{500nm} >0.1	AOD _{500nm} >0.2	AOD _{500nm} >0.3				
380 nm	1.9 (25.0)	1.2 (47.2)	0.5 (59.8)				
440 nm	1.0 (13.5)	0.8 (32.0)	0.5 (57.6)				
500 nm	0.6 (8.0)	0.5 (18.7	0.3 (39.3)				
870 nm	0.3 (4.1)	0.2 (6.4)	0.1 (14.0)				

5 Aerosol forward scattering within the FOV of various instruments and calculated AOD was 6 investigated some decades ago by Grassl (1971) who determined that at AOD=1 the circumsolar radiation 7 increases by >10% the incoming radiation. Russell et al (2004), using dust and marine aerosols data, 8 quantified the effect of diffuse light for common sun photometer FOV. They reported that the correction to 9 AOD is negligible (<1% of AOD) for sun photometers with narrow FOV (< 2°), which is greater than the 10 Cimel FOV and slightly smaller than the PFR FOV (2.5°). Sinyuk et al. (2012) assessed the impact of the 11 forward scattering aerosol on the uncertainty of the AERONET AOD, concluding that only dust aerosol 12 with high AOD and low solar elevation could cause a significant bias in AOD (> 0.01).

GAW-PFR has double the FOV (2.5°; Wehrli (2000)) compared to the AERONET-Cimel (1.3° \pm 4.8 13 14 %; Torres et al., 2013), so it is reasonable to expect that it is more affected by the circumsolar radiation 15 than the AERONET-Cimel radiometer. Taking advantage of the fact that Saharan dust intrusions regularly affect IZO, we provide a detailed analysis on the impact that dust forward scattering causes on the AOD 16 retrieval of the two radiometers with different FOV, explaining the AOD differences under moderate-to-17 18 high dust load (AOD > 0.1) conditions. For this purpose we have used a forward Monte Carlo model (see 19 section 4.4) with which we perform simulations that include accurate dust aerosol near-forward scattering 20 effects.

21 Dust aerosol single-scattering properties were computed using Mie theory, assuming a refractive index 22 of 1.47+0.0025i at the wavelengths 380 nm, 440 nm and 500 nm and 1.46+0.012i at 870 nm, based on 23 AERONET measurements at IZO. Seven values of aerosol effective radius (re) in the range 0.2 to 3.0 µm 24 were considered, and a lognormal size distribution with a geometric standard deviation of 2 was assumed. 25 A mid-latitude summer atmospheric profile starting from the Izaña altitude (2.4 km a.s.l.) was assumed, 26 with the aerosol layer located at 5-6 km a.s.l. (typical of summertime). A spectrally uniform surface albedo 27 of 0.11 was employed. Computations were performed for nine AOD values (AOD= 0, 0.1, 0.2, 0.3, 0.4, 28 0.5, 0.6, 0.8, and 1.0) and for five solar elevation angles (θ =80°, 60°, 45°, 30° and 20°). The Monte Carlo

1 model assumes a plane-parallel atmosphere, so the air mass factor is $m=1/\sin\theta$. Ten million photons were

2 used for each case and wavelength.

Supplement S14 shows the ratio of scattered to direct radiation for cases with AOD up to 0.5. The ratio increases with increasing r_e , as the aerosol forward-scattering peak grows stronger. In the case of Saharan dust intrusions at IZO, the median r_e determined from both AERONET data inversion and the in-situ aerodynamic particle sizer (APS) analyzer is ~1.5 µm. This value agrees with the dust size distribution found during SAMUM-2 during long-range transport regime (Weinzierl et al., 2011). For this particle size, the ratio of scattered to direct radiation is ~3 times larger for FOV of 2.5° than FOV of 1.3°.

9 The error in the retrieved AOD due to scattered radiation within the instrument FOV was evaluated by 10 comparing the apparent AODs, defined as:

12
$$AOD_{app,PFR} = -\frac{1}{m} ln \frac{F_{PFR}}{F_{PFR}(AOD=0)}$$
(15)

13
$$AOD_{app,Cimel} = -\frac{1}{m} ln \frac{F_{Cimel}}{F_{Cimel}(AOD=0)}$$
(16)

14 with the true AOD

15
$$AOD_{true} = -\frac{1}{m} ln \frac{F_{dir}}{F_{dir}(AOD=0)}.$$
 (17)

Here, F_{dir} is the irradiance due to direct (i.e., non-scattered) radiation, and F_{PFR} (F_{Cimel}) is the total irradiance that would be measured by the PFR (Cimel) radiometer, considering the instrument FOV and the FOV angular function. The relative error in AOD depends strongly on the particle size but it is fairly constant for each r_e value considered (see Supplement S15). For $r_e \sim 1.5 \mu m$, the relative error in AOD at 380 nm (500 nm) is ~1.6% (1.0%) for Cimel, and ~5% (~3%) for PFR. These errors are in good agreement with those estimated by Russell at al. (2004), and slightly higher than the relative AOD error of 0.7% due to coarse dust aerosol forward scattering reported by Eck et al. (1999).

23 The Monte-Carlo-simulated relative differences in retrieved AOD (in %) that would result from the 24 scattered radiation within the FOV of the PFR and Cimel instruments, and the difference in retrieved AOD 25 between PFR and Cimel as a function of the AOD retrieved with PFR, for 380 nm and 500 nm, are shown 26 in Figure 8. The main results of these simulations are: 1) the higher FOV of the PFR, compared to that of 27 the Cimel, results in lower AOD values for the PFR; 2) the fractional AOD difference related to the different 28 FOVs of PFR and Cimel is fairly constant for any aerosol effective radius, but increases with increasing 29 the effective radius; and 3) this fact might explain at least some of the systematic differences seen in Fig. 30 7. Note that, as lower AOD values derived from the PFR are expected based on its larger FOV, the linear 31 fits in Fig. 7 have been calculated for those data points with the Cimel-PFR AOD differences> 0. In this 32 way, we discard those pairs of AOD data whose difference is not only due to the different FOV between 33 both instruments, obtaining in this way a better approximation to quantify this effect.

The slopes of the fitting lines of the Cimel-PFR AOD differences vs. PFR AOD for AOD> 0.1 (dusty conditions) are 2.7% for 380 nm and 2.3% for 500 nm (Figure 7), which are quite consistent with the percentage differences of AOD between Cimel and PFR for an effective radius of 1.5 μ m (Figures 8a and

- 1 8b). These percentages correspond to absolute AOD differences of 0.016 at 380 nm, and 0.011 at 500 nm
- 2 for AOD=0.5 (Figures 8c and 8d), that are of sufficient magnitude to cause an appreciable number of 1-
- 3 minute AOD data outside the U95 limits, as indicated in Table 10.



Figure 8. Panels a) and b): the simulated relative differences in retrieved AOD (in %) that would result from the scattered radiation within the FOV of the PFR and Cimel instruments. The red (blue) dots show the differences between the AOD that would be retrieved using PFR (Cimel) and the actual AOD, and the grey dots the difference between PFR and Cimel, at wavelengths (a) 380 nm and (b) 500 nm. Panels c) and d): the difference in retrieved AOD between PFR and Cimel, plotted as a function of the AOD retrieved with PFR, for seven values of aerosol effective radius between 0.2 and 3.0 µm, at (c) 380 nm and (d) 500 nm.

10

11

4 5

6 7

8 9

13 If we apply the corresponding corrections to the 1-minute AOD PFR data > 0.1 assuming an effective 14 radius of 1.5μ m, + 3.3% at 380nm and + 2.2.% at 500 nm, it turns out that the slopes of the fitting lines of 15 the Cimel-PFR AOD differences vs. PFR AOD become practically zero (Figure 9). Moreover, the number of AOD data outside the U95 limits is reduced by approximately 53% for 380 nm and by 13% for 500 nm.
It must be taken into account that the percentage of AOD data for AOD> 0.1 outside the U95 limits, before
the corrections, is only 8% at 500 nm, while at 380 nm it is a significant value (24%).

This AOD "correction" reduces the Cimel-PFR AOD differences substantially but does not eliminate them completely. The main reason is the inherent limitation of data correction using the percentage difference in AOD obtained by model simulation for a fixed effective radius.

We have assumed an effective radius of 1.5 μ m but, in reality, the radius of dust particles varies. A reasonable range of dust particle radius is between 0.1 and 3 μ m (Balkanski et al., 1996; Denjean et al., 2016; Mahowald et al., 2014). So, depending on the distance from the dust source to IZO and the size of the emitted dust, the effective radius could vary slightly between dust episodes. As can be seen in Figures 8a and 8b, the percentage differences in AOD between Cimel and PFR for a 1-2 μ m effective radius interval, the PFR-Cimel AOD relative difference at 380 nm (500 nm) might change between ~-1.8% (-1.1%) to -4.9% (3.3%).

14 A similar analysis has been carried out for AERONET V3 (see Supplement S16), where we observe 15 that the corrections obtained are not as good as those obtained for V2. The effect of FOV on AOD retrieval 16 should be taken into account for those radiometers with a relatively high FOV ($>3^\circ$) measuring in regions with relatively high AOD (> 0.2) for most of the year, as is the case in many sites of Northern Africa, the 17 18 Middle East and East Asia (Basart et al., 2009; Cuevas et al., 2015; Eck et al., 1999; Kim et al., 2007). This 19 effect leads to AOD underestimation, and the variable number of high AOD episodes in each season of the 20 year might affect the AOD long-term trends. AOD measurements under these conditions would be 21 especially affected for optical air mass < 3.

22



Figure 9. The same as Figure 7 after "correcting" the PFR AOD data by adding + 3.3% at 380nm and+ 2.2.% at 500 nm to the 1-minute PFR AOD data > 0.1.

4

5 5.5. Angström exponent comparison

We have performed a comparison of the AE provided by GAW-PFR and AERONET-Cimel using in
both cases the AOD data obtained from the four common channels (380 nm, 440 nm, 500 nm and 870 nm)
with a total of 70716 data-pairs. The PFR-AOD values have been ordered from lowest to highest by

- 1 grouping them in intervals of 500 values for which the averages (and corresponding standard deviations)
- 2 of the PFR-Cimel AE differences have been calculated (Figure 10a). In a similar way we proceeded with
- 3 the PFR-AE values (Figure 10b).



8

Figure 10. (a) PFR-Cimel AE mean absolute differences (and corresponding standard deviations) vs
PFR mean AOD_{500nm} in 500 data intervals (b) and vs PFR mean AE in 500 data intervals. AE has been
computed for both PFR and Cimel using the four common channels (380, 440, 500 and 870nm).

AE differences > 0.2 increase exponentially for AOD < 0.02, reaching AE differences of up to 1.6 under pristine conditions (Figure 10a). For very low AOD the provided instruments uncertainty is the source of the sharp increase in AE, and at the same time AE becomes very sensitive to slight AOD changes. However, for AOD < 0.02 the atmospheric aerosol load is practically zero and so, its characterization with AE have in practice relatively minor importance.

In addition, the AE differences remain < 0.1 when AE_{PFR} values are < 1 (Figure 10b), which shows that these differences are small in most of the possible atmospheric scenarios. For 1 < AE_{PFR} < 1.2 the AE differences increase slightly to values < 0.2, and for AE_{PFR} > 1.2 (very fine particles or pristine conditions) the AE differences increase sharply to reach values of ~ 1.2. In our case, the non-pristine conditions, or those with a high content of mineral dust, have associated AOD > 0.03 and AE < 1, where the AE differences remain < 0.1. In case of pristine conditions AOD \leq 0.03 and AE \geq 1 the AE differences can reach a maximum of 1.6. Wagner and Silva (2008) estimated the usual maximum AE error by error propagation using a pair of spectral channels in which AOD is measured. Their results show that for clean optical conditions (AOD_{440nm}= 0.06) the maximum AE error is 1.17, and for hazy conditions (AOD_{440nm}=0.17) the error is 0.17, assuming an underlying AE of 1.5. These values decrease to 0.73 and 0.11, respectively, if AE=0. The AE differences found between GAW-PFR and AERONET-Cimel lie within the estimated errors reported by Wagner and Silva (2008).

8 Table 11. Uncertainty in AE determination for three typical atmospheric situations.

	Uncertainty
	in AE
Normal pristine conditions	≥1
AOD _{500nm} = 0.03 and AE = 1.4	
Hazy conditions	≥ 0.2
$AOD_{500nm} = 0.14 \text{ and } AE = 1.15$	
Strong dust intrusion	~ 0
$AOD_{500nm} = 0.3 \text{ and } AE = 0.3$	

9

In any case, as in our study the AE has been determined from AOD measured in the four common channels of GAW-PFR and AERONET-Cimel, we estimated the uncertainty in the calculation of the AE for three typical aerosol scenarios at Izaña. Following the methodology shown by Wagner and Silva (2008), the AE uncertainty estimations have been calculated using AOD measurements at four wavelengths and AOD uncertainty error propagation (Table 11). The AE derived from more than 2 wavelengths is less affected by AOD uncertainties than AE calculated with pairs of wavelengths, since the latter are calculated from the ratio of AOD at two channels (Cachorro et al., 2008).

17 The AE differences of our study (Figure 10) are within the AE uncertainty estimated for each type of 18 atmospheric condition (pristine, hazy and heavily dust loaded). However, although AE is a quantitative 19 parameter, it is only used in a qualitative way to estimate the range of sizes (fine, medium, coarse) of the 20 predominant aerosol in the inevitable mixture of aerosols that we observe. With this parameter, and together 21 with the information that is available in the measurement site about the most frequent types of aerosols and 22 their concentration, we can estimate the type of aerosols that are being measured. There are many 23 publications with different thresholds of AE and AOD in order to classify different types of aerosols (e.g. 24 Basart et al., 2009; Cuevas et al., 2015; Dubovik et al., 2002; Guirado et al., 2014; Holben et al., 2001; Kim 25 et al., 2007; Todd et al., 2007; Toledano et al., 2007; Wang et al., 2004). However, there is no consensus 26 on these thresholds since at each site there are different mixtures of aerosols, and each type of aerosol shows

- 1 specific frequencies of appearance and different concentrations. An alternative way of analyzing the degree
- of agreement in AE between GAW-PFR and AERONET-Cimel is to verify to what extent both networks
 provide the same information regarding the type of aerosol they observe in a certain site.
- 4 Considering the AE criteria established by Cuevas et al. (2015) and Berjón et al. (2019), we have 5 identified the following four main categories according to the AE_{PFR} and AE_{Cimel} values:
- 6 1. $AE_{PFR} \& AE_{Cimel} > 0.6$: Pristine conditions.

2. $0.25 < AE_{PFR} \& AE_{Cimel} \le 0.6$: Hazy, mineral dust being the main aerosol component.

- 8 3. AE_{PFR} & $AE_{Cimel} \le 0.25$: Pure dust.
- 9

7

4. AE_{PFR} and AE_{Cimel} do not fit any of the previous categories.

In 94.9 % of the cases, GAW-PFR and AERONET-Cimel V2 match the AE intervals of each aerosol scenario. Similar results (93.4%) were obtained when comparing with AERONET V3. Most of the agreement (>80 %) occurs in the predominant scenario of pristine conditions despite the AE uncertainty under pristine conditions being \geq 1. See Supplement S17 for more details. Note that the choice of these categories is not relevant since this is only used to examine the long-term agreement in AE between GAW-PFR and AERONET-Cimel in different atmospheric conditions. **6.** Summary and Conclusions

While GAW-PFR is the WMO-defined global AOD reference, being directly linked to WMO / CIMO, and was specifically designed to detect long-term AOD trends, AERONET-Cimel is the densest AOD measurement network globally, and the network most frequently used for aerosol characterization and for model and satellite observation evaluation.

20 An AERONET-Cimel 11-year AOD data series at IZO was obtained using a large number of 21 radiometers. A total of 13 Reference instruments were used in the period 2005-2009 which means that 22 every 4 and a half months, approximately, an instrument was replaced by another one to be calibrated. Their 23 calibrations were performed during their respective measurement time periods at IZO. Therefore, these 24 calibrations were not in any way linked with those of the instruments that preceded or replaced them, nor 25 with GAW-PFR reference. These facts led us to investigate the homogeneity of the AERONET-Cimel AOD 26 data series and their intercomparability with the much more homogeneous AOD data series from GAW-27 PFR (3 instruments in 11 years). The traceability concept for AOD suggested by WMO consists in 28 determining whether the AOD difference of the AERONET CIMELs vs the GAW PFRs lie within the U95 29 limits. We have used uncertainty limits for AOD traceability established by WMO (2005) for these type of 30 instruments with finite FOV. The acceptable traceability is when 95 % of the absolute AOD differences lie 31 within these limits, in which case both data populations are considered equivalent. It should be clarified 32 that "traceability" is not used in a strict metrological sense. This study has addressed the comparison of the 33 GAW-PFR dataset with the two versions of AERONET (V2 and V3) in the period 2005-2015. An excellent agreement between V2 and V3 for the four analysed channels ($R^2 > 0.999$) has been obtained. 34

More than 70000 synchronous GAW-PFR (PFR) and AERONET-Cimel (Cimel) 1-minute data-pairs in each channel in the period 2005-2015 were analysed. An excellent traceability of AOD from the AERONET-Cimel (V2 and V3) is found for 440 nm, 500 nm and 870 nm, and fairly good results for 1 380 nm. The lowest percentage of traceable AOD data is registered in 380 nm with 92.7 % of the 1-minute

2 data within the WMO limits, and the highest in 870 nm with 98.0 % of the data.

3 The different possible causes of non-traceability in AOD were investigated as follows:

• Absolute AOD measurements synchronization.

Analyzing 1-minute AOD variability we concluded that its impact on the AOD differences is quite small
as only ~0.8% of the AOD data has a variability larger than 0.005 in all spectral ranges.

• Sun tracking misalignments.

8 Sun tracking misalignments constitute a serious problem and a major cause of non-traceability of AOD 9 data-pairs as demonstrated by the AOD data outside the U95 limits from the period 2005-2009 as a 10 consequence of episodic problems with the sun-tracker of the GAW-PFR radiometer. For the 2010-2015 11 period the percentage of traceable data-pairs improves to 93.5% (380 nm), 97.4% (440 nm), 97.2% (500 12 nm) and 99.1% (870 nm). However, most of these cases could be identified and excluded from the analysis.

• Cloud screening failure by both network algorithms.

According to our observations, the simultaneous failure of both cloud screening algorithms might occur only under the presence of large and stable cirrus. In these cases, the radiometers interpret these clouds as aerosol layers and might provide values very different from the real AOD. For the comparison at IZO, however, this effect is negligible since GAW-PFR and AERONET-Cimel cloud screening algorithms provide successful cloud identification on clear direct-sun conditions during cloudy skies (FCS < 40 %) for 99.75 % of the cases.

• Pressure measurements related errors.

Since the accuracy of the new barometers built into new radiometers is about 3 hPa, and only errors in atmospheric pressure > 30 hPa might produce an impact on Rayleigh scattering, the AOD non-traceability due to errors in Rayleigh scattering is negligible.

24

34

4

7

• Total column ozone input uncertainty.

25 The largest influence of total ozone data uncertainty on ozone absorption occurs mainly at 500 nm. Total 26 ozone needs to be determined to ± 30 DU or 10 % of typical values, to ensure an uncertainty of ± 0.001 27 ozone absorption at 500 nm. In the case of the GAW-PFR / AERONET-Cimel comparison, despite the very 28 different methods in which both networks obtained O_3 values for their corresponding corrections, large 29 ozone differences were found (> 40 DU) only on 2.4 % of the days, resulting in a difference in the ozone 30 optical depth slightly above ~ 0.001 . The potential contribution to non-traceable AOD values between the 31 two networks is negligible. However, in mid or high latitude stations where fast O_3 variations of several 32 tens of DU might be registered, the correction of 1-minute AOD measurements by ozone absorption might 33 be an issue to be considered.

• Total column NO₂ input uncertainty.

The differences in NO₂ absorption caused by differences in daily total NO₂ between GAW-PFR and AERONET-Cimel is of the order of 10^{-3} for 380 nm and 440 nm channels, while, for 500 nm channel, it is even lower, of the order of 10^{-4} . So, differences in NO₂ absorption are negligible in the 1-minute AOD nontraceability of our study. However, NO₂ absorption might have some impact on AOD in highly polluted regions, such as in large industrial cities, where column NO₂ values are much larger than the climatological ones.

Taking into account the corrections for Rayleigh scattering and for the absorptions by O₃ and NO₂, we have calculated the combined effect of all of them on the non-traceability of the 1-minute AOD values. The highest impact occurs in the 380 nm channel, in which 25 % of the AOD data outside the *U*95 limits (~2% of the total compared data) are due to significant differences in pressure, and in O₃ and NO₂ absorption. The 1-minute AOD data outside the *U*95 limits by these corrections is negligible in the 870 nm channel.

9 • Impact of dust forward scattering in AOD retrieval uncertainty for different instrument FOVs

Since GAW-PFR has almost double the FOV ($\sim 2.5^{\circ}$) compared to the AERONET-Cimel ($\sim 1.3^{\circ}$), and 10 11 direct solar irradiance measurements are biased by the amount of aureole radiation that is assumed to be 12 direct solar radiation, it is reasonable to expect that GAW-PFR is more affected by the circumsolar 13 irradiance than AERONET-Cimel radiometer when AOD is relatively high. Modelling the dust forward 14 scattering we have shown that a non-negligible percentage of the non-traceable 1-minute AOD data for 15 AOD > 0.1, ranging from ~ 0.3 % at 870 nm to ~ 1.9 % at 380 nm is caused by the different FOV. Due to 16 this effect, the GAW-PFR provides AOD values which are ~3% lower at 380 nm, and ~2% lower at 500 nm, compared with AERONET-Cimel. However, AOD underestimation could only have some relevance 17 18 in dusty regions if radiometers with relatively large FOV are used.

19 A comparison of the AE provided by GAW-PFR and AERONET-Cimel has been performed using in both 20 cases AOD data obtained from the four nearby common channels with a total of 70716 data-pairs. This is 21 a very strict AE calculation since it is necessary that AOD be accurately measured by the four channels 22 simultaneously. AE differences > 0.2 increase exponentially under very pristine conditions (AOD < 0.03and $AE \ge 1$), reaching AE differences of up to 1.6. However, for these conditions the atmospheric aerosol 23 24 load is practically zero and so, its characterization with AE does not have any importance in practice. Under 25 non-pristine conditions or those with a high mineral dust content (associated AOD > 0.03 and AE < 1), the 26 AE differences remain < 0.1.

27 Summarizing, we have presented for the first time a long-term (2005-2015) 1-minute AOD comparison 28 among different types of radiometers belonging to different aerosol global networks. This comparison is a 29 very demanding test of both GAW-PFR and AERONET-Cimel validated AOD datasets since aerosol 30 scenarios correspond to extreme conditions: either very low aerosol loading, a "pristine" scenario that reveals small uncertainties in the calibration and in the cloud screening, or large dust load, which leads to 31 32 a significant increase in the forward scattering aerosol with AOD, resulting in a slightly higher AOD 33 underestimation by the GAW-PFR. From this comprehensive comparison, we can conclude that both AOD 34 datasets are representative of the same AOD population, which is a remarkable fact for the global aerosol 35 community. It should be noted that AOD traceability at 380 nm (92.7 %) does not reach 95 % of the 36 common data, the percentage recommended by WMO U95 criterion, so more efforts should be made to 37 improve AOD in the UV range. In this study we have also investigated the data that are outside of the WMO

U95 limits in order to understand their causes and to be eventually able to correct the small inconsistencies
 detected in instrumental and methodological aspects in the future.

3 Our results suggest that WMO/CIMO traceability limits could be redefined as a function of 4 wavelength, and the recommended radiometer FOV range should be reconsidered. The widely deployed 5 AERONET-Cimel and GAW-PFR datasets play a crucial role in understanding long-term AOD changes 6 and detecting trends, so it would be desirable for both networks to be linked to the same GAW-WMO 7 related reference.

8

9 *Competing interests.* The authors declare that they have no conflict of interest.

10

11 Acknowledgements. The authors thank Luc Blarel and Philippe Goloub (LOA, CNRS-University of Lille, 12 France) for supervising the periodic calibrations of the Cimel Reference instruments. This study has been 13 performed in the frame of the WMO CIMO Izaña Testbed for Aerosols and Water Vapour Remote Sensing 14 Instruments. The work was supported by the project "The Global Atmosphere Watch Precision Filter 15 Radiometer (GAW-PFR) Network for Aerosol Optical Depth long term measurements" funded by the Federal Office of Meteorology and Climatology MeteoSwiss International Affairs Division, Swiss GCOS 16 17 Office. Part of the AERONET-Cimel radiometers have been calibrated at Izaña Observatory by 18 AERONET- EUROPE Calibration Service, financed by the European Community specific programs for 19 Integrating Activities: Research Infrastructure Action under the Seventh Framework Programme 20 (FP7/2007-2013), ACTRIS grant agreement No. 262254, and Horizon 2020 Research and Innovation 21 Program, ACTRIS-2 grant agreement No. 654109. This research has received funding from the European 22 Union's Horizon 2020 Research and Innovation Programme under grant agreement No. 654109 (ACTRIS-23 2). The funding by MINECO (CTM2015-66742-R) and Junta de Castilla y León (VA100P17) is also 24 gratefully acknowledged. We thank the staff of the Izaña Observatory for their effort and dedication in 25 maintaining the instruments. We acknowledge the constructive comments of the anonymous referees. Our 26 colleague Celia Milford has improved the English language of the paper. In memory of Prof Klaus Fröhlich, 27 former director of PMOD-WRC, who initiated the AOD measurements programme at the Izaña 28 Observatory in 1984 within the WMO Background Atmospheric Pollution Monitoring Network 29 (BAPMoN).

30

31 **References**

Amiridis, V., Marinou, E., *Tsekeri*, A., Wandinger, U., Schwarz, A., Giannakaki, E., Mamouri, R.,
Kokkalis, P., Binietoglou, I., Solomos, S., Herekakis, T., Kazadzis, S., Gerasopoulos, E., Proestakis,
E., Kottas, M., Balis, D., Papayannis, A., Kontoes, C., Kourtidis, K., Papagiannopoulos, N., Mona, L.,
Pappalardo, G., Le Rille, O., and Ansmann, A.: LIVAS: a 3-D multi-wavelength aerosol/cloud
database based on CALIPSO and EARLINET, Atmos. Chem. Phys., 15, 7127-7153,
https://doi.org/10.5194/acp-15-7127-2015, 2015.

Angstrom, A.: On the atmospheric transmission of sun radiation and on dust in the air, Geografiska Annaler,
 11, 156–166, 1929. Balkanski, Y., Schulz, M., Marticorena, B., Bergametti, G.,Guelle, W., Dulac, F.,

- Moulin, C.,and Lambert, C.E.: Importance of the source term and of the size distribution to model the
 mineral dust cycle, in The Impact of Desert Dust Across the Mediterranean, edited by S. Guerzoni and
 R. Chester, pp. 69–76, Kluwer Academic Pub., Boston, MA. 2, 9, 2, 1996.
- Barker, H. W.: Solar radiative transfer through clouds possessing isotropic variable extinction coefficient,
 Quart. J. Roy. Meteor. Soc., 118, 1145-1162, doi: 10.1002/qj.49711850807, 1992.
- Barker, H. W.: Estimating cloud field albedo using one-dimensional series of optical depth, J. Atmos. Sci.,
 53, 2826-2837, doi: 10.1175/1520-0469(1996)053<2826:ECFAUO>2.0.CO;2}, 1996.
- Barreto, A., Cuevas, E., Pallé, P., Romero, P. M., Guirado, C., Wehrli, C. J., and Almansa, F.: Recovering
 long-term aerosol optical depth series (1976-2012) from an astronomical potassium-based resonance
 scattering spectrometer, Atmospheric Measurement Techniques, 7, 5 4103–4116,
 https://doi.org/10.5194/amt-7-4103-2014, 2014.
- 12 Barreto, A., Cuevas, E., Granados-Muñoz, M.-J., Alados-Arboledas, L., Romero, P. M., Gröbner, J., Kouremeti, N., Almansa, A. F., Stone, T., Toledano, C., Román, R., Sorokin, M., Holben, B., Canini, 13 14 M., and Yela, M.: The new sun-sky-lunar Cimel CE318-T multiband photometer a comprehensive Techniques, 631–654. 15 performance evaluation. Atmospheric Measurement 9. https://doi.org/10.5194/amt-9-63110 2016, https://www.atmos-meas-tech.net/9/631/2016/, 2016. 16
- Basart, S., Pérez, C., Cuevas, E., Baldasano, J. M., and Gobbi, G. P.: Aerosol characterization in Northern
 Africa, Northeastern Atlantic, Mediterranean Basin and Middle East from direct-sun AERONET
 observations, Atmospheric Chemistry and Physics, 9, 8265–8282, https://doi.org/10.5194/acp-9-82652009, https://www.atmos-chem-phys.net/9/8265/2009/, 2009.
- Basart, S., Pérez, C., Nickovic, S., Cuevas, E., and Baldasano, J.: Development and evaluation of the BSC DREAM8b dust regional model over Northern Africa, the Mediterranean and the Middle East, Tellus
 B: Chemical and Physical Meteorology, 64, 18539, https://doi.org/10.3402/tellusb.v64i0.18539, 2012.
- Benedetti, A., Reid, J. S., Knippertz, P., Marsham, J. H., Di Giuseppe, F., Rémy, S., Basart, S., Boucher,
 O., Brooks, I. M., Menut, L., Mona, L., Laj, P., Pappalardo, G., Wiedensohler, A., Baklanov, A.,
 Brooks, M., Colarco, P. R., Cuevas, E., da Silva, A., Escribano, J., Flemming, J., Huneeus, N., Jorba,
 O., Kazadzis, S., Kinne, S., Popp, T., Quinn, P. K., Sekiyama, T. T., Tanaka, T., and Terradellas, E.:
 Status and future of numerical atmospheric aerosol prediction with a focus on data requirements,
 Atmospheric Chemistry and Physics, 18, 10615–10643, https://doi.org/10.5194/acp-18-10615-2018,
 https://www.atmos-chem-phys.net/18/10615/2018/, 2018.
- Berjón, A., Barreto, A., Hernández, Y., Yela, M., Toledano, C., and Cuevas, E.: A 10-year characterization
 of the Saharan Air Layer lidar ratio in the subtropical North Atlantic, Atmos. Chem. Phys. Discuss.,
 https://doi.org/10.5194/acp-2018-1315, in review, 2019.
- Blanc, P., Espinar, B., Geuder, N., Gueymard, C., R., M., Pitz-Paal, R., Reinhardt, B., Renné, D., M., S.,
 Wald, L., and Wilbert, S.: Direct normal irradiance related definitions and applications: The
 circumsolar issue, Solar Energy, 110, 561 577,
 https://doi.org/10.1016/j.solener.2014.10.001, 2014.
- Bodhaine, B. A., Wood, N. B., Dutton, E. G., and Slusser, J. R.: On Rayleigh optical depth calculations,
 Journal of Atmospheric and Oceanic Technology, 16, 1854–1861, 1999.
- Böhm-Vitense, E.: Introduction to stellar astrophysics, volume 2: stellar atmospheres, Cambridge
 University Press, 260 pp., 1989.
- Bokoye, A. I., Royer, A., O'Neill, N. T., Cliche, P., Fedosejevs, G., Teillet, P. M., and McArthur, L. J. B.:
 Characterization of atmospheric aerosols across Canada from a ground-based sunphotometer network:
 AEROCAN, Atmosphere-Ocean, 39, 429–456, https://doi.org/10.1080/07055900.2001.9649687,
 2001.
- Cachorro, V., Toledano, C., Sorribas, M., Berjón, A., De Frutos, A., and Laulainen, N.: An "in situ"
 calibration-correction procedure (KCICLO) based on AOD diurnal cycle: Comparative results between

- AERONET and reprocessed (KCICLO method) AOD-alpha data series at El Arenosillo, Spain, Journal
 of Geophysical Research: Atmospheres, 113, https://doi.org/https://doi.org/10.1029/2007JD009001,
 2008.
- Cachorro, V. E., Romero, P. M., Toledano, C., Cuevas, E., and de Frutos, A. M.: The fictitious diurnal
 cycle of aerosol optical depth: A new approach for "in situ" calibration and correction of AOD data
 series, Geophysical Research Letters, 31, https://doi.org/10.1029/2004GL019651, 2004.
- Campanelli, M., Nakajima, T. and Olivieri, B.: Determination of the Solar Calibration Constant for a SunSky Radiometer: Proposal of an In-Situ Procedure, Applied optics, 43, 3, 651-659,
 doi:10.1364/AO.43.000651, 2004.
- 10 Campanelli, M., Nakajima, T., and Olivieri, B., 2004: Determination of the solar calibration constant for a sun - sky radiometer: proposal of an in situ procedure. Appl. Opt., 43(1), 651-659. Carlund, T., 11 12 Kouremeti, N., Kazadzis, S., and Gröbner, J.: Aerosol optical depth determination in the UV using a Atmos. radiometer. 13 four-channel precision filter Meas. Tech., 10. 905-923. 14 https://doi.org/10.5194/amt-10-905-2017, 2017.
- Carrillo, J., Guerra, J. C., and Cuevas, E.and Barrancos, J.: Characterization of the Marine Boundary Layer
 and the Trade-Wind Inversion over the Sub-tropical North Atlantic, Boundary-Layer Meteorology,
 158, 311–330, https://doi.org/10.1007/s10546-015-0081-1, https://doi.org/10.1007/s10546-015-0081-1
 1, 2016.
- Chance, K. and Kurucz, R.: An improved high-resolution solar reference spectrum for earth's atmosphere
 measurements in the ultraviolet, visible, and near infrared, Journal of quantitative spectroscopy and
 radiative transfer, 111, 1289–1295, https://doi.org/https://doi.org/10.1016/j.jqsrt.2010.01.036,
 http://www.sciencedirect.com/science/article/pii/S0022407310000610, 2010.
- Che, H., Zhang, X.-Y., Xia, X., Goloub, P., Holben, B., Zhao, H., Wang, Y., Zhang, X.-C., Wang, H.,
 Blarel, L., Damiri, B., Zhang, R., Deng, X., Ma, Y., Wang, T., Geng, F., Qi, B., Zhu, J., Yu, J., Chen,
 Q., and Shi, G.: Ground-based aerosol climatology of China: aerosol optical depths from the China
 Aerosol Remote Sensing Network (CARSNET) 2002–2013, Atmos. Chem. Phys., 15, 7619–7652,
 https://doi.org/10.5194/acp15-7619-2015, 2015.
- Chedin, A., Capelle, V., and Scott, N.: Detection of IASI dust AOD trends over Sahara: How many years
 of data required?, Atmospheric Research, 212, 120–129,
 https://doi.org/https://doi.org/10.1016/j.atmosres.2018.05.004,
 http://www.sciencedirect.com/science/article/pii/ S0160800517310566, 2018
- 31 <u>http://www.sciencedirect.com/science/article/pii/</u>S0169809517310566, 2018.
- Chubarova, N. Y., Poliukhov, A. A., and Gorlova, I. D.: Long-term variability of aerosol optical thickness
 in Eastern Europe over 2001– 2014 according to the measurements at the Moscow MSU MO
 AERONET site with additional cloud and NO₂ correction, Atmospheric Measurement Techniques, 9,
 313–334, https://doi.org/10.5194/amt-9-313-2016, https://www.atmos-meas-tech.net/9/313/2016/,
 2016.
- 37 Cuevas, E., González, Y., Rodríguez, S., Guerra, J. C., Gómez-Peláez, A. J., Alonso-Pérez, S., Bustos, J., 38 and Milford, C.: Assessment of atmospheric processes driving ozone variations in the subtropical North 39 free troposphere, Atmospheric Chemistry and Physics, 13. Atlantic 1973-1998. 40 https://doi.org/10.5194/acp-13-1973-2013, https://www.atmos-chem-phys.net/13/1973/2013/, 2013.
- Cuevas, E., Camino, C., Benedetti, A., Basart, S., Terradellas, E., Baldasano, J. M., Morcrette, J. J.,
 Marticorena, B., Goloub, P., Mortier, A., Berjón, A., Hernández, Y., Gil-Ojeda, M., and Schulz, M.:
 The MACC-II 2007-2008 reanalysis: atmospheric dust evaluation and characterization over northern
 Africa and the Middle East, Atmospheric Chemistry and Physics, 15, 3991–4024,
 https://doi.org/10.5194/acp-153991-2015, https://www.atmos-chem-phys.net/15/3991/2015/, 2015.
- Cuevas, E., Gómez-Peláez, A., Rodríguez, S., Terradellas, E., Basart, S., García, R., García, O., and Alonso Pérez, S.: The pulsating nature of large-scale Saharan dust transport as a result of interplays between

- mid-latitude Rossby waves and the North African Dipole Intensity, Atmospheric Environment, 167,
 586–602, https://doi.org/https://doi.org/10.1016/j.atmosenv.2017.08.059,
- 3 <u>http://www.sciencedirect.com/</u>20 science/article/pii/S1352231017305757, 2017a.
- Cuevas, E., Milford, C., Bustos, J. J., del Campo-Hernández, García, O., D., G. R., Gómez-Peláez, GuiradoFuentes, C., Marrero, C., Prats, N., Ramos, R., Redondas, A., Reyes, E., Rodríguez, S., RomeroCampos, P., Scheneider, M., Belmonte, J., Yela, M., Almansa, F., Barreto, A., López-Solano, C.,
 Basart, S., Terradellas, E., Afonso, S., Bayo, C., Berjón, A., Bethencourt, J., Carreño, V., Castro, N. J.,
 Cruz, A. M., Damas, M., De Ory-Ajamil, F., García, M. I., Gómez-Trueba, V., González, Y.,
 Hernández, C., Hernández, Y., Hernández-Cruz, B., Jover, M., León, S., López-Fernández, R., López-
- 10 Solano, J., Rodríguez, E., Rodríguez-Franco, J., Rodríguez-Valido, M., Sálamo, C., Sanromá, E.,
- 11 Santana, D., Santo-Tomás, F., Sepúlveda, E., Sierra, M., and Sosa, E.: Izaña Atmospheric Research
- 12 Center Activity Report 2015-2016, State Meteorological Agency (AEMET), 2017b.
- Denjean, C., Cassola, F., Mazzino, A., Triquet, S., Chevaillier, S., Grand, N., Bourrianne, T., Momboisse,
 G., Sellegri, K., Schwarzenbock, A., Freney, E., Mallet, M., and Formenti, P.: Size distribution and
 optical properties of mineral dust aerosols transported in the western Mediterranean, Atmos. Chem.
 Phys., 16, 1081-1104, https://doi.org/10.5194/acp-16-1081-2016, 2016.
- Driemel, A., Augustine, J., Behrens, K., Colle, S., Cox, C., Cuevas-Agulló, E., Denn, F. M., Duprat, T., 17 18 Fukuda, M., Grobe, H., Haeffelin, M., Hodges, G., Hyett, N., Ijima, O., Kallis, A., Knap, W., Kustov, 19 V., Long, C. N., Longenecker, D., Lupi, A., Maturilli, M., Mimouni, M., Ntsangwane, L., Ogihara, H., 20 Olano, X., Olefs, M., Omori, M., Passamani, L., Pereira, E. B., Schmithüsen, H., Schumacher, S., Sieger, R., Tamlyn, J., Vogt, R., Vuilleumier, L., Xia, X., Ohmura, A., and König-Langlo, G.: Baseline 21 22 Surface Radiation Network (BSRN): structure and data description (1992–2017), Earth System Science 23 Data, 10, 1491–1501, https://doi.org/10.5194/essd-10-1491-2018, https://www.earth-syst-sci-24 data.net/10/1491/2018/, 2018.
- Dubovik, O., Holben, B., Eck, T. F., Smirnov, A., Kaufman, Y. J., King, M. D., Tanré, D., and Slutsker, I.:
 Variability of absorption and optical properties of key aerosol types observed in worldwide locations,
 Journal of the atmospheric sciences, 59, 590–608, https://doi.org/https://doi.org/10.1175/15200469(2002)059<0590:VOAAOP>2.0.CO;2, 2002.
- Eck, T., Holben, B., Reid, J., Dubovik, O., Smirnov, A., O'neill, N., Slutsker, I., and Kinne, S.: Wavelength
 dependence of the optical depth of biomass burning, urban, and desert dust aerosols, Journal of
 Geophysical Research: Atmospheres, 104, 31333–31349, https://doi.org/10.1029/1999JD900923,
 1999.
- Eissa, Y., Blanc, P., Wald, L., and Ghedira, H.: Can AERONET data be used to accurately model the
 monochromatic beam and circumsolar irradiances under cloud-free conditions in desert environment?,
 Atmospheric Measurement Techniques, 8, 5099–5112, https://doi.org/DOI = 10.5194/amt-8-5099 2015, 2015.
- Eissa, Y., Blanc, P., Ghedira, H., Oumbe, A., and Wald, L.: A fast and simple model to estimate the
 contribution of the circumsolar irradiance to measured broadband beam irradiance under cloud-free
 conditions in desert environment, Solar Energy, 163, 497–509,
 https://doi.org/https://doi.org/10.1016/j.solener.2018.02.015, 2018.
- Eskes, H. J. and Boersma, K. F.: Averaging kernels for DOAS total-column satellite retrievals, Atmospheric
 Chemistry and Physics, 3, 1285–1291, https://doi.org/10.5194/acp-3-1285-2003, https://www.atmoschem-phys.net/3/1285/2003/, 2003.
- Freidenreich, S. M., and Ramaswamy, V.: A new multiple-band solar radiative parameterization for general
 circulation models, J. Geophys. Res., 104, 31389-31409, doi: 10.1029/1999JD900456, 1999.
- 46 García, R. D., Cuevas, E., García, O. E., Cachorro, V. E., Pallé, P., Bustos, J. J., Romero-Campos, P. M., 47 and de Erutos, A. M.: Paconstruction of global solar radiation time series from 1022 to 2012 at the
- 47 and de Frutos, A. M.: Reconstruction of global solar radiation time series from 1933 to 2013 at the

- Izaña Atmospheric Observatory, Atmospheric Measurement Techniques, 7, 3139–3150, https://doi.org/10.5194/amt-7-3139-2014, https://www.atmos-meas-tech.net/7/3139/2014/, 2014.
- García, R. D., Barreto, A., Cuevas, E., Gröbner, J., García, O. E., Gómez-Peláez, A., Romero-Campos, P.
 M., Redondas, A., Cachorro, V. E., and Ramos, R.: Comparison of observed and modeled cloud-free
 longwave downward radiation (2010–2016) at the high mountain BSRN Izaña station, Geoscientific
 Model Development, 11, 2139–2152, https://doi.org/10.5194/gmd-11-2139-2018, https://www.geoscimodel-dev.net/11/2139/2018/, 2018.
- García, R. D., Cuevas, E., Ramos, R., Cachorro, V. E., Redondas, A., and Moreno-Ruiz, J. A.: Description
 of the Baseline Surface Radiation Network (BSRN) station at the Izaña Observatory (2009–2017):
 measurements and quality control/assurance procedures, Geosci. Instrum. Method. Data Syst., 8, 7796, https://doi.org/10.5194/gi-8-77-2019, 2019.
- Giles, D. M., Sinyuk, A., Sorokin, M. G., Schafer, J. S., Smirnov, A., Slutsker, I., Eck, T. F., Holben, B.
 N., Lewis, J. R., Campbell, J. R., Welton, E. J., Korkin, S. V., and Lyapustin, A. I.: Advancements in
 the Aerosol Robotic Network (AERONET) Version 3 database automated near-real-time quality
 control algorithm with improved cloud screening for Sun photometer aerosol optical depth (AOD)
 measurements, Atmos. Meas. Tech., 12, 169-209, https://doi.org/10.5194/amt-12-169-2019, 2019.
- Goloub, P., Li, Z., Dubovik, O., Blarel, L., Podvin, T., Jankowiak, I., Lecoq, R., Deroo, C., Chatenet, B.,
 Morel, J., Cuevas, E., and Ramos, R.: PHOTONS/AERONET sunphotometer network overview:
 description, activities, results, in: Fourteenth International Symposium on Atmospheric and Ocean
 Optics/Atmospheric Physics, vol. 6936, p. 69360V, International Society for Optics and Photonics,
 2007.
- Grassl, H.: Calculated Circumsolar Radiation as a Function of Aerosol Type, Field of View, Wavelength,
 and Optical Depth, Applied Optics, Vol. 10, No. 11, 2543, 1971.
- Gueymard, C.: SMARTS2: a simple model of the atmospheric radiative transfer of sunshine: algorithms
 and performance assessment, Florida Solar Energy Center Cocoa, FL, 1995.
- Guirado, C., Cuevas, E., Cachorro, V. E., Toledano, C., Alonso-Pérez, S., Bustos, J. J., Basart, S., Romero,
 P. M., Camino, C., Mimouni, M., Zeudmi, L., Goloub, P., Baldasano, J. M., and de Frutos, A. M.:
 Aerosol characterization at the Saharan AERONET site Tamanrasset, Atmospheric Chemistry and
 Physics, 14, 11753–11773, https://doi.org/10.5194/acp-14-11753-2014, https://www.atmos-chemphys.net/ 14/11753/2014/, 2014.
- Holben, B., Eck, T., Slutsker, I., Tanré, D., Buis, J., Setzer, A., Vermote, E., Reagan, J., and Kaufman, Y.:
 Multi-band automatic sun and sky scanning radiometer system for measurement of aerosols, pp. 75–
 83, CNES, Proceedings of 6th International Symposium on Physical Measurements and Signatures in
 Remote Sensing, 1994.
- Holben, B., Eck, T., Slutsker, I., Tanré, D., Buis, J., Setzer, A., Vermote, E., Reagan, J., Kaufman, Y.,
 Nakajima, T., Lavenu, F., Jankowiak, I., and Smirnov, A.: AERONET—A Federated Instrument
 Network and Data Archive for Aerosol Characterization, Remote Sensing of Environment, 66, 1 16,
 https://doi.org/https://doi.org/10.1016/S0034-4257(98)00031-5, 1998. Holben, B., Tanre, D., Smirnov,
 A., Eck, T., Slutsker, I., Abuhassan, N., Newcomb, W., Schafer, J., Chatenet, B., Lavenu, F., et al.: An
- 40 emerging ground-based aerosol climatology: Aerosol optical depth from AERONET, Journal of
- 41
 Geophysical
 Research:
 Atmospheres,
 106,
 12067–12097,

 42
 https://doi.org/https://doi.org/10.1029/2001JD900014, 2001.
 106,
 12067–12097,
- Huijnen, V. and Eskes, H.: Skill scores and evaluation methodology for the MACC II project, MACC-II
 Deliverable D_85, 2, http://www. gmesatmosphere.eu/documents/maccii/deliverables/val/MACCII_VAL_DEL_D_85.2_ScoringReport01_2
 0120222.pdf, 2012.

- Huneeus, N., Basart, S., Fiedler, S., Morcrette, J.-J., Benedetti, A., Mulcahy, J., Terradellas, E., Pérez
 García-Pando, C., Pejanovic, G., Nickovic, S., Arsenovic, P., Schulz, M., Cuevas, E., Baldasano, J.
 M., Pey, J., Remy, S., and Cvetkovic, B.: Forecasting the northern African dust outbreak towards
 Europe in April 2011: a model intercomparison, Atmospheric Chemistry and Physics, 16, 4967–4986,
- 5 15 https://doi.org/10.5194/acp-16-4967-2016, https://www.atmos-chem-phys.net/16/4967/2016/,
 6 2016.
- 7 IPCC: The Physical Science Basis. Intergovernmental Panel on Climate Change, 8 https://doi.org/doi:10.1017/CBO9781107415324, 2013. Jarosyawski, J., Krzyscin, J.W., Puchalski, S., 9 Sobolewski, P.: On the optical thickness in the UV range: analysis of the ground-based data taken at 10 Belsk, Poland. Journal of Geophysical Research 108 (D23), doi: 10.1029/2003JD003571, 2003.
- Kahn, R. A. and Gaitley, B. J.: An analysis of global aerosol type as retrieved by MISR, Journal of
 Geophysical Research: Atmospheres, 120, 4248–4281, https://doi.org/10.1002/2015JD023322,
 https://agupubs.onlinelibrary.wiley.com/doi/abs/10.1002/2015JD023322, 2015.
- Kalnay, E., Kanamitsu, M., Kistler, R., Collins, W., Deaven, D., Gandin, L., Iredell, M., Saha, S., White,
 G., Woollen, J., et al.: The 20 NCEP/NCAR 40-year reanalysis project, Bulletin of the American
 meteorological Society, 77, 437–472, 1996.
- Kasten, F.: A new table and approximation formula for the relative optical air mass, Archiv für
 Meteorologie, Geophysik und Bioklimatologie, Serie B, 14, 206–223,
 https://doi.org/https://doi.org/10.1007/BF02248840, 1966.
- 20Kasten, F.andYoung, A.T.:Revisedoptical airmasstablesand21approximation formula,Appl.Opt.,28,4735–4738,
- 22 https://doi.org/10.1364/AO.28.004735, 1989.
- Kazadzis, S., Veselovskii, I., Amiridis, V., Gröbner, J., Suvorina, A., Nyeki, S., Gerasopoulos, E.,
 Kouremeti, N., Taylor, M., Tsekeri, A., and Wehrli, C.: Aerosol microphysical retrievals from
 precision filter radiometer direct solar radiation measurements and comparison with AERONET,
 Atmospheric Measurement Techniques, 7, 2013–2025, https://doi.org/10.5194/amt-7-2013-2014,
 https://www.atmos-meas-tech.net/7/2013/2014/, 2014.
- Kazadzis, S., Kouremeti, N., Diémoz, H., Gröbner, J., Forgan, B. W., Campanelli, M., Estellés, V., Lantz,
 K., Michalsky, J., Carlund, T., Cuevas, E., Toledano, C., Becker, R., Nyeki, S., Kosmopoulos, P. G.,
 Tatsiankou, V., Vuilleumier, L., Denn, F. M., Ohkawara, N., Ijima, O., Goloub, P., Raptis, P. I., Milner,
 M., Behrens, K., Barreto, A., Martucci, G., Hall, E., Wendell, J., Fabbri, B. E., and Wehrli, C.: Results
 from the Fourth WMO Filter Radiometer Comparison for aerosol optical depth measurements,
 Atmospheric Chemistry and Physics, 18, 3185–3201, https://doi.org/10.5194/acp-18-3185-2018,
 2018a.
- Kazadzis, S., Kouremeti, N., Nyeki, S., Gröbner, J., and Wehrli, C.: The World Optical Depth Research
 and Calibration Center (WORCC) quality assurance and quality control of GAW-PFR
 AOD measurements, Geoscientific Instrumentation, Methods and Data Systems, 7, 39–53,
 https://doi.org/10.5194/gi-739-2018, https://www.geosci-instrum-method-data-syst.net/7/39/2018/,
 2018b.
- Kentarchos, A., Roelofs, G.J., Lelieveld, J., and Cuevas, E.: On the origin of elevated surface ozone
 concentrations at Izaña Observatory during the last days of March 1996: a model study, Geophys. Res.
 Lett., Vol. 27, 22, 3,699-3,702, 2000.
- Kim, S.-W., Jefferson, A., Soon-Chang, Y., Dutton, E., Ogren, J., Valero, F., Kim, J., and Holben, B.:
 Comparisons of aerosol optical depth and surface shortwave irradiance and their effect on the aerosol
 surface radiative forcing estimation, Journal of Geophysical Research: Atmospheres, 110,
 https://doi.org/10.1029/2004JD004989, 2005.

- Kim, S.-W., Yoon, S.-C., Kim, J., and Kim, S.-Y.: Seasonal and monthly variations of columnar aerosol optical properties over east Asia determined from multi-year MODIS, LIDAR, and AERONET Sun/sky radiometer measurements, Atmospheric Environment, 41, 1634–1651, https://doi.org/https://doi.org/10.1016/j.atmosenv.2006.10.044, 2007.
- Kim, S.-W., Yoon, S.-C., Dutton, E., Kim, J., and Wehrli, C.and Holben, B.: Global surface-based sun 5 photometer network for long-term observations of column aerosol optical properties: intercomparison 6 7 of aerosol optical depth, Aerosol Science and Technology, 42, 1-9. https://doi.org/https://doi.org/10.1080/02786820701699743, 2008. 8
- Klingmüller, K., Pozzer, A., Metzger, S., Stenchikov, G. L., and Lelieveld, J.: Aerosol optical depth trend
 over the Middle East, Atmospheric Chemistry and Physics, 16, 5063–5073,
 https://doi.org/10.5194/acp-16-5063-2016, 2016.
- Komhyr, W.: Dobson spectrophotometer systematic total ozone measurement error, Geophysical Research
 Letters, 7, 161–163, 1980.
- Komhyr, W. D., Grass, R. D., and Leonard, R. K.: Dobson spectrophotometer 83: A standard for total ozone
 measurements, 1962–1987, Journal of Geophysical Research: Atmospheres, 94, 9847–9861,
 https://doi.org/10.1029/JD094iD07p09847, 1989. Mahowald, N., M., Albani, S. Kok, J. F.,
 Engelstaeder, S., Scanza, R., Ward, D. S., Flanner, M. G.: The size distribution of desert dust aerosols
 and its impact on the Earth system, Aeolian Research, 15, 53-71, 2014.
- McArthur, L. J. B., Halliwell, D. H., Niebergall, O. J., O'Neill, N. T., Slusser, J. R., and Wehrli, C.: Field
 comparison of network Sun photometers, Journal of Geophysical Research: Atmospheres, 108,
 https://doi.org/10.1029/2002JD002964, 2003.
- McPeters, R., Frith, S., and Labow, G.: OMI total column ozone: extending the long-term data record,
 Atmospheric Measurement Techniques, 8, 4845–4850, https://doi.org/https://doi.org/10.5194/amt-8 4845-2015, 2015.
- Mitchell, R. and Forgan, B.: Aerosol measurement in the Australian outback: Intercomparison of sun
 photometers, Journal of Atmospheric 20 and Oceanic Technology, 20, 54–66,
 https://doi.org/https://doi.org/10.1175/1520-0426(2003)020<0054:AMITAO>2.0.CO;2, 2003.
- Mitchell, R. M., Forgan, B. W., and Campbell, S. K.: The Climatology of Australian Aerosol, Atmos.
 Chem. Phys., 17, 5131-5154, <u>https://doi.org/10.5194/acp-17-5131-2017</u>, 2017.
- Nakajima, T., Yoon, S. C., Ramanathan, V., Shi, G. Y., Takemura, T., Higurashi, A., Takamura, T., Aoki,
 K., Sohn, B. J., Kim, S. W., Tsuruta, H., Sugimoto, N., Shimizu, A., Tanimoto, H., Sawa, Y., Lin, N.
 H., Lee, C. T., Goto, D., and Schutgens, N.: Overview of the Atmospheric Brown Cloud East Asian
 Regional Experiment 2005 and a study of the aerosol direct radiative forcing in east Asia, J. Geophys.
 Res., 112, D24S91, doi:10.1029/2007JD009009, 2007.
- Nyeki, S., Halios, C., Baum, W., Eleftheriadis, K., Flentje, H., Gröbner, J., Vuilleumier, L., and Wehrli, C.:
 Ground-based aerosol optical depth trends at three high-altitude sites in Switzerland and southern
 Germany from 1995 to 2010, Journal of Geophysical Research: Atmospheres, 117,
 https://doi.org/https://doi.org/10.1029/2012JD017493, 2012.
- Nyeki, S., Gröbner, J., and Wehrli, C.: Ground-based aerosol optical depth inter-comparison campaigns at
 European EUSAAR super-sites, 25 vol. 1531, pp. 584–587, https://doi.org/10.1063/1.4804837, 2013.

Nyeki, S., Wehrli, C., Gröbner, J., Kouremeti, N., Wacker, S., Labuschagne, C., Mbatha, N., and Brunke,
E.-G.: The GAW-PFR aerosol optical depth network: The 2008–2013 time series at Cape Point Station,
South Africa, Journal of Geophysical Research: Atmospheres, 120, 5070–5084, 2015. Räisänen, P.,
Isaac, G. A., Barker, H. W. and Gultepe, I.: Solar radiative transfer for stratiform clouds with horizontal
variations in liquid-water path and droplet effective radius, Quart. J. Roy. Meteor. Soc., 129, 21352149, doi: 10.1256/qj.02.149, 2003.

- Rodríguez, S., González, Y., Cuevas, E., Ramos, R., Romero, P. M., Abreu-Afonso, J., and Redondas, A.:
 Atmospheric nanoparticle observations in the low free troposphere during upward orographic flows at
 IzaÃña Mountain Observatory, Atmospheric Chemistry and Physics, 9, 6319–6335,
 https://doi.org/10.5194/acp-9-6319-2009, 2009.
- Rodríguez, S., Alastuey, A., Alonso-Pérez, S., Querol, X., Cuevas, E., Abreu-Afonso, J., Viana, M., Pérez,
 N., Pandolfi, M., and de la Rosa, J.: Transport of desert dust mixed with North African industrial
 pollutants in the subtropical Saharan Air Layer, Atmospheric Chemistry and Physics, 11, 6663–6685,
 https://doi.org/10.5194/acp-11-6663-2011, 2011.
- Rodríguez, S., Cuevas, E., Prospero, J. M., Alastuey, A., Querol, X., López-Solano, J., García, M. I., and
 Alonso-Pérez, S.: Modulation of Saharan dust export by the North African dipole, Atmospheric
 Chemistry and Physics, 15, 7471–7486, https://doi.org/10.5194/acp-157471-2015, 2015.
- Rodriguez-Franco, J. J., and Cuevas, E.: Characteristics of the subtropical tropopause region based on long term highly-resolved sonde records over Tenerife, J. Geophys. Res. Atmos., 118,
 doi:10.1002/jgrd.50839, 2013.
- Romero, P. M. and Cuevas, E.: Variación diurna del espesor óptico de aerosoles:¿ ficción o realidad?,
 3 Asamblea Hispano Portuguesa de Geofísica y Geodesia. Valencia, 2002.
- Romero-Campos, P., Cuevas, A., Kazadzis, S., Kouremeti, N., García, R., and Guirado-Fuentes, C.:
 Análisis de la trazabilidad en los valores del AOD obtenidos a partir de las medidas de las redes
 AERONET-CIMEL y GAW-PFR durante el período 2005-2015 en el Observatorio Atmosférico de
 Izaña, 2017.
- Russell, P. B., Livingston, J. M., Dubovik, O., Ramirez, S. A., Wang, J., Redemann, J., Schmid, B., Box,
 M., and Holben, B. N.: Sunlight transmission through desert dust and marine aerosols: Diffuse light
 corrections to Sun photometry and pyrheliometry, J. Geophys. Res., 109, D08207,
 10.1029/2003JD004292, 2004.
- Sakerin, S. M., Kabanov, D. M., Panchenko, M. V., Pol'kin, V. V., Holben, B. N., Smirnov, A. V.,
 Beresnev, S. A., Gorda, S. Y., Kornienko, G. I., Nikolashkin, S. V., Poddubnyi, V. A., and Tashchilin,
 M. A.: Monitoring of atmospheric aerosol in the Asian part of Russia in 2004 within the framework of
 AEROSIBNET program, Atmos. Oceanic Optics, 18, 871–878, 2005.
- Sayer, A. M., Hsu, N. C., Bettenhausen, C., Jeong, M., Holben, B. N., and Zhang, J.: Global and regional
 evaluation of over-land spectral aerosol optical depth retrievals from SeaWiFS, Atmospheric
 Measurement Techniques, 5, 1761–1778, https://doi.org/10.5194/amt-51761-2012, 2012.
- Sayer, A. M., Hsu, N. C., Bettenhausen, C., and Jeong, M.: Validation and uncertainty estimates for MODIS
 Collection 6 "Deep Blue" aerosol 10 data, Journal of Geophysical Research: Atmospheres, 118, 7864–
 7872, https://doi.org/10.1002/jgrd.50600,
 https://agupubs.onlinelibrary.wiley.com/doi/abs/10.1002/jgrd.50600, 2013.
- Schmid, B. and Wehrli, C.: Comparison of Sun photometer calibration by use of the Langley technique and
- 37
 the standard lamp, Appl. Opt., 34, 4500–4512, https://doi.org/10.1364/AO.34.004500,

 38
 http://ao.osa.org/abstract.cfm?URI=ao-34-21-4500, 1995.
- Schmid, B., Michalsky, J., Halthore, R., Beauharnois, M., Harrison, L., Livingston, J., Russell, P., Holben,
 B., Eck, T., and Smirnov, A.: Comparison of aerosol optical depth from four solar radiometers during
 the fall 1997 ARM intensive observation period, Geophysical Research Letters, 26, 2725–2728,
 https://doi.org/10.1029/1999GL900513, https://agupubs.onlinelibrary.wiley.com/doi/abs/10.1029/
 1999GL900513, 1999.
- Sinyuk, A., Holben, B.N., Smirnov, A., Eck, T.F., Slutsker, I., Schafer, J.S., Giles, D.M., and Sorokin, M.:
 Assessment of error in aerosol optical depth measured by AERONET due to aerosol forward scattering,
 Geophys. Res. Lett., 39, 23806, doi:10.1029/2012GL053894, 2012.

- Smirnov, A., Holben, B., Eck, T., Dubovik, O., and Slutsker, I.: Cloud-screening and quality control algorithms for the AERONET database, Remote sensing of environment, 73, 337–349, https://doi.org/https://doi.org/10.1016/S0034-4257(00)00109-7, http://www.sciencedirect. 20 com/science/article/pii/S0034425700001097, 2000.
- Takamura, T., and Nakajima, T.: Overview of SKYNET and its activities, Opt. Pura Apl., 37, 3303-3308,
 2004.
- Thomason, L.W., Herman, B. M., Schotland, R.M., and Reagan, J.A.: Extraterrestrial solar flux
 measurement limitations due to a Beer's law assumption and uncertainty in local time, Appl. Opt.,
 21, 1191–1195, https://doi.org/10.1364/AO.21.001191, 1982.
- Todd, M. C., Washington, R., Martins, J. V., Dubovik, O., Lizcano, G., M'bainayel, S., and Engelstaedter,
 S.: Mineral dust emission from the Bodélé Depression, northern Chad, during BoDEx 2005, Journal of
 Geophysical Research: Atmospheres, 112, https://doi.org/https://doi.org/10.1029/2006JD007170,
 2007.
- Toledano, C., Cachorro, V. E., Berjón, A., de Frutos, A. M., Sorribas, M., de la Morena, B. A., and Goloub,
 P.: Aerosol optical depth and Ångström exponent climatology at El Arenosillo AERONET site
 (Huelva, Spain), Quarterly Journal of the Royal Meteorological Society, 133, 795–807,
 https://doi.org/10.1002/qj.54, https://rmets.onlinelibrary.wiley.com/doi/abs/10.1002/qj.54, 2007.
- Toledano, C., Cachorro, V. E., Berjon, A., de Frutos, A. M., Fuertes, D., Gonzalez, R., Torres, B., Rodrigo,
 R., Bennouna, Y., Martin, L., and Guirado, C.: RIMA-AERONET network: long-term monitoring of
 aerosol properties, Opt. Pura Apl., 44, 629–633, 2011.
- Toledano, C., Cachorro, V., Gausa, M., Stebel, K., Aaltonen, V., Berjón, A., de Galisteo, J. P. O., de Frutos,
 A. M., Bennouna, Y., Blindheim, S., Myhre, C. L., Zibordi, G., Wehrli, C., Kratzer, S., Hakansson,
 B., Carlund, T., de Leeuw, G., Herber, A., and Torres, B.: Overview of Sun Photometer Measurements
 of Aerosol Properties in Scandinavia and Svalbard, Atmospheric Environment, 52, 18–28, 30
 https://doi.org/10.1016/j.atmosenv.2011.10.022, 2012.
- Toledano, C., González, R., Fuertes, D., Cuevas, E., Eck, T. F., Kazadzis, S., Kouremeti, N., Gröbner, J.,
 Goloub, P., Blarel, L., Román, R., Barreto, A., Berjón, A., Holben, B. N., and Cachorro, V. E.:
 Assessment of Sun photometer Langley calibration at the high-elevation sites Mauna Loa and Izaña,
 Atmospheric Chemistry and Physics, 18, 14555–14567, https://doi.org/10.5194/acp-18-14555-2018,
 2018.
- Torres, B., Toledano, C., Berjón, A., Fuertes, D., Molina, V., Gonzalez, R., Canini, M., Cachorro, V. E.,
 Goloub, P., Podvin, T., Blarel, L., Dubovik, O., Bennouna, Y., and de Frutos, A. M.: Measurements
 on pointing error and field of view of Cimel-318 Sun photometers in the scope of AERONET,
 Atmospheric Measurement Techniques, 6, 2207–2220, https://doi.org/10.5194/amt-6-2207-2013,
 2013.
- Wagner, F. and Silva, A. M.: Some considerations about Angström exponent distributions, Atmospheric
 Chemistry and Physics, 8, 481–489, https://doi.org/10.5194/acp-8-481-2008, 2008.
- Wang, J., Xia, X., Wang, P., and Christopher, S.: Diurnal variability of dust aerosol optical thickness and
 Ångströ⁻⁻m exponent over dust source regions in China, Geophysical Research Letters, 31,
 https://doi.org/10.1029/2004GL019580, 2004.
- Weinzierl, B., Sauer, D., Esselborn, M., Petzold, A., Veira, A., Rose, M., Mund, S., Wirth, M., Ansmann,
 A., Tesche, M., Gross, S., and Freudenthaler, V.: Microphysical and optical proper-ties of dust and
 tropical biomass burning aerosol layers in the Cape Verde region-an overview of the airborne in situ
 and lidar measurements during SAMUM-2, Tellus B, 63, 589–618, doi:10.1111/j.16000889.2011.00566.x, 2011.

- Wehrli, C.: Calibrations of filter radiometers for determination of atmospheric optical depth, Metrologia,
 37, 419, http://stacks.iop.org/ 0026-1394/37/i=5/a=16, 2000. Wehrli, C.: GAWPFR: A network of
 aerosol optical depth observations with precision filter radiometers, Global Atmosphere Watch 2005.
- Wehrli, C.: Precision Filter Radiometer Documentation, Version 4.0, 38 pp., Davos Dorf, 2008a.Wehrli,
 C.: Remote sensing of aerosol optical depth in a global surface network, Ph.D. thesis, ETH Zurich,
 https://www.research-collection.ethz.ch/bitstream/handle/20.500.11850/150574/eth-30693-02.pdf,
 2008b.
- 8 Wilbert, S.: Determination of circumsolar radiation and its effect on concentrating solar power, Ph.D. thesis,
 9 Hochschulbibliothek der Rheinisch-Westfälischen Technischen Hochschule Aachen, https://d 10 nb.info/1059537710/34, 2014.
- WMO: Recent Progress in Sunphotometry. Determination of the aerosol optical depth, Environmental
 Pollution Monitoring and Research Programme, N° 43, 21 pp, November,
 https://library.wmo.int/pmb_ged/wmo-td_143.pdf, 1986.
- WMO: Aerosol measurement procedures, guidelines and recommendations, GAW Report No. 153, WMO
 TD No. 1178, https://library.wmo. int/pmb_ged/wmo-td_1178.pdf, 2003.
- WMO: WMO/GAW Experts Workshop on a Global Surface-Based Network for Long Term Observations
 of Column Aerosol Optical Properties, GAW Report No. 162, WMO TD No. 1287,
 https://library.wmo.int/pmb_ged/wmo-td_1287.pdf, 2005.
- WMO: Abridged final report with resolutions and recommendations, GAW Report WMO TD No. 1019,
 WMO-CIMO Fourteenth session Geneva 7–14 December 2006, 2007. WMO: WMO/GAW Aerosol
 Measurement Procedures, Guidelines and Recommendations, 2nd Edition, WMO No 1177, GAW
 Report No. 227, 93 pp, <u>https://library.wmo.int/doc_num.php?explnum_id=3073</u>, 2016.
- Young, A. T.: Revised depolarization corrections for atmospheric extinction, Appl. Opt., 19, 3427– 3428,https://doi.org/10.1364/AO.19.003427, 1980.°