



Determination of n-alkanes, PAHs and hopanes in atmospheric aerosol: evaluation and comparison of thermal desorption GC-MS and solvent extraction GC-MS approaches

5 Meng Wang^{1, 2, 3}, Ru-Jin Huang^{1, 2}, Junji Cao^{1, 2}, Wenting Dai^{1, 2}, Jiamao Zhou^{1, 2}, Chunshui Lin^{1, 2}, Haiyan, Ni^{1, 2}, Jing Duan^{1, 2}, Ting Wang^{1, 2}, Yang Chen⁴, Yongjie Li⁵, Qi Chen⁶, Imad El Haddad⁷ and Thorsten Hoffmann⁸

¹State Key Laboratory of Loess and Quaternary Geology (SKLLQG) and Key Laboratory of Aerosol Chemistry & Physics (KLACP), Institute of Earth Environment, Chinese Academy of Sciences, Xi'an 710061, China

10 ²CAS Center for Excellence in Quaternary Science and Global Change, Xi'an, Chinese Academy of Sciences, 710061, China

³University of Chinese Academy of Sciences, Beijing 100049, China

⁴Key Laboratory of Reservoir Aquatic Environment of CAS, Chongqing Institute of Green and Intelligent Technology, Chinese Academy of Sciences, Chongqing 400714, China

15 ⁵Department of Civil and Environmental Engineering, Faculty of Science and Technology, University of Macau, Taipa, Macau, China

⁶State Key Joint Laboratory of Environmental Simulation and Pollution Control, College of Environmental Sciences and Engineering, Peking University, Beijing 100871, China

⁷Laboratory of Atmospheric Chemistry, Paul Scherrer Institute, 5232 Villigen PSI, Switzerland

20 ⁸Institute of Inorganic and Analytical Chemistry, Johannes Gutenberg University of Mainz, Duesbergweg 10–14, 55128 Mainz, Germany

Correspondence to: Ru-Jin Huang (rujin.huang@ieecas.cn)

Abstract. Organic aerosol (OA) constitutes a large fraction of fine particulate matter (PM) in the urban air. However, the chemical nature and sources of OA are not well constrained. Quantitative analysis of OA is essential for understanding the sources and atmospheric evolution of fine PM, which requires accurate quantification of some organic compounds (e.g., markers). In this study, two analytical approaches, i.e., thermal desorption (TD) gas chromatography-mass spectrometry (GC-MS) and solvent extract (SE) GC-MS were evaluated for the determination of n-alkanes, polycyclic aromatic hydrocarbons (PAHs), and hopanes in ambient aerosol. For the SE approach, the recovery obtained is 89.3–101.5%, the limits of detection (LOD) are 0.05–1.1 ng (1.5–33.9 ng m⁻³), repeatability is 3.5–14.5% and reproducibility is 1.2–10.9%. For the TD approach, the recovery is 57.2–109.8%, the LODs are 0.1–1.9 ng (0.04–0.9 ng m⁻³), repeatability is 2.1–19.4% and reproducibility is 1.1–12.9%. Ambient aerosol samples were collected from Beijing, Chengdu, Shanghai, and Guangzhou, during the winter of 2013 and were analyzed by the two methods. After considering the recoveries, the two methods show a good agreement with a high correlation efficient ($R^2 > 0.98$) and a slope close to unity. The concentrations of n-alkanes, PAHs, and hopanes are found to be much higher in Beijing than those in Chengdu, Shanghai, and Guangzhou, most likely due to emissions from coal combustion for wintertime heating in Beijing.



1 Introduction

Fine particulate matter (with aerodynamic diameters $\leq 2.5 \mu\text{m}$, $\text{PM}_{2.5}$) has significant impacts on human health (Cao et al., 2012; Lelieveld et al., 2015; Zheng et al., 2015; Cohen et al., 2017), visibility (Lin et al., 2014; Shen et al., 2015) and climate (Kanakidou et al., 2005; Gustafson et al., 2011; IPCC, 2014). It is composed of a large variety of inorganic and organic compounds, in which the organic fraction is much less constrained because of a very large chemical space with respect to molecular weight, functional groups and polarity (Zhang et al., 2007; Jimenez et al., 2009; Huang et al., 2014; Wang et al., 2018). Molecular-level characterization of the organic compounds in $\text{PM}_{2.5}$ is therefore essential for better understanding of the chemical nature, sources, atmospheric processes and impacts of $\text{PM}_{2.5}$ (Surratt et al., 2008; Chan et al., 2011; Fu et al., 2016). However, only 10–30% of the particulate organic matter has been identified as specific compounds despite years of effort and the use of the most sophisticated techniques available (Hoffmann et al., 2011).

The conventional approach for molecular analysis of organic compounds often requires the extraction of aerosol collected on filter into solvents followed by evaporation, separation, and detection by e.g. gas chromatography-mass spectrometry (GC-MS) (Kourtchev et al., 2009; Nozière et al., 2015; Maenhaut et al., 2017; Yu et al., 2017). However, solvent extraction (SE) method is prone to contamination and/or analyte lost during sample pre-treatment and is environmentally unfriendly because of the use of a large amount of solvent (Samy et al., 2010; Giri et al., 2013; Urban et al., 2014; Yang et al., 2017). In addition, the SE method is often time-consuming. Thermal desorption (TD) is an alternative approach which uses an elevated temperature to evaporate organic analytes from the filter for subsequent GC-MS detection (Ho and Yu, 2004; Chow et al., 2007; Xu et al., 2013). Compared to SE, the TD method is convenient and solvent free (Graham et al., 2010; Lambe et al., 2010; Ramírez et al., 2010; Kim et al., 2016). The application of TD/GC-MS on the qualitative and quantitative analysis of ambient organic compounds such as polycyclic aromatic hydrocarbons (PAHs) in $\text{PM}_{2.5}$ samples has demonstrated its analytical potential (Orecchio, 2011; Cheruiyot et al., 2015; Wang et al., 2015; Yang et al., 2017). However, the evaluation and comparison of TD/GC-MS with the well-established SE/GC-MS are still needed.

Particulate air pollution has become a serious environmental problem in China over the past decades (Chan and Yao, 2008; Zhao et al., 2009; Zhang et al., 2012; Jiang et al., 2015). Understanding the chemical composition and sources of fine particles is therefore essential for mitigating particulate air pollution. This is particularly true for organic aerosol (OA) which often constitutes more than 50% of fine PM mass in urban areas during haze pollution events (Takegawa et al., 2006; Srivastava et al., 2018) and is much less constrained compared to the inorganic fraction. In this study, we evaluated and compared the TD/GC-MS and SE/GC-MS methods in terms of their analytical performance including sensitivity, recovery, limit of detection (LOD), repeatability, and reproducibility. Then the two methods were applied for the determination of alkanes, PAHs, and hopanes in ambient aerosol samples collected in four Chinese megacities including Beijing, Shanghai, Guangzhou, and Chengdu.



2 Experimental

2.1 Sampling

Twenty-four-hour integrated PM_{2.5} samples were collected on pre-backed (780 °C, 3 h) quartz fiber filter (8 inch×10 inch) using a high-volume sampler at a flow rate of 1.05 m³ min⁻¹ from December 2013 to January 2014 in the four megacities in
5 China (Table S1). Eleven to twelve filters were collected from each sampling site, including 2 field blanks. The filters were wrapped in aluminum foils and stored at -20 °C until laboratory analysis.

2.2 SE/GC-MS analysis

A portion (4.34 cm²) of the sample or blank filter was cut with a clean stainless-steel punch, then extracted with a mixture of dichloromethane (DCM) (99.9%, LC grade, Mallinckrodt Laboratory Chemicals, Phillipsburg, NJ, USA) and methanol
10 (99.9%, LC grade, Mallinckrodt Laboratory Chemicals, Phillipsburg, NJ, USA) (3:1, v/v) under ultra-sonication for 15 min and filtered through quartz wool packed in a Pasteur pipette. The extraction procedure was repeated for 3 times. Then, the extracts were concentrated using a rotary evaporator under a vacuum condition and blown to < 0.5 mL with a gentle stream of pure nitrogen. The remaining solution was transferred to a 1.5 mL brown vial. Finally, 25 μL n-C₂₄D₅₀ (20 μg mL⁻¹) and 25 μL fluoranthene-d₁₀ (20 μg mL⁻¹) were added as internal standards. The final volume of the solution was fixed at 1.0 mL. For
15 the analysis of standards, a mixture of standards was spiked onto the pre-cleaned blank filters and then followed the same procedure as for ambient samples. The extracts were kept in a refrigerator until GC-MS analysis.

The GC-MS analysis was performed using an Agilent 7890 GC coupled with an Agilent 5975C mass spectrometer detector (MSD). The GC oven program was set at an initial value of 50 °C, held at this temperature for 1 min, programmed at a rate of 25 °C min⁻¹ to 140 °C and 10 °C min⁻¹ to 300 °C, and then held at the final temperature of 300 °C for 5 min. The column was
20 an HP-5MS (5% diphenyl/95% dimethylsiloxane, 30 m×0.25 mm×0.25 μm). The carrier gas was helium which was held at a constant pressure of 7.7 psi. The MSD was operated at 280 °C and 70 eV for electron ionization. The scan range of mass-to-charge ratio (m/z) was from 50 to 650 atomic mass unit (amu).

2.3 TD/GC-MS analysis

The filter punch of 0.53 cm² was cut with the stainless-steel punch. Two internal standards 25 μL n-C₂₄D₅₀ (20 μg mL⁻¹) and
25 25 μL fluoranthene-d₁₀ (20 μg mL⁻¹) with dichloromethane were spiked onto the filter punch. After air-drying for a few seconds, the punch was divided into four roughly equal portions with a razor blade, then loaded into the TD tube. The TD tube is a Pyrex glass tube, and it was cleaned with methanol under ultra-sonication for more than 30 min, then baked at 550 °C for at least 10 hours before use. A small amount of pre-baked glass wool was used to make two plugs for holding the filter pieces in the middle of the tube. After the loaded TD tube was placed in the injector port, the septum cap was closed and the injector-
30 port temperature was raised from 50 °C to 280 °C for desorption (about 8–9 min). During this period, the GC oven temperature



was initially at 30 °C, held at this temperature for 2 min, programmed at a rate of 20 °C min⁻¹ to 120 °C and 10 °C min⁻¹ to 300 °C, and then held at the final temperature of 300 °C for 10 min. The analysis time of each sample was about 52 minutes. The injector-port temperature was optimized by testing a set of temperature programs (see below). The injector was set in the splitless mode for the first 2 min in the GC temperature program, then switched to the split mode, and returned to the splitless mode at the end of the GC run. When the analysis was finished, the cooling program automatically started which took about 35 minutes. When the injection temperature was down to 50 °C, the next sample was put into the injector. A new TD tube was used for each analysis to avoid any potential contamination carry-over to the next analysis. The GC column was an HP-5MS (5% diphenyl/95% dimethylsiloxane, 30 m×0.25 mm×0.25 μm). The carrier gas was helium which was held at 7.7 psi. The MSD was operated at 280 °C and 70 eV for electron ionization. The mass scan range (m/z) was the same with the SE method (50–650 amu).

2.4 Quality assurance and control (QA/QC)

Twenty-four n-alkanes (C₁₄–C₃₇) (99%, Aldrich, Milwaukee, WI, USA), twelve PAHs (99%, Aldrich, Milwaukee, WI, USA) (Acenaphthene (Ace), Fluorene (Fl), Phenanthrene (Phe), Anthracene (Ant), Fluoranthene (Fla), Pyrene (Pyr), Benz(a)anthracene (BaA), Chrysene (Chry), Benzo(b)fluoranthene (BbF), Benzo(k)fluoranthene (BkF), Benzo(a)pyrene (BaP), Indeno(1,2,3-cd)pyrene (IcdP), Dibenz(a,h)anthracene (DahA), Benzo(g,h,i)perylene (BghiP)) and eight hopanes (99%, Supelco, Bellefonte, PA, USA) (22,29,30-Trisnorhopane (TH), C₂₉αβ-hopane (C₂₉αβH), C₂₉αα-+βα-hopane (C₂₉αα-+βαH), C₃₀αβ-hopane (C₃₀αβH), C₃₀αα-hopane (C₃₀ααH), C₃₀βα-hopane (C₃₀βαH), C₃₁αβS-hopane (C₃₁αβSH), C₃₁αβR-hopane (C₃₁αβRH)) were quantified using both the TD and SE methods. Internal standards (n-C₂₄D₅₀ (98%, Aldrich, Milwaukee, WI, USA) and fluoranthene-d₁₀ (98%, Aldrich, Milwaukee, WI, USA)) were added during the pre-treatment process for two methods. The calibration curves were established by the standard solution of different concentrations for individual target compounds, and the correlation coefficients (R²) for linear regressions of the calibration curves were > 0.99 (Table S2). Repeatability and reproducibility were also investigated for both methods. Repeatability refers to the variation in repeat measurements made under identical conditions, e.g., by the same method and the same individual while reproducibility refers to the variation in measurements made by same method but by different individuals (Bartlett and Frost, 2008). In this study, seven repeat measurements were made on n-alkanes (1000 ng mL⁻¹), PAHs (1000 ng mL⁻¹), and hopanes (1000 ng mL⁻¹) standards spiked on blank filters for both repeatability and reproducibility studies.

For every ten samples, one replicate analysis was performed; and for every 5 samples, one blank was analyzed to check any potential contamination. In addition, ambient samples spiked with known amounts of internal and external standards were analyzed to check the potential interference. All data reported here were corrected for the blanks.



3 Results and discussion

3.1 Temperature effect on thermal desorption

The effect of temperature in the injector port on thermal desorption is investigated by analyzing a mixture of n-alkane standards (C_{14} – C_{37}) with TD/GC-MS and compared with SE/GC-MS. The vapor pressure of these n-alkanes (molecular weight, MW, 198–520) at 275 °C ranges from 0.5 kPa to 177.6 kPa, covering the vapor pressure of PAHs (0.7–97.3 kPa, MW 154–278) and hopanes (0.2–2.4 kPa, MW 370–426) investigated here. Therefore, only n-alkanes were tested to investigate the temperature effect. A range of temperature in the injector port (from 255 °C to 300 °C) was studied and the results are compared with those from the SE method (see Fig. 1). It can be seen from the figure that at lower TD temperatures (255 °C and 260 °C, see Fig. 1b and c), lighter n-alkanes have higher signal abundances than heavier ones, with the abundance of C_{16} being the highest. However, the abundances of individual n-alkanes measured at these lower TD temperatures are only approximately 20% of those from the SE method, likely due to incomplete thermal desorption (Ho and Yu, 2004). With increased TD temperature (275 °C, Fig. 1d), the signal abundances for all n-alkanes species increase by 2–5 times, indicating a better thermal desorption efficiency. When the TD temperature increases further to 280 °C, the signal abundances increase slightly, particularly for those lighter n-alkanes (C_{14} – C_{19}). It should be noted that the signal abundance patterns change with the TD temperature (Fig. 1b–e), with those mid-chain n-alkanes being more abundant at 275 °C and 280 °C. However, as the TD temperature further increased to 290 °C and 300 °C (Fig. 1f and g), the signal abundances show a decreasing trend, for example, > 30% decrease in abundance for mid-chain n-alkanes at 300 °C was observed. The decrease in signal abundances at higher temperatures (e.g., 290 °C and 300 °C) is likely associated with pyrolysis. Therefore, 280 °C was selected as the optimized TD temperature to compromise for the thermal desorption efficiency and pyrolysis effect. Note that even at the optimized TD temperature, the abundances of individual n-alkanes (Fig. 1e) show some differences from the SE method. In particular, the abundances of lighter n-alkanes (C_{14} – C_{19}) from TD method are about 50% lower than those from SE method, and for heavier n-alkanes (C_{31} – C_{37}) the decreases are more than 50%. However, the abundances for mid-chain n-alkanes are very similar between TD and SE methods.

3.2 Comparison of the analytical performance

Table 1 shows the comparison of repeatability, reproducibility, recovery, and LODs of SE/GC-MS and TD/GC-MS for n-alkanes, PAHs, and hopanes. Repeatability and reproducibility are two measures of the precision of the analytical methods. Repeatability is determined by the relative standard deviation (RSD) through measuring the identical samples ($n = 7$) with the same procedure while reproducibility represents the consistency between the repeated measurements by different individuals. The repeatability of SE/GC-MS is 3.5–14.5%, 4.4–10.6%, and 3.5–7.7% for n-alkanes, PAHs, and hopanes, respectively, which are similar to those obtained from TD/GC-MS (3.2–19.3%, 2.7–9.5%, and 2.1–8.4%, correspondingly). The reproducibility is also consistent for these two methods (1.2–10.9% for SE versus 1.1–12.9% for TD). In general, the precision



of both SE/GC-MS and TD/GC-MS methods are satisfactory ($RSD < 20\%$) considering the multiple sources of the errors (e.g., the error of sample pre-treatment and error from GC-MS detection).

The recoveries for the two methods were determined by analyzing ambient samples spiked with known amounts of n-alkanes (1000 ng mL⁻¹), PAHs (1000 ng mL⁻¹), and hopanes (1000 ng mL⁻¹) standards, respectively. The recoveries of SE/GC-MS are in the range of 89.4–98.6% for n-alkanes (Table 1) which are better than those obtained from TD/GC-MS (57.2–109.8%). The good recoveries provided by SE/GC-MS indicate a good extraction efficiency and low sample matrix effect. For TD/GC-MS method, only lighter n-alkanes (C₁₄–C₁₈) and heavier n-alkanes (C₃₂–C₃₇) show relatively lower recoveries (57.2–80.6%) (see Fig. 2), likely due to insufficient thermal desorption efficiency and/or pyrolysis effect as discussed above. The recoveries of PAHs and hopanes in TD/GC-MS method are 89.4–106.1% and 90.4–99.8%, respectively, which are as good as those obtained from SE/GC-MS method (89.7–101.5% and 90.3–97.2%, respectively).

The LODs (the concentration with a signal to noise ratio of three) of the SE method are 0.05–0.9 ng for n-alkanes, 0.3–1.1 ng for PAHs, and 0.3–1.0 ng for hopanes. These LODs are very similar to those from the TD method, i.e., 0.09–0.8 ng for n-alkanes, 0.1–1.9 ng for PAHs, and 0.2–0.7 ng for hopanes.

3.3 Comparison based on ambient aerosol samples

The comparison between TD and SE methods discussed above are made base on standard compounds. When analyzing ambient samples, however, the analytical performance may be affected by e.g., matrix effect. We evaluate these potential effects by analyzing 45 ambient PM_{2.5} samples collected in Beijing, Chengdu, Shanghai, and Guangzhou. Fig. 3 shows the correlation of quantitative results between SE and TD methods for n-alkanes (Fig. 3a), PAHs (Fig. 3b), and hopanes (Fig. 3c). Note that the data shown in Fig.3 are corrected for individual recoveries. The good correlations between these two methods for n-alkanes ($R^2 = 0.99$), PAHs ($R^2 = 0.98$), and hopanes ($R^2 = 0.98$) with the slopes close to unity (0.94–0.98) suggest that both TD and SE methods can be used for quantitative measurements of ambient aerosol samples.

The LODs of a method are very important for ambient measurements, particularly in regions with low concentrations or sampling with low-volume samplers. Our results show that, for example, for samples collected with a high-volume sampler at 1.1 m³ min⁻¹ for 24 h on 8" × 10" filters, 0.53 cm² filter punch (TD method) is required to reach LODs of 0.04–0.4 ng m⁻³ for n-alkanes, 0.06–0.9 ng m⁻³ for PAHs, and 0.09–0.4 ng m⁻³ for hopanes; and 4.3 cm² filter punch (SE method) is required to reach LODs of 1.5–27.4 ng m⁻³ for n-alkanes, 9.8–33.9 ng m⁻³ for PAHs, and 8.1–29.8 ng m⁻³ for hopanes. In this regard, the TD method is superior to the SE method because the LODs of the TD method are about two orders of magnitude lower than the SE method and meanwhile it requires about 10 times lower sample material.

3.4 n-alkanes, PAHs, and hopanes in wintertime PM_{2.5} in Chinese megacities

Figure 4 shows the average concentrations of individual n-alkanes species (C₁₄–C₃₇) during the winter time in Beijing, Chengdu, Shanghai, and Guangzhou. Short chain n-alkanes (C_n ≤ C₂₆) are mainly derived from anthropogenic emissions while



long chain n-alkanes ($C_n > C_{26}$) are typical of the biogenic source (Xu et al., 2013). The short chain n-alkanes accounts for 72.2% (262.0 ng m^{-3}) of total n-alkanes in Beijing, 68.4% (246.7 ng m^{-3}) in Chengdu, 64.6% (111.6 ng m^{-3}) in Shanghai, and 55.8% (102.6 ng m^{-3}) in Guangzhou. Specifically, C_{25} exhibits the highest concentration (66.1 ng m^{-3}) in Beijing while C_{24} (48.0 ng m^{-3}) is the highest in Chengdu, C_{22} (19.8 ng m^{-3}) in Shanghai, and C_{26} (19.3 ng m^{-3}) in Guangzhou. These results indicate that the major source of n-alkanes is from anthropogenic emissions (e.g., traffic and/or coal combustion) in Beijing, Chengdu and Shanghai while the biogenic source is more important in Guangzhou. In addition, the short chain n-alkanes concentrations are 1.1–2.6 times higher in Beijing than in the other 3 cities, further supporting the higher anthropogenic emissions in Beijing. This is consistent with the higher traffic fleets and larger coal usage in Beijing than in the other 3 cities studied here (Huang et al., 2014). In fact, Beijing is the only city equipped with residential heating in the four cities studied here. In contrast, higher contribution from the biogenic source in Guangzhou is likely associated with the higher temperature and vegetation coverage in southern China.

The average concentrations of individual PAHs species in the four cities are shown in Fig. 5. The contributions of PAHs are site dependent, indicating different sources in different cities (Okuda et al., 2006; Dockery and Stone, 2007). For example, BbF is the most abundant PAHs in Beijing with an average concentration of 7.3 ng m^{-3} while Phe (6.9 ng m^{-3}) exhibits the highest concentration in Chengdu, BbF (3.5 ng m^{-3}) is the highest in Shanghai and BaA (3.0 ng m^{-3}) in Guangzhou. BbF is mainly emitted from coal combustion with high emission factor while BghiP is mainly derived from traffic emission (Katsoyiannis et al., 2011; Tobiszewski and Namieśnik, 2012). The BbF concentration (7.3 ng m^{-3}) in Beijing is 2–3 times as much as that in the other three cities (Chengdu 3.6 ng m^{-3} , Shanghai 3.5 ng m^{-3} , and Guangzhou 2.2 ng m^{-3}), supporting its large emissions from residential coal combustion in Beijing. The BghiP concentrations Beijing and Chengdu ($3\text{--}4 \text{ ng m}^{-3}$) are rather similar, indicating similar levels of traffic emission in the two cities. In addition, the BaP concentrations are similar in Beijing and Chengdu ($3\text{--}4 \text{ ng m}^{-3}$) and similar in Shanghai and Guangzhou ($2.0\text{--}2.5 \text{ ng m}^{-3}$). They all exceed the guideline from the World Health Organization (WHO) by a factor of 2–4, indicating potential health risk as BaP is a carcinogenic indicator.

Hopanes are pentacyclic hydrocarbons with triterpene group and are abundant in crude oils and lubricant oils (Oros and Simoneit, 2000; Schauer et al., 2002; Zhang et al., 2008). As shown in Fig. 6, Beijing exhibits the highest concentrations of hopanes among the four cities studied. $C_{30}\alpha\beta\text{H}$ is the most abundant hopane species in the four cities, with the highest concentration in Beijing (2.0 ng m^{-3}) and similar concentrations in Chengdu, Shanghai, and Guangzhou ($\sim 0.6 \text{ ng m}^{-3}$). Such a large difference could be attributed to coal combustion emissions in Beijing.

Generally, the total n-alkane concentrations rank in the order Shanghai < Guangzhou < Chengdu < Beijing, but PAHs and hopanes concentrations are in the order of Guangzhou < Shanghai < Chengdu < Beijing (Fig. 7). The average n-alkanes concentration in Beijing and Chengdu ($\sim 400 \text{ ng m}^{-3}$) are 2 times that of Guangzhou and Shanghai ($\sim 200 \text{ ng m}^{-3}$). The average concentrations of PAHs in Beijing are $46.4\text{--}47.1 \text{ ng m}^{-3}$, which are nearly two times higher than those in Guangzhou ($20.8\text{--}21.3 \text{ ng m}^{-3}$), 1.7 times higher than those in Shanghai ($26.2\text{--}26.7 \text{ ng m}^{-3}$), and 1.3 times higher than those in Chengdu (35.0--



35.6 ng m⁻³). For hopanes, the concentrations in Beijing (~6 ng m⁻³) are 3 times higher than those in Chengdu, around 4 times higher than those in Shanghai and Guangzhou.

4 Conclusion

This study evaluated and compared the analytical performances of SE/GC-MS and TD/GC-MS for the determination of n-alkanes, PAHs and hopanes by analyzing both standards and ambient PM_{2.5} samples from four Chinese megacities. For the SE method, the recovery obtained is 89.4–101.5%, the limits of detection (LOD) are 0.05–1.1 ng (1.5–33.9 ng m⁻³), repeatability is 3.5–14.5% and reproducibility is 1.2–10.9%, while for the TD approach, the recovery is 57.2–109.8%, the limits of detection are 0.09–1.9 ng (0.04–0.9 ng m⁻³), repeatability is 2.1–19.4% and reproducibility is 1.1–12.9%. Ambient aerosol samples were simultaneously collected from Beijing, Chengdu, Shanghai, and Guangzhou during the winter of 2013 and were analyzed by the two methods. Although TD/GC-MS exhibit relatively lower recoveries for lighter n-alkanes and heavier n-alkanes, after correction the two methods show a good agreement with a high correlation efficient ($R^2 > 0.98$) and a slope close to unity. Moreover, n-alkanes, PAHs, and hopanes concentrations in wintertime are found to be much higher in Beijing than those in Chengdu, Shanghai, and Guangzhou, likely due to the coal combustion for heating in Beijing.

Data availability

Raw data used in this study are archived at the Institute of Earth Environment, Chinese Academy of Sciences, and are available on request by contacting the corresponding author.

The Supplement related to this article is available online at

Competing interests

The authors declare that they have no conflict of interest.

20 Acknowledgments

This work was supported by the National Natural Science Foundation of China (NSFC) under grant No. 91644219, No. 41877408 and 4171101096), the National Key Research and Development Program of China (No. 2017YFC0212701).

References

- Bartlett, J. W., and Frost, C.: Reliability, repeatability and reproducibility: analysis of measurement errors in continuous variables, 31, 466-475, <http://doi:10.1002/uog.5256>, 2008.
- Cao, J., Xu, H., Xu, Q., Chen, B., and Kan, H.: Fine particulate matter constituents and cardiopulmonary mortality in a heavily polluted Chinese city, Environ. Health Perspect., 120, 373-378, <https://doi.org/10.1289/ehp.1103671>, 2012.



- Chan, C. K., and Yao, X.: Air pollution in mega cities in China, *Atmos. Environ.*, 42, 1-42, <https://doi.org/10.1016/j.atmosenv.2007.09.003>, 2008.
- Chan, M. N., Surratt, J. D., Chan, A. W. H., Schilling, K., Offenberg, J. H., Lewandowski, M., Edney, E. O., Kleindienst, T. E., Jaoui, M., Edgerton, E. S., Tanner, R. L., Shaw, S. L., Zheng, M., Knipping, E. M., and Seinfeld, J. H.: Influence of aerosol acidity on the chemical composition of secondary organic aerosol from β -caryophyllene, *Atmos. Chem. Phys.*, 11, 1735-1751, <https://doi.org/10.5194/acp-11-1735-2011>, 2011.
- Cheruiyot, N. K., Lee, W. J., Mwangi, J. K., Wang, L. C., Lin, N. H., Lin, Y. C., Cao, J., Zhang, R., and Guo-Ping, C. C.: An Overview: Polycyclic Aromatic Hydrocarbon Emissions from the Stationary and Mobile Sources and in the Ambient Air, *Aerosol Air Qual. Res.*, 15, 2730-2762, <https://doi.org/10.4209/aaqr.2015.11.0627>, 2015.
- Chow, J. C., Yu, J. Z., Watson, J. G., Hang Ho, S. S., Bohannon, T. L., Hays, M. D., and Fung, K. K.: The application of thermal methods for determining chemical composition of carbonaceous aerosols: A review, *J. Environ. Sci. Health, Part A*, 42, 1521-1541, <https://doi.org/10.1080/10934520701513365>, 2007.
- Cohen, A. J., Brauer, M., Burnett, R., Anderson, H. R., Frostad, J., Estep, K., Balakrishnan, K., Brunekreef, B., Dandona, L., Dandona, R., Feigin, V., Freedman, G., Hubbell, B., Jobling, A., Kan, H., Knibbs, L., Liu, Y., Martin, R., Morawska, L., Pope, C. A., Shin, H., Straif, K., Shaddick, G., Thomas, M., van Dingenen, R., van Donkelaar, A., Vos, T., Murray, C. J. L., and Forouzanfar, M. H.: Estimates and 25-year trends of the global burden of disease attributable to ambient air pollution: an analysis of data from the Global Burden of Diseases Study 2015, *Lancet*, 389, 1907-1918, [https://doi.org/10.1016/S0140-6736\(17\)30505-6](https://doi.org/10.1016/S0140-6736(17)30505-6), 2017.
- Dockery, D. W., and Stone, P. H.: Cardiovascular Risks from Fine Particulate Air Pollution, *N. Engl. J. Med.*, 356, 511-513, <https://doi.org/10.1056/NEJMe068274>, 2007.
- Fu, P., Aggarwal, S. G., Chen, J., Li, J., Sun, Y., Wang, Z., Chen, H., Liao, H., Ding, A., Umarji, G. S., Patil, R. S., Chen, Q., and Kawamura, K.: Molecular Markers of Secondary Organic Aerosol in Mumbai, India, *Environ. Sci. Technol.*, 50, 4659-4667, <https://doi.org/10.1021/acs.est.6b00372>, 2016.
- Giri, B., Patel, K. S., Jaiswal, N. K., Sharma, S., Ambade, B., Wang, W., Simonich, S. L. M., and Simoneit, B. R.: Composition and sources of organic tracers in aerosol particles of industrial central India, *Atmos. Res.*, 120, 312-324, <https://doi.org/10.1016/j.atmosres.2012.09.016>, 2013.
- Graham, L. A., Tong, A., Poole, G., Ding, L., Ke, F., Wang, D., Caravaggio, G., Charland, J.-P., MacDonald, P., and Hall, A.: A comparison of direct thermal desorption with solvent extraction for gas chromatography-mass spectrometry analysis of semivolatile organic compounds in diesel particulate matter, *International Journal of Environmental and Analytical Chemistry*, 90, 511-534, <https://doi.org/10.1080/03067310903108352>, 2010.
- Gustafson, W. I., Qian, Y., and Fast, J. D.: Downscaling aerosols and the impact of neglected subgrid processes on direct aerosol radiative forcing for a representative global climate model grid spacing, *J. Geophys. Res.: Atmos.*, 116, <https://doi.org/10.1029/2010JD015480> 2011.



- Ho, S. S. H., and Yu, J. Z.: In-injection port thermal desorption and subsequent gas chromatography–mass spectrometric analysis of polycyclic aromatic hydrocarbons and n-alkanes in atmospheric aerosol samples, *J. Chromatogr. A*, 1059, 121-129, <https://doi.org/10.1016/j.chroma.2004.10.013>, 2004.
- Hoffmann, T., Huang, R.-J., and Kalberer, M.: Atmospheric Analytical Chemistry, *Anal. Chem.*, 83, 4649-4664, <https://doi.org/10.1021/ac2010718>, 2011.
- Huang, R. J., Zhang, Y., Bozzetti, C., Ho, K. F., Cao, J. J., Han, Y., Daellenbach, K. R., Slowik, J. G., Platt, S. M., Canonaco, F., Zotter, P., Wolf, R., Pieber, S. M., Bruns, E. A., Crippa, M., Ciarelli, G., Piazzalunga, A., Schwikowski, M., Abbaszade, G., Schnelle-Kreis, J., Zimmermann, R., An, Z., Szidat, S., Baltensperger, U., Haddad, I. E., and Prévôt, A. S. H.: High secondary aerosol contribution to particulate pollution during haze events in China, *Nature*, 514, 218, <https://doi.org/10.1038/nature13774>, 2014.
- IPCC: Climate change 2013: the physical science basis: Working Group I contribution to the Fifth assessment report of the Intergovernmental Panel on Climate Change, Cambridge University Press, 2014.
- Jiang, Q., Sun, Y. L., Wang, Z., and Yin, Y.: Aerosol composition and sources during the Chinese Spring Festival: fireworks, secondary aerosol, and holiday effects, *Atmos. Chem. Phys.*, 15, 6023-6034, <https://doi.org/10.5194/acp-15-6023-2015>, 2015.
- Jimenez, J. L., Canagaratna, M. R., Donahue, N. M., Prevot, A. S. H., Zhang, Q., Kroll, J. H., DeCarlo, P. F., Allan, J. D., Coe, H., Ng, N. L., Aiken, A. C., Docherty, K. S., Ulbrich, I. M., Grieshop, A. P., Robinson, A. L., Duplissy, J., Smith, J. D., Wilson, K. R., Lanz, V. A., Hueglin, C., Sun, Y. L., Tian, J., Laaksonen, A., Raatikainen, T., Rautiainen, J., Vaattovaara, P., Ehn, M., Kulmala, M., Tomlinson, J. M., Collins, D. R., Cubison, M. J., Dunlea, J., Huffman, J. A., Onasch, T. B., Alfarra, M. R., Williams, P. I., Bower, K., Kondo, Y., Schneider, J., Drewnick, F., Borrmann, S., Weimer, S., Demerjian, K., Salcedo, D., Cottrell, L., Griffin, R., Takami, A., Miyoshi, T., Hatakeyama, S., Shimono, A., Sun, J. Y., Zhang, Y. M., Dzepina, K., Kimmel, J. R., Sueper, D., Jayne, J. T., Herndon, S. C., Trimborn, A. M., Williams, L. R., Wood, E. C., Middlebrook, A. M., Kolb, C. E., Baltensperger, U., and Worsnop, D. R.: Evolution of Organic Aerosols in the Atmosphere, *Science*, 326, 1525-1529, <https://doi.org/10.1126/science.1180353>, 2009.
- Kanakidou, M., Seinfeld, J. H., Pandis, S. N., Barnes, I., Dentener, F. J., Facchini, M. C., Van Dingenen, R., Ervens, B., Nenes, A., Nielsen, C. J., Swietlicki, E., Putaud, J. P., Balkanski, Y., Fuzzi, S., Horth, J., Moortgat, G. K., Winterhalter, R., Myhre, C. E. L., Tsigaridis, K., Vignati, E., Stephanou, E. G., and Wilson, J.: Organic aerosol and global climate modelling: a review, *Atmos. Chem. Phys.*, 5, 1053-1123, <https://doi.org/10.5194/acp-5-1053-2005>, 2005.
- Katsoyiannis, A., Sweetman, A. J., and Jones, K. C.: PAH Molecular Diagnostic Ratios Applied to Atmospheric Sources: A Critical Evaluation Using Two Decades of Source Inventory and Air Concentration Data from the UK, *Environ. Sci. Technol.*, 45, 8897-8906, <https://doi.org/10.1021/es202277u>, 2011.
- Kim, J. W., Kim, Y.-M., Moon, H. M., Hosaka, A., Watanabe, C., Teramae, N., Choe, E. K., and Myung, S.-W.: Comparative study of thermal desorption and solvent extraction–gas chromatography–mass spectrometric analysis for the quantification of phthalates in polymers, *J. Chromatogr. A*, 1451, 33-40, <https://doi.org/10.1016/j.chroma.2016.05.014>, 2016.



- Kourtchev, I., Copolovici, L., Claeys, M., and Maenhaut, W.: Characterization of Atmospheric Aerosols at a Forested Site in Central Europe, 43, 4665-4671, <https://doi.org/10.1021/es803055w>, 2009.
- Lambe, A. T., Chacon-Madrid, H. J., Nguyen, N. T., Weitkamp, E. A., Kreisberg, N. M., Hering, S. V., Goldstein, A. H., Donahue, N. M., and Robinson, A. L.: Organic aerosol speciation: Intercomparison of thermal desorption aerosol GC/MS (TAG) and filter-based techniques, *Aerosol Sci. Technol.*, 44, 141-151, <https://doi.org/10.1080/02786820903447206>, 2010.
- 5 Lelieveld, J., Evans, J. S., Fnais, M., Giannadaki, D., and Pozzer, A.: The contribution of outdoor air pollution sources to premature mortality on a global scale, 525, 367, <https://doi.org/10.1038/nature15371>, 2015.
- Lin, Y., Huang, K., Zhuang, G., Fu, J. S., Wang, Q., Liu, T., Deng, C., and Fu, Q.: A multi-year evolution of aerosol chemistry impacting visibility and haze formation over an Eastern Asia megacity, Shanghai, *Atmos. Environ.*, 92, 76-86, <https://doi.org/10.1016/j.atmosenv.2014.04.007>, 2014.
- 10 Maenhaut, W., Chi, X., Wang, W., Cafmeyer, J., Yasmeeen, F., Vermeylen, R., Szmigielska, K., Janssens, I. A., and Claeys, M.: Contribution from Selected Organic Species to PM_{2.5} Aerosol during a Summer Field Campaign at K-Pusztá, Hungary, *Atmosphere*, 8, 221, <https://doi.org/10.3390/atmos8110221>, 2017.
- Nozière, B., Kalberer, M., Claeys, M., Allan, J., D'Anna, B., Decesari, S., Finessi, E., Glasius, M., Grgić, I., Hamilton, J. F., Hoffmann, T., Iinuma, Y., Jaoui, M., Kahnt, A., Kampf, C. J., Kourtchev, I., Maenhaut, W., Marsden, N., Saarikoski, S., Schnelle-Kreis, J., Surratt, J. D., Szidat, S., Szmigielski, R., and Wisthaler, A.: The Molecular Identification of Organic Compounds in the Atmosphere: State of the Art and Challenges, *Chem. Rev.*, 115, 3919-3983, [10.1021/cr5003485](https://doi.org/10.1021/cr5003485), 2015.
- 15 Okuda, T., Naoi, D., Tenmoku, M., Tanaka, S., He, K., Ma, Y., Yang, F., Lei, Y., Jia, Y., and Zhang, D.: Polycyclic aromatic hydrocarbons (PAHs) in the aerosol in Beijing, China, measured by aminopropylsilane chemically-bonded stationary-phase column chromatography and HPLC/fluorescence detection, *Chemosphere*, 65, 427-435, <https://doi.org/10.1016/j.chemosphere.2006.01.064>, 2006.
- 20 Orecchio, S.: Polycyclic aromatic hydrocarbons (PAHs) in indoor emission from decorative candles, *Atmos. Environ.*, 45, 1888-1895, <https://doi.org/10.1016/j.atmosenv.2010.12.024>, 2011.
- Oros, D. R., and Simoneit, B. R. T.: Identification and emission rates of molecular tracers in coal smoke particulate matter, *Fuel*, 79, 515-536, [https://doi.org/10.1016/S0016-2361\(99\)00153-2](https://doi.org/10.1016/S0016-2361(99)00153-2), 2000.
- 25 Ramírez, N., Cuadras, A., Rovira, E., Borrull, F., and Marcé, R. M.: Comparative study of solvent extraction and thermal desorption methods for determining a wide range of volatile organic compounds in ambient air, *Talanta*, 82, 719-727, <https://doi.org/10.1016/j.talanta.2010.05.038>, 2010.
- Samy, S., Mazzoleni, L. R., Mishra, S., Zielinska, B., and Hallar, A. G.: Water-soluble organic compounds at a mountain-top site in Colorado, USA, *Atmos. Environ.*, 44, 1663-1671, <https://doi.org/10.1016/j.atmosenv.2010.01.033>, 2010.
- 30 Schauer, J. J., Kleeman, M. J., Cass, G. R., and Simoneit, B. R. T.: Measurement of Emissions from Air Pollution Sources. 5. C₁-C₃₂ Organic Compounds from Gasoline-Powered Motor Vehicles, *Environ. Sci. Technol.*, 36, 1169-1180, <https://doi.org/10.1021/es0108077>, 2002.



- Shen, X. J., Sun, J. Y., Zhang, X. Y., Zhang, Y. M., Zhang, L., Che, H. C., Ma, Q. L., Yu, X. M., Yue, Y., and Zhang, Y. W.: Characterization of submicron aerosols and effect on visibility during a severe haze-fog episode in Yangtze River Delta, China, *Atmos. Environ.*, 120, 307-316, <https://doi.org/10.1016/j.atmosenv.2015.09.011>, 2015.
- Srivastava, D., Favez, O., Perraudin, E., Villenave, E., and Albinet, A.: Comparison of measurement-based methodologies to apportion secondary organic carbon (SOC) in PM_{2.5}: a review of recent studies, *Atmosphere*, 9, 452, 2018.
- Surratt, J. D., Gómez-González, Y., Chan, A. W. H., Vermeylen, R., Shahgholi, M., Kleindienst, T. E., Edney, E. O., Offenberg, J. H., Lewandowski, M., Jaoui, M., Maenhaut, W., Claeys, M., Flagan, R. C., and Seinfeld, J. H.: Organosulfate Formation in Biogenic Secondary Organic Aerosol, *J. Phys. Chem. A*, 112, 8345-8378, <https://doi.org/10.1021/jp802310p>, 2008.
- 10 Takegawa, N., Miyakawa, T., Kondo, Y., Jimenez, J., Zhang, Q., Worsnop, D., and Fukuda, M.: Seasonal and diurnal variations of submicron organic aerosol in Tokyo observed using the Aerodyne aerosol mass spectrometer, *J. Geophys. Res.: Atmos.*, 111, 2006.
- Tobiszewski, M., and Namieśnik, J.: PAH diagnostic ratios for the identification of pollution emission sources, *Environ. Pollut.*, 162, 110-119, <https://doi.org/10.1016/j.envpol.2011.10.025>, 2012.
- 15 Urban, R., Alves, C., Allen, A. G., Cardoso, A. A., Queiroz, M., and Campos, M.: Sugar markers in aerosol particles from an agro-industrial region in Brazil, *Atmos. Environ.*, 90, 106-112, <https://doi.org/10.1016/j.atmosenv.2014.03.034>, 2014.
- Wang, J., Ho, S. S. H., Cao, J., Huang, R., Zhou, J., Zhao, Y., Xu, H., Liu, S., Wang, G., and Shen, Z.: Characteristics and major sources of carbonaceous aerosols in PM_{2.5} from Sanya, China, *Sci. Total Environ.*, 530, 110-119, <https://doi.org/10.1016/j.scitotenv.2015.05.005>, 2015.
- 20 Wang, K., Zhang, Y., Huang, R.-J., Cao, J., and Hoffmann, T.: UHPLC-Orbitrap mass spectrometric characterization of organic aerosol from a central European city (Mainz, Germany) and a Chinese megacity (Beijing), 189, 22-29, <https://doi.org/10.1016/j.atmosenv.2018.06.036>, 2018.
- Xu, H. M., Tao, J., Ho, S. S. H., Ho, K. F., Cao, J. J., Li, N., Chow, J. C., Wang, G. H., Han, Y. M., Zhang, R. J., Watson, J. G., and Zhang, J. Q.: Characteristics of fine particulate non-polar organic compounds in Guangzhou during the 16th Asian
- 25 Games: Effectiveness of air pollution controls, *Atmos. Environ.*, 76, 94-101, <https://doi.org/10.1016/j.atmosenv.2012.12.037>, 2013.
- Yang, T.-T., Hsu, C.-Y., Chen, Y.-C., Young, L.-H., Huang, C.-H., and Ku, C.-H.: Characteristics, sources, and health risks of atmospheric PM_{2.5}-bound polycyclic aromatic hydrocarbons in Hsinchu, Taiwan, *Aerosol Air Qual. Res.*, 17, 563-573, <https://doi.org/10.4209/aaqr.2016.06.0283>, 2017.
- 30 Yu, R., Duan, L., Jiang, J., and Hao, J.: An optimized two-step derivatization method for analyzing diethylene glycol ozonation products using gas chromatography and mass spectrometry, *J. Environ. Sci.*, 53, 313-321, <https://doi.org/10.1016/j.jes.2016.02.021>, 2017.



- Zhang, Q., Jimenez, J. L., Canagaratna, M., Allan, J., Coe, H., Ulbrich, I., Alfarra, M., Takami, A., Middlebrook, A., and Sun, Y.: Ubiquity and dominance of oxygenated species in organic aerosols in anthropogenically-influenced Northern Hemisphere midlatitudes, *Geophys. Res. Lett.*, 34, 2007.
- Zhang, Q., He, K., and Huo, H.: Cleaning China's air, 484, 161, <https://doi.org/10.1038/484161a>, 2012.
- 5 Zhang, Y., Schauer, J. J., Zhang, Y., Zeng, L., Wei, Y., Liu, Y., and Shao, M.: Characteristics of Particulate Carbon Emissions from Real-World Chinese Coal Combustion, *Environ. Sci. Technol.*, 42, 5068-5073, <https://doi.org/10.1021/es7022576>, 2008.
- Zhao, X., Zhang, X., Xu, X., Xu, J., Meng, W., and Pu, W.: Seasonal and diurnal variations of ambient PM_{2.5} concentration in urban and rural environments in Beijing, *Atmos. Environ.*, 43, 2893-2900, <https://doi.org/10.1016/j.atmosenv.2009.03.009>, 2009.
- 10 Zheng, S., Pozzer, A., Cao, C. X., and Lelieveld, J.: Long-term (2001–2012) concentrations of fine particulate matter (PM_{2.5}) and the impact on human health in Beijing, China, *Atmos. Chem. Phys.*, 15, 5715-5725, <https://doi.org/10.5194/acp-15-5715-2015>, 2015.



Table 1. Analytical performance of SE and TD/GC-MS methods for determination of n-alkanes, PAHs, and hopanes.

	SE method					TD method				
	Repeatability (%)	Reproducibility (%)	Recovery (%)	LOD (ng)	LOD* (ng m ⁻³)	Repeatability (%)	Reproducibility (%)	Recovery (%)	LOD (ng)	LOD* (ng m ⁻³)
n-Alkanes	3.5–14.5	1.2–10.9	89.4–98.6	0.05–0.9	1.5–27.4	3.2–19.3	3.3–12.9	57.2–109.8	0.09–0.8	0.04–0.4
PAHs	4.4–10.6	1.6–6.5	89.7–101.5	0.3–1.1	9.8–33.9	2.7–9.5	4.7–8.4	89.4–106.1	0.1–1.9	0.06–0.9
Hopanes	3.5–7.7	1.2–5.8	90.3–97.2	0.3–1.0	8.1–29.8	2.1–8.4	1.1–4.6	90.4–99.8	0.2–0.7	0.09–0.4

*For analyzing a 0.526 cm² (TD) and 4.34 cm² (SE) punch of filters collected with high-volume samplers (sampling at 1.1 m³ min⁻¹ for 24 h on 8" × 10" filters).

5

10




Figure 1. Comparison of GC-MS chromatograms from solvent extraction (SE, a) and thermal desorption (TD) at different injector-
port temperatures of 255 °C (b), 265 °C (c), 275 °C (d), 280 °C (e), 290 °C (f), and 300 °C (g).




Figure 2. The recoveries of n-alkanes (a) (C_n is the straight-chain n-alkanes which carbon number from 14–37), PAHs (b) and hopanes (c) by SE and TD/GC-MS. The full names of individual abbreviations are shown in SI.




Figure 3. Comparisons of n-alkanes (a), PAHs (b), and hopanes (c) determined by SE/GC-MS and TD/GC-MS methods for the ambient PM_{2.5} samples collected simultaneously from Beijing, Shanghai, Guangzhou, and Chengdu during winter of 2013.




Figure 4. Average concentration of n-alkanes (n-C₁₄–n-C₃₇) determined by SE/GC-MS and TD/GC-MS methods during wintertime from December 2013 to January 2014 in Beijing (a), Chengdu (b), Shanghai (c), and Guangzhou (d). Error bar is one standard deviation of individual n-alkanes.




Figure 5. Average concentration of individual PAHs determined by SE/GC-MS and TD/GC-MS methods during wintertime from December 2013 to January 2014 in Beijing (a), Chengdu (b), Shanghai (c), and Guangzhou (d). Error bar is one standard deviation of individual PAHs.




Figure 6. Average concentration of individual hopanes determined by SE/GC-MS and TD/GC-MS methods during wintertime from December 2013 to January 2014 in Beijing (a), Chengdu (b), Shanghai (c), and Guangzhou (d). Error bar is one standard deviation of individual hopanes.




Figure 7. Box plots of n-alkanes (a), PAHs (b), and hopanes (c) concentrations (ng m^{-3}) in Beijing, Guangzhou, Shanghai, and Chengdu during winter, 2013. The boxes represent the 25th percentile (lower edge), median (solid line), mean (cube mark), and 75th percentile (higher edge). The whiskers represent the minimum and maximum. The star mark represented one outlier for the samples collected in Guangzhou.

5