

Interactive comment on “Long-term reliability of the Figaro TGS 2600 solid-state methane sensor under low Arctic conditions at Toolik lake, Alaska”

by Werner Eugster et al.

Anonymous Referee #2

Received and published: 29 January 2020

This manuscript presents the results from a field deployment of a pair of low-cost metal-oxide sensors. The sensors were co-located with a reference instrument, allowing the researchers to train various calibration models to predict methane concentrations. These calibration models relied on the signals from the low-cost sensors as well as other sensors (i.e., temperature and humidity). Researchers then assessed the performance of and potential for these sensors using the predicted signals.

This manuscript is especially relevant to the field of low-cost sensor research and readers of Atmospheric Measurement Techniques for two reasons: (1) it provides an example of a long-term (multi-year) field deployment of low-cost metal-oxide sensors, and

[Printer-friendly version](#)

[Discussion paper](#)



Interactive
comment

(2) it provides an example of VOC sensors deployed to predict ambient methane levels - two areas that would benefit from further study. Furthermore, the deployment of the sensors in a remote area with little potential for the presence of confounding pollutants provides useful information on the potential ability of this sensor to be used for methane detection. Though a few revisions (listed below) are recommended prior to publication.

1. Please clarify throughout whether the results for the linear model being discussed in the text are based on the model that was fitted to the complete data set or the model which was fitted to the shorter training data set. Additionally, the training and testing periods defined for the linear model (in Table 1) and for the ANN (in the end of Section 2.4) appear to be different. Could the authors comment on the rationale for this choice and whether the use of these different periods might affect the comparability of the results for these two models presented in Table 1?
2. In Section 2.3, please provide information on any additional processing of the sensor data that may have occurred (e.g., filtering outliers, or removing sensor “warm-up” periods), or state that the data did not undergo additional filtering or processing.
3. Suggest moving the description of the motivation and development of the model for heat loss to an earlier point in the manuscript (e.g., after the description of the linear model in Section 3.0). This would assist the reader in their interpretation of the results in Table 1. Though the discussion of how this approach could be improved should remain in Section 3.5.
4. Could the authors provide additional information or discuss how the parameters of the model were selected (Eq. 2), for example, did this model yield substantial improvements over a simpler linear model?
5. Suggest expanding on the point made in Section 3.5 (Lines 253-254) to explain in what ways laboratory conditions over-simplify real-world conditions. This observation has been demonstrated in other studies [1, 2] and it could be valuable to highlight the challenges that may be associated with laboratory calibrations of sensors for this

[Printer-friendly version](#)[Discussion paper](#)

particular application.

6. Could the authors provide additional detail on the potential or likelihood for confounding pollutants, in particular carbon monoxide (Section 2.2)? For example, are there any towns nearby where emissions from wintertime heating may be a concern, or did any major wildfires occur in the area throughout the deployment period?
7. Is there any concern that the temperature/humidity sensor described in Section 2.2 might itself experience any issues with drift or aging over such a long field deployment?
8. Line 38: add an 's', "assessment of low-cost sensor[s]"
9. Line 66: delete 'e.g.,', "in an area like e.g., the arctic"
10. Line 246-247: change the color of the red text to black
11. Line 254: delete 'it', "as it would be required"

References

[1] Castell, N., Dauge, F., Schneider, P., Vogt, M., Lerner, U., Fishbain, B., . . . Bartonova, A. (2017). Can commercial low-cost sensor platforms contribute to air quality monitoring and exposure estimates? *Environment International*, 99, 293-302.

[2] Piedrahita, R., Xiang, Y., Masson, N., Ortega, J., Collier, A., Jiang, Y., . . . Shang, L. (2014). The next generation of low-cost personal air quality sensors for quantitative exposure monitoring. *Atmospheric Measurement Techniques Discussions*, 7(2), 2425-2457.

Interactive comment on *Atmos. Meas. Tech. Discuss.*, doi:10.5194/amt-2019-402, 2019.

[Printer-friendly version](#)

[Discussion paper](#)

