

## Supplementary Material

# Development of an Automatic Linear Calibration Method for High Resolution Single Particle Mass Spectrometry: Improved Chemical Species Identification for Atmospheric Aerosols

*Shengqiang Zhu<sup>1</sup>, Lei Li<sup>2</sup>, Shurong Wang<sup>1</sup>, Mei Li<sup>2</sup>, Yaxi Liu<sup>1</sup>, Xiaohui Lu<sup>1</sup>, Hong Chen<sup>1</sup>, Lin Wang<sup>1,3</sup>, Jianmin Chen<sup>1,3</sup>, Zhou Zhen<sup>2</sup>, Xin Yang<sup>\*1,3</sup> and Xiaofei Wang<sup>\*1,3</sup>*

[1] Shanghai Key Laboratory of Atmospheric Particle Pollution and Prevention, Department of Environmental Science and Engineering, Fudan University, Shanghai 200433, China

[2] Institute of Mass Spectrometer and Atmospheric Environment, Jinan University, Guangzhou, 510632, China

[3] Shanghai Institute of Pollution Control and Ecological Security, Shanghai 200092, China

## Atmospheric Measurement Techniques Discussions

\*To whom correspondence should be addressed.

Correspondence to:

Xiaofei Wang

Email: [xiaofeiwang@fudan.edu.cn](mailto:xiaofeiwang@fudan.edu.cn) Tel: +86-21-31242526

Xin Yang

Email: [yangxin@fudan.edu.cn](mailto:yangxin@fudan.edu.cn) Tel: +86-21-31245272



Fig.S1 the location of the Fengxian and Fudan university






Fig.S2 the explanation of the m/z bin value in the single particle mass spectra




(a) positive ions




(b) negative ions

Fig.S3 slope-intercept of the sea spray aerosol calibration



(a) positive ions



(b) negative ions

Fig.S4 slope-intercept of the ambient aerosol calibration




Figure.S5 Classification results of the LR-SPAMS by ART-2a




Figure.S6 Classification results of the HR-SPAMS by ART-2a

Table S1 deviation comparison between the averaged positive and negative mass spectra of calibration and non-calibration for sea spray aerosol

positive species	Theo (m/z)	Theo-bin(m/z)	Noncal-bin(m/z)	$\Delta m/z$	Cal-bin(m/z)	$\Delta m/z$	negative species	Theo (m/z)	Theo-bin(m/z)	Noncal-bin(m/z)	$\Delta m/z$	Cal-bin(m/z)	$\Delta m/z$
Na <sup>+</sup>	22.98977	22.993	23.0167	0.0237	22.993	0	CN <sup>-</sup>	26.00307	25.9994	25.9658	-	26.0078	0.0084
Mg <sup>+</sup>	23.98505	23.9829	24.0072	0.0243	23.9829	0	Cl <sup>-</sup>	34.96885	34.9641	34.9349	-	34.9641	0
K <sup>+</sup>	38.96371	38.9672	38.998	0.0288	38.9672	0	Cl <sup>-</sup>	36.9659	36.9622	36.9321	-	36.9622	0
Na <sub>2</sub> Cl <sup>+</sup>	80.94839	80.9438	80.9883	0.0445	80.9438	0	CNO <sup>-</sup>	41.99799	41.9971	41.965	-	41.9864	0.0107
Na <sub>2</sub> Cl <sup>+</sup>	82.94544	82.9421	82.9871	0.045	82.9421	0	NaCl <sup>-</sup>	57.95862	57.9574	57.9448	-	57.9574	0
K <sub>2</sub> Cl <sup>+</sup>	112.89627	112.898	112.951	0.053	112.898	0	MgCl <sub>3</sub> <sup>-</sup>	128.89161	128.883	128.939	0.056	128.901	0.018
K <sub>2</sub> Cl <sup>+</sup>	114.89332	114.885	114.938	0.053	114.885	0	MgCl <sub>3</sub> <sup>-</sup>	130.88866	130.894	130.932	0.038	130.894	0
							MgCl <sub>3</sub> <sup>-</sup>	132.88571	132.884	132.941	0.057	132.884	0

Table S2 deviation comparison between the averaged positive and negative mass spectra of calibration and non-calibration for ambient aerosol

positive species	Theo (m/z)	Theo-bin(m/z)	Noncal-bin(m/z)	$\Delta m/z$	Cal-bin(m/z)	$\Delta m/z$	negative species	Theo (m/z)	Theo-bin(m/z)	Noncal-bin(m/z)	$\Delta m/z$	Cal-bin(m/z)	$\Delta m/z$
C <sup>+</sup>	12	11.9991	12.0105	0.0114	12.0048	0.0057					-		
Na <sup>+</sup>	22.98977	22.993	23.0167	0.0237	22.993	0	CN <sup>-</sup>	26.00307	25.9994	25.9658	0.0336	26.0078	0.0084
							Cl <sup>-</sup>	34.96885	34.9641	34.9349	-	34.9641	0
											0.0292		
C <sub>3</sub> <sup>+</sup>	36	36.0023	36.032	0.0297	35.9925	-	NO <sub>2</sub> <sup>-</sup>	45.99291	45.9897	45.9673	-	45.9897	0
						0.0098					0.0224		
K <sup>+</sup>	38.9672	38.9672	38.998	0.0308	38.9672	0	NO <sub>3</sub> <sup>+</sup>	61.98783	61.9938	61.9678	-0.026	61.9808	-0.013
Fe <sup>+</sup>	55.93494	55.932	55.9813	0.0493	55.9443	0.0123	SO <sub>4</sub> <sup>-</sup>	95.95175	95.9502	96.031	0.0808	95.9663	0.0161
Pb <sup>+</sup>	207.97664	207.967	208.014	0.047	207.967	0	HSO <sub>4</sub> <sup>-</sup>	96.95958	96.9546	96.9709	0.0163	96.9546	0
Pb <sup>+</sup>	205.97446	205.976	206.023	0.047	205.976	0							
Pb <sup>+</sup>	206.97589	206.97	207.018	0.048	206.994	0.024							

Table S3 deviation comparison between the averaged positive and negative mass spectra of calibration and non-calibration for ambient aerosol with additional marker ions

positive species	Theo (m/z)	Theo-bin(m/z)	Noncal-bin(m/z)	$\Delta m/z$	Cal-bin(m/z)	$\Delta m/z$	negative species	Theo (m/z)	Theo-bin(m/z)	Noncal-bin(m/z)	$\Delta m/z$	Cal-bin(m/z)	$\Delta m/z$
C <sup>+</sup>	12	11.9991	12.0105	0.0114	11.9877	-	CN <sup>-</sup>	26.00307	25.9994	25.9658	-	26.0078	0.0084
Na <sup>+</sup>	22.98977	22.993	23.0167	0.0237	22.993	0	Cl <sup>-</sup>	34.96885	34.9641	34.9349	-	34.9641	0
C <sub>3</sub> <sup>+</sup>	36	36.0023	36.032	0.0207	36.0023	0	NO <sub>2</sub> <sup>-</sup>	45.99291	45.9897	45.9673	-	45.9897	0
K <sup>+</sup>	38.9672	38.9672	38.998	0.0308	38.9672	0	NO <sub>3</sub> <sup>+</sup>	61.98783	61.9938	61.9678	-0.026	61.9808	-0.013
Fe <sup>+</sup>	55.93494	55.932	55.9813	0.0493	55.9566	0.0246	SO <sub>4</sub> <sup>-</sup>	95.95175	95.9502	96.0148	0.0646	95.9663	0.0161
Pb <sup>+</sup>	207.97664	207.967	208.038	0.071	207.967	0	HSO <sub>4</sub> <sup>-</sup>	96.95958	96.9546	96.9709	0.0163	96.9546	0
Pb <sup>+</sup>	205.97446	205.976	206.047	0.071	205.976	0							
Pb <sup>+</sup>	206.97589	206.97	207.041	0.071	206.994	0.024							

Notes: Theo(m/z): Theoretical m/z value; Theo-bin(m/z): Theoretical m/z bin value;  
 Noncal-bin(m/z): Non-calibration m/z bin value. Cal bin(m/z): Calibration m/z bin value.

