# Letter to Editor

Dear Dr. Stutz,

We thank you very much for your consideration of our article titled: "Effect of Polyoxymethylene (POM-H Delrin) offgassing within Pandora head sensor on direct sun and multi-axis formaldehyde column measurements in 2016 - 2019" for pub-

5 lication in AMT. We found the reviewers' comments and suggestions very valuable and helpful in improving our manuscript. We have made revisions according to the reviewers' comments and suggestions, as described below. Reviewers' comments are given in regular font, authors responses are in italics font. We are also attaching a marked up version showing the changes.

# Best Regards, Elena Spinei and co-authors.

10

15

#### **Responses to comments by Anonymous Referee 1**

#### 0.0.1 General comments:

I would recommend adding one or two tables summarizing 1) which instruments have been used when/for which part of this work, to help the reader following which instruments have been participating to several steps;

We added table 1 describing Pandoras used in this study with their appropriate modifications and part of this study where they have been used.

Ν	Owner	Manufactured	<b>Relevant Hardware Components</b>	Contribution to This Study
2	NASA	2011	upgrade in summer 2019: Nylon parts,	Temperature (Section 3); Field study
			temperature sensor, wedged window	(direct sun, Section 5.1)
21	NASA	2011	upgrade in 2016: ARC window;	Laboratory tests of HCHO emissions
			POM-H Delrin parts	(Section 4)
32	NASA	2016	ARC window; POM-H Delrin parts	Field study (direct sun, Section 5.1)
46	NASA	2015	upgrade in 2016 ARC window;	Laboratory tests of HCHO emissions
			POM-H Delrin parts	(Section 4)
108	ECCC	2016	ARC window; POM-H Delrin parts	Field measurements (Section 2)
118	KNMI	2016	ARC window; POM-H Delrin parts	Laboratory tests of HCHO emissions
				(Section 4)
148	Virginia Tech	2018	temperature sensor (April 2019), wedged window; POM-H Delrin parts	Temperature (Section 3); Laboratory
				tests of HCHO emissions (Section 4);
				Field study (MAX-DOAS, Section 5.2)
155	Boston University	2019	temperature sensor; wedged window;	Temperature (Section 3)
			POM-H Delrin parts	Temperature (Section 5)
165	EPA	summer 2019	Nylon parts; temperature sensor;	Laboratory tests of HCHO emissions
			wedged window	(Section 4)
167	EPA	summer 2019	Nylon parts; temperature sensor;	Laboratory tests of HCHO emissions
			wedged window	(Section 4)
168	EPA	summer 2019	Nylon parts; temperature sensor;	Laboratory tests of HCHO emissions
			wedged window	(Section 4)

#### Table 1. Pandora instruments used in this study

ECCC: Environment and Climate Change Canada; KNMI: Royal Netherlands Meteorological Institute; EPA: US Environmental Protection Agency; NASA: US National Aeronautics and Space Administration

2) the different steps of the Pandora HCHO "history" with some links to published HCHO Pandora papers: - pre2016: no HCHO possible; - spring 2016: window replacement and KORUS-AQ and CINDI-2 but issue of coherence between MAX-DOAS and direct-sun results; - april 2019: added temperature sensors for a few instruments and highlight of the potential link

of HCHO creation with temperature; - lab measurements to confirm the findings; - summer 2019: new head sensor without 20 POM-H Delrin to solve the problem; - measurements campaigns illustrating the findings.

We added table 2 describing Pandora design modifications including references

Period	Hardware components	Impact on HCHO	HCHO Data Used
2007 - winter 2016	parallel window, POM-H Delrin	window caused etalloning in direct sun measurements, HCHO emissions from POM-H Delrin - <b>direct sun HCHO is not correctable</b>	MAX-DOAS: Pinardi et al., 2013; Direct Sun: Park et. al., 2018
spring 2016 - 2017	anti reflective coating on parallel widow, POM-H Delrin	ARC degrades within 1 year of installation, temperature dependent HCHO internal emission from POM-H Delrin (disagreement between direct sun total column and MAX-DOAS tropospheric column), can be corrected for functioning ARC	MAX-DOAS: Kreher et al., 2020; Direct Sun: Spinei et al., 2018, Herman et al., 2018; Spinei et al., 2020
2018 - spring 2019	wedged window*, POM-H Delrin	temperature dependent HCHO internal emission from POM-H Delrin (disagreement between direct sun total column and MAX-DOAS tropospheric column), can be corrected	MAX-DOAS: Nowak et al. (2020)
summer 2019-	wedged widow, nylon	believed not to have any interference caused by design (confirmed by extensive laboratory studies)	

 Table 2. History of Pandora hardware changes related to direct sun HCHO measurements

Note: HCHO from direct sun is not a standard PGN product and was not provided by the NASA and Luftblick PGN groups outside of KORUS-AQ study (Spinei et al., 2018, Herman et al., 2018). Park et al., 2018 performed HCHO analysis independently and were not aware of any PGN discoveries.

\* wedged windows are installed on new instruments, if the instruments are not returned to NASA or SciGlob - they are not upgraded, therefore some instruments probably still have degrading ARC windows

In the light of the "pre-2016" first point, there should also be some comments of Pandora HCHO measurements performed before 2016 (e.g., in Pinardi et al. 2013, for measurements in 2009 – is there any other ?).

Table 2 contains periods of operation and the publications used data from that period to the best of our knowledge

I would also include some more references in the introduction on satellite HCHO (only one is mentioned, while at least 8 are known - and given below) and on satellite validation.

# We added the recommended references

In the abstract and/or conclusions, it would also be nice to also convert the findings from DU to molec/cm2, as this unit is usually used for HCHO retrievals.

30 We added "2.69 molecules/ $cm^2$ " to the abstract. There is a definition of DU in text p.2, line 42

# 0.0.2 Specific comments and technical corrections:

- P2, line 29-33: give some references for the other satellites, as done for SCIAMACHY. Several are suggested below.

We added the recommended references

- P2, line 37: give more references for other validation studies. Several are suggested below.

We added the recommended references

35

P2, last line: Lamsal et al. 2014 and Kollonige et al. 2017 are not given in the reference list. Consider adding other recent works with several Pandora NO2 measurements used for validation (Herman et al., 2019; Pinardi et al., 2020; Verhoelst et al., 2020).

# - We added the recommended references

40 Please check completeness of the reference list in the whole manuscript (e.g., Reed et al., 2015, Gronoff et al., 2019 are also missing in the references).

We have checked the references

- P3, line 65 to 71: it would be nice to link the different steps, with the manuscript sections. Or at least add the link with the sections in the proposed additional table n2.

# 45 We added reference to the corresponding sections

- P3, line 74: Retrieval of weak absorbers such as HCHO was not possible from 'the pre-2016 Pandora direct sun measurements due to the telescope assembly front window etaloning'': comment Pinardi et al., 2013 measurements of 2009.

Pinardi et al. 2013 reference MAX-DOAS HCHO only, no HCHO direct sun data from Pandora are presented. Only coherent light (e.g direct solar beam) causes etalloning by the window.

50 - P4, figure 1: please revisit the legend. What are the light blue points?

Both blue and light blue points represent HCHO emissions from the same instrument P46 but at different temperature ramp rates. We believe this is appropriate.

- P4, line 95: . . . the reference spectrum is \*often\* taken within. . . (not for the CINDI-2, as mentioned later in the paper)

MAX-DOAS data analysis for profiles is done with the scan reference. CINDI-2 MAX-DOAS data for profiles was also analysed with the scan reference (Tirpitz et al., 2020). Local noon reference was used for  $\Delta S$  intercomparison not the final inversion. Zenith  $\Delta S$  were subtracted from the  $\Delta S$  at all other elevation angles

- P5, line 108: "A baffle holding tube, two filter wheels, and a dark filter" -> The baffle holding tube, the two filter wheels, and the dark filter. . .

corrected

60 - P7, line 149: remove parenthesis in "Figure (4). . ."

corrected

- P7, line 153: consider changing "follow-up work" to the relevant section of the manuscript.

We removed this sentence

- P7, line 160: add a comma before 118

65 *corrected* 

55

- P7, lines 147 to 148 and lines 160 to 161: as proposed in the general comments, consider adding a table with the Pandora numbers, their construction year, and in what part of this work they appear.

Table 1 was added

- P8, line 161: "Three other Pandora" -> The other three Pandora

# 70 *corrected*

- P8, line 169: consider adding here the information on the 1000 W FED lamp given in line 186 and the LED lamp (cf line 2018 and 2016), or removing this information from line 169.

# We removed this sentence

- P8, line 170: what is the meaning/purpose of the "The head sensor was not disturbed during the measurements" ?

75

We intended to emphasize that the instrument position was not changed. We removed it.

- P8, line 175 to 179: these DOAS settings seems the generic ones used in the laboratory study (332-360nm), but actually in P10 line 250, these are different (300-350nm) (and also different than the in-field settings, P13 line 396, 332-359nm), so this is a bit perturbing. Is there any reason why changing wavelength ranges, polynomial and offset order? For the inclusion of trace gases, this is clear/mentioned when relevant.

80 We performed identical tests for all pandora instruments. The new pandoras had no detectable HCHO absorption in multiple fitting windows. Since the results were not impacted by the window choice one single fitting window was chosen for both SO2 and HCHO retrieval. We also wanted to emphasise the low fitting residual for 50 nm fitting window.

P9, line 193: "using the pre-installed Bosch BME280 digital humidity, pressure and temperature sensor on Spark-Fun Atmospheric Sensor Breakout Board" -> this is a detail of the new head sensor. It would be maybe better to introduce it in Sect
 2.3 when introducing the Pandora internal head sensor temperature?

We added the information about the internal temperature sensor to Section 2.3

- P9, lines 208 to 210: the Pandora 118 was also tested with a LED lamp in addition to the FEL lamp. Is this shown somewhere? Why is this done? What was expected (differently) from this additional test?

Some literature suggested that UV and VIS radiation (200-800 nm) was responsible for Photo-oxidation at those wavelengths. We were curious whether more HCHO will be emitted while using FEL vs LED light source. We did not see any statistically meaningful difference between the retrievals.

- P10 and P11 (figure 6): POM-free and Delrin-free. These are used as synonyms? In P13, line 281, also POM-H is used. Use one name everywhere.

We corrected all to POM-H Delrin free

95 - P12, line 268 and elsewhere in the manuscript: replace Kreher et al., 2019 by Kreher et al., 2020.

corrected

- P12, line 269: there are two point at the end of the line.

corrected

- P14, line 314: "contribution" seems to be in bold
- 100 *corrected* 
  - P17, line 348: add reference to Fig 5 for the "small generation rates inside Pandora 148 head sensor (a = 0.0041 DU)"?

corrected

- P17, lines 53: give an estimation and a reference (Pinardi et al., 2013 ?) of the "fitting noise for most DOAS instruments" (usually in molec/cm2!).

105 We added:  $(0.3 \text{ DU} = 8 \times 10^{15} \text{ molecules/cm}^2$ , Table 7 in Kreher et al., 2020.

- P17, line 357: add "Sept. 2016" after the coordinates and update Kreher reference.

added

- P18, line 378: if I followed well all the Pandora numbers, 32 was involved in the GSFC field campaign, and 31 and 46 in the lab measurements wrt temperature. A table would definitely help follow which instrument has been used when!

110 Table 1 was added

- P20: suggestion to add in each bullet conclusions the Pandora numbers relevant to each bullet (or not if it is already clear in an additional table), and the relevant Figure supporting each conclusion (e.g., Fig 5 for the 1rst point). Also add molec/cm2 estimation in addition to DU values.

Since Table 1 was added we have not modified the Pandora listing in conclusions. We added Figure references to the conclusions.

- P20, point 5: maybe cite publications that were made with Pandora direct sun HCHO data between 2016 and 2019, and that should not be considered "valid"?

We added references to those publications in Table 2. We replaced "build" with "operational" to emphasise that any other Pandora data whether part of a field campaign or routine measurements cannot be used in the current form

120

# **Response to comments by Anonymous Referee 2**

# 0.0.3 General comments:

The manuscript provides a detailed timelines and descriptions for multiple events such as Pandora sensor head design changes; laboratory testing for POM and non-POM units; and case-studies for real-life deployments/colocations. I found myself flipping back and fourth through the manuscript, I therefore think that the manuscript will be improved by inclusion of a master table summarizing the types of tests, dates, intercomparison campaigns, identification numbers of units, modifications to the units,

125

130

etc. This table can be placed at the end of Section 1.

We have added Table 1 (see the attached manuscript) that summarizes the Pandora instrument information and how they contributed to this study. We also added the following sentence at the end of Section 1: "Table 1 lists the Pandora instruments description and contribution to this study.". We did not add the actual field campaigns that each pandora participated in to make sure we do not give a wrong impression that only these instruments are impacted.

Can authors develop a correction factor that can be applied for direct sun HCHO data collected during 2016-2019 to correct for HCHO production, so the dataset can be utilized by the scientific community? For example, recommending temperature ranges during which data would be usable, and showing examples of intercomparison with in-situ techniques (if available) or with satellite data showing a reasonable agreement.

We have performed detailed evaluation of Pandora head sensor heat transfer and estimation of HCHO production amplitude. We are preparing a separate manuscript: Spinei et al. 2020 (in preparation). We added the following reference to point 5 of the conclusions: " Considering that Pandora head sensors have almost identical design from material, shape and thermodynamics point of view data between 2016 and 2019 can be corrected based on (a) meteorological data (temperature and wind) to estimate internal head sensor temperature and (b) on ΔS measurements to estimate 140 HCHO production amplitude (Spinei et al. 2020 in preparation)"

# 0.0.4 Specific comments:

Lines 10-14: define cold and warm temperature ranges. Remove quotation marks from "cold"

We added the temperature information: Measurements in winter, during colder (<10°C) days in general and at high solar zenith angles (> 75 °) were minimally impacted. Measurements during hot days (>28°C)

145 Figure 1: define light blue, gray and green lines

We added the following information to the figure caption: (green: box with a median mixing layer height (MLH), grey: box with a measured MLH; light blue: box+exponential profile with a median MLH, and black: box+exponential profile with a measured MLH,

Lines 91-98: add explanation for which spectra are used for direct sun and multi-axis DOAS retrievals.

150 We have made the following modification: DOAS implementation of multi-axis retrieval is significantly less sensitive to instrumental changes. This is due to the fact that single scan sky scattered solar spectra are analysed using a zenith reference spectrum taken within maximum 10-15 minutes from the scan measurements. Direct sun spectra, on the other hand, are analyzed using a single reference spectrum that was potentially taken months apart from the rest of the spectra.

Figure 2: add ambient temperature to the figure

155 added

Line 160: add a coma before 118

added

Line 161: add a coma before and 148

added

160 Line 171: remove quotation marks from, "mimicking"

removed

Line 314: remove bold face from contribution

removed

Lines 404-405: The statement "Pandora HCHO measurements derived from direct sun observations between 2016 and 2019
 165 cannot be used in the current form. Results presented here most likely are representative of other instruments build between 2016 and 2019" is very drastic. Authors should consider adding recommendations on possible corrective approaches, so the data could be utilized by scientific community.

We are preparing a publication to describe such corrections. We added: "Considering that Pandora head sensors have almost identical design from material, shape and thermodynamics point of view measurements between 2016 and 2019
 170 can be corrected based on (a) meteorological observations (temperature and wind) to estimate internal head sensor temperature and (b) on ΔS measurements to estimate HCHO production amplitude (Spinei et al. 2020 in preparation)

# Effect of Polyoxymethylene (POM-H Delrin) offgassing within Pandora head sensor on direct sun and multi-axis formaldehyde column measurements in 2016 - 2019

Elena Spinei<sup>1</sup>, Martin Tiefengraber<sup>2,3</sup>, Moritz Müller<sup>2,3</sup>, Manuel Gebetsberger<sup>2</sup>, Alexander Cede<sup>2</sup>, Luke Valin<sup>4</sup>, James Szykman<sup>4</sup>, Andrew Whitehill<sup>4</sup>, Alexander Kostakis<sup>5</sup>, Fernando Santos<sup>6</sup>, Nader Abbuhasan<sup>7</sup>, Xiaoyi Zhao<sup>8</sup>, Vitali Fioletov<sup>8</sup>, Sum Chi Lee<sup>8</sup>, and Robert Swap<sup>9</sup>

<sup>1</sup>Center for Space Science And Engineering Research, Virginia Polytechnic Institute and State University, Blacksburg, VA, USA

<sup>2</sup>LuftBlick, Innsbruck, Austria

<sup>3</sup>Department of Atmospheric and Cryospheric Sciences, University of Innsbruck, Innsbruck, Austria

<sup>4</sup>United States Environmental Protection Agency, Durham, NC, USA

<sup>5</sup>Universities Space Research Association, Columbia, MD, USA

<sup>6</sup>Earth System Science Interdisciplinary Center, University of Maryland, College Park, MD, USA

<sup>7</sup>Joint Center for Earth Systems Technology, University of Maryland, Baltimore County, Baltimore, USA

<sup>8</sup>Air Quality Research Division, Environment and Climate Change Canada, Toronto, M3H 5T4, Canada

<sup>9</sup>NASA Goddard Space Flight Center, Greenbelt, MD, USA

Correspondence: Elena Spinei: eslind@vt.edu

# Abstract.

Analysis of formaldehyde measurements by the Pandora spectrometer systems between 2016 and 2019 suggested that there was a temperature dependent process inside Pandora head sensor that emitted formaldehyde. Some parts in the head sensor were manufactured from thermal plastic polyoxymethylene homopolimer (E.I. Du Pont de Nemour & Co., USA: POM-H Delrin®)

5 and were responsible for formaldehyde production. Laboratory analysis of the four Pandora head sensors showed that internal formaldehyde production had exponential temperature dependence with a damping coefficient of  $0.0911 \pm 0.0024$  °C<sup>-1</sup> and the exponential function amplitude ranging from 0.0041 DU to 0.049 DU. No apparent dependency on the head sensor age and heating/cooling rates was detected.

The total amount of formaldehyde internally generated by the POM-H Delrin components and contributing to the direct sun 10 measurements were estimated based on the head sensor temperature and solar zenith angle of the measurements. Measurements in winter, during colder (<10°C) days in general and at high solar zenith angles (> 75 °) were minimally impacted. Measurements during hot days (>28°C) and small solar zenith angles had up to 1 DU (2.69 molecules/cm<sup>2</sup>) contribution from POM-H Delrin parts. Multi-axis differential slant column densities were minimally impacted (< 0.01 DU) due to the reference spectrum collected within a short time period with a small difference in head sensor temperature.

15 Three new POM-H Delrin free Pandora head sensors (manufactured in summer 2019) were evaluated for temperature dependent attenuation across the entire spectral range (300 to 530 nm). No formaldehyde or any other absorption above the instrumental noise was observed across the entire spectral range.

# 1 Introduction

- 20 The importance of formaldehyde (HCHO) in tropospheric chemistry arises from its participation in radical formation and recycling, including HO<sub>x</sub> (HO + HO<sub>2</sub>) (Liu et al., 2007; Alicke, 2002). The HO<sub>x</sub> budget, in its tern, determines the oxidation capacity of the atmosphere and formation of photochemical smog (O<sub>3</sub>) in the lower troposphere in the presence of NO<sub>x</sub> and sunlight. Since the major path of HCHO into the atmosphere is through oxidation of non-methane volatile organic compounds (NMVOC) and daylight removal is mostly through the photolysis and oxidation by HO, daytime HCHO abundances above background levels are mainly indicative of local emissions and the local oxidizing capacity of the atmosphere.
- The most efficient way to monitor geographical and temporal variability of HCHO on a global scale is from satellite platforms. Satellite observations of HCHO from sun synchronous polar orbits started with GOME in 1995. Since then several instruments provided global coverage with various spatial and temporal resolution (Zhu et al. (2016a), Wang et al. (2017): SCIAMACHY (10:00 hr LT, 32x215 km<sup>2</sup>, 2002 - 2012, Wittrock et al., 2006, De Smedt et al. (2008)); GOME-2A and -2B (09:30 hr LT, 40x40 km<sup>2</sup>, since 2007 to the date of publication Pinardi et al. (2020), De Smedt et al. (2015), De Smedt et al.
- 30 (09:30 hr LT, 40x40 km<sup>2</sup>, since 2007 to the date of publication Pinardi et al. (2020), De Smedt et al. (2015), De Smedt et al. (2012), Hewson et al. (2013)), OMI (13:30 hr LT, 13x24 km<sup>2</sup>, since 2004 to the date of publication, Herman et al. (2018), Pinardi et al. (2020), De Smedt et al. (2015)); TROPOMI (13:30 hr LT, 3.5x7 km<sup>2</sup>, since 2018 to the date of publication, Verhoelst et al. (2020), De Smedt et al. (2018), Vigouroux et al. (2018)). Next generation air quality instruments, positioned in geostationary orbit, will provide unprecedented temporal coverage over Asia (GEMS, since February 18, 2020, Kwon et al.
- (2019); North America (TEMPO, estimated launch in early 2022); Europe (Sentinel-4, estimated launch in 2023). Current and future satellite HCHO observations require routine and systematic validation through the use of independent measurements to assess biases and uncertainties and encourage full utilization of satellite data to support both science and applications. Validation of satellite HCHO products, however, is challenging due to spatial and temporal sampling differences among the satellite, ground-based (e.g. FTIR (Vigouroux et al., 2009), DOAS (Chan et al., 2020)) and airborne platforms (Zhu
- 40 et al., 2016b). Ideally, it involves data from ground-based networks of identical instrumentation with continuous measurements and uniform data analysis and wide global distribution. To meet current and future satellite validation needs (e.g. TEMPO, TROPOMI) ground-based HCHO column measurements should have an accuracy better than 0.1 DU (1 DU =  $2.69 \cdot 10^{16}$ molecules/cm<sup>2</sup>, expected nominal TEMPO precision over 3 hr is  $1.95 \cdot 10^{15}$  molecules/cm<sup>2</sup>).

Pandonia Global Network (PGN) is a NASA and ESA sponsored ground-based network of standardized and homogeneously
calibrated Pandora instruments focused on air quality and atmospheric composition measurements. The main objective of PGN is to provide systematic data processing and data dissemination to the greater global community in support of in situ and remotely sensed AQ monitoring (Szykman et al., 2019). One of the PGN's major objectives is the validation of satellite-based UV-visible sensors, specifically, Sentinel 5P, TEMPO, GEMS and Sentinel 4. PGN is focused on providing measurements of the total column and vertically resolved concentrations of a range of trace gases (e.g., NO<sub>2</sub>, O<sub>3</sub>, HCHO, SO<sub>2</sub>). Pandora total NO<sub>2</sub>

50 column measurements have been extensively used for OMI validation (Herman et al. (2009); Pinardi et al. (2020); Herman

et al. (2019); Verhoelst et al. (2020)) and atmospheric composition studies during multi-agency supported field campaigns such as DISCOVER-AQ (Reed et al., 2015), KORUS-AQ (Spinei et al., 2018), OWLETS (Gronoff et al., 2019)), and LISTOS (https://www-air.larc.nasa.gov/missions/listos).

- The Pandora spectrometer system, deployed within PGN, is a cost-effective ground-based instrument, operating on the principle of the passive UV-visible differential optical absorption spectroscopy technique (DOAS). Pandoras undergo extensive laboratory characterization and have a robust data acquisition and analysis software package, Blick Software Suite (Cede, 2019). Pandora instruments are fully automated and fully programmable to perform all types of DOAS observation geometries (e.g., direct sun, multi-axis, and target) from sunrise to sunset and overnight for moon measurements. Pandoras have no consumables and are designed for unattended operation in outdoor environments. Measured spectra are automatically collected
- 60 and submitted to the PGN server via an Internet connection for centralized uniform real-time processing by the Blick Software Suite. Pandora instrument consists of a small Avantes low stray light spectrometer (280 - 530 nm with 0.6 nm full width at half maximum spectral resolution) connected to a telescope assembly by a 400-micron core diameter single strand fiber optic cable. The telescope assembly (head sensor) is attached to a 2-axis positioner (SciGlob), capable of accurate pointing ( $\pm 0.1^{\circ}$ ). A diffuser is included in the optical path to minimize the effect of small pointing errors for direct sun measurements with a
- 65 2.5° full width half maxima (FWHM) field of view (FOV). Pandoras measure scattered solar photons without the diffuser with 1.5° FWHM FOV.

Here we present (a) the discovery of the Pandora instrument artifact due to POM-H Delrin plastic offgassing impacting the Pandora HCHO measurements up to 2019 (Section 2); (b) a laboratory and field characterization of the interference on direct sun and multi-axis results (Sections 3, 4, 5); (c) characterization of the interference following an engineered solution (Sections

70 3, 4). Results presented in this study show that plastic related HCHO offgasing significantly impacted direct sun total columns and minimally impacted multi-axis retrievals (Section 5). Due to strong temperature dependence of HCHO offgasing the largest interference was observed in summer. Extensive analysis of the Pandora instruments, after the engineering solution was implemented, shows no interference and strengthens confidence in future direct sun measurements. Table 1 lists the Pandora instrument description and contribution to this study.

# 75 2 Pandora HCHO measurements

Pandora instruments were first field deployed in 2006 with the main focus on direct sun  $O_3$  and  $NO_2$  total columns measurements. Retrieval of weak absorbers such as HCHO was not possible from the pre-2016 Pandora direct sun measurements due to the telescope assembly front window etaloning. Window introduced interference was larger than the background HCHO absorption level and is not correctable in the pre-2016 measurements (Park et al., 2018).

80 Pandora instrumentation has undergone several design changes that significantly improved HCHO direct sun measurements (e.g. 64 row CCD, new tracker). In spring 2016, the telescope assembly front window was replaced with a window containing an anti-reflective coating (ARC). This reduced the etaloning interference and improved the ability to retrieve formaldehyde

#### Table 1. Pandora instruments used in this study

N	Owner	Manufactured	<b>Relevant Hardware Components</b>	<b>Contribution to This Study</b>
2	NASA	2011	upgrade in summer 2019: Nylon parts,	Temperature (Section 3); Field study
			temperature sensor, wedged window	(direct sun, Section 5.1)
21	NASA	2011	upgrade in 2016: ARC window;	Laboratory tests of HCHO emissions
			POM-H Delrin parts	(Section 4)
32	NASA	2016	ARC window; POM-H Delrin parts	Field study (direct sun, Section 5.1)
46	NASA	2015	upgrade in 2016 ARC window;	Laboratory tests of HCHO emissions
			POM-H Delrin parts	(Section 4)
108	ECCC	2016	ARC window; POM-H Delrin parts	Field measurements (Section 2)
118	KNMI	2016	ARC window; POM-H Delrin parts	Laboratory tests of HCHO emissions
				(Section 4)
148	Virginia Tech	2018	temperature sensor (April 2019), wedged window; POM-H Delrin parts	Temperature (Section 3); Laboratory
				tests of HCHO emissions (Section 4);
				Field study (MAX-DOAS, Section 5.2)
155	Boston University	2019	temperature sensor; wedged window;	Temperature (Section 3)
			POM-H Delrin parts	
165	EPA	summer 2019	Nylon parts; temperature sensor;	Laboratory tests of HCHO emissions
			wedged window	(Section 4)
167	EPA	summer 2019	Nylon parts; temperature sensor;	Laboratory tests of HCHO emissions
			wedged window	(Section 4)
168	EPA	summer 2019	Nylon parts; temperature sensor;	Laboratory tests of HCHO emissions
			wedged window	(Section 4)

ECCC: Environment and Climate Change Canada; KNMI: Royal Netherlands Meteorological Institute; EPA: US Environmental Protection Agency; NASA: US National Aeronautics and Space Administration

columns from the Pandora direct sun measurements. Due to ARC degradation, the front window was again replaced in 2018 with a wedged window which practically removed the etaloning interference.

- The May-June 2016 Korea–United States Air Quality Study (KORUS-AQ) offered the first opportunity to evaluate direct-sun observations of HCHO total column densities with the improved Pandoras (ARC window and 64 row CCDs, Spinei et al., 2018; Herman et al., 2018). Comparison between the HCHO total columns derived from the Pandora direct sun measurements and the integrated in situ aircraft measurements by Spinei et al., 2018 (Fig. 1) showed that Pandoras overestimated the aircraft derived columns by 16% on average, with an offset of 0.22 DU. However, a point-to-point comparison shows that the measurements
- 90 agreed on a cold and breezy day (4-May-2016) and on most mornings. Pandoras measured up to 0.8 DU larger columns than DC-8 on hot days during early afternoon hours (12 to 16 hours local time, 1). Measured surface concentrations scaled-up to

#### Table 2. History of Pandora hardware changes related to direct sun HCHO measurements

Daniad	Hardware	Impact on HCHO	HCHO Data Usad
renou	components	impact on HCHO	ICHO Data Oscu
2007 - winter 2016	parallel window, POM-H Delrin	window caused etalloning in direct sun measurements, HCHO emissions from POM-H Delrin - direct sun HCHO is not correctable	MAX-DOAS: Pinardi et al. (2013); Direct Sun: Park et al. (2018)
spring 2016 - 2017	anti reflective coating on parallel widow, POM-H Delrin	ARC degrades within 1 year of installation, temperature dependent HCHO internal emission from POM-H Delrin (disagreement between direct sun total column and MAX-DOAS tropospheric column), can be corrected for functioning ARC	MAX-DOAS: Kreher et al. (2020); Direct Sun: Spinei et al. (2018); Herman et al. (2018); Spinei et al. (2020)
2018 - spring 2019	wedged window*, POM-H Delrin	temperature dependent HCHO internal emission from POM-H Delrin (disagreement between direct sun total column and MAX-DOAS tropospheric column), can be corrected	MAX-DOAS: Nowak et al. (2020)
summer 2019-	wedged widow, nylon	believed not to have any interference caused by design (confirmed by extensive laboratory studies)	

Note: HCHO from direct sun is not a standard PGN product and was not provided by the NASA and Luftblick PGN groups outside of KORUS-AQ study (Spinei et al. (2018); Herman et al. (2018)). Park et al. (2018) performed HCHO analysis independently and were not aware of any PGN discoveries.

\* wedged windows are installed on new instruments, if the instruments are not returned to NASA or SciGlob - they are not upgraded, therefore some instruments probably still have degrading ARC windows

the total columns, assuming different profiles (black, grey and green lines on 1), and mixing layer height from Ceilometer data agreed with DC-8 measurements better than with Pandora measurements.

- Pandora direct sun HCHO total columns were also larger than the multi-axis measured columns during short-term field campaigns (e.g., CINDI-2, LISTOS 2018) and during summer versus winter comparisons for long-term routine observations. Since
  HCHO is mostly located in the lower troposphere, multi-axis and direct sun measurements should result in HCHO columns that closely match (assuming sampling of the same air masses). DOAS implementation of multi-axis retrieval is significantly less sensitive to instrumental changes. This is due to the fact that single scan sky scattered solar spectra are analyzed using a zenith reference spectrum taken within maximum 10-15 minutes from the scan measurements. Direct sun spectra, on the other
- 100 hand, are analyzed using a single reference spectrum that was potentially taken months apart from the rest of the spectra. Figure 2 shows an example of HCHO columns derived from Pandora 108 direct sun and multi-axis measurements in Egbert, Canada (44.23°N, -79.78°W) from May 2018 to March 2019. Significant differences (up to 1.5 DU) were observed in retrievals by direct sun measurements in summer during hot days versus the multi-axis measurements.

Comparisons of multiple Pandora data sets covering a variety of ambient conditions led us to conclude that there was an 105 intrinsic property of Pandora that interfered with its HCHO measurements at higher ambient temperatures. The most likely



**Figure 1.** HCHO vertical columns during KORUS-AQ near Olympic Park, South Korea derived from direct sun Pandora measurements, DC-8 in situ measurements integrated from surface to 8 km and surface measurements scaled up to tropopause assuming various profile shapes (green: box with a median mixing layer height (MLH), grey: box with a measured MLH; light blue: box+exponential profile with a median MLH, and black: box+exponential profile with a measured MLH, modified from Spinei et al., 2018).



**Figure 2.** Top panel: HCHO direct sun total columns (red) and multi axis tropospheric columns (blue) as retrieved from Pandora 108 near Toronto, Canada, for a 10 months time series 2018/2019. The significant overestimation of direct sun HCHO in summer months is evident. Reference spectrum for direct sun DOAS analysis was collected during a cold winter day. Multi-axis analysis was done with zenith reference spectra measured within 2-3 min of the rest of the spectra. Bottom panel: mean diurnal temperature and maximum wind speed measured near Pandora 108.

source of the observed interference was the Pandora telescope assembly (further referred to as a head sensor). This is the only part of the instrument that was consistently exposed to the ambient conditions without any temperature control.



Figure 3. Pandora head sensor design (electronics boards are not shown).

# 2.1 Pandora head sensor

The main purpose of the Pandora head sensor is to collect light within a specific field of view, transmit light through optical

- 110 filters (e.g. U340 to block visible part of solar spectrum), and focus it onto the fiber optics patch cable for transmission to the spectrometer. The Pandora head sensor consists of several components: sealed aluminum cylindrical housing, wedged fused silica entrance window (25 mm in diameter), two filter wheels with motors, baffle holding tubes, lens, fiber optics cable, electronics board (Fig. 3) and a desiccator bag. The baffle holding tube, the two filter wheels, and the dark filter wheel plug were machined from POM-H Delrin, a trade name for polyoxymethylene, engineering thermoplastic up to March of
- 115 2019. The desiccant bags (McMaster-CARR model 2189K76, manufacturer Multisorb Technologies, model name MINIPAX) contain activated carbon (43 48% by weight) and silica gel (43 48% by weight) enclosed in Tyvec material (high-density polyethylene fiber, 5 15% by weight) and are designed to remove moisture as well as some VOCs including HCHO.

# 2.2 Polyoxymethylene (POM) Pandora head sensor parts

Polyoxymethylene (POM) has a wide range of applications due to (1) excellent mechanical (high tensile strength, rigidity
and toughness) and electrical properties at temperatures from -40 to 130°C (short-term); (2) low reactivity with and low permeability to many chemicals including organic solvents, fuels and gases at room temperature; and (3) ease of processing on standard thermoplastics equipment (Luftl et al., 2014). Six major manufacturers produce about 70% of POM worldwide, and each has its trade name (e.g., Ticona GmbH, Germany:, Celcon®; Polyplastics Co., Ltd., Japan: POM-C Duracon®, Tepcon®; E.I. Du Pont de Nemour & Co., USA: POM-H Delrin®).

- POM-H Delrin used for Pandora head sensor parts is a homopolymer POM (POM-H, Ensinger Hyde: black Delrin ecetal resin II150ebk602sheet <sup>3</sup>/<sub>4</sub> x 6 x 6) purchased from McMaster Carr, part numbers: 8575K145 and 8576K21. POM-H is produced by polymerization of purified gaseous formaldehyde in an organic liquid reaction medium in the presence of polymerization catalysts. The resulting polymer has a crystalline granular structure with macromolecules ending in at least one unstable hydroxyl group. These hydroxyl groups are responsible for POM-H thermal instability. POM deterioration occurs mainly due to the following processes:
  - 1. Depolymerization (unzipping);

- 2. Auto-oxidative scission;
- 3. Degradation by secondary products of the auto-oxidative scission;
- 4. Hydrolysis and acidolysis;
- 135 5. Photo-oxidation at wavelength 200-800 nm
  - 6. Thermal degradation.

HCHO is a byproduct of most POM degradation processes. Considering the function of the head sensor, we suspect that more than one degradation process will impact POM-H Delrin Pandora head sensor components over their lifetime (several years). POM deterioration studies are typically performed at elevated temperatures (> 90 °C) and focus on mass loss and physical

140 and mechanical property degradation measurements (Grajales et al., 2015). Review of such literature during the initial Pandora design stage led to the assumption by the NASA and SciGlob teams that HCHO was not emitted from POM-H Delrin under ambient conditions.

It may be worth noting that paraformaldehyde and high-purity  $\alpha$ -polyoxymethylene have been used to generate known concentrations of HCHO in gaseous mixtures for various applications (Ho, 1985). They are commonly used in permeation

tubes and other permeation devices at elevated temperatures (50-80°C) as a stable source of HCHO for instrument calibrations (Gilpin et al., 1997; Ho, 1985). At elevated temperatures (> 50°C) paraformaldehyde or  $\alpha$ -polyoxymethylene thermally depolymerizes to produce HCHO vapor that diffuses through the permeation tube membrane.

#### 2.3 Pandora internal head sensor temperature

- We hypothesized that the thermal instability of POM-H Delrin resulted in HCHO release at higher temperature and was the source of the temperature-dependent formaldehyde interference in Pandora direct sun HCHO. To test this hypothesis, we added an internal temperature sensor in April 2019 to monitor the internal head sensor temperature in a few instruments. We have evaluated the range of internal head sensor temperatures measured at various sites: Pandora 2 (Greenbelt, MD), 148 (Blacksburg, VA; Rotterdam and Cabauw, the Netherlands) and 155 (Boston, MA) (with an emphasis on the USA East coast where several intergovernmental field campaigns took place, Table 1). Figure 4 shows that internal head sensor temperatures ranged between 20-25°C during the nighttime hours and up to 45-50 °C during the daytime hours in summer months. During
- colder months the temperature ranged between 0 to 25  $^{\circ}$ C. These data suggested that HCHO generation was potentially relevant at internal to head sensor temperatures between 20 and 50  $^{\circ}$ C.

Internal to head sensor temperatures are determined by the following heat transfer processes between the head sensor and the surrounding environment: (a) Convective heat transfer (natural and forced) due to wind; (b) Radiant heat transfer due to

160 short wave absorption, long wave emission and long wave absorption; (c) Conduction between the head sensor and the tracker brackets; (d) Energy generation inside the head sensor.



Figure 4. Internal head sensor temperatures for Pandora 2 (Greenbelt, MD), 148 (Blacksburg, VA; Rotterdam and Cabauw, the Netherlands) and 155 (Boston, MA)

#### 3 Laboratory studies: HCHO columns as a function of Pandora head sensor internal temperature

165

170

Since multiple processes are potentially responsible for POM-H Delrin deterioration leading to HCHO generation (see Sec. 2.2) over the instrument lifetime, four Pandora head sensors of various ages were evaluated: 21 (made in 2011), 46 (made in 2015), 118 (made in 2016), and 148 (made in 2018). Pandora 148 was equipped with the internal temperature sensor in April 2019. The other three Pandora head sensors had no internal temperature measurements.

To evaluate the HCHO columns inside the Pandora head sensor as a function of internal temperature, the Pandora head sensor was placed inside a temperature-controlled enclosure ( $\pm 0.1$  °C) with a window. Enclosure temperature was varied from 10°C up to 45-55°C, kept at 45-55°C for 0.5-1 hours, and back to 10 °C at different rates. Since POM-H Delrin thermal degradation at higher temperatures is heating rate dependent (e.g., Pielichowska (2015)) we tested atmospheric relevant heating

and cooling rates: 3, 3.5, 5, 8, 8.2 °C/hr. Pandora 148 also was tested at 50 °C for 9 hr to determine time needed to reach steady state conditions.

The DOAS technique was used to analyse laboratory measurements to determine differential HCHO columns at various temperatures with respect to the lowest temperature. The experiments were mimicking temperature ranges and DOAS analysis

175 during a typical summer field campaign but without the actual atmosphere. Since the laboratory measurements were performed under tightly controlled conditions (e.g., stable smooth FEL source, constant spectrometer temperature, single gas), the DOAS equation is simplified to Eq. (1) for such conditions:

$$\ln\left(\frac{I_o(\lambda, T_o)}{I(\lambda, T) - \text{offset}(\lambda)}\right) = \sigma(\lambda, 298\text{K})_{\text{HCHO}} \cdot \Delta S_{\text{HCHO}}^{\text{T}} + P_L$$
(1)

Where,  $I(\lambda)$  are the measured attenuated FEL intensities (corrected for instrumental properties) within the wavelength fitting window between 332 to 360 nm at an internal head sensor temperature T;  $I_o(\lambda)$  are the measured FEL intensities at internal head sensor temperature  $T_o$  corresponding to ambient temperature 10 °C;  $\sigma(\lambda, 298\text{K})$  is HCHO molecular absorption cross-section at temperature 298K (Meller and Moortgat, 2000); polynomial order  $P_L = 5$ ; and offset order 1. This approach estimates differential slant column densities of HCHO along the head sensor length (153.5mm) from the front window to the lens ( $\Delta S_{\text{HCHO}}^{\text{T}} = S_{\text{HCHO}}^{\text{T}} - S_{\text{HCHO}}^{\text{T}_o}$ ).

#### 185 3.1 Experimental setup

The head sensor collimator was protruded through the enclosure window to avoid measuring any potential HCHO outgassing inside the enclosure itself (e.g., paint). The laboratory hosting the measurements was temperature-controlled (20-23°C), has both an air supply and air intake, and the door to the room was open to improve ventilation.

The collimator was pointed at the Gooch and Housego 1000W FEL lamp controlled by the current precision source (OL 410-1000). The FEL lamp was operated at 8A. The distance between the FEL and the collimator was 50 cm.

The enclosure temperature was controlled by NesLab 7 recirculating bath and a LYTRON heat exchanger with two fans to  $\pm 0.1^{\circ}$ C. The NesLeb 7 temperature sensor was placed near the Pandora head sensor. Enclosure air temperature near the head sensor, front plate Pandora head sensor temperature, and Pandora head side temperature were recorded during all the measurements using fast response stick-on surface thin film PT100 RTD elements (3-Wire, The Sensor Connection). An ADAM-4015

- 6-channel RTD Module with Modbus digitized the RTD signal. The temperature inside the newer generation head sensor (since April 2019) was measured using the pre-installed Bosch BME280 digital humidity, pressure and temperature sensor on Spark-Fun Atmospheric Sensor Breakout Board. PT100 RTD elements were inter calibrated. They agreed within the manufacture specifications (< 0.15 °C). The accuracy of BME280 was harder to verify due to Pandora head internal power generation of 2W (manufacturer reported accuracy is ±1.0 °C between 0 and 65 °C).</p>
- 200 Since only one of the tested Pandora head sensors was equipped with the internal temperature sensor, we determined an outside measurement that is the most representative of the internal temperature. This was done by comparing surface temperature measurements by the PT100 RTD elements at various locations on the Pandora head sensor versus the internal to head sensor temperature. As expected, there is some time delay in response between the surface measurements and the internal temperature. This delay is rate specific. Strong linear correlation between the surface measurements and internal temperature
- 205 measurements (accounting for transient heat transfer) was observed for both front plate (slope = 0.970, intercept =  $4.74^{\circ}$ C, RMSE = 0.059 °C) and side (slope = 0.999, intercept =  $6.79^{\circ}$ C, RMSE =  $0.087^{\circ}$ C). Since the electronics board heat sink is connected to the front, we use the front plate surface temperature as the proxy for the internal temperature.

The Pandora spectrometer temperature was controlled using a Pandora thermoelectric controller at the set temperature of 15°C. The measurements were averaged over 40 seconds sequentially switching between open, plug, U340, plug filter wheel positions to simulate direct sun measurements.

In addition to the FEL lamp one head sensor (Pandora 118) was also analyzed using a 300 nm LED (Thorlabs M300L4) controlled by a high precision LED driver (Thorlabs DC2200). In the case of the LED source, the Pandora collimator pointed into an 8.3 cm Labsphere Spectralon® reflectance material integrating sphere illuminated by the LED.

#### 3.2 Post-summer 2019: "POM-H Delrin free" Pandora head sensors

- 215 Since Summer 2019, new Pandora head sensors are POM-H Delrin free. POM-H Delrin was replaced with Molybdenum Disulfide ( $MoS_2$ ) filled Easy-to-Machine Wear-Resistant Cast Nylon 6/6 also purchased from McMaster (Tecamid 66 MO, polyamide > 90% by weight,  $MoS_2 < 10$ % by weight, manufactured by ENSINGER INC). To evaluate potential thermal oxidation of polyamide and  $MoS_2$  by air oxygen three new head sensors (Pandora 165, 167 and 168, manufactured at the end of 2019) upgraded with the Nylon parts were tested using FEL (1000W) and 300 nm LED sources. The enclosure temperature
- varied from 10°C to 55°C over 8 hours, with 1 hour at 55°, and 8 hours cooling from 55°C to 10°C. This translated to internal temperatures from 17°C to 60°C. The FEL current was set at 7.5A. Pandora spectra were binned within 40s for measurements with no filters (open, single spectrum integration time 2.4 ms, about 12550 cycles per measurement and dark 2320 cycles); 240 s with U340 filter (integration time 12.9 ms, about 15335 cycles per measurement and dark 2835 cycles); and 240 s with BP300 filter (integration time 117 ms, about 1730 cycles per measurement and dark 320 cycles). The spectrometer electronics
  board temperature was maintained at 12.9 °C (controller set temperature 5 °C). We also repeated the test with 300 nm LED
  - at a constant current of 350 mA with no filters for 100 s total integration time. The experiments were designed to ensure low noise in case of small emissions of HCHO or presence of other species.

The reference spectrum was collected at the lowest internal head sensor temperature (about  $17^{\circ}$ C). DOAS fitting included only HCHO and sulfur dioxide, SO<sub>2</sub>, molecular absorption cross-sections (Vandaele et al., 1998), the polynomial order for the

230 broadband attenuation ( $P_L$ ) was set to 5, and a 0 order polynomial represented the offset. Only the dark current spectrum was subtracted from the data since all other parameters are expected to be constant (e.g., no wavelength shift in the temperaturecontrolled lab, no need for non-linearity correction, no need for pixel response non-uniformity correction, etc.). The DOAS fitting was performed using QDOAS v3.2 program. SO<sub>2</sub> was fitted as a precaution since sulfur containing in MoS<sub>2</sub> has been reported to oxidise to SO<sub>2</sub> at temperatures higher than 140 °C (https://core.ac.uk/download/pdf/10884897.pdf).

#### 235 4 Laboratory studies results

# 4.1 POM Pandora head sensor HCHO dynamics rates

Analysis of Pandora 148 data over 9-hr period showed that the equilibrium between HCHO generation and removal processes inside the head sensor is reached almost instantaneously (at the DOAS fitting accuracy). Investigation of the actual process mechanisms are outside of the scope of this paper. However, it is probably also controlled by desiccant activated carbon adsorption as a function of temperature, in addition to "pure" solid POM-H - vapor phase processes (Sec. 2.2).

240



**Figure 5.** HCHO formation/deposition inside Pandora 118, 148, 21 and 46 as a function of front panel outside surface temperature relative to the individual head sensor measurements at 16 °C. Pandora 46 was tested twice with a combination of temperature change rates (3 and  $8.2 \,^{\circ}\text{C} \cdot \text{hr}^{-1}$  light blue and  $3.5 \,^{\circ}\text{C} \cdot \text{hr}^{-1}$  dark blue)

Figure 5 shows that HCHO columns inside Pandora head sensors follow exponential dependence on temperature irrelevant of heating or cooling rates for all four tested head sensors (Eq. 2). This temperature dependence does not show any hysteresis at the time scales relevant to this study. HCHO columns in all four sensors had the same exponential function damping  $(b = 0.0911 \pm 0.0024 \,^{\circ}\text{C}^{-1})$  but different amplitudes (and most likely absolute offsets).

245 
$$\Delta S(\Delta T_{\rm hs}) = a \left[ \exp(b \cdot T_{\rm hs}) - \exp(b \cdot T_{\rm hs}^{\rm ref}) \right]$$
(2)

The newest head sensor (Pandora 148) produced the lowest amount of HCHO (about 0.3 DU) at an internal temperature of about 50 °C (front panel external temperature 45 °C). Pandora 118 generated the largest - about 3 DU at the same temperature. Pandora 46 and 21 were in between. No clear trend was observed between the age of the instruments and the HCHO amount produced in the head sensor. Pandora 148 head sensor was evaluated for temperature dependence of HCHO several times over five months and did not show any difference in the HCHO generation during that period.

250

# 4.2 **POM-H Delrin free** Pandora head sensors - no HCHO production

Initially we conducted DOAS fitting of HCHO and SO<sub>2</sub> absorption within their standard fitting windows 332 - 359 nm and 307 - 328 nm (Spinei et al., 2010) respectively. No HCHO or SO<sub>2</sub> was detected above the optical depth rms noise level of 5.10<sup>-5</sup> from the new "POM-H Delrin free" head sensors as a function of internal head sensor temperature relative to 17°C.
255 To consolidate HCHO/SO<sub>2</sub> results for both species and to evaluate residuals we have done DOAS fitting at a broader fitting window: 300 - 350 nm. Figure 6 shows a retrieval example for Pandora 167 (300 - 350 nm fitting window from spectra collected



Figure 6.  $\Delta S$  of HCHO and SO<sub>2</sub> retrieved from POM-H Delrin free Pandora 167 head sensor at internal temperatures from 17°C to 61°C using FEL at 7.5 A (see text for information about thermal rates). Individual spectra integration time 12.9 ms, total exposure per measurement 240 s. Fitting window 300 to 350 nm. Fitting residual optical depth rms was  $4.77^{-5} \pm 4.29^{-6}$  during the entire measurements period.

with U340 filter). The spectra were also evaluated for any absorption across the entire instrument wavelength range from 300 to 530 nm by only taking the radiance ratio but not fitting any trace gases. We did not see any signatures above the instrumental noise level.

# 260 5 Effect of internally generated HCHO on direct sun and multi-axis Pandora HCHO measurements.

265

In general, DOAS analysis will cancel any instrumental "artifacts" if they are the same in the reference spectrum and the rest of the spectra. As applied to the internally generated HCHO, it will cancel if the reference spectrum and the rest of the spectra are measured at the same head sensor temperature. DOAS fitting results from direct sun or multi-axis measurements  $(\Delta S(\mu, \Delta T_{hs}, t)_{HCHO})$  accounting for internal to the head sensor HCHO  $(S(T_{hs}, t)_{HCHO}^{hs})$  can be described by the following equation:

$$\Delta S(\mu, \Delta T_{\rm hs}, t) = S(T_{\rm hs}, t)^{\rm hs} + S(\mu, t)^{\rm atm} - S(T_{\rm hs}^{\rm ref}, t^{\rm ref})^{\rm hs} - S(\mu^{\rm ref}, t^{\rm ref})^{\rm atm}$$
$$= S(\mu, t)^{\rm atm} - S(\mu^{\rm ref}, t^{\rm ref})^{\rm atm} + a \cdot \exp(b \cdot T_{\rm hs}) - a \cdot \exp(b \cdot T_{\rm hs}^{\rm ref})$$
$$= \Delta S(\mu, t)^{\rm atm} + a \left[\exp(b \cdot T_{\rm hs}(t)) - \exp(b \cdot T_{\rm hs}^{\rm ref})\right]$$
(3)

Where, a is a head sensor dependent amplitude and is not known for instruments not tested in the laboratory; b is constant damping for all tested systems and is  $\approx 0.10 \,^{\circ}\text{C}^{-1}$ .

For multi-axis observations where spectra measured at low elevation angles ( $\mu$ ) are analyzed using a zenith ( $\mu^{ref}$ ) reference 270 spectrum measured within a few minutes when  $T_{hs} \approx T_{hs}^{ref}$ , HCHO amount due to POM-H Delrine emission is about the same and mostly cancels (see Sec. 5.2). During the CINDI-2 campaign, however, the data analysis protocol for  $\Delta S_{\rm HCHO}$ , that were inter-compared between the instruments, was to use a reference spectrum collected around the local noon for all spectra measured throughout the entire day (Kreher et al., 2020). In this case  $T_{\rm hs} \neq T_{\rm hs}^{\rm ref}$  and the retrieved  $\Delta S_{\rm HCHO}$  are impacted by the internally generated HCHO (see Sec. 5.3).

In the case of direct sun measurements, a single reference spectrum at a specific temperature is applied to analyse the data over extended periods. In this case  $T_{hs}(t) \neq T_{hs}^{ref}$  and total vertical column (*C*) derived from direct sun measurements is impacted differently depending on the actual head sensor temperature and air mass factor (*AMF*), according to Eq. (4). It is assumed that calibration approach called 'Minimum Langley Extrapolation Method' (Herman et al., 2009) is capable to estimate the amount in the reference spectrum (including the head sensor amount).

$$C = \frac{\Delta S(\mu, \Delta T_{\rm hs}, t) + S(\mu^{\rm ref}, t^{\rm ref})}{\rm AMF}(\mu)}{\rm AMF}(\mu) = \frac{\Delta S(\mu, t)^{\rm atm} + \Delta S(T_{\rm hs}, t)^{\rm hs} + S(T_{\rm hs}^{\rm ref}, t^{\rm ref})^{\rm hs} + S(\mu^{\rm ref}, t^{\rm ref})^{\rm atm}}{\rm AMF}(\mu)}{\rm AMF}(\mu)$$

$$\approx \frac{\Delta S(\mu, t)^{\rm atm} + S(\mu^{\rm ref}, t^{\rm ref})^{\rm atm}}{\rm AMF}(\mu)}{\rm AMF}(\mu) + \frac{a \cdot \exp(b \cdot T_{\rm hs}(t))}{\rm AMF}(\mu)}{\rm AMF}(\mu)$$
(4)

#### 5.1 Case study: direct sun measurements using collocated POM- and POM-H Delrin free instruments.

Since direct measurements of HCHO in the head sensor is harder to perform during routine direct sun observations, we evaluate the head sensor HCHO production effect using two instruments: Pandora 32 and Pandora 2 (neither was evaluated in the laboratory according to Sec. 3) during outdoor operation. Both instruments operated side-by-side at the NASA Goddard Space Flight Center in Greenbelt, MD (38.9926°, -76.8396°, 90 m a.s.l.) in direct sun mode during July 2019 - January 2020. Pandora 2, originally built in 2009 with the standard POM-H Delrin components, was upgraded in June 2019 with POM-H Delrin free components and internal temperature sensor (see section 3.2). The Pandora 32 head sensor, originally built in 2012, still contains the original POM-H Delrin components and does not have an internal temperature sensor. To evaluate the effect of internally generated HCHO on the direct sun total column measurements during a "typical" field campaign study, we used 1.5 month of data from August 30 to October 15, 2019 when both instruments had minimal instrumental issues.

The evaluation consists of several steps:

1. Use Pandora 32 and 2 data to estimate the exponential HCHO production amplitude inside Pandora 32 during selected 1.5 months;

275

280

$$a = \operatorname{median}\left[\frac{\Delta S(\mu, \Delta T_{hs}, t) - \Delta S(\mu, t)^{\operatorname{atm}}}{\exp(b \cdot T_{hs}(t)) - \exp(b \cdot T_{hs}^{\operatorname{ref}})}\right] = \operatorname{median}\left[\frac{\Delta S(\mu, \Delta T_{hs}, t)^{\operatorname{P32}} - \Delta S(\mu, t)^{\operatorname{P2}}}{\exp(b \cdot T_{hs}(t)) - \exp(b \cdot T_{hs}^{\operatorname{ref}})}\right]$$
(5)

2. Calculate HCHO column produced in the head sensor knowing Pandora 2 head sensor temperature and exponential damping and amplitude for Pandora 32 head sensor ( $S_{\text{HCHO}} = a \cdot \exp(b \cdot T_{\text{hs}})$ ) for 7 months (17 July 2019 - 7 Feb 2020);

3. Apply air mass factor to the amount in the head sensor to evaluate diurnal and seasonal contribution to the total column measurements from direct sun data during 7 months (17 July 2019 - 7 Feb 2020).

The assumption about the same internal temperature for Pandora 32 and 2 is based on almost identical head sensor designs, collocation and the same mode of operation. The derived HCHO production amplitude for Pandora 32 is 0.0133 DU. DOAS analysis to calculate HCHO columns was performed in the fitting window 332 -359 nm with  $P_L = 4$  and an offset and wavelength shift of polynomial order 1. In addition to HCHO at 298 K (Meller and Moortgat, 2000), absorption by ozone (O<sub>3</sub>, 305 at 223 and 243 K, Malicet et al., 1995), nitrogen dioxide (NO<sub>2</sub>, linear temperature model, Vandaele et al., 1998), oxygen collision complex (O<sub>2</sub>O<sub>2</sub>, at 294 K Thalmann and Volkamer 2013), and bromine monoxide (BrO, at 223 K Fleischmann et al., 2004) was fitted. Their high-resolution molecular absorption cross-sections were convolved with the Pandora instrument transfer function prior to DOAS fitting (for convolution details see Cede (2019)). The reference spectrum was created by averaging all spectra within  $\pm 5^{\circ}$  of the minimum SZA on a cloud-free day 15 October 2020 with an average internal head sensor

temperature of  $\approx 29 \ ^{\circ}C$ . 310

300

315

Figure 7 shows a linear correlation between  $\Delta S_{\rm HCHO}$  measured by Pandora 32 and differential columns estimated from Pandora 2 measurements and HCHO produced by the Pandora 32 head sensor. The linear regression analysis between these data sets shows that the exponential function represents a reasonable estimation of the internally generated HCHO by Pandora 32 head sensor measurement during direct sun measurements (slope = 1.00, intercept = -0.03DU and  $R^2 = 0.92$ ). Deviations between the true Pandora 32 measurements and simulated from Pandora 2 measurements and internally produced HCHO are also due to small differences in Pandora 32 and 2 fields of view, diffusers, and pointing accuracy.

Figure 8 shows estimated HCHO column density inside Pandora 32 head sensor (red) based on the exponential function coefficients (a = 0.0133 DU, b = 0.0911  $^{\circ}C^{-1}$ ) and collocated Pandora 2 internal temperature. Internally generated HCHO amount is smaller during the winter months (< 0.2 DU) and reaches up to 1.15 DU during hot summer days for Pandora 32 320 head sensor. Since the total  $S_{\rm HCHO}$  is divided by the direct sun air mass factor, the head sensor contribution to the total vertical column is also solar zenith angle-dependent in addition to the internal head sensor temperature (this should not be confused with the actual amount in the head sensor). Its contribution is the largest during the middle of the day near the summer solstice and smallest at large solar zenith angles (80 ° in this study). Due to colder temperatures and larger AMFs during winter months over non-tropical regions, head sensor generated HCHO contribution to the vertical column is small (< 0.1 DU for this instrument). Figure 9 shows that during the cooler and windy summer days (e.g., 23 August 2019), head sensor HCHO 325 can result in a relatively small amount contributing to the total direct sun column. Similar behavior was observed during the KORUS-AQ field campaign (Spinei et al., 2018) when altitude integrated in situ aircraft measurements mostly agreed with Pandora columns in the morning at higher solar zenith angles and disagreed during the middle of the day. Figure 10 (a) shows that warm season measurements at solar zenith angle  $< 60^{\circ}$  are most impacted. Figure 10 (b) shows that a significant number

of measurements have a contribution to the order of background level ( $\approx 0.5$  DU) or higher, which significantly impacts the 330 accuracy of direct sun observations.



Figure 7. Linear regression analysis of the estimated and measured  $\Delta S_{\text{HCHO}}$  by Pandora 32, including true atmospheric and POM-H Delrine emitted HCHO. Pandora 2 (POM free) measured only atmospheric HCHO. Pandora 32 measured both true atmospheric abundance and POM-H Delrine emitted HCHO. Instruments were collocated at NASA/GSFC and made direct sun measurements from 30 August to 15 October 2019. The reference spectrum was collected around local noon on 2019/10/15:  $\Delta S_{\text{HCHO}}^{\text{hs}} = 0.0133 \cdot \exp(0.0911 \cdot T_{\text{hs}}) \text{ DU}$ 



**Figure 8.** Estimated HCHO column density inside Pandora 32 head sensor during deployment at GSFC/NASA, Greenbelt, MD. Estimation is based on the exponential function amplitude derived from Pandora 32 and Pandora 2 direct sun measurements of HCHO (0.0133 DU) and exponential function damping coefficient derived from the laboratory measurements of four other instruments (0.091). Direct sun air mass factors used to calculate HCHO are limited to solar zenith angles smaller than  $80^{\circ}$ 



**Figure 9.** Estimated HCHO column density inside Pandora 32 head sensor during deployment at GSFC/NASA, Greenbelt, MD. Estimation is based on the exponential function amplitude derived from Pandora 32 and Pandora 2 direct sun measurements of HCHO (0.0133 DU) and exponential function damping coefficient derived from the laboratory measurements of four other instruments (0.091). Direct sun air mass factors used to calculate HCHO are limited to solar zenith angles smaller than  $80^{\circ}$ 



**Figure 10.** Estimated HCHO column density inside Pandora 32 head sensor during deployment at GSFC/NASA, Greenbelt, MD (17 July 2019 - 7 Feb 2020). Estimation is based on the exponential function amplitude derived from Pandora 32 and Pandora 2 direct sun measurements of HCHO (0.0133 DU) and exponential function damping coefficient derived from the laboratory measurements of four other instruments (0.091). Direct sun air mass factors used to calculate HCHO are limited to solar zenith angles smaller than 80°

# 5.2 Case study: effect of internally generated HCHO on multi-axis measurements during TROLIX'19 field campaign

In this section we evaluate the variability of internally generated HCHO, and its effect on MAX-DOAS retrievals using "closest" in time (< 10 minutes) zenith reference spectra during TROLIX'19 campaign. Pandoras 148 and 118 participated in TROLIX'19 campaign. The main goal of TROLIX'19 was validation of TROPOMI L2 main data products including UVAI, Aerosol Layer Height, NO<sub>2</sub>, O<sub>3</sub>, and HCHO under a wide range of atmospheric conditions. Pandora 148 has been tested in the laboratory (see Sec. 4.1) three times over the period of 5 months and showed no changes in internally produced HCHO as a function of temperature. We use Pandora 148 data collected during the TROLIX'19 campaign to estimate the effect of internally produced HCHO on the multi-axis  $\Delta S_{\rm HCHO}$  retrieved with individual scan reference. Pandora 148 was equipped with an

internal temperature sensor and had well characterized internal HCHO temperature dependence before deployment in western Rotterdam metropolitan area (51.9172°, 4.4066°, 7 m above sea level) during September 2019. Pandora 118 was characterized for temperature dependent HCHO production in December 2019, three months after the TROLIX'19 deployment.

The effect of internally generated HCHO on multi-axis  $\Delta S_{\text{HCHO}}$  using closest in time reference zenith spectrum collected maximum 10 min apart from the rest of the scan spectra are evaluated according to the following steps:

- 1. Calculating the head sensor produced amount,  $S_{\text{HCHO}}(T_{hs}(t))$ , based on the head sensor temperature and Pandora 148 exponential function coefficients: a = 0.0041 DU and b = 0.0911 °C<sup>-1</sup>;
- 2. Calculating the head sensor HCHO produced amount at the individual scan reference spectrum time (maximum 10 min) based on the head sensor temperature and Pandora 148 exponential function coefficients: a = 0.0041 DU;  $b = 0.0911 \,^{\circ}\text{C}^{-1}$  ( $S_{\text{HCHO}}(T_{hs}^{ref})$ ;
- 3. Calculate the amount of HCHO due to POM:  $\Delta S_{HCHO}^{hs} = a \left[ \exp(b \cdot T_{hs}(t)) \exp(b \cdot T_{hs}^{ref}) \right]$

350

Figure 11 shows that Pandora 148 internally generated HCHO contribution to the multi-axis  $\Delta S_{\text{HCHO}}$  while using single scan reference is very small (< 0.005 DU). As expected, this small contribution is mostly due to lower internal temperature variations within 10 min period and partially due to small generation rates inside Pandora 148 head sensor (a = 0.0041 DU, see Fig. 5). Since Pandora 118 was only characterized once for temperature dependence of HCHO production we do not have high confidence in its temperature dependence "stability". If we assume that the exponential function amplitude was the same in September as in December (a = 0.049 DU), Pandora 118 head sensor contributed almost 10 times more than Pandora 148 to multi-axis  $\Delta S_{HCHO}$ . Even in this case, the resulting amount is smaller than 0.05 DU, which is lower than the DOAS fitting noise for most DOAS instruments (0.3 DU = 8 × 10<sup>15</sup> molecules/cm<sup>2</sup>, Table 7 in Kreher et al. (2020).)

# 360 5.3 Case study: effect of Pandora internally generated HCHO on the CINDI-2 $\Delta S_{\text{HCHO}}$ intercomparison with other DOAS instruments

Five Pandoras participated in the Second Cabauw Intercomparison campaign for Nitrogen Dioxide measuring Instruments (CINDI-2) that took place at Cabauw, The Netherlands (51.97° N, 4.93° E, September 2016) (Kreher et al., 2020). A formal



Figure 11. Estimated contribution from internally generated HCHO on the multi-axis  $\Delta S_{\text{HCHO}}$  measured by Pandora 148 during TROLIX'19 using individual scan reference spectrum about 10 min apart from the lowest viewing angle.

- semi-blind intercomparison exercise was performed to compare ΔS<sub>gas</sub> of NO<sub>2</sub>, HCHO, O<sub>2</sub>O<sub>2</sub>, and O<sub>3</sub> measured by 36
  spectroscopic systems from 24 institutes during 17 days in September 2016. To limit any variability due to differences in temporal sampling by each instrument for semi-blind intercomparison exercise, all multi-axis daily scans were analyzed using that day's local noon spectra. This type of analysis results in a stronger contribution of the internally generated HCHO on the Pandora ΔS<sub>HCHO</sub> that were compared with the rest of DOAS instruments. Since none of the Pandoras in September 2016 were equipped with an internal temperature sensor, we use Pandora 148 data during TROLIX'19 measurements as a surrogate for the CINDI-2 campaign. Pandora 148 was deployed at a location about 38 km south-west of the CINDI-2 site during the same month of the year as CINDI-2. While differences in atmospheric conditions are expected between the sites and years (2019 vs.)
  - 2016), we assume that general trends in internal Pandora head sensor temperature are similar over the measurement periods.Only one Pandora (Pandora 118) was tested for internal HCHO generation, but more than three years later. We assume thatPandora 32 exponential amplitude is more representative of a "typical" Pandora rate than Pandora 118.
- To evaluate the effect of internally generated HCHO on  $\Delta S_{\text{HCHO}}$  used for semi-blind intercomparison during CINDI-2, a single reference spectrum was used to analyze the entire day of multi-axis data during TROLIX'19 by:
  - 1. Calculating the head sensor produced amount based on Pandora 148 head sensor temperature and Pandora 32 exponential function coefficients: a = 0.013 DU;  $b = 0.0911 \,^{\circ}\text{C}^{-1}$ ;
  - 2. Calculating the head sensor produced amount at the reference spectrum time (local noon at minimum solar zenith angle);
  - 380 3. Calculate  $\Delta S_{\text{HCHO}}^{\text{hs}}$  due to POM.

Figure 12 shows that the internally generated HCHO contribution to  $\Delta S$  is negative before local noon, positive during early afternoon and negative again during later afternoon. Since the internal temperatures did not vary by more than 15°C during



Figure 12. Estimated  $\Delta S_{\text{HCHO}}$  contributing to the multi-axis measurements while local noon reference is used. This is relevant only to DOAS instrument intercomparison campaigns such as CINDI-2 since standard data inversion requires individual scan reference. Exponential production amplitude applied is 0.0133 DU, however the actual amplitude observed in the lab was between 0.0041 and 0.049.  $\Delta S_{\text{HCHO}}$  is calculated based on Pandora 148 head sensor temperature during TROLIX'19 campaign west of Rotterdam, the Netherlands during August 31 to 5 October 2019.

daily measurements and maximum did not exceed 40°C the overall effect is in general small < 0.1 DU with slightly negative bias (Fig. 13) for an instrument similar to Pandora 32, 21 and 46. While we do not know the exact HCHO internal generation rates for the Pandoras deployed during CINDI-2 we can assume that the minimum corresponds to Pandora 148 and maximum to Pandora 118, which is about 3.2 times smaller or 3.7 times larger than in Fig. 12 and 13.

Note, that DOAS analysis using daily noon zenith reference spectra was implemented only for the formal semi-blind intercomparison of  $\Delta SCD$  exercise (Kreher et al., 2020). Full data processing and inversion to the final products, tropospheric columns and profiles, was done using individual scan zenith spectra not daily noon zenith spectra (Tirpitz et al., 2020).

#### 390 6 Conclusions

385

Pandora direct sun measurements of HCHO were impacted by the internally generated HCHO inside head sensor due to thermal degradation of **POM-H Delrin** plastic parts up until summer 2019. Direct sun measurements before spring 2016 were also effected by the etaloning off the front window surfaces. Pandora multi-axis measurements of HCHO were significantly less impacted by the internally generated HCHO. The following list represents the major findings of this work:

HCHO in Pandora head sensors (up to summer 2019): exponential temperature dependence of HCHO production was observed for four tested head sensors with a damping coefficient of 0.0911 ± 0.0024 °C<sup>-1</sup>. The exponential function amplitude ranged from 0.0041 DU for P148 to 0.049 DU for P118. No apparent dependency on the head sensor age and heating/cooling rates was observed (Fig. 5);



Figure 13. Estimated  $\Delta S_{\text{HCHO}}$  contributing to the multi-axis measurements while local noon reference is used. This is relevant only to DOAS instrument intercomparison campaigns such as CINDI-2 since standard data inversion requires individual scan reference. Exponential production amplitude applied is 0.0133 DU, however the actual amplitude observed in the lab was between 0.0041 and 0.049.  $\Delta S_{\text{HCHO}}$  is calculated based on Pandora 148 head sensor temperature during TROLIX'19 campaign west of Rotterdam, the Netherlands during 31 August to 5 October 2019.

Three new POM-H Delrin free Pandora head sensors (starting from summer 2019) were evaluated for temperature dependent attenuation across the entire spectral range. The noise was minimized by reducing spectrometer temperature to 5 °C set temperature and averaging more than 10000 spectra per measurement to allow detection of smaller absorption (Fig. 6). No HCHO or SO<sub>2</sub> were detected under the measurement conditions. No other absorptions above the instrumental noise were observed across the entire spectral range;

400

405

410

- 3. Evaluation of  $\Delta S_{HCHO}$  measured from two collocated Pandora instruments, one with POM-H Delrin and one without POM-H Delrin parts, operating in direct sun mode allowed for derivation of exponential production amplitude (0.0133 DU, Fig. 7);
  - 4. The total amount of HCHO internally generated by the POM-H Delrin components and contributing to the direct sun measurements were estimated based on temperature and solar zenith angle of the measurements. Measurements in winter, during colder days in general and at high solar zenith angles (> 75 °) were minimally impacted. Measurements during hot days and small solar zenith angles had up to 1 DU contribution from POM-H Delrin parts (Fig. 8, 9, 10).
- 5. Pandora HCHO measurements derived from Pandora direct sun observations between 2016 and 2019 cannot be used in the current form for any scientific conclusions about atmospheric HCHO. Results presented here most likely are representative of other Pandora instruments operational between 2016 and summer 2019.

- Considering that Pandora head sensors have almost identical design from material, shape and thermodynamics point of view measurements between 2016 and 2019 can be corrected based on (a) meteorological observations (temperature and wind) to estimate internal head sensor temperature and (b) on  $\Delta S$  measurements to estimate HCHO production amplitude (Spinei et al. 2020 in preparation)
  - 6. Multi-axis measurements had a minimal contribution (< 0.01 DU) to  $\Delta S_{HCHO}$  due to the scan reference spectrum and the rest of the scan spectra collected within a short time period with small difference in head sensor temperature (Fig. 11).
  - 7. CINDI-2 instrument intercomparison data analysis (Kreher et al., 2020) is not representative of the final multi-axis data processing (leading to profile inversion described by Tirpitz et al. (2020)) since the noon reference spectrum was used for DOAS fitting and no subsequent subtraction of the scan zenith was done. This resulted in higher internal head sensor contribution to  $\Delta S_{HCHO}$  that were inter-compared with the other instruments (Fig. 12, 13).
- 425 Data availability. All Pandora data are available from lb3.pandonia.net

*Author contributions*. Elena Spinei (ES) has conceived the idea of the laboratory testing, performed laboratory data collection and analysis and taken the lead role in writing the manuscript. ES operated Pandora 148 during TROLIX-19 campaign.

Manuel Gebetsberger and Moritz Mueller assured Pandora network data quality and data analysis from multiple instruments, and edited the manuscript.

430 Martin Tiefengraber and Alexander Cede lead the pre head sensor HCHO discovery effort to identify HCHO inconsistency issues and edited the manuscript Alexander Kostakis and Fernando Santos contributed to introduction writing, manuscript editing and manuscript logical flow improvement.

Nader Abbuhasan provided custom modifications to the Pandora head sensor, closely worked with ES to ensure Pandora equipment availability.

435 Luke Valin, James Szykman, and Andrew Whitehill participated in early problem identification and edited the the manuscript Xiaoyi Zhao, Vitali Fioletov, and Sum Chi Lee operated and managed the Canadian Pandora network, which provided the early field evidence of the problem.

Robert Swap closely worked with ES to ensure Pandora equipment availability for laboratory testing and participated in strategy development.

440 All co-authors discussed the results.

420

Competing interests. No competing interests are present

*Acknowledgements.* NASA Tropospheric Composition branch for providing Pandora instrumentation for testing. Michel Van Roozendael, BIRA, for a suggestion to look into the plastic components used in the Pandora head sensor. Ankie Piters, KNMI, for providing Pandora 118 for testing. Jeffrey Geddes for contributing Pandora 155 head sensor temperature data. Nabil Nowak, Yun Dong and Nash Kocur, Virginia

<sup>445</sup> Tech graduate students, for helping with early testing of Pandora 148. Jonathan Davies and Ihab Abboud, ECCC, for technical supports to Pandora 108.

#### References

455

465

Alicke, B.: Impact of nitrous acid photolysis on the total hydroxyl radical budget during the Limitation of Oxidant Production/Pianura 

450 //doi.wiley.com/10.1029/2000JD000075, 2002.

- Cede, A.: Manual for Blick Software Suite 1.7, https://www.pandonia-global-network.org/wp-content/uploads/2019/11/BlickSoftwareSuite Manual v1-7.pdf, 2019.
- Chan, K. L., Wiegner, M., Alberti, C., and Wenig, M.: MAX-DOAS measurements of tropospheric NO2; and HCHO in Munich and the comparison to OMI and TROPOMI satellite observations, preprint, Gases/Remote Sensing/Data Processing and Information Retrieval, https://doi.org/10.5194/amt-2020-35, https://www.atmos-meas-tech-discuss.net/amt-2020-35/, 2020.
- De Smedt, I., Müller, J.-F., Stavrakou, T., Van Der A, R., Eskes, H., and Van Roozendael, M.: Twelve years of global observations of

formaldehyde in the troposphere using GOME and SCIAMACHY sensors, 2008.

- De Smedt, I., Van Roozendael, M., Stavrakou, T., Müller, J.-F., Lerot, C., Theys, N., Valks, P., Hao, N., and van der A, R.: Improved retrieval of global tropospheric formaldehyde columns from GOME-2/MetOp-A addressing noise reduction and instrumental degrada-
- 460 tion issues, Atmospheric Measurement Techniques, 5, 2933-2949, https://doi.org/10.5194/amt-5-2933-2012, https://amt.copernicus.org/ articles/5/2933/2012/, 2012.
  - De Smedt, I., Stavrakou, T., Hendrick, F., Danckaert, T., Vlemmix, T., Pinardi, G., Theys, N., Lerot, C., Gielen, C., Vigouroux, C., Hermans, C., Fayt, C., Veefkind, P., Müller, J.-F., and Van Roozendael, M.: Diurnal, seasonal and long-term variations of global formaldehyde columns inferred from combined OMI and GOME-2 observations. Atmospheric Chemistry and Physics, 15, 12519–12545. https://doi.org/10.5194/acp-15-12519-2015, https://acp.copernicus.org/articles/15/12519/2015/, 2015.
- De Smedt, I., Theys, N., Yu, H., Danckaert, T., Lerot, C., Compernolle, S., Van Roozendael, M., Richter, A., Hilboll, A., Peters, E., Pedergnana, M., Loyola, D., Beirle, S., Wagner, T., Eskes, H., van Geffen, J., Boersma, K. F., and Veefkind, P.: Algorithm theoretical baseline for formaldehyde retrievals from S5P TROPOMI and from the QA4ECV project, Atmospheric Measurement Techniques, 11, 2395–2426, https://doi.org/10.5194/amt-11-2395-2018, https://amt.copernicus.org/articles/11/2395/2018/, 2018.
- 470 Gilpin, T., Apel, E., Fried, A., Wert, B., Calvert, J., Genfa, Z., Dasgupta, P., Harder, J. W., Heikes, B., Hopkins, B., Westberg, H., Kleindienst, T., Lee, Y.-N., Zhou, X., Lonneman, W., and Sewell, S.: Intercomparison of six ambient [CH2O] measurement techniques, Journal of Geophysical Research: Atmospheres, 102, 21 161–21 188, https://doi.org/10.1029/97JD01314, http://doi.wiley.com/10.1029/97JD01314, 1997.
- Grajales, E. J., Alarcón, E. A., and Villa, A. L.: Kinetics of depolymerization of paraformaldehyde obtained by thermogravimet-475 ric analysis, Thermochimica Acta, 609, 49-60, https://doi.org/10.1016/j.tca.2015.04.016, https://linkinghub.elsevier.com/retrieve/pii/ \$0040603115001549, 2015.
  - Gronoff, G., Robinson, J., Berkoff, T., Swap, R., Farris, B., Schroeder, J., Halliday, H. S., Knepp, T., Spinei, E., Carrion, W., Adcock, E. E., Johns, Z., Allen, D., and Pippin, M.: A method for quantifying near range point source induced O3 titration events using Colocated Lidar and Pandora measurements, Atmospheric Environment, 204, 43–52, https://doi.org/10.1016/j.atmosenv.2019.01.052, https://doi.org/10.1052, https://doi //linkinghub.elsevier.com/retrieve/pii/S1352231019300846, 2019.
- 480
  - Herman, J., Cede, A., Spinei, E., Mount, G., Tzortziou, M., and Abuhassan, N.: NO2 column amounts from ground-based Pandora and MF-DOAS spectrometers using the direct-sun DOAS technique: Intercomparisons and application to OMI validation, Journal of Geophysical Research, 114, D13 307, https://doi.org/10.1029/2009JD011848, http://doi.wiley.com/10.1029/2009JD011848, 2009.

- Herman, J., Spinei, E., Fried, A., Kim, J., Kim, J., Kim, W., Cede, A., Abuhassan, N., and Segal-Rozenhaimer, M.: NO2 and HCHO
   measurements in Korea from 2012 to 2016 from Pandora spectrometer instruments compared with OMI retrievals and with aircraft measurements during the KORUS-AQ campaign, Atmospheric Measurement Techniques, 11, 4583–4603, https://doi.org/10.5194/amt-11-4583-2018, https://amt.copernicus.org/articles/11/4583/2018, 2018.
  - Herman, J., Abuhassan, N., Kim, J., Kim, J., Dubey, M., Raponi, M., and Tzortziou, M.: Underestimation of column NO2; amounts from the OMI satellite compared to diurnally varying ground-based retrievals from multiple PANDORA spectrometer instruments, Atmospheric
- 490 Measurement Techniques, 12, 5593–5612, https://doi.org/10.5194/amt-12-5593-2019, https://amt.copernicus.org/articles/12/5593/2019/, 2019.
  - Hewson, W., Bösch, H., Barkley, M. P., and De Smedt, I.: Characterisation of GOME-2 formaldehyde retrieval sensitivity, Atmospheric Measurement Techniques, 6, 371–386, https://doi.org/10.5194/amt-6-371-2013, https://amt.copernicus.org/articles/6/371/2013/, 2013.
- Ho, M. H.: Generation of Standard Gaseous Formaldehyde in Test Atmospheres, in: Formaldehyde, edited by Turoski, V., vol. 210, pp. 193–
   214, American Chemical Society, Washington, D.C., https://doi.org/10.1021/ba-1985-0210.ch013, https://pubs.acs.org/doi/abs/10.1021/
- ba-1985-0210.ch013, 1985. Kreher, K., Van Roozendael, M., Hendrick, F., Apituley, A., Dimitropoulou, E., Frieß, U., Richter, A., Wagner, T., Lampel, J., Abuhassan,
  - N., Ang, L., Anguas, M., Bais, A., Benavent, N., Bösch, T., Bognar, K., Borovski, A., Bruchkouski, I., Cede, A., Chan, K. L., Donner, S., Drosoglou, T., Fayt, C., Finkenzeller, H., Garcia-Nieto, D., Gielen, C., Gómez-Martín, L., Hao, N., Henzing, B., Herman, J. R., Hermans,
- 500 C., Hoque, S., Irie, H., Jin, J., Johnston, P., Khayyam Butt, J., Khokhar, F., Koenig, T. K., Kuhn, J., Kumar, V., Liu, C., Ma, J., Merlaud, A., Mishra, A. K., Müller, M., Navarro-Comas, M., Ostendorf, M., Pazmino, A., Peters, E., Pinardi, G., Pinharanda, M., Piters, A., Platt, U., Postylyakov, O., Prados-Roman, C., Puentedura, O., Querel, R., Saiz-Lopez, A., Schönhardt, A., Schreier, S. F., Seyler, A., Sinha, V., Spinei, E., Strong, K., Tack, F., Tian, X., Tiefengraber, M., Tirpitz, J.-L., van Gent, J., Volkamer, R., Vrekoussis, M., Wang, S., Wang, Z., Wenig, M., Wittrock, F., Xie, P. H., Xu, J., Yela, M., Zhang, C., and Zhao, X.: Intercomparison of NO2, O2O2; and HCHO slant column
- 505 measurements by MAX-DOAS and zenith-sky UV-visible spectrometers during CINDI-2, Atmospheric Measurement Techniques, 13, 2169–2208, https://doi.org/10.5194/amt-13-2169-2020, https://amt.copernicus.org/articles/13/2169/2020/, 2020.
  - Kwon, H.-A., Park, R. J., González Abad, G., Chance, K., Kurosu, T. P., Kim, J., De Smedt, I., Van Roozendael, M., Peters, E., and Burrows, J.: Description of a formaldehyde retrieval algorithm for the Geostationary Environment Monitoring Spectrometer (GEMS), Atmospheric Measurement Techniques, 12, 3551–3571, https://doi.org/10.5194/amt-12-3551-2019, https://www.atmos-meas-tech.net/12/3551/2019/, 2010
- 510 2019.
  - Liu, L., Andreani-Aksoyoglu, S., Keller, J., Ordóñez, C., Junkermann, W., Hak, C., Braathen, G. O., Reimann, S., Astorga-Llorens, C., Schultz, M., Prévôt, A. S. H., and Isaksen, I. S. A.: A photochemical modeling study of ozone and formaldehyde generation and budget in the Po basin, Journal of Geophysical Research, 112, D22 303, https://doi.org/10.1029/2006JD008172, http://doi.wiley.com/10.1029/ 2006JD008172, 2007.
- 515 Luftl, S., P. M., V., and Sarathchandran, C., eds.: Polyoxymethylene handbook: structure, properties, applications and their nanocomposites, John Wiley & Sons, Scrivener Publishing, Hoboken, New Jersey : Salem, MA, 2014.
  - Meller, R. and Moortgat, G. K.: Temperature dependence of the absorption cross sections of formaldehyde between 223 and 323 K in the wavelength range 225-375 nm, Journal of Geophysical Research: Atmospheres, 105, 7089–7101, https://doi.org/10.1029/1999JD901074, http://doi.wiley.com/10.1029/1999JD901074, 2000.

- 520 Park, J., Lee, H., Kim, J., Herman, J., Kim, W., Hong, H., Choi, W., Yang, J., and Kim, D.: Retrieval Accuracy of HCHO Vertical Column Density from Ground-Based Direct-Sun Measurement and First HCHO Column Measurement Using Pandora, Remote Sensing, 10, 173, https://doi.org/10.3390/rs10020173, http://www.mdpi.com/2072-4292/10/2/173, 2018.
  - Pielichowska, K.: Thermooxidative degradation of polyoxymethylene homo- and copolymer nanocomposites with hydroxyapatite: Kinetic and thermoanalytical study, Thermochimica Acta, 600, 7–19, https://doi.org/10.1016/j.tca.2014.11.016, https://linkinghub.elsevier.com/retrieve/pii/S0040603114005231, 2015.
  - Pinardi, G., Van Roozendael, M., Abuhassan, N., Adams, C., Cede, A., Clémer, K., Fayt, C., Frieß, U., Gil, M., and Herman, J.: MAX-DOAS formaldehyde slant column measurements during CINDI: intercomparison and analysis improvement., Atmospheric Measurement Techniques, 6, 2013.
  - Pinardi, G., Van Roozendael, M., Hendrick, F., Theys, N., Abuhassan, N., Bais, A., Boersma, F., Cede, A., Chong, J., Donner, S., Drosoglou,
- 530 T., Frieß, U., Granville, J., Herman, J. R., Eskes, H., Holla, R., Hovila, J., Irie, H., Kanaya, Y., Karagkiozidis, D., Kouremeti, N., Lambert, J.-C., Ma, J., Peters, E., Piters, A., Postylyakov, O., Richter, A., Remmers, J., Takashima, H., Tiefengraber, M., Valks, P., Vlemmix, T., Wagner, T., and Wittrock, F.: Validation of tropospheric NO<sub&gt;2&lt;/sub&gt; column measurements of GOME-2A and OMI using MAX-DOAS and direct sun network observations, preprint, Gases/Remote Sensing/Validation and Intercomparisons, https://doi.org/10.5194/amt-2020-76, https://amt.copernicus.org/preprints/amt-2020-76/, 2020.
- 535 Reed, A. J., Thompson, A. M., Kollonige, D. E., Martins, D. K., Tzortziou, M. A., Herman, J. R., Berkoff, T. A., Abuhassan, N. K., and Cede, A.: Effects of local meteorology and aerosols on ozone and nitrogen dioxide retrievals from OMI and pandora spectrometers in Maryland, USA during DISCOVER-AQ 2011, Journal of Atmospheric Chemistry, 72, 455–482, https://doi.org/10.1007/s10874-013-9254-9, http://link.springer.com/10.1007/s10874-013-9254-9, 2015.
- Spinei, E., Carn, S. A., Krotkov, N. A., Mount, G. H., Yang, K., and Krueger, A.: Validation of ozone monitoring instrument SO
- 540 2 measurements in the Okmok volcanic cloud over Pullman, WA, July 2008, Journal of Geophysical Research, 115, D00L08, https://doi.org/10.1029/2009JD013492, http://doi.wiley.com/10.1029/2009JD013492, 2010.
  - Spinei, E., Whitehill, A., Fried, A., Tiefengraber, M., Knepp, T. N., Herndon, S., Herman, J. R., Müller, M., Abuhassan, N., Cede, A., Richter, D., Walega, J., Crawford, J., Szykman, J., Valin, L., Williams, D. J., Long, R., Swap, R. J., Lee, Y., Nowak, N., and Poche, B.: The first evaluation of formaldehyde column observations by improved Pandora spectrometers during the KORUS-AQ field study, Atmospheric
- 545 Measurement Techniques, 11, 4943–4961, https://doi.org/10.5194/amt-11-4943-2018, https://www.atmos-meas-tech.net/11/4943/2018/, 2018.
  - Szykman, J., Swap, R., Lefer, B., Valin, L., Lee, S., Fioletov, V., Zhao, X., Davies, J., Williams, D., Abuhassan, N., Shalaby, L., Cede, A., Tiefengraber, M., Mueller, M., Kotsakis, A., Santos, F., and Robinson Pandora, J.: Pandora: Connecting in-situ and Satellite Monitoring in Support of the Canada - U.S. Air Quality Agreement, Air and Waste Management Association's Magazine for Environmental Managers, 2019.
- 550

525

- Tirpitz, J.-L., Frieß, U., Hendrick, F., Alberti, C., Allaart, M., Apituley, A., Bais, A., Beirle, S., Berkhout, S., Bognar, K., Bösch, T., Bruchkouski, I., Cede, A., Chan, K. L., den Hoed, M., Donner, S., Drosoglou, T., Fayt, C., Friedrich, M. M., Frumau, A., Gast, L., Gielen, C., Gomez-Martín, L., Hao, N., Hensen, A., Henzing, B., Hermans, C., Jin, J., Kreher, K., Kuhn, J., Lampel, J., Li, A., Liu, C., Liu, H., Ma, J., Merlaud, A., Peters, E., Pinardi, G., Piters, A., Platt, U., Puentedura, O., Richter, A., Schmitt, S., Spinei, E., Stein Zweers, D., Strong,
- 555 K., Swart, D., Tack, F., Tiefengraber, M., van der Hoff, R., van Roozendael, M., Vlemmix, T., Vonk, J., Wagner, T., Wang, Y., Wang, Z., Wenig, M., Wiegner, M., Wittrock, F., Xie, P., Xing, C., Xu, J., Yela, M., Zhang, C., and Zhao, X.: Intercomparison of MAX-DOAS

vertical profile retrieval algorithms: studies on field data from the CINDI-2 campaign, preprint, Gases/Remote Sensing/Validation and Intercomparisons, https://doi.org/10.5194/amt-2019-456, https://www.atmos-meas-tech-discuss.net/amt-2019-456/, 2020.

- Vandaele, A., Hermans, C., Simon, P., Carleer, M., Colin, R., Fally, S., Mérienne, M., Jenouvrier, A., and Coquart, B.: Measurements of 560 the NO2 absorption cross-section from 42 000 cm-1 to 10 000 cm-1 (238-1000 nm) at 220 K and 294 K, Journal of Quantitative Spectroscopy and Radiative Transfer, 59, 171-184, https://doi.org/10.1016/S0022-4073(97)00168-4, http://linkinghub.elsevier.com/retrieve/ pii/S0022407397001684, 1998.
  - Verhoelst, T., Compernolle, S., Pinardi, G., Lambert, J.-C., Eskes, H. J., Eichmann, K.-U., Fjæraa, A. M., Granville, J., Niemeijer, S., Cede, A., Tiefengraber, M., Hendrick, F., Pazmiño, A., Bais, A., Bazureau, A., Boersma, K. F., Bognar, K., Dehn, A., Donner, S., Elokhov,
- 565 A., Gebetsberger, M., Goutail, F., Grutter de la Mora, M., Gruzdev, A., Gratsea, M., Hansen, G. H., Irie, H., Jepsen, N., Kanaya, Y., Karagkiozidis, D., Kivi, R., Kreher, K., Levelt, P. F., Liu, C., Müller, M., Navarro Comas, M., Piters, A. J. M., Pommereau, J.-P., Portafaix, T., Puentedura, O., Ouerel, R., Remmers, J., Richter, A., Rimmer, J., Rivera Cárdenas, C., Saavedra de Miguel, L., Sinvakov, V. P., Strong, K., Van Roozendael, M., Veefkind, J. P., Wagner, T., Wittrock, F., Yela González, M., and Zehner, C.: Ground-based validation of the Copernicus Sentinel-5p TROPOMI NO<sub&gt;2&lt;/sub&gt; measurements with the NDACC ZSL-DOAS, MAX-DOAS and
- 570 Pandonia global networks, preprint, Gases/Remote Sensing/Validation and Intercomparisons, https://doi.org/10.5194/amt-2020-119, https://doi.org/10, https/ //amt.copernicus.org/preprints/amt-2020-119/, 2020.
  - Vigouroux, C., Hendrick, F., Stavrakou, T., Dils, B., Smedt, I. D., Hermans, C., Merlaud, A., Scolas, F., Senten, C., and Vanhaelewyn, G.: Ground-based FTIR and MAX-DOAS observations of formaldehyde at Réunion Island and comparisons with satellite and model data. Atmospheric Chemistry and Physics, 9, 9523–9544, 2009.
- Vigouroux, C., Bauer Aquino, C. A., Bauwens, M., Becker, C., Blumenstock, T., De Mazière, M., García, O., Grutter, M., Guarin, C., 575 Hannigan, J., Hase, F., Jones, N., Kivi, R., Koshelev, D., Langerock, B., Lutsch, E., Makarova, M., Metzger, J.-M., Müller, J.-F., Notholt, J., Ortega, I., Palm, M., Paton-Walsh, C., Poberovskii, A., Rettinger, M., Robinson, J., Smale, D., Stavrakou, T., Stremme, W., Strong, K., Sussmann, R., Té, Y., and Toon, G.: NDACC harmonized formaldehyde time series from 21 FTIR stations covering a wide range of column abundances, Atmospheric Measurement Techniques, 11, 5049-5073, https://doi.org/10.5194/amt-11-5049-2018, https://www. atmos-meas-tech.net/11/5049/2018/, 2018.
- 580
  - Wang, Y., Lampel, J., Xie, P., Beirle, S., Li, A., Wu, D., and Wagner, T.: Ground-based MAX-DOAS observations of tropospheric aerosols, NO2, SO2 and HCHO in Wuxi, China, from 2011 to 2014, 2017.
  - Zhu, L., Jacob, D. J., Kim, P. S., Fisher, J. A., Yu, K., Travis, K. R., Mickley, L. J., Yantosca, R. M., Sulprizio, M. P., De Smedt, I., González Abad, G., Chance, K., Li, C., Ferrare, R., Fried, A., Hair, J. W., Hanisco, T. F., Richter, D., Jo Scarino, A., Walega, J., Weibring,
- 585 P., and Wolfe, G. M.: Observing atmospheric formaldehyde (HCHO) from space: validation and intercomparison of six retrievals from four satellites (OMI, GOME2A, GOME2B, OMPS) with SEAC4RS aircraft observations over the southeast US, Atmospheric Chemistry and Physics, 16, 13477-13490, https://doi.org/10.5194/acp-16-13477-2016, https://www.atmos-chem-phys.net/16/13477/2016/, 2016a.
  - Zhu, L., Jacob, D. J., Kim, P. S., Fisher, J. A., Yu, K., Travis, K. R., Mickley, L. J., Yantosca, R. M., Sulprizio, M. P., De Smedt, I., González Abad, G., Chance, K., Li, C., Ferrare, R., Fried, A., Hair, J. W., Hanisco, T. F., Richter, D., Jo Scarino, A., Walega, J., Weibring,
- 590 P. and Wolfe, G. M.: Observing atmospheric formaldehyde (HCHO) from space: validation and intercomparison of six retrievals from four satellites (OMI, GOME2A, GOME2B, OMPS) with SEAC<sup&gt;4&lt;/sup&gt;RS aircraft observations over the southeast US, Atmospheric Chemistry and Physics, 16, 13 477-13 490, https://doi.org/10.5194/acp-16-13477-2016, https://acp.copernicus.org/articles/ 16/13477/2016/, 2016b.