Evaluation of optical particulate matter sensors under realistic conditions of strong and mild urban pollution

Adnan Masic¹, Dzevad Bibic¹, Boran Pikula¹, Almir Blazevic¹, Jasna Huremovic², and Sabina Zero² ¹Faculty of Mechanical Engineering, University of Sarajevo, 71000 Sarajevo, Bosnia-Herzegovina ²Faculty of Science, University of Sarajevo, 71000 Sarajevo, Bosnia-Herzegovina

Correspondence: Adnan Masic (masic@mef.unsa.ba)

Abstract. In this paper we evaluate characteristics of three optical particulate matter sensors/sizers (OPS): high-end spectrometer 11-D (Grimm, Germany), low-cost sensor OPC-N2 (Alphasense, United Kingdom) and in-house developed MAQS (Mobile Air Quality System) which is based on another low-cost sensor – PMS5003 (Plantower, China), under realistic conditions of strong and mild urban pollution. Results were compared against a reference gravimetric system, based on Gemini

- 5 (Dadolab, Italy), 2.3 m³/h air sampler, with two channels (simultaneously measuring $PM_{2.5}$ and PM_{10} concentrations). The measurements were performed in Sarajevo, the capital of Bosnia-Herzegovina, from December 2019 until May 2020. This interval is divided into period 1 strong pollution and period 2 mild pollution. The city of Sarajevo is one of the most polluted cities in Europe in terms of aerosols: the average concentration of $PM_{2.5}$ during the period 1 was 83 μ g/m³, with daily average values exceeding 500 μ g/m³. During period 2, the average concentration of $PM_{2.5}$ was 20 μ g/m³. These conditions represent
- 10 a good opportunity to test optical devices against reference instrument in a wide range of ambient particulate matter (PM) concentrations. The effect of an in-house developed diffusion dryer for 11-D is discussed as well. In order to analyze the mass distribution of particles, a scanning mobility particle sizer (SMPS), which together with the 11-D spectrometer gives the full spectrum from nanoparticles of diameter 10 nm to coarse particles of diameter 35 μ m, was used. All tested devices showed excellent correlation with the reference instrument in period 1, with R^2 values between 0.90 and 0.99 for daily average PM
- 15 concentrations. However, in period 2, where the range of concentrations was much narrower, R^2 values decreased significantly, to values from 0.28 to 0.92. We have also included results of a 13.5 month long-term comparison of our MAQS sensor with a nearby beta attenuation monitor (BAM) 1020 (Met One Instruments, USA) operated by the United States Environmental Protection Agency (US EPA), which showed similar correlation and no observable change of performance over time.

1 Introduction

20 Analysis of particulate matter represents a key element for the studies of air pollution. Various studies shed light on their effect on health (Downward et al., 2018) and climate (Zhao et al., 2019). In many cases particulate matter is a dominant pollutant among other components of pollution. Therefore, developing a strategy for reliable quantification of particulate matter in ambient air is necessary. The traditional and most accurate approach to measuring the particulate matter concentration in the air is the reference method, based on gravimetric measurements, after the collection of particulate matter by air samplers.

- 25 The typical time resolution of such measurements is 24 hours. Although there are portable air samplers, these measurements are usually performed at fixed locations, such as research supersites. Reference systems are expensive and require a lot of laboratory work. Results are not immediately available, because of the time-consuming process of filter treatment. Taking that into account, various governmental institutions usually opt for more affordable and easier to use and maintain equivalent methods. These are usually fixed, semi-automatic stations equipped with beta attenuation monitors (BAMs). The typical time
- 30 resolution of such stations is 1 hour. If maintained and calibrated properly, the equivalent methods should achieve an acceptable level of agreement with the reference. For example, one long-term comprehensive study (Hafkenscheid and Vonk, 2014) performed at 14 different locations across Netherlands, showed that a linear correction y = 0.91x - 1.6, applied on raw readings from BAM, was necessary to achieve the requirements of the Guide to the Demonstration of Equivalence (ECWG, 2010).
- Newer methods, based on optical particle sensors (OPS), are nowadays increasingly more popular, particularly low-cost
 variants (Zheng et al., 2019; Mukherjee et al., 2019; Tanzer et al., 2019; Morawska et al., 2018). Their typical time resolution is between 1 s and 1 min, and because of their price and size, they can be used in networks to provide better spatial coverage (Martin et al., 2019; Li et al., 2019). Furthermore, they provide information about multiple mass fractions of particulate matter simultaneously, unlike the concentration of single fraction in gravimetric system or BAM. However, there are concerns about their suitability for measuring mass concentrations of ambient PM, since there is a significant measurement uncertainty arising
- 40 from the principles of their operation.

45

50

Most commercially available OPS use the Mie scattering theory (Mie, 1908) to determine the size and number of particles within the unit volume of air. The Mie theory provides the solution of the Maxwell equations for the scattering of plane waves on spherical particles. The Mie solution is rather complex, but in order to illustrate the non-linearity of the theory, it will suffice to consider the case where particles are much smaller than the wavelength (of light, since a red laser is commonly used in practice). In that case the intensity of scattered radiation is given by

$$I = I_0 \frac{1 + \cos^2 \theta}{2R^2} \left(\frac{2\pi}{\lambda}\right)^4 \left(\frac{d}{2}\right)^6 \left|\frac{m^2 - 1}{m^2 + 2}\right|^2,$$
(1)

where I_0 is intensity of the incident radiation, θ is the scattering angle, R is the distance between the particle and the observing point, λ is the wavelength, d is the particle diameter and m is the refractive index of the particle. Thus, in order to calculate the diameter of the particle by measuring the intensity of the scattered radiation, one must assume a value for the refractive index of the particle. If the particle absorbs nothing from incoming radiation, its refractive index will be real, otherwise it is written in the form

$$m = n + i\kappa,\tag{2}$$

where κ is called the extinction coefficient and is related to the absorption coefficient α :

$$\alpha = \frac{4\pi\kappa}{\lambda}.\tag{3}$$

55 Once the size distribution is calculated accross K channels (bins), the total mass concentration of particles will be

$$c_m = \sum_{i=1}^{K} w_i \rho_i V_i N_i,\tag{4}$$

where V_i is the (average) volume, ρ_i is the density of the particles, N_i is the number of particles per unit volume and w_i is the weighting factor for channel *i*. Here we have another cause of OPS uncertainty: the density of particles must be assumed. Regarding the weighting factors, sensor manufacturers calculate values to correct for certain effects, such as the fact that OPS

60 cannot detect particles which are too small.

70

Laboratory tests and calibrations of OPS are performed under controlled conditions with known particles, such as polystyrene latex spheres (Walser et al., 2017; Bezantakos et al., 2018), continuously changing monodisperse, particles (Kuula et al., 2017, 2020) or multi-modal particles (Cai et al., 2019). Burning chamber is used in some investigations as well (Wang et al., 2015). However, equation (1) is strongly non-linear in terms of refractive index, and in most practical cases corrections for different

65 particles' optical properties are impossible to implement. Furthermore, densities appearing in equation (4) are not known *a priori*. That explains why it is difficult to calibrate OPS for realistic ambient PM concentration measurements: any laboratory calibration may or may not be applicable to the changing outdoor conditions (Tryner et al., 2020; Crilley et al., 2020).

For outdoor applications, there is an additional problem: hygroscopic growth of particles (Jayaratne et al., 2018; Granados-Muñoz et al., 2015; Di Antonio et al., 2018), which leads to overshoots of OPS if the ambient air humidity is (too) high. An obvious solution is to dry the air. However, any proper drying system would cost more than many models of OPS and it is

rarely seen in combination with low-cost sensors. Analytical corrections are often used: humidity sensors are used to measure the relative humidity of ambient air and some analytical model, like Kohler's theory (Castarède and Thomson, 2018) or Hänel equation (Hänel, 1976), is applied. Later in this paper we will make some observations on this issue.

Due to all the above-mentioned factors, it is always interesting to check how OPS perform in different realistic scenarios.

- Numerous papers deal with laboratory calibrations and outdoor evaluations of OPS (Karagulian et al., 2019; Borghi et al., 2018; Chatzidiakou et al., 2019; Magi et al., 2020; Sousan et al., 2016b; Malings et al., 2020; Kelly et al., 2017; Sayahi et al., 2019; Crilley et al., 2018; Zheng et al., 2018; Tasic et al., 2012; Cavaliere et al., 2018; Mukherjee et al., 2017; Sousan et al., 2016a; Zhang et al., 2018; Holstius et al., 2014; Badura et al., 2018). Reported results vary depending on the composition of particulate matter pollution, range of concentrations and meteorological factors. In (Mukherjee et al., 2017) OPC-N2, PMS7003 and 11-R
- 80 were compared against BAM-1020 during 12 weeks in the Cuyama Valley, California, USA. Grimm 11-R performed well, while both OPC-N2 and PMS7003 (which is a miniaturized version of PMS5003) produced mediocre performance with heavy low bias. PurpleAir (PMS5003) was tested in (Tryner et al., 2020) using laboratory and field tests. High bias of PMS5003 was observed. In (Magi et al., 2020) PurpleAir (PMS5003) was analyzed for 16 months in Charlotte, North Carolina, USA against BAM-1022, high bias of PMS5003 that increases with humidity was reported. High mean bias of PurpleAir (PMS5003) was
- 85 reported in (Kosmopoulos et al., 2020) as well.

The novelty of this research is a unique combination of instruments and conditions of extremely high urban pollution. The city of Sarajevo is situated in a valley and is affected by strong temperature inversions that appear typically 150 m-300 m above ground level with a very strong temperature gradient in the inversion layer, exceeding 30 K/km (Masic et al., 2019). The inversion episodes were present during most of January 2020. As a consequence, the average monthly concentration of

90 $PM_{2.5}$ was very high: 167.3 μ g/m³. In contrast to that, monthly average values for March and April 2020 were 21.6 μ g/m³ and 19.6 μ g/m³, respectively. This presented an excellent opportunity to test the performance of OPS in very different pollution

levels. Simultaneously with OPS and reference gravimetric measurements, an SMPS was employed to detect nanoparticles. It can detect particles with diameters from few nm up to 1 μ m. While SMPS can count very small particles, 11-D can count larger particles, from 0.25 μ m to 35 μ m in diameter. When they work simultaneously, they can detect (almost) the full range of

particles' diameters, with a span of more than three orders of magnitude. This will give detailed insights on the mass distribution

95

of particles.

2 Methodology and experimental setup

The experimental facility was located at the Faculty of Mechanical Engineering in the central part of Sarajevo valley (N 43.85424, E 18.39607, 540 m above sea level) and represents well the overall conditions in the city. The reference instrument for measurements of PM concentrations was a Dadolab Gemini air sampler (Figure 1). It is a single device with two completely independent channels (PM_{2.5} and PM₁₀ in this campaign). The filter preparation and gravimetric analysis are performed in separate laboratory of Faculty of Science, Department of Chemistry. The air sampler, gravimetric laboratory and all filter procedures satisfied requirements of the standard EN 12341:2014. According to requirements of the standard, all filters were conditioned at relative humidity between 45% and 50%, and temperature between 19 and 21 ^oC.

- Grimm 11-D is a high-end optical particle sizer, with sophisticated construction and ability to count individual particles from 250 nm to 35 μ m in 31 equidistant (on logarithmic scale) channels. It uses a proprietary algorithm and the manufacturer does not share information about the refractive index, density or weighting factors. It was factory calibrated, and equipped with firmware version 12.50. Data was recorded in 1 minute intervals (6 seconds is also possible). Since we use the common term OPS occasionally, it should be noted that 11-D belongs to different category of devices (in comparison to low-cost sensors).
- Alphasense OPC-N2 belongs to the category of low-cost sensors. The manufacturer transparently shared most specifications. It has a much simpler construction than 11-D: instead of regulated pump, air flow is provided by 25 mm fan. The device has 16 channels, from 380 nm to 17 μ m. Firmware version 18.2 was used. Refractive index was n = 1.50 + i0 and density was 1.65 g/cm³. All other parameters, including weighting factors, were used as firmware default values.

The Plantower PMS5003 could be termed a very low-cost sensor, since its price is lower by an order of magnitude than that of the OPC-N2. Limited specifications don't reveal all operating parameters. From the specification sheet we can conclude that the device uses Mie scattering theory, with detection limit of 300 nm, and has 6 channels. It uses red semiconductor laser, photodetector at 90⁰ scattering angle (Kuula et al., 2020) and 32-bit processor (Cypress CY8C4245, 48 MHz). According to (Tanzer et al., 2019) PMS5003 is a nephelometer, not the particle counter. Air flow is provided by a 20 mm fan. The PMS5003 has two data outputs, one is called SM (standard material, CF = 1), and another AE (atmospheric environment).

120 The latter mode is used in our work, since the manufacturer recommends AE mode for ambient air measurements, without further explanation. Figure 2 shows results from our laboratory test using the incense scents as the source of PM. Based on

these results, the relationships between SM and AE modes are

$$SM_{PM2.5} = \begin{cases} AE_{PM2.5} \text{ for } AE_{PM2.5} \leq 30 \\ \text{nonlinear for } 30 < AE_{PM2.5} \leq 50 \\ 1.5 \times AE_{PM2.5} \text{ for } AE_{PM2.5} > 50 \\ AE_{PM10} \text{ for } AE_{PM10} \leq 43 \\ \text{nonlinear for } 43 < AE_{PM10} \leq 77 \\ 1.5 \times AE_{PM10} \text{ for } AE_{PM10} > 77 \end{cases}$$
(5)

Based on PMS5003, we have designed MAQS (Mobile Air Quality System) smart sensor. Essentially, it is a modular platform for PMS5003, with options for additional sensors (pressure, temperature, humidity, carbon dioxide, wind speed), GNSS receiver, flash memory, Wi-Fi module and 3D-printed enclosure. Eight MAQS sensors were made and tested prior to the main campaign in order to evaluate consistency between units. Figure 1 shows the results of preliminary outdoor measurements for a batch of 8 MAQS sensors. They showed very good consistency: the coefficient of determination, R², between any two
sensors from the batch was greater than 0.99 and average readings from all sensors are within ±10% from the average value of the batch of sensors. Data was recorded every minute on a local SD card and remote cloud server simultaneously. Recording

interval can be as short as 1 second, but there was no need for that.

Grimm 11-D and Alphasense OPC-N2 could not be used outdoors without shelter, while MAQS has a special case which provides basic protection for outdoor use. Furthermore, netbook PC was used to record data from the OPC-N2. Outdoor shelter

135 had to be constructed to accommodate 11-D with power supply, OPC-N2 with PC and SPI adapter, and MAQS (for better protection). Stevenson screen like wooden structure was designed for that purpose. Another MAQS sensor was used at a remote location, for reasons that will be explained later.

For low-cost sensors (OPC-N2 and MAQS) there was no air dryer or heater, since they are typically used in such conditions. We have designed and constructed diffusion dryer for application on 11-D, which consists of porous stainless steel tube

140 surrounded with 1 kg of silica gel. The dryer is compact, 25 cm in length with 8 cm external diameter and does not reduce the mobility of the instrument. It was installed only during the period of mild pollution.

Meteorological parameters were measured using Vantage Pro2 (Davis Instruments, USA) weather station with recording intervals of 15 minutes.

- SMPS is a complex system which consists of a condensation particle counter (CPC), a differential mobility analyzer (DMA) and a charge conditioner (often inadequately called "neutralizer"). Depending on the characteristics of the DMA, the SMPS can be configured for certain span of particle diameters. We have used Grimm 5.416 high-end SMPS with long DMA which is able to separate particles from 10 nm to 1000 nm in 129 channels, equidistant on a logarithmic scale. Despite the fact that particles with diameter below 10 nm play an important role in nucleation and growth studies (Tiszenkel et al., 2019), their
- 150 SMPS. A soft X-ray device was used as the charge conditioner. Scanning mode (alternating upscan and downscan) was used for all measurements. One scan takes about 4 minutes (8 minutes for both upscan and downscan). When working parallelly,

contribution to the mass budget is negligible. Another (larger) in-house developed diffusion dryer was installed at the inlet of

SMPS and 11-D form a powerful wide-range spectrometer, which covers a range of particle diameters from 10 nm to $35 \,\mu$ m in 160 channels. Additionally, there is an overlapping area between 250 nm and 1000 nm where we can see how well these two instruments match. The complex SMPS system was kept indoors (an unavoidable necessity, since both X-ray charger and

DMA use very high operating voltages). The air was sampled from outside using a conductive tube of shortest possible length,

155

to avoid particle losses. It was running continuously, except for the periods of maintenance.

A rigorous data validation procedure was used. All instruments were inspected periodically and data logs were analyzed thoroughly. When calculating daily average values, complete and consistent data series were required.

3 Results and discussion

- 160 During this campaign 296 filters were used in the reference air sampler. After the removal of several blank filters used for periodic verification and those with incomplete sampling (pneumatic system of air sampler failed to load new filters automatically couple of times), 288 filters remained: 143 $PM_{2.5}$ and 145 PM_{10} samples. Figure 2 shows $PM_{2.5}$ and PM_{10} daily average concentrations, together with hourly and daily values of ambient air temperature and relative humidity.
- Some modifications of the shelter for 11-D and OPC-N2 were necessary, making those instruments unavailable periodically during December and January. Additionally, more frequent maintenance, such as cleaning of 11-D, was needed when working in extreme conditions. The same stands for SMPS, which was maintained according to the recommendations of the manufacturer. Taking into account difficult operating conditions, the amount of data collected is satisfactory during the period of strong pollution and excellent during the period of mild pollution. The lower limit of detection (LLoD) of PM_{2.5} concentration for evaluated optical aerosol devices is estimated based on their actual field performance. Standard deviation (*σ*) was calculated
- 170 for periods with near-zero ambient PM concentration and average value of 3σ is estimated LLoD. For PMS5003 our final estimation is 5 µg/m³. The same value is an estimation of (Magi et al., 2020), calculated by averaging segmented regressions and (Bulot et al., 2019) by combining results from several previous studies. This method applied on OPC-N2 yields LLoD of 2 µg/m³ and 1 µg/m³ for 11-D. For reference gravimetric system LLoD was calculated using the blank filters, which were treated exactly the same way as real samples (except the sampling of particulate matter), and the calculated value of LLoD is 0.7 µg/m³. All measurements below LLoD were discarded during the quality assurance phase.

3.1 Strong urban pollution

During the period of strong urban pollution (12/2/2019–3/12/2020), the average value of PM_{2.5} concentration was 82.9 μ g/m³, with minimum daily average value 1.3 μ g/m³ and maximum value 504.9 μ g/m³. In the same period, the average PM₁₀ concentration was 95.5 μ g/m³, with minimum value 3.6 μ g/m³ and maximum value 549.0 μ g/m³. The ratio of average values of

180 concentrations $PM_{2.5}/PM_{10}$ was 0.87. Very good correlations were observed for all three OPS against the reference instrument (Figure 5). Such a range of ambient PM concentrations was favorable for achievement of high R^2 values, but non-linear effects of low-cost sensors were observed too. Grimm 11-D produced results with R^2 values 0.988 and 0.985 for $PM_{2.5}$ and PM_{10} concentrations, respectively. Absolute values were larger than the reference, on average 17.6% for $PM_{2.5}$ and 25.5% for PM_{10} . The average ratio $PM_{2.5}/PM_{10}$

- 185 measured by 11-D was 0.93. Mean absolute error (MAE) was $13.4 \,\mu g/m^3$ for PM_{2.5} and $10.8 \,\mu g/m^3$ for PM₁₀. Alphasense OPC-N2 undershoots with respect to the reference values, on average 31.0% for PM_{2.5} and 36.8% for PM₁₀, but R^2 coefficients are relatively high: 0.903 and 0.920 for PM_{2.5} and PM₁₀ respectively. The OPC-N2 measured the ratio PM_{2.5}/PM₁₀ to be 0.97. MAE for this sensor was $29.4 \,\mu g/m^3$ for PM_{2.5} and $34.8 \,\mu g/m^3$ for PM₁₀. MAQS sensor produced surprisingly good R^2 values of 0.975 for PM_{2.5} and 0.950 for PM₁₀. In terms of absolute values, it overshoots by 31.9% for PM_{2.5} and 49.3% for
- 190 PM_{10} (on average). The calculated ratio $PM_{2.5}/PM_{10}$ was 0.76. MAE was 35.9 μ g/m³ for $PM_{2.5}$ and 55.2 μ g/m³ for PM_{10} . It seems that the Plantower PMS5003 can not accurately determine the PM_{10} fraction. One possible explanation is provided by a laboratory test of PMS5003, where it was found that its size bin [2.5 μ m-10 μ m] is noisy and inaccurate (Kuula et al., 2020). Further investigation of this behavior would be useful.

None of the tested OPS were equipped with an air dryer, and this certainly contributes to overprediction. Yet, Alphasense OPC-N2 with default firmware settings underpredicts values, despite the particle hygroscopic growth effect.

3.2 Mild urban pollution

195

The correlation coefficients changed dramatically in the period of mild pollution (3/13/2020–5/4/2020), as Figure 6 shows. The much narrower range of particulate matter concentrations plays an important role, and even the reference method is less accurate, since the mass difference of loaded and blank filters becomes very small (smaller than 1 mg for 24 h sampling period

- 200 if PM concentration is below $18 \,\mu g/m^3$). The average concentration of PM_{2.5} was $19.7 \,\mu g/m^3$ with a minimum daily average value of 7.1 $\mu g/m^3$ and a maximum value of $39.3 \,\mu g/m^3$. During this period, the average value of PM₁₀ concentration was 24.2 $\mu g/m^3$, with minimum and maximum values of 7.6 $\mu g/m^3$ and 48.8 $\mu g/m^3$, respectively. The ratio PM_{2.5}/PM₁₀ was 0.81 on average.
- This time Grimm 11-D was equipped with a dryer, whose effects will be discussed in the next subsection. The device produced relatively high R^2 values 0.868 for PM_{2.5} and 0.917 for PM₁₀. The absolute readings underestimated concentrations of PM_{2.5} by 16.3% on average, while PM₁₀ were underestimated by 10.9% on average. The PM_{2.5}/PM₁₀ ratio was 0.87. This test clearly shows that 11-D is a completely different class of instrument (in comparison to low-cost sensors). When equipped with dryer, 11-D shows level of performance comparable to BAM, at least those reported by Hafkenscheid and Vonk (2014). MAE was 3.0 μ g/m³ for PM_{2.5} and 4.1 μ g/m³ for PM₁₀.
- Alphasense OPC-N2 did not perform well during the period of mild pollution. Coefficients of determination, R^2 , were only 0.284 for PM_{2.5} and 0.525 for PM₁₀. Absolute readings are worrying: the OPC-N2 underpredicted PM_{2.5} by 67.6% and PM₁₀ by 71.6% on average. The ratio PM_{2.5}/PM₁₀ was 0.73. MAE was 13.8 μ g/m³ for PM_{2.5} and 15.8 μ g/m³ for PM₁₀.

MAQS sensor demonstrated mediocre performance, with R^2 values of 0.730 for PM_{2.5} and 0.718 for PM₁₀. On average, this sensor overpredicted PM_{2.5} by 30.5% and PM₁₀ by 32.6%. The PM_{2.5}/PM₁₀ ratio was 0.83, very close to the reference value

215 (contrary to the performance of the sensor in the period of strong pollution). MAE was $7.1 \,\mu\text{g/m}^3$ for PM_{2.5} and $8.2 \,\mu\text{g/m}^3$ for PM₁₀.

It would be interesting to test low-cost sensors with a proper dryer as well, but that combination is rarely seen in practice.

3.3 Humidity influence

One of the important factors in ambient measurements of PM concentrations is humidity, since the particles reflect more light (i.e. appear larger) during measurements due to hygroscopic growth. This can be described using Hänel equation:

$$f_{\zeta}(RH) = \left(\frac{1 - RH}{1 - RH_{\text{ref}}}\right)^{-\gamma},\tag{6}$$

where f_{ζ} is enhancement factor for particle property ζ . Here *RH* represents the relative humidity and *RH*_{ref} is a reference relative humidity:

$$f_{\zeta}(RH) = \frac{\zeta(RH)}{\zeta(RH_{\text{ref}})}.$$
(7)

225 It is important to note that the coefficient γ , which is an indicator of the hygroscopicity of particles, depends on the type of particles (and changes whenever composition of ambient particles is changed).

If we compare results produced by 11-D relative to the reference, during period 1 (without dryer) and period 2 (with dryer), we can see that readings of the 11-D were reduced by more than 30%. However, we can not conclude whether it was the effect of the dryer or the consequence of significantly different ambient conditions. Unfortunately, we have only one 11-D,

- so we couldn't measure simultaneously with and without dryer (that's the reason why we used the instrument with the dryer only in one of the two periods). If we take into account two intervals with similar ambient conditions, with and without dryer we get following values: from 2/27/2020 to 3/12/2020 average ambient concentration of PM_{2.5} was 21.1 μ g/m³ while 11-D (without dryer) measured 21.5% more. In the second interval, from 3/13/2020 to 4/1/2020, ambient concentration was similar, 21.0 μ g/m³ while 11-D (with dryer) measured 1.4% smaller value. This comparison indicates that the effect of the dryer could
- 235 be around 23%. A similar analysis for PM₁₀ concentrations gives an estimate of about 20% for the dryer effect. Grimm 11-D has a very useful feature: internal temperature and humidity sensor. Figure 7 shows self-heating and diffusion dryer effect on 11-D, by comparing internal and external measurements of temperature and humidity. The average ambient air temperature from 2/27/2020 to 4/1/2020 was 7.02 °C while the average 11-D internal temperature was 14.27 °C, which shows a significant difference of 7.25 °C. This self-heating effect reduces internal humidity significantly, and we can see that it
- rarely goes beyond 50%. Once the dryer is installed, internal relative humidity is further reduced: the average value of internal humidity without dryer (2/27/2020-3/12/2020) was 36.2% and with dryer (3/13/2020-4/1/2020) was 21.8% (the ambient air humidity also dropped in the later period, but nevertheless the effect of the dryer is evident).

After roughly a month, the dryer's performance degraded and the silica gel needed a regeneration (but it wasn't performed since we didn't want to interrupt measurements when the end of the campaign was near).

Figure 8 shows the long-term (13.5 months) comparison of MAQS and BAM-1020 with time resolution of 1 hour, together with measured values of ambient air humidity. By averaging all this data we can estimate the influence of humidity on the MAQS sensor: if we sort the measurements by humidity, subset of points where humidity is below 50% has average bias of 14.3%, for humidity range 50%-70%, bias is 16.5%; for humidity range 70%-85% bias is 31.6% and for humidity rang 85%-

100% bias is 37.3%. If we subtract bias of least humidity subset from bias of highest humidity subset, we can estimate that humidity influence adds up to 23% on PM2.5 readings from MAQS sensor, which is similar result to the analysis of humidity influence on our 11-D with dryer installed. While this influence can not be neglected, it is still relatively modest. Reason for this is the composition of particles, where we have mostly fine particles below 300 nm, for which hygroscopic growth is less pronounced (Kosmopoulos 2020).

3.4 SMPS data and wide-range spectrometer

- 255 The wide-range spectrometer (SMPS+11D) produced very valuable results. Figure 6 shows the continuous concentration and mass distributions. It is created from hourly average measurements from SMPS and 11-D. A relative density of 1.65 was applied in SMPS software (based on LabVIEW) for the mass calculation. No other corrections were performed and all settings were factory defaults. Selected histograms (hourly average values) are shown in Figure 7. What we can see from figures 6 and 7 is that in the period of strong pollution the dominant mass contribution comes from particles with diameters around 300 nm.
 260 In terms of concentrations, particles around 100 nm appear in greatest numbers, with occasional secondary peaks coming from
- even smaller particles.

In the period of mild pollution, however, we can see that particles larger than $2.5 \,\mu$ m often appear on histograms (usually about 3 μ m in diameter). Number concentrations still have peaks about 100 nm, but sometimes the distribution is different in favor of even smaller particles, as Figure 7 shows. Again, the largest mass contribution comes from particles around 300 nm.

265

In the overlapping area, SMPS and 11-D matched very well, almost perfectly for concentrations. Their match was not as good for mass calculations, but that is understandable, taking into account all the factors explained in section 1. Overall, the combination of SMPS and 11-D worked very well and gave the full spectrum of particles, both for number concentrations and mass distribution.

The obtained mass distribution of particles, especially during the period of strong pollution, rises the question on suitability of OPS for measurements of mass concentrations, and resolving different fractions, since they cannot detect small particles that significantly contribute to the total mass. For example, the Alphasense OPC-N2 has a detection limit of 380 nm, and is likely to miss the particles around 300 nm which form the dominant contribution to the mass budget. The Grimm 11-D, with a detection limit of 250 nm, has a far better potential to resolve mass fractions.

3.5 OPS histograms and Aralkum Desert dust

All tested OPS have data bins, with different number of channels, as described in section 2. Figure 11 shows histograms that compare data bins from 11-D, OPC-N2 and MAQS on 1/18/2020 (strong pollution) and 4/16/2020 (mild pollution). It should be noted that we compare here data bins from devices with different specifications and category. As expected, 11-D has ability to count particles below 300 nm, which appear in greatest numbers. Counting efficiency of OPC-N2 is investigated in laboratory conditions using PSL particles in (Sousan et al., 2016a), and the results were good for particles larger than 0.8 μm while for particles with diameter of 0.5 μm OPC-N2 the device showed lower detection efficiency (detection limit of OPC-N2 is 0.38)

 μ m). In our realistic scenario, dominant contribution to the mass comes from particles much smaller than 0.8 μ m (Figures 9 and 10) which is not favorable to OPC-N2.

Contrary to OPC-N2, PMS5003 has problems with coarse particle, as indicated in laboratory test (Kuula et al., 2020). If the fraction of coarse particles is small and steady, PMS5003 performs much better. Ambient conditions in Bosnia-Herzegovina are such most of the time, since the primary source of PM is combustion of coal and biomass. That could explain why PMS5003 285 performs better than OPC-N2 most of the time. However, different conditions were observed on 3/27/2000 when the dust from Aralkum Desert covered part of Europe, including our test location. During this episode, OPC-N2 performed much better than PMS5003, which wasn't able to determine large fraction of coarse particles correctly (Figure 11). Similar observation about PMS5003 was reported by (Kosmopoulos et al., 2020), when Sahara dust covered Greece.

3.6 Long-term performance 290

Another question about OPS, especially low-cost types, is the drift of performance over time. The PMS5003 sensor uses a semiconductor laser (diode laser) which has a limited lifetime. We have some long-term comparisons of the MAQS sensor with MetOne BAM-1020 operated at a nearby location by the US EPA. Strictly speaking, their station is not collocated with our equipment, but for the distance of only 300 m it is reasonable to assume that the air composition is very similar at these two

- points, since they are located in the same neighborhood. In order to verify that assumption, we have installed another MAOS 295 sensor at the location of Faculty of Electrical Engineering, University of Sarajevo, which is in immediate vicinity of the US EPA site, and at the same distance from us (about 300 m). Figure 8 shows long-term comparisons of MAOS sensor and BAM-1020, and additional verification of correlation between readings of two MAOS sensors, which was very high $(R^2 = 0.970, R^2)$ $MAE = 4.7 \mu g/m^3$ for hourly average values and $R^2 = 0.994$, $MAE = 2.9 \mu g/m^3$ for daily average values) confirming our assumption that these two locations share the same air, in terms of PM concentrations and properties.
- 300

Based on 13.5 months of continuous comparison of MAOS and BAM-1020, hourly average values give R^2 coefficient 0.919 and MAE 16.7 μ g/m³. Daily average values produce R^2 coefficient 0.980 and MAE 12.2 μ g/m³, while the monthly average values give $R^2 = 0.998$ and MAE = $11.4 \,\mu$ g/m³ (Figure 8).

- This leads us to the conclusion that time averaging reduces a lot the influence of variation of PM composition and meteorological variations. If we use a longer time average period, we lose one of the major advantages of low-cost sensors (time 305 resolution), but it is a more natural approach to correcting readings compared to using artificial algorithms like neural networks (Badura et al., 2019) or machine learning (Si et al., 2020). An excellent viewpoint of this issue is given by Hagler et al. (2018). The calibration of a larger number of low-cost sensors can be simplified if they show similar relative performance (to each other) in laboratory and field (Sousan et al., 2018). Floating corrections, even physically justifiable interventions, such as the
- instantaneous correction for humidity growth of particles, insert a lot of noise, and the benefit is questionable. Depending on 310 ambient conditions, self-heating of the sensor and some other factors, relative humidity may not be accurately determined. Even if we have a very accurate humidity measurement, the hygroscopic growth coefficient will change whenever the composition of PM changes, inevitably injecting noise into the results.

We can also see strong non-linear effects at very high concentrations, above $500 \,\mu g/m^3$. In that case a quadratic regression fit will be more suitable.

During this period of 13.5 months of continuous outdoor operation, the MAQS sensor worked flawlessly without performance drifts. Designed enclosure sufficiently protected the sensor outdoors, while not obstructing air sampling.

4 Conclusions

A comprehensive experimental study was carried out with the aim of evaluating the performance of three very different OPS: high-end Grimm 11-D, low-cost Alphasense OPC-N2 and in-house developed MAQS sensor, which is based on another lowcost sensor, the Plantower PMS5003 sensor. The study was performed in realistic conditions of strong and mild urban pollution. The reference instrument was a dual-channel air sampler with gravimetric analysis in separate laboratory. In total 288 filters were collected from 12/2/2019 to 5/4/2020.

During the period of strong urban pollution all three instruments produced very high R^2 values. However, during the period of mild urban pollution, these correlation factors dropped significantly, especially for Alphasense OPC-N2 sensor measuring PM2.5 parameter. The OPC-N2 underestimated the mass concentrations badly, especially during the period of mild pollution. MAQS sensor overshoots PM_{2.5} concentrations by approximately 30% on average, which is likely to be partially caused by hygroscopic growth.

The wide-range spectrometer, which consists of SMPS and 11-D, produced valuable information about distribution of particles, both in number and mass concentrations. Particles with diameters around 100 nm (and sometimes below) represent the dominant fraction in pure numbers, while particles with diameter of around 300 nm give the highest contribution to mass. In the period of mild pollution, particles larger then 2.5 μ m gave a larger contribution than in the period of strong pollution.

Grimm 11-D performed well in all conditions, and when equipped with dryer, it performed at a comparable level to the beta attenuation monitor. For the calibration of low-cost sensors, especially those based on PMS5003, we propose a linear or quadratic correction (in case of high pollution levels) with steady coefficients, since the instantaneous corrections insert noise

335

into results.

Future measurements should further investigate characteristics of OPS in different ambient conditions, influence of humidity and effect of micro-dryers specifically designed for low-cost sensors and mass distributions by means of wide-range spectrometer.

340 Data availability. The underlying datasets for this publication are available at https://doi.org/10.5281/zenodo.3897379 Furthermore, data from the BAM measurements by US EPA are available at https://cfpub.epa.gov/airnow *Author contributions*. AM participated in all phases of this research and wrote the manuscript with contributions from all co-authors. DB and BP performed field work together with AM. AB designed, manufactured and analysed the performance of two diffusion dryers and evaluated the influence of humidity on the readings of 11-D. SZ and JH performed gravimetric measurements, including preconditioning and treatment of the filters, ensuring strict fulfilment of the standard EN 12341:2014

345

Competing interests. The authors declare that they have no conflict of interest.

Acknowledgements. We would like to thank to the Embassy of Sweden in Bosnia-Herzegovina for supporting this research and United States Environmental Protection Agency for sharing data publicly.

References

- 350 Badura, M., Batog, P., Drzeniecka-Osiadacz, A., and Modzel, P.: Evaluation of low-cost sensors for ambient PM2.5 monitoring, Journal of Sensors, https://doi.org/10.1155/2018/5096540, 2018.
 - Badura, M., Batog, P., Drzeniecka-Osiadacz, A., and Modzel, P.: Regression methods in the calibration of low-cost sensors for ambient particulate matter measurements, SN Applied Sciences, 1, https://doi.org/10.1007/s42452-019-0630-1, 2019.
- Bezantakos, S., Schmidt-Ott, F., and Biskos, G.: Performance evaluation of the cost-effective and lightweight Alphasense
 optical particle counter for use onboard unmanned aerial vehicles, Aerosol Science and Technology, 52, 385–392, https://doi.org/10.1080/02786826.2017.1412394, 2018.
 - Borghi, F., Spinazze, A., Campagnolo, D., Rovelli, S., Cattaneo, A., and Cavallo, D. M.: Precision and accuracy of a direct-reading miniaturized monitor in PM2.5 exposure assessment, Sensors, 18, https://doi.org/10.3390/s18093089, 2018.
 - Bulot, F. M. J., Johnston, S. J., Basford, P. J., Easton, N. H. C., Apetroaie-Cristea, M., Foster, G. L., Morris, A. K. R., Cox, S. J., and Loxham,
- 360 M.: Long-term field comparison of multiple low-cost particulate matter sensors in an outdoor urban environment, Scientific Reports, 9, https://doi.org/10.1038/s41598-019-43716-3, 2019.
 - Cai, C., Stebounova, L. V., Peate, D. W., and Peters, T. M.: Evaluation of a Portable Aerosol Collector and Spectrometer to measure particle concentration by composition and size, Aerosol Science and Technology, 53, 675–687, https://doi.org/10.1080/02786826.2019.1600654, 2019.
- 365 Castarède, D. and Thomson, E. S.: A thermodynamic description for the hygroscopic growth of atmospheric aerosol particles, Atmospheric Chemistry and Physics, 18, 14 939–14 948, https://doi.org/10.5194/acp-18-14939-2018, 2018.
 - Cavaliere, A., Carotenuto, F., Di Gennaro, F., Gioli, B., Gualtieri, G., Martelli, F., Matese, A., Toscano, P., Vagnoli, C., and Zaldei, A.: Development of Low-Cost Air Quality Stations for Next Generation Monitoring Networks: Calibration and Validation of PM2.5 and PM10 Sensors, Sensors, 18, https://doi.org/10.3390/s18092843, 2018.
- 370 Chatzidiakou, L., Krause, A., Popoola, O. A. M., Di Antonio, A., Kellaway, M., Han, Y., Squires, F. A., Wang, T., Zhang, H., Wang, Q., Fan, Y., Chen, S., Hu, M., Quint, J. K., Barratt, B., Kelly, F. J., Zhu, T., and Jones, R. L.: Characterising low-cost sensors in highly portable platforms to quantify personal exposure in diverse environments, Atmospheric Measurement Techniques, 12, 4643–4657, https://doi.org/10.5194/amt-12-4643-2019, 2019.
- Crilley, L. R., Shaw, M., Pound, R., Kramer, L. J., Price, R., Young, S., Lewis, A. C., and Pope, F. D.: Evaluation of a low-cost
- optical particle counter (Alphasense OPC-N2) for ambient air monitoring, Atmospheric Measurement Techniques, 11, 709–720, https://doi.org/10.5194/amt-11-709-2018, 2018.
 - Crilley, L. R., Singh, A., Kramer, L. J., Shaw, M. D., Alam, M. S., Apte, J. S., Bloss, W. J., Hildebrandt Ruiz, L., Fu, P., Fu, W., Gani, S., Gatari, M., Ilyinskaya, E., Lewis, A. C., Ng'ang'a, D., Sun, Y., Whitty, R. C. W., Yue, S., Young, S., and Pope, F. D.: Effect of aerosol composition on the performance of low-cost optical particle counter correction factors, Atmospheric Measurement Techniques, 12, 1181, 1102, https://doi.org/10.5104/cmt.12, 1181, 2020, 2020.
- 380 13, 1181–1193, https://doi.org/10.5194/amt-13-1181-2020, 2020.
 - Di Antonio, A., Popoola, O. A. M., Ouyang, B., Saffell, J., and Jones, R. L.: Developing a relative humidity correction for low-cost sensors measuring ambient particulate matter, Sensors, 18, https://doi.org/10.3390/s18092790, 2018.
 - Downward, G. S., van Nunen, E. J. H. M., Kerckhoffs, J., Vineis, P., Brunekreef, B., Boer, J. M. A., Messier, K. P., Roy, A., Verschuren, W. M. M., van der Schouw, Y. T., Sluijs, I., Gulliver, J., Hoek, G., and Vermeulen, R.: Long-term exposure to ultrafine particles and

385 incidence of cardiovascular and cerebrovascular disease in a prospective study of a Dutch cohort, Environmental Health Perspectives, 126, https://doi.org/10.1289/EHP3047, 2018.

ECWG: Guide to the demonstration of equivalence of ambient air monitoring methods, 2010.

410

- Granados-Muñoz, M. J., Navas-Guzmán, F., Bravo-Aranda, J. A., Guerrero-Rascado, J. L., Lyamani, H., Valenzuela, A., Titos, G., Fernández-Gálvez, J., and Alados-Arboledas, L.: Hygroscopic growth of atmospheric aerosol particles based on active remote sensing and radiosound-
- 390 ing measurements: selected cases in southeastern Spain, Atmospheric Measurement Techniques, 8, 705–718, https://doi.org/10.5194/amt-8-705-2015, 2015.
 - Hafkenscheid, T. L. and Vonk, J.: Evaluation of equivalence of the MetOne BAM-1020 for the measurement of PM2.5 in ambient air, 2014.
 Hagler, G. S. W., Williams, R., Papapostolou, V., and Polidori, A.: Air Quality Sensors and Data Adjustment Algorithms: When Is It No Longer a Measurement?, Environmental Science & Technology, 52, 5530–5531, https://doi.org/10.1021/acs.est.8b01826, 2018.
- 395 Hänel, G.: The properties of atmospheric aerosol particles as functions of the relative humidity at thermodynamic equilibrium with the surrounding moist air, vol. 19 of *Advances in Geophysics*, pp. 73–188, Elsevier, https://doi.org/10.1016/S0065-2687(08)60142-9, 1976.
 - Holstius, D. M., Pillarisetti, A., Smith, K. R., and Seto, E.: Field calibrations of a low-cost aerosol sensor at a regulatory monitoring site in California, Atmospheric Measurement Techniques, 7, 1121–1131, https://doi.org/10.5194/amt-7-1121-2014, 2014.

Jayaratne, R., Liu, X., Thai, P., Dunbabin, M., and Morawska, L.: The influence of humidity on the performance of a low-cost air particle

- 400 mass sensor and the effect of atmospheric fog, Atmospheric Measurement Techniques, 11, 4883–4890, https://doi.org/10.5194/amt-11-4883-2018, 2018.
 - Karagulian, F., Barbiere, M., Kotsev, A., Spinelle, L., Gerboles, M., Lagler, F., Redon, N., Crunaire, S., and Borowiak, A.: Review of the Performance of Low-Cost Sensors for Air Quality Monitoring, Atmosphere, 10, https://doi.org/10.3390/atmos10090506, 2019.
- Kelly, K. E., Whitaker, J., Petty, A., Widmer, C., Dybwad, A., Sleeth, D., Martin, R., and Butterfield, A.: Ambient and laboratory evaluation
 of a low-cost particulate matter sensor, Environmental Pollution, 221, 491–500, https://doi.org/10.1016/j.envpol.2016.12.039, 2017.
- Kosmopoulos, G., Salamalikis, V., Pandis, S. N., Yannopoulos, P., Bloutsos, A. A., and Kazantzidis, A.: Low-cost sensors for measuring airborne particulate matter: Field evaluation and calibration at a South-Eastern European site, Science of The Total Environment, 748, 141 396, https://doi.org/10.1016/j.scitotenv.2020.141396, 2020.

Kuula, J., Makela, T., Hillamo, R., and Timonen, H.: Response characterization of an inexpensive aerosol sensor, Sensors, 17, https://doi.org/10.3390/s17122915, 2017.

- Kuula, J., Mäkelä, T., Aurela, M., Teinilä, K., Varjonen, S., González, O., and Timonen, H.: Laboratory evaluation of particle-size selectivity of optical low-cost particulate matter sensors, Atmospheric Measurement Techniques, 13, 2413–2423, https://doi.org/10.5194/amt-13-2413-2020, 2020.
- Li, H. Z., Gu, P., Ye, Q., Zimmerman, N., Robinsona, E. S., Subramanian, R., Apte, J. S., Robinsona, A. L., and Presto, A. A.: Spatially
- 415 dense air pollutant sampling: Implications of spatial variability on the representativeness of stationary air pollutant monitors, Atmospheric Environment: X, 2, https://doi.org/10.1016/j.aeaoa.2019.100012, 2019.
 - Magi, B. I., Cupini, C., Francis, J., Green, M., and Hauser, C.: Evaluation of PM2.5 measured in an urban setting using a low-cost optical particle counter and a Federal Equivalent Method Beta Attenuation Monitor, Aerosol Science and Technology, 54, 147–159, https://doi.org/10.1080/02786826.2019.1619915, 2020.
- 420 Malings, C., Tanzer, R., Hauryliuk, A., Saha, P. K., Robinson, A. L., Presto, A. A., and Subramanian, R.: Fine particle mass monitoring with low-cost sensors: Corrections and long-term performance evaluation, Aerosol Science and Technology, 54, 160–174, https://doi.org/10.1080/02786826.2019.1623863, 2020.

- Martin, R. V., Brauer, M., van Donkelaar, A., Shaddick, G., Narain, U., and Dey, S.: No one knows which city has the highest concentration of fine particulate matter, Atmospheric Environment: X, 3, https://doi.org/10.1016/j.aeaoa.2019.100040, 2019.
- 425 Masic, A., Bibic, D., Pikula, B., Dzaferovic-Masic, E., and Musemic, R.: Experimental study of temperature inversions above urban area using unmanned aerial vehicle, Thermal Science, 23, 3327–3338, https://doi.org/10.2298/TSCI180227250M, 2019.
 - Mie, G.: Beiträge zur Optik trüber Medien, speziell kolloidaler Metallösungen (contributions to the optics of diffuse media, especially colloid metal solutions, Annalen der Physik, 25, 377–445, https://doi.org/10.1002/andp.19083300302, 1908.
 - Morawska, L., Thai, P. K., Liu, X., Asumadu-Sakyi, A., Ayoko, G., Bartonova, A., Bedini, A., Chai, F., Christensen, B., Dunbabin, M.,
- 430 Gao, J., Hagler, G. S. W., Jayaratne, R., Kumar, P., Lau, A. K. H., Louie, P. K. K., Mazaheri, M., Ning, Z., Motta, N., Mullins, B., Rahman, M. M., Ristovski, Z., Shafiei, M., Tjondronegoro, D., Westerdahl, D., and Williams, R.: Applications of low-cost sensing technologies for air quality monitoring and exposure assessment: How far have they gone?, Environment International, 116, 286–299, https://doi.org/10.1016/j.envint.2018.04.018, 2018.
- Mukherjee, A., Stanton, L. G., Graham, A. R., and Roberts, P. T.: Assessing the utility of low-cost particulate matter sensors over a 12-week
 period in the Cuyama Valley of California, Sensors, 17, https://doi.org/10.3390/s17081805, 2017.
 - Mukherjee, A., Brown, S. G., McCarthy, M. C., Pavlovic, N. R., Stanton, L. G., Snyder, J. L., D'Andrea, S., and Hafner, H. R.: Measuring Spatial and Temporal PM2.5 Variations in Sacramento, California, Communities Using a Network of Low-Cost Sensors, Sensors, 19, https://doi.org/10.3390/s19214701, 2019.
 - Sayahi, T., Butterfield, A., and Kelly, K. E.: Long-term field evaluation of the Plantower PMS low-cost particulate matter sensors, Environmental Pollution, 245, 932–940, https://doi.org/10.1016/j.envpol.2018.11.065, 2019.
 - Si, M., Xiong, Y., Du, S., and Du, K.: Evaluation and calibration of a low-cost particle sensor in ambient conditions using machine-learning methods, Atmospheric Measurement Techniques, 13, 1693–1707, https://doi.org/10.5194/amt-13-1693-2020, 2020.
 - Sousan, S., Koehler, K., Hallett, L., and Peters, T. M.: Evaluation of the Alphasense optical particle counter (OPC-N2) and the Grimm portable aerosol spectrometer (PAS-1.108), Aerosol Science and Technology, 50, 1352–1365, https://doi.org/10.1080/02786826.2016.1232859, 2016a.
- 445 2016a.

440

- Sousan, S., Koehler, K., Thomas, G., Park, J. H., Hillman, M., Halterman, A., and Peters, T. M.: Inter-comparison of lowcost sensors for measuring the mass concentration of occupational aerosols, Aerosol Science and Technology, 50, 462–473, https://doi.org/10.1080/02786826.2016.1162901, 2016b.
- Sousan, S., Gray, A., Zuidema, C., Stebounova, L., Thomas, G., Koehler, K., and Peters, T.: Sensor selection to improve estimates of
 particulate matter concentration from a low-cost network, Sensors, 18, https://doi.org/10.3390/s18093008, 2018.
- Tanzer, R., Malings, C., Hauryliuk, A., Subramanian, R., and Presto, A. A.: Demonstration of a low-cost multi-pollutant network to quantify intra-urban spatial variations in air pollutant source impacts and to evaluate environmental justice, International Journal of Environmental Research and Public Health, 16, https://doi.org/10.3390/ijerph16142523, 2019.
- Tasic, V., Jovasevic-Stojanovic, M., Vardoulakis, S., Milosevic, N. Kovacevic, R., and Petrovic, J.: Comparative assessment of a real-time
 particle monitor against the reference gravimetric method for PM10 and PM2.5 in indoor air, Atmospheric Environment, 54, 358–364, https://doi.org/10.1016/i.atmosenv.2012.02.030, 2012.
 - Tiszenkel, L., Stangl, C., Krasnomowitz, J., Ouyang, Q., Yu, H., Apsokardu, M. J., Johnston, M. V., and Lee, S.-H.: Temperature effects on sulfuric acid aerosol nucleation and growth: initial results from the TANGENT study, Atmospheric Chemistry and Physics, 19, 8915–8929, https://doi.org/10.5194/acp-19-8915-2019, 2019.

- 460 Tryner, J., L'Orange, C., Mehaffy, J., Miller-Lionberg, D., Hofstetter, J. C., Wilson, A., and Volckens, J.: Laboratory evaluation of low-cost PurpleAir PM monitors and in-field correction using co-located portable filter samplers, Atmospheric Environment, 220, https://doi.org/10.1016/j.atmosenv.2019.117067, 2020.
 - Walser, A., Sauer, D., Spanu, A., Gasteiger, J., and Weinzierl, B.: On the parametrization of optical particle counter response including instrument-induced broadening of size spectra and a self-consistent evaluation of calibration measurements, Atmospheric Measurement Techniques, 10, 4341–4361, https://doi.org/10.5194/amt-10-4341-2017, 2017.
 - Wang, Y., Li, J., Jing, H., Zhang, Q., Jiang, J., and Biswas, P.: Laboratory evaluation and calibration of three low-cost particle sensors for particulate matter measurement, Aerosol Science and Technology, 49, 1063–1077, https://doi.org/10.1080/02786826.2015.1100710, 2015.

465

470

Zhang, J., Marto, J. P., and Schwab, J. J.: Exploring the applicability and limitations of selected optical scattering instruments for PM mass measurement, Atmospheric Measurement Techniques, 11, 2995–3005, https://doi.org/10.5194/amt-11-2995-2018, 2018.

- Zhao, A., Bollasina, M. A., Crippa, M., and Stevenson, D. S.: Significant climate impacts of aerosol changes driven by growth in energy use and advances in emission control technology, Atmospheric Chemistry and Physics, 19, 14517–14533, https://doi.org/10.5194/acp-19-14517-2019, 2019.
 - Zheng, T., Bergin, M. H., Johnson, K. K., Tripathi, S. N., Shirodkar, S., Landis, M. S., Sutaria, R., and Carlson, D. E.: Field evaluation of
- low-cost particulate matter sensors in high- and low-concentration environments, Atmospheric Measurement Techniques, 11, 4823–4846, https://doi.org/10.5194/amt-11-4823-2018, 2018.
 - Zheng, T., Bergin, M. H., Sutaria, R., Tripathi, S. N., Caldow, R., and Carlson, D. E.: Gaussian process regression model for dynamically calibrating and surveilling a wireless low-cost particulate matter sensor network in Delhi, Atmospheric Measurement Techniques, 12, 5161–5181, https://doi.org/10.5194/amt-12-5161-2019, 2019.



(a) Air sampler and Stevenson screen.

(b) Devices under test.



(c) SMPS with dryer.

Figure 1. Experimental setup: a) colocated air sampler and Stevenson screen, b) devices under test inside of Stevenson screen: 11-D with dryer, OPC-N2 with SPI adapter and (white-orange) enclosure, MAQS (white enclosure with grey front panel), and c) indoors SMPS with dryer.



Figure 2. AE and SM modes of PMS5003 sensor.



Figure 3. Preliminary test of 8 MAQS sensors, outdoor measurements.



PM concentrations - reference method (daily average values)

Figure 4. Reference PM concentrations with ambient air temperature and humidity.



Figure 5. OPS performance during the period of strong pollution (12/2/2019-3/12/2020).



Figure 6. OPS performance during the period of mild pollution (3/13/2020–5/4/2020).



Self heating of Grimm 11-D, hourly average temperature

Figure 7. Self-heating and diffusion dryer effect of Grimm 11-D. Dryer was installed on 3/12/2020 15:30.



Figure 8. Long-term comparisons of MAQS sensor with BAM operated by US EPA at the nearby location: hourly, daily, monthly average values and comparison of hourly and daily average values of two MAQS sensors: first one (MAQS) at our main facility and second one (MAQS-FEE) at Faculty of Electrical Engineering in the immediate vicinity of BAM-1020.



Figure 9. Wide-range spectrometer (SMPS+11D), hourly average values. Relative density 1.65 was applied on SMPS to calculate mass of particles.



Figure 10. Wide-range histograms, hourly average values. Relative density 1.65 was applied on SMPS to calculate mass of particles.



Figure 11. OPS histograms and Aralkum Desert dust episode, hourly average values.