

Introducing the Extended Volatility Range Proton-Transfer-Reaction Mass Spectrometer (EVR PTR-MS)

Felix Piel^{1,2,+}, Markus Müller¹, Klaus Winkler¹, Jenny Skytte af Sättra^{3,*} and Armin Wisthaler^{2,3}

5 ¹ IONICON Analytik, Innsbruck, Austria

² Institute for Ion Physics and Applied Physics, University of Innsbruck, Innsbruck, Austria

³ Department of Chemistry, University of Oslo, Oslo, Norway

⁺ Now at: Department of Chemistry, University of Oslo, Oslo, Norway

^{*} Now at: Norwegian Environment Agency, Oslo, Norway

10 *Correspondence to:* Armin Wisthaler (armin.wisthaler@kjemi.uio.no)

Abstract. Proton-transfer-reaction mass spectrometry (PTR-MS) is widely used in atmospheric sciences for measuring volatile organic compounds in real time. In the most widely used type of PTR-MS instruments, air is directly introduced into a chemical ionization reactor via an inlet capillary system. The reactor has a volumetric exchange time of ~ 0.1 s enabling PTR-MS analyzers to measure at a frequency of 10 Hz. The time response does, however, deteriorate if low-volatility analytes interact with surfaces in the inlet or in the instrument. Herein, we present the “Extended Volatility Range” (EVR) PTR-MS instrument which mitigates this issue. In the EVR configuration, inlet capillaries are made of passivated stainless steel and all wetted metal parts in the chemical ionization reactor are surface-passivated with a functionalized hydrogenated amorphous silicon coating. Heating the entire set-up (up to 120°C) further improves the time-response performance.

We carried out time-response performance tests on a set of 29 analytes having saturation mass concentrations C^0 in the range between 10^{-3} and $10^5 \mu\text{g m}^{-3}$. $1/e$ -signal decay times after instant removal of the analyte from the sampling flow were between 0.2 and 90 s for gaseous analytes. We also tested the EVR PTR-MS instrument in combination with the CHARON particle inlet, and $1/e$ -signal decay times were in the range between 5 and 35 s for particulate analytes. We show on a set of exemplary compounds that the time-response performance of the EVR PTR-MS instrument is comparable to that of fastest flow tube chemical ionization mass spectrometers that are currently in use. The fast time response can be used for rapid (~ 1 min equilibration time) switching between gas and particle measurements. The CHARON EVR PTR-MS instrument can thus be used for real-time monitoring of both gaseous and particulate organics in the atmosphere. Finally, we show that the CHARON EVR PTR-MS instrument also rapidly detects highly oxygenated species (with up to eight oxygen atoms) in particles formed by limonene ozonolysis.

1 Introduction

The Earth's atmosphere contains a plethora of organic compounds, both in the gas and in the particulate phase (Goldstein and Galbally, 2007). Atmospheric organic compounds vary widely in their physico-chemical properties (*e.g.*, volatility, polarity, solubility), which makes their comprehensive measurement challenging (Heald and Kroll, 2020).

35 Proton-transfer-reaction mass spectrometry (PTR-MS) is widely used in atmospheric sciences for measuring volatile organic compounds (Hansel et al., 1995; de Gouw and Warneke, 2007; Yuan et al., 2017). In PTR-MS, air is directly introduced into an ion-molecule reactor wherein organic molecules ionize in collisions with hydronium (H_3O^+) ions. An electric field is applied across the ion-molecule-reactor, which is thus commonly referred to as the drift tube. Reagent and analyte ions are extracted from the drift tube and analyzed in a mass spectrometer.

40 One of the main advantages of PTR-MS is its rapidness. The drift tube has a volumetric exchange time of ~ 0.1 s enabling PTR-MS analyzers to measure at a frequency of 10 Hz (Müller et al., 2010). The time response does, however, deteriorate if low-volatility analytes interact with surfaces in the inlet or in the instrument. PTR-MS users have mitigated this problem by i) operating the drift tube at elevated temperature (Mikoviny et al., 2010), ii) increasing the flow through the drift tube (Breitenlechner et al., 2017; Krechmer et al., 2018) and iii) minimizing the wall collisions of analyte molecules (Breitenlechner et al., 2017). The problem of analyte adsorption becomes even more pronounced when particles are analyzed with the *Chemical Analysis of Aerosols Online* (CHARON) inlet (Eichler et al., 2015; Müller et al., 2017). After particle vaporization, low-volatility gases adsorb onto surfaces in the vaporizer, in the transfer line from the vaporizer to the drift tube and in the drift tube itself. This slows down the instrumental response significantly (Piel et al., 2019).

A second major advantage of PTR-MS is that H_3O^+ ion chemistry detects a wide spectrum of organic analytes, from non-oxidized to highly oxidized species. It must, however, be ensured that the analyte molecules reach the ionization region and are not lost in the inlet line (*e.g.* Pagonis et al., 2017; Deming et al., 2019) or onto the drift tube walls.

Herein we will demonstrate how the use of heated inlet capillaries made of passivated stainless steel (SS) and of a heated drift tube with passivated metal surfaces significantly improves the time response performance of PTR-MS analyzers. We will show that the heated and passivated instrument responds fast to low-volatility analytes, both for gas-phase and particle-phase measurements. The instrument set-up described herein has thus been named “*Extended Volatility Range*” (EVR) configuration.

2 Experimental

2.1 The EVR PTR-MS instrument

The PTR-MS instrument has been described in detail elsewhere (Yuan et al., 2017; and references therein). The data presented herein were obtained with two state-of-the-art CHARON PTR-MS analyzers (models PTR-TOF 4000X2 and PTR-TOF 6000X2) produced by Ionicon Analytik (Innsbruck, Austria). In their conventional set-up, these analyzers include inlet capillaries made of polyether ether ketone (PEEK) and a drift tube plus ion funnel consisting of electropolished SS drift rings and polytetrafluoroethylene (PTFE) spacers. PEEK and especially SS are known to adsorb certain analytes. In an effort to optimize the instrumental time response, we have eliminated all surfaces that are prone to analyte adsorption (Fig. 1). All PEEK capillaries were replaced by surface-treated SS capillaries. The surface treatment consisted in an application of functionalized hydrogenated amorphous silicon coating. The same surface passivation was applied to all wetted SS parts in the drift tube. In the CHARON inlet, the sampler and the vaporizer as well as the transfer tube from the vaporizer to the drift tube were surface-passivated. Fig. 1 shows all surface-treated parts in blue. As for conventional PTR-MS analyzers, the drift tube and inlet lines were housed in a temperature-controlled enclosure that can be operated from room temperature to 120°C. The enclosure temperature is referred to as the drift tube temperature (T_{drift}).

70 2.2 Performance assessment of the EVR PTR-MS instrument

A laboratory study was carried out to measure signal decay times in the EVR PTR-MS analyzer for 29 analytes listed in the Supplement (Tab. S1). A single analyte was supplied in steady concentration to the analyzer and instantly removed from the inlet. $\tau_{1/e}$ is the time it took the analyte signal to decay to 1/e (36.8%) of its initial value. τ_{90} is the time it took the analyte signal to decay to 10% of its initial value. A stable gaseous analyte concentration, which is herein denoted with the subscript (g), was generated by placing a spatula tip of the solid sample into a 100 ml glass vial. The vial was heated and flushed with zero air (RH ~ 30%). Heating temperatures ranged from 50 to 120°C, depending upon the melting point of the analyte. The dynamic headspace of the vial was sampled through the gas inlet of the PTR-MS analyzer. Instrument and inlet were zeroed by overflowing the inlet with zero air (Fig. S1). A stable particulate analyte concentration, which is herein denoted with the subscript (p), was generated by dissolving an aliquot of the solid sample in HPLC-grade water (Sigma-Aldrich Chemie GmbH, Taufkirchen, Germany). The solution was then nebulized with a home-built nebulizer. The nebulizer outflow was dried with two home-built diffusion dryers and gases were removed with an activated charcoal denuder (NovaCarb F, Mast Carbon International Ltd., Guilford, UK). The CHARON inlet and PTR-MS instrument were zeroed by diverting the sample flow through a high efficiency particulate air (HEPA) filter (Fig. S1).

Exemplary data from a field study were taken to show the signal response of the EVR PTR-MS analyzer when switching between the CHARON particle inlet and the gas inlet. The data were collected during a measurement campaign at the TROPOS Research Station Melpitz (Spindler et al., 2013) in Germany in February 2019.

A laboratory study was carried out for investigating the capability of the CHARON EVR PTR-MS analyzer to detect highly oxidized organic molecules in particles. For this purpose, we reacted ozone and limonene in a flow reactor to form secondary organic aerosol (SOA). The reactor outflow was passed through an activated charcoal denuder (NovaCarb F, Mast Carbon International Ltd., Guilford, UK) for removing gaseous organics and subsequently injected into a 210 l steel barrel (Wilai GmbH, Wiedemar, Germany). The CHARON EVR PTR-MS analyzer sampled from this reservoir. The instrument was operated at low reduced electric field strength ($E/N = 30 \text{ Td}$; $1 \text{ Td} = 10^{-17} \text{ Vcm}^{-2}$) and with NH_4^+ as the reagent ion (Müller et al., 2020). With these instrumental settings, ionic fragmentation is largely suppressed and highly oxidized organic molecules are detected in their ammonium adduct form (Zaytsev et al., 2019).

95 3 Results and Discussion

3.1 Effect of surface passivation on time response

Fig. 2 shows data from two experiments wherein the signal decay of *cis*-pinonic acid_(g) was measured with the same PTR-MS instrument, before and after being modified to the EVR configuration. T_{drift} was set to 100°C in both experiments; all other experimental parameters were also kept identical. In the conventional configuration, the analyte signal exhibited a long tailing while it rapidly dropped to near-zero levels in the EVR configuration. $\tau_{1/e}$ was 181 s for the conventional set-up and 4 s in the EVR configuration. These data clearly demonstrate that surface passivation greatly improves the time response of PTR-MS analyzers.

Fig. 3 shows the response of the EVR PTR-MS analyzer ($T_{\text{drift}} = 120^\circ\text{C}$) to an exemplary set of analytes, measured via the gas inlet (upper panel) and via the CHARON particle inlet (lower panel). 4-nitrocatechol_(g) (in dark yellow) exhibited a rather slow decay ($\tau_{1/e} = 27 \text{ s}$). *cis*-Pinonic acid_(g) (in red) decayed in a few seconds ($\tau_{1/e} = 2.4 \text{ s}$). For 2-tridecanone (in blue), $\tau_{1/e}$ was 0.3 s, which is close to the volumetric exchange time of the drift tube (~ 0.1 s). The three exemplary compounds shown here cover the full three orders of magnitude span in $\tau_{1/e}$ (10^{-1} to 10^2 s) that was observed for gaseous analytes. For particulate analytes, $\tau_{1/e}$ ranged from a few seconds to a few tens of seconds. The lower panel of Fig. 3 shows exemplary data obtained for 2,7-

dihydroxynaphthalene_(p) (in dark yellow; $\tau_{1/e} = 15$ s), levoglucosan_(p) (in red; $\tau_{1/e} = 8.1$ s) and nitrate_(p) (in blue; $\tau_{1/e} = 4.4$ s).

110 The nitrate_(p) signal originated from ammonium nitrate particles.

Fig. 4 summarizes the $\tau_{1/e}$ values measured for 21 gaseous analytes (upper panel) and 15 particulate analytes (lower panel).

The color-coding and sizing are explained in the figure legend. Mixing ratios were quantified according to the procedure outlined in the Supplement of Müller et al. (2017) and were typically in the 0.1–10 ppbv range. For some compounds (vanillin_(g), 2-tridecanone_(g), 2,6-dimethoxyphenol_(g), 4-nitroguaiacol_(g), ammonia_(g)) it was difficult to generate mixing ratios

115 in the low-ppbv range, which is why we also included mixing ratios up to 100 ppbv in our analysis. We typically measured $\tau_{1/e}$ at three different mixing ratios for each compound. [Three data points are thus typically shown for each analyte. Since these](#)

[were in a rather narrow range, Only small changes in the instrumental time response were observed because analyte concentrations typically only spanned a relatively narrow range.](#) We will thus not discuss any concentration dependence of $\tau_{1/e}$

here. We observed an increase in $\tau_{1/e}$ with decreasing saturation mass concentrations ($\log C^0$) as a general trend, although with

120 significant deviations for some compounds (Fig. S2). Glucose_(g), for example, exhibited a much faster response than the sugar alcohols (xylitol_(g), arabitol_(g)) despite having a similar $\log C^0$. Structural effects may play a role here, since glucose is a cyclic molecule, while the sugar alcohols are both linear. It is also important to note that the SIMPOL.1 method has not been validated

for saccharides and that the calculated $\log C^0$ may be inaccurate. 4-nitrocatechol, with a relatively high $\log C^0$ of 4.2, was among the slowest responding gaseous analytes. This observation remains unexplained. With the CHARON inlet connected,

125 $\tau_{1/e}$ was in the 5 to 20 s range for most analytes (Fig. 4, lower panel). Exceptions were diglycolic acid_(p) and tartaric acid_(p) with $\tau_{1/e}$ up to ~35 s. Notably, no obvious dependence of $\tau_{1/e}$ on $\log C^0$ was observed. Levoglucosan, 2,7-dihydroxynaphthalene, stearic acid, azelaic acid, diglycolic acid and vanillic acid were studied in both phases. The instrumental response was typically

~5 s slower in the particle measurements. The CHARON inlet has a larger surface area and a lower sample flow than the gas inlet. Stearic acid and azelaic acid responded faster with the CHARON inlet, which remains unexplained. The reader is

130 cautioned that the $\tau_{1/e}$ values presented in Fig. 4 should not be taken as absolute and generally applicable values. The reported numbers should be seen as indicative [estimates-figures](#) for the time response of state-of-the-art IONICON EVR PTR-MS instruments. It is well known that analytes compete for surface adsorption with other matrix constituents such as water or other

surface-affine compounds. All of our experiments were carried out with a single compound at one humidity level. We consider it beyond the scope of this work to investigate a matrix dependence of $\tau_{1/e}$. [Especially for low-volatility analytes, repeated](#)

135 [sampling may passivate remaining active sites, which in turn improves the time response. This effect probably explains the relatively large variations in \$\tau_{1/e}\$ observed for the slow-responding analytes, even if all three experiments were carried out at similar mixing ratios.](#) In previous work, we anecdotally observed that basic analytes exhibited a significantly slower time response when acidic samples had been sampled before. The sampling history was not considered in our study.

[Especially for low-volatility analytes, repeated sampling may passivate remaining active sites, which in turn improves the time response. This effect probably explains the relatively large variations in \$\tau_{1/e}\$ observed for the slow-responding analytes, even if all three experiments were carried out at similar mixing ratios.](#)

In previous work, we anecdotally observed that basic analytes exhibited a significantly slower time response when acidic samples had been sampled before. The sampling history was not considered in our study.

The sampling history was not considered in our study.

3.2 Drift tube temperature effects

140 For long-chain ketones (Pagonis et al., 2017; Krechmer et al., 2018) and carboxylic acids (Fig. S2), $\tau_{1/e}$ exhibits the expected increase with decreasing $\log C^0$. Since $\log C^0$ depends upon temperature, changes in T_{drift} should lead to predictable changes in $\tau_{1/e}$. We thus measured $\tau_{1/e}$ for 2-tridecanone_(g) and *cis*-pinonic acid_(g) at variable T_{drift} (Fig. 5). In the case of 2-tridecanone_(g)

(upper panel), an increase in T_{drift} from room temperature to 60 °C decreased $\tau_{1/e}$ from 24 s to 2 s. At $T_{\text{drift}} = 100^\circ\text{C}$, $\tau_{1/e}$ was 0.3 s approaching the volumetric exchange time of the drift tube. In the case of *cis*-pinonic acid_(g) (lower panel), we only

145 investigated temperatures of 80°C and above because the solid analyte sample was kept at 70°C. When increasing T_{drift} from 80 to 120°C, $\tau_{1/e}$ dropped from 6.5 to 2.4 s. For the two compounds investigated, the decrease in $\tau_{1/e}$ can be explained by the increase in $\log C^0$ with temperature. According to Epstein et al. (2010), a 15°C temperature rise increases $\log C^0$ by ~1 in the 0 to 50°C temperature range. The effect becomes less pronounced at higher temperatures. For *cis*-pinonic acid, the SIMPOL.1 method yields a $\log C^0$ increase from 5.2 to 6.2 in the 80 to 120°C temperature range. At 45°C, 2-tridecanone has roughly the

increase in $\log C^0$ with temperature. According to Epstein et al. (2010), a 15°C temperature rise increases $\log C^0$ by ~1 in the 0 to 50°C temperature range. The effect becomes less pronounced at higher temperatures. For *cis*-pinonic acid, the SIMPOL.1 method yields a $\log C^0$ increase from 5.2 to 6.2 in the 80 to 120°C temperature range. At 45°C, 2-tridecanone has roughly the

At 45°C, 2-tridecanone has roughly the

150 same $\log C^0$ as *cis*-pinonic acid at 120°C (Fig. S3). Consistently, we observed a similar time response for the two compounds at $T_{\text{drift}} = 45^\circ\text{C}$ and $T_{\text{drift}} = 120^\circ\text{C}$, respectively. We conclude that increasing T_{drift} to 120°C is an effective way for reducing $\tau_{1/e}$ in EVR PTR-MS analyzers.

Exposing the sample gas to heated surfaces in an analyzer, may thermally degrade some analytes. It is important to note that none of the acids studied in this work decarboxylated at $T_{\text{drift}} = 120^\circ\text{C}$. It may, however, be necessary to use a lower T_{drift} when
155 more thermally labile analytes are targeted.

3.3 Response times of different online CIMS instruments

Fig. 6 compares the signal decay times (here reported as τ_{90} and not as $\tau_{1/e}$) as observed in different online CIMS instruments. The plot includes data from this study, which were obtained with two state-of-the-art IONICON PTR-MS analyzers. The data points in yellow were obtained with a conventional instrument, while the data points in dark red were collected with an EVR-
160 type analyzer. Both instruments were operated at $T_{\text{drift}} = 120^\circ\text{C}$. The figure also includes literature data obtained with an old quadrupole PTR-MS instrument (qPTR-MS; in blue: Pagonis et al., 2017), with a PTR-ToF-MS analyzer from a different manufacturer (VOCUSTM; in orange: Krechmer et al., 2018) and with two iodide (I⁻) CIMS instruments (in green: Liu et al., 2019; in light blue: Palm et al., 2019), respectively. The latter four instruments were all operated at room temperature (RT).
 τ_{90} is plotted against $\log C^0$ of the respective analyte, which was calculated using the SIMPOL.1 method (Pankow and Asher,
165 2008). The upper horizontal axis lists the compound names; the lower horizontal axis shows in which $\log C^0$ range molecules are classified as volatile organic compounds (VOCs), intermediate volatility organic compounds (IVOCs) and semi-volatile organic compounds (SVOCs). Since the instruments did not measure the same SVOCs, we only use our carboxylic acid data for the $\log C^0$ -based comparison. For the acids, $\log C^0$ appears to be the main determinant of τ_{90} while for the saccharides and substituted phenols other properties affect τ_{90} (Fig. S2). For VOCs and the more volatile IVOCs ($\log C^0 \geq 5$), τ_{90} is close to the
170 volumetric exchange time of the drift tube for both the EVR and the conventional PTR-ToF-MS instrument, if the drift tube and inlet are heated to 120°C. Conventional unheated PTR-MS instruments have a much slower time response. For the less volatile IVOCs ($\log C^0 \leq 3$; *cis*-pinonic acid) surface passivation reduces τ_{90} by two orders of magnitude, even if the drift tube is heated to 120°C in both the conventional and the EVR PTR-ToF-MS analyzer. For the SVOCs, the heated EVR PTR-ToF-MS has a similar time response as an optimized I⁻ CIMS instrument with reduced instrument wall interactions (Palm et al.,
175 2019).

3.4 Rapid switching between the gas inlet and the CHARON particle inlet

For certain applications, it is desirable to periodically switch between gas and particle measurements. Fig. 7 shows exemplary data collected by a CHARON EVR PTR-MS instrument during ambient air measurements at a rural background station in Germany.

180 The reader should focus on the transition from background to ambient particle measurements at 06:15:40 and the transition from particle to gas measurements at 06:21:40, which is when the slow signal response becomes most evident. The upper panel shows the time evolution of the NH_4^+ (m/z 18.034) and NO_2^+ (m/z 45.996) signals. In the CHARON inlet, ammonium nitrate particles evaporate to yield gaseous ammonia (NH_3) and nitric acid (HNO_3). Ammonia is detected in its protonated form, while protonated nitric acid dehydrates upon protonation to yield the nitronium ion (NO_2^+). Both ammonia and nitric acid are
185 particularly prone to adsorptive losses on stainless steel (Neuman et al., 1999; Nowak et al., 2007). In the EVR-type instrument, wetted surfaces do not include any untreated stainless steel, which results in a fast instrumental response to both compounds. Both the NH_4^+ and the NO_2^+ signal equilibrated within 1 minute when switching from HEPA to ambient CHARON measurements and from particle to gas measurements, respectively. The lower panel shows the evolution of the $\text{C}_4\text{H}_5\text{O}^+$ (m/z 69.033) and $\text{C}_6\text{H}_9\text{O}_4^+$ (m/z 145.049) signals within one measurement cycle. $\text{C}_6\text{H}_9\text{O}_4^+$ is the main ionic fragment from

190 levoglucosan (Leglise et al., 2019). $C_4H_5O^+$ is believed to be a fragment of larger furanoid compounds in particles and protonated furan in the gas measurement. The instrumental response to these analytes was somewhat slower than to ammonium nitrate but equilibration still occurred within 1 minute. The instrumental response to levoglucosan is similar to what we observed in single compound measurements in the laboratory, suggesting that the presence of a complex matrix does not negatively affect instrumental response times.

195 **3.5 Detection of highly oxidized organic compounds**

The CHARON PTR-MS instrument has been successfully used for measuring the chemical composition of SOA (*e.g.*, Gkatzelis et al., 2018). The detection of highly oxidized species has, however, been hampered by the adsorption of low-volatility species onto instrumental surfaces. In an attempt to demonstrate the improved detection capability of the CHARON EVR PTR-MS instrument, we sampled SOA generated from limonene ozonolysis. Up to eight oxygen atoms were detected in
200 the analyte ions (Fig S4). Fig. 8 shows the rapid decay ($\tau_{1/e} = 15$ s) of an O_8 -compound ($C_{19}H_{30}O_8$, detected as ammonium adduct) when switching from the CHARON particle inlet to the gas inlet. The rapidness of the decay suggests that surface adsorption effects play a minor role even for such highly oxidized species. Hammes et al. (2019) claimed that the $C_{19}H_{30}O_8$ molecule formed from limonene ozonolysis is a multifunctional compound containing three keto groups, one carboxyl group, one hydroperoxy group and one ester group. A calculation using SIMPOL.1 (Pankow and Asher, 2008) yields a $\log C^0$ of -5.9
205 for this molecule, suggesting that even extremely low-volatility organic compounds (ELVOCs) can be monitored in real time by CHARON EVR PTR-MS. The fact that the detected $C_{19}H_{30}O_8$ molecule contains a hydroperoxy group is notable, because hydroperoxides are known to efficiently decompose on metal surfaces in conventional PTR-MS instruments (Rivera-Rios et al., 2014). An additional benefit of surface passivation may thus be that the metal-catalyzed decomposition of peroxides and hydroperoxides is suppressed. Preliminary laboratory tests with cumene hydroperoxide and dicumylperoxide (*i.e.*, the only
210 peroxides that are commercially available, although with significant impurities) indicate that these compounds do indeed not decompose in EVR-type instruments. More work is needed to confirm this preliminary finding.

4 Conclusion

We have described and characterized the novel EVR PTR-MS instrument, which exhibits a significantly improved time-response performance as compared to conventional IONICON PTR-MS analyzers. The time response of this optimized
215 instrument is comparable to that of fastest flow tube CIMS instruments that are currently in use. This allows to rapidly switch between gas and particle measurements, making the CHARON EVR PTR-MS instrument the only direct sample introduction CIMS instrument that can monitor gaseous and particulate organics in the atmosphere in real time. Besides being faster, the EVR PTR-MS instrument also allows to target new analyte classes such as highly oxygenated organic molecules and potentially also hydroperoxides. We believe that the CHARON EVR PTR-MS instrument will be a valuable tool for
220 overcoming current challenges in the measurement of atmospheric organic carbon (Heald and Kroll, 2020).

Author contribution

KW developed the EVR system and performed initial tests. MM, FP and AW designed the experimental studies. MM, FP and JSS carried out the experiments. FP performed the data analysis, with support from JSS and MM. FP and MM drafted the manuscript. AW wrote the final manuscript. All authors commented and accepted the final version of the manuscript.

225 **Conflicts of interest**

MM, FP and KW work for IONICON Analytik. IONICON Analytik commercializes PTR-MS instruments (both in the conventional and in the EVR configuration) and the CHARON particle inlet. MM and AW profit from a license agreement (CHARON inlet) between the University of Innsbruck and IONICON Analytik.

Acknowledgements


230 We like to thank Laurent Poulain and Gerald Spindler from TROPOS for their support during the measurements in Melpitz. Measurements in Melpitz were supported by the European Union's Horizon 2020 research and innovation programme under grant agreement N°654109 (ACTRIS-2). FP has received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement N°674911 (IMPACT). Special thanks go to Magda Claeys for providing many of the substances studied in this work.

235

References

- Breitenlechner, M., Fischer, L., Hainer, M., Heinritzi, M., Curtius, J. and Hansel, A.: PTR3: An Instrument for Studying the Lifecycle of Reactive Organic Carbon in the Atmosphere, *Anal. Chem.*, 89(11), 5824–5831, doi:10.1021/acs.analchem.6b05110, 2017.
- 240 Deming, B. L., Pagonis, D., Liu, X., Day, D. A., Talukdar, R., Krechmer, J. E., de Gouw, J. A., Jimenez, J. L. and Ziemann, P. J.: Measurements of delays of gas-phase compounds in a wide variety of tubing materials due to gas–wall interactions, *Atmos. Meas. Tech.*, 12(6), 3453–3461, doi:10.5194/amt-12-3453-2019, 2019.
- Eichler, P., Müller, M., D’Anna, B. and Wisthaler, A.: A novel inlet system for online chemical analysis of semi-volatile submicron particulate matter, *Atmos. Meas. Tech.*, 8(3), 1353–1360, doi:10.5194/amt-8-1353-2015, 2015.
- 245 Epstein, S. A., Riipinen, I. and Donahue, N. M.: A Semiempirical Correlation between Enthalpy of Vaporization and Saturation Concentration for Organic Aerosol, *Environ. Sci. Technol.*, 44(2), 743–748, doi:10.1021/es902497z, 2010.
- Goldstein, A. H. and Galbally, I. E.: Known and Unexplored Organic Constituents in the Earth’s Atmosphere, *Environ. Sci. Technol.*, 41(5), 1514–1521, doi:10.1021/es072476p, 2007.
- de Gouw, J. and Warneke, C.: Measurements of volatile organic compounds in the earth’s atmosphere using proton-transfer-reaction mass spectrometry, *Mass Spectrom. Rev.*, 26(2), 223–257, doi:10.1002/mas.20119, 2007.
- 250 Hammes, J., Lutz, A., Mentel, T., Faxon, C. and Hallquist, M.: Carboxylic acids from limonene oxidation by ozone and OH radicals: Insights into mechanisms derived using a FIGAERO-CIMS, *Atmos. Chem. Phys.*, 19(20), 13037–13052, doi:10.5194/acp-19-13037-2019, 2019.
- Hansel, A., Jordan, A., Holzinger, R., Prazeller, P., Vogel, W. and Lindinger, W.: Proton transfer reaction mass spectrometry: on-line trace gas analysis at the ppb level, *Int. J. Mass. Spectrom.*, 149–150, 609–619, doi:10.1016/0168-1176(95)04294-U, 1995.
- 255 Heald, C. L. and Kroll, J. H.: The fuel of atmospheric chemistry: Toward a complete description of reactive organic carbon, *Sci. Adv.*, 6(6), eaay8967, doi:10.1126/sciadv.aay8967, 2020.
- Krechmer, J., Lopez-Hilfiker, F., Koss, A., Hutterli, M., Stoermer, C., Deming, B., Kimmel, J., Warneke, C., Holzinger, R., Jayne, J., Worsnop, D., Fuhrer, K., Gonin, M. and de Gouw, J.: Evaluation of a New Reagent-Ion Source and Focusing Ion–Molecule Reactor for Use in Proton-Transfer-Reaction Mass Spectrometry, *Anal. Chem.*, 90(20), 12011–12018, doi:10.1021/acs.analchem.8b02641, 2018.
- 260 Leglise, J., Müller, M., Piel, F., Otto, T. and Wisthaler, A.: Bulk organic aerosol analysis by PTR-MS: an improved methodology for the determination of total organic mass, O:C and H:C elemental ratios and the average molecular formula, *Anal. Chem.*, acs.analchem.9b02949, doi:10.1021/acs.analchem.9b02949, 2019.
- 265 Liu, X., Deming, B., Pagonis, D., Day, D. A., Palm, B. B., Talukdar, R., Roberts, J. M., Veres, P. R., Krechmer, J. E., Thornton, J. A., de Gouw, J. A., Ziemann, P. J. and Jimenez, J. L.: Effects of gas–wall interactions on measurements of semivolatile compounds and small polar molecules, *Atmos. Meas. Tech.*, 12(6), 3137–3149, doi:10.5194/amt-12-3137-2019, 2019.
- 270 Mikoviny, T., Kaser, L. and Wisthaler, A.: Development and characterization of a High-Temperature Proton-Transfer-Reaction Mass Spectrometer (HT-PTR-MS), *Atmos. Meas. Tech.*, 3(3), 537–544, doi:10.5194/amt-3-537-2010, 2010.
- Müller, M., Graus, M., Ruuskanen, T. M., Schnitzhofer, R., Bamberger, I., Kaser, L., Titzmann, T., Hörtnagl, L., Wohlfahrt, G., Karl, T. and Hansel, A.: First eddy covariance flux measurements by PTR-TOF, *Atmos. Meas. Tech.*, 3(2), 387–395, doi:10.5194/amt-3-387-2010, 2010.
- 275 Müller, M., Eichler, P., D’Anna, B., Tan, W. and Wisthaler, A.: Direct Sampling and Analysis of Atmospheric Particulate Organic Matter by Proton-Transfer-Reaction Mass Spectrometry, *Anal. Chem.*, 89(20), 10889–10897, doi:10.1021/acs.analchem.7b02582, 2017.
- Müller, M., Piel, F., Gutmann, R., Sulzer, P., Hartungen, E. and Wisthaler, A.: A novel method for producing NH₄⁺ reagent ions in the hollow cathode glow discharge ion source of PTR-MS instruments, *Int. J. Mass. Spectrom.*, 447, 116254, doi:10.1016/j.ijms.2019.116254, 2020.
- 280 Neuman, J. A., Huey, L. G., Ryerson, T. B. and Fahey, D. W.: Study of Inlet Materials for Sampling Atmospheric Nitric Acid, *Environ. Sci. Technol.*, 33(7), 1133–1136, doi:10.1021/es980767f, 1999.

- Nowak, J. B., Neuman, J. A., Kozai, K., Huey, L. G., Tanner, D. J., Holloway, J. S., Ryerson, T. B., Frost, G. J., McKeen, S. A. and Fehsenfeld, F. C.: A chemical ionization mass spectrometry technique for airborne measurements of ammonia: AIRBORNE CIMS NH₃ MEASUREMENTS, *J. Geophys. Res.*, 112(D10), doi:10.1029/2006JD007589, 2007.
- Pagonis, D., Krechmer, J. E., de Gouw, J., Jimenez, J. L. and Ziemann, P. J.: Effects of gas-wall partitioning in Teflon tubing and instrumentation on time-resolved measurements of gas-phase organic compounds, *Atmos. Meas. Tech.*, 10(12), 4687–4696, doi:10.5194/amt-10-4687-2017, 2017.
- Palm, B. B., Liu, X., Jimenez, J. L. and Thornton, J. A.: Performance of a new co-axial ion-molecule reaction region for low-pressure chemical ionization mass spectrometry with reduced instrument wall interactions, *Atmos. Meas. Tech.*, 12(11), 5829–5844, doi:10.5194/amt-12-5829-2019, 2019.
- Pankow, J. F. and Asher, W. E.: SIMPOL.1: a simple group contribution method for predicting vapor pressures and enthalpies of vaporization of multifunctional organic compounds, *Atmos. Chem. Phys.*, 8(10), 2773–2796, doi:10.5194/acp-8-2773-2008, 2008.
- Piel, F., Müller, M., Mikoviny, T., Pusede, S. E. and Wisthaler, A.: Airborne measurements of particulate organic matter by proton-transfer-reaction mass spectrometry (PTR-MS): a pilot study, *Atmos. Meas. Tech.*, 12(11), 5947–5958, doi:10.5194/amt-12-5947-2019, 2019.
- Rivera-Rios, J. C., Nguyen, T. B., Crouse, J. D., Jud, W., St. Clair, J. M., Mikoviny, T., Gilman, J. B., Lerner, B. M., Kaiser, J. B., Gouw, J., Wisthaler, A., Hansel, A., Wennberg, P. O., Seinfeld, J. H. and Keutsch, F. N.: Conversion of hydroperoxides to carbonyls in field and laboratory instrumentation: Observational bias in diagnosing pristine versus anthropogenically controlled atmospheric chemistry, *Geophys. Res. Lett.*, 41(23), 8645–8651, doi:10.1002/2014GL061919, 2014.
- Spindler, G., Grüner, A., Müller, K., Schlimper, S. and Herrmann, H.: Long-term size-segregated particle (PM₁₀, PM_{2.5}, PM₁) characterization study at Melpitz -- influence of air mass inflow, weather conditions and season, *J. Atmos. Chem.*, 70(2), 165–195, doi:10.1007/s10874-013-9263-8, 2013.
- Yuan, B., Koss, A. R., Warneke, C., Coggon, M., Sekimoto, K. and de Gouw, J. A.: Proton-Transfer-Reaction Mass Spectrometry: Applications in Atmospheric Sciences, *Chem. Rev.*, 117(21), 13187–13229, doi:10.1021/acs.chemrev.7b00325, 2017.
- Zaytsev, A., Breitenlechner, M., Koss, A.R., Lim, C. Y., Rowe, J. C., Kroll, J. H. and Keutsch, F. N.: Using Collision-Induced Dissociation to Constrain Sensitivity of Ammonia Chemical Ionization Mass Spectrometry (NH₄⁺ CIMS) to Oxygenated Volatile Organic Compounds. *Atmos. Meas. Tech.*, 12(3), 1861–1870, doi:10.5194/amt-12-1861-2019, 2019.



315

Figure 1. Scheme of the EVR PTR-MS instrument including a gas inlet and the CHARON particle inlet. The parts in blue were passivated with a functionalized hydrogenated amorphous silicon coating. MFC: mass flow controller; PC: pressure controller, ADL: aerodynamic lens; ToF-MS: time-of-flight mass spectrometer

320






Figure 2. Signal decay as observed in a conventional PTR-MS instrument ($T_{\text{drift}} = 100^{\circ}\text{C}$; in blue) and in an EVR PTR-MS analyzer ($T_{\text{drift}} = 100^{\circ}\text{C}$; in red) after a steady supply of gaseous *cis*-pinonic acid was instantly switched off a $t = 0$ s.



330 **Figure 3.** Signal decay as observed in an EVR PTR-MS analyzer ($T_{\text{drift}} = 120^{\circ}\text{C}$) after a steady supply of analyte was instantly switched off at $t = 0$ s. The decay of gaseous analytes is shown in the upper panel, while the lower panel refers to particulate analytes. Initial steady-state mixing ratios were as follows: 2-tridecanone_(g): 1 ppbv, *cis*-pinonic acid_(g): 2.2 ppbv, 4-nitrocatechol_(g): 30 ppbv, nitrate: 3 ppbv; levoglucosan_(p): 1.2 ppbv, 2,7-dihydroxynaphthalene_(p): 0.3 ppbv. Signals were fitted using a double exponential decay function (see insert in the upper panel). All fitting parameters are listed in Table S2.



335 **Figure 4.** Signal decay times ($\tau_{1/e}$) measured for 21 gaseous analytes (upper panel) and 15 particulate analytes (lower panel). Analytes
 were grouped and color-coded into six classes: saccharides (in dark blue), carboxylic acids (in red), substituted phenols (in orange),
 nitroaromatics (in magenta), small polar molecules (in light blue) and others (in black). The size of the dots indicates the initial
 steady-state mixing ratio (0.1–100 ppbv) used in the respective experiment. The color code of the data points indicates the saturation
 mass concentrations ($\log C^0$) of the analytes as calculated using the SIMPOL.1 method (Pankow and Asher, 2008).

340 (i) originating from ammonium nitrate, (ii) originating from ammonium sulfate




Figure 5. Signal decay as observed in an EVR PTR-MS analyzer for 2-tridecanone_(g) (upper panel) and *cis*-pinonic acid_(g), respectively, at different drift tube temperatures. Initial steady-state mixing ratios were as follows: 2-tridecanone_(g): 1–1.3 ppbv; *cis*-pinonic acid: 2.2 ppbv. Signals were fitted using a double exponential decay function (see insert in the lower panel). All fitting parameters are listed in Table S3.

345





Figure 6. Signal decay times (here reported as τ_{90}) as observed in different online CIMS instruments for a set of ketones, carboxylic acids and dihydroxycarbonyls. τ_{90} is plotted as a function of the SIMPOL.1-derived saturation mass concentration ($\log C^0$). [Note that the figure legend shows the different operating temperatures of the instruments used \(RT: room temperature\).](#)



355 **Figure 7.** Time series showing various analyte signals as recorded during a 10-minute measurement cycle of ambient air. The 10-min measurement cycle included: i) 1 min of instrumental background measurements with the CHARON inlet including a HEPA filter, ii) 4 mins of particulate measurements at an E/N of 60 Td, iii) 1 min of particulate measurements at an E/N of 100 Td and iv) 1 min of gas measurements at an E/N of 100 Td. The benefit of measuring particles at 60 and 100 Td is explained in Leglise et al. (2019) and is not discussed here. The CHARON inlet enriched the particle concentration by a factor of ~ 20 , which explains the higher signal intensities in the particle measurement (in blue) as compared to the gas measurement (in yellow).




Figure 8: Decay of the $\text{NH}_4^+(\text{C}_{19}\text{H}_{30}\text{O}_8)$ signal as recorded by a CHARON EVR PTR-MS analyzer when O_3 /limonene SOA was sampled and the instrument was switched from the CHARON particle inlet to the gas inlet at $t=0$.