

# A single-beam photothermal interferometer for in-situ measurements of aerosol light absorption

Bradley Visser<sup>1</sup>, Jannis Röhrbein<sup>1</sup>, Peter Steigmeier<sup>1</sup>, Luka Drinovec<sup>2,3</sup>, Griša Močnik<sup>2,3,4</sup>, Ernest Weingartner<sup>1</sup>

5 <sup>1</sup> University of Applied Sciences Northwestern Switzerland, Windisch, Switzerland

<sup>2</sup> Jozef Stefan Institute, Ljubljana, Slovenia

<sup>3</sup> Haze Instruments d.o.o., Ljubljana, Slovenia

<sup>4</sup> University of Nova Gorica, Ajdovščina, Slovenia

10 **Abstract.** We have developed a novel single-beam photothermal interferometer and present here its application for the measurement of aerosol light absorption. The use of only a single laser beam allows for a compact optical set up and significantly easier alignment compared to standard dual-beam photothermal interferometers, making it ideal for field measurements. Due to a unique configuration of the reference interferometer arm, light absorption by aerosols can be determined directly even in the presence of light absorbing gases. The instrument can be  
15 calibrated directly with light absorbing gases, such as NO<sub>2</sub>, and can be used to calibrate other light absorption instruments. The detection limits ( $1\sigma$ ) for absorption for ten and sixty second averaging times were determined to be 14.6 Mm<sup>-1</sup> and 7.4 Mm<sup>-1</sup>, respectively, which for a mass absorption cross-section of 10 m<sup>2</sup> g<sup>-1</sup> leads to equivalent black carbon concentration detection limits of 1460 ng m<sup>-3</sup> and 740 ng m<sup>-3</sup>, respectively. The detection limit could be reduced further by improvements to the isolation of the instrument and the signal detection and processing  
20 schemes employed.

## 1. Introduction

According to estimates from the World Health Organization (WHO), particulate air pollution contributes to about seven million premature deaths each year, making it one of the leading causes of early mortality worldwide (WHO, 2014). Studies of short-term health effects suggest that black carbon (BC) particles, a component of carbonaceous  
25 aerosols, are a better indicator of the effect harmful particulate substances from combustion sources exert on human health than any other metric (Janssen et al., 2011; Janssen et al., 2012), and it is acknowledged that BC poses tremendous harm to public health. New estimates based on re-evaluation of data from across Europe suggest that air pollution leads to more than double the number of deaths than previously thought (Lelieveld et al., 2019).

Besides its health relevance, aerosol black carbon also significantly affects Earth's climate (Bond et al., 2013; Myhre et al., 2013). Aerosols influence our climate by their ability to scatter and absorb solar radiation (Myhre et al., 2013). As BC particles are highly efficient light absorbers, they are considered to be the second most important anthropogenic climate forcer after CO<sub>2</sub> (Bond et al., 2013). However, the uncertainty of the BC warming effect is still very high: the best estimate of the radiative forcing of BC is +1.1 W m<sup>-2</sup> (90% uncertainty bounds +0.17 to +2.1 W m<sup>-2</sup>) (Bond et al., 2013).

35 Aerosol light absorption is quantified using the wavelength dependent absorption coefficient  $b_{abs}(\lambda)$ , which is defined as the attenuation of light due to absorption in the medium per unit length. The total attenuation of light passing through a sample is determined by the absorption and scattering ( $b_{scat}$ ) coefficients using the Beer-Lambert law

$$I = I_0 e^{-(b_{abs}+b_{scat})x} \quad (1)$$

40 where  $I$  is the intensity of light remaining after transmission through a medium of length  $x$  given an initial intensity  $I_0$ . In order to relate aerosol light absorption to a mass concentration of (absorbing) aerosol particles the mass absorption cross-section (MAC) of the aerosol is required:  $m = \frac{b_{abs}}{MAC}$ , where  $m$  is the mass concentration of the light absorbing aerosol component. For typical ambient BC aerosols measured at  $\lambda = 637$  nm the MAC is approximately  $10 \text{ m}^2 \text{ g}^{-1}$  (Petzold et al., 2002), however the uncertainty in this value is large due to the unavailability of traceable reference methods (Zanatta et al., 2016). In order to clarify the quantity that is measured in such experiments, Petzold et al. (2013) recommend to use the term equivalent black carbon (eBC) when its mass is derived by optical measurements.

Aerosol light absorption is commonly measured *ex-situ* using filter-based devices, such as the Aethalometer (Drinovec et al., 2015), Multi Angle Absorption Photometer (MAAP) (Petzold et al., 2002), Particle Soot Absorption Photometer (PSAP) (Bond et al., 1999) and Continuous Light Absorption Photometer (CLAP) (Ogren et al., 2017). In such measurements, the aerosol particles are deposited into a filter and the light transmission through the sample-laden filter is measured relative to the unloaded filter. The advantage of filter-based techniques is that they are straightforward, allow for unattended operation, and are relatively inexpensive. In addition, they have low detection limits due to the accumulation of the absorbing species on the filter over time: the detection limits can reach  $b_{abs} < 0.05 \text{ Mm}^{-1}$  when the sample is collected over a sufficiently long time (Springston and Sedlacek, 2007; Backman et al., 2017). The durability and sensitivity of filter-based instruments have led to their employment in environmental monitoring stations. These methods have significant drawbacks however, as they suffer from large systematic errors caused by the modification of particle properties upon deposition in the filter (Weingartner et al., 2003; Lack et al., 2008; Drinovec et al., 2015; Drinovec et al., 2017; Lee, 2019). In addition, various optical interactions between the deposited particles and the filter medium can enhance or lower the measured absorption. One major issue is the cross sensitivity to scattering material embedded in the filter, which enhances the apparent absorption (Arnott et al., 2005; Collaud Coen et al., 2010).

Due to the artefacts inherent in filter-based measurements, it is advantageous to measure the aerosol absorption with the particles suspended in the air. Several *in-situ* measurement techniques exist, amongst which the most common being the “extinction minus scattering” and photoacoustic methods. In the “extinction minus scattering” method, light extinction and light scattering are measured separately, with light absorption defined as the difference between the measured quantities. The measurements can be very accurate, but encounter difficulties for aerosols featuring high single-scattering albedo  $\frac{b_{scat}}{b_{scat}+b_{abs}}$  (above approx. 0.75), in which extinction and scattering are both large and almost equal quantities (Bond et al., 1999; Schnaiter et al., 2005). Instruments based on the photoacoustic effect measure the light absorption of the sample directly as a pressure wave generated after the absorption of light in the aerosol and subsequent heating of the gas. As the photoacoustic signal is only generated by light absorption, artefacts from light scattering are completely eliminated. The photoacoustic method encounters a significant bias when measuring hygroscopic aerosols in elevated relative humidity (RH) or samples with volatile coatings – this artefact arises from the loss of the latent heat of these particle-bound volatile species as they evaporate from the heated particles, reducing the apparent acoustic signal (Arnott et al., 2003; Raspert et al., 2003; Murphy, 2009; Langridge et al., 2013). Some photoacoustic instruments can achieve detection limits of  $b_{abs} \approx 0.1 \text{ Mm}^{-1}$  (with 60 s averaging) (Lack et al., 2006), though most instruments have considerably higher detection limits (Linke et al., 2016).

80 *In situ* absorption methods have a further advantage over traditional measurements: the ability to traceably calibrate the instrument using an absorbing gaseous species such as NO<sub>2</sub> (Arnott et al., 2000; Nakayama et al., 2015) or O<sub>3</sub> (Lack et al., 2006; Lack et al., 2012; Davies et al., 2018). Such internal primary calibration standards are unavailable for filter-based instrumentation, which rely on comparative measurements with reference instruments and reference aerosols; not only are such calibration processes prone to biases, they also cannot be performed in the field, requiring the instrument to be shipped to the calibration facility.

85 Photothermal interferometry (PTI) is an *in-situ* direct absorption measurement technique originally developed for measurements of trace gases (Davis and Petuchowski, 1981; Fulghum and Tilleman, 1991; Mazzoni and Davis, 1991) that has also been applied to aerosol measurements (Davis and Petuchowski, 1981; Fluckiger et al., 1985; Lin and Campillo, 1985; Moosmüller and Arnott, 1996; Sedlacek, 2006; Sedlacek and Lee, 2007; Moosmüller et al., 2009; Lack et al., 2014; Lee and Moosmüller, 2020). In PTI, the light absorption  $b_{abs}$  of a sample is measured  
90 by probing light absorption induced changes of refractive index using interferometry. Previous realisations of PTI require two lasers, one of high power that is modulated and absorbed by the sample (pump), and a second CW interferometry laser (probe). Upon absorption of pump beam light by the sample, energy is transferred via heat conduction to the buffer gas, which results in highly localised heating and thus a refractive index change within the pump beam volume, which is measured by the probe beam. At the shot noise limit, the theoretical detection  
95 limit of PTI has been calculated to be  $b_{abs} < 0.01 \text{ Mm}^{-1}$  (30 s integration time) (Sedlacek, 2006), though for aerosol measurements the practical detection limit is consistently considerably higher (e.g.  $> 0.2 \text{ Mm}^{-1}$  from Sedlacek and Lee (2007)).

The primary difficulties associated with achieving the theoretical detection limits are the sensitivity of interferometric measurements to external noise sources, the difficulty of maintaining the optimal alignment of the  
100 pump and probe beams, and measurement artefacts due to cross-sensitivity to other absorbing species, such as NO<sub>2</sub>, volatile organic compounds (VOC) and O<sub>3</sub>. Improvements to isolation systems and adoption of folded interferometer configurations has led to a significant reduction in noise (Moosmüller and Arnott, 1996). The alignment of the laser beams requires the instrument to be calibrated using a species of known absorption and this calibration must be repeated periodically to track the sensitivity (beam overlap) of the PTI. Otherwise, uncorrected  
105 changes in sensitivity are interpreted as changes of the aerosol light absorption. Measurement artefacts due to the cross sensitivity of PTI to other absorbing species have typically been compensated for by simultaneously measuring a filtered sample stream and subtracting the offset.

In order to address the aforementioned difficulties associated with the PTI technique, a new PTI instrument employing a single laser and unique beam configuration has been developed. This instrument greatly simplifies  
110 the alignment of the interferometer, maximises the sensitivity of the measurement and enables artefact free measurement of aerosol absorption in the presence of absorbing gaseous species. We report here on the experimental realisation of this instrument, which we have termed the modulated single-beam PTI (MSPTI) configuration, its initial characterisation with NO<sub>2</sub> and first laboratory measurements of carbonaceous aerosols. Future improvement of the sensitivity and durability of the MSPTI is planned, enabling its use as a field monitoring  
115 instrument.

## 2. Standard pump-probe PTI

In the PTI technique, the absorption based induced change of the refractive index of the buffer gas is detected via the relative phase shift of light waves passing through the sample versus those traversing the reference medium within the interferometer. Substances (particles, molecules, etc.) that absorb light at the pump laser wavelength transfer the absorbed energy to the surrounding buffer gas via heat conduction, resulting in a local increase of the buffer gas temperature. As the refractive index of a gas depends upon its density, which is itself temperature dependent, modulation of the pump laser intensity in the presence of a light absorbing substance results in the local modulation of the refractive index. Light passing through this volume experiences a periodic phase shift, which can be measured via interferometry. Figure 1 shows a schematic of this process.

The measured phase shift  $\Delta\varphi$  is related the absorption coefficient  $b_{abs}$  via the following relation (Moosmüller et al., 1997; Sedlacek, 2006):

$$\Delta\varphi = \frac{2\pi(n-1)}{\lambda_{probe}T\rho C_p} \frac{lP_{pump}}{A} b_{abs}\Delta t \quad (2)$$

The first term can be considered constant for a given temperature, where  $n$ ,  $T$ ,  $\rho$  and  $C_p$  are the refractive index, temperature, density and heat capacity of the air, respectively.  $\lambda_{probe}$  is the interferometer laser wavelength and  $\Delta t$  denotes the heating time within the modulation cycle. The second term can be defined as the sensitivity of the PTI measurement, where  $l$  is the length of the overlap of the beams within the sample volume,  $P_{pump}$  is the modulation amplitude of the pump beam power and  $A$  is the effective cross-sectional area of the laser beams. Therefore, maximum sensitivity is achieved by maximising the length of the interaction and the pump laser power and minimising the cross-sectional area of the beams (though requiring that the cross sections overlap).

In standard realisations of PTI, a single sample chamber is placed either in the measurement arm alone or across both arms of the interferometer. One potential realisation of a standard PTI set up is shown in Figure 2. The modulated pump beam is set to overlap the probe beam within the sample chamber in the measurement (lower) arm. As interferometric measurements are relative, the resultant signal contains an modulated component associated with light absorption and an offset due to the optical path length difference between the two arms of the interferometer.

Interferometric detection schemes use the wave properties of light to measure miniscule changes in optical path lengths. An interferometer typically consists of two optical paths (designated measurement and reference arms) and two output beams in which light from the respective paths are combined. The light waves interfere in the output beams, causing the measured intensities of the output beams to vary with the phase difference between the measurement and reference light waves (which is determined by the difference in the optical path length between the two arms). The light intensity in each output beam (measured with the respective detectors) can then be defined in terms of this phase difference  $\varphi_{meas} - \varphi_{ref}$ , which we will assign as the interferometric phase  $\varphi$  and is given

$$I_1 = I_0 \sin^2\left(\frac{\varphi}{2}\right) = \frac{1}{2}I_0(1 + \cos(\varphi)) \quad (3)$$

$$I_2 = I_0 \cos^2\left(\frac{\varphi}{2}\right) = \frac{1}{2}I_0(1 - \cos(\varphi)) \quad (4)$$

where  $I_0$  is the intensity of the laser before the initial beam-splitter.

Interference of the beams leads to the limits of totally constructive  $I_x = I_0$  and totally destructive interference  $I_x = 0$  at each detector. These limits are shifted by  $\pi$  radians for  $D_2$  compared to  $D_1$ , thus maintaining conservation

of energy. The relationship between the intensity measured at the detectors and the phase difference between the interfering light waves for an ideal interferometer is shown in Figure 3.

For the measurement of a small time-dependent phase shift  $\Delta\varphi(t)$ , such as that produced via light absorption in PTI, it is necessary to consider the interferometric phase  $\varphi$ . Due to the sinusoidal relationship between the measured signals and  $\varphi$ , the relationship between a small change in the phase difference between the waves  $\Delta\varphi(t)$  and the change of the measured intensity  $\Delta I$  is not constant and depends on  $\varphi$ . This property of the measurement is shown in Figure 3. At  $t = (2n + 1)\frac{\pi}{2}$ , where  $n = 0, 1, 2, \dots$ , the slope  $\left|\frac{\Delta I}{\Delta\varphi}\right|$  and thus the sensitivity of the measurement is maximised. In order to take advantage of this, PTI measurements are typically performed in phase quadrature by actively regulating  $\varphi$  to  $(2n + 1)\frac{\pi}{2}$ . Previous PTI instruments (Moosmüller and Arnott, 1996; Sedlacek, 2006; Sedlacek and Lee, 2007) have used a number of different methods to regulate the phase difference between the two arms of the interferometer, with the application of a piezo-electric element to move one of the interferometer optical elements being the most common solution.

At the quadrature points the relationship between a sufficiently small phase change  $\Delta\varphi(t)$  (e.g. for  $\sin \Delta\varphi \cong \Delta\varphi$ ) and the measured intensities can be approximated by:

$$\Delta\varphi(t) \cong \frac{I_1 - I_2}{I_1 + I_2} \quad (5)$$

where  $I_1$  and  $I_2$  are the intensities of light measured by the detectors in the two outputs of the interferometer as a function of time. In the ideal case of a light source of constant intensity and an optically thin medium,  $I_1 + I_2$  is constant.

Phase shifts may additionally arise from sources other than the photothermal effect, such as from acoustic noise and changes of the length of the interferometer due to vibrations. Low frequency noise can be separated from phase shifts due to the photothermal effect by modulating the pump laser at a higher frequency and restricting the detection bandwidth to this frequency. This is typically performed experimentally using a lock-in amplifier. Unwanted variations in phase (phase noise) can be reduced through the choice of the interferometer geometry. The current preferred interferometer geometry for aerosol measurements is a folded one, for example a folded Jamin interferometer (Moosmüller and Arnott, 1996; Sedlacek, 2006). This design minimises the influence of interferometric noise by placing both arms of the interferometer in parallel and close proximity and through the use of an etalon and retroreflector. The use of these two optical components ensures that any noise that is coupled into the interferometer affects both arms equally and thus cancels out.

### 3. Modulated single-beam PTI

Here we present a new PTI configuration, which we have named the MSPTI, in which the pump and probe beams are replaced by a single modulated laser beam. This beam has the same optical path as the probe beam in a conventional PTI set up and is modulated between two sufficiently different intensity levels. A larger intensity difference between the levels leads to an increased PTI signal, however signal to noise limitations restrict the choice of  $I_{low}$  in the case that measurements are made during the laser low phase. In this study  $I_{low}$  was set to  $\frac{1}{10}I_{high}$ , which allowed a qualitative indication of the signal response during the cooling phase. The major advantage of the modulated configuration is the simplified optical alignment of the system. As a single beam fulfils both pump and probe functions, the pump-probe co-incident volume is, by definition, the entire beam volume. This ensures maximum heating and detection sensitivities and enables the use of significantly lower pump

intensities, thus reducing the potential for the destruction of the sample and measurement noise arising from heating of the interferometer optical components.

The MSPTI configuration requires a different approach to signal evaluation than standard PTI. As the probe beam is not maintained at constant intensity, additional data analysis steps are required in order to optimally extract the PTI signal encoded on the modulated laser beam. The MSPTI configuration also places additional constraints on the single laser employed. The standard requirements for the interferometric probe beam of low noise and significant coherence length remain, but high CW power is additionally required. For the MSPTI prototype, we have chosen to employ a diode pumped solid state laser operating at 532 nm, modulated with an external acousto-optic modulator to improve rise and fall times as well as laser stability.

The measured phase shift in the MSPTI is equivalent to the two-beam case, with only slight modification required to the formulae. The pump and probe subscripts are dropped from Equation 2 and  $P_{pump}$  is replaced by  $\Delta P$ , the modulated laser power in the measurement chamber (for the case of a 50:50 beam splitter and no optical losses  $\Delta P$  is half of the modulated power exiting the laser). Performing these substitutions results in Equation 6, which is valid for the simplified case of a laser beam with constant diameter.

$$\Delta\varphi = \frac{2\pi(n-1)}{\lambda T \rho C_p} \frac{\Delta P}{A} b_{abs} \Delta t \quad (6)$$

For the case of a focused laser beam with a Gaussian intensity distribution with focal point in the middle of measurement chamber, it can be shown (see the Supplementary Information) that the phase change due to the PTI effect is:

$$\Delta\varphi = \frac{2\pi(n-1)}{\lambda T \rho C_p} \frac{2\Delta P}{\lambda} \tan^{-1}\left(\frac{a}{z_r}\right) b_{abs} \Delta t \quad (7)$$

where  $2a$  is the length of the measurement chamber and  $z_r$  is the Rayleigh distance for the modulated beam focused in the middle of the measurement chamber. For the case where the length of the chamber is twice the Rayleigh distance  $a = z_r$ , and the sensitivity becomes  $\cong 0.79 \cdot \frac{2\Delta P}{\lambda}$ . The sensitivity of the measurement approaches the limit  $\tan^{-1}\left(\frac{a}{z_r}\right) \approx \frac{\pi}{2} \cdot \frac{2\Delta P}{\lambda}$ , for  $a > 2 \cdot z_r$ . Thus, the maximum sensitivity achievable for an arbitrary interaction (chamber) length is limited and cannot be further increased through improved focusing.

Unlike for the case of standard PTI, in which the small phase changes are measured from a stable CW laser intensity, the MSPTI signal is dominated by the modulation of the laser intensity. Thus, lock-in detection cannot be directly performed with the difference signal of the interferometer outputs and a normalisation step is required. This is performed by normalising the time-dependent difference signal from the detectors by the total light intensity (refer to Equation 5 above). This step accounts for the dependence of the phase change signal on the total light intensity of the interferometer beam. An example of this normalisation step is shown in Figure 4.

Thus, a new PTI signal processing method was developed to address the additional complications of the MSPTI method, when compared to standard pump-probe PTI. Signal processing was performed in software, after digitising the raw signals from the photodetectors. As the desired quantity from the measurement is the magnitude of the light absorption by the aerosol, it was sufficient to analyse the heating (or high) phases alone. To ensure maximum signal-to-noise ratio and avoid reducing the data further, the heating curves were analysed in full. An example heating curve calculated from Eq. 5 for approximately  $100 \mu\text{g m}^{-3}$  (eBC) of electrical discharge soot is shown in Figure 5. From Eq. 6 it could be expected that the phase shift due to light absorption should increase linearly with the duration of the heating phase  $\Delta t$ . However, the example heating curve shown in Figure 5 deviates considerably from linearity as the heating phase progresses. This is due to the loss of absorbed energy in the form

of heat out of the sensing (laser) volume with time. Taken to the limit of  $\Delta t \rightarrow \infty$  for a non-modulated laser beam, Eq. 6 implies that  $\Delta\phi \rightarrow \infty$ , however in reality,  $\Delta\phi$  approaches equilibrium as the heat arising from absorption of the laser beam is balanced by the heat flowing out of the detection volume. Thus, Equation 6 is only valid for heating times  $\Delta t$  shorter than a characteristic time.

Empirically, the best fit to the data was found to be an exponential of the form:

$$\Delta\phi(t_{heat}) = a \left(1 - e^{-\frac{t}{\tau}}\right) + c \quad (8)$$

where  $a$  is a parameter representing the limit of the phase change due to the temperature increase of the sample volume due to light absorption and temperature loss outside of the laser beam volume,  $\tau$  is the mean lifetime of the cooling process and is dependent on the beam geometry and  $c$  is the absolute offset from phase quadrature. Equation 8 is closely related to Newton's Law of Cooling adapted to be expressed in terms of phase shift and with the addition of a heating term due to light absorption during the heating cycle. An example of the least squares best fit of this form is shown as the dashed line in Figure 5.

As the characteristic cooling time  $\tau$  is predominantly dependent on the geometry of the heating/sensing volume and this does not change during measurements, further simplification of the fit to obtain  $\Delta\phi(t_{heat})$  is possible. It was found that for a specific range of heating times that the exponential fit could be approximated with a linear one, with slope  $\frac{d\Delta\phi(t_{heat})}{dt}$ . An example least squares linear fit to a heating curve is shown as the dotted line in Figure 5. The quality of the linear fit to the data appears poor; however, it still contains the required information for the calculation of  $\Delta\phi(t_{heat})$  and  $b_{abs}$  when calibrated using a species of known absorption. The interested reader is directed to the Supplementary Information for additional details.

It should be noted that this analysis only holds for the specific cases of a single-beam modulated interferometer and a two-beam interferometer with exactly equal perfectly overlapped pump and probe volumes. For the general case of significantly different pump and probe beam geometries, the dynamics of the system will differ considerably from those obtained in this work (Monson et al., 1989).

#### 4. Modulated single-beam PTI experimental setup

The physical layout of the interferometer is based on the folded Jamin interferometers of Moosmüller and Sedlacek (Moosmüller and Arnott, 1996; Sedlacek, 2006) and is shown in Figure 6. The etalon in the Jamin design has been replaced by separate beam splitter and mirror optics, which are mounted in a solid metal block. The overlap of the interfering beams can be adjusted with the positioning of the mirror by way of thumbscrews. The resulting layout is a folded Mach-Zehnder interferometer. As both the reference and measurement beams are incident on the beam splitter and mirror, the effects of mechanical (vibrational) noise are reduced when compared to standard Michelson or Mach-Zehnder designs. The insensitivity to mechanical noise is not as complete as for the Jamin design, as the two optical elements are able to move with respect to each other, but the design does allow for flexibility in the design of the aerosol chamber.

The MSPTI design additionally requires a different aerosol chamber design compared to standard PTI instruments. In the case of MSPTI, the single modulated beam is present in both reference and measurement arms of the interferometer and therefore a difference in aerosol composition between the reference and measurement arms is required. In the current MSPTI prototype the aerosol chamber consists of three isolated cells, one for the

270 measurement arm (sample) and two for the reference arm. A HEPA-grade absolute filter separates the sample and reference cells. A schematic of the flow set up for the MSPTI instrument is shown in Figure 7.

The effective perfect beam overlap for the MSPTI in both sample and reference chambers confers an additional advantage – the ability to directly subtract absorption by gaseous species. As the light absorbing gaseous species are present in the same concentration in both arms of the interferometer, the photothermal effect due to these  
275 gaseous species is the same and the net phase difference is zero. This compensation of the gas absorption requires both equal laser intensity in the sample and reference arms of the interferometer and equal sensitivity due to the positioning of the two focal points. Equal intensity can be achieved with a 50:50 beam splitter (at the laser wavelength). Equal sensitivities also require that the Rayleigh distance from the focal points lie entirely within the respective chambers. Fulfilment of these prerequisites thus enables the determination of the light absorption of the  
280 aerosol only, even though the complex sample mixture may additionally feature absorbing gases and light scattering aerosol.

In the MSPTI instrument, phase quadrature is actively regulated using a pressure cell in the reference beam path. As the refractive index of a gas depends on pressure, the optical path length of the reference path can be adjusted by varying the pressure in the cell (positioned between the reference arm and the retroreflector in Figure 6). The  
285 quadrature regulation is performed at frequencies below 1 Hz in order to counteract slow changes in the optical path lengths, such as from thermal drifts or changes in refractive index of the gas. The advantage of this method is its simplicity (does not require the production of custom optics) and the lack of moving parts.

The alignment of the interferometer is comparatively simple. The modulated beam is coupled into the interferometer block at an angle of 45 degrees relative to the beam splitter, after which the retroreflector is adjusted  
290 so that the returning beams pass through the cells in the aerosol chamber as required. The focusing lens is then inserted before the interferometer block and its position adjusted so that the focus is centred within the aerosol chamber. Finally, the overlap of the interfering beams is optimised by adjusting the mirror in the interferometer block until it is parallel with the beam splitter and maximum interferometric contrast is acquired. No further adjustment of beam overlap is required.

295 The PTI instrument is mounted on an optical breadboard (Thorlabs, B60120A). Solid borne vibrations are damped using a set of passive vibration isolators (Thorlabs, PWA090). The interferometer is housed within a metal box lined with acoustic foam for isolation from air currents and external noise sources.

The laser source is a diode pumped solid-state laser (Laser Quantum, GEM 450 mW) at 532 nm. The laser power was regulated at 200 mW in this study. Before entering the interferometer the beam first passes through a half-  
300 wave plate (Thorlabs, WPH05M-532), which rotates the polarisation to vertical. Intensity modulation of the beam is performed with an acousto-optic modulator (AOM) (AA Optoelectronic, AA.MT110-A1,5-VIS) and the 0<sup>th</sup> order output is selected in order to maximise the available laser power for PTI. Subsequently, the beam is expanded by a factor of three by a Galilean beam expander and the polarisation rotated by 45° for optimal splitting at the non-polarising beam splitter in the interferometer. All mirrors employed in the interferometer are broadband  
305 dielectric mirrors designed for use at visible wavelengths (Thorlabs, BB1-E02).

The modulated interferometer in this work is of folded Mach-Zehnder design and consists of a broadband dielectric mirror (Thorlabs, BBSQ2-E02), a 50:50 amplitude splitting beam splitter optimised for 532 nm (Thorlabs, BSW4R-532) and a 50.8 mm diameter retroreflector (Edmund Optics, #49-666). The mirror and beam splitter are mounted into a single custom machined metal block in order to reduce relative movements of the optics and shift  
310 mechanical vibrations to higher frequencies. The mirror tilt is adjustable in two planes via thumbscrews in order



to align the beams and achieve maximum interferometric contrast (97% typical). The laser focus inside the interferometer was checked using a CCD camera (Basler, acA1300-30um) and optimised by adjusting the position of the focusing lens (Thorlabs, LA1908-A).

315 Custom-built aerosol and pressure chambers are situated within the interferometer. Each of the chambers consists of three individual cells, which are separately sealed using optical windows (Edmund Optics, #46-100) with O-rings. The reference arm of the interferometer consists of the two outer cells of each chamber. The central sample cell comprises the measurement arm, through which the laser passes twice. Samples are introduced into the central cell and then are either exhausted (NO<sub>2</sub> calibration measurements) or flow through an absolute filter and then each outer cell in series (standard aerosol measurements) as shown in Figure 5. The gas flow is subsequently measured  
320 using a flowmeter (WISAG, 1000 series). One outer cell of the pressure cell is connected to a regulated pressure valve (Parker, 980-005101-015) and is controlled by a software based proportional-integral (PI) controller.

In the detection component, the diverging beams are refocused using biconvex lenses (Thorlabs, LB1945-A-ML) and the optical power reduced with ND filters (Thorlabs, NE10A). Detection of the interfering laser beams is performed with a photodiode in each interferometer output (Thorlabs, DET36A) operating in photoconductive  
325 mode. The use of two detectors allows the rejection of false signals, such as changes in laser intensity. The detected photocurrents are converted into voltages using a 1.2 kΩ resistor in parallel to the photodiode and subsequently digitised (National Instruments, NI USB-6356).

Carbonaceous aerosols are generated with a spark discharge soot generator (PALAS, GFG 1000). Argon (Messer, 4.8) is used as the inert carrier gas for the discharge and subsequent transport of the generated particles.  
330 Comparison measurements of eBC concentrations are performed using an Aethalometer (Aerosol d.o.o., AE33). PTI is an *in situ* light absorption measurement technique and as such, it is possible to use an absorbing gas to calibrate the sensitivity of the instrument (Lack et al., 2006). In the visible range, NO<sub>2</sub> and O<sub>3</sub> gases have the highest absorption cross-sections and NO<sub>2</sub> was chosen as the calibration gas for this study. The initial characterisation of the MSPTI was additionally performed with NO<sub>2</sub>, in order to determine the optimal operating  
335 conditions for the instrument. The optimal operating frequency can be determined by investigating the relationship between the duration of the laser high period (heating time) and the resultant PTI signal. NO<sub>2</sub> concentrations are prepared from a mixture of 1 ppm NO<sub>2</sub> in synthetic air (Messer) in excess synthetic air (Messer, 5.6) using mass flow controllers (Voegtlin Instruments, red-y GSC-B9SA-DD23 and -DD26) and a NO<sub>2</sub> monitor (Horiba, APNA-370) was available for reference concentration measurements.

## 340 5. Results

As can be seen in Equation 6, the measured phase shift for PTI is linearly dependent on the heating time. This however, is only valid for sufficiently short heating times. If the heating time exceeds a characteristic value  $\tau > \frac{w^2}{D}$ , dependent on the beam radius  $w$  and the gas thermal diffusivity  $D$ , then this equality is no longer maintained as heat flows out of the sensing volume during the measurement and no longer contributes to the signal (Monson  
345 et al., 1989). This is observed as a flattening of the heating curves with increasing heating time and results in the apparent reduction of absorption. The calculations of Monson *et al.* were performed for collimated laser beams and under the condition that the probe beam diameter was much smaller than that of the pump beam. In the current case, the pump and probe diameters are equal and vary with  $z$ , leading to a non-cylindrically symmetric temperature distribution along  $z$ . If we however assume an average beam radius of 0.1 mm within the sample cell,

350 then  $\frac{w^2}{D}$  becomes 0.53 ms for a gas thermal diffusivity of 19 mm<sup>2</sup> s<sup>-1</sup>. This is in reasonable agreement with the heating curve shown in Figure 5, in which the deviation of the observed phase shift from linearity begins at ~ 0.7 ms.

In order to determine the optimal modulation frequency of the laser beam, the PTI signal for 1 ppm NO<sub>2</sub> was measured for a range of different modulation frequencies. The results of these measurements are shown in Figure 355 8 as a function of heating time. In order to enable comparison with previous works, the measured signals have been converted to phase shifts with units of radians. At shorter heating times the PTI signal is observed to increase linearly as a function of the heating time, in agreement with Equation 6. The transition out of the linear regime occurs at heating time of ~5 ms (~100 Hz modulation frequency). This result is consistent with the value previously reported by Sedlacek (Sedlacek, 2006) for a two beam experimental PTI set up.

360 It must be noted that all heating curves recorded in this study were non-linear, even for heating times within the linear regime. The linear relationship between PTI signal and heating time for shorter heating times only implies that the shape of the heating curves remains constant for heating times within this range. Heating curves of a constant shape are evaluated consistently by the chosen linear-fit mechanism, thus allowing the transfer of calibration measurements from one heating time to another. Outside of the linear regime, calibration measurements 365 cannot easily be transferred from one heating time to another, however measurements performed for an arbitrary heating time are still in good agreement with Eq. 6 as long as the shape of the heating curves remains constant (e.g. with concentration of the light absorbing species).

A modulation frequency of 91 Hz was chosen in this study in order to avoid a significant noise band around 100 Hz, which was observed during the initial laboratory measurements. Operation at 91 Hz ensured a larger signal 370 than operation at frequencies above 100 Hz, but meant that the MSPTI was operated outside of the afore-described linear regime. Therefore, calibration measurements were also performed at 91 Hz such that the application of Eq. 6 was valid for this modulation frequency. This is explained further in the following paragraphs detailing the calibration procedure employed in this work.

The sensitivity of the instrument was experimentally determined from the MSPTI signal dependence on NO<sub>2</sub> 375 concentration. For these measurements, the aerosol chamber was connected so that the NO<sub>2</sub> flowed through the sample chamber and then was exhausted (see Figure 7a, calibration configuration). The reference chambers were filled with synthetic air for the calibration procedure. Measurements were performed for 0.2 to 1 ppm of NO<sub>2</sub> in synthetic air at a flow rate of 0.5 l min<sup>-1</sup> and the results are plotted in Figure 9. The two data sets represent two separate measurements, where the concentration of NO<sub>2</sub> was firstly increased stepwise from 0 ppm to 1 ppm and 380 then decreased back to 0 ppm. No obvious measurement hysteresis was observed between the data sets. The offset of the measurement is attributed to PTI signals generated in the optical components in the interferometer.

The data show a clear linear relationship between the PTI signal and NO<sub>2</sub> mixing concentration set in the flow system and show the validity of the developed signal analysis. From the slope of the concentration curve and the absorption cross-section of NO<sub>2</sub> reported in literature the sensitivity ( $\frac{I\Delta P}{A}$ ) for the MSPTI instrument can be 385 calculated with Equation 6. The absorption cross-section of NO<sub>2</sub> used was 1.47x10<sup>-19</sup> cm<sup>2</sup> molecule<sup>-1</sup>, which was obtained by convoluting the data of Vandaele et al. (2002) (accessed from the MPI-Mainz UV-VIS Spectral Atlas (Keller-Rudek et al., 2013)) with a Gaussian function at the reported laser wavelength of 532.075 nm and spectral bandwidth of 30 GHz. Using this value and typical literature values for air at standard temperature and pressure, a sensitivity ( $\frac{I\Delta P}{A}$ ) of 6.80 kW m<sup>-1</sup> is calculated for the MSPTI system. This value is approximately two orders of

390 magnitude lower than the theoretical value determined by measuring the properties of the laser beam at its focus  
(see Supplementary Information for the calculation). One major reason for this is the application of a linear fit to  
the heating curves in this study (see Figure 5 and accompanying text), which is affected by heat loss out of the  
laser volume during the measurement. As Equations 6 and 7 do not account for the heat loss out of the measurement  
volume during the measurement, any signal loss due to this process negatively impacts the magnitude of  $\frac{I\Delta P}{A}$   
395 determined from the calibration measurements. This is true for measurements performed at modulation frequencies  
both within and outside the linear range of Figure 8, but the effect is larger for modulation frequencies outside of  
the linear range. The reason for this can be seen in the linear fit of Figure 5. The non-linear signal due to diffusional  
loss of heat outside the probe volume suppresses measurement of the total amount of energy deposited into the  
system. An exponential fit to the heating curves and subsequent use of  $\frac{a}{\tau}$  for the determination of the phase change  
400 due to light absorption would lead the magnitude of the sensitivity determined from the calibration to approach  
the value obtained by measuring the beam parameters.

Besides the ability to perform a primary calibration of the MSPTI with NO<sub>2</sub>, a further advantage of the instrumental  
design over existing PTI instruments is the ability to directly differentiate aerosol absorption from absorption from  
gaseous species during a standard measurement. This removes the need to intermittently determine the absorption  
405 background from gases using filtered measurements as per other techniques. In the standard measurement  
configuration, aerosol enters the sample chamber, passes an absolute filter and then flows through the reference  
chamber. The absolute filter traps the aerosol particles, but transmits the gas, which then flows through the  
reference chambers. To demonstrate the relative nature of PTI measurements and the advantages of the MSPTI set  
up in separating gas and aerosol absorption, comparison measurements of NO<sub>2</sub> were performed using the  
410 calibration and standard flow set ups. The results of these measurements for a flow rate of 0.5 l min<sup>-1</sup> are presented  
in Figure 10. For the calibration flow set up, the NO<sub>2</sub> is only present in the sample cell and a PTI signal is measured.  
In the standard flow set up, NO<sub>2</sub> is present at the same concentration in the sample and reference cells and no  
signal is observed within experimental error. Thus, with the new MSPTI configuration, aerosol absorption can be  
measured independent of gas absorption, which reduces possible artefacts in the determination of aerosol  
415 absorption in ambient measurements. This is a significant advantage over previous PTI designs, which rely on  
either periodic measurements of the background gas absorption, NO<sub>2</sub> denuders or the measurements of other  
sensors to determine the aerosol absorption from the total absorption.

The detection limit dependence on the averaging time was determined from a baseline measurement with no flow,  
during which no effort was made to control noise in the laboratory. The raw data from the baseline measurement  
420 show a linear drift in apparent absorption most probably due to a slow change in splitting ratio of the beam splitter  
over time, causing a change in the laser power in both arms of the interferometer and therefore the absorption by  
optical elements in the each arm. The exact source of the drift is yet to be confirmed as it is complicated by positive  
and negative contributions by optics in either interferometer arm. The drift, however, could be easily corrected by  
subtracting a linear fit from the data. The data and linear fit for evaluation of the drift are shown in the Supporting  
425 Information. In standard measurements, the baseline is determined at regular intervals with filtered air to account  
for such baseline drifts. The experimentally measured standard deviations for the raw and drift corrected data are  
shown in Figure 11. The Allan deviation of the same data is provided for comparison purposes in Figure S8 in the  
supplementary information. The standard deviation of the raw data is dominated by the observed drift and improves  
very little for longer averaging durations. The standard deviation of the drift corrected data reduces at a rate of

430 approximately  $t^{-\frac{1}{2}}$  up to an averaging time of 60 seconds, after which the rate of reduction reduces. For the calculation of the detection limits of the instrument ( $1\sigma$ ), standard deviations for averaging times of 1, 10, 60 and 300 seconds were compared to the NO<sub>2</sub> calibration measurement in Figure 9. Baseline drift corrected detection limits of our current instrument are summarized in terms of MSPTI signal (rad s<sup>-1</sup>),  $b_{abs}$ , NO<sub>2</sub> concentration and eBC (MAC = 10 m<sup>2</sup> g<sup>-1</sup>) for selected integration times in Table 1.

435 Initial measurements of aerosols were performed by sampling from a reservoir pre-filled with graphitic soot produced using a spark discharge source and diluted with laboratory air. The aerosol was sampled from the volume at a rate of 0.25 l min<sup>-1</sup> through the PTI in the measurement flow configuration using a pump connected to the outlet of the interferometer flow system. Comparison measurements were made using an Aethalometer AE33 sampling at 2 l min<sup>-1</sup>, which was connected separately to the sampling volume in order to reduce the contribution  
 440 from the AE33 pump to the MSPTI signal noise. The AE33 reported loading corrected values. eBC measured at  $\lambda = 520$  nm were converted into absorption coefficients according to

$$b_{abs} = \text{eBC} \cdot \text{MAC}_{instr} \cdot \frac{C_{instr}}{C_{new}} \quad (9)$$

where  $\text{MAC}_{instr} = 13.14$  m<sup>2</sup> g<sup>-1</sup> and  $C_{instr} = 1.57$  are the MAC and filter multiple-scattering enhancement  
 445 parameter values used by the instruments' firmware, respectively. This conversion was corrected with the updated value of  $C_{new}=2.6$  (value from (WMO, 2016), normalized as in (Drinovec et al., 2015)).

The MSPTI signal was converted to  $b_{abs}$  using the sensitivity calculated from the NO<sub>2</sub> calibration (*vide supra*). The results of a typical measurement are presented in Figure 12. The AE33 signal shows strong loading artefacts at high soot concentrations, which are not completely corrected by the internal correction algorithm (Drinovec et al., 2015). Also seen are the automated filter changes of the AE33, indicated by black arrows, which are triggered  
 450 when the light extinction through the sample spot exceeds a pre-specified value. A discontinuity can be seen at around 15:15 in the MSPTI data, indicated by a red arrow, which is assigned to a laser mode hop. This leads to a change in the background level and as such can be corrected using regular background measurements. This artefact has been left uncorrected in order to show the effects of laser noise in the measurement. Improvements to the laser  
 455 cooling system will be made in future development of the interferometer in order to stabilise the temperature and thus lasing mode of the DPSS laser.

The noise visible in the MSPTI signal can be attributed to several sources. Large outliers in the MSPTI signal are attributed to mechanical shocks that are transported to the interferometer. Other obvious outliers can be assigned to imperfect isolation of the interferometer to external acoustic noise sources (pumps, discussions in the laboratory)  
 460 and laser instability. Additionally, noise from the pump in the AE33 is coupled through the sampling reservoir and adds to the baseline noise of the PTI measurement. This can be seen at 17:30 in the measurement data, when an absolute filter was inserted into the sampling line to determine the background signal level, thus better isolating the MSPTI from noise transported from the sampling volume through the sampling line. The standard deviation of the data collected with the filter is a factor of 6.5 lower than the aerosol measurement and shows the need for  
 465 an improvement in decoupling the MSPTI from external noise sources via the sampling line. For this measurement with the associated noise sources, the MSPTI instrument was determined to have a higher detection limit ( $1\sigma$ ) of approximately  $b_{abs} = 17$  Mm<sup>-1</sup> (eBC  $\sim 1.7$   $\mu\text{g m}^{-3}$ ) for an averaging period of 10 seconds, which agrees well with the values determined for the measurements of NO<sub>2</sub> summarised in Table 1.

Future improvements to the MSPTI set up are primarily targeted at the reduction of noise in the measurements.  
 470 The laboratory tests show that an improvement in the isolation of the interferometer is required when operating in

noisy environments. New outer and inner enclosures for the MSPTI are currently being evaluated for this purpose. Improvements to the data analysis system to reduce the detection bandwidth and thus improve noise rejection are ongoing and are expected to bring a significant improvement in the detection limit. First measurements with 400 mW (200 mW per interferometer arm) laser power show a two-fold signal increase with no associated increase in noise. This implies that the system is interferometer noise limited and a further increase of the laser power would lead to an equal improvement in the detection limit. Potential improvement is foreseen by employing a balanced photodiode detector and amplifying the difference signal to better employ the full range of the ADC.

## 6. Conclusion

We have demonstrated a new PTI prototype utilising only a single laser with a significantly improved ease of alignment compared to existing PTI instruments. The MSPTI design also allows for the direct measurement of aerosol absorption in the presence of absorbing gases, which would normally require a complicated correction, a scrubber or secondary measurement of the gas absorption for other *in-situ* aerosol absorption measurements.

With a detection limit of aerosol absorption of  $\sim 10\text{-}20 \text{ Mm}^{-1}$  for an integration time of 10 seconds the MSPTI set up does not currently improve upon the best reported detection limits for PTI measurements of aerosols (see e.g. (Sedlacek and Lee, 2007)), but simplifies its operational use in the field. Improvements to the isolation of the interferometer and data handling and analysis are expected to reduce the detection limit to the point where unattended field measurements of ambient aerosol concentrations are possible.

The improvement in handling and alignment of the interferometer is significant and of great advantage when operated by non-experts in field measurements. Furthermore, the ability to directly measure aerosol absorption without bias from light absorbing gaseous species further reduces the potential for measurement artefacts due to concentration changes of these species. It also opens up the potential to employ the MSPTI in emission measurements, where the concentrations of absorbing gaseous species can be significant and fluctuate rapidly.

## Author contributions

BV, JR, PS and EW designed and developed the MSPTI. BV and JR carried out the experiments and analysed the data. LD and GM provided input into experimental design. BV wrote the manuscript, with JR, EW, LD and GM providing valuable additions.

## Competing interests

The authors declare that they have no conflict of interest. LD and GM are employed at the potential future manufacturer of the instrument.

## Acknowledgements

The research leading to these results has received funding from the Swiss National Science Foundation (SNSF) grant no. 200021\_172649, the 16ENV02 Black Carbon project of the European Union through the European Metrology Programme for Innovation and Research (EMPIR), and the EUROSTARS grant E!11386 IMALA. We thank J. Lee for his help and encouragement in the initial stages of the project, M. Gysel (PSI) for the use of the

505 AE33, C. Hüglin (EMPA) for the loan of the NO<sub>2</sub> Monitor, A. Sedlacek, M. Krejci and H. Looser for fruitful discussions, S. Sjögren for the initial work on evaluating PTI configurations and A. Meier, M. Wipf, D. Egli, P. Specht for their assistance and technical expertise. We thank AMES d.o.o. for technical support.

A patent application was filed on the described technology.

510

## References

515 Arnott, W. P., Moosmüller, H., and Walker, J. W.: Nitrogen dioxide and kerosene-flame soot calibration of photoacoustic instruments for measurements of light absorption by aerosols, *Rev. Sci. Instrum.*, 71, 4545-4552, <https://doi.org/10.1063/1.1322585>, 2000.

Arnott, W. P., Moosmüller, H., Sheridan, P. J., Ogren, J. A., Raspet, R., Slaton, W. V., Hand, J. L., Kreidenweis, S. M., and Collett, J. L.: Photoacoustic and filter-based ambient aerosol light absorption measurements: Instrument comparisons and the role of relative humidity, *J. Geophys. Res. Atmos.*, 108, <https://doi.org/10.1029/2002jd002165>, 2003.

520 Arnott, W. P., Hamasha, K., Moosmüller, H., Sheridan, P. J., and Ogren, J. A.: Towards aerosol light-absorption measurements with a 7-wavelength Aethalometer: Evaluation with a photoacoustic instrument and 3-wavelength Nephelometer, *Aerosol Sci. Technol.*, 39, 17-29, <https://doi.org/10.1080/027868290901972>, 2005.

525 Backman, J., Schmeisser, L., Virkkula, A., Ogren, J. A., Asmi, E., Starkweather, S., Sharma, S., Eleftheriadis, K., Uttal, T., Jefferson, A., Bergin, M., Makshtas, A., Tunved, P., and Fiebig, M.: On Aethalometer measurement uncertainties and an instrument correction factor for the arctic, *Atmos. Meas. Tech.*, 10, 5039-5062, <https://doi.org/10.5194/amt-10-5039-2017>, 2017.

Bond, T. C., Anderson, T. L., and Campbell, D.: Calibration and intercomparison of filter-based measurements of visible light absorption by aerosols, *Aerosol Sci. Technol.*, 30, 582-600, <https://doi.org/10.1080/027868299304435>, 1999.

530 Bond, T. C., Doherty, S. J., Fahey, D. W., Forster, P. M., Berntsen, T., DeAngelo, B. J., Flanner, M. G., Ghan, S., Karcher, B., Koch, D., Kinne, S., Kondo, Y., Quinn, P. K., Sarofim, M. C., Schultz, M. G., Schulz, M., Venkataraman, C., Zhang, H., Zhang, S., Bellouin, N., Guttikunda, S. K., Hopke, P. K., Jacobson, M. Z., Kaiser, J. W., Klimont, Z., Lohmann, U., Schwarz, J. P., Shindell, D., Storelvmo, T., Warren, S. G., and Zender, C. S.: Bounding the role of black carbon in the climate system: A scientific assessment, *J. Geophys. Res. Atmos.*, 118, 5380-5552, <https://doi.org/10.1002/jgrd.50171>, 2013.

535 Collaud Coen, M., Weingartner, E., Apituley, A., Ceburnis, D., Fierz-Schmidhauser, R., Flentje, H., Henzing, J. S., Jennings, S. G., Moerman, M., Petzold, A., Schmid, O., and Baltensperger, U.: Minimizing light absorption measurement artifacts of the Aethalometer: Evaluation of five correction algorithms, *Atmos. Meas. Tech.*, 3, 457-474, <https://doi.org/10.5194/amt-3-457-2010>, 2010.

540 Davies, N. W., Cotterell, M. I., Fox, C., Szpek, K., Haywood, J. M., and Langridge, J. M.: On the accuracy of aerosol photoacoustic spectrometer calibrations using absorption by ozone, *Atmos. Meas. Tech.*, 11, 2313-2324, <https://doi.org/10.5194/amt-11-2313-2018>, 2018.

Davis, C. C., and Petuchowski, S. J.: Phase fluctuation optical heterodyne spectroscopy of gases, *Appl. Opt.*, 20, 2539-2554, <https://doi.org/10.1364/ao.20.002539>, 1981.

545 Drinovec, L., Močnik, G., Zotter, P., Prévôt, A. S. H., Ruckstuhl, C., Coz, E., Rupakheti, M., Sciare, J., Müller, T., Wiedensohler, A., and Hansen, A. D. A.: The "dual-spot" Aethalometer: An improved measurement of aerosol black carbon with real-time loading compensation, *Atmos. Meas. Tech.*, 8, 1965-1979, <https://doi.org/10.5194/amt-8-1965-2015>, 2015.

- 550 Drinovec, L., Gregorič, A., Zotter, P., Wolf, R., Bruns, E. A., Prévôt, A. S. H., Petit, J.-E., Favez, O., Sciare, J., Arnold, I. J., Chakrabarty, R. K., Moosmüller, H., Filep, A., and Močnik, G.: The filter-loading effect by ambient aerosols in filter absorption photometers depends on the coating of the sampled particles, *Atmos. Meas. Tech.*, 10, 1043-1059, <https://doi.org/10.5194/amt-10-1043-2017>, 2017.
- 555 Fluckiger, D. U., Lin, H. B., and Marlow, W. H.: Composition measurement of aerosols of submicrometer particles by phase fluctuation absorption-spectroscopy, *Appl. Opt.*, 24, 1668-1681, <https://doi.org/10.1364/AO.24.001668>, 1985.
- Fulghum, S. F., and Tilleman, M. M.: Interferometric calorimeter for the measurement of water-vapor absorption, *J. Opt. Soc. Am.*, 8, 2401-2413, <https://doi.org/10.1364/josab.8.002401>, 1991.
- Janssen, N. A., Gerlofs-Nijland, M. E., Lanki, T., Salonen, R. O., Cassee, F., Hoek, G., Fischer, P., Brunekreef, B., and Krzyzanowski, M.: Health effects of black carbon, World Health Organization, 2012.
- 560 Janssen, N. A. H., Hoek, G., Simic-Lawson, M., Fischer, P., van Bree, L., ten Brink, H., Keuken, M., Atkinson, R. W., Anderson, H. R., Brunekreef, B., and Cassee, F. R.: Black carbon as an additional indicator of the adverse health effects of airborne particles compared with PM10 and PM2.5, *Environ. Health Perspect.*, 119, 1691-1699, <https://doi.org/10.1289/ehp.1003369>, 2011.
- 565 Keller-Rudek, H., Moortgat, G. K., Sander, R., and Sörensen, R.: The MPI-Mainz UV/VIS spectral atlas of gaseous molecules of atmospheric interest, *Earth Syst. Sci. Data*, 5, 365-373, <https://doi.org/10.5194/essd-5-365-2013>, 2013.
- Lack, D. A., Lovejoy, E. R., Baynard, T., Pettersson, A., and Ravishankara, A. R.: Aerosol absorption measurement using photoacoustic spectroscopy: Sensitivity, calibration, and uncertainty developments, *Aerosol Sci. Technol.*, 40, 697-708, <https://doi.org/10.1080/02786820600803917>, 2006.
- 570 Lack, D. A., Cappa, C. D., Covert, D. S., Baynard, T., Massoli, P., Sierau, B., Bates, T. S., Quinn, P. K., Lovejoy, E. R., and Ravishankara, A. R.: Bias in filter-based aerosol light absorption measurements due to organic aerosol loading: Evidence from ambient measurements, *Aerosol Sci. Technol.*, 42, 1033-1041, <https://doi.org/10.1080/02786820802389277>, 2008.
- 575 Lack, D. A., Richardson, M. S., Law, D., Langridge, J. M., Cappa, C. D., McLaughlin, R. J., and Murphy, D. M.: Aircraft instrument for comprehensive characterization of aerosol optical properties, part 2: Black and brown carbon absorption and absorption enhancement measured with photo acoustic spectroscopy, *Aerosol Sci. Technol.*, 46, 555-568, <https://doi.org/10.1080/02786826.2011.645955>, 2012.
- 580 Lack, D. A., Moosmueller, H., McMeeking, G. R., Chakrabarty, R. K., and Baumgardner, D.: Characterizing elemental, equivalent black, and refractory black carbon aerosol particles: A review of techniques, their limitations and uncertainties, *Anal. Bioanal. Chem.*, 406, 99-122, <https://doi.org/10.1007/s00216-013-7402-3>, 2014.
- Langridge, J. M., Richardson, M. S., Lack, D. A., Brock, C. A., and Murphy, D. M.: Limitations of the photoacoustic technique for aerosol absorption measurement at high relative humidity, *Aerosol Sci. Technol.*, 47, 1163-1173, <https://doi.org/10.1080/02786826.2013.827324>, 2013.
- 585 Lee, J.: Performance test of MicroAeth® AE51 at concentrations lower than 2 µg/m<sup>3</sup> in indoor laboratory, *Appl. Sci.*, 9, 2766, 2019.
- Lee, J., and Moosmuller, H.: Measurement of light absorbing aerosols with folded-Jamin photothermal interferometry, *Sensors (Basel)*, 20, <https://doi.org/10.3390/s20092615>, 2020.
- 590 Lelieveld, J., Klingmuller, K., Pozzer, A., Poschl, U., Fnais, M., Daiber, A., and Munzel, T.: Cardiovascular disease burden from ambient air pollution in Europe reassessed using novel hazard ratio functions, *Eur. Heart J.*, 40, 1590-1596, <https://doi.org/10.1093/eurheartj/ehz135>, 2019.
- Lin, H. B., and Campillo, A. J.: Photothermal aerosol absorption-spectroscopy, *Appl. Opt.*, 24, 422-433, <https://doi.org/10.1364/AO.24.000422>, 1985.

- Linke, C., Ibrahim, I., Schleicher, N., Hitzenberger, R., Andreae, M. O., Leisner, T., and Schnaiter, M.: A novel single-cavity three-wavelength photoacoustic spectrometer for atmospheric aerosol research, *Atmos. Meas. Tech.*, 9, 5331–5346, <https://doi.org/10.5194/amt-9-5331-2016>, 2016.
- Mazzoni, D. L., and Davis, C. C.: Trace detection of hydrazines by optical homodyne interferometry, *Appl. Opt.*, 30, 756-764, <https://doi.org/10.1364/AO.30.000756>, 1991.
- Monson, B., Vyas, R., and Gupta, R.: Pulsed and CW photothermal phase shift spectroscopy in a fluid medium: Theory, *Appl. Opt.*, 28, 2554-2561, <https://doi.org/10.1364/AO.28.002554>, 1989.
- Moosmüller, H., and Arnott, W. P.: Folded Jamin interferometer: A stable instrument for refractive-index measurements, *Opt. Lett.*, 21, 438-440, <https://doi.org/10.1364/ol.21.000438>, 1996.
- Moosmüller, H., Arnott, W. P., and Rogers, C. F.: Methods for real time, in situ measurement of aerosol light absorption, *J. Air Waste Manag. Assoc.*, 47, 157-166, <https://doi.org/10.1080/10473289.1997.10464430>, 1997.
- Moosmüller, H., Chakrabarty, R. K., and Arnott, W. P.: Aerosol light absorption and its measurement: A review, *J. Quant. Spectros. Radiat. Transfer*, 110, 844-878, <https://doi.org/10.1016/j.jqsrt.2009.02.035>, 2009.
- Murphy, D. M.: The effect of water evaporation on photoacoustic signals in transition and molecular flow, *Aerosol Sci. Technol.*, 43, 356-363, <https://doi.org/10.1080/02786820802657392>, 2009.
- Myhre, G., Shindell, D., Bréon, F.-M., Collins, W., Fuglested, J., Huang, J., Koch, D., Lamarque, J.-F., Lee, D., Mendoza, B., Nakajima, T., Robock, A., Stephens, G., Takemura, T., and Zhang, H.: Anthropogenic and natural radiative forcing. In: *Climate change 2013: The physical science basis. Contribution of working group I to the fifth assessment report of the Intergovernmental Panel on Climate Change*, Cambridge University Press, Cambridge, United Kingdom and New York, NY, USA, 2013.
- Nakayama, T., Suzuki, H., Kagamitani, S., Ikeda, Y., Uchiyama, A., and Matsumi, Y.: Characterization of a three wavelength Photoacoustic Soot Spectrometer (PASS-3) and a Photoacoustic Extinctionmeter (PAX), *J. Meteorol. Soc. Japan*, 93, 285-308, <https://doi.org/10.2151/jmsj.2015-016>, 2015.
- Ogren, J. A., Wendell, J., Andrews, E., and Sheridan, P. J.: Continuous light absorption photometer for long-term studies, *Atmos. Meas. Tech.*, 10, 4805-4818, <https://doi.org/10.5194/amt-10-4805-2017>, 2017.
- Petzold, A., Kramer, H., and Schönlinner, M.: Continuous measurement of atmospheric black carbon using a multi-angle absorption photometer, *Environ. Sci. Pollut. Res.*, 9, 78-82, 2002.
- Petzold, A., Ogren, J. A., Fiebig, M., Laj, P., Li, S. M., Baltensperger, U., Holzer-Popp, T., Kinne, S., Pappalardo, G., Sugimoto, N., Wehrli, C., Wiedensohler, A., and Zhang, X. Y.: Recommendations for reporting "black carbon" measurements, *Atmos. Chem. Phys.*, 13, 8365-8379, <https://doi.org/10.5194/acp-13-8365-2013>, 2013.
- Raspet, R., Slaton, W. V., Arnott, W. P., and Moosmuller, H.: Evaporation-condensation effects on resonant photoacoustics of volatile aerosols, *J. Atmos. Oceanic Technol.*, 20, 685-695, <https://doi.org/10.1175/1520-0426>, 2003.
- Schnaiter, M., Horvath, H., Mohler, O., Naumann, K. H., Saathoff, H., and Schock, O. W.: UV-VIS-NIR spectral optical properties of soot and soot-containing aerosols, *J. Aerosol Sci.*, 34, 1421-1444, [https://doi.org/10.1016/S0021-8502\(03\)00361-6](https://doi.org/10.1016/S0021-8502(03)00361-6), 2003.
- Schnaiter, M., Schmid, O., Petzold, A., Fritzsche, L., Klein, K. F., Andreae, M. O., Helas, G., Thielmann, A., Gimmler, M., Mohler, O., Linke, C., and Schurath, U.: Measurement of wavelength-resolved light absorption by aerosols utilizing a UV-VIS extinction cell, *Aerosol Sci. Technol.*, 39, 249-260, <https://doi.org/10.1080/027868290925958>, 2005.
- Sedlacek, A., and Lee, J.: Photothermal interferometric aerosol absorption spectrometry, *Aerosol Sci. Technol.*, 41, 1089-1101, <https://doi.org/10.1080/02786820701697812>, 2007.
- Sedlacek, A. J.: Real-time detection of ambient aerosols using photothermal interferometry: Folded Jamin interferometer, *Rev. Sci. Instrum.*, 77, <https://doi.org/10.1063/1.2205623>, 2006.




Springston, S. R., and Sedlacek, A. J.: Noise characteristics of an instrumental particle absorbance technique, *Aerosol Sci. Technol.*, 41, 1110-1116, <https://doi.org/10.1080/02786820701777457>, 2007.

640 Vandaele, A. C., Hermans, C., Fally, S., Carleer, M., Colin, R., Mérienne, M. F., Jenouvrier, A., and Coquart, B.: High-resolution fourier transform measurement of the NO<sub>2</sub> visible and near-infrared absorption cross sections: Temperature and pressure effects, *J. Geophys. Res. Atmos.*, 107, ACH 3-1-ACH 3-12, <https://doi.org/10.1029/2001JD000971>, 2002.


645 Weingartner, E., Saathoff, H., Schnaiter, M., Streit, N., Bitnar, B., and Baltensperger, U.: Absorption of light by soot particles: Determination of the absorption coefficient by means of Aethalometers, *J. Aerosol Sci.*, 34, 1445-1463, [https://doi.org/10.1016/S0021-8502\(03\)00359-8](https://doi.org/10.1016/S0021-8502(03)00359-8), 2003.

WHO: 7 million premature deaths annually linked to air pollution: <http://www.who.int/mediacentre/news/releases/2014/air-pollution/en/>, access: 18 June 2020, 2014.


650 Zanatta, M., Gysel, M., Bukowiecki, N., Müller, T., Weingartner, E., Areskoug, H., Fiebig, M., Yttri, K. E., Mihalopoulos, N., Kouvarakis, G., Beddows, D., Harrison, R. M., Cavalli, F., Putaud, J. P., Spindler, G., Wiedensohler, A., Alastuey, A., Pandolfi, M., Sellegri, K., Swietlicki, E., Jaffrezo, J. L., Baltensperger, U., and Laj, P.: A European aerosol phenomenology-5: Climatology of black carbon optical properties at 9 regional background sites across Europe, *Atmos. Environ.*, 145, 346-364, <https://doi.org/10.1016/j.atmosenv.2016.09.035>, 2016.



655 **Figure 1 – Scheme of PTI signal generation and measurement. For standard pump-probe PTI measurements the laser power at the low level is  $L = 0$ .**



660 **Figure 2 – A potential realisation of a standard two-beam photothermal interferometer, similar to published configurations (see e.g. (Lee and Moosmüller, 2020)). The pump laser is set to overlap the probe beam in the measurement arm of the interferometer within the aerosol chamber. BS are beam splitters, M a mirror, DM a dichroic mirror, F a band pass filter, BD a beam dump, AC an aerosol chamber and D are detectors.**



665

Figure 3 – The intensity of light measured by each detector in the interferometer as a function of the interferometric phase. The dashed circles indicate the quadrature points, where the light intensity falling on each detector is approximately equal. At these points the relationship between a small phase shift  $\Delta\phi$  and the measured intensities are approximately linear and the sensitivity of the measurement is maximal.

670






Figure 4 – (a) Typical raw signals from the interferometer outputs ( $I_1, I_2$ ; blue and red dotted lines in (a), respectively) measured during a modulation cycle with a strongly absorbing sample. Over the course of the heating (high) phase, an intensity difference arises between the outputs. (b) This effect is seen more clearly in the resultant phase shift, which is calculated by normalising the difference of the raw signals by the total intensity (Eq. 5). The baseline offset has been subtracted.

675




680

Figure 5 – Measured phase shift during the heating phase averaged over a 1 second interval for a strongly absorbing aerosol (approximately  $100 \mu\text{g m}^{-3}$  eBC) in argon. The black dashed line represents the best least squares exponential fit, while the solid line is the linear fit to the data.




685

Figure 6 – Schematic of the MSPTI prototype. The marking M denotes mirrors, while L, AOM, BS, PC, RR, ND and PD denote focusing lenses, acousto-optic modulator, beam splitter, pressure chamber, retroreflector, neutral density filter and photodiode, respectively. The dashed circles show the positions of the beam focus in each arm of the interferometer. An image of the experimental set up is shown in Figure S6 in the Supplementary Information.




690

Figure 7 – Gas flow system for the MSPTI prototype. Calibration measurements are performed as in (a) by filling the sample cell with the calibration gas and the reference cells with non-absorbing synthetic air. All three cells are connected for standard measurements as in (b), with the filtered sample flowing through both reference chambers in sequence.



695


Figure 8 – MSPTI signal dependence on the heating time for 1 ppm of NO<sub>2</sub> in synthetic air. For shorter heating times the PTI signal is linearly dependent on the heating time (dotted line). Error bars represent one standard deviation of the 10 s integrated data.



700

Figure 9 – PTI signal measured for two consecutive measurement series (filled squares for increasing concentration series, filled circles for the decreasing series) with NO<sub>2</sub> concentrations between 0.2 and 1 ppm measured at a flow rate of 0.5 lpm. The data are presented in internal interferometric units (primary y-axis) and absorption units (secondary y-axis, conversion via calibration using literature absorption values for NO<sub>2</sub>. See main text for details). Error bars indicate one standard deviation of the data integrated over 10 seconds. The dashed line represents the best linear fit to the data set.

705



710

Figure 10 - Demonstration of the relative nature of the MSPTI measurement. Circles represent averages over 10 seconds, whereas the blue line is the 100 s moving average. In the measurement flow configuration NO<sub>2</sub> is present in both sample and reference chambers and no longer contributes to the MSPTI signal.

715

Table 1: Limits of detection for the MSPTI for different integration times. Measured phase shifts were converted into absorption coefficients using the conversion factor from the NO<sub>2</sub> calibration presented in Figure 9. The eBC concentration was calculated with a MAC value of 10 m<sup>2</sup> g<sup>-1</sup>. This is lower than the value of 12 m<sup>2</sup> g<sup>-1</sup> obtained by transferring the measurements of ambient BC particles by Zanatta et al. (2016) to 532 nm using an Angström exponent of 1. Electrical discharge generated BC has been shown to have a MAC value lower than that of ambient soot (Schnaiter et al., 2003). As the MAC of the BC from the employed BC generator has not been measured directly, the limits of detection for eBC should be taken as a reference only.

Averaging time [s]	Standard deviation of MSPTI signal [ $10^{-3}$ rad s <sup>-1</sup> ]	Limit of detection		
		b <sub>abs</sub> [Mm <sup>-1</sup> ]	NO <sub>2</sub> [ppb]	eBC [ng m <sup>-3</sup> ]
1	4.13	49.6	165	4960
10	1.22	14.6	49	1460
60	0.61	7.35	25	735
300	0.44	5.3	18	530

720




Figure 11 – The standard deviation of the baseline for drift uncorrected (filled circles) and drift corrected data (filled triangles). The drift corrected data approximately follows a square root dependence (grey line) on the averaging interval up to an averaging interval of 60 seconds. Note the logarithmic scales.

725




Figure 12 – Comparison of measured light absorption for the MSPTI (blue) and AE33 (green) sampling from a common pre-prepared aerosol reservoir filled with graphitic soot prepared with an electrical discharge source. Data points represent a 9 second running average of 1 second data. Black arrows indicate automated filter changes for the AE33 and the red arrow indicates a suspected mode hop by the MSPTI laser.

730