

Intercomparison and characterization of 23 Aethalometers under laboratory and ambient air conditions: Procedures and unit-to-unit variabilities

Andrea Cuesta-Mosquera¹, Griša Močnik^{2,3,4}, Luka Drinovec^{2,3}, Thomas Müller¹, Sascha Pfeifer¹, María Cruz Minguillón⁵, Briel Björn⁶, Paul Buckley⁷, Vadimas Dudoitis⁸, Javier Fernández-García⁹, María Fernández-Amado¹⁰, Joel Ferreira De Brito¹¹, Veronique Riffault¹¹, Harald Flentje⁶, Eimear Heffernan⁷, Nikolaos Kalivitis¹², Athina-Cerise Kalogridis¹³, Hannes Keernik^{14,15}, Luminita Marmureanu¹⁶, Krista Luoma¹⁷, Angela Marinoni¹⁸, Michael Pikridas¹⁹, Gerhard Schauer²⁰, Norbert Serfozo²¹, Henri Servomaa²², Gloria Titos²³, Jesús Yus-Díez^{5,24}, Natalia Ziola²⁵, Alfred Wiedensohler¹

¹Department of Experimental Aerosol and Cloud Microphysics, Leibniz Institute for Tropospheric Research, Leipzig, 04318, Germany

²Department of Condensed Matter Physics, Jožef Stefan Institute, Ljubljana, 1000, Slovenia

³Haze Instruments d.o.o., Ljubljana, 1000, Slovenia

⁴Center for Atmospheric Research, University of Nova Gorica, Ajdovščina, 5270, Slovenia

⁵Institute of Environmental Assessment and Water Research (IDAEA), CSIC, Barcelona, 08034, Spain

⁶Deutscher Wetterdienst (DWD), Meteorologisches Observatorium Hohenpeißenberg, Hohenpeißenberg, 82383, Germany

⁷School of Chemistry, Centre for Research into Atmospheric Chemistry & Environmental Research Institute, University College Cork, Cork, T23 XE10, Ireland

⁸Department of Environmental Research, SRI Center for Physical Sciences and Technology, Vilnius, 10257, Lithuania

⁹Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas, Madrid, 28040, Spain

¹⁰Instituto Universitario de Medio Ambiente-Grupo QANAP, Universidade da Coruña, Oleiros, 15179, Spain

¹¹IMT Lille Douai, Institut Mines-Télécom, Univ. Lille, Centre for Energy and Environment, F-59000, Lille, France

¹²Department of Chemistry, University of Crete, Heraklion, 70013, Greece

¹³Environmental Radioactivity Laboratory, National Centre of Scientific Research, Aghia Paraskevi, 15310, Greece

¹⁴Air Quality Management Department, Estonian Environmental Research Centre, Tallinn, 10617, Estonia

¹⁵Department of Software Sciences, Tallinn University of Technology, Tallinn, 12616, Estonia

¹⁶National Institute of Research and Development for Optoelectronics, Măgurele, 077125, Romania

¹⁷Institution for Atmospheric and Earth System Research, University of Helsinki, Helsinki, 00014, Finland

¹⁸Institute of Atmospheric Sciences and Climate, National Research Council of Italy, Bologna, 40129, Italy

¹⁹Climate, Atmosphere Research Centre (CARE-C), The Cyprus Institute, Nicosia, 1645, Cyprus

²⁰Sonnblick Observatory, Central Institute for Meteorology and Geodynamics (ZAMG), Salzburg, 5020, Austria

²¹Global Change Research Institute, Brno, 60300, Czech Republic

²²Finnish Meteorological Institute, Helsinki, FI-00101, Finland

²³Andalusian Institute for Earth System Research, University of Granada, Granada, 18006, Spain

²⁴Departament of Applied Physics - Meteorology, University of Barcelona, Barcelona, 08028, Spain

²⁵Department of Air Protection, Institute of Environmental Engineering of the Polish Academy of Sciences, Zabrze, 41-819, Poland

Correspondence to: Andrea Cuesta-Mosquera (cuesta@tropos.de)

Key words: filter-based absorption photometer, aethalometer, intercomparison

40 **Abstract.** Aerosolized black carbon is monitored worldwide to quantify its impact on air quality and climate. Given its importance, measurements of black carbon mass concentrations must be conducted with instruments operating in a quality checked and assured conditions, to generate data which are reliable and comparable temporally and geographically. In this study, we report the results from the largest characterization and intercomparison of filter-based absorption photometers -aethalometers model AE33-, belonging to several European monitoring networks. Under controlled laboratory conditions, a
45 total of 23 instruments measured mass concentrations of black carbon from three well-characterized aerosol sources: synthetic soot, nigrosin particles and ambient air from the urban background of Leipzig, Germany. The objective was to investigate the individual performance of the instruments and their comparability; we analyzed the response of the instruments to the different aerosol sources, and the impact caused by the use of obsolete filter materials and the application of maintenance activities. Differences in the instrument-to-instrument variabilities from eBC concentrations reported at 880 nm were determined before
50 maintenance activities (average deviation from total least square regression: -2.0 %, range: -16 % to 7 %, for soot measurements; average deviation: 0.4 %, range: -15 % to 17 %, for nigrosin measurements), and after they were carried out (average deviation: -1.0 %, range: -14 % to 8 %, for soot measurements; average deviation: 0.5 %, range: -12 % to 15 %, for nigrosin measurements). The deviations are in most of the cases explained by the type of filter material employed by the instruments, the total particles load on the filter and the flow calibration.
55 The results of this intercomparison activity show that relatively small unit-to-unit variability of AE33-based particle light absorbing measurements is possible with well-maintained instruments. It is crucial to follow the guidelines for maintenance activities and the use of the proper filter tape in the AE33, to assure high quality and comparable BC measurements among international observational networks.

1. Introduction

60 The impact of black carbon (BC) on climate, health and human activities prioritizes the observation of BC mass concentration and its optical properties in different environments. In the atmosphere, BC absorbs solar radiation from the visible to the infrared optical spectrum, causing visibility degradation and making it the second most important radiative forcer (Ramanathan and Carmichael, 2008). Black carbon particles modify the lifetime, distribution, and formation processes of clouds, because they can act as cloud condensation nuclei, ice nuclei (predominantly in the cirrus temperature range), and modify clouds
65 internal mixing state, therefore alters clouds albedo (Koch et al., 2011; Bond et al., 2013; Chen et al., 2018; Wex et al., 2019). BC is also well-known as an air pollutant, affecting human health since it serves as a carrier of multiple toxic substances, which are harmful for the respiratory system, the cardiac function and the immune system (Janssen et al., 2011; WHO, 2012). In consequence, networks for observation of atmospheric black carbon are growing and need to be maintained worldwide. BC measurements provide base information to develop and track strategies aimed to reduce and manage air pollution and climate
70 change.

The understanding of the spatial and temporal variability of BC and its collateral effects, requires reliable, highly time-resolved, and long-term observations. To achieve this, three main aspects must be fulfilled during BC monitoring: (i) appropriate performance and quality check of the monitoring instruments, (ii) standardized use and maintenance by the operators, and (iii) reliable transmission and validation of data. The non-compliance of these requirements challenges the accuracy and comparability of BC observations.

Defined as the most refractory portion of particles produced in combustion processes, with a strong light absorption capacity (Petzold et al., 2013), diverse techniques are available to measure black carbon in the atmosphere. Depending on the measurement technique, BC may be addressed by different terminologies. When thermal methods are used, black carbon is measured as the non-volatized carbon remaining after applying specific high temperature to the sample, therefore BC is called elemental carbon (EC). In laser-induced incandescence techniques, the sample is heated to vaporization temperatures using an infrared laser and the thermal radiation emitted by incandescent black carbon is measured, then converted to mass concentration; here, we measure BC as refractory black carbon (rBC). When optical methods are used, the mass concentration of black carbon is indirectly retrieved from optical measurements of light attenuation caused by the aerosol particles – the determined quantity is equivalent to the mass concentration and therefore called equivalent black carbon (eBC; see Petzold et al., 2013). This method employs an external conversion factor known as the Mass Absorption Cross Section (MAC), to estimate the eBC mass concentrations. Other techniques used to measure black carbon include chemical oxidation and Raman spectroscopy (Bond et al., 2013; Lack et al., 2014).

In optical methods, the aerosol particle light absorption is measured either on the particles collected on a filter (filter-based absorption photometers), or measured directly in the aerosols suspended in a sample of air (photo-thermal spectrometers). In field monitoring, the filter-based absorption photometers (named in this document as FBAP) are widely used to perform long-term BC measurements, because these are robust, require relatively low human intervention, and no laboratory analysis are needed to process the sample. The aethalometer (Hansen et al., 1982), an instrument quantifying the transmission of light through a filter where the aerosol particles are collected, is one of the commonly used FBAP instruments. The difference between the light from an internal source transmitted through the sample-laden filter relative to the clean part of the filter is used to calculate the attenuation coefficient. The attenuation is transformed to absorption and later to eBC mass concentration using the black carbon mass absorption cross section and filter properties; these two last steps involve the use of fixed correction factors and a compensation algorithm. Further description about the functioning of the instrument is given in section 2.1.

According to EBAS database (Tørseth et al., 2012), in the last 10 years (2011-2020) a total of 57 European stations or sites have reported data from particle light absorption measurements using filter-based photometers, including aethalometers. These measurements contribute to networks and projects such as ACTRIS (41 sites) and EMEP (European Monitoring and Evaluation Programme, 5 sites) among others, and some stations may contribute to more than one network at the time. Data from stations using FBAP in 29 non-European countries are also available in EBAS. The COST Action CA16109 Chemical On-Line

cOmpoSition and Source Apportionment of fine aerosoL COLOSSAL, reports in its catalogue the cooperation with 49 sites
105 using FBAP in Europe.

Despite its wide use, the FBAP and particularly the aethalometers, feature inherent artifacts increasing the uncertainty in the measurements (Collaud Coen et al., 2010; Müller et al., 2011; Saturno et al., 2017). In first place, these instruments do quantify directly neither the absorption nor the eBC mass concentration, these are instead estimated from the measurements of light attenuation caused by the aerosol particles. The absorption coefficients and concentrations are calculated based on different
110 parametrizations and corrections for the absorption enhancement due to light scattering in the sample-laden filter matrix. In fact, the filter material used in the aethalometer and the particles collected in it, scatter a portion of the incident light reducing the transmission of it through the filter (Weingartner et al., 2003). Therefore, a reduction in the light transmitted may be taken as a higher absorption, with an additional small cross-sensitivity to scattering. A second artifact is caused by the loading effect produced by the aerosol particles accumulated in the filter matrix (Weingartner et al., 2003; Virkkula et al., 2007). After
115 particles are deposited, the detection of changes in the attenuation decreases, causing an underestimation of black carbon absorption and in consequence, lower eBC concentrations (Drinovec et al., 2015; Drinovec et al., 2017).

The characterization of aethalometers is therefore required to understand and reduce the variability and uncertainty in the measurements of BC, and this can be done by comparison experiments (EEA, 2013; WMO, 2016). The intercomparison consists in placing two or more instruments to measure the same sample under equal conditions and time. By intercomparing,
120 it is possible to study the instrument sensitivities to different aerosol sources and concentrations, the deviations caused by the type of filter material and numerical corrections used by the instruments, and the effects from different operational and maintenance procedures.

One of the first documented aethalometers intercomparison was performed by Ruoss et al. (1993), contrasting ambient air measurements performed by the DLR aethalometer (DLR Research Centre) and the Hansen-type aethalometer (Magee
125 Scientific); the authors found a significant up to 50 % variability among both instruments. In 1999, Hitztenberger et al. intercompared the absorption measurements carried out by an integrating sphere and one aethalometer (Hansen et al., 1984) using different filter materials; the aethalometer underestimated the absorption coefficients in the range of -26 % to -66 % when using quartz fiber filters; while using glass fiber and membrane filters, the aethalometer overestimated the absorption up to 34 %. Hitztenberger et al. (2006) compared a wide range of optical and thermal methods measuring ambient air aerosols.
130 The aethalometer AE9 reported higher eBC concentrations when compared against a multiangle absorption photometer (MAAP) (11.5 %) and a thermal-optical transmission (TOT) method (19 %). In the same study, when compared against an integrating sphere, the aethalometer showed lower concentrations (-5 %). Collaud Coen et al. (2010) compared the absorption coefficients reported by aethalometers (AE10, AE16, and AE31) and MAAP in four different locations; depending on the algorithms used to correct the loading effect in the aethalometers, these instruments overestimated the absorption coefficients
135 reported by the MAAP by 1 % to 33 %.

Although widely used, few experiments have been performed in order to characterize and compare the most recent generation of filter-based absorption photometers used in BC monitoring (Drinovec et al., 2015). Extensive intercomparisons are

fundamental to determine the variability among instruments from stations supporting international collaborative projects. They contribute to identify and quantify the factors influencing the instruments performance.

140 In this investigation, the authors present the results from the largest intercomparison of aethalometers model AE33, where 23 instruments were characterized and measured BC mass concentrations from three different aerosol sources. The main goal is to determine the unit-to-unit variabilities and their tendencies throughout the spectral range covered by the AE33. Also, we studied the influence of the maintenance activities and accessories used by the instruments on the reported eBC concentrations. In the end, we provide a series of simple recommendations for the instrument operation and maintenance.

145 2. Materials and Methods

The intercomparisons of aethalometers were conducted in three laboratory workshops at the World Calibration Centre for Aerosol Physics (WCCAP) in Leipzig, Germany. During the first workshop (14th to 25th January 2019), the characterization of seventeen AE33 part of the COST action CA16109 COLOSSAL and ACTRIS (Table 1) was performed. In this first experiment, the instruments were divided in four separated groups (A, B, C, D), due to space limitations in the laboratory
150 which did not permit to perform a simultaneous intercomparison. The instruments from each group completed 2.5 to 3 days of measurements. In the second workshop (7th to 12th June 2019), two aethalometers AE33, designated as group E, were intercompared. Finally, four aethalometers, comprising the group F, were intercompared during the third workshop (18th to 20th June 2019). Instruments in groups E and F does not form part of COLOSSAL, they belong to German research and regional monitoring organizations. The same WCCAP reference instrumentation setup was used in all three workshops.

155 2.1 The aethalometer AE33 and the compensation of eBC concentrations

The Aethalometer AE33 (Drinovec et al., 2015) uses a dual spot system to compensate the loading effect artifact. It calculates the absorption and the compensated eBC concentrations from measurements of light transmission at seven wavelengths from the near-UV to the near-IR (370, 470, 520, 590, 660, 880 and 950 nm).

The operation principle of the aethalometer consists in the continuous collection of aerosols on a filter, forming a sample-laden
160 spot. A light source illuminates the spot on the filter and, on the opposite side, a sensor measures the intensity of light transmitted through it (signal I). The light transmission is also measured through a sample-free area on the filter and is used as reference (signal I_0). By using both signals it is possible to calculate an attenuation (ATN; Eq. (1)).

$$\text{ATN}(\lambda) = -100 * \ln\left(\frac{I}{I_0}\right), \quad (1)$$

where the factor 100 is present for convenience only and ATN/100 should be used in further calculations. In the aethalometer,
165 the change in the attenuation with time is assumed to be caused by the increase of black carbon mass deposited on the filter. However, it was demonstrated that the correlation between the attenuation and the aerosol load on the filter is not linear at high attenuations (Gundel et al., 1984). Over time, the particles accumulated on the filter “shadow” each other, reducing the optical

path length, saturating the signal and therefore the measurement of light transmission. This condition is known as the loading effect and causes an underestimation of eBC concentrations. It is a cumulative property that needs to be calculated in real time to accurately report eBC mass concentrations (Drinovec et al., 2015).

To overcome the loading effect, the AE33 collects the aerosol sample in two spots (S_1 and S_2) on the filter. Each spot is collected with different airflow rate, and the spot with higher flow becomes more loaded with sample (Fig. 1). The instrument measures the light transmission through both spots and calculates two attenuations (ATN_1 and ATN_2 , using Eq. (1)), for the seven wavelengths of the instrument light source. The dual system allows to estimate a *compensation parameter* (k), based in the proportionality from the loading of both spots, to their airflows (F_1 from S_1 and F_2 from S_2), as shown in Eq. (2),

$$\frac{F_2}{F_1} = \frac{\ln(1-k*ATN_2)}{\ln(1-k*ATN_1)}, \quad (2)$$

where the compensation parameter k , representing the loading effect, will be equivalent for both spots as they are loaded with the same sample of aerosols. Equation (2) is used to calculate the instantaneous compensation for each wavelength $k(\lambda)$.




Figure 1. Schematic representation of the optical chamber in the Aethalometer AE33.

The calculation of the absorption coefficient is the intermediate step between the measured ATN and the calculated eBC mass. The absorption coefficient non-compensated for loading ($b_{abs}(\lambda)^{non\ comp.}$), is calculated as shown in Eq. (3),

$$b_{abs}(\lambda)^{non\ comp.} = \frac{s*(\Delta ATN_1(\lambda)/100)}{F_1*(1-\zeta)*C*\Delta t}, \quad (3)$$

where s is the spot area (constant, 0.785 cm^2), F_1 is the air flow through spot 1 (measured), ζ is the leakage factor (constant adjustable, depends on the filter material and the leakage test), $ATN_1(\lambda)$ is the attenuation calculated for spot 1, $\Delta ATN_1(\lambda)$ is the change in the attenuation at each wavelength in a given time step Δt , and C is the correction parameter for the multiple scattering enhancement. The particles and the filter may scatter a portion of light incident from the light source, increasing the optical path of the light in the filter and increasing the probability of light being absorbed, i.e., a light absorption enhancement. In the AE33 the user sets a constant value of C ; there are specific values available for each type of filter tape (Magee Scientific, 2018); nevertheless, multiple studies have shown this C factor depends also on the source of the aerosols measured (Collaud Coen et al., 2010; Ajtai et al., 2019), but this topic is out of the scope of this investigation.

The compensation parameter k is used to correct the loading effect and calculate the compensated absorption coefficient ($b_{abs}(\lambda)^{comp.}$) as follows:

$$b_{abs}(\lambda)^{comp.} = \frac{b_{abs}(\lambda)^{non\ comp.}}{(1-k(\lambda)*ATN_1(\lambda))}, \quad (4)$$

195 Finally, the compensated absorption coefficient and the BC mass absorption cross section ($\sigma_{air}(\lambda)$, external fixed constants in the instrument), are used to calculate the compensated eBC mass concentrations at seven wavelengths ($eBC(\lambda)^{comp.}$), as shown in Eq. (4):

$$eBC(\lambda)^{comp.} = \frac{b_{abs}(\lambda)^{comp.}}{\sigma_{air}(\lambda)}, \quad (5)$$

2.2 Intercomparison procedure

200 The workshops were performed in three sessions (Fig. 2):

1. Initial comparison: During the first day the instruments were connected to the mixing chamber and started the measurements of urban background aerosols from Leipzig conserving the internal configuration and accessories provided by the operators. The instrument time was synchronized, and only the measurement time resolution (1 min), the flow reporting standard (AMCA: 21.1 °C, 1013 hPa) and the maximum attenuation limit (ATN=120), were modified. After approximately 1 hour of ambient air measurements, synthetic soot produced with a miniCAST 5203 (Table 2) was supplied to the mixing chamber, followed by nigrosin particles (see section 2.4), particle-free air and ambient air. The initial comparison was performed with the aim to: (i) allow the adjustment of the internal compensation parameters $k(\lambda)$ to the local conditions and, (ii) determine the initial variability and deviation of the aethalometers before the maintenance and calibration procedures.

210




Figure 2. The intercomparison procedure.

2. Maintenance and calibration: The maintenance included a series of procedures performed by following the instructions given in the AE33 user manual - version 1.57 (Magee Scientific, 2018):

- Flow verification test, using an externally calibrated flowmeter (Table 2).
- Cleaning of the optical chamber.
- Flow calibration, performed only in those instruments with non-acceptable results from the flow verification test (deviations >10 %).
- Leakage test.
- Replacement of the filter tape, performed for instruments using a different filter tape to the one recommended currently (M8060).

3. Final comparison: The aethalometers measured the same three aerosol sources used in the initial comparison (synthetic soot, nigrosin particles, ambient), and particle-free air. The goal was to determine the new instrument-to-instrument variabilities after maintenance.

2.3 Experimental set-up

A mixing chamber (0.5 m³) with an internal fan was used to distribute well-mixed samples of aerosols to the aethalometers (Fig. 3). The instruments intercompared measured eBC mass concentrations from ambient air, synthetic soot and nigrosin particles.

The aethalometer AE33 used as reference belongs to the WCCAP; it receives frequent maintenance, and is operated with the correct accessories (filter tape M8060). The flow of this AE33 is calibrated with an externally calibrated flowmeter model 4140 F, TSI Inc. The reported absorption coefficients of this reference AE33, have been compared with the absorption calculated by a reference set up from the WCCAP, consisting in one nephelometer Aurora 3000, EcoTech, measuring the aerosols light scattering coefficients, and one CAPS PMex Monitor, Aerodyne Research, Inc, which measures the aerosols optical extinction. The absorption from the reference set up is calculated as *absorption = extinction – scattering*, at 450, 525 and 635 nm; the measurements at 450 nm and 635 nm are extrapolated to 470 and 660 nm, respectively. The results from this comparison are shown in Figure S1 in the supplementary material.

A Mobility Particle Size Spectrometer was used to quantify the particle number size distribution of the different aerosol samples used during the workshops. Table 2 presents a list of the auxiliary instruments used during the intercomparisons.




Figure 3. Experimental set-up used in the intercomparisons.

Table 1. List of aethalometers AE33 intercompared.

| Group | N° | ID | Serial | Intercomparison | | Operating conditions |
|-------|----|-----|-----------|-----------------|------------|---|
| | | | | Date start | Date end | |
| A | 1 | A01 | S02-00246 | 14/01/2019 | 16/01/2019 | Measurement time resolution: 1 min ATN max: 120 Filter tape: see table 3 Inlet flow: 5 L min ⁻¹ Flow reporting conditions: AMCA, 21 °C, 1013 hPa |
| | 2 | A02 | S07-00618 | | | |
| B | 3 | B01 | S02-00170 | 16/01/2019 | 18/01/2019 | |
| | 4 | B02 | S01-00080 | | | |
| | 5 | B03 | S07-00767 | | | |
| | 6 | B04 | S04-00387 | | | |
| | 7 | B05 | S0200267 | | | |
| | 8 | B06 | S02-00204 | | | |
| C | 9 | C01 | S01-00113 | 21/01/2019 | 23/01/2019 | |
| | 10 | C02 | S01-00114 | | | |
| | 11 | C03 | S06-00560 | | | |
| | 12 | C04 | S07-00729 | | | |
| D | 13 | D01 | S07-00669 | 23/01/2019 | 26/01/2018 | |
| | 14 | D02 | S00-00049 | | | |
| | 15 | D03 | S02-00258 | | | |
| | 16 | D04 | S00-00055 | | | |
| | 17 | D05 | S02-00156 | | | |
| E | 18 | E01 | S02-00202 | 07/06/2019 | 12/06/2019 | |
| | 19 | E02 | S07-00737 | | | |
| F | 20 | F01 | S07-00705 | 19/06/2019 | 21/06/2019 | |
| | 21 | F02 | S07-00706 | | | |
| | 22 | F03 | S05-00443 | | | |
| | 23 | F04 | S06-00578 | | | |

Table 2. Instruments employed during the workshops.

| Instrument | Measurement | Operating conditions |
|--|--|---|
| <i>Intercomparison</i> | | |
| Aethalometer used as reference: model AE33, Magee Scientific (TROPOS), S/N: S02-00163 | eBC concentration at 7 wavelengths: 370, 470, 520, 590, 660, 880 & 950 nm | Measurement time resolution: 1 min ATN max: 120 Filter tape: M8060 Inlet flow: 5 L min ⁻¹ Flow reporting conditions: AMCA*, 21.1 °C, 1013 hPa |
| MPSS (Mobility particle size spectrometer): CPC model 3010 from TSI Inc. & DMA from TROPOS (WCCA reference) | Particle number size distribution. Aerodynamic diameter range: 10 – 800 nm | Measurement time resolution: 5 min Inlet flow: 1 L min ⁻¹ Flow reporting conditions: Standard, 0 °C, 1013.25 hPa |
| Soot Generator miniCAST: model: 5203 Type C, Jing Ltd. | | Diffusion flame conditions: <ul style="list-style-type: none">• Propane: 105 mL min⁻¹;• Oxidation air: 3.6 L min⁻¹;• Dilution air: 20 L min⁻¹;• Quench gas N₂: 20 L min⁻¹ |
| Customized particle nebulizer: built using a Constant Output Atomizer model 3076, TSI Inc. | | Nigrosin: <ul style="list-style-type: none">• CAS: 8005-03-6• Molecular weight: 202.2 g mol⁻¹• Concentration of the solution: 0.5-0.8 g L⁻¹ |
| <i>Maintenance</i> | | |
| Mass flow meter: model 4140 F, TSI Inc. | | Measurement time resolution: 1 sec Operative range: 0.01 – 20 L min ⁻¹ Flow reporting conditions: AMCA*, 21.1 °C, 1013 hPa |

*AMCA (Air Movement and Control Association International) are the default standard conditions used by the flow sensors in the AE33, to report the measured mass flow (Magee Scientific, 2018).

2.4 Aerosol sources

- 250 The aethalometers measured eBC concentrations from three aerosol sources:
1. Synthetic soot particles produced with a miniCAST (Jing Ltd, 2013), using a fuel-lean mixture (fuel-to-air equivalence ratio, $\phi < 1$). Table 2 shows the operating conditions used in the miniCAST during the intercomparisons.

2. Black particles created by the nebulization of a Nigrosin solution (Table 2).
3. Ambient air aerosols from the urban background in the city of Leipzig, Germany. Concentrations correspond to early-morning periods (3:00 to 9:00 am), during winter (workshop 1) and summer time (workshop 2 and 3).

Figure 4 presents the average particle number size distributions of the aerosol sources measured during the workshops.




Figure 4. Particle number size distributions of the aerosol sources used in the intercomparison.

2.5 Data processing and analysis

Equivalent black carbon concentrations were measured every minute. Subsequent to the workshops, the data from the aethalometers were cleaned based on the instrument status codes. The AE33 reports a series of status codes representing the operational state, internal procedures in course, warning alerts or errors in the instrument, e.g., 0: normal measurements, 1: filter tape advance, 8: Check flow status history, 384: tape error (tape not moving, end of tape) (Magee Scientific, 2018). As more than one condition may occur at the same time, these statuses are built under a binary system therefore they may be numerically combined, e.g., 9: Check flow status history (warning) + Tape advance (procedure), 387: Tape error (tape not moving, end of tape) + Stopped. Therefore, data from the few valid statuses available were kept and then used in the data analysis (see table S1 in the supplementary material containing the list of valid statuses used in this study). Invalid data was removed for data analysis.

To compare the measurements made by the aethalometers and the reference AE33, we used Deming total least squares regression analysis (R package “Deming” (Therneau, 2018)), to account for the independent observational errors from each instrument and the reference. Deming regression finds the best fitting line by minimizing the sum of the distances in both x and y directions, simultaneously (Cornbleet and Gochman, 1979). The processes of data cleaning and analysis were performed in the software R, version 4.0.0.

Section 3.1 presents a detailed analysis of the instruments characterized in the group D, as a case of study illustrating the wide range of deviations observed in real monitoring networks. In this group we have intercompared aethalometers using three different filter materials, and one of them presented unacceptable results from the flow verification test before maintenance. A summary and analysis of the results obtained for the total 23 units intercompared is given in section 3.2; additionally, the Figures S2 to S11 in the supplementary material, present the time series of the measurements performed by the instruments in the groups A, B, C, E and F, before and after maintenance.

3.1 Unit-to-unit variability for study case group D

The experiments for instruments in group D were divided in three sections: an initial comparison before maintenance, an intermediate comparison after a partial maintenance, and a final comparison after filter tape change.

Initial comparison

Figure 5 presents the time series of the one-minute eBC mass concentrations measured in the initial comparison by the aethalometers within group D. The gray areas represent the periods when the different aerosol sources and particle free air were supplied in to the mixing chamber; the red bars indicate times when tape advances occurred in more than one instrument at the same time.

The variability in the measurements of eBC observed in group D, and in general in the six groups, were significantly higher few minutes before and after a tape advance (TA). Continuous supply of CAST soot with concentrations above 15 to 20 $\mu\text{g}\cdot\text{m}^{-3}$ led to the instrument reaching the maximum attenuation limit ($\text{ATN}_{\text{TA}}=120$) after ~ 30 min, inducing a TA in the majority of the instruments. The differences in the reported eBC mass concentrations close the TAs reached up to 25% among the AE33. During nigrosin measurements, constant concentrations of 10 $\mu\text{g}\cdot\text{m}^{-3}$ led to the maximum attenuation limit after ~ 120 min. Offsets in the concentrations measured by the aethalometers were up to 25% during the nigrosin supply.

Before the maintenance and calibration, the highest deviations in group D were seen in three aethalometers (Fig. 5): D03, D04 and D05. The instruments D04 and D05 underestimated the eBC concentrations by 11% and 20% with respect to the reference, respectively. These two instruments used an older version of filter tape M8050 (also known as TX40), a glass fiber filter on a woven backing. The M8050 filter tape was distributed during a short period from 2016 to 2017, and according to the manufacturer it was substituted because of evidence of unsatisfactory performance. On the other hand, the instrument D03 reported slightly higher concentrations than our reference aethalometer, overestimating the eBC concentrations by up to 6 and 8%, while measuring soot and nigrosin particles, respectively. The aethalometer D03 used the T60A20 filter tape (also known as M8020 or AE33-FT), made from TFE-coated glass fibers; this was the first filter used in the AE33 (Drinovec et al., 2015), and was available from 2014 to 2016.

The different materials used, and the structure of the filters, give them specific light scattering properties responsible of light absorption enhancement (Petzold et al., 1997). Therefore, the correction factor C accounting for the light scattering effects of the filter and the particles collected in it, take different values for each type of filter tape. It has been demonstrated that the apparent C correction factor is in addition susceptible to the type of aerosols measured (Collaud Coen et al., 2010). However, the estimation of a source-dependent C factor was not in the scope of this study; our interest is limited to the correct use of the C factors associated to each filter tape. Standard C factor for the filters previously available, M8050 and T60A20, is 1.57, relative to the value 2.14 determined in Weingartner et al. (2003). To guarantee the comparability within monitoring stations, these two filter tapes are no longer recommended to be used in the AE33 (Magee Scientific, 2018). The new filter M8060 must be used instead, and its corresponding multiple scattering parameter C has to be set as the internal parameter of the instrument ($C_{M8060} = 1.39$).




Figure 5. Time series of eBC mass concentrations at 880 nm before maintenance in group D. Gray panes show the periods of clean air and aerosols supply, red bars indicate times when tape advances occurred in more than one instrument simultaneously.


320 General maintenance and intermediate comparison

In group D the maintenance was divided in two phases to observe separately the influence of the essential servicing activities, i.e., cleaning of the optical chamber, flow verification and calibration, and leakage test, all performed in the first phase, and the replacement of the older filter tapes, which comprises the second phase. An intermediate intercomparison was performed in the middle of both maintenance phases.

325 During the first phase of maintenance, the instrument D03 showed unsatisfactory results from the flow verification and leakage tests. In average, the flow sensors detected a 30% less airflow than the reference flowmeter (see Table 3), requiring a flow

calibration. In the aethalometer AE33, if the flow verification test indicates a deviation of $\pm 10\%$ in any of the three flow rates (flow through spot 1, spot 2 and common flow), a flow calibration must be carried out (Magee Scientific, 2018). From the leakage test, this instrument reported a leakage of 9%. This result indicates that almost a ten percent of the inlet air flow in the aethalometer (F_{in}), is being tangentially lost across the edges of the filter tape. The results from the flow verification and leakage tests for the other instruments in group D were satisfactory (Table 3).

After the first maintenance phase, the instrument D03 went from overestimating eBC to underestimate the mass concentration by 33% and 18% during soot and nigrosin measurements, respectively (Fig. 6a). During the same period, the instruments D04 and D05 continued underestimating the eBC concentrations while measuring soot (25%) and nigrosin (27%).






Figure 6. Time series of (a) eBC mass concentrations at 880 nm and (b) compensation parameters k_{880} , after the initial phase of maintenance in group D. Gray panes show the periods of clean air and aerosols supply, red bars indicate times when tape advances occurred in more than one instrument simultaneously.

The shape of the sample spots also demonstrates problems in an aethalometer. This was evident in the instrument D03 whose sample spots presented an irregular shape with heterogeneous saturations/streaks (Fig. 7a); under optimal operating conditions the spots formed on the filter tape have a circular and well-defined shape which assures a correct measurement of the attenuation change (Fig. 7b). However, it was shown that the streaks (Fig. 7a) do not impact the measurement (John Ogren, personal communication).




Figure 7. Shapes of sample spots observed during maintenance in group D. (a) instrument D03, (b) instrument D01. These instruments used different filter tapes (D03: T60A20, D01: M8060). The spot size was not measured during the workshops. Irregular or diffuse edges of the filter spot can indicate leakage.

The real time calculation of the eBC mass concentrations in the AE33 is based on the compensation parameters k . Figure 6b illustrates the time series of the k_{880} values during the intermediate comparison (after flow and leakage adjustments). The

comparison of the compensation parameters is significant as these respond to the changes of aerosol sources and concentrations in the experiments, independently of instrument correction factors. Once the flow verification and calibration were implemented, it would be expected to have similar k values among the instruments. Nevertheless, as evidenced in Figure 6b, the differences in the k from the reference aethalometer and the instruments in group D ranged from -40 % until > +60 %. The main reasons explaining these deviations are the time at which the TAs occurred in the instruments, and the filter materials they used. Although the instantaneous k values are calculated using Eq. (2), the instrument uses specific k depending on a weighting method based on the values of the attenuation as follows (supplementary material, Drinovec et al., 2015):

360

- **ATN₁ < 30:** Under low BC concentrations as well as after one tape advance, a low attenuation change will take place in the aethalometers. While ATN₁ < 30 (attenuation from in spot 1, channel 1), the last k values from the previous spot (k_{old}), will be used to compensate the eBC mass concentrations. Accordingly,

$$k = k_{old}, \quad (6)$$

365

- **30 < ATN₁ < 120:** For this attenuation range, the k is calculated using the k_{old} , and the instantaneous k calculated from Eq. (2):

$$k = k_{weighted} = \frac{(ATN_{TA} - ATN_1) * k_{old} + (ATN_1 - ATN_{f2}) * k_{inst}}{(ATN_{TA} - ATN_{f2})}, \quad (7)$$

where ATN_{TA} is the maximum limit of attenuation triggering a tape advance and ATN_{f2} is the upper limit of attenuation for the fitting range, ATN_{f2} = 30.

370

- **ATN₁ = 120 = ATN_{TA}:** Once the spot 1 is completely loaded and the threshold attenuation is reached, the k becomes equals to the instantaneous k calculated with Eq. (2).

The k values also depend on the filter type as the different materials determine the filter loading rate and thus the time when the threshold attenuation (ATN_{TA}) is attained. In addition, the k values are susceptible to the type of aerosols measured (composition and size) and their mixing state (Drinovec et al., 2017). If the equation 4 is rearranged and expressed in terms of eBC, the k values could be defined as a function of the non-compensated and compensated black carbon mass concentration:

375

$$k = \frac{1}{ATN} \left(\frac{eBC^{comp.} - eBC^{non\ comp.}}{eBC^{comp.}} \right), \quad (8)$$


According to Eq. 8, for a given attenuation, if the compensated eBC is larger than the uncompensated, k will be positive and inversely proportional to the eBC mass. This was observed for the instruments D04 and D05, with higher and positive deviations from our reference aethalometer (Fig. 6b). Through this last period, the k from both instruments decreased constantly, meaning this couple of aethalometers detected a different attenuation range (between 30 and 120), instead of a low attenuation associated to low ambient air concentrations taking place. Along the intermediate comparison, the compensation parameters exhibited a significant variability in group D, ranging from 0.0045 to 0.0115. When performing an intercomparison, the change of the sample requires a “conditioning run” of a sample spot up to ATN_{TA} to obtain the source specific value of the parameter k . Alternatively, all data needs to be manually reprocessed.

385

Final maintenance and comparison

In the final step of maintenance, the filter tape was replaced in three instruments from the group D: D03, D04 and D05, which were using older versions of filter tapes which are no longer recommended for the AE33. The final comparison was performed during two days. As observed in Figure 8a, the deviations among the eBC mass concentrations reported by the instruments reduced significantly for all aerosol sources ($<10\%$), in comparison with the initial and intermediate comparisons performed in group D. Significant deviations ($>10\%$) were observed only at the beginning of this final stage, after the first tape advance while the instruments measured BC mass from the soot source; these higher deviations are associated with the initial adjustment required by the compensation algorithm to a new aerosol source and the effect of a filter tape advance, as mentioned earlier.

The internal algorithm uses the previous spot compensation parameter and, if this parameter was determined from measurements of a completely different sample, the value does not apply for measurements of the new sample (Drinovec et al., 2015). Figure 8b shows the time series of the compensation parameters k at 880 nm corresponding to the final comparison. The new k ranged now from 0.005 to 0.008, and their deviations from the reference aethalometer ranged from -20% to -10% during soot and nigrosin. However, substantial differences took place during ambient air measurements mainly in the aethalometer D05; once the supply in the mixing chamber changed from clean air to ambient air at midnight, the compensation parameters from D05 jumped abruptly between > 0.012 to negative values within 1 hour. This situation is associated with the response of the instrument to changes in pressure: the airflow pressure in the mixing chamber affects the flow rates, which is directly related with the instantaneous k . During maintenance, the aethalometer D05 presented unusual behavior associated with the flow ratio appearing in a certain extent even after flow calibration. As recommended by the manufacturer, the ratio of F_1 (higher sample airflow through spot 1) to F_2 (lower sample airflow through spot 2), should be about 1.75 to 2.5, however this relation was not always accomplished in this instrument. As result, it was recommended to the operator to send the instrument to the manufacturer for advanced maintenance.



410 **Figure 8. Time series of (a) eBC mass concentrations at 880 nm and (b) compensation parameters k_{880} , after the final phase of maintenance in group D. Gray panes show the periods of clean air and aerosols supply, red bars indicate times when tape advances occurred in more than one instrument simultaneously.**

Statistical summary


The numerical unit-to-unit variabilities were calculated using data from measurements at 880 nm before and after maintenance, via total least squares regression with intercepts forced through zero. In the case of group D, maintenance includes flow, leakages and filter tape adjustments.

Figures 9, 10 and 11, present the scatterplots from the comparison of the five instruments in group D against our reference aethalometer, while measuring soot, nigrosin and ambient air, respectively. For soot and nigrosin, the correlations were built using measurements after the majority of the instruments performed one or two tape advances, to avoid the bias caused by the effects previously described in this section. In the case of ambient air, the measurements were averaged to 5 minutes as the ambient concentrations of BC were low ($< 1 \mu\text{g m}^{-3}$) after maintenance.

The results from the regressions demonstrate acceptable agreement between the aethalometers in group D, in general improved after maintenance activities. In soot measurements, the total average slope ranged from 0.94 (6 % deviation) before service, to 0.95 (5 % deviation) after service. According to Figure 9, the aethalometers D01, D02 and D03 performed better before maintenance. The deviations of these three instruments (how much the slopes differ from 1, bias), are certainly higher after service, even if these deviations are all relatively low ($< 10\%$). As mentioned earlier, the scatter plots and the Deming regressions were calculated using measurements performed after one or two tape advances, to avoid the effects produced by the filter advance and the initial adjustment of the compensation algorithm. To evaluate the agreements among the instruments, it is fundamental to observe as well the complete time series (Figures 5, 6 and 8), which give a broader perspective of the performance of the instruments; from this observation it is evident that offsets and deviations are significantly higher before maintenance, more remarkable in the case of the instruments D03, D04 and D05.

In the case of nigrosin, the total average slope was 0.99 (1 % deviation) before service, to 0.94 (6 % deviation) after service. From the ambient air observations, low variabilities were estimated in group D however, it is inaccurate to state an improvement or not regarding the measurements before and after maintenance, as these concentrations were very low and rather stable, making the interpretation of the regression imprecise. In ambient air measurements, the instrumental noise (described in section 3.3), becomes critical in sites with low BC concentrations. Before maintenance in group D, the one-minute average eBC concentration measured at 880 nm by the reference aethalometer in the early morning time, was $1.82 \mu\text{g m}^{-3}$, and the average noise calculated in this group was $0.038 \mu\text{g m}^{-3}$, representing 2% of the reported eBC concentration. After maintenance, the average eBC was $0.65 \mu\text{g m}^{-3}$ (880 nm), and the average noise (see section 3.3 below) was $0.037 \mu\text{g m}^{-3}$, representing 6 % of the average eBC.

Table 4 presents a summary for the regression analysis in group D and the other 18 aethalometers intercompared showing the relative slope (AE33 vs. AE33 REF) and the determination coefficient R^2 before and after maintenance.



445 **Figure 9.** Regression for the comparison of the instruments in group D and the reference aethalometer, before and after maintenance during soot measurements. The intercept was forced through zero. BM: Before maintenance, AM: After maintenance.




Figure 10. Regression for the comparison of the instruments in group D and the reference aethalometer, before and after maintenance during nigrosin measurements. The intercept was forced through zero. BM: Before maintenance. AM: After maintenance.




Figure 11. Regression for the comparison of the instruments in group D and the reference aethalometer, before and after maintenance during ambient air measurements. The intercept was forced through zero. BM: Before maintenance. AM: After maintenance.

The Figures S12, S13 and S14 in the supplementary material, present the scatter plots from the intercomparison among the instruments D01 to D05, for soot, nigrosin and ambient air, respectively.

3.2 Total unit-to-unit variability of the 23 aethalometers

In general, the aethalometers in groups A to F showed acceptable agreement when compared against our reference aethalometer, and this improved in most of the cases after the maintenance activities. Out of the 23 aethalometers intercompared, five instruments from groups B and D exhibited the highest unit-to-unit deviation. Regarding the eBC concentrations (880 nm), the total average deviation from the 23 instruments was -2.0 % before maintenance and -1.0 % after maintenance, for soot measurements. In nigrosin, the total average deviation changed from 0.4 % to 0.5 % before and after maintenance, respectively. A fair comparison for ambient air measurements is not possible to calculate, considering the

significantly low and stable concentrations measured during some days in the urban background in Leipzig, and the fluctuations of concentrations in the workshops. The deviations calculated for the 23 aethalometers at 880 nm are summarized in Table 4.

465 Flow verification and leakages tests results were acceptable for most of the 23 instruments and are listed in Table 2; both results from the flow verification before and after maintenance are shown only for those instruments whose initial results were unsatisfactory and whose tests had to be repeated after maintenance. The flow calibration procedure was applied to five instruments; these presented initially higher deviation in the lower flow tested (1 L min^{-1} ; 16 % in average), followed by the high flow (3 L min^{-1} , 9 % in average) and the total flow (5 L min^{-1} ; 7 % in average). After maintenance the new results from

470 the flow verification tests showed average deviations of 1.8 %, 0.9 % and 0.1 %, for the airflows of 1, 3 and 5 L min^{-1} , respectively.

From the six groups, only three required a change of filter tape: D03, D04 and D05 (Table 2).

Wavelength-dependency of the unit-to-unit variability

475 The change in the unit-to-unit variabilities according to wavelength was analyzed, by calculating total least squares regressions of BC concentrations measured at the seven channels of the twenty-three aethalometers intercompared against the reference AE33. The influence of the maintenance in these variabilities was also investigated. Figure 12 shows the boxplots representing the range of the deviations calculated for the 23 instruments; the figure includes the average values of the slope, bias, RMSE and the coefficient of determination R^2 , before and after maintenance. Slope equaling to 1 (red line) indicates the instruments

480 are 1:1 in respect with the reference (0 % of deviation).

As seen in Fig. 12, no significant changes in the variabilities throughout the spectral range was seen during soot measurements, but slightly negative deviations were observed at 660 nm. The total range of variabilities also had a slight reduction after maintenance and the median slope values were closer to 1 during this period. A reduction in the values of RMSE was remarked at wavelengths 660, 880 and 950 nm. For nigrosin measurements, no significant changes were observed in the slopes with

485 wavelength, but the spread of the values of slope increased after maintenance. A significant reduction in the values of RMSE was observed at wavelengths 880 and 950 nm during nigrosin measurements. The variabilities from measurements of eBC from ambient air, showed a slightly reduced range after maintenance. No clear trend was observed with wavelengths, but negative deviations were seen at 590, 660 and 880 nm, for ambient air data.




Figure 12. Statistics from the unit-to-unit variabilities of the 23 units intercompared with the reference AE33, as function of wavelength for soot, nigrosin and ambient air: (a) slope, (b) R^2 , (c) bias, (d) RMSE. The black horizontal line inside the boxes represents the median value of the slope, R^2 , bias and RMSE (statistical variables), respectively; the lower and upper borders of the boxes are the first and third quartiles, on which the middle 50% of the statistical variables are located; the whiskers correspond to ranges for the bottom 25% and the top 25% without outliers (black points). The red line represents the slope and R^2 equaling to 1 (0 % of deviation) in the subplots (a) & (b), respectively, and the bias and RMSE equaling to 0 in the subplots (b) & (c), respectively. BM: Before maintenance. AM: After maintenance.

Table 3. Summary of results from maintenance tests.

| Group | Instrument | Flow verification test ¹ (%) | | | | | | | Flow calibration | Leakage test (%) | | Initial type of filter tape | Change of filter tape |
|-------|------------|---|------------------------|-----|-----|------------------------|-----|-----|------------------|------------------|-----|-----------------------------|-----------------------|
| | | BM or AM | F1 | | | FC | | | | | | | |
| | | | 1 | 3 | 5 | 1 | 3 | 5 | | | | | |
| | | | (L min ⁻¹) | | | (L min ⁻¹) | | | | BM | AM | | |
| A | A01 | BM | 81 | 98 | 99 | 84 | 99 | 99 | ✓ | 2 | - | M8060 | - |
| | | AM | 104 | 101 | 101 | 102 | 101 | 101 | | | | | |
| | A02 | BM | 101 | 101 | 101 | 99 | 101 | 101 | - | 1.4 | - | M8060 | - |
| B | B01 | BM | 84 | 90 | 94 | 80 | 89 | 93 | ✓ | 1.9 | - | M8060 | - |
| | | AM | 99 | 98 | 99 | 95 | 97 | 98 | | | | | |
| | B02 | BM | 98 | 100 | 101 | 95 | 100 | 100 | - | 1.8 | - | M8060 | - |
| | B03 | BM | 99 | 101 | 102 | 97 | 101 | 102 | - | 2.4 | - | M8060 | - |
| | B04 | BM | 94 | 99 | 100 | 98 | 100 | 101 | - | 1.6 | - | M8060 | - |
| | | AM | 93 | 98 | 100 | 93 | 98 | 100 | | | | | |
| | B05 | BM | 82 | 93 | 95 | 80 | 92 | 95 | - | 1.9 | - | M8060 | - |
| | | AM | 95 | 99 | 101 | 91 | 98 | 100 | | | | | |
| | B06 | BM | 88 | 89 | 89 | 81 | 87 | 89 | ✓ | 6.5 | ND | T60A20 ² | - |
| | | AM | 101 | 100 | 100 | 101 | 100 | 100 | | | | | |
| C | C01 | BM | 99 | 102 | 103 | 101 | 103 | 103 | - | 1.9 | - | M8060 | - |
| | C02 | BM | 97 | 98 | 99 | 95 | 98 | 100 | - | 1.7 | - | M8060 | - |
| | C03 | BM | 101 | 100 | 100 | 97 | 99 | 100 | - | 1.8 | - | M8060 | - |
| | C04 | BM | 93 | 100 | 101 | 90 | 100 | 101 | ✓ | ND | 1.8 | M8060 | - |
| | | AM | 98 | 99 | 100 | 99 | 99 | 100 | | | | | |
| D | D01 | BM | 103 | 101 | 101 | 100 | 101 | 102 | - | 1.9 | - | M8060 | - |
| | D02 | BM | 98 | 99 | 99 | 101 | 99 | 99 | - | 8.4 | - | M8060 | - |
| | D03 | BM | 68 | 70 | 72 | 69 | 70 | 72 | ✓ | 9.2 | 1.5 | T60A20 | ✓ |
| | | AM | 102 | 100 | 100 | 102 | 100 | 100 | | | | | |
| | D04 | BM | 103 | 101 | 102 | 102 | 100 | 101 | - | 0.9 | - | M8050 | ✓ |
| | D05 | BM | 101 | 100 | 101 | 98 | 100 | 101 | - | 0.7 | - | M8050 | ✓ |
| E | E01 | BM | 98 | 99 | 100 | 99 | 99 | 100 | - | 1.8 | - | M8060 | - |
| | E02 | BM | 100 | 100 | 100 | 101 | 100 | 100 | - | 2.7 | - | M8060 | - |
| F | F01 | BM | 97 | 100 | 100 | 97 | 99 | 100 | - | 1.8 | - | M8060 | - |
| | F02 | BM | 96 | 101 | 101 | 96 | 101 | 101 | - | 1.7 | - | M8060 | - |
| | F03 | BM | 97 | 101 | 101 | 97 | 102 | 101 | - | 2 | - | M8060 | - |
| | F04 | BM | 101 | 101 | 101 | 99 | 101 | 101 | - | 2 | - | M8060 | - |

¹Flow reporting standard: AMCA (21 °C, 1013 hPa). ²This filter tape was not replaced because of operational reasons.

500 BM: Before maintenance. AM: After maintenance. ND: No data.

Table 4. Relative slope and correlation coefficients for total least squares regression forced through the origin, for eBC mass concentrations (880 nm), before and after maintenance.

| Instrument | Aerosol source | Slope relative to AE33 REF | | Adjusted R^2 | |
|------------|----------------|----------------------------|------|----------------|-------|
| | | BM | AM | BM | AM |
| Group A | | | | | |
| A01 | Soot | 1.03 | 0.95 | 0.999 | 1 |
| | Nigrosin | 0.93 | 1.03 | 1 | 0.999 |
| | Ambient Air | 0.86 | 0.72 | 0.973 | 0.976 |
| A02 | Soot | 1.01 | 0.99 | 0.995 | 1 |
| | Nigrosin | 1.05 | 1.15 | 1 | 0.999 |
| | Ambient Air | 1.12 | 0.94 | 0.988 | 0.979 |
| Group B | | | | | |
| B01 | Soot | 0.91 | 0.97 | 0.999 | 1 |
| | Nigrosin | 0.99 | 1.01 | 1 | 1 |
| | Ambient Air | 0.95 | ND | 0.993 | ND |
| B02 | Soot | 1.05 | 1.04 | 1 | 1 |
| | Nigrosin | 1.03 | 1.05 | 1 | 1 |
| | Ambient Air | 0.97 | 0.90 | 0.992 | 0.94 |
| B03 | Soot | 1.06 | 1.06 | 1 | 0.998 |
| | Nigrosin | 0.99 | 0.97 | 0.999 | 1 |
| | Ambient Air | 0.98 | 1.11 | 0.986 | 0.896 |
| B04 | Soot | 1.00 | 1.02 | 1 | 0.999 |
| | Nigrosin | 0.92 | 0.92 | 0.999 | 1 |
| | Ambient Air | 0.86 | 0.81 | 0.993 | 0.935 |
| B05 | Soot | 1.07 | 1.03 | 1 | 1 |
| | Nigrosin | 1.07 | 1.03 | 0.999 | 1 |
| | Ambient Air | 0.66 | 1.13 | 0.967 | 0.948 |
| B06 | Soot | 1.02 | 0.86 | 1 | 0.996 |
| | Nigrosin | 1.14 | 1.09 | 1 | 0.998 |
| | Ambient Air | 0.81 | 0.75 | 0.987 | 0.933 |
| Group C | | | | | |
| C01 | Soot | 0.94 | 0.93 | 1 | 1 |
| | Nigrosin | 0.95 | 0.98 | 0.999 | 0.999 |
| | Ambient Air | 0.98 | 0.99 | 0.997 | 1 |
| C02 | Soot | 0.96 | 0.98 | 0.998 | 1 |
| | Nigrosin | 0.96 | 1.04 | 0.999 | 0.999 |
| | Ambient Air | 1.04 | 1.03 | 0.998 | 1 |
| C03 | Soot | 1.04 | 1.06 | 1 | 1 |
| | Nigrosin | 1.07 | 1.06 | 0.999 | 1 |

| Instrument | Aerosol source | Slope relative to AE33 REF | | Adjusted R^2 | |
|----------------|----------------|----------------------------|------|----------------|-------|
| | | BM | AM | BM | AM |
| | Ambient Air | 1.07 | 1.06 | 0.999 | 0.999 |
| C04 | Soot | 0.92 | 0.99 | 0.991 | 0.997 |
| | Nigrosin | 1.03 | 1.09 | 0.998 | 1 |
| | Ambient Air | 1.07 | 1.02 | 1 | 0.999 |
| Group D | | | | | |
| D01 | Soot | 0.96 | 0.94 | 0.997 | 1 |
| | Nigrosin | 1.05 | 0.94 | 1 | 1 |
| | Ambient Air | 0.95 | 0.99 | 1 | 0.999 |
| D02 | Soot | 1 | 1.06 | 0.999 | 0.999 |
| | Nigrosin | 0.99 | 1.03 | 1 | 1 |
| | Ambient Air | ND | 1.04 | ND | 0.999 |
| D03 | Soot | 1.03 | 0.93 | 1 | 1 |
| | Nigrosin | 1.17 | 0.88 | 0.999 | 1 |
| | Ambient Air | 1.17 | 0.95 | 1 | 0.999 |
| D04 | Soot | 0.84 | 0.93 | 1 | 0.998 |
| | Nigrosin | 0.89 | 0.91 | 0.997 | 0.999 |
| | Ambient Air | 0.87 | 1.05 | 1 | 0.994 |
| D05 | Soot | 0.89 | 0.91 | 0.999 | 0.999 |
| | Nigrosin | 0.85 | 0.93 | 0.999 | 1 |
| | Ambient Air | 0.96 | 0.99 | 0.999 | 0.999 |
| Group E | | | | | |
| E01 | Soot | 0.95 | 0.98 | 0.999 | 1 |
| | Nigrosin | ND | ND | ND | ND |
| | Ambient Air | 1.01 | 1.02 | 0.995 | 0.992 |
| E02 | Soot | 0.88 | 1.00 | 0.998 | 1 |
| | Nigrosin | ND | ND | ND | ND |
| | Ambient Air | 1.15 | 0.93 | 0.996 | 0.993 |
| Group F | | | | | |
| F01 | Soot | 1.01 | 1.00 | 1 | 1 |
| | Nigrosin | ND | ND | ND | ND |
| | Ambient Air | 1.07 | 0.87 | 0.996 | 0.994 |
| F02 | Soot | 0.99 | 1.01 | 1 | 1 |
| | Nigrosin | ND | ND | ND | ND |
| | Ambient Air | 1.10 | 0.87 | 0.996 | 0.995 |
| F03 | Soot | 1.03 | 1.04 | 1 | 1 |
| | Nigrosin | ND | ND | ND | ND |
| | Ambient Air | 1.02 | 1.02 | 0.998 | 0.995 |

| Instrument | Aerosol source | Slope relative to AE33 REF | | Adjusted R^2 | |
|------------|----------------|----------------------------|------|----------------|-------|
| | | BM | AM | BM | AM |
| F04 | Soot | 1.05 | 1.08 | 1 | 1 |
| | Nigrosin | ND | ND | ND | ND |
| | Ambient Air | 1.05 | 1.02 | 0.998 | 0.994 |

BM: Before maintenance. AM: After maintenance. ND: No data.

The instrumental noise defined as the single standard deviation of the eBC mass concentration, was calculated with measurements of dry filtered air (particle-free, RH < 40 %), reported with a time resolution of one-minute. Measurements were performed during 4 ± 1 hours in average. The noise dependency on the wavelength was also studied for each one of the twenty-three instruments, as well as the influence of the maintenance activities. The results from this analysis are summarized in Fig. 13 and Table S2 from the supplementary material.

In general, the instrumental noise decreased after service, with more significant changes in the lower and middle wavelengths. The average noise at 1 min time resolution calculated at 370 nm, ranged from $0.030 \mu\text{g m}^{-3}$ before maintenance, to $0.023 \mu\text{g m}^{-3}$ after maintenance, which means a decrease of 32 %. At 660 nm, the average value of the instrumental noise dropped from 0.046 to $0.033 \mu\text{g m}^{-3}$, implicating a reduction of 40 %, which is the highest average reduction for the 7 wavelengths of the AE33. In the near-IR wavelengths the noise remained almost constant; at 880 nm the average noise did not change, passing from 0.032 to $0.031 \mu\text{g m}^{-3}$ before and after service, respectively; at 950 nm the average noise was $0.032 \mu\text{g m}^{-3}$ before and after maintenance. Larger noise values in the near-UV in multi-wavelength aethalometers have been also found in previous intercomparison exercises when aethalometers AE31 were intercompared (Müller et al., 2011).

From the average concentrations of ambient eBC (880 nm) measured every minute, the calculated instrumental noise represented between 1 and 10 % of the concentrations measured in the urban background in Leipzig.




Figure 13. Average instrument noise at the measurement wavelengths at 1 min time resolution. The black horizontal line inside the boxes represents the median noise value; the lower and upper borders of the boxes are the first and third quartiles, on which the middle 50% of the noise values are located; the whiskers correspond to ranges for the bottom 25% and the top 25% of the noise without outliers, which are represented by the black points. The red dashed line represents the noise equaling to zero. BM: Before maintenance. AM: After maintenance.

3.4 Wavelength-dependency of the light absorption


The absorption Ångström exponents (α) were calculated for soot and ambient air measurements, by applying a power law fitting describing the wavelength (λ) dependency of the aerosol light absorption (b_{abs}):

$$b_{abs} = A \lambda^{-\alpha}, \quad (9)$$

Where A is a constant. The absorption coefficients b_{abs} were first determined from Eq. 5, using the 5 minute-averaged eBC mass concentrations, and the default values of the mass absorption cross sections (σ_{air}) used by the AE33 for each wavelength (370 nm: 18.47 m²·g⁻¹, 470 nm: 14.54 m²·g⁻¹, 520 nm: 13.14 m²·g⁻¹, 590 nm: 11.58 m²·g⁻¹, 660 nm: 10.35 m²·g⁻¹, 880 nm: 7.77 m²·g⁻¹, 950 nm: 7.19 m²·g⁻¹; Magee Scientific, 2018).

Figure 14 shows the histograms of α estimated for each instrument in group D and the reference aethalometer, before and after maintenance. During soot measurements (Fig. 14a), the median absorption Ångström exponents before maintenance, ranged from 1.19 to 1.30; after maintenance, the median values fluctuated from 1.21 to 1.29. For ambient air (Fig. 14b), the median α before maintenance varied from 1.43 to 1.77; after maintenance, the median α ranged from 1.34 to 1.4. For both aerosol sources, the variability of the absorption Ångström exponents were reduced after maintenance (soot: Interquartile range IQR before maintenance = 0.08, IQR after maintenance = 0.05; ambient air: IQR before maintenance = 0.10, IQR after maintenance = 0.07). Values of α larger than 1, may indicate the presence of organic compounds in the aerosol particles of soot and ambient air. It has been demonstrated that α is also dependent in the aerosol size and coating (Liu et al., 2018; Virkkula, 2020).

The values of the absorption Ångström exponents shown in Fig. 14 were calculated using the absorption from channels 1 to 7 (370 nm to 950 nm); some studies suggest the omission of $b_{abs, 370 \text{ nm}}$ reduces the uncertainty in the estimation of the absorption Ångström exponent and their use in source apportionment models (Zotter et al., 2017). We have revised the impact of calculating α with measurements from six channels (470 nm to 950 nm), but no significant advantage or improvement was found from the omission of $b_{abs, 370 \text{ nm}}$ while calculating α from the samples measured in this study.



550 **Figure 14.** Absorption Ångström exponent (α) calculated using power law fitting of the absorption from 370 to 950 nm, for the instruments in group D and the reference aethalometer, before and after maintenance during measurements of (a) soot, and (b) ambient air. BM: Before maintenance. AM: After maintenance.

4. Summary, conclusions and recommendations

In this study, we presented the methodology and results from a comprehensive characterization and intercomparison workshop for aethalometers model AE33 (Magee Scientific). Twenty-three instruments were intercompared at the World Calibration Centre for Aerosol Physics (WCCAP) in Germany, measuring and reporting eBC mass concentrations of laboratory-produced aerosols and ambient air at an urban background site. The instruments received maintenance and were compared against the reference WCCAP aethalometer. The influence of maintenance activities, the filter material and different aerosol sources, in the instrumental variabilities were investigated.

560 The average unit-to-unit variability in the measurements of eBC mass concentrations (880 nm) reported by the 23 instruments was -2.0 % for soot and 0.4 % for nigrosin, before maintenance. After the maintenance activities, the average variabilities were

-1.0 % and 0.5 % for soot and nigrosin, respectively. The average variabilities were calculated using data from measurements performed after one or two filter tape advances to stabilize the internal correction algorithm, as variabilities increased few minutes before a tape advance, when the attenuation is close to 120; in some cases, the offsets among the instruments reached up to > 25 %. Tape advances are also crucial since the instruments need to fully adjust to the new aerosol sources and local conditions, and calculate appropriate values of the compensation parameters k . The aerosol composition and the filter material exert an influence on the rate of attenuation change and the k values. The combination of these factors directly influences the compensated eBC mass concentrations. Therefore, it is recommended to allow one or two tape advances in the aethalometers before the valid data are obtained for intercomparison purposes, or when the instrument is moved to a new location. For ambient air, the calculation of the total average variability may be biased, as the concentrations measured in the workshops were low (some days < 0.30 $\mu\text{g m}^{-3}$) and stable during the intercomparisons, and unequal among the different groups of instruments. Nevertheless, the results within groups were satisfactory even for very low concentrations measured with a time resolution of one minute.

Scientists performing laboratory experiments need to understand the functioning of the internal compensation algorithm – the compensation parameter used immediately after the tape is advanced and measurements are restarted, is the parameter from the previous measurements. If the sample being measured has changed (in our case measurements of soot after measurements of nigrosin, for example), the compensation algorithm will use the wrong compensation parameter, which was determined from the previous sample spot. The users need to wait for a new tape advance before using the automatically compensated data. If the experimental setup does not allow for this, manual compensation is necessary.

One of the most important characteristics of the aethalometer AE33 is the reporting of eBC mass concentrations at seven wavelengths. From the intercomparison data analysis, no significant influence of the wavelength in the unit-to-unit variabilities was seen. This fact is important as the spectral range covered by this instrument is usually employed in source apportionment studies. However, the instrumental noise calculated was slightly higher before maintenance in lower (370 nm: 0.030 $\mu\text{g m}^{-3}$, and 470 nm: 0.041 $\mu\text{g m}^{-3}$) and middle wavelengths (660 nm: 0.046 $\mu\text{g m}^{-3}$), and improved significantly after maintenance (370 nm: 0.023 $\mu\text{g m}^{-3}$; 470 nm: 0.031 $\mu\text{g m}^{-3}$; 660 nm: 0.033 $\mu\text{g m}^{-3}$). For higher wavelengths, the instrumental noise was lower and remained almost constant before and after maintenance (880 nm: 0.031 $\mu\text{g m}^{-3}$ and 950 nm: 0.032 $\mu\text{g m}^{-3}$). The instrumental noise was calculated as the single standard deviation of the eBC mass concentrations measured from dry particle-free air; it is important in clean environments with ambient air concentrations similar to those measured during the intercomparisons (0.2 to 3 $\mu\text{g m}^{-3}$). Noise accounted up to 10 % of the average ambient eBC mass concentration reported. The instrumental noise contributes to the uncertainty of the measurements, and must be considered when comparing BC observations intra-sites and inter-sites, in monitoring networks. The unit-to-unit variabilities also contributes to the uncertainties in the reported eBC inside monitoring networks; these will account for uncertainties associated with instrument flow calibration, leakages, internal correction factors and filter material.

To ensure the measurements of eBC made with filter-based absorption photometers are comparable, reliable and traceable on time, the performance of intercomparisons and maintenance activities are crucial; the type of filter material employed is also

very important, as each filter has specific optical properties affecting the measurements of attenuation, used to calculate black carbon mass concentrations. The utilization of a different filter material may result in differences up to 30 %, even after completing the standard maintenance activities. In the AE33 it is strongly recommended to use the most recent version of filter tape (M8060), and avoid the use of older versions. However, if an aethalometer operates with an older type of filter, it is absolutely necessary to check the use of the corresponding multiple scattering parameter C in the internal settings. This value needs to be checked and confirmed each time the filter tape is changed. Both, the filter type and the correction parameter C , should be reported alongside the measurement data when submitting the measurements to databases like EBAS. The change of tape from one (T60A20, also referred to as M8020) type to the current one (M8060), requires the user to change the multiple-scattering parameter C and the leakage factor Zeta. In our experience, this is the most important systematic error in the measurements that the user can make in a field or laboratory measurement campaign. The operators must perform maintenance following the frequencies and instructions given in the user manual. Flow verification, a leakage test and a check of the spot shape are good starting points to verify the instrument performance. When carrying out a flow verification and calibration, an externally calibrated flowmeter should be used and care should be taken on the flow reporting conditions, otherwise the test and calibration are not reliable. Cleaning the optical chamber and checking the absence of blocking materials in the airflow sample lines, is also vital and should be done more frequently in polluted environments.

Data availability. Experiments data will be available at the public data repository EBAS managed by NILU <http://ebas.nilu.no/default.aspx>, following the data policies from EBAS, COLOSSAL and ACTRIS.

Supplement. The supplement related to this article is available online at:

Author contributions. LD, ACM, TM, GM, SP and AW planned and designed the study. All co-authors participated in the experiments. ACM processed the data and prepared the manuscript with inputs from LD, TM, GM and AW.

Acknowledgements.

- ACTRIS (Aerosols, Clouds and Trace gases Research Infrastructure) and COLOSSAL (COST Action CA16109 Chemical On-Line cOmpoSition and Source Apportionment of fine aerosoL), for their financial and logistical support in the preparation and execution of the workshops.
- The WCCAP for hosting the experiments in their laboratories.
- Aerosol d.o.o., Slovenia, the manufacturer of the Aethalometer AE33, for troubleshooting support during the workshops.
- M. Fernández-Amado acknowledges to Ministerio de Ciencia e Innovación (PTA2017-13607-I).

→ Instrument S04-00387 is operated and maintained thanks to the Spanish Ministry of Science and innovation through the projects CGL2016-81092-R, CGL2017-42 90884REDT and RTI2018.101154.A.I00.
→ G. Titos is funded by Juan de la Cierva-Incorporación postdoctoral program (IJC-2016-29838).
→ IMT Lille Douai acknowledges financial support from the Labex CaPPA project, which is funded by the French National Research Agency (ANR) through the PIA (Programme d'Investissement d'Avenir) under contract ANR-11-LABX-0005-01, and the CLIMIBIO project, both financed by the Regional Council "Hauts-de-France" and the European Regional Development Fund (ERDF).
→ Instruments S00-00055 and S02-00156 operation is funded by projects CRISOL (CGL2017-85344-R MINECO/AEI/FEDER, UE) and TIGAS-CM (Y2018/EMT-5177).

Competing interests. LD and GM have been in the past, but not during the described work or its planning, employed by the manufacturer of the Aethalometer AE33; they are now employed in part by Haze Instruments d.o.o., Slovenia.

REFERENCES

- Ajtai, T., Kiss-Albert, G., Utry, N., Tóth, A., Hoffer, A., Szabó, G., Bozóki, Z.: Diurnal variation of aethalometer correction factors and optical absorption assessment of nucleation events using multi-wavelength photoacoustic spectroscopy, *J Environ Sci*, 83, 96-109, doi:10.1016/j.jes.2019.01.022, 2019.
- Bond, T. C., Doherty, S. J., Fahey, D. W., Foster, P. M., Berntsen, T., DeAngelo, B. J., Flanner, M. G., Ghan, S., Kärcher, B., Koch, D., Kinne, S., Kondo, Y., Quinn, P. K., Sarofim, M. C., Schultz, M. G., Schulz, M., Venkataraman, C., Zhang, H., Zhang, S., Bellouin, N., Guttikunda, S. K., Hopke, P. K., Jacobson, M. Z., Kaiser, J. W., Klimont, Z., Lohmann, U., Schwarz, J. P., Shindell, D., Storelvmo, T., Warren, S. G., and Zender, C. S.: Bounding the role of black carbon in the climate system: A scientific assessment, *J Geophys Res-Atmos*, 118(11), 5380–5552, doi:10.1002/jgrd.50171, 2013.
- Chen, J., Wu, Z., Augustin-Bauditz, S., Grawe, S., Hartmann, M., Pei, X., Liu, Z., Ji, D. and Wex, H.: Ice-nucleating particle concentrations unaffected by urban air pollution in Beijing, China, *Atmos. Chem. Phys.*, 18(5), 3523–3539, doi:10.5194/acp-18-3523-2018, 2018.
- Collaud Coen, M., Weingartner, E., Apituley, A., Ceburnis, D., Fierz-Schmidhauser, R., Flentje, H., Henzing, J. S., Jennings, S. G., Moerman, M., Petzold, A., Schmid, O. and Baltensperger, U.: Minimizing light absorption measurement artifacts of the Aethalometer: Evaluation of five correction algorithms, *Atmos. Meas. Tech.*, 3(2), 457–474, doi:10.5194/amt-3-457-2010, 2010.
- Cornbleet, P. J. and Gochman, N.: Incorrect least-squares regression coefficients in method-comparison analysis, *Clin. Chem.*, 25, 432-438, doi: 10.1093/clinchem/25.3.432, 1979.

- Drinovec, L., Močnik, G., Zotter, P., Prévôt, A. S. H., Ruckstuhl, C., Coz, E., Rupakheti, M., Sciare, J., Müller, T.,
660 Wiedensohler, A. and Hansen, A. D. A.: The “dual-spot” Aethalometer: An improved measurement of aerosol black carbon
with real-time loading compensation, *Atmos. Meas. Tech.*, 8(5), 1965–1979, doi:10.5194/amt-8-1965-2015, 2015.
- Drinovec, L., Gregoric, A., Zotter, P., Wolf, R., Anne Bruns, E., Bruns, E. A., Prevot, A. S. H., Favez, O., Sciare, J., Arnold,
I. J., Chakrabarty, R. K., Moosmüller, H., Filep, A. and Mocnik, G.: The filter-loading effect by ambient aerosols in filter
absorption photometers depends on the coating of the sampled particles, *Atmos. Meas. Tech.*, 10(3), 1043–1059,
665 doi:10.5194/amt-10-1043-2017, 2017.
- EEA: EEA Technical report No 18/2013 Status of black carbon monitoring in ambient air in Europe. European Environment
Agency, Luxembourg, Luxembourg, 2013.
- Gundel, L. A., Dod, R. L., Rosen, H. and Novakov, T.: The relationship between optical attenuation and black carbon, *Sci.
Total Environ.*, 36, 197–202, doi:10.1016/0048-9697(84)90266-3, 1984.
- 670 Hansen, A. D. A., Rosen, H. and Novakov, T.: Real-time measurement of the absorption coefficient of aerosol particles, *Appl.
Opt.*, 21(17), 3060–3062, 1982.
- Hansen, A. D. A., Rosen, H., and Novakov, T.: The Aethalometer - an instrument for the real-time measurement of optical
absorption by aerosol particles, *Sci. Total Environment*, 36, 191-196, doi:10.1016/0048-9697(84)90265-1, 1984.
- Hitzenberger, R., Jennings, S. G., Larson, S. M., Dillner, A., Cachier, H., Galambos, Z., Rouc, A. and Spain, T. G.:
675 Intercomparison of measurement methods for black carbon aerosols, *Atmos. Environ.*, 33(17), 2823–2833,
doi:10.1016/S1352-2310(98)00360-4, 1999.
- Hitzenberger, R., Petzold, A., Bauer, H., Ctyroky, P., Pouresmaeil, P., Laskus, L. and Puxbaum, H.: Intercomparison of thermal
and optical measurement methods for elemental carbon and black carbon at an urban location, *Environ. Sci. Technol.*, 40(20),
6377–6383, doi:10.1021/es051228v, 2006.
- 680 Janssen, N. A. H., Hoek, G., Simic-Lawson, M., Fischer, P., van Bree, L., Brink, H. Ten, Keuken, M., Atkinson, R. W., Ross
Anderson, H., Brunekreef, B. and Cassee, F. R.: Black carbon as an additional indicator of the adverse health effects of airborne
particles compared with PM10 and PM2.5, *Environ. Health Perspect.*, 119(12), 1691–1699, doi:10.1289/ehp.1003369, 2011.
- Jing Ltd: Real Soot Generator Model 5203 Type C miniCAST Instruction Manual, Jing Aerosol, Switzerland, 2013.
- Koch, D., Bauer, S. E., Del Genio, A., Faluvegi, G., McConnell, J. R., Menon, S., Miller, R. L., Rind, D., Ruedy, R., Schmidt,
685 G. A., and Shindell, D.: Coupled aerosol-chemistry–climate twentieth-century transient model investigation: trends in short-
lived species and climate responses; *J. Climate*, 24(11), 2693-2714, doi:10.1175/2011JCLI3582.1, 2011
- Lack, D. A., Moosmüller, H., McMeeking, G. R., Chakrabarty, R. K. and Baumgardner, D.: Characterizing elemental,
equivalent black, and refractory black carbon aerosol particles: A review of techniques, their limitations and uncertainties,
Anal. Bioanal. Chem., 406(1), 99–122, doi:10.1007/s00216-013-7402-3, 2014.
- 690 Liu, C., Chung, C. E., Yin, Y. and Schnaiter, M.: The absorption Ångström exponent of black carbon: From numerical aspects,
Atmos. Chem. Phys., 18(9), 6259–6273, doi:10.5194/acp-18-6259-2018, 2018.
- Magee Scientific: Aethalometer Model AE33 User Manual Version 1.57, Aerosol d.o.o., Ljubljana, Slovenia, 2018.

- Müller, T., Henzing, J. S., De Leeuw, G., Wiedensohler, A., Alastuey, A., Angelov, H., Bizjak, M., Collaud Coen, M., Engström, J. E., Gruening, C., Hillamo, R., Hoffer, A., Imre, K., Ivanow, P., Jennings, G., Sun, J. Y., Kalivitis, N., Karlsson, H., Komppula, M., Laj, P., Li, S. M., Lunder, C., Marinoni, A., Martins Dos Santos, S., Moerman, M., Nowak, A., Ogren, J. A., Petzold, A., Pichon, J. M., Rodriguez, S., Sharma, S., Sheridan, P. J., Teinilä, K., Tuch, T., Viana, M., Virkkula, A., Weingartner, E., Wilhelm, R. and Wang, Y. Q.: Characterization and intercomparison of aerosol absorption photometers: Result of two intercomparison workshops, *Atmos. Meas. Tech.*, 4(2), 245–268, doi:10.5194/amt-4-245-2011, 2011.
- Petzold, A., Kopp, C. and Niessner, R.: The dependence of the specific attenuation cross-section on black carbon mass fraction and particle size, *Atmos. Environ.*, 31(5), 661–672, doi:10.1016/S1352-2310(96)00245-2, 1997.
- Petzold, A., Ogren, J. A., Fiebig, M., Laj, P., Li, S. M., Baltensperger, U., Holzer-Popp, T., Kinne, S., Pappalardo, G., Sugimoto, N., Wehrli, C., Wiedensohler, A. and Zhang, X. Y.: Recommendations for reporting black carbon measurements, *Atmos. Chem. Phys.*, 13(16), 8365–8379, doi:10.5194/acp-13-8365-2013, 2013.
- Ramanathan, V. and Carmichael, G.: Global and regional climate changes due to black carbon, *Nat. Geosci.*, 1, 221–227, doi:10.1038/ngeo156, 2008.
- Ruoss, K., Dlugi, R., Weigl, C., Hänel, G.: Intercomparison of different aethalometers with an absorption technique: laboratory calibrations and field measurements, *Atmos. Environ.*, 27A(8), 1221–1228, doi:10.1016/0960-1686(92)90472-W, 1993.
- Saturno, J., Pöhlker, C., Massabò, D., Brito, J., Carbone, S., Cheng, Y., Chi, X., Ditas, F., Hrab De Angelis, I., Morán-Zuloaga, D., Pöhlker, M. L., Rizzo, L. V., Walter, D., Wang, Q., Artaxo, P., Prati, P. and Andreae, M. O.: Comparison of different Aethalometer correction schemes and a reference multi-wavelength absorption technique for ambient aerosol data, *Atmos. Meas. Tech.*, 10(8), 2837–2850, doi:10.5194/amt-10-2837-2017, 2017.
- Therneau, T.: Deming, Theil-Sen, Passing-Bablok and Total Least Squares Regression. R package version 1.4, <https://CRAN.R-project.org/package=deming>, 2018.
- Tørseth, K., Aas, W., Breivik, K., Fjæraa, A. M., Fiebig, M., Hjellbrekke, A. G., Lund Myhre, C., Solberg, S. and Yttri, K. E.: Introduction to the European Monitoring and Evaluation Programme (EMEP) and observed atmospheric composition change during 1972–2009, *Atmos. Chem. Phys.*, 12(12), 5447–5481, doi:10.5194/acp-12-5447-2012, 2012.
- Weingartner, E., Saathoff, H., Schnaiter, M., Streit, N., Bitnar, B. and Baltensperger, U.: Absorption of light by soot particles: Determination of the absorption coefficient by means of aethalometers, *J. Aerosol Sci.*, 34(10), 1445–1463, doi:10.1016/S0021-8502(03)00359-8, 2003.
- Virkkula, A., Mäkelä, T., Hillamo, R., Yli-Tuomi, T., Hirsikko, A., Hämeri, K., and Koponen, I. K.: A simple procedure for correcting loading effects of aethalometer data, *J. Air & Waste Manage.*, 57, 1214–1222, doi:10.3155/1047-3289.57.10.1214, 2007.
- Virkkula, A.: Modeled source apportionment of black carbon particles coated with a light-scattering shell, *Atmos. Meas. Tech. Discuss.* [preprint], doi:10.5194/amt-2020-438, in review, 2020.
- Wex, H., Huang, L., Zhang, W., Hung, H., Traversi, R., Becagli, S., Sheesley, R. J., Moffett, C. E., Barrett, T. E., Bossi, R., Skov, H., Hünerbein, A., Lubitz, J., Löffler, M., Linke, O., Hartmann, M., Herenz, P. and Stratmann, F.: Annual variability of

ice nucleating particle concentrations at different Arctic locations, *Atmos. Chem. Phys. Discuss.*, 1–31, doi:10.5194/acp-2018-1274, 2018.

730 WHO: Health effects of black carbon. World Health Organization Regional Office for Europe, Copenhagen, Denmark, 86 pp., 2012.

WHO: GAW Report No. 227, WMO/GAW Aerosol Measurement Procedures, Guidelines and Recommendations, World Meteorological Organization and Global Atmosphere Watch, Geneva, Switzerland, 93 pp., 2016.

735 Zotter, P., Herich, H., Gysel, M., El-Haddad, I., Zhang, Y., Mocnik, G., Hüglin, C., Baltensperger, U., Szidat, S. and Prévôt, A. S. H.: Evaluation of the absorption Ångström exponents for traffic and wood burning in the Aethalometer-based source apportionment using radiocarbon measurements of ambient aerosol, *Atmos. Chem. Phys.*, 17(6), 4229–4249, doi:10.5194/acp-17-4229-2017, 2017.