

Impact of ozone and inlet design on the ~~detection~~ quantification of isoprene-derived organic nitrates by thermal dissociation cavity ring-down spectroscopy (TD-CRDS)

Patrick Dewald¹, Raphael Dörich¹, Jan Schuladen¹, Jos Lelieveld¹, and John N. Crowley¹

5 ¹Atmospheric Chemistry Department, Max-Planck-Institut für Chemie, 55128 Mainz, Germany

Correspondence to: John N. Crowley (john.crowley@mpic.de)

Abstract. We present measurements of isoprene-derived organic nitrates (ISOP-NITs) generated in the reaction of isoprene with the nitrate radical (NO₃) in a 1 m³ Teflon reaction chamber. Detection of ISOP-NITs is achieved via their thermal dissociation to nitrogen dioxide (NO₂), which is monitored by cavity ring-down spectroscopy (TD-CRDS). Using thermal dissociation inlets (TDIs) made of quartz, the temperature-dependent dissociation profiles (thermograms) of ISOP-NITs measured in the presence of ozone (O₃) are broad (350 to 700 K), which contrasts the narrower profiles previously observed for e.g. isopropyl nitrate (iPN) or peroxy acetyl nitrate (PAN) under the same conditions. The shape of the thermograms varied with the TDI's surface to volume ratio and with material of the inlet walls, providing clear evidence that ozone and quartz surfaces catalyse the dissociation of unsaturated organic nitrates leading to formation of NO₂ at temperatures well below 475 K, impeding the separate detection of alkyl nitrates (ANs) and peroxy nitrates (PNs). ~~We present a simple, viable solution to this problem and discuss the potential for interference by the thermolysis of nitric acid (HNO₃), nitrous acid (HONO) and O₃.~~ The use of a TDI consisting of a non-reactive material suppresses the conversion of isoprene-derived ANs at 473 K, thus allowing selective detection of PNs. The potential for interference by the thermolysis of nitric acid (HNO₃), nitrous acid (HONO) and O₃ is assessed.

20 **1 Introduction**

Understanding the atmospheric fate of nitrogen oxide (NO) and nitrogen dioxide (NO₂) is critical as both trace-gases have a great impact on air quality and human health (Crutzen and Lelieveld, 2001; Lelieveld et al., 2015). Ambient measurement of trace-gases that function as NO_x reservoirs or sinks (where NO_x = NO + NO₂) are thus needed to provide insight into NO_x removal and transport. Organic compounds with nitrate functionality can serve as NO_x reservoirs in the troposphere (Thornton et al., 2002; Horowitz et al., 2007) and are generally categorised as peroxy nitrates (PNs, RO₂NO₂, with peroxy acetyl nitric anhydride (PAN) being its most abundant representative in the troposphere) and alkyl (aliphatic) nitrates (ANs, RONO₂). PNs are formed via the reaction of organic peroxy radicals (RO₂) with NO₂ (R1); ANs are formed via the minor (termolecular) channel of the reaction of RO₂ with NO (R2a). The competitive bimolecular process leads to alkoxy radicals (RO, R2b). During

the daytime, RO₂ is formed mainly by the oxidation of volatile organic compounds (VOCs) by hydroxyl radicals (OH) in air

30 (R3), with ozonolysis important at night (R4):



At nighttime, when OH radicals and NO are significantly less abundant, the NO₃ radical **can** initiate the oxidation of many VOCs **that contain a double-bond** (Ng et al., 2017). NO₃, formed in the oxidation of NO₂ by O₃ (R5), is photolysed rapidly by sunlight (R6) and also reacts efficiently with NO (R7) so that it is generally of minor importance during the day (Wayne et al., 1991).




NO₃ readily undergoes reactions with many unsaturated organic trace-gases of biogenic origin including isoprene, monoterpenes and sesqui-terpenes to form organic nitrates in high yields (Ng et al., 2017; Wennberg et al., 2018; Mellouki et al., 2020). The focus of this work is the formation of organic nitrates in its reaction (in air) with the C₅-di-ene isoprene (ISOP, R8)



where ISOP-NIT represents an isoprene-derived nitrate. Isoprene is **with a total global emission of $\approx 500 \text{ Tg yr}^{-1}$ significantly released to the most abundant non-methane VOC** in the atmosphere (Guenther et al., 2012) and a large fraction of the organic nitrates formed at night-time is attributed to the reaction between NO₃ and isoprene (R8) (Carlton et al., 2009). The atmospheric oxidation of isoprene involving OH, O₃ and NO₃ as oxidizing agents is complex and leads to a huge variety of products (Ng et al., 2017; Wennberg et al., 2018) including multifunctional, unsaturated nitrates such as O₂NOCH₂C(CH₃)=CHCHO, O₂NOCH₂C(CH₃)=CHCH₂OH or O₂NOCH₂C(CH₃)=CHCH₂OOH among other secondary oxidation products like dinitrates or epoxides (Wu et al., 2020).

Studies of the NO₃-induced oxidation of isoprene in air report AN yields between 60 and 100 % (Barnes et al., 1990; Berndt and Boge, 1997; Perrig et al., 2009; Kwan et al., 2012; Schwantes et al., 2015; IUPAC, 2017; Wu et al., 2020; Brownwood et al., 2021) and the NO₃-induced oxidation of isoprene is responsible for a dominant fraction of organic nitrates observed in rural environments with strong biogenic emissions (Beaver et al., 2012). The major fate of isoprene-derived organic nitrates formed in the boundary layer is deposition onto particulate matter to form nitric acid (HNO₃) or secondary organic aerosols (SOA) leading to largely irreversible removal of NO_x from the gas phase (Ng et al., 2008; Rollins et al., 2009; Fry et al., 2018; 60 Hamilton et al., 2021).

Individual isoprene nitrates have been measured selectively in the atmosphere by mass-spectrometric methods (Wolfe et al., 2007; Wu et al., 2020). An alternative detection scheme, in which the sum of all atmospheric PNs (Σ PNs) and ANs (Σ ANs) are separately measured, takes advantage of their different C-N bond energy by combining thermal dissociation to NO_2 (R9 and R10) with detection of the latter with means of laser-induced fluorescence (TD-LIF) or cavity ring-down spectroscopy (TD-CRDS).



Several instruments using thermal dissociation inlets consisting of fused silica (quartz) with residence times between tens to hundreds of milliseconds have been described in the literature (Day et al., 2002; Paul et al., 2009; Wild et al., 2014; Sobanski et al., 2016; Thieser et al., 2016; Keehan et al., 2020). In these instruments, quantitative conversion of PAN is reported for temperatures between 375 and 420 K. Generally, the temperature dependence of the instrument's response to ANs has been tested using mostly saturated organic nitrates such as isopropyl nitrate (iPN) and isobutyl nitrate, which are dissociated to NO_2 at temperatures between 500 and 675 K. The temperatures at which PNs and ANs are quantitatively converted to NO_2 thus differ by ~ 200 K and are largely independent of their organic backbone (Kirchner et al., 1999; Wild et al., 2014) allowing separate measurement of the sum of all alkyl nitrates (Σ ANs) and of the sum of all peroxy nitrates (Σ PNs). These observations have provided the basis for analysis of field data in which an unknown mixture of PNs and ANs are present. We note however, that most of the first generation ANs formed in the $\text{NO}_3 +$ isoprene system still contain a double-bond (Barnes et al., 1990; Skov et al., 1992; Schwantes et al., 2015) which renders them more reactive towards oxidizing agents than e.g. iPN. A well characterised thermogram for aliphatic nitrates derived from the oxidation of e.g. isoprene is thus a pre-requirement for extracting the mixing ratios of PNs and ANs from ambient measurements when using a TD-inlet. To date, only one such thermogram has been presented (Brownwood et al., 2021) which appears to be the result of a single experiment (i.e. no variation of experimental conditions) using a sample that was not stable over time. The thermogram also features slopes before and after the ANs transition temperature which is consistent with the ideal behaviour of e.g. iPN.

In this study, we generated ISOP-NITs by reacting isoprene and NO_3 in a Teflon simulation chamber and used a custom-built, five-channel Cavity-Ring-Down Spectrometer (CRDS) (Sobanski et al., 2016) to analyse the organic nitrates formed. In the presence of O_3 we find that ISOP-NIT does not behave like the saturated analogue iPN in our quartz TD-inlet and we characterised the processes (both gas-phase and surface-catalysed) that lead to the observed behaviour. We also examined the potential role of surface-catalysed dissociation of HNO_3 and nitrous acid (HONO) to NO_2 as well as the effect of humidity as a potential bias to measurements of PNs and ANs.

90 2 Experimental

2.1 Simulation Chamber

In order to analyse organic nitrates formed from the NO_3 + isoprene system under realistic operational conditions for the 5-Channel-CRDS (e.g. normal sample flow rates), we constructed a dynamic, flow-through simulation chamber SCHARK (Simulation CHamber for Atmospheric Reactions and Kinetics) of volume 1 m^3 (cubic, all sides $\sim 1 \text{ m}$ long) made of PFA foil of 0.005 in. ($\sim 0.13 \text{ mm}$) thickness (Ingeniven). The chamber is operated at ambient pressure and temperature; a magnetically coupled, Teflon coated propeller-type stirrer situated in the centre of the chamber floor ensures continuous mixing of the air. The trace-gas inlets and sampling ports were located at opposite corners of the cubic chamber to reduce the potential of sampling gas that had not yet mixed. The PFA foil is surrounded by a $120 \times 120 \times 120 \text{ cm}$ cube constructed of four Perspex and two steel walls, the interspace (0.7 m^3) is permanently flushed with 1 SLPM (L (STP) min^{-1}) of dry synthetic “zero-air” in order to avoid contamination through permeation of trace-gases present in the laboratory air. The Perspex walls serve as observation windows and were covered with light-tight material during the experiments described here.

Zero air was provided by passing pressurized air through a commercial air purifier (CAP 180, Fuhr GmbH). Humidification of the air was achieved with a permeation source (MH-110-24-F-4, Perma Pure LLC) filled with deionized water. Typical total flow rates of 15 or 23 SLPM zero air into the chamber result in exchange rates k_{exch} of 2.7 or $4.2 \times 10^{-4} \text{ s}^{-1}$, i.e. lifetimes of gases in the chamber of ~ 40 -60 minutes. Note that in “flow-through” operation, the concentrations of trace gases in the chamber are controlled both by chemical processes and by the rate of flow into (and out of) the chamber so that “steady-state” is achieved on the order of hours.

Ozone mixing ratios in the SCHARK were measured by sampling 2 SLPM through a $\sim 3 \text{ m}$ long section of 0.25 in. (outer diameter, OD) PFA-tubing to a commercial ozone monitor (2B Technologies, model 205) with a detection limit of ~ 1 part per billion by volume (ppbv) and 5 % uncertainty. O_3 measurements were also used to establish the time required (under standard flow conditions) to ~~acquire~~ achieve complete mixing within the chamber (< 1 minute) and to derive the exchange rate by monitoring the exponential rise or decay of O_3 when its supply was switched on or off (Fig. S1 in the Supplement). O_3 (up to 600 ppbv) was generated by passing a fraction of the air flowing into the chamber through a UV-transparent cuvette ($\sim 70 \text{ cm}^3$) illuminated by a low-pressure Hg-lamp (PenRay) that dissociated O_2 (to O atoms and thus O_3) at 185 nm.

. A known flow of isoprene entered the chamber as a dilute sample from a 12 L stainless steel storage canister (Landefeld GmbH) which was prepared manometrically from evaporation of pure isoprene (Acros Organics, 98 %) and mixing with helium (5.0, Westfalen). The isoprene concentration in the storage canister was quantified indirectly by measuring the NO_3 reactivity via flowtube-CRDS (Liebmann et al., 2017; Dewald et al., 2020) and was found to be 46.5 ppmv, in agreement (within 15 %) with the manometrically derived mixing ratio. A gas sample of isopropyl nitrate (Sigma Aldrich, 58 ppmv in N_2 5.0, Westfalen) was prepared in a similar fashion.

Two methods of in-situ NO_3 generation were employed. In the first, NO_3 was produced in the chamber via the reaction of NO_2 with O_3 (R5), whereby O_3 was generated as described above and NO_2 was taken from a bottled sample (Air Liquide, 1 ppmv

in N₂). Typical concentrations of NO₂ and O₃ were 6-10 and 100-160 ppbv, respectively. Alternatively, NO₃ was generated in the thermal decomposition of N₂O₅ (R11) which was eluted into the chamber by passing a regulated flow of N₂ over N₂O₅ crystals held at temperatures between -78 and -70°C.



N₂O₅ was synthesised by the sequential, gas-phase oxidation of NO (5% in N₂, Westfalen) in an excess of O₃ (Davidson et al., 1978) and trapped at -78°C (acetone and dry ice). In this case, O₃ was obtained by electrical discharge through oxygen (5.0, Westfalen) using a commercial generator (Ozomat Com, Anseros). Note that, the latter method enables us to generate NO₃ in the chamber in an O₃-free environment.

2.2 Detection of organic nitrates by cavity ring-down spectroscopy (CRDS)

Simultaneous measurements of the mixing ratios of NO₂, NO₃, N₂O₅, ΣANs and ΣPNs in the SCHARK chamber were made using a five-channel cavity ring-down spectrometer (CRDS) that has been described in detail (Sobanski et al., 2016) and only a brief summary of key features of the instrument are given here. Each of the five cavities consists of FEP-coated (FEPD 121, DuPont) stainless steel tubes which are equipped with two high-reflectivity mirrors (see below) supported 90 cm apart (*L*). The volumes in front of the mirrors are purged with dry synthetic air, which results in a reduction of the effective optical path length from 90 cm to 62.1 cm (*d*). The standard expression Eq. (1) is used to derive in-cavity concentrations [X] from the difference in ring-down constant in the absence (*k*₀) and presence (*k*) of an absorber X:

$$[\text{X}] = \frac{L}{d} \cdot \frac{1}{c\sigma_{eff}} \cdot (k - k_0) \quad (1)$$

where *c* is the speed of light and σ_{eff} is the effective cross-section derived from the overlap of the laser emission and the NO₂ (Vandaele et al., 1998) or NO₃ absorption spectrum (Orphal et al., 2003).

Three of the cavities are operated at 409 nm for detection of NO₂ whereby 409 nm light is provided by a square-wave-modulated (2500 Hz) laser-diode. The three 409 nm cavities, thermostated to 303 K and typically operated at a pressure of ~ ~~550 Torr~~ (733 hPa mbar), sampled from the SCHARK at a total flow rate of 6 SLPM, which initially passes through a 2.3 m long PFA inlet (1.5 m with OD 0.25 inch. and 0.8 m with OD 0.125 inch) before being split into three equal flows. One flow is directed to a cavity via an unheated, 60 cm long PFA tube (0.375 inch OD) to measure NO₂. The other two flows are directed through thermal dissociation inlets (TDIs) in which PN_s and AN_s are converted to NO₂. **At the given conditions (i.e. flow rate, pressure, residence time in the heated section), keeping our A TDI at temperatures close to 448 K, results in quantitative conversion of PN_s to NO₂ so that the cavity sampling via this inlet measures the sum of PN_s + NO₂. A hotter TDI (~ 650 K) Heating our second TDI to ~ 650 K results in the additional complete conversion of AN_s to NO₂ so that the sum of AN_s + PN_s + NO₂ can be measured as described in the literature cited in the introduction. The choice of material for these TD-inlets has a profound influence on the results obtained, as described below.**

The standard deviation (2σ) of consecutive baseline measurements define the limits of detection (LOD) which are 38, 44 and 90 pptv for [NO₂], [ΣPN_s] and [ΣAN_s] respectively under laboratory conditions. The total uncertainty for the NO₂

155 measurement is 9 % which includes uncertainty in the (effective) NO₂ cross-sections. For the measurements of ANs and PNs the associated uncertainties are highly dependent on the concentrations of other trace gases and the corrective procedure accounting for radical recombination effects (Sobanski et al., 2016).:

For simplicity, we refer to the three cavities as the “NO₂ cavity” (room-temperature inlet), the “PNs cavity” (TD-inlet at circa 473 K in which ΣPNs + NO₂ are measured) and the “ANs cavity” (TD-inlet at circa 673 K in which ΣANs + ΣPNs + NO₂ are measured).

The remaining two cavities of the CRDS were operated at 662 nm (laser modulation at 625 Hz) for detection of NO₃. While one cavity is thermostated to 303 K (and detects NO₃ only), the second one (as well as an FEP-coated glass reactor located upstream) is thermostated to 373 K so that N₂O₅ is stoichiometrically converted NO₃ and the summed mixing ratio of NO₃ and N₂O₅ is obtained. The two 662 nm cavities sampled air from the SCHARK at a total flow rate of 15 SLPM through a ~1.5 m
165 ¼ in. (OD) PFA tube. Corrections to the mixing ratios were made to account for loss of NO₃ and N₂O₅ during transport to and through the cavities. Using the method described in Sobanski et al. (2016), NO₃ transmission was found to be 89 % in both cavities. The NO₃ and N₂O₅ measurements are not central to this study, but allowed the quantitative surveillance of NO₃ (and indirect N₂O₅) consumption by isoprene.

Figure 1 shows three types of thermal dissociation inlets (TDIs) used to convert organic nitrates to NO₂. In the original version of this instrument (Sobanski et al., 2016) the ΣPNs and ΣANs cavities sampled via 12 mm ID quartz TD-inlets (TDI-1), with a length of 55 cm, the first ~10 cm of which was wrapped with heating wire. In order to reduce bias caused e.g. by the reformation of the organic nitrate after its thermal dissociation, this section was filled with glass beads (Sigma-Aldrich G9268, ø ~ 0.5 mm) to provide a surface for heterogeneous loss of radicals. **The glass beads were supported on a 2 cm thick glass frit and reduce the pressure downstream by ≈ 28 hPa compared to TDI-1.** Problems associated with temperature-dependent flow resistance through these small beads and the need for an extra filter (upstream) to prevent their transport into the inlet lines when flows were temporarily reversed (e.g. during instrument shut-down), led us to switch to larger beads (Merck, ø = 3 mm) and these were used throughout this study in TDI-1. TDI-2 is made of a quartz glass tube with the same dimensions as TDI-1, but features a longer heated section (20 cm) and is free of additional surfaces like glass beads or frits. TDI-2 is thus similar to many other thermal dissociation inlets described in the literature (see above). TDI-3 is constructed from a 55 cm long PFA
175 tube (0.375 inch OD), where the first 20 cm are heated. The melting point of PFA is lower than the temperature required to thermally dissociate ANs, so TDI-3 could only be used for the measurement of PNs + NO₂. The temperature of the external wall of the TD-inlets were measured with a K-type thermocouple situated at the centre of the heated section, which was insulated with mineral wool. At a flow rate of 2 SLPM and an operating pressure of ~~550 Torr~~ **733 mbar**, approximate residence times in the inlets without glass-beads are 0.20 s (in TDI-2 at 650 K) and 0.13 s (TDI-3 at 450 K) when assuming a
185 homogeneous temperature distribution equal to that measured on the outer wall of the tubing.

3 Results and Discussion

Figure 2 shows the result of an experiment in which 150 sccm NO₂ (1 ppmv in air) was flowed into the SCHARK along with isoprene (7 sccm of 46.5 ppmv in He) and 24 SLPM zero air of which 5 SLPM were passed over the low-pressure Hg-lamp zero-air to generate O₃ (~ 96 ppbv). NO₂ was sampled (as usual) via the room-temperature PFA inlet, the ΣPNs (473 K) and ΣANs cavities (673 K) both sampled via TDI-1 (quartz tube with glass-beads). O₃ was added at 09:30 and NO₂ at 10:00 (all times are local times, LT).

Just prior to the addition of isoprene at 12:00 LT, the system is close to steady-state with ~ 5 ppbv NO₂ and 92 ppbv O₃. After subtraction of the measured N₂O₅ mixing ratios, a residual signal of ~ 100 pptv is detected in both PNs and ANs channel, which may be caused, as discussed below, by interference of HNO₃ in the ANs channel and a memory effect of the glass beads in the PNs channel (section 3.2). Note that after addition of isoprene, both NO₃ and N₂O₅ are reduced drastically (NO₃ ~3 pptv, N₂O₅ ~ 5 pptv) and the thermal dissociation of N₂O₅ no longer contributes to NO₂ signals in the PNs and ANs channels (Sobanski et al., 2016; Thieser et al., 2016).

At 14:00 LT, the cavity sampling from the 673 K TD-inlet indicated ~ 610 pptv for the summed mixing ratio of (ΣANs + ΣPNs), whereas the cavity sampling from the 473 K TD-inlet (ΣPNs) indicated ~ 400 pptv. Since the signal in the ΣANs channel includes both the contribution of peroxy and alkyl nitrates this implies that only 210 pptv (34 %) of the detected products can be attributed to alkyl nitrates, which is inconsistent with the high yields (60-100 %) of ANs that result from the reaction of NO₃ with isoprene (Barnes et al., 1990; Berndt and Boge, 1997; Perring et al., 2009; Rollins et al., 2009; Kwan et al., 2012; Schwantes et al., 2015; IUPAC, 2017; Brownwood et al., 2021). Compared to ANs, we expect the mixing ratios of PNs (e.g. PAN, O₂NOCH₂C(CH₃)=CHC(O)O₂NO₂ or methacryloyl peroxy nitrate (MPAN)) in this system to be negligible as their precursors such as O₂NOCH₂C(CH₃)=CHCHO (Jenkin et al., 2015) or methacrolein (Kwok et al., 1996; Berndt and Boge, 1997; Schwantes et al., 2015) are oxidized only inefficiently in the dark. The formation of PNs only takes place once isoprene has been reduced to very low concentrations and depleted so that secondary oxidation of the above-mentioned aldehydes by OH or NO₃ leading to further acyl-peroxy radicals (which form PNs) become at least competitive to the primary oxidation of isoprene. This is however never the case in the present experiments as isoprene is continuously flowed into the chamber and remains according to model calculations (see below) at a level of ≈ 11.4 ppbv. Given the high abundance of O₃ and isoprene in this system, ozonolysis of the latter together with the associated formation and decomposition of Criegee intermediates to acetylperoxy radicals CH₃C(O)O₂ (Nguyen et al., 2016; Vansco et al., 2020) should make PAN the most important, potential contributor to a signal in the PNs cavity. However, according to the branching ratios given in Nguyen et al. (2016), this reaction path is a minor one and CH₃C(O)O₂ (and thus PAN) should be formed in negligible amounts.

In order to identify the origin of the unexpectedly high ΣPNs signal when NO₃ and isoprene are mixed in the dark, thermograms of the NO₃ + isoprene system were recorded in an experiment where ~ 2.8 ppbv of isoprene-derived nitrates (as measured with TDI-2 at 625 K) were generated by flowing NO₂ (200 sccm of 1 ppmv) and isoprene (9.8 sccm of 46.5 ppmv) in 15 SLPM dry synthetic air (with 5 SLPM over the Hg lamp for generation of ~ 150 ppbv O₃). Similar to the experiment in Fig. 2, N₂O₅

220 mixing ratios are expected to be suppressed to a few pptv under these conditions so that its thermal dissociation (to NO₂) did not contribute to the ΣPNs and ΣANs signals. Using this chemical system, we simultaneously measured ISOP-NIT thermograms once steady-state established using TDI-1 (quartz, glass beads, 10 cm heated section) and TDI-2 (quartz, no glass beads, 20 cm heated section) both initially held at 703 K. Subsequently, both TDIs were cooled to ambient temperature over a period of ~ 1.75 h. The ΣANs signals from this experiment are plotted against the inlet temperature in Fig. 3a to generate the ISOP-NIT thermogram. This is displayed along with an isopropyl-nitrate thermogram (iPN, red data points) measured using 225 the same inlets under the same flow-conditions but using iPN diluted to 5.5 ppbv in dry synthetic air sampled directly through a PFA line (together with 1.5 ppbv NO₂ impurity) to the instrument.

For iPN, we observe a well-defined onset of thermal dissociation at ~ 525 K with a plateau (maximum conversion) at ~ 650 K as reported previously for this setup (Sobanski et al., 2016). When measuring iPN, TDI-1 results in a slightly steeper thermogram than TDI-2 in the 575-650 K range, which may be related to changes in gas-flow and heat-transfer within the inlet 230 caused by the glass beads. Neither TD-inlet type results in dissociation to NO₂ at temperatures < 500 K. In contrast, the ISOP-NIT thermograms (normalised to the signal at the plateau at 625 K of TDI-2) indicate formation of NO₂ over a much broader range of temperatures (350-700 K).

The effect of humidifying the air was examined in an almost identical experiment conducted with NO₂ (150 sccm of 1 ppmv) and isoprene (7 sccm of 46.5 ppmv) in 15 SLPM synthetic air with relative humidity (in the SCHARK) of 33.5 % at 22°C. In 235 this case, ~ 2.3 ppbv ISOP-NIT was formed. The thermograms obtained with TDI-1 and TDI-2 under these conditions are depicted in Fig. 3b. The broad thermogram measured with TDI-1 is very similar to that obtained under dry conditions (Fig. 3a) although even at room temperature an additional NO₂ signal of 500 pptv is detected. Sampling via TDI-2, yields an ISOP-NIT thermogram that has similar features to that obtained under dry conditions, although the peak at ~400 K has well defined minima on both flanks and is shifted to higher temperatures. In separate experiments, humidified synthetic air (RH = 40 %, 240 23°C) and NO₂ (10.8 ppbv) were sampled through a PFA line directly to the instrument. The thermogram (in the absence of isoprene or ISOP-NIT) using TDI-2 was recorded and can be found in the Supplement (Fig. S2) revealing that the presence of water and NO₂ in the inlet is sufficient to reproduce some features displayed in Fig. 3b with TDI-2. It is well known that H₂O and NO₂ can react on surfaces to form HONO and HNO₃ (Pitts et al., 1984; Finlayson-Pitts et al., 2003) and their formation in the SCHARK was verified in section 3.1. In the presence of H₂O, the efficiency of conversion of ISOP-NIT to NO₂ drops 245 to about 5% at ~ 460 K. This is much less than under dry conditions whereby 20 % conversion of ISOP-NIT was observed between 375 and 475 K (Fig. 3a). Within a framework for surface-catalysed conversion of ISOP-NIT to NO₂ presented below, this observation can be interpreted as arising from the competitive adsorption to the surface of nitrated hydroperoxides and H₂O, i.e. H₂O (which is vastly more abundant) reduces the surface coverage of the organic nitrate at the surface.

The results in Fig.2 and Fig.3 show that separate detection of ANs and PNs based on their thermal dissociation can be 250 problematic for the NO₃ + isoprene system. Identifying the cause of this and providing potential solutions to circumvent the problem is the aim of this work. To do this, we first focus on the “dry” experiment and highlight two regions of the thermograms in which large deviations from the expected behaviour are observed.

Region I ($T > 648$ K)

255 Figure 3a indicates that, at temperatures above 648 K (shaded region I), the behaviour of the two TDIs diverges significantly: While use of TDI-1 (glass beads) results in an increase in NO_2 with increasing temperature, the use of TDI-2, leads to a decrease in the NO_2 signal in the same temperature range. The increase in NO_2 continues at temperatures above that required to convert ANs to NO_2 , which implies the presence of a NO_2 -containing trace-gas where the NO_2 -moiety is more strongly bound than in ANs.

260 In order to assess to which extent this behaviour is potentially caused by inorganic trace-gases that are not directly related to isoprene oxidation, an experiment with only NO_2 (2.75 ppbv) and O_3 (146 ppbv) in 23 SLPM dry synthetic air was performed. The steady-state concentration of $\text{N}_2\text{O}_5 + \text{NO}_3$ was measured as 78 pptv. The resulting thermograms using TDI-1 and TDI-2 and after subtraction of the signal from the NO_2 cavity (i.e. unheated inlet) are depicted in Fig. 4. No significant additional signal is observed below 475 K (region I) in either of the inlets. In region II ($T > 675$ K) on the other hand, we observed an

265 increase (by ~ 500 pptv) in the signal to at 703 K with TDI-1, whereas ~ 50 pptv are lost in TDI-2. In order to identify the trace-gas(es) responsible for the signals observed in the system without isoprene, an Iodide-Chemical-Ionization Mass Spectrometer (I-CIMS (Eger et al., 2019)) **described in the Supplement (S8)** was coupled to the experiment. As shown in the Supplement (Fig. S3) both HNO_3 and nitrous acid (HONO) were observed as soon as O_3 and NO_2 were present in the chamber and their formation is enhanced in the presence of water vapour, which is a common phenomenon in Teflon chambers (Pitts et al., 1984).

270 We also found that reversing the flows and sampling the air into the CIMS *after* passing through the TDI-1 or -2 (at 475 K) resulted in removal of the HNO_3 . Sampling through TDI-1 also led to **quantitative** loss of HONO.

Various TD-CRDS and TD-LIF instruments report the detection of HNO_3 as NO_2 following thermal dissociation at temperatures around 700 K (Day et al., 2002; Wild et al., 2014; Thieser et al., 2016). The sensitivity of the present set-up to HNO_3 was investigated by sampling nitric acid from a calibrated permeation source (Friedrich et al., 2020) via TDI-1 and TDI-

275 2 simultaneously. In these experiments, 22 ppbv HNO_3 (with 780 pptv NO_2 impurity) in dry synthetic air was delivered to the TDIs along with 350 ppbv O_3 . **The HNO_3 mixing ratio was derived using a known permeation rate (Friedrich et al. 2021) and dilution factor.** Figure 5 shows the temperature-dependent conversion efficiency of HNO_3 to NO_2 in the presence of ozone (squares) with TDI-1 (black solid symbols) and TDI-2 (open symbols). Conversion of HNO_3 to NO_2 starts at ~ 550 K and increases **with rising temperature as the temperature is ramped to 680 K.** At 680 K, the conversion efficiency is 23 % for TDI-

280 1 and 8 % for TDI-2. No significant conversion of HNO_3 to NO_2 was observed when ozone was absent (blue datapoints) and was drastically reduced when the synthetic air was humidified to $\text{RH} = 55$ % at room temperature (red datapoints). The effect of water vapour is consistent with previous observations on the effect of humidity (Sobanski et al., 2016; Thieser et al., 2016; Friedrich et al., 2020).

The decomposition of HNO_3 to NO_2 thus only occurs in the presence of ozone under dry conditions and its rate increases

285 greatly at $T > 650$ K. This is consistent with the observations in Fig. 3b for TDI-1 and represents a likely explanation for the increase in signal when sampling from the SCHARK to investigate the $\text{NO}_3 + \text{isoprene}$ system. The apparently more efficient

(~ factor three) conversion of HNO₃ to NO₂ in TDI-21 than in TDI-2 is explained by the loss of NO₂ at high temperatures in TDI-2 through the reaction with O-atoms (see section 3.3). In TDI-1 this is prevented by the removal scavenging of O-atoms by the glass beads (e.g. via scavenging or surface-catalysed recombination to O₂).

290 The ozone-assisted conversion of HNO₃ to NO₂ cannot be explained by known gas-phase processes as the reaction between HNO₃ and O(³P) (R13) has a low rate coefficient ($k_{13} < 3 \times 10^{-17} \text{ cm}^3\text{molecule}^{-1}\text{s}^{-1}$ at 298 K (Burkholder et al., 2016)) and results mainly in the formation of OH and NO₃ (R13). The more efficient conversion of HNO₃ to NO₂ in TDI-1 (with glass beads) compared to TDI-2 indicates that a surface-catalysed process involving either ozone or O(³P) is involved (R14).



295 $\text{HNO}_3 + \text{O}_3 \text{ or } \text{O}(\text{}^3\text{P}) + \text{surface} \rightarrow \text{NO}_2 + \text{products} \quad (\text{R14})$

Assuming that, in a Langmuir-Hinshelwood type process, the first step in the surface catalysed reaction is physis-adsorption of HNO₃ to the surface, the strong reduction in conversion of HNO₃ to NO₂ under humid conditions is explained by the competitive adsorption of HNO₃ and H₂O, the latter favoured by its much larger concentrations. i.e. H₂O drives HNO₃ from the surface and thus protects it from surface reactions.

300

Region II ($T = 350\text{-}475 \text{ K}$)

In region II (350-475 K, shaded area in Fig.3), instead of the near-zero signal expected in the absence of significant amounts of PN_s or N₂O₅ we observe a monotonic increase in NO₂ with the temperature which is a factor of ~ 2 steeper in TDI-1 than in TDI-2. The signal in TDI-1 at 475 K, where only PN_s are expected to dissociate, is ~50 % of the maximum signal at 650 K.

305 There are several potential explanations for this behaviour which include: 1) the formation and detection of thermally less stable AN_s (including e.g. di-nitrates), which dissociate at lower temperatures than e.g. iPN, 2) the formation of non-acyl, isoprene-derived peroxy-nitrates (RO₂NO₂) that are sufficiently long-lived to build up to appreciable concentrations in the SCHARK, 3) chemical processes taking place in the TD-inlets that convert ISOP-NIT to NO₂. Scenario 1) appears unlikely as several studies have shown that the ~~C-N~~ O-N bond-strength in various alkyl-nitrates is very similar (Hao et al., 1994; Wild et al., 2014). We also note that the formation of di-nitrates (in the absence of NO) only takes place when isoprene levels are very low and the first-generation nitrates formed in the NO₃ + isoprene reaction can react with a further NO₃. This can be ruled out for the present experiments in which the isoprene mixing ratio is always much larger than that of the first generation nitrates formed, which in any case react much more slowly with NO₃ than does isoprene. The second explanation requires that RO₂ formed in the initial reaction between NO₃ and isoprene react with NO₂ to form RO₂NO₂. Given our experimental conditions, 315 we would indeed expect that the main fate of any RO₂ formed in the reaction between NO₃ and isoprene is reaction with NO₂, which will dominate over self-reaction or reaction with NO₃, other RO₂ or HO₂. Non-acyl RO₂NO₂ are however generally highly thermally unstable, with lifetimes (at room temperature) of seconds or minutes, with respect to re-dissociation to RO₂ + NO₂.

For isoprene-derived RO₂NO₂ to contribute to the signal observed in region II would require that the RO₂ – NO₂ bond strength 320 be comparable to those of acyl nitrates such as PAN. The dominant 1,4-peroxy radical formed when NO₃ reacts with isoprene

which has a nitrate group separated by two carbon atoms from the peroxy carbon. It seems unlikely that this could have a stabilising effect on the C-N O-N bond in RO_2NO_2 in the same way that an α -carbonyl group does. Indeed, chamber experiments investigating the products of the $\text{NO}_3 +$ isoprene reaction in detail (Barnes et al., 1990; Wu et al., 2020) have failed to identify neither acyl- nor non-acyl- RO_2NO_2 as stable or semi-stable products formed from primary oxidation.

325 In support of scenario 3, sections 3.2 to 3.4 describe the evidence for chemical reactions leading to NO_2 formation that bypass the thermodynamic barrier for direct NO_2 formation but are surface-catalysed, require the presence of O_3 in either the SCHARK or in the inlet. These processes are peculiar to alkyl nitrates with a C=C double bond and thus have not been observed in TD-inlets tested only with saturated alkyl nitrates such as the frequently used isopropyl nitrate.

3.2 The role of O_3

330 To further investigate the conversion of ISOP-NITs to NO_2 at low temperatures in the TD-inlets, we generated NO_3 via the room-temperature thermal decomposition of N_2O_5 thus ruling out chemical processes that were initiated or catalysed by O_3 . In these experiments, N_2O_5 was transported into the SCHARK by passing a flow of 0.1 SLPM dry synthetic air over a crystalline N_2O_5 sample cooled to -78°C with further dilution in a 15 SLPM flow of zero-air. The combined concentration of $\text{N}_2\text{O}_5 + \text{NO}_3$ that remains after the reaction with in the presence of ~ 22 ppbv isoprene (7 sccm of 46.5 ppmv) was measured
335 as 40.5 pptv. The use of high isoprene concentrations guarantees that the thermal decomposition of N_2O_5 does not contribute significantly to the thermograms. Under these conditions several ppbv of ISOP-NIT were formed. A constant flow (2 SLPM) of zero air was passed over a low-pressure Hg-lamp and added between the sampling port of the SCHARK and the inlets and TD-inlets of the NO_2 , PNs and ANs cavities. This way the O_3 mixing ratio in the TD-inlets could be varied without affecting the chemistry in the chamber.

340 Figure 6 presents the results of one such experiment in which ISOP-NIT was sampled from the SCHARK using TDI-1 and TDI-2, both initially held at 703 K with thermograms obtained by decreasing the temperature of both inlets in 25 K steps. At each temperature step (periods of 20 min), after recording the signal under O_3 free conditions (black squares), a low (40-54 ppbv, green triangles), medium (97-111 ppbv, blue triangles) and high (185-219 ppbv, orange circles) mixing ratio of O_3 was added in front of the TD-inlets. Before cooling to the next temperature, the signal without O_3 was measured again and agreed
345 within 30-150 pptv to the value at the beginning of the corresponding period. To enable comparison with “ideal” behaviour, the thermogram of isopropyl nitrate (iPN, red circles) recorded from an experiment while flowing 162 ppbv O_3 (and 3 ppbv NO_2 impurity) through the SCHARK is also plotted.

Figure 6a displays a thermogram obtained when sampling ISOP-NITs from the SCHARK via TDI-1 (glass beads). In the presence of O_3 , the thermograms are very broad with substantial NO_2 formation between 350 and 475 K (shaded region II)
350 and in this sense are comparable to those displayed in Fig. 3, where NO_3 was obtained from the reaction of NO_2 with O_3 . In region I, the effect of going from ~ 50 to ~ 200 ppb of ozone is to increase the NO_2 generated drastically. This is the opposite of that observed when sampling via TDI-2 and thus in agreement with the results of the experiment depicted in Fig. 3, where the increase in signal was assigned to detection of HNO_3 . For the experiments in which NO_3 was generated in the room-

temperature thermal dissociation of N_2O_5 , HNO_3 can arise from reactions of N_2O_5 with moisture on the walls and is present as impurity in the N_2O_5 sample. As described above, the presence of glass beads has two effects which operate in the same direction in this temperature regime: The conversion of HNO_3 to NO_2 is catalysed by the surface provided by the glass-beads and at the same time the loss of NO_2 (via reaction with $\text{O}(^3\text{P})$) is reduced as $\text{O}(^3\text{P})$ is scavenged by the glass surface.

For TDI-1, the thermograms obtained without ozone (black squares) differ greatly to those in which ozone was present. Without ozone, NO_2 is not generated at temperatures lower than 550 K but its concentration increases rapidly at temperatures above ~ 600 K with no indication of a plateau being reached. The thermograms obtained in this NO_3 -isoprene system using TDI-1 in the absence of ozone bears little resemblance to that of iPN. Furthermore, during periods without ozone or heat in TDI-1, compounds of lower volatility like ISOP-NITs or HNO_3 appear to deposit on the glass beads and frit. This would form an explanation for a “memory effect” observed for TDI-1, whereby after exposure to HNO_3 or organic nitrates during unheated periods, an increase in the NO_2 signal followed by a slow decrease taking several hours is observed as soon as pure synthetic air and ozone were added to the flow through the heated inlet. An example of this phenomenon is shown in Fig. S4 in which (at peak signal) 60 ppbv of NO_2 was detected just by heating TDI-1 to 703 K in the presence of O_3 in synthetic air. In order to avoid bias in results caused by this effect, a cleaning procedure was adopted prior to all experiments whereby the inlet was heated to 703 K and exposed to ozone in synthetic air until a constant, low residual signal, usually between 20 and 200 pptv, was established. This memory effect seen for TDI-1 is also observed when a thermogram of ISOP-NIT (generated in a system similar to the experiment in Fig. 3) is measured by going from cold to hot temperatures (Fig. S5).

In Fig. 6b we present the results of the same experiment using TDI-2. The non-zero signal (~ 250 pptv) at temperatures between ~ 320 and 450 K) results from instability in the baseline. In the absence of ozone, the organic nitrates generated in the NO_3 -initiated oxidation of isoprene follow a well-defined thermogram (black squares and solid line) between 475 and 650 K, which is very similar to the thermogram measured for iPN (red circles). The addition of ozone does not result in the formation of NO_2 in region II, but does induce NO_2 losses for temperatures above 650 K (region I). The fact that ISOP-NIT was not converted to NO_2 at temperatures < 475 K suggests that the observation of a large signal in Fig. 3a (region II, white squares) is linked to O_3 -induced chemistry in the SCHARK which will be discussed in more detail in section 3.3. The loss of NO_2 in region I increases with increasing amounts of O_3 with $\sim 35\%$ of the NO_2 formed at 625 K lost when O_3 was increased to ~ 200 ppbv. The same behaviour is observed in the thermogram of iPN which confirms that this process is independent of the nature of the nitrate but solely linked to thermal decomposition of O_3 and subsequent reactions of $\text{O}(^3\text{P})$ with NO_2 .

3.2.1 Thermal dissociation and gas-phase reactions of O_3 and $\text{O}(^3\text{P})$

An important clue to the underlying chemical process that lead to the conversion of ANs to NO_2 at temperatures lower than those required to break the C-N O-N bond is the fact that the thermogram of iPN (measured with TDI-1) is not significantly affected by the presence of 163 ppbv O_3 whereas thermograms of ISOP-NIT, the vast majority of which are unsaturated, are greatly broadened when O_3 is present.

It is well known that O(³P) reacts rapidly (via electrophilic addition) to C=C double bonds (Leonori et al., 2015) and we thus assessed the potential impact of NO₂ formation via reactions of O₃ or O(³P) (formed in the thermal dissociation of O₃ in the TDIs) with ISOP-NIT.

390 The concentration of O(³P) in the TDIs depends on the concentration and rate of thermal decomposition of O₃ and thus on the gas-temperature as well as its rate of recombination with O₂, reactions with O₃, NO₂, isoprene, isoprene nitrates and loss to the walls:



The contribution (in addition to the thermal dissociation of ISOP-NIT) to NO₂ formation via reactions (R20) and (R21) in TDI-2 were assessed via numerical simulation (FACSIMILE/CHEKMAT release H010 (Curtis and Sweetenham, 1987)). The rate coefficients for the most important reactions are listed in Tab. 1; the complete reaction scheme is listed in the supplementary information (S79). Reaction times in heated and unheated section of TDI-2 were calculated from the temperature, internal diameter of the quartz tube and the flow rate. The rate coefficients for the gas-phase reactions of isoprene (IUPAC, 2017) and 2-methyl-2-butene (Herron and Huie, 1973) with O(³P) and O₃ were used as surrogates for the reactions of ISOP-NIT for which data is not available. The temperature-dependent dissociation rate coefficient of n-propyl nitrate was used to account for NO₂ from the thermal dissociation of isoprene-derived nitrates (Morin and Bedjanian, 2017). Wall loss of O(³P) in the instrument was estimated to be 90 s⁻¹ using the method as described in Thieser et al. (2016) and implemented in
410 the model run.

The initial conditions for the simulation were 1 ppbv ISOP-NIT, 10 ppbv isoprene and 5 ppbv NO₂ at a cavity pressure of ~~543 Torr~~ 724 hPa and a temperature of 298 K. The results obtained are shown by black curves in Fig.7. For temperatures up to 575 K the simulated thermograms with and without O₃ are almost identical, whereas at higher temperatures, the amount of NO₂ exiting the inlet decreases because of its reaction with O(³P) (R18), in broad agreement with the experiments carried out using
415 TDI-2 as shown in Fig. 3b. The model simulations show (Fig. 7), that almost no O(³P) is formed by the thermal dissociation of O₃ at the lower temperatures of region II. Only at higher temperatures, is a significant fraction of O₃ (27 % at 703 K) converted to O(³P) with the majority subsequently lost at the inlet walls. These calculations underline the observation that the

low temperature formation of NO₂ from ISOP-NIT seen when using TDI-1 cannot be explained with known gas-phase chemistry.

420 In summary, the experimental observations and the numerical simulations indicate that the presence of O₃ is required in the inlet for TDI-1 and in the chamber for TDI-2 to generate NO₂ from isoprene-derived nitrates at temperatures less than 475 K. We have shown that the generation of NO₂ from alkyl-nitrates at low temperatures using TDI-1 requires that the organic nitrate has a double-bond and that, while gas-phase reactions of O(³P) are responsible for the loss of NO₂ at high temperatures, they are NOT responsible for the conversion of isoprene-derived nitrates at lower temperatures in neither of the TD-inlets. The
425 presence of glass beads (large surface area) favours the formation of NO₂ from ISOP-NIT at low temperatures. Altogether, these observations indicate that a surface-catalysed reaction involving ozone is the process most likely to be responsible for the conversion of ISOP-NIT to NO₂ at temperatures below those required for the gas-phase thermolysis.

3.2.3 Surface catalysed reactions with ozone

Quartz tubes contain impurities and surface defects that can provide reactive sites (RS) at elevated temperatures and the surface
430 catalysed chemistry of ozone on e.g. mineral silicates is well known (Bulanin et al., 1994; Hanisch and Crowley, 2003a,b; Usher et al., 2003). O₃ can be surface-catalytically converted to O₂ (R25) by the formation and loss of reactive, oxygenated surface sites (RSO) via (R23) and (R24).



In order to test for ozone loss in our setup, O₃ in synthetic air was passed through the TDIs. The O₃ mixing ratio prior to entering the inlets was measured continuously using the ozone monitor (UV absorption). The ozone exiting the TD-inlets was converted to NO₂ (by addition of 1 ppmv NO (R4) in a 1.5 m long PFA tubing with 0.5 inch OD and a residence time of 5.2
440 s) and then measured in the 409 nm cavities. Figure 8a shows the results of such an experiment for TDI-1. The concentration of ozone before entering the inlets was 20.5 ppbv (open dots) which was detected as 17.1 ppbv after passing through the inlet at 298 K (black squares, 11:15 to 11:55). This represents a conversion efficiency of 0.83 which matches **exactly** that expected when considering the reaction time, NO concentration and the rate coefficient ($k_4 = 1.9 \times 10^{-14} \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$, IUPAC, 2021).

445 At \approx 12:00 LT, upon heating the inlet to 473 K the ozone exiting TDI-1 was depleted by up to 27 %, while heating to 573 K results in further loss (up to 40 %). An analogous experiment using TDI-2 (Fig. 8b) showed that while some O₃ was also lost (e.g. 4 % at 573 K) this was much less than in TDI-1. As the gas-phase thermal decomposition of O₃ is negligible under these conditions (0.6 % at 575 K, Fig. 7) the loss of O₃ when passing through the TD-inlets indicates that surface catalysed ozone decomposition takes place (R23 and R24), especially in TDI-1 where larger surface areas are available.

450 We now consider the possibility that the conversion of isoprene-derive nitrates to NO₂ can be catalysed by surfaces in the presence of O₃. We note that previous work has shown that the heterogeneous ozonolysis of alkenes on glass or other surfaces can be more efficient than its analogous, gas-phase process (Dubowski et al., 2004; Stokes et al., 2008; Ray et al., 2013) and now consider the possibility that RSO is the mediating reactive species in TDI-1 in our experiments.



455 (R26)

We first examine the possible contributors to ISOP-NIT formed in the reaction of NO₃ with isoprene, in which the dominant initial step (in air) is a sequential 1,4 Addition of NO₃ and O₂ to form δ- and β-peroxy radicals e.g. O₂NOCH₂C(CH₃)=CHCH₂OO (Schwantes et al., 2015). In the presence of NO₃, RO₂ or HO₂, the peroxy radicals react further to form “first generation” isoprene nitrates which contain carbonyl, hydroperoxidic and alcoholic groups such as
460 O₂NOCH₂C(CH₃)=CHCHO, O₂NOCH₂C(CH₃)=CHCH₂OOH and O₂NOCH₂C(CH₃)=CHCH₂OH, respectively. Note that most of the known, first-generation organic nitrates retain a C=C double-bond.

A hypothetical ISOP-NIT degradation scheme involving the initial attachment of RS-O to the remaining double bond is given in Fig. 9. We consider only the fate of the most stable surface-adducts, i.e. tertiary radical in case of δ-products and secondary radicals in case of β-products. Both radical-adducts will react with O₂ to form organic peroxy radicals which may undergo H-
465 shifts (via five- or six-membered rings) resulting in formation of a radical with its unpaired electron in the direct vicinity of the nitrate functionality. Such unimolecular processes may become competitive to bimolecular reactions under atmospheric conditions (Møller et al., 2019). A possible fate of this radical is decomposition to form a carbonyl compound via NO₂ elimination (Hjorth et al., 1990; Berndt and Boge, 1995; Vereecken et al., 2021). For δ-products, the tertiary product may also eliminate NO₂ under the formation of an epoxide.

470 With TDI-2, the conversion of ISOP-NIT to NO₂ at low temperatures (region I) is only observed when O₃ is present in the chamber (Fig. 3a), but not when it is added only to the inlet (Fig. 6b) implying that the presence of O₃ as an oxidizing agent in the SCHARK is the main difference between these two reaction systems. This suggests that the ozonolysis of isoprene may play an important role. As the reaction between isoprene and O₃ leads to the formation of OH and additional HO₂ (Zhang et al., 2002; Malkin et al., 2010; Cox et al., 2020), the fraction of RO₂ reacting with HO₂ in this system is enhanced when
475 compared to the experiments in which NO₃ was generated from N₂O₅. The major product from the reaction between nitrated δ- and β-peroxy radicals with HO₂ are hydroperoxides, such as O₂NOCH₂C(CH₃)=CHCH₂OOH (Schwantes et al., 2015). With the intention of assessing the effect of O₃ on hydroperoxide yields, further model calculations were performed using the Framework for 0-D Atmospheric Modelling (F0AM) (Wolfe et al., 2016). The NO₃ isoprene oxidation scheme is still subject of current research and uncertain which is why the Master Chemical Mechanism (MCM, version 3.3.1,
480 <http://mcm.leeds.ac.uk/MCM>) (Jenkin et al., 2015) as well as the Reduced Caltech Isoprene Mechanism Plus (RCIMP, version 5, <https://data.caltech.edu/records/247>) (Wennberg et al., 2018; Bates and Jacob, 2019) were applied. Figure 10 shows the calculated fraction of hydroperoxides to ISOP-NIT using both models and mixing 22 ppbv isoprene with either 10.8 ppbv NO₂ and 150 ppbv O₃ (runs 1 and 3) or 3 ppbv N₂O₅ (runs 2 and 4) to achieve NO₃ in a flow-through chamber with an exchange

rate of $2.73 \times 10^{-4} \text{ s}^{-1}$. Generally, the MCM predicts higher hydroperoxide yields than RCIMP. A possible reason might be that
485 MCM has higher rates of HO₂ reformation from reactions of isoprene-derived alkoxy radicals than RCIMP does. For that
reason, MCM only predicts an increase in the hydroperoxide yield by a factor of 1.33 when the NO₃ source is changed from
N₂O₅ to NO₂ + O₃ (model runs 1 and 2), whereas RCIMP predicts a factor of 2.56 (model runs 3 and 4). The calculations thus
broadly support the hypothesis that unsaturated, nitrated hydroperoxides are involved in the low temperature (surface catalysed)
formation of NO₂ from ISOP-NIT. **Indeed, hydroperoxides not only have high affinity for surfaces but also have a rather weak**
490 **O-O bond with a dissociation energy of $\approx 45 \text{ kcal/mol}$ (Bach and Schlegel, 2020)** and several routes to surface catalysed
elimination of NO₂ appear feasible. Decomposition of isoprene-derived hydroperoxides within the sampling line has been
observed for other instruments (Rivera-Rios et al., 2014). A possible degradation mechanism for an isoprene-derived
hydroperoxide which is abundant from the NO₃ + isoprene system (Schwantes et al., 2015) is depicted in Fig.11. In this case,
O₂NOCH₂C(CH₃)=CHCH₂OOH coordinates via hydrogen bonds to the SiO₂ surface prior to cleavage of the O-O bond to form
495 OH and an alkoxy radical. The latter may dissociate via unimolecular reactions to form closed-shell products under elimination
of NO₂ (Wennberg et al., 2018; Vereecken et al., 2021).

3.4 Elimination of O₃ and surface catalysed conversion of ANs at low temperatures

Our findings clearly show that that the combination of heated quartz surfaces and ozone catalyse the decomposition of the
unsaturated nitrates formed in the reaction between NO₃ and isoprene at temperatures below 473 K. While the exemplary,
500 surface catalysed processes depicted in Fig.9 and Fig.11 fulfil the requirement of conversion of ISOP-NIT to NO₂ we stress
that these are purely hypothetical and we cannot state with any certainty that they are the reactions responsible for our
observations. While it would be highly interesting to investigate such processes using different (e.g. surface sensitive)
techniques this is clearly beyond the scope of this paper and of the experimental capabilities of this research project. Instead,
we take the more pragmatic approach and indicate potential methods to eliminate such unwanted reactions when using TD-
505 inlets.

In principal, as the process that convert ISOP-NIT to NO₂ are clearly surface catalysed (involving formation of RSO) their
impact can be reduced by using a surface that does not support formation of RSO. We therefore tested a third TD-inlet (TDI-
3), consisting of 55 cm PFA tubing (0.375 in. OD) with a 20 cm heated section. PFA-tubing has been routinely used as TDI
for measurement of e.g. PAN, as it is relatively unreactive to the peroxy radicals formed (Phillips et al., 2013). The C-F-bond
510 of PFA is very strong and nonpolar which should reduce the formation of RSO as well as adsorption of ISOP-NIT to the
surface. The performance of a TDI made of PFA (TDI-3) was examined by performing an experiment with the SCHARK
analogous to the one shown in Fig. 2. The resulting time-series of NO₂, ΣPNs (using TDI-3 heated to 448 K) and ΣANs (using
TDI-2 heated to 648 K) are depicted in Fig.12. At first, O₃ (5 SLPM synthetic air passed over a Hg lamp) and NO₂ (200 sccm
of 1 ppmv) in 25 SLPM dry synthetic air were constantly introduced into the SCHARK. After detectable amounts of NO₃ and
515 N₂O₅ had been formed, isoprene (9.8 sccm of 46.5 ppmv) was added leading (as expected) to almost quantitative depletion of
NO₃ and N₂O₅.

During the following 3 hours the signal in the Σ ANs cavity (TDI-2) increased to ~ 1.2 ppbv, while the signal in the Σ PNs channel (~ 40 pptv) was close to the detection limit. A CIMS measurement obtained during this experiment validates that PAN mixing ratios are lower than 50 pptv. This result already confirms that 1) the ANs derived from NO_3 + isoprene reaction are not detected at $T < 448$ K when PFA is used instead of quartz and 2) as expected, there is no significant generation of peroxy nitrates in these experiments which would be detectable with both ovens at $T = 448$ K (see Fig. S6 in the Supplement).

This result not only confirms that the previous detection of ISOP-NITs at low TD-inlet temperatures when using TDI-2 and especially TDI-1 were caused by heterogeneous reactions at the quartz surface, but also provides a solution to the problem of the separate measurement of Σ ANs and Σ PNs using TD-inlets.

We recall however, that the reason for adding glass beads to the inlet was to suppress recombination reactions of NO_2 with peroxy radicals by providing a surface to scavenge the latter. The use of a PFA tube rather than quartz will certainly exacerbate this effect, as the rate of loss of RO_2 to PFA surfaces is expected to be lower than on quartz (Wooldridge et al., 2010). This implies that corrective procedures based on numerical simulation may be necessary in some environments as shown by Thieser et al. (2016) and this may limit the useful deployment of the method to regions where NO_x levels are sufficiently low that reaction of RO_2 with NO and NO_2 become insignificant. **In addition, similar to the observations of Sobanski et al. (2016), ClNO_2 can interfere with the detection of ANs. The experiment shown in Fig. S7 reveals that ClNO_2 is detected with TDI-2 at 698 K, but not with TDI-3 at 448 K.**

4 Summary, Conclusions and implications for atmospheric measurements of Σ PNs and Σ ANs

We have shown that the detection of isoprene-derived organic nitrates via its thermal dissociation in quartz / glass inlets can (in the presence of O_3) be accompanied by undesirable side reactions which broaden the thermograms and thus impede the separation of PN and AN signals by sampling through TD-inlets at different temperatures as is commonly practised. While our experiments deal with the nitrates formed in the NO_3 -initiated oxidation of isoprene, it is very likely that similar broadening of thermograms would also occur with nitrates formed from the oxidation of other terpenoids, as some organic nitrates derived from nighttime oxidation of e.g. limonene still contain a double bond (Fry et al., 2011) and/or contain hydroperoxy groups.

Specifically, we find that the presence of O_3 in either the quartz TD-inlet or in a Teflon simulation chamber results in the generation of NO_2 from isoprene-derived nitrates in TD-inlets made of quartz at temperatures less than 475 K **and that this only occurs when the organic nitrate either has a double-bond or a hydroperoxy group (or both)**. The formation of NO_2 from ISOP-NIT was accelerated in the presence of glass-beads which indicates that a surface-catalysed reaction involving ozone and reactive surface sites is the process most likely to be responsible both for the conversion of ISOP-NIT to NO_2 at low temperatures (375-475 K) and the conversion of HNO_3 to NO_2 at high temperatures (> 550 K). By avoiding the use of O_3 or using a non-quartz TD-inlet, we were able to show that the ISOP-NIT thermogram is entirely consistent with those of saturated alkyl nitrates.

We show that surface-catalysed reactions on quartz TD-inlets involving O_3 represent a potential source of bias in measurements of ΣANs and ΣPNs during field observations, especially when isoprene is abundant. For example, we previously reported results from two campaigns carried out using the TD-CRDS system described here on the same rural mountain site (Kleiner Feldberg) and season (but in different years). We found that the average, relative abundance of ΣPNs and ΣANs was quite different with $\Sigma ANs \approx \Sigma PNs$ during the PARADE campaign in 2011 and $\Sigma PNs > \Sigma ANs$ during the NOTOMO campaign in 2015 (Sobanski et al., 2017). In 2011, the TD-inlet was a quartz tube (i.e. similar to TDI-2) (Thieser et al., 2016), whereas in 2015 the inlet contained glass beads (i.e. similar to TDI-1) (Sobanski et al., 2016). With our present understanding of the role of surfaces and O_3 in TD-inlets, we cannot rule out that the observations during 2015 were biased to lower values for ΣANs and higher values for ΣPNs , although, as discussed by (Sobanski et al., 2017) there are other, meteorological factors which would have contributed.

For the detection of PNs we have shown that (at the lower temperature required to thermally dissociate PNs to NO_2) surface catalytic effects that convert ANs (or other species) to NO_2 can be completely avoided by using a TD-inlet made of a non-reactive material like PFA (TDI-3). In this case, O_3 does not appear to have any impact.

When using a quartz TD-inlet for conversion of ANs + PNs to NO_2 at higher temperatures the surface reactions that shift thermograms to lower temperatures are of less significance as, in any case, the role of the TD-inlet is to convert all ANs and all PNs to NO_2 . However, in order to avoid detection of HNO_3 or HONO other materials may be more suitable than quartz. Sapphire, commonly used in microwave discharge generated plasmas owing to its high purity and non-reactive surface, may represent a useful alternative. Under humid conditions, some of the observed interferences become negligible: No HNO_3 appears to interfere with the ANs measurement and the interference of ANs to the PNs measurement is reduced in TDI-2, but not in TDI-1.

This study emphasizes the importance of characterising thermal dissociation inlets under conditions which are similar to those found in the atmosphere. We recognise that the impact of surface catalytic processes will vary from one inlet to the next (even if made from the same material) and all quartz-TDIs will not necessarily exhibit the same degree of conversion of BVOC-derived ANs at low temperature. For this reason, thermograms should be measured using trace-gases that are abundant in the atmosphere and the effect of e.g. O_3 and water vapour should be thoroughly investigated.

Data availability. The data underlying the figures is available on request from the corresponding author.

575

Author contributions. PD conducted the experiments, analysed the data and wrote the manuscript. RD was responsible for the CIMS measurements. JS designed and built the SCHARK. JNC designed the experiments and together with JL contributed to the manuscript.

580 *Competing interests.* The authors declare that they have no conflict of interest.

References

- Bach, R. D., and Schlegel, H. B.: Bond Dissociation Energy of Peroxides Revisited, *J. Phys. Chem. A*, **124**, 4742-4751, doi:10.1021/acs.jpca.0c02859, 2020.
- 585 Barnes, I., Bastian, V., Becker, K. H., and Tong, Z.: Kinetics and products of the reactions of NO₃ with monoalkenes, dialkenes, and monoterpenes, *J. Phys. Chem.*, **94**, 2413-2419, doi:10.1021/j100369a041, 1990.
- Bates, K. H., and Jacob, D. J.: A new model mechanism for atmospheric oxidation of isoprene: global effects on oxidants, nitrogen oxides, organic products, and secondary organic aerosol, *Atmos. Chem. Phys.*, **19**, 9613-9640, doi:10.5194/acp-19-9613-2019, 2019.
- 590 Beaver, M. R., St Clair, J. M., Paulot, F., Spencer, K. M., Crouse, J. D., LaFranchi, B. W., Min, K. E., Pusede, S. E., Wooldridge, P. J., Schade, G. W., Park, C., Cohen, R. C., and Wennberg, P. O.: Importance of biogenic precursors to the budget of organic nitrates: observations of multifunctional organic nitrates by CIMS and TD-LIF during BEARPEX 2009, *Atmos. Chem. Phys.*, **12**, 5773-5785, doi:10.5194/acp-12-5773-2012, 2012.
- Berndt, T., and Boge, O.: Products and Mechanism of the Reaction of NO₃ with Selected Acyclic Monoalkenes, *J. Atmos. Chem.*, **21**, 275-291, doi:10.1007/Bf00696759, 1995.
- 595 Berndt, T., and Boge, O.: Gas-phase reaction of NO₃ radicals with isoprene: A kinetic and mechanistic study, *Int. J. Chem. Kinet.*, **29**, 755-765, doi:10.1002/(sici)1097-4601(1997)29:10<755::aid-kin4>3.0.co;2-1, 1997.
- 600 Brownwood, B., Turdziladze, A., Hohaus, T., Wu, R., Mentel, T. F., Carlsson, P. T. M., Tsiligiannis, E., Hallquist, M., Andres, S., Hantschke, L., Reimer, D., Rohrer, F., Tillmann, R., Winter, B., Liebmann, J., Brown, S. S., Kiendler-Scharr, A., Novelli, A., Fuchs, H., and Fry, J. L.: Gas-Particle Partitioning and SOA Yields of Organonitrate Products from NO₃-Initiated Oxidation of Isoprene under Varied Chemical Regimes, *ACS Earth Space Chem.*, **5**, 785-800, doi:10.1021/acsearthspacechem.0c00311, 2021.
- Bulanin, K. M., Alexeev, A. V., Bystrov, D. S., Lavalley, J. C., and Tsyganenko, A. A.: IR Study of Ozone Adsorption on SiO₂, *J. Phys. Chem.*, **98**, 5100-5103, doi:10.1021/j100070a026, 1994.
- 605 Burkholder, J. B., Sander, S. P., Abbatt, J., Barker, J. R., Huie, R. E., Kolb, C. E., Kurylo, M. J., Orkin, V. L., Wilmouth, D. M., and Wine, P. H.: Chemical Kinetics and Photochemical Data for Use in Atmospheric Studies, Evaluation No. 18, "JPL Publication 15-10, Jet Propulsion Laboratory, Pasadena, <http://jpldataeval.jpl.nasa.gov>, 2016.
- Carlton, A. G., Wiedinmyer, C., and Kroll, J. H.: A review of Secondary Organic Aerosol (SOA) formation from isoprene, *Atmos. Chem. Phys.*, **9**, 4987-5005, doi:10.5194/acp-9-4987-2009, 2009.
- 610 Cox, R. A., Ammann, M., Crowley, J. N., Herrmann, H., Jenkin, M. E., McNeill, V. F., Mellouki, A., Troe, J., and Wallington, T. J.: Evaluated kinetic and photochemical data for atmospheric chemistry: Volume VII - Criegee intermediates, *Atmos. Chem. Phys.*, **20**, 13497-13519, doi:10.5194/acp-20-13497-2020, 2020.
- Crutzen, P. J., and Lelieveld, J.: Human impacts on atmospheric chemistry, *Annu. Rev. Earth Planet. Sci.*, **29**, 17-45, doi:10.1146/annurev.earth.29.1.17, 2001.
- Curtis, A. R., and Sweetenham, W. P.: Facsimile, Atomic Energy Research Establishment, Report R-12805, Harwell Laboratory, Oxfordshire, UK, 1987.
- 615 Davidson, J. A., Viggiano, A. A., Howard, C. J., Dotan, I., Fehsenfeld, F. C., Albritton, D. L., and Ferguson, E. E.: Rate Constants for Reactions of O₂⁺, NO₂⁺, NO⁺, H₃O⁺, CO₃⁺, NO₂⁺ and Halide Ions with N₂O₅ at 300 K, *J. Chem. Phys.*, **68**, 2085-2087, doi:10.1063/1.436032, 1978.
- 620 Day, D. A., Wooldridge, P. J., Dillon, M. B., Thornton, J. A., and Cohen, R. C.: A thermal dissociation laser-induced fluorescence instrument for in situ detection of NO₂, peroxy nitrates, alkyl nitrates, and HNO₃, *J. Geophys. Res.-Atmos.*, **107**, doi:10.1029/2001jd000779, 2002.

- Dewald, P., Liebmann, J. M., Friedrich, N., Shenolikar, J., Schuladen, J., Rohrer, F., Reimer, D., Tillmann, R., Novelli, A., Cho, C. M., Xu, K. M., Holzinger, R., Bernard, F., Zhou, L., Mellouki, W., Brown, S. S., Fuchs, H., Lelieveld, J., and Crowley, J. N.: Evolution of NO₃ reactivity during the oxidation of isoprene, *Atmos. Chem. Phys.*, 20, 10459-10475, doi:10.5194/acp-20-10459-2020, 2020.
- 625 Dubowski, Y., Vieceli, J., Tobias, D. J., Gomez, A., Lin, A., Nizkorodov, S. A., McIntire, T. M., and Finlayson-Pitts, B. J.: Interaction of gas-phase ozone at 296 K with unsaturated self-assembled monolayers: A new look at an old system, *J. Phys. Chem. A*, 108, 10473-10485, doi:10.1021/jp046604x, 2004.
- Eger, P. G., Helleis, F., Schuster, G., Phillips, G. J., Lelieveld, J., and Crowley, J. N.: Chemical ionization quadrupole mass spectrometer with an electrical discharge ion source for atmospheric trace gas measurement, *Atmos. Meas. Tech.*, 12, 1935-1954, doi:10.5194/amt-12-1935-2019, 2019.
- 630 Finlayson-Pitts, B. J., Wingen, L. M., Sumner, A. L., Syomin, D., and Ramazan, K. A.: The heterogeneous hydrolysis of NO₂ in laboratory systems and in outdoor and indoor atmospheres: An integrated mechanism, *Phys. Chem. Chem. Phys.*, 5, 223-242, doi:10.1039/B208564J, 2003.
- Friedrich, N., Tadic, I., Schuladen, J., Brooks, J., Darbyshire, E., Drewnick, F., Fischer, H., Lelieveld, J., and Crowley, J. N.: Measurement of NO_x and NO_y with a thermal dissociation cavity ring-down spectrometer (TD-CRDS): instrument characterisation and first
635 deployment, *Atmos. Meas. Tech.*, 13, 5739-5761, doi:10.5194/amt-13-5739-2020, 2020.
- Fry, J. L., Kiendler-Scharr, A., Rollins, A. W., Brauers, T., Brown, S. S., Dorn, H. P., Dube, W. P., Fuchs, H., Mensah, A., Rohrer, F., Tillmann, R., Wahner, A., Wooldridge, P. J., and Cohen, R. C.: SOA from limonene: role of NO₃ in its generation and degradation, *Atmos. Chem. Phys.*, 11, 3879-3894, doi:10.5194/acp-11-3879-2011, 2011.
- Fry, J. L., Brown, S. S., Middlebrook, A. M., Edwards, P. M., Campuzano-Jost, P., Day, D. A., Jimenez, J. L., Allen, H. M., Ryerson, T.
640 B., Pollack, I., Graus, M., Warneke, C., de Gouw, J. A., Brock, C. A., Gilman, J., Lerner, B. M., Dubé, W. P., Liao, J., and Welti, A.: Secondary organic aerosol (SOA) yields from NO₃ radical + isoprene based on nighttime aircraft power plant plume transects, *Atmos. Chem. Phys.*, 18, 11663-11682, doi:10.5194/acp-18-11663-2018, 2018.
- Guenther, A. B., Jiang, X., Heald, C. L., Sakulyanontvittaya, T., Duhl, T., Emmons, L. K., and Wang, X.: The Model of Emissions of Gases and Aerosols from Nature version 2.1 (MEGAN2.1): an extended and updated framework for modeling biogenic emissions, *Geosci. Model. Dev.*, 5, 1471-1492, doi:10.5194/gmd-5-1471-2012, 2012.
645
- Hamilton, J. F., Bryant, D. J., Edwards, P. M., Ouyang, B., Bannan, T. J., Mehra, A., Mayhew, A. W., Hopkins, J. R., Dunmore, R. E., Squires, F. A., Lee, J. D., Newland, M. J., Worrall, S. D., Bacak, A., Coe, H., Percival, C., Whalley, L. K., Heard, D. E., Slater, E. J., Jones, R. L., Cui, T. Q., Surratt, J. D., Reeves, C. E., Mills, G. P., Grimmond, S., Sun, Y. L., Xu, W. Q., Shi, Z. B., and Rickard, A. R.:
650 Key Role of NO₃ Radicals in the Production of Isoprene Nitrates and Nitrooxyorganosulfates in Beijing, *Environ. Sci. Technol.*, 55, 842-853, doi:10.1021/acs.est.0c05689, 2021.
- Hanisch, F., and Crowley, J. N.: Ozone destruction on Saharan dust: An experimental investigation, *Atmos. Chem. Phys.*, 3, 119-130, doi:10.5194/acp-3-119-2003, 2003a.
- Hanisch, F., and Crowley, J. N.: Heterogeneous reactivity of NO and HNO₃ on mineral dust in the presence of ozone, *Phys. Chem. Chem. Phys.*, 5, 883-887, doi:10.1039/B211503D, 2003b.
- 655 Hao, C. S., Shepson, P. B., Drummond, J. W., and Muthuramu, K.: Gas-chromatographic detector for selective and sensitive detection of atmospheric organic nitrates, *Anal. Chem.*, 66, 3737-3743, doi:10.1021/ac00093a032, 1994.
- Herron, J. T., and Huie, R. E.: Rate Constants for the Reactions of Atomic Oxygen (O³P) with Organic Compounds in the Gas Phase, *J. Phys. Chem. Ref. Data*, 2, 467-518, doi:10.1063/1.3253125, 1973.
- 660 Hjorth, J., Lohse, C., Nielsen, C. J., Skov, H., and Restelli, G.: Products and Mechanisms of the Gas-Phase Reactions between NO₃ and a Series of Alkenes, *J. Phys. Chem.*, 94, 7494-7500, doi:10.1021/j100382a035, 1990.

- Horowitz, L. W., Fiore, A. M., Milly, G. P., Cohen, R. C., Perring, A., Wooldridge, P. J., Hess, P. G., Emmons, L. K., and Lamarque, J.-F.: Observational constraints on the chemistry of isoprene nitrates over the eastern United States, *J. Geophys. Res.*, 112, D12S08-, doi: 10.1029/2006JD007747, 2007.
- 665 IUPAC: Task Group on Atmospheric Chemical Kinetic Data Evaluation, edited by: Ammann, M., Cox, R.A., Crowley, J.N., Herrmann, H., Jenkin, M.E., McNeill, V.F., Mellouki, A., Rossi, M. J., Troe, J. and Wallington, T. J., available at: <http://iupac.pole-ether.fr/index.html>, last access: 23 April 2021.
- Jenkin, M. E., Young, J. C., and Rickard, A. R.: The MCM v3.3.1 degradation scheme for isoprene, *Atmos. Chem. Phys.*, 15, 11433-11459, doi:10.5194/acp-15-11433-2015, 2015.
- 670 Keehan, N. I., Brownwood, B., Marsavin, A., Day, D. A., and Fry, J. L.: A thermal-dissociation-cavity ring-down spectrometer (TD-CRDS) for the detection of organic nitrates in gas and particle phases, *Atmos. Meas. Tech.*, 13, 6255-6269, doi:10.5194/amt-13-6255-2020, 2020.
- Kirchner, F., Mayer-Figge, A., Zabel, F., and Becker, K. H.: Thermal stability of peroxy nitrates, *Int. J. Chem. Kinet.*, 31, 127-144, doi: 10.1002/(SICI)1097-4601(1999)31:2<127::AID-KIN6>3.0.CO;2-L, 1999.
- 675 Kwan, A. J., Chan, A. W. H., Ng, N. L., Kjaergaard, H. G., Seinfeld, J. H., and Wennberg, P. O.: Peroxy radical chemistry and OH radical production during the NO₃-initiated oxidation of isoprene, *Atmos. Chem. Phys.*, 12, 7499-7515, doi:10.5194/acp-12-7499-2012, 2012.
- Kwok, E. S. C., Aschmann, S. M., Arey, J., and Atkinson, R.: Product formation from the reaction of the NO₃ radical with isoprene and rate constants for the reactions of methacrolein and methyl vinyl ketone with the NO₃ radical, *Int. J. Chem. Kinet.*, 28, 925-934, doi:10.1002/(SICI)1097-4601(1996)28:12<925::AID-KIN10>3.0.CO;2-B, 1996.
- 680 Lelieveld, J., Evans, J. S., Fnais, M., Giannadaki, D., and Pozzer, A.: The contribution of outdoor air pollution sources to premature mortality on a global scale, *Nature*, 525, 367-371, doi:10.1038/nature15371, 2015.
- Leonori, F., Balucani, N., Nevrlý, V., Bergeat, A., Falcinelli, S., Vanuzzo, G., Casavecchia, P., and Cavallotti, C.: Experimental and Theoretical Studies on the Dynamics of the O(³P) + Propene Reaction: Primary Products, Branching Ratios, and Role of Intersystem Crossing, *J Phys Chem C*, 119, 14632-14652, doi:10.1021/jp512670y, 2015.
- 685 Liebmann, J. M., Schuster, G., Schuladen, J. B., Sobanski, N., Lelieveld, J., and Crowley, J. N.: Measurement of ambient NO₃ reactivity: Design, characterization and first deployment of a new instrument, *Atmos. Meas. Tech.*, 2017, 1241-1258, doi:10.5194/amt-2016-381, 2017.
- Malkin, T. L., Goddard, A., Heard, D. E., and Seakins, P. W.: Measurements of OH and HO₂ yields from the gas phase ozonolysis of isoprene, *Atmos. Chem. Phys.*, 10, 1441-1459, doi:10.5194/acp-10-1441-2010, 2010.
- 690 Mellouki, A., Ammann, M., Cox, R. A., Crowley, J. N., Herrmann, H., Jenkin, M. E., McNeill, V. F., Troe, J., and Wallington, T. J.: Evaluated kinetic and photochemical data for atmospheric chemistry: Volume VIII - gas phase reactions of organic species with four, or more, carbon atoms (≥ C₄), *Atmos. Chem. Phys. Discuss.*, 2020, 1-22, doi:10.5194/acp-2020-940, 2020.
- Møller, K. H., Bates, K. H., and Kjaergaard, H. G.: The Importance of Peroxy Radical Hydrogen-Shift Reactions in Atmospheric Isoprene Oxidation, *J. Phys. Chem. A*, 123, 920-932, doi:10.1021/acs.jpca.8b10432, 2019.
- 695 Morin, J., and Bedjanian, Y.: Thermal decomposition of n-propyl and n-butyl nitrates: Kinetics and products, *J. Anal. Appl. Pyrolysis*, 124, 576-583, doi:10.1016/j.jaap.2017.01.014, 2017.
- Ng, N. L., Kwan, A. J., Surratt, J. D., Chan, A. W. H., Chhabra, P. S., Sorooshian, A., Pye, H. O. T., Crounse, J. D., Wennberg, P. O., Flagan, R. C., and Seinfeld, J. H.: Secondary organic aerosol (SOA) formation from reaction of isoprene with nitrate radicals (NO₃), *Atmos. Chem. Phys.*, 8, 4117-4140, doi:10.5194/acp-8-4117-2008, 2008.

- 700 Ng, N. L., Brown, S. S., Archibald, A. T., Atlas, E., Cohen, R. C., Crowley, J. N., Day, D. A., Donahue, N. M., Fry, J. L., Fuchs, H., Griffin, R. J., Guzman, M. I., Herrmann, H., Hodzic, A., Iinuma, Y., Jimenez, J. L., Kiendler-Scharr, A., Lee, B. H., Luecken, D. J., Mao, J., McLaren, R., Mutzel, A., Osthoff, H. D., Ouyang, B., Picquet-Varrault, B., Platt, U., Pye, H. O. T., Rudich, Y., Schwantes, R. H., Shiraiwa, M., Stutz, J., Thornton, J. A., Tilgner, A., Williams, B. J., and Zaveri, R. A.: Nitrate radicals and biogenic volatile organic compounds: oxidation, mechanisms, and organic aerosol, *Atmos. Chem. Phys.*, 17, 2103-2162, doi:10.5194/acp-17-2103-2017, 2017.
- 705 Nguyen, T. B., Tyndall, G. S., Crounse, J. D., Teng, A. P., Bates, K. H., Schwantes, R. H., Coggon, M. M., Zhang, L., Feiner, P., Miller, D. O., Skog, K. M., Rivera-Rios, J. C., Dorris, M., Olson, K. F., Koss, A., Wild, R. J., Brown, S. S., Goldstein, A. H., de Gouw, J. A., Brune, W. H., Keutsch, F. N., Seinfeld, J. H., and Wennberg, P. O.: Atmospheric fates of Criegee intermediates in the ozonolysis of isoprene, *Phys. Chem. Chem. Phys.*, 18, 10241-10254, doi:10.1039/c6cp00053c, 2016.
- Orphal, J., Fellows, C. E., and Flaud, P. M.: The visible absorption spectrum of NO₃ measured by high-resolution Fourier transform spectroscopy, *J. Geophys. Res.-Atmos.*, 108, Art. Nr. 4077, doi:10.1029/2002JD002489, 2003.
- 710 Paul, D., Furgeson, A., and Osthoff, H. D.: Measurements of total peroxy and alkyl nitrate abundances in laboratory-generated gas samples by thermal dissociation cavity ring-down spectroscopy, *Rev. Sci. Instrum.*, 80, Art. 114101, doi:10.1063/1.3258204, 2009.
- Perring, A. E., Wisthaler, A., Graus, M., Wooldridge, P. J., Lockwood, A. L., Mielke, L. H., Shepson, P. B., Hansel, A., and Cohen, R. C.: A product study of the isoprene+NO₃ reaction, *Atmos. Chem. Phys.*, 9, 4945-4956, doi: 10.5194/acp-9-4945-2009, 2009.
- 715 Phillips, G. J., Pouvesle, N., Thieser, J., Schuster, G., Axinte, R., Fischer, H., Williams, J., Lelieveld, J., and Crowley, J. N.: Peroxyacetyl nitrate (PAN) and peroxyacetic acid (PAA) measurements by iodide chemical ionisation mass spectrometry: first analysis of results in the boreal forest and implications for the measurement of PAN fluxes, *Atmospheric Chemistry and Physics*, 13, 1129-1139, doi:10.5194/acp-13-1129-2013, 2013.
- Pitts, J. N., Sanhueza, E., Atkinson, R., Carter, W. P. L., Winer, A. M., Harris, G. W., and Plum, C. N.: An investigation of the dark formation of nitrous acid in environmental chambers, *Int. J. Chem. Kinet.*, 16, 919-939, doi:10.1002/kin.550160712, 1984.
- 720 Ray, D., Malongwe, J. K., and Klan, P.: Rate Acceleration of the Heterogeneous Reaction of Ozone with a Model Alkene at the Air-Ice Interface at Low Temperatures, *Environ. Sci. Technol.*, 47, 6773-6780, doi:10.1021/es304812t, 2013.
- Rivera-Rios, J. C., Nguyen, T. B., Crounse, J. D., Jud, W., St Clair, J. M., Mikoviny, T., Gilman, J. B., Lerner, B. M., Kaiser, J. B., de Gouw, J., Wisthaler, A., Hansel, A., Wennberg, P. O., Seinfeld, J. H., and Keutsch, F. N.: Conversion of hydroperoxides to carbonyls in field and laboratory instrumentation: Observational bias in diagnosing pristine versus anthropogenically controlled atmospheric chemistry, *Geophys. Res. Lett.*, 41, 8645-8651, doi:10.1002/2014gl061919, 2014.
- 725 Rollins, A. W., Kiendler-Scharr, A., Fry, J. L., Brauers, T., Brown, S. S., Dorn, H. P., Dubé, W. P., Fuchs, H., Mensah, A., Mentel, T. F., Rohrer, F., Tillmann, R., Wegener, R., Wooldridge, P. J., and Cohen, R. C.: Isoprene oxidation by nitrate radical: alkyl nitrate and secondary organic aerosol yields, *Atmos. Chem. Phys.*, 9, 6685-6703, doi:10.5194/acp-9-6685-2009, 2009.
- 730 Schwantes, R. H., Teng, A. P., Nguyen, T. B., Coggon, M. M., Crounse, J. D., St Clair, J. M., Zhang, X., Schilling, K. A., Seinfeld, J. H., and Wennberg, P. O.: Isoprene NO₃ Oxidation Products from the RO₂ + HO₂ Pathway, *J. Phys. Chem. A*, 119, 10158-10171, doi:10.1021/acs.jpca.5b06355, 2015.
- Skov, H., Hjorth, J., Lohse, C., Jensen, N. R., and Restelli, G.: Products and mechanisms of the reactions of the nitrate (NO₃) with isoprene, 1,3-butadiene and 2,3-dimethyl-1,3-butadiene in air, *Atmos. Environ., Part A*, 26, 2771-2783, doi:10.1016/0960-1686(92)90015-d, 1992.
- 735 Sobanski, N., Schuladen, J., Schuster, G., Lelieveld, J., and Crowley, J. N.: A five-channel cavity ring-down spectrometer for the detection of NO₂, NO₃, N₂O₅, total peroxy nitrates and total alkyl nitrates, *Atmos. Meas. Tech.*, 9, 5103-5118, doi:10.5194/amt-9-5103-2016, 2016.
- Sobanski, N., Thieser, J., Schuladen, J., Sauvage, C., Song, W., Williams, J., Lelieveld, J., and Crowley, J. N.: Day and Night-time Formation of Organic Nitrates at a Forested Mountain-site in South West Germany, *Atmos. Chem. Phys.*, 17, 4115-4130, doi: 10.5194/acp-17-4115-2017, 2017.

- 740 Stokes, G. Y., Buchbinder, A. M., Gibbs-Davis, J. M., Scheidt, K. A., and Geiger, F. M.: Heterogeneous Ozone Oxidation Reactions of 1-Pentene, Cyclopentene, Cyclohexene, and a Menthhol Derivative Studied by Sum Frequency Generation, *J. Phys. Chem. A*, 112, 11688-11698, doi:10.1021/jp803277s, 2008.
- Thieser, J., Schuster, G., Phillips, G. J., Reiffs, A., Parchatka, U., Pöhler, D., Lelieveld, J., and Crowley, J. N.: A two-channel, thermal dissociation cavity-ringdown spectrometer for the detection of ambient NO₂, RO₂NO₂ and RONO₂, *Atmos. Meas. Tech.*, 9, 553-576, doi:10.5194/amt-9-553-2016, 2016.
- 745 Thornton, J. A., Wooldridge, P. J., Cohen, R. C., Martinez, M., Harder, H., Brune, W. H., Williams, E. J., Roberts, J. M., Fehsenfeld, F. C., Hall, S. R., Shetter, R. E., Wert, B. P., and Fried, A.: Ozone production rates as a function of NO_x abundances and HO_x production rates in the Nashville urban plume, *J. Geophys. Res.-Atmos.*, 107, doi:10.1029/2001JD000932, 2002.
- Usher, C. R., Michel, A. E., and Grassian, V. H.: Reactions on Mineral Dust, *Chem. Rev.*, 103, 4883-4940, doi:10.1021/cr020657y, 2003.
- 750 Vandaele, A. C., Hermans, C., Simon, P. C., Carleer, M., Colin, R., Fally, S., Merienne, M. F., Jenouvrier, A., and Coquart, B.: Measurements of the NO₂ absorption cross-section from 42 000 cm⁻¹ to 10 000 cm⁻¹ (238-1000 nm) at 220 K and 294 K, *J. Quant. Spectrosc. Radiat. Transfer*, 59, 171-184, doi: 10.1016/S0022-4073(97)00168-4, 1998.
- Vansco, M. F., Caravan, R. L., Zuraski, K., Winiberg, F. A. F., Au, K., Trongsiwat, N., Walsh, P. J., Osborn, D. L., Percival, C. J., Khan, M. A. H., Shallcross, D. E., Taatjes, C. A., and Lester, M. I.: Experimental Evidence of Dioxole Unimolecular Decay Pathway for Isoprene-Derived Criegee Intermediates, *J. Phys. Chem. A*, 124, 3542-3554, doi:10.1021/acs.jpca.0c02138, 2020.
- 755 Vereecken, L., Carlsson, P. T. M., Novelli, A., Bernard, F., Brown, S. S., Cho, C., Crowley, J. N., Fuchs, H., Mellouki, W., Reimer, D., Shenolikar, J., Tillmann, R., Zhou, L., Kiendler-Scharr, A., and Wahner, A.: Theoretical and experimental study of peroxy and alkoxy radicals in the NO₃-initiated oxidation of isoprene, *Phys. Chem. Chem. Phys.*, 23, 5496-5515, doi:10.1039/D0CP06267G, 2021.
- Wayne, R. P., Barnes, I., Biggs, P., Burrows, J. P., Canosamas, C. E., Hjorth, J., Lebras, G., Moortgat, G. K., Perner, D., Poulet, G., Restelli, G., and Sidebottom, H.: The Nitrate Radical - Physics, Chemistry, and the Atmosphere, *Atmos. Environ.*, 25, 1-203, doi:10.1016/0960-1686(91)90192-A, 1991.
- 760 Wennberg, P. O., Bates, K. H., Crounse, J. D., Dodson, L. G., McVay, R. C., Mertens, L. A., Nguyen, T. B., Praske, E., Schwantes, R. H., Smarte, M. D., St Clair, J. M., Teng, A. P., Zhang, X., and Seinfeld, J. H.: Gas-Phase Reactions of Isoprene and Its Major Oxidation Products, *Chem. Rev.*, 118, 3337-3390, doi:10.1021/acs.chemrev.7b00439, 2018.
- 765 Wild, R. J., Edwards, P. M., Dube, W. P., Baumann, K., Edgerton, E. S., Quinn, P. K., Roberts, J. M., Rollins, A. W., Veres, P. R., Warneke, C., Williams, E. J., Yuan, B., and Brown, S. S.: A measurement of total reactive nitrogen, NO_y, together with NO₂, NO, and O₃ via cavity ring-down spectroscopy, *Environ. Sci. Technol.*, 48, 9609-9615, doi:10.1021/es501896w, 2014.
- Wolfe, G. M., Thornton, J. A., McNeill, V. F., Jaffe, D. A., Reidmiller, D., Chand, D., Smith, J., Swartzendruber, P., Flocke, F., and Zheng, W.: Influence of trans-Pacific pollution transport on acyl peroxy nitrate abundances and speciation at Mount Bachelor Observatory during INTEX-B, *Atmos. Chem. Phys.*, 7, 5309-5325, doi:10.5194/acp-7-5309-2007, 2007.
- 770 Wolfe, G. M., Marvin, M. R., Roberts, S. J., Travis, K. R., and Liao, J.: The Framework for 0-D Atmospheric Modeling (F0AM) v3.1, *Geosci. Model. Dev.*, 9, 3309-3319, doi:10.5194/gmd-9-3309-2016, 2016.
- Wooldridge, P. J., Perring, A. E., Bertram, T. H., Flocke, F. M., Roberts, J. M., Singh, H. B., Huey, L. G., Thornton, J. A., Wolfe, G. M., Murphy, J. G., Fry, J. L., Rollins, A. W., LaFranchi, B. W., and Cohen, R. C.: Total Peroxy Nitrates (SPNs) in the atmosphere: the Thermal Dissociation-Laser Induced Fluorescence (TD-LIF) technique and comparisons to speciated PAN measurements, *Atmos. Meas. Tech.*, 3, 593-607, doi:10.5194/amt-3-593-2010, 2010.
- 775 Wu, R., Vereecken, L., Tsiligiannis, E., Kang, S., Albrecht, S. R., Hantschke, L., Zhao, D., Novelli, A., Fuchs, H., Tillmann, R., Hohaus, T., Carlsson, P. T. M., Shenolikar, J., Bernard, F., Crowley, J. N., Fry, J. L., Brownwood, B., Thornton, J. A., Brown, S. S., Kiendler-Scharr, A., Wahner, A., Hallquist, M., and Mentel, T. F.: Molecular composition and volatility of multi-generation products formed from isoprene oxidation by nitrate radical, *Atmos. Chem. Phys. Discuss.*, 2020, 1-37, doi:10.5194/acp-2020-1180, 2020.
- 780

Figures and Tables

785 **Table 1:** Central reactions used for numerical simulation of the thermal dissociation of organic nitrates in TDI-2.

Reaction	Rate constant	Reference
ISOP-NIT + M → RO + NO ₂	$3.16 \times 10^{15} \exp(-19676/T) \text{ s}^{-1}$	Morin and Bedjanian, 2017
O ₃ → O(³ P) + O ₂	Pressure dependent (see text)	Peukert et al., 2013
O(³ P) + O ₂ + M → O ₃ + M	$6.0 \times 10^{-34} (T/300)^{-2.6} [M] \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$	IUPAC, 2021
O(³ P) + O ₃ → 2O ₂	$8.0 \times 10^{-12} \exp(-2060/T) \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$	IUPAC, 2021
O(³ P) + ISOP-NIT → products + NO ₂	$3.9 \times 10^{-12} \exp(680/T) \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$	Herron and Huie, 1973
O(³ P) + ISOP → products	$3.5 \times 10^{-11} \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$ (298 K)	Paulson et al., 1995
O(³ P) + wall →	90 s ⁻¹	See text
O(³ P) + NO ₂ → NO + O ₂	$5.1 \times 10^{-12} \exp(198/T) \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$	IUPAC, 2021
O ₃ + ISOP → products	$1.05 \times 10^{-14} \exp(-2000/T) \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$	IUPAC, 2021
O ₃ + ISOP-NIT → products + NO ₂	$1.05 \times 10^{-14} \exp(-2000/T) \text{ cm}^3 \text{ molecule}^{-1} \text{ s}^{-1}$ ²	IUPAC, 2021

Notes: ¹Value for 2-methyl-2-butene used. ²Value for isoprene used.






Figure 1: Schematic diagram of the thermal dissociation inlets TDI-1, TDI-2 and TDI-3.



800 **Figure 2:** Evolution of the mixing ratios of NO_2 , ΣPNs , $\Sigma\text{PNs} + \Sigma\text{ANs}$, N_2O_5 , and O_3 when flowing 150 sccm NO_2 (from 1 ppmv bottle), 7 sccm isoprene (from a 46.5 ppmv cylinder) and 24 SLPM of zero-air (of which 5 SLPM were passed over a Penray lamp) into the SCHARK. Isoprene was added at 12:00 after the system was close to steady-state. Note that the NO_2 and N_2O_5 mixing ratios were subtracted from the organic nitrate signals (middle panel).



805

Figure 3: (a) Thermograms (relative to the signal at 650 K) of isoprene-derived organic nitrates (ISOP-NIT) and isopropyl nitrate (iPN) obtained with TDI-1 (solid symbols) and TDI-2 (open symbols) under dry conditions. (b) Absolute thermograms of ISOP-NIT obtained with TDI-1 (solid symbols) and TDI-2 (open symbols) under humid conditions (RH = 34 %, 22°C). Regions (I and II) with unexpected detection of NO_2 are shaded yellow. Error bars denote standard deviation (1σ , 30 s) of the signal.




Figure 4: Temperature-dependent NO_2 detection when sampling 2.75 ppbv NO_2 and 146 ppbv O_3 in 23 SLPM dry synthetic air from the SCHARK through TDI-1 and TDI-2. The signal from the NO_2 cavity (no TDI) has been subtracted from both datasets. The mixing ratios of $\text{N}_2\text{O}_5 + \text{NO}_3$ (with associated uncertainty) is indicated by the blue area.





Figure 5: Thermograms of nitric acid (22 ppbv) in either dry (black) or humidified air (RH = 55 % at 23°C, red) obtained with TDI-1 (closed squares) and TDI-2 (open squares). The O₃ mixing ratio was either zero or 350 ppbv in the non-humidified experiment and 350 ppbv in the humidified experiment. The error bars denote standard deviation (1 σ , 1 min) of the signal.



825 **Figure 6:** (a) Thermograms of isoprene-derived nitrates generated by mixing N₂O₅ (as NO₃ source) and ~ 20 ppbv isoprene in the SCHARK under dry conditions and sampling via TDI-1. The mixing ratio of O₃ (added only to the inlets) was varied from 0 to 202 ppbv. A thermogram of isopropyl nitrate (iPN) in the presence of 162 ppbv O₃ (and 3 ppbv NO₂ impurity) sampled through the same TDI is shown for comparison. The black solid line is a Boltzmann sigmoidal fit of the ozone free experiment. The error bars show the standard deviation of the signal (1σ).
 830 (b) Same as (a) but using TDI-2.





Figure 7: Black lines and data points: Simulated, fractional conversion of n-propyl nitrate to NO₂ for TDI-2 without and with O₃ (200 ppbv). Red lines and data points: Simulated fractional conversion of O₃ to O(³P) within the heated section of TDI-2 with and without wall loss of O(³P). Error bars denote the standard deviation (1σ) of the fractional conversion changing over time passed in the heated section.

835



845 **Figure 8:** Time-series of O₃ mixing ratios before and after passing through TDI-1 (a) or TDI-2 (b).






Figure 9: Possible mechanism for surface-catalysed conversion of first generation ISOP-NIT in TDI-1. RS-O represents a reactive surface site.




855 **Figure 10:** Simulated fractional contribution of hydroperoxy nitrates to ISOP-NIT using schemes taken from the MCM (red) and from RCIMP (black). The simulations reproduce the experimental conditions in Fig.3 and Fig.6, with 22 ppbv of isoprene together with either 3 ppbv N₂O₅ as NO₃ source (runs 1 and 3) or 10.8 ppbv NO₂ + 150 ppbv O₃ (runs 2 and 4).

860



865 **Figure 11:** Potential degradation pathways of isoprene-derived hydroperoxides on a heated quartz surface.



870 **Figure 12:** Mixing ratios of NO_2 , ΣPNs with TDI-3 and $\Sigma\text{PNs} + \Sigma\text{ANs}$ with TDI-2 obtained from constant introduction of 200 sccm NO_2 , and 9.8 sccm isoprene in 25 SLPM (of which 5 SLPM were passed over a Penray lamp to generate O_3) dry synthetic air into the SCHARK. A CIMS measurement of PAN is appended (blue squares).

875






Figure S1: (a) Time-series of O_3 in the SCHARK at a flow rate of 15 SLPM dry zero-air. The Penray lamp used to generate O_3 was switched on at ~12:50 LT and switched off at 13:28 LT. The growth and depletion of the O_3 signal are reproduced by exponential expressions to derive the exchange constant (k_{exch}) (b) Dependence of k_{exch} on the total flow rate into the SCHARK. The linear regression indicates $k_{exch} = 1.82 \times 10^{-5} \text{ s}^{-1} \text{ SLPM}^{-1}$.

885



890 **Figure S2:** Dependence of NO₂ detection on TD temperature when sampling a mixture of 10.8 ppbv NO₂ in 15 SLPM humidified synthetic air (RH = 40%, 23°C) through TDI- 2. The NO₂ signal measured in the NO₂ cavity (unheated inlet) has been subtracted. The error bars represent standard deviation (1σ , 1 min) of the signal.

895



900 **Figure S3:** CIMS measurement of HNO₃ and HONO during an experiment in which NO₂ (10.8 ppbv) and O₃ (150 ppbv) were mixed in the SCHARK. At 16:30 LT the zero-air was humidified (RH = 23%, 23°C). Periods during which air from the SCHARK was first passed through TDI-1 or -2 prior to entering the CIMS are marked in blue and orange respectively. When the air was heated to 400 K while flowing through TDI-2 (yellow hatched) no effect was observed in the signal.




Figure S4: Unexpected detection of NO₂ in dry synthetic air when adding 120 ppbv O₃ to TDI-1. Initially, TDI-1 was at room temperature, with heating (to 703 K) of the TDI starting at ~14:20. When sampling via TDI-1 ~ 60 ppbv of NO₂ were observed, while no signal was registered when sampling into the NO₂ cavity through a cold inlet.




Figure S5: Mixing ratios of ISOP-NIT when sampling via TD-1 and TD-2 and different temperatures. ISOP-NIT was formed by flowing NO₂ (150 sccm of 1 ppmv) isoprene (9.8 sccm of 46.5 ppmv) and O₃ in 15 SLPM synthetic air into the SCHARK. The solid and dotted lines in the upper panel indicate the temperature of each TDI. At ~15:15 LT (when the signals were stable and the system in steady-state) the TDI temperatures were decreased to obtain a thermogram. For 30 min, starting at 16:30 LT both TDIs were kept close to room temperature before being re-heated to ~700 K at 17:00 LT. The strong increase in signal (above that observed at e.g. 15:00) observed when sampling via TD-1 indicates that ISOP-NIT can deposit on the glass beads and frit at lower temperatures. This effect is not observed for TDI-2 and is clearly related to the use of the glass-beads.

930




Figure S6: Thermograms of peroxy acetyl nitrate (PAN) obtained when sampling via TDI-2 and TDI-3. PAN (together with ~ 1.2 ppbv NO₂) was produced from a photochemical source as described recently (Eger et al., 2019) using the 285 nm photolysis of acetone in the presence of NO (Warneck and Zerbach, 1992; Flocke et al., 2005). The error bars represent standard deviation (1 σ , 1 min) of the signal.

940

945




Figure S7: Mixing ratios of NO₂, PNs (measured with TDI-3 at 448 K) and ClNO₂ (measured with TDI-2 at 698 K) recorded with the TD-CRDS setup during an experiment in which Cl₂ (1-10 sccm of 50 ppmv in N₂) was passed through a glass bulb with sodium nitrate and diluted in 10 SLPM dry synthetic air. Variation of the amounts of ClNO₂ and NO₂ was achieved by changing the flow of Cl₂. The NO₂ signal measured with an unheated inlet was subtracted from the channels with TDIs. Note that, as expected, no signal additional to NO₂ was observed with TDI-3 at 448 K. Using TDI-2 at 698 K, ClNO₂ was detectable implying that it interferes detection of ANs.

950

955

960

S8 Detection of PAN, HONO and HNO₃ with I-CIMS

965 A chemical ionization mass spectrometer using iodide primary ions (I-CIMS) described recently (Dörich et al., 2021) was deployed to detect PAN, HONO and HNO₃. The I-CIMS was coupled to the SCHARK via \approx 2m of ¼ inch PFA tubing heated to 40 °C. The flow rate through the PFA-tubing was \sim 2.1 SLM. Iodide anions were generated by passing 4 sccm of 400 ppmv methyl iodide (CH₃I in N₂) diluted in 750 sccm N₂ (Westfalen, 5.0) through a 370 MBq polonium (²¹⁰Po) source.

PAN was thermally dissociated in an heated inlet (PFA tube at 170°C, residence time of 40 ms) to peroxy acetyl radicals which are detected as acetate ions CH₃CO₂⁻ (*m/z* 59) after reaction with I⁻ (Phillips et al., 2013). Calibration was performed using a photochemical PAN source (Warneck and Zerbach, 1992). HNO₃ was detected as the I(HNO₃) cluster ion at *m/z* 190 and
970 calibrated using an HNO₃ permeation source characterised by optical absorption. HONO was detected as NO₂⁻ (*m/z* 46) using acetate anions (Veres et al., 2008) generated by adding a high concentration of PAN to the TD-inlet.

975

980

985

990

995

S7 S9 Reaction scheme used for numerical simulations in inlets

```

1000 * Thermolysis of ozone and alkyl nitrates ;
* ===== ;
variable N2O5 NO3 NO OH HO2 HNO3 H2O2 HONO ;
HO2NO2 H2O ISOP O PROD PROD1 PROD2 ISOPNIT O3 NO2 ;
* ----- ;
1005 * INITIAL CONCENTRATIONS ;
* ----- ;
parameter O3i 3.53e12 ;
parameter NO2i 8.83e10 ;
1010 parameter ISOPi 1.8e11 ;
parameter NO3i 8.82e7 ;
parameter N2O5i 8.82e8 ;
parameter NOi 0.0e10 ;
parameter O2 ;
1015 parameter k1 k2 k2 k3 k4 k5 k6 k7 k8 k9 k10 k11 k12 k13 k14 ;
parameter k15 k16 k17 k18 k19 k20 k21 k22 k23 k24 k25 k26 ;
parameter k27 k28 k29 k30 k31 k32 k33 k34 k35 ;
parameter M ;
parameter ISOPNITi 1.8E10 ;
parameter kWall 90 ;
1020 * ----- ;
parameter P 543.478 ;
parameter T 298 ;
* ----- ;
* ;
1025 COMPILE GENERAL ;
M = P * 3.24E16 * 298/T ;
O2 = 0.2*M ;
** ;
* ;
1030 COMPILE INITIAL ;
N2O5 = N2O5i ;
NO2 = NO2i ;
O3 = O3i ;
O2 = O2i ;
NO = NOi ;
1035 NO3 = NO3i ;
ISOP = ISOPi ;
ISOPNIT = ISOPNITi ;
** ;
* ;
1040 COMPILE EQUATIONS ;
* ----- ;
* HOx and NOx chemistry ;
* ----- ;
% k1 : N2O5 = NO3 + NO2 ;
1045 % k2 : NO2 + NO3 = N2O5 ;
% k3 : NO + NO3 = NO2 + NO2 ;
% k4 : NO2 + O3 = NO3 + O2 ;
% k5 : NO + O3 = NO2 + O2 ;
% k6 : OH + O3 = HO2 + O2 ;
1050 % k7 : HO2 + O3 = OH + O2 + O2 ;
% k8 : OH + HO2 = H2O + O2 ;
% k9 : OH + NO2 = HNO3 ;
% k10 : OH + NO3 = HO2 + NO2 ;
% k11 : HO2 + NO3 = OH + NO2 + O2 ;
1055 % k12 : OH + H2O2 = HO2 + H2O ;
% k13 : HO2 + HO2 = H2O2 + O2 ;
% k14 : OH + NO = HONO ;
% k15 : HO2 + NO = OH + NO2 ;
% k16 : HO2 + NO2 = HO2NO2 ;
1060 % k17 : HO2NO2 + OH = NO2 + H2O + O2 ;
% k18 : OH + HONO = NO2 + H2O ;
% k19 : OH + HNO3 = NO3 + H2O ;
% k20 : HO2NO2 = HO2 + NO2 ;
* ----- ;
1065 * ISOPRENE oxidation ;
* ----- ;
% k21*0.74 : ISOP + O3 = PROD ;
% k21*0.26 : ISOP + O3 = OH ; IUPAC recommended yield
% k22 : ISOP + NO3 = PROD ;

```



```

1070 % k23           : ISOP + OH = PROD           ;
*-----;
* Ox chemistry  ;
*-----;
% k24           : O3 = O + O2                 ; Peukert et al.
% k25           : O + O3 = O2 + O2             ; IUPAC
1075 % k26           : O + O2 = O3               ; IUPAC
% k27           : O + H2O = OH + OH           ;
% kWall        : O =                           ;
% k28           : NO2 + O = O2 + NO           ; bimolecular
% k29           : NO2 + O = NO3               ; termolecular
1080 % k30           : NO3 + O = O2 + NO2       ;
% k31           : O + NO = NO2               ;
*-----;
* Thermolysis of alkyl nitrates
*-----;
1085 % k32           : ISOPNIT = NO2 + PROD1     ;
% k33           : ISOPNIT + O = NO2 + PROD2 + O2 ;
% k34           : ISOP + O = PROD            ;
% k21          : ISOPNIT + O3 = NO2 + PROD2 + OH ;
*-----;
1090 *Rate equations
k1 = ((1.3e-3*(T/300)^-3.5*exp(-11000/T))*M*(9.7e14*(T/300)^0.1*exp(-11080/T)))/((1.3e-3*(T/300)^-3.5*exp(-11000/T))*M+(9.7e14*(T/300)^0.1*exp(-11080/T)))*10@(log10(0.35)/(1+(log10((1.3e-3*(T/300)^-3.5*exp(-11000/T))*M/(9.7e14*(T/300)^0.1*exp(-11080/T))))/(0.75-1.27*log10(0.35)))@2) ; N2O5 decomp IUPAC
k2 = ((3.6e-30*(T/300)^-4.1)*M*(1.9e-12*(T/300)^0.2))/((3.6e-30*(T/300)^-4.1)*M+(1.9e-12*(T/300)^0.2))*10@(log10(0.35)/(1+(log10((3.6e-30*(T/300)^-4.1)*M/(1.9e-12*(T/300)^0.2))/(0.75-1.27*log10(0.35)))@2) ; NO2 + NO3 IUPAC
1100 k3 = 1.8E-11*exp(110/T) ; IUPAC NO + NO3 (220-420)K
k4 = 1.4E-13 * exp (-2470/T) ; IUPAC NO2 + O3 (230-360)K
k5 = 2.07E-12 * exp (-1400/T) ; IUPAC NO + O3 (195-310)K
1105 k6 = 1.70E-12*EXP(-940/T) ; IUPAC OH + O3 (220-450)K
k7 = 2.03E-16*(T/300)^4.57*EXP(693/T) ; IUPAC HO2 + O3 (250-340)K
k8 = 4.8E-11*EXP(250/T) ; IUPAC HO2 + OH (250-400)K
k9 = ((3.2e-30*(T/300)^-4.5)*M*(3.0e-11))/((3.2e-30*(T/300)^-4.5)*M+(3.0e-11))*10@(log10(0.41)/(1+(log10((3.2e-30*(T/300)^-4.5)*M/(3.0e-11))/(0.75-1.27*log10(0.41)))@2) ; IUPAC HNO3 formation
1110 k10 = 2E-11 ; IUPAC OH + NO3 (298K)
k11 = 4E-12 ; IUPAC HO2 + NO3 (298K)
k12 = 2.9E-12*exp(-160/T) ; IUPAC OH + H2O2 (240-460)K
1115 k13 = 2.20E-13*(1+(1.40E-21*EXP(2200/T)*H2O))*EXP(600/T) ; IUPAC H2O2 formation (230-420)K
k14 = ((7.4e-31*(T/300)^-2.4)*M*(3.3e-11*(T/300)^-0.3))/((7.4e-31*(T/300)^-2.4)*M+(3.3e-11*(T/300)^-0.3))*10@(log10(0.81)/(1+(log10((7.4e-31*(T/300)^-2.4)*M/(3.3e-11*(T/300)^-0.3))/(0.75-1.27*log10(0.81)))@2) ; IUPAC HONO formation
1120 k15 = 3.45E-12*exp(270/T) ; IUPAC HO2 + NO (200-400)K
k16 = ((1.4e-31*(T/300)^-3.1)*M*(4.0e-12))/((1.4e-31*(T/300)^-3.1)*M+(4.0e-12))*10@(log10(0.4)/(1+(log10((1.4e-31*(T/300)^-3.1)*M/(4.0e-12))/(0.75-1.27*log10(0.4)))@2) ; IUPAC HO2 + NO2
1125 k17 = 3.2e-13*EXP(690/T) ; IUPAC HO2NO2 + OH (210-300)K
k18 = 2.5e-12*EXP(260/T) ; IUPAC HONO + OH (290-380)K
k19 = 2.40E-14*EXP(460/T) + ((6.50E-34*EXP(1335/T))*M)/(1+(6.50E-34*EXP(1335/T))*M/2.70E-17*EXP(2199/T))) ; IUPAC OH + HNO3
1130 k20 = ((4.1e-5*exp(-10650/T))*M*(6.0e15*exp(-11170/T)))/((4.1e-5*exp(-10650/T))*M+(6.0e15*exp(-11170/T))*10@(log10(0.4)/(1+(log10((4.1e-5*exp(-10650/T))*M/(6.0e15*exp(-11170/T)))/(0.75-1.27*log10(0.4)))@2) ; IUPAC HO2NO2 decomposition
k21 = 1.05E-14*exp(-2000/T) ; O3 + ISOP - IUPAC
k22 = 2.95E-12*exp(-450/T) ; NO3 + ISOP - IUPAC
1135 k23 = 2.1E-11*exp(465/T) ; OH + ISOP - IUPAC
k24 = (37800*(T^-4.37)*EXP(-13738/T))*M*(1370000000000000*(T^-0.67)*EXP(-13080/T))*10@(LOG10((1-0.647)*EXP(-T/0.000391)+0.6417*EXP(-T/8680.74)+EXP(-6060.75/T))/(1+(LOG10((37800*(T^-4.37)*EXP(-13738/T))*M/(1370000000000000*(T^-0.67)*EXP(-13080/T)))/(0.75-1.27*LOG10((1-0.647)*EXP(-T/8680.74)+EXP(-6060.75/T))))@2) /
1140 EXP(-T/8680.74)+EXP(-6060.75/T)))@2) /

```

```

((37800*(T@-4.37)*EXP(-13738/T))*M+(1370000000000000*(T@-0.67)*
EXP(-13080/T)))
; O3 thermolysis (T>700K), Peukert et al., 2013
1145 k25 = 8E-12*exp(-2060/T) ; IUPAC O + O3 (200-400)K
k26 = 6.0E-34*((T/300)@-2.6)*M ; IUPAC O + O2 (200-300)K
k27 = 8.44E-14*(T@0.946)*exp(-8571/T) ; O(3P) + H2O (250-2400)K Lifshitz et al., 1990
k28 = 5.1E-12*exp(198/T) ; IUPAC NO2 + O (220-420)K
k29 = ((1.3e-31*(T/300)@-1.5)*M*(2.3e-11*(T/300)@0.24))/
((1.3e-31*(T/300)@-1.5)*M+(2.3e-11*(T/300)@0.24))*10@(log10(0.6)/
(1+(log10((1.3e-31*(T/300)@-1.5)*M/(2.3E-11*(T/300)@0.24))/
(0.75-1.27*log10(0.6)))@2)) ; IUPAC termol NO2 + O
1150 k30 = 1.7E-11 ; IUPAC NO3 + O (298K)
k31 = ((1.0e-31*(T/300)@-1.6)*M*(5e-11*(T/300)@-0.3))/
((1.0e-31*(T/300)@-1.6)*M+(5e-11*(T/300)@-0.3))*10@(log10(0.85)/
(1+(log10((1.0e-31*(T/300)@-1.6)*M/(5e-11*(T/300)@-0.3))/
(0.75-1.27*log10(0.85)))@2)) ; IUPAC O + NO
1155
k32 = 7.34E15*exp(-19676/T) ; thermolysis n-propylnitrate
k33 = 3.9E-12*exp(680/T) ; 2-methyl-2-butene - Herron & Huie, 1973 (298-400K)
1160 k34 = 3.5E-11 ; Paulson et al., 1995 (298K)
** ;
COMPILE INSTANT ;
open 7 "thermo.sim" new ;
** ;
1165 COMPILE BLOCK 3 ;
PSTREAM 3 ;
** ;
COMPILE BLOCK 4 ;
1170 T = 425 ;
** ;
COMPILE BLOCK 5 ;
T = 298 ;
** ;
1175 COMPILE BLOCK 6 ;
T = 303 ;
** ;
PSTREAM 3 7 ;
time O3 O NO2 M T PROD1 PROD2 ISOPNIT ;
** ;
1180 when ;
1) time = 0 + 0.01*128 call block 3 ;
2) time = 0.007 call block 4 restart ;
3) time = 0.335 call block 5 restart ;
1185 4) time = 1.08 call block 6 restart ;
** ;
*hmax 0.01 ;
BEGIN ;
STOP ;
1190
1195
1200
1205
1210

```

1215 **References**

- Dörich, R., Eger, P., Lelieveld, J., and Crowley, J. N.: Iodide-CIMS and m/z 62: The detection of HNO₃ as NO₃ in the presence of PAN, peracetic acid and O₃, *Atmos. Meas. Tech. Discuss.*, 2021, 1-26, doi:10.5194/amt-2021-57, 2021.
- 1220 Eger, P. G., Helleis, F., Schuster, G., Phillips, G. J., Lelieveld, J., and Crowley, J. N.: Chemical ionization quadrupole mass spectrometer with an electrical discharge ion source for atmospheric trace gas measurement, *Atmos. Meas. Tech.*, 12, 1935-1954, doi:10.5194/amt-12-1935-2019, 2019.
- Flocke, F. M., Weinheimer, A. J., Swanson, A. L., Roberts, J. M., Schmitt, R., and Shertz, S.: On the measurement of PANs by gas chromatography and electron capture detection, *J. Atmos. Chem.*, 52, 19-43, 2005.
- 1225 Phillips, G. J., Pouvesle, N., Thieser, J., Schuster, G., Axinte, R., Fischer, H., Williams, J., Lelieveld, J., and Crowley, J. N.: Peroxyacetyl nitrate (PAN) and peroxyacetic acid (PAA) measurements by iodide chemical ionisation mass spectrometry: first analysis of results in the boreal forest and implications for the measurement of PAN fluxes, *Atmospheric Chemistry and Physics*, 13, 1129-1139, doi:10.5194/acp-13-1129-2013, 2013.
- 1230 Veres, P., Roberts, J. M., Warneke, C., Welsh-Bon, D., Zahniser, M., Herndon, S., Fall, R., and de Gouw, J.: Development of negative-ion proton-transfer chemical-ionization mass spectrometry (NI-PT-CIMS) for the measurement of gas-phase organic acids in the atmosphere, *Int. J. Mass Spectrom.*, 274, 48-55, doi:10.1016/j.ijms.2008.04.032, 2008.
- 1235 Warneck, P., and Zerbach, T.: Synthesis of peroxyacetyl nitrate in air by acetone photolysis, *Environ. Sci. Technol.*, 26, 74-79, doi:10.1021/es00025a005, 1992.