Effects of Aerosol Size and Coating Thickness on the Molecular Detection using Extractive Electrospray Ionization

Chuan Ping Lee¹, Mihnea Surdu¹, David M. Bell¹, Houssni Lamkaddam¹, Mingyi Wang^{2,3}, Farnoush Ataei⁶, Victoria Hofbauer^{2,3}, Brandon Lopez^{2,4}, Neil M. Donahue^{2,3,4,5}, Josef Dommen¹, Andre S. H. Prévôt¹, Jay G. Slowik^{1,*}, Dongyu Wang^{1,*}, Urs Baltensperger¹, Imad El Haddad^{1,*}

¹Laboratory of Atmospheric Chemistry, Paul Scherrer Institute, 5232 Villigen, Switzerland ²Center for Atmospheric Particle Studies, Carnegie Mellon University, Pittsburgh, PA 15213, USA ³Department of Chemistry, Carnegie Mellon University, Pittsburgh, PA 15213, USA ⁴Department of Chemical Engineering, Carnegie Mellon University, Pittsburgh, PA 15213, USA

⁵Department of Engineering and Public Policy, Carnegie Mellon University, Pittsburgh, PA 15213, USA ⁶Leibniz-Institute for Tropospheric Research, 04318 Leipzig, Germany

Correspondence to: Imad El Haddad (imad.el-haddad@psi.ch), Dongyu Wang (dongyu.wang@psi.ch), Jay G. Slowik (jay.slowik@psi.ch)

- 15 Abstract. Extractive electrospray ionization (EESI) is a well-known technique for high throughput online molecular characterization of chemical reaction products and intermediates, detection of native biomolecules, in vivo metabolomics, and environmental monitoring with negligible thermal and ionization-induced fragmentation for over two decades. However, the EESI extraction mechanism remains uncertain. Prior studies disagree whether particles between 20 and 400 nm diameter are fully extracted or if the extraction is limited to the surface layer. Here, we examined the analyte extraction mechanism by
- 20 assessing the influence of particle size and coating thickness on the detection of the molecules therein. We find that particles are extracted fully: Organics-coated NH₄NO₃ particles with a fixed core volume (156 and 226 nm in diameter without coating) showed constant EESI signals for NH₄NO₃ independent of the shell coating thickness, while the signals of the secondary organic molecules comprising the shell varied proportionally to the shell volume. We also found that the EESI sensitivity exhibited a strong size dependence, with an increase in sensitivity by one to three orders of magnitude as particles size
- 25 decreased from 300 nm to 30 nm. This dependence varied with the electrospray (ES) droplet size, the particles size and residence time for coagulation in the EESI inlet, suggesting that the EESI sensitivity was influenced by the coagulation coefficient between particles and ES droplets. Overall, our results indicate that, in the EESI, particles are fully extracted by the ES droplets regardless of the chemical composition, when they are collected by the ES droplets. However, their coalescence is not complete and depends strongly on their size. This size-dependence is especially relevant when EESI is used to probe
- 30 size-varying particles as is the case in aerosol formation and growth studies with size ranges below 100 nm.

1 Introduction

5

Atmospheric aerosols are suspended particles in the air ranging from a few nanometers (nm) to several micrometers (μ m) in diameter. Fine particles (< 1 μ m) comprise nucleation, Aitken and accumulation mode particles, and can account for 50-70 % of the total particulate matter (PM) mass concentration in polluted environments (Yue et al., 2009). They can affect the earth's

- 35 radiative balance either directly, by interacting with solar radiation, or indirectly by acting as cloud condensation nuclei (CCN), influencing cloud albedo and lifetime (Seinfeld and Pandis, 2016). Exposure to PM is one of the leading causes for premature death, accounting for ~8.9 million deaths, or ~10% of total global burden of mortality in 2015 (Burnett et al., 2018), though the underlying mechanisms remain uncertain (Daellenbach et al., 2020). PM can be emitted as primary aerosol or produced in the atmosphere after chemical reactions via nucleation or condensation of gas-phase products (Berndt et al., 2005; Clarke et al., 2005).
- 40 al., 1984; Hoffmann et al., 1997; Jimenez et al., 2009; Kalberer et al., 2004; Kirkby et al., 2011). Heterogeneous reactions may also further increase the complexity of ambient aerosol mixtures (George and Abbatt, 2010; Ditto et al., 2020).

Online molecular characterization of atmospheric aerosols is required to resolve the spatiotemporal variability of PM molecular composition and to identify PM sources. Progress has been made with the development of chemical ionization interfaces such as the Filter Inlet for Gases and AEROsols (FIGAERO) (Lopez-Hilfiker et al., 2014), Thermal Desorption Differential

- 45 Mobility Analyzer (TD-DMA) (Holzinger et al., 2010; Wagner et al., 2018), and Chemical Analysis of Aerosol Online (CHARON) (Eichler et al., 2015) coupled to a mass spectrometer. However, these techniques suffer from thermal decomposition of the analyte prior to ionization and/or ionization-induced fragmentation, impeding molecular speciation (Müller et al., 2017; Stark et al., 2017). To complement these instruments, an extractive electrospray (ES) ionization time-offlight mass spectrometer (EESI-TOF) was developed to enable molecular characterization of organic aerosol at 1 Hz time
- 50 resolution with ng m⁻³ level detection limit, and minimal thermal and ionization-induced fragmentation (Lopez-Hilfiker et al., 2019). The EESI-TOF was further developed to enable online tandem mass spectrometry for molecular structure elucidation, and to characterize water-soluble metals (Giannoukos et al., 2020; Lee et al., 2020). Several studies on topics such as the extraction of macromolecules from colloidal solution (Chen et al., 2006), electron-transfer-catalyzed dimerization (Marquez et al., 2008), and gas plume mixing in the charged droplets (Cheng et al., 2008)
- 55 reported that the ionization of EESI mainly happens in the liquid phase via interaction between ES charged droplets and neutral analyte molecule. For clarity, we refer to our analytes (here introduced in aerosol form) as "particles" prior to their interaction with ES droplets, and as "analyte-laden droplets" afterwards. If this liquid-phase extraction of EESI occurs via total coalescence between particles and ES droplets, the measured EESI signal should be proportional to the total analyte mass concentration, i.e. full extraction of particles by ES droplets as demonstrated by several studies (Law et al., 2010; Lopez-
- 60 Hilfiker et al., 2019). In contrast, prior studies suggested that the particles may be only partially probed, limiting the full quantification of the extracted analyte with extractive electrospray ionization (Wang et al., 2012, Kumbhani et al., 2018). Kumbhani et al. (2018) suggested that only the surface of particles with a diameter of approximately 100 nm was extracted by comparing infusion ESI-MS with EESI-MS using coated chemical standards (Kumbhani et al., 2018). Using other techniques such as phase Doppler anemometer, Wang et al. (2012) suggested that the extraction happens via fragmentation of the analyte
- droplets and ES droplets as the result of droplet-droplet collisions (Wang et al., 2012). Finally, other studies proposed that the EESI extraction efficiency could depend on the analyte volatility and size (Meier et al., 2011a; Pagonis et al., 2020). Since all these studies only probed simple systems i.e. individual chemical standards using one kind of experimental and EESI ionization source, these discrepancies could be inherently attributed to their differences of ES ionization geometries, experimental conditions, irreproducible ES Taylor cone conditions and perhaps the choices of chemicals.
- 70 Without reconciling the discrepancies of these reported EESI mechanisms, EESI quantification must be regarded as highly uncertain when the technique is used to probe varying size distributions of particles that exist in different mixing states and are comprised of different molecular polarity, volatility, and sizes. Here, we took advantage of recent advancements in particle generation and chemical analysis to evaluate the extraction mechanism of EESI using three different methods for particles generation and several online mass spectrometers for aerosol chemical speciation. First, we characterized the EESI extraction
- 75 efficiency with particles containing atmospherically relevant standard compounds and mixtures, size-selected in the range of 30-500 nm using an aerosol aerodynamic classifier. We elucidated the influence of ES operating parameters and the residence time of ES droplets and particles within the ionization source using two different EESI sources. Second, we assessed whether the EESI extraction efficiency depends on the analyte chemical composition, by comparing EESI-TOF and a chemical ionization (CI) TOF-MS equipped with a Filter Inlet for Gases and AEROsols (FIGAERO) sampling manifold (FIGAERO-
- 80 CI-ToF-MS) measurements of α-pinene secondary organic aerosol (SOA) generated in the CLOUD (Cosmics Leaving OUtdoor Droplets) chamber at CERN, Switzerland (Kirkby et al., 2016; Tröstl et al., 2016; Dias et al., 2017). Third, we determined whether particles are fully extracted or if extraction is limited to the coated surface by coating monodisperse NH₄NO₃ particles at a fixed size with variable amounts of oxidation products in an oxidation flow tube reactor.

2 Experimental

85 2.1 Materials

Acetonitrile (Sigma-Aldrich, UV grade), sodium iodide (Sigma-Aldrich, 99.7% purity) and milli-Q water (18 MΩ cm) were used to prepare the electrospray (ES) and chemical standard nebulization solution. Polyimide-coated fused silica capillary (inner diameter (i.d.): 75 µm, outer diameter (o.d.): 369 µm; BGB Analytik, Boeckten, Switzerland), HEPA capsule filter (Pall Corporation), PEEK tubing (i.d.: 500 µm, o.d.: 1/16 inch; BGB Analytik, Boeckten, Switzerland) and charcoal denuders

90 (Ionicon GmbH, Austria) were used for the electrospray ionization inlet. As chemical standards, α-pinene (Sigma-Aldrich, 99% purity), levoglucosan (Sigma-Aldrich, 99% purity), sucrose (Sigma-Aldrich, 99% purity) and ammonium nitrate (Sigma-Aldrich, 98% purity) were used.

2.2 Electrospray ionization configuration

- Two designs of the EESI sources with a factor of 2 difference in their residence time in the electrospray ionization region were used in this work, coupled to a high-resolution TOF mass spectrometer (HTOF, Tofwerk AG, Switzerland). EESI source A (Lopez-Hilfiker et al., 2019) and B were developed initially for Tofwerk TOF and Thermo Scientific Orbitrap mass analyzers (Figure S1), respectively, though EESI source B is compatible with both mass analyzers, as described in detail elsewhere (Lee et al., 2020). Source A was used throughout the study and Source B was only used in size-selection experiments shown Figure 2. Two electrospray (ES) solutions were used to generate charged ES droplets: (1) acetonitrile/H₂O (50/50 v/v); and (2) 100%
- H₂O (Table S1). Both solutions were doped with 100 ppm NaI. A potential difference of around 2.6-2.9 kV relative to the MS interface was applied to the ES solution, and an air pressure difference of 120 to 800 mbar was applied to the ES solution bottle reservoir, delivering 0.3 23 μl min⁻¹ of ES solution via a polyimide fused silica capillary (o.d.: 369 μm and i.d.: 50, 75 and 100 μm, BGB Analytik, Switzerland). Different ES operating parameters with estimated ES parent droplet size ranging from 0.7 to 5.66 um are tabulated in Table S2. The ES droplets intersected with the particles before entering the heated TOF-
- 105 capillary kept at 275 °C (<1 ms residence time), undergoing a Coulomb explosion as the ES droplets evaporated. The ions generated from organic molecules were detected predominantly (> 95 % relative abundance) as sodiated adducts ([M+Na]⁺) in the positive ionization mode by the HTOF. Ammonium nitrate (NH₄NO₃), an inorganic salt, was detected as [NaNO₃+Na]⁺. The raw mass spectra (1 Hz) were post-averaged every 10 seconds using Tofware (version 2.5.13). All measured analyte signals were normalized by the most abundant electrospray ion (i.e. [NaI+Na]⁺) to account for the variation of the electrospray
- 110 signal (± 5 %).

2.3 Particle size selection

Figures S2 and S3 show two experimental setups for the investigation of the size-dependence of the particle extraction efficiency using EESI. Chemical standards were used in the first experimental setup (Figure S2). Three individual aqueous solutions containing 4000 ppm of levoglucosan, sucrose and ammonium nitrate, respectively, were nebulized separately at 1.4

- 115 L min⁻¹, which was then mixed with a 1.6 L min⁻¹ make-up zero air. The output particles were dried with a custom-made drier containing silica gel (< 5 % RH) and subsequently size-selected using an aerosol aerodynamic classifier (AAC; Cambustion, United Kingdom) to produce monodisperse particles (Tavakoli and Olfert, 2013, 2014; Tavakoli et al., 2014). The size selection was implemented by centrifugal separation of the particles according to their mass. Unlike size selection using differential mobility analyzers (Lopez-Hilfiker et al., 2019), size selection using the AAC does not require electrical charging,</p>
- 120 thereby avoiding multi-charging artifacts. The possible multi-charging of particles might affect the normal extraction condition by EESI where the particles are assumed to be neutral (Kebarle and Verkcerk, 2009, 2012). In addition, doubly charged particles could result in underestimation of particles size and mass concentration. Therefore, the new experimental setup we used here was well suited to study the size dependence of EESI sensitivity.

After particle size selection, the sample was drawn through a multichannel charcoal denuder to strip gas-phase constituents

- 125 before entering the EESI-TOF inlet manifold. The sample was also characterized immediately upstream of the electrospray region by a nano-scanning mobility particle sizer (size range 2.5 - 239 nm, nano-SMPS, TSI Inc., USA), a scanning mobility particle sizer (size range 16 - 638 nm, TSI SMPS, TSI Inc., USA) and an aerosol mass spectrometer equipped with a long time-of-flight mass analyzer (AMS-LTOF, Aerodyne Research Inc., USA) (Figure S2). The high concentration of the chemical solutions ensured that sufficiently high analyte concentrations (> 3 μ g m⁻³) remain after size-selection by the AAC using the
- 130 highest possible sheath flow (15 L min⁻¹ at an aerodynamic diameter $D_{ae} > 150$ nm) to produce highly monodisperse particles distributions (Tavakoli and Olfert, 2014) at 30 - 40 % RH. A HEPA filter was used for the background measurements after each particle size-selection.

In the second configuration (Figure S3), we investigated the size-dependent EESI sensitivity towards biogenic SOA produced from α -pinene oxidation in the Cosmic Leaving OUtdoor Droplets (CLOUD) chamber at CERN, Switzerland (Kirkby et al.,

- 135 2011, Dias et al., 2017), at -50 to -30 °C with 20% and 60% RH (Simon et al., 2020). The EESI-TOF signals of individual αpinene oxidation products (C₁₀H₁₆O₃₋₈) were compared to FIGAERO-CI-ToF-MS (Lopez-Hilfiker et al., 2014). The FIGAERO-CI-ToF-MS measured both the gas and particle phases. Here, particles were first collected onto a 24-mm diameter PTFE filter via a dedicated port with a sampling flow rate of 6 L min⁻¹. Then, 2.7 L min⁻¹ of ultra high purity (UHP) N₂ was heated progressively to thermally desorb and vaporize the collected particles during a 14-minute desorption period, with the
- 140 filter temperature varying from 20 to 150 °C at a rate of 10 °C min⁻¹. The desorbed vapor analytes were sampled into an ionmolecule reactor at 150 mbar and chemically ionized by iodide (I⁻) ions generated by passing a UHP N₂ gas stream containing CH₃I through a ²¹⁰Polonium radioactive source before entering an LTOF mass analyzer for separation. The signal of the FIGAERO-CI-ToF-MS was integrated over the period of particle desorption. The organic analytes were detected predominantly in the form of iodide adducts [M+I]⁻ (> 95% relative abundance). The sample collection efficiency of the filter
- 145 used by the FIGAERO-CI-ToF-MS is expected to be higher than 99.9 % for particles above 10 nm (Hilfiker-Lopez et al., 2014). The volume-weighted geometric mean diameters were determined using an SMPS (size range 9 834 nm, Leibniz Institute for Tropospheric Research, Germany). The SMPS(s) used for the measurements of chemical standards and α-pinene SOA were calibrated using size standards of polystyrene latex beads.

2.4 Particle surface coating

- 150 A 104 cm long Pyrex flow tube of 7.4 cm inner diameter with a total volume of approximately 5 L (Molteni et al., 2018) was used for particle surface coating experiments (Figure S4). A 1000 ppm NH₄NO₃ solution in pure water was nebulized at 1.4 L min⁻¹ and dried before size-selection by the AAC (< 5% RH). The resulting NH₄NO₃ particles passed through the charcoal denuders before entering concentrically into the flow tube with a laminar zero air sheath flow of 10 L min⁻¹ at 20 °C and 60% RH. Measurements were performed downstream of the flow tube. Particle composition and size were measured by the EESI-
- TOF and SMPS (16 615 nm), respectively. Two different core sizes (155.8 and 226.4 nm) of NH₄NO₃ particles were used for the coating experiments. 4.7 ± 0.4 ppm α-pinene, as measured by a quadrupole proton-transfer-reaction mass spectrometer (Q-PTR), was injected into the flow tube from a glass vial with a zero air carrier flow (1 L min⁻¹). To generate ozone, 20 - 200 mL min⁻¹ zero air (60% RH at 20 °C) was irradiated by an amalgam lamp (185 and 254 nm; WISAG GmbH, Switzerland). Ozone was mixed with α-pinene to produce ozonolysis products which condensed onto, i.e. "coated" the NH₄NO₃ particles
- 160 inside the flow tube. Note that depending on the conditions, this coating may either result in a core-shell structure or in formation of a homogeneous single phase, though the exact morphology does not affect the conclusion regarding surface extraction, as discussed below. The coating period in the flow tube was approximately 26 ± 0.5 s. The coating thickness was controlled by varying the ozone concentration in the presence of excess α -pinene, which was measured by a Thermo 49A ozone analyzer (Thermo Fisher Scientific, US) to be 20 - 310 ppb. This ozone concentration range was optimized before
- 165 injecting the NH₄NO₃ particles to ensure that no nucleation occurred which would have resulted in particles consisting only of

SOA. At the beginning of each ozone concentration step, the EESI-TOF sampled the gas and aerosol mixture through a bypass channel without denuder to ensure that all oxidation product signals reached steady state (< 20 min). Afterwards, routine sampling alternated between filtered background (5 min) and particle-phase measurements (10 min). This coating experiment was carefully designed to achieve high condensational growth rates of about 0.8 nm s⁻¹ with negligible nucleation.

170 3 Results and discussion

175

3.1 Influence of particles size on EESI-TOF detection

Figure 1a shows a typical measurement of the EESI-TOF and SMPS for size-selected sucrose particles. Two sheath flow rates (5 and 15 with L min⁻¹) at 1.4 L min⁻¹ of particle flow rate were used to generate size-selected sucrose particle distribution with geometric standard deviation σ_g of 1.4 and 1.2, respectively. A comparison of the signals in the red windows in Figure 1b shows that the sucrose signal did not increase commensurately with the mass concentration measured by the SMPS (regardless of σ_g), as the volumetric geometric mean diameter of the particles increased. To quantify this effect, we define the size-dependent sensitivity $S(D_P)$ as

$$\mathbf{S}(D_P) = \frac{I(D_P)}{M(D_P)},\tag{1}$$

where $I(D_P)$ is the intensity of the analyte (Figure 2) or the total intensity of fitted organic ions (Figure 3a) that is normalized

180 by the most abundant electrospray ion ([NaI+Na]⁺) to account for the ES fluctuation (< 5%); $M(D_p)$ is the mass concentration of the particles measured by the SMPS or/and by the AMS-LTOF as a function of the volumetric geometric mean mobility diameter D_P . To show the relative change of the sensitivity as a function of D_P , $S(D_P)$ is normalized by its value at 100 nm electrical mobility diameter, defined as the normalized sensitivity, $S_{100 nm}$

$$S_{100 nm} = \frac{S(D_P)}{S(D_P = 100 nm)}$$
(2)

185 The normalization by the sensitivity at 100 nm, $S(D_p=100nn)$ was chosen to accommodate and compare all datasets in this study. The value of $S(D_p=100nn)$ was interpolated by fitting a 3-parameter function $S(D_P, P_1, P_2, P_3) = P_1 \cdot D_p \wedge (P_2) + P_3$ to the $S(D_p)$.

We investigated the normalized sensitivities of the EESI-TOF for levoglucosan, sucrose and NH_4NO_3 (tracers of biomass and anthropogenic activities in the ambient atmosphere) using different EESI ionization sources and ES operating parameters that

- resulted in different ES parent droplet diameters as estimated in Tables S1, S2 and S3. Figure 2 shows the normalized sensitivity of size-selected particles, *S*_{100nm} (Eq. 2) as a function of the volumetric geometric mean diameter of the particles generated using both pure component and mixed solutions detected under different ES conditions (see also Figure S6, Tables S1-S3). The *S*_{100 nm} for different types of particles decreased by up to 3 orders of magnitude as the volumetric geometric mean diameter increased from 30 to 300 nm, with some approaching a plateau at larger particle sizes. The size-dependent sensitivity is observed for both single compounds and compound mixtures (Figure S6).
- Assuming that the detected ions from the size-selected particles by EESI are generated after coagulation and extraction between the particles and ES droplets, the normalized sensitivity $S_{100 nm}$ should be proportional to the total coagulated mass. The determination of the total coagulated mass requires Brownian coagulation coefficients (*BCC*, computed from the particle and ES droplet sizes), number concentrations and residence time. However, the actual ES droplet size distribution could not be
- 200 measured using other physical processes because these additional processes could alter the ES droplets properties and affect the electrospray ionization. Therefore, we could only calculate the *BCC* for different size-selected monodisperse particles assuming ES parent droplet sizes of 0.5, 1.5 or 5 μm. These three chosen ES parent droplet sizes represent the likely range of the actual ES droplet sizes, which is theoretically estimated from our ES operating parameters as summarized in Table S2, based on SI Eq. 3-6 and Figure S7. The calculated *BCC* values were normalized to the *BCC* for 100 nm monodisperse particles,

- 205 denoted as $BCC_{100 nm}$, as shown in Figure S6a, analogous to the normalization for $S_{100 nm}$. Most normalized sensitivities (i.e. normalized total coagulated masses) correlate well with the $BCC_{100 nm}$, as shown in Figure 2. Smaller particles have higher BCC and are collected more efficiently, contributing a higher percentage of their total mass to the analyte-laden ES droplets during extraction. Furthermore, the plateaus of $S_{100 nm}$ at larger particle sizes could be explained by the suggested behavior of $BCC_{100 nm}$ when the size of the particle is similar to the actual ES droplet size or partly to the estimated ES parent droplet size
- 210 in our study. The high deviation of size-dependent sensitivity (~50 %) for $D_p > 200$ nm is likely due to the variation of the actual ES droplet size distribution in different calibration runs, which can deviate from the estimated ES parent droplet size. Knowledge of the actual ES droplet size distribution is needed to further explain the variabilities but are beyond the scope of the current study.

It is intuitive that the total coagulated mass for extraction is also dependent on the residence time for coagulation between the particles and the ES droplets during electrospray ionization. Longer residence time would allow for higher percentage of the

- 215 particles and the ES droplets during electrospray ionization. Longer residence time would allow for higher percentage of the particle total mass to be extracted, i.e. the coagulation of smaller particles would saturate, while the coagulation of larger particles would continue, which would result in a smaller range of size-dependent total coagulated mass, i.e. shallower sensitivity size dependence. We examined this hypothesis by using an EESI source B which provides a factor of 2 longer residence time in the electrospray ionization region. As shown in Figure 2, the sensitivity size dependence resulting from EESI
- 220 source B (yellow markers), which has twice the residence time as EESI source A, is significantly shallower than the one from EESI source A (blue markers), consistent with our hypothesis. Overall, Figure 2 suggests that the size-dependent sensitivity (total coagulated mass) is dependent on the Brownian coagulation coefficient, which varies with the ES droplet size (and therefore ES operating parameters), as well as the residence time for coagulation. Such size dependence suggests that the ionization of analyte particles in the EESI proceeds through coagulation at a certain size-dependent efficiency, e.g. partial
- 225 coalescence between particles and ES droplets, as reported by the previous studies (Wang et al., 2012; Kumbhani et al., 2018; Pagonis et al., 2020).

Konermann et al. (2013) reported that the electrospray droplet evaporation can be affected by the size and the polarity of analyte molecules, while Meier et al. (2011) suggested that the extraction efficiency of EESI can depend on the volatility of analyte molecules. We investigated the EESI sensitivity size dependence for a complex mixture of internally mixed α -pinene

- 230 oxidation products formed in the CLOUD chamber, to evaluate whether such dependence varies with analyte volatilities, e.g. if volatile species preferentially evaporate from smaller particles before their subsequent ionization. We generated unimodal size distributions of secondary organic aerosol (SOA) with volumetric geometric mean diameters ranging from 17 to 137 nm (Figure S8 and Figure S9). Figure 3a shows the normalized sensitivity of the sum of the organic ions measured by EESI-TOF after high-resolution peak fitting, *S*_{100nm}, as a function of the measured particle size. *S*_{100nm} decreases from a value of 6 at D_p =
- 17 nm to ~1 at $D_p = 110$ nm. The change in normalized sensitivity is similar to the results obtained for individual chemical standards presented in Figure 2 for EESI source A. To examine whether there is a composition dependence on the EESI extraction, we compared the signals of C₁₀H₁₆O₃₋₈ compounds measured by the EESI-TOF and the FIGAERO-CI-ToF-MS from SOA produced at different temperatures and RH in the CLOUD chamber as shown in Figure 3b (see also Figure S10). The linear behaviors of different measured species between the EESI-TOF and the FIGAERO-CI-TOF-MS for 17 < D_p < 80
- 240 nm in Figure 3b show that the relative abundances of the sampled aerosol chemical composition are similar and comparable for both instruments with negligible re-volatilization of particles at two different sampling points. Thermal decomposition may affect the absolute quantification of particle-phase compounds by the FIGAERO-CI-ToF-MS (Stark et al., 2017). However, to the best of our knowledge, no size dependence has been reported in the literature for this thermal artefact, which should be cancelled after sensitivity normalization comparison in relative scale for each species of FIGAERO-CI-ToF-MS. The
- sensitivity size dependence appears to be similar for $C_{10}H_{16}O_{3.8}$ compounds with estimated saturation vapor concentrations ranging from $10^{-8.6}$ to $10^{1.6}$ µg m⁻³. Both results from size-selected chemical standards (Figure 2) and chemical resolution comparison between EESI-TOF and FIGAERO-CI-ToF-MS using a complex SOA mixture indicate that the EESI sensitivity

size dependence is a function of the Brownian coagulation coefficient rather than molecular size, polarity, or volatility. Aside from the size dependence, we did not observe any systematic RH influence on the EESI sensitivity for size-selected chemical

250 standards (30 - 40 % RH) and α-pinene SOA (20% and 60% RH). This is consistent with the findings by Lopez-Hilfiker et al. (2019), where RH does not systematically affect EESI sensitivity, but instead shows molecule-dependent effects where, within an internally mixed particle, the sensitivity of certain molecules may increase with RH while others decrease. The enhancement in EESI sensitivity for wet aerosol over dry aerosol was reported in a previous study (Kumbhani et al., 2018). If EESI extraction is limited to the surface of the analyte aerosol, the aerosol water content may mobilize surface species to facilitate dissolution.

255 However, the lack of RH dependence for our EESI setup indicates that such surface extraction limitation is absent in our study.

260

3.2 Influence of particles coating thickness on EESI sensitivity

Limited surface extraction, approximately 2-4 nm in depth, of the particles was reported for some ESI source designs (Kumbhani et al., 2018; Wingen and Finlayson-Pitts, 2019). If such an effect were present in the EESI-TOF design used in the current study, it could also appear as a size-dependent sensitivity. This would mean that a smaller fraction of the analyte volume is extracted as the particle diameter increases, and that the EESI sensitivity scales with the particle surface area rather than the volume. To determine the potential contribution by surface extraction to the observed sensitivity size dependence, we investigated the extraction efficiency of NH_4NO_3 particles of 156 and 226 nm in diameter before being coated by the α -pinene oxidation products using source A. Source A was chosen because it has the greatest extent of size-dependent sensitivity in

- comparison to source B. The core NH₄NO₃ sizes were chosen as the size-dependent sensitivity decreases by less than 15 %
 from 155 nm to 250 nm for single and mixed component particles (Figure 2). The coating thickness on the NH₄NO₃ particles ranged between 12 and 26 nm, with a coated organic mass concentration up to 31 µg m⁻³ (Figure S11).
- If extraction were limited to the particle surface, the EESI signal for NH₄NO₃, i.e. [NaNO₃+Na]⁺, should decrease similar to the size-dependent sensitivity (Figure 2) that is exhibited by the source A. For instance, if the coated particles were of coreshell morphology, then the extraction of the NH₄NO₃ core would be limited by the thickness of the organic coating and the ES extraction depth. Alternatively, if the coated particles were homogeneous inorganic-organic mixtures, then the detected
- NH_4NO_3 signal would still decrease in proportion to the decreasing NH_4NO_3 mass fraction as the condensed organic mass increases. In Figure 4, we show the signals of NH_4NO_3 and selected organic molecules with low volatility as a function of the coating volume (normalized to their respective minimum coating volume separately for each of the NH_4NO_3 particle core sizes). The coating signal from $C_{10}H_{16}O_{6-8}$ is proportional to the coating volume, whereas the NH_4NO_3 particle signal remains
- 275 constant with increasing coating thickness for both core sizes (see also Figure S12). This proportionality also demonstrates that the condensable species as coating substance is not limited by the mean oxidation states of oxidation products because there is no decrease of the $C_{10}H_{16}O_6$ for an increase of $C_{10}H_{16}O_8$. Our results suggest that there is no surface extraction limitation for particles up to at least 250 nm in diameter for the EESI inlet designs used in the current study. Prior reporting of surface extraction limitation may stem from the specific EESI configuration or experimental method used, which relied on the 280 comparison of EESI and ESI measurements (Kumbhani et al., 2018), while the differences in dissolution/extraction timescale
- and sample preparation between EESI and ESI techniques could contribute to the discrepancies observed.

4 Conclusion

285

We explored the dependence of EESI sensitivity on particle size using individual chemical standards and chemical mixtures with two different EESI source designs. We show that the EESI sensitivity decreases as the size of the particles increases. The sensitivity size dependence correlates with the Brownian coagulation coefficient and the residence time for coagulation. The results suggest that the particles undergo coalescence with the ES charged droplets as suggested in previous studies (Law et al., 2010; Wang et al., 2012), but the efficiency of the coalescence is limited by the coagulation coefficient, which depends on

the particle and ES charged droplet sizes. From a comparison with the FIGAERO-CI-ToF-MS online measurements, we show that the EESI sensitivity size dependence is also present for internally mixed secondary organic aerosol made of molecules

- 290 with volatilities varying by approximately 10 orders of magnitudes. While the total extracted mass is related to the sizedependent Brownian coagulation coefficients (i.e. not all particles of different size can coalesce with all the electrospray droplets), coating experiments show that the volume of particles, once coagulated with the ES droplet, is fully extracted up to a size of 250 nm for our EESI configuration instead of limited surface extraction reported by the previous work (Kumbhani et al., 2018). Future work should investigate the EESI response to coarse mode particles (with $D_p > 1 \mu m$), elucidate the
- 295 relationship between size-dependent sensitivity behavior and different chemical mixtures, and search for an optimal residence time for coagulation in the EESI source to achieve the least steep size-dependent sensitivity for the particle size range of interest. EESI users should be cognizant of the size-dependent sensitivity during their experiment design and data analysis. Such size dependence is especially relevant when studying aerosol formation and growth, or when studying external mixtures of particles of distinct sizes. However, such effect is not expected to substantially influence the detection of ambient aerosols
- 300 dominated by well-mixed accumulation mode particles.

Data availability. Data presented in this study can be obtained at the Zenodo online repository hosted by CERN (https://doi.org/10.5281/zenodo.5094651, Lee et al., 2021). Raw data can be obtained from the corresponding authors upon reasonable request.

305

Author contributions. C.P.L. and M.S. designed the experiment. M.S., C.P.L., D.W., H.L, M.W., F.A., V. H., B. L., performed the experiments. C.P.L., M.S., M.W., F.A. analyzed the data. C.P.L., I.E., M.S., D.W., H. L., J. D., J.S. and U.B. interpreted the compiled results. C.P.L. prepared the manuscript. All authors contributed to the discussion and revision of the manuscript.

310 Competing interests. The authors declare that they have no conflict of interest.

Acknowledgements. We thank our technician Pascal Andre Schneider for technical support throughout our experiments. We thank Martin Gysel for the scientific discussion. Special thanks to the CLOUD collaboration and CERN facilities for providing us the possibilities and resources to realize our investigation.

315

Financial support. This work was financially supported by the Swiss National Science Foundation (20020_172602, BSSGI0_155846), the European Union's Horizon 2020 research and innovation program under the Marie Skłodowska-Curie actions (No. 701647, No. 764991), US National Science Foundation (AGS1801574, AGS1531284) and EUROCHAMP-2020 projects and infrastructure (No. 730997).




Figure 1. (a) Number-weighted size distribution of sucrose particles as measured by the SMPS after size selection using the AAC at two different settings of the AAC sheath flow. The solid and dotted black lines denote the geometric mean and ±1 geometric standard deviation, σ_g, of the number-weighted size distributions, respectively. The geometric standard deviation of the size-selected particle distribution of sucrose is lower at the higher sheath flow of the AAC as expected. Data points of particle counts lower than 1 cm⁻³
were removed. See Figure S5 for the size-selection performance of the AAC as measured by the SMPS. (b) A representative EESI-TOF measurement that shows the time series of sucrose normalized to the [NaI+Na]⁺ signal (most abundant ES ion) and the corresponding integrated particle mass concentration measured by the SMPS for size-selected sucrose particles (using the integrated volume concentration and a density of 1.59 g cm⁻³). Red windows indicate periods with the similar EESI signal intensity of sucrose but different particle size and mass concentration.




Figure 2. Sensitivities of EESI-TOF towards various standards normalized to their respective values at 100 nm as a function of the particle volumetric geometric mean diameter. Blue and yellow markers indicate EESI sources A and B which were initially developed for TOF and Orbitrap mass analyzers, respectively (Lopez-Hilfiker et al., 2019; Lee et al., 2020). Different marker types (●), (■), (◆) denote levoglucosan, NH₄NO₃ and sucrose, respectively. The Brownian coagulation coefficients are calculated using the range of ES parent droplet sizes estimated based on our ES operating parameters (Figures S6 and S7). Note: Some of the data points may overlap at the similar volumetric geometric mean diameter due to repetitions of the same experiment settings.




Figure 3. (a) Normalized sensitivity of EESI-TOF calculated from the sum of high-resolution fitted organic ion signals and the total particle mass concentrations as a function of the volumetric geometric mean diameter over the course of new particle formation and growth of α -pinene (AP) ozonolysis SOA. See Figure S8, Figure S9 and Figure S10 for more information. (b) Normalized sensitivity of EESI-TOF intensity divided by the FIGAERO-CI-ToF-MS intensity for C₁₀H₁₆O₃₋₈ molecules in the particle phase as a function of the volumetric geometric mean diameter. Different marker types \Box , *, O, + indicate different SOA formation conditions, i.e. at -30 °C and 20% RH, -50 °C and 20% RH, -50 °C and 60% RH, -30 °C and 60% RH, respectively. The saturation concentration was estimated as $log_{10}(C^0) = (n_c^0 - n_c)b_c - n_0b_0 - 2 \cdot b_{c0}(n_cn_0)/(n_c + n_0)$ from the number of carbon and oxygen (n_c, n_0) , and coefficients (b_c, b_0, b_{c0}) provided in Donahue et al. (2011).






Figure 4. Relative changes of α -pinene ozonolysis products coated on NH4NO3 particles at 156 (\bigcirc) and 226 (\Box) nm core sizes. The coating volume (*x*-axis) measured by an SMPS is normalized by the smallest coating volume, and the coating signals (*y*-axis) of C₁₀H₁₆O₆₋₈ molecules as measured by the EESI-TOF are normalized by the signals observed at the smallest coating volume. The largest coating thicknesses are 19.8 and 26.8 nm for 156 and 226 nm core sizes of the NH4NO3 particles, respectively. The black-dotted line denotes a relative change of the coating signal and volume equal to 1.

References

400

Berndt, T., Böge, O., Stratmann, F., Heintzenberg, J. and Kulmala, M.: Rapid formation of sulfuric acid particles at nearatmospheric conditions, Science, 307(5710), 698–700, doi:10.1126/science.1104054, 2005.

Burnett, R., Chen, H., Szyszkowicz, M., Fann, N., Hubbell, B., Pope, C. A., Apte, J. S., Brauer, M., Cohen, A., Weichenthal,

- 360 S., Coggins, J., Di, Q., Brunekreef, B., Frostad, J., Lim, S. S., Kan, H., Walker, K. D., Thurston, G. D., Hayes, R. B., Lim, C. C., Turner, M. C., Jerrett, M., Krewski, D., Gapstur, S. M., Diver, W. R., Ostro, B., Goldberg, D., Crouse, D. L., Martin, R. V., Peters, P., Pinault, L., Tjepkema, M., Van Donkelaar, A., Villeneuve, P. J., Miller, A. B., Yin, P., Zhou, M., Wang, L., Janssen, N. A. H., Marra, M., Atkinson, R. W., Tsang, H., Thach, T. Q., Cannon, J. B., Allen, R. T., Hart, J. E., Laden, F., Cesaroni, G., Forastiere, F., Weinmayr, G., Jaensch, A., Nagel, G., Concin, H. and Spadaro, J. V.: Global estimates of mortality
- 365 associated with longterm exposure to outdoor fine particulate matter, Proc. Natl. Acad. Sci. U. S. A., 115(38), 9592–9597, doi:10.1073/pnas.1803222115, 2018.

Chen, H., Venter, A. and Cooks, R. G.: Extractive electrospray ionization for direct analysis of undiluted urine, milk and other complex mixtures without sample preparation, Chem. Commun., 0(19), 2042–2044, doi:10.1039/b602614a, 2006.

Cheng, C. Y., Yuan, C. H., Cheng, S. C., Huang, M. Z., Chang, H. C., Cheng, T. L., Yeh, C. S. and Shiea, J.: Electrosprayassisted laser desorption/ionization mass spectrometry for continuously monitoring the states of ongoing chemical reactions in organic or aqueous solution under ambient conditions, Anal. Chem., 80(20), 7699–7705, doi:10.1021/ac800952e, 2008.

Clarke, A. G., Willison, M. J. and Zeki, E. M.: Aerosol Neutralization by Atmospheric Ammonia, in Physico-Chemical Behaviour of Atmospheric Pollutants, pp. 331–338, Springer Netherlands., 1984.

Daellenbach, K. R., Uzu, G., Jiang, J., Cassagnes, L. E., Leni, Z., Vlachou, A., Stefenelli, G., Canonaco, F., Weber, S., Segers,
A., Kuenen, J. J. P., Schaap, M., Favez, O., Albinet, A., Aksoyoglu, S., Dommen, J., Baltensperger, U., Geiser, M., El Haddad,
I., Jaffrezo, J. L. and Prévôt, A. S. H.: Sources of particulate-matter air pollution and its oxidative potential in Europe, Nature, 587(7834), 414–419, doi:10.1038/s41586-020-2902-8, 2020.

Dias, A., Ehrhart, S., Vogel, A., Williamson, C., Almeida, J., Kirkby, J., Mathot, S., Mumford, S. and Onnela, A.: Temperature uniformity in the CERN CLOUD chamber, Atmos. Meas. Tech., 10(12), 5075–5088, doi:10.5194/amt-10-5075-2017, 2017.

380 Ditto, J. C., Joo, T., Slade, J. H., Shepson, P. B., Ng, N. L. and Gentner, D. R.: Nontargeted Tandem Mass Spectrometry Analysis Reveals Diversity and Variability in Aerosol Functional Groups across Multiple Sites, Seasons, and Times of Day, Environ. Sci. Technol. Lett., 7(2), 60–69, doi:10.1021/acs.estlett.9b00702, 2020.

Donahue, N. M., Epstein, S. A., Pandis, S. N. and Robinson, A. L.: A two-dimensional volatility basis set: 1. organic-aerosol mixing thermodynamics, Atmos. Chem. Phys., 11(7), 3303–3318, doi:10.5194/acp-11-3303-2011, 2011.

385 Eichler, P., Müller, M., D'Anna, B. and Wisthaler, A.: A novel inlet system for online chemical analysis of semi-volatile submicron particulate matter, Atmos. Meas. Tech., 8(3), 1353–1360, doi:10.5194/amt-8-1353-2015, 2015.

George, I. J. and Abbatt, J. P. D.: Heterogeneous oxidation of atmospheric aerosol particles by gas-phase radicals, Nat. Chem., 2(9), 713–722, doi:10.1038/nchem.806, 2010.

 Giannoukos, S., Lee, C. P., Tarik, M., Ludwig, C., Biollaz, S., Lamkaddam, H., Baltensperger, U., Henry Prevot, A. S. and
 Slowik, J.: Real-Time Detection of Aerosol Metals Using Online Extractive Electrospray Ionization Mass Spectrometry, Anal. Chem., 92(1), 1316–1325, doi:10.1021/acs.analchem.9b04480, 2020.

Hoffmann, T., Odum, J. R., Bowman, F., Collins, D., Klockow, D., Flagan, R. C. and Seinfeld, J. H.: Formation of organic aerosols from the oxidation of biogenic hydrocarbons, J. Atmos. Chem., 26(2), 189–222, doi:10.1023/A:1005734301837, 1997.

395 Holzinger, R., Williams, J., Herrmann, F., Lelieveld, J., Donahue, N. M. and Röckmann, T.: Aerosol analysis using a Thermal-Desorption Proton-Transfer-Reaction Mass Spectrometer (TD-PTR-MS): A new approach to study processing of organic aerosols., 2010.

Jimenez, J. L., Canagaratna, M. R., Donahue, N. M., Prevot, A. S. H., Zhang, Q., Kroll, J. H., DeCarlo, P. F., Allan, J. D., Coe, H., Ng, N. L., Aiken, A. C., Docherty, K. S., Ulbrich, I. M., Grieshop, A. P., Robinson, A. L., Duplissy, J., Smith, J. D., Wilson, K. R., Lanz, V. A., Hueglin, C., Sun, Y. L., Tian, J., Laaksonen, A., Raatikainen, T., Rautiainen, J., Vaattovaara, P.,

Ehn, M., Kulmala, M., Tomlinson, J. M., Collins, D. R., Cubison, M. J., Dunlea, E. J., Huffman, J. A., Onasch, T. B., Alfarra, M. R., Williams, P. I., Bower, K., Kondo, Y., Schneider, J., Drewnick, F., Borrmann, S., Weimer, S., Demerjian, K., Salcedo, D., Cottrell, L., Griffin, R., Takami, A., Miyoshi, T., Hatakeyama, S., Shimono, A., Sun, J. Y., Zhang, Y. M., Dzepina, K., Kimmel, J. R., Sueper, D., Jayne, J. T., Herndon, S. C., Trimborn, A. M., Williams, L. R., Wood, E. C., Middlebrook, A. M.,

405 Kolb, C. E., Baltensperger, U. and Worsnop, D. R.: Evolution of organic aerosols in the atmosphere, Science, 326(5959), 1525–1529, doi:10.1126/science.1180353, 2009.

Kalberer, M., Paulsen, D., Sax, M., Steinbacher, M., Dommen, J., Prevot, A. S. H., Fisseha, R., Weingartner, E., Frankevich, V., Zenobi, R. and Baltensperger, U.: Identification of Polymers as Major Components of Atmospheric Organic Aerosols, Science, 303(5664), 1659–1662, doi:10.1126/science.1092185, 2004.

410 Kebarle, P. and Verkcerk, U. H.: Electrospray: From Ions in solution to Ions in the gas phase, what we know now, Mass Spectrom. Rev., 28(6), 898–917, doi:10.1002/mas.20247, 2009.

Kebarle, P. and Verkerk, U. H.: On the Mechanism of Electrospray Ionization Mass Spectrometry (ESIMS), Electrospray MALDI Mass Spectrom. Fundam. Instrumentation, Pract. Biol. Appl. Second Ed., 82(11), 1–48, doi:10.1002/9780470588901.ch1, 2012.

- 415 Kirkby, J., Curtius, J., Almeida, J., Dunne, E., Duplissy, J., Ehrhart, S., Franchin, A., Gagné, S., Ickes, L., Kürten, A., Kupc, A., Metzger, A., Riccobono, F., Rondo, L., Schobesberger, S., Tsagkogeorgas, G., Wimmer, D., Amorim, A., Bianchi, F., Breitenlechner, M., David, A., Dommen, J., Downard, A., Ehn, M., Flagan, R. C., Haider, S., Hansel, A., Hauser, D., Jud, W., Junninen, H., Kreissl, F., Kvashin, A., Laaksonen, A., Lehtipalo, K., Lima, J., Lovejoy, E. R., Makhmutov, V., Mathot, S., Mikkilä, J., Minginette, P., Mogo, S., Nieminen, T., Onnela, A., Pereira, P., Petäjä, T., Schnitzhofer, R., Seinfeld, J. H., Sipilä,
- 420 M., Stozhkov, Y., Stratmann, F., Tomé, A., Vanhanen, J., Viisanen, Y., Vrtala, A., Wagner, P. E., Walther, H., Weingartner, E., Wex, H., Winkler, P. M., Carslaw, K. S., Worsnop, D. R., Baltensperger, U. and Kulmala, M.: Role of sulphuric acid, ammonia and galactic cosmic rays in atmospheric aerosol nucleation, Nature, 476(7361), 429–435, doi:10.1038/nature10343, 2011.
 - Kirkby, J., Duplissy, J., Sengupta, K., Frege, C., Gordon, H., Williamson, C., Heinritzi, M., Simon, M., Yan, C., Almeida, J.,
- 425 Trostl, J., Nieminen, T., Ortega, I. K., Wagner, R., Adamov, A., Amorim, A., Bernhammer, A. K., Bianchi, F., Breitenlechner, M., Brilke, S., Chen, X., Craven, J., Dias, A., Ehrhart, S., Flagan, R. C., Franchin, A., Fuchs, C., Guida, R., Hakala, J., Hoyle, C. R., Jokinen, T., Junninen, H., Kangasluoma, J., Kim, J., Krapf, M., Kurten, A., Laaksonen, A., Lehtipalo, K., Makhmutov, V., Mathot, S., Molteni, U., Onnela, A., Perakyla, O., Piel, F., Petaja, T., Praplan, A. P., Pringle, K., Rap, A., Richards, N. A. D., Riipinen, I., Rissanen, M. P., Rondo, L., Sarnela, N., Schobesberger, S., Scott, C. E., Seinfeld, J. H., Sipila, M., Steiner,
- 430 G., Stozhkov, Y., Stratmann, F., Tomé, A., Virtanen, A., Vogel, A. L., Wagner, A. C., Wagner, P. E., Weingartner, E., Wimmer, D., Winkler, P. M., Ye, P., Zhang, X., Hansel, A., Dommen, J., Donahue, N. M., Worsnop, D. R., Baltensperger, U., Kulmala, M., Carslaw, K. S. and Curtius, J.: Ion-induced nucleation of pure biogenic particles, Nature, 533(7604), 521–526, doi:10.1038/nature17953, 2016.

Konermann, L., Ahadi, E., Rodriguez, A. D. and Vahidi, S.: Unraveling the mechanism of electrospray ionization, Anal. 435 Chem., 85(1), 2–9, doi:10.1021/ac302789c, 2013.

Kumbhani, S., Longin, T., Wingen, L. M., Kidd, C., Perraud, V. and Finlayson-Pitts, B. J.: New Mechanism of Extractive Electrospray Ionization Mass Spectrometry for Heterogeneous Solid Particles, Anal. Chem., 90(3), 2055–2062, doi:10.1021/acs.analchem.7b04164, 2018.

Law, W. S., Wang, R., Hu, B., Berchtold, C., Meier, L., Chen, H. and Zenobi, R.: On the mechanism of extractive electrospray ionization, Anal. Chem., 82(11), 4494–4500, doi:10.1021/ac100390t, 2010.

Lee, C. P., Riva, M., Wang, D., Tomaz, S., Li, D., Perrier, S., Slowik, J. G., Bourgain, F., Schmale, J., Prevot, A. S. H., Baltensperger, U., George, C. and El Haddad, I.: Online Aerosol Chemical Characterization by Extractive Electrospray Ionization-Ultrahigh-Resolution Mass Spectrometry (EESI-Orbitrap), Environ. Sci. Technol., 54(7), 3871–3880, doi:10.1021/acs.est.9b07090, 2020.

445 Lopez-Hilfiker, Mohr, C., Ehn, M., Rubach, F., Kleist, E., Wildt, J., Mentel, T. F., Lutz, A., Hallquist, M., Worsnop, D. and Thornton, J. A.: A novel method for online analysis of gas and particle composition: Description and evaluation of a filter inlet for gases and AEROsols (FIGAERO), Atmos. Meas. Tech., 7(4), 983–1001, doi:10.5194/amt-7-983-2014, 2014.

Lopez-Hilfiker, Pospisilova, V., Huang, W., Kalberer, M., Mohr, C., Stefenelli, G., Thornton, J. A., Baltensperger, U., Prevot,

A. S. H. and Slowik, J. G.: An extractive electrospray ionization time-of-flight mass spectrometer (EESI-TOF) for online
 measurement of atmospheric aerosol particles, Atmos. Meas. Tech., 12(9), 4867–4886, doi:10.5194/amt-12-4867-2019, 2019.

Marquez, C. A., Wang, H., Fabbretti, F. and Metzger, J. O.: Electron-transfer-catalyzed dimerization of trans-anethole: Detection of the distonic tetramethylene radical cation intermediate by extractive electrospray ionization mass spectrometry, J. Am. Chem. Soc., 130(51), 17208–17209, doi:10.1021/ja806791c, 2008.

Meier, L., Schmid, S., Berchtold, C. and Zenobi, R.: Contribution of liquid-phase and gas-phase ionization in extractive
electrospray ionization mass spectrometry of primary amines, Eur. J. Mass Spectrom., 17(4), 345–351, doi:10.1255/ejms.1146, 2011.

Molteni, U., Bianchi, F., Klein, F., El Haddad, I., Frege, C., Rossi, M. J., Dommen, J. and Baltensperger, U.: Formation of highly oxygenated organic molecules from aromatic compounds, Atmos. Chem. Phys., 18(3), 1909–1921, doi:10.5194/acp-18-1909-2018, 2018.

460 Müller, M., Eichler, P., D'Anna, B., Tan, W. and Wisthaler, A.: Direct Sampling and Analysis of Atmospheric Particulate Organic Matter by Proton-Transfer-Reaction Mass Spectrometry, Anal. Chem., 89(20), 10889–10897, doi:10.1021/acs.analchem.7b02582, 2017.

Pagonis, D., Campuzano-Jost, P., Guo, H., Day, D. A., Schueneman, M. K., Brown, W. L., Nault, B. A., Stark, H., Siemens, K., Laskin, A., Piel, F., Tomsche, L., Wisthaler, A., Coggon, M. M., Gkatzelis, G. I., Halliday, H. S., Krechmer, J. E., Moore,

465 R. H., Thomson, D. S., Warneke, C., Wiggins, E. B. and Jimenez, J. L.: Airborne extractive electrospray mass spectrometry measurements of the chemical composition of organic aerosol, Atmos. Meas. Tech., 14(2), 1545–1559, doi:10.5194/amt-14-1545-2021, 2021.

Seinfeld, J. H. and Pandis, S. N.: Atmospheric Chemistry and Physics: From Air Pollution to Climate Change, 3rd ed., Wiley., 2016.

- 470 Simon, M., Dada, L., Heinritzi, M., Scholz, W., Stolzenburg, D., Fischer, L., Wagner, A., Kürten, A., Rörup, B., He, X.-C., Almeida, J., Baalbaki, R., Baccarini, A., Bauer, P., Beck, L., Bergen, A., Bianchi, F., Bräkling, S., Brilke, S., Caudillo, L., Chen, D., Chu, B., Dias, A., Draper, D., Duplissy, J., El Haddad, I., Finkenzeller, H., Frege, C., Gonzalez-Carracedo, L., Gordon, H., Granzin, M., Hakala, J., Hofbauer, V., Hoyle, C., Kim, C., Kong, W., Lamkaddam, H., Lee, C., Lehtipalo, K., Leiminger, M., Mai, H., Manninen, H., Marie, G., Marten, R., Mentler, B., Molteni, U., Nichman, L., Nie, W., Ojdanic, A.,
 475 Onnela, A., Partoll, F., Petäjä, T., Pfaifer, J., Philippov, M., Ouéléver, L., Baniithkumar, A., Pissanen, M., Schallhart, S.,
- Onnela, A., Partoll, E., Petäjä, T., Pfeifer, J., Philippov, M., Quéléver, L., Ranjithkumar, A., Rissanen, M., Schallhart, S., Schobesberger, S., Schuchmann, S., Shen, J., Sipilä, M., Steiner, G., Stozhkov, Y., Tauber, C., Tham, Y., Tomé, A., Vazquez-Pufleau, M., Vogel, A., Wagner, R., Wang, M., Wang, D., Wang, Y., Weber, S., Wu, Y., Xiao, M., Yan, C., Ye, P., Ye, Q., Zauner-Wieczorek, M., Zhou, X., Baltensperger, U., Dommen, J., Flagan, R., Hansel, A., Kulmala, M., Volkamer, R., Winkler, P., Worsnop, D., Donahue, N., Kirkby, J. and Curtius, J.: Molecular understanding of new-particle formation from alpha-pinene between -50 °C and 25 °C, Atmos. Chem. Phys., 40(January), 1–42, doi:10.5194/acp-2019-1058, 2020.

Stark, H., Yatavelli, R. L. N., Thompson, S. L., Kang, H., Krechmer, J. E., Kimmel, J. R., Palm, B. B., Hu, W., Hayes, P. L., Day, D. A., Campuzano-Jost, P., Canagaratna, M. R., Jayne, J. T., Worsnop, D. R. and Jimenez, J. L.: Impact of Thermal Decomposition on Thermal Desorption Instruments: Advantage of Thermogram Analysis for Quantifying Volatility Distributions of Organic Species, Environ. Sci. Technol., 51(15), 8491–8500, doi:10.1021/acs.est.7b00160, 2017.

485 Tavakoli, F. and Olfert, J. S.: An instrument for the classification of aerosols by particle relaxation time: Theoretical models of the aerodynamic aerosol classifier, Aerosol Sci. Technol., 47(8), 916–926, doi:10.1080/02786826.2013.802761, 2013.

Tavakoli, F. and Olfert, J. S.: Determination of particle mass, effective density, mass-mobility exponent, and dynamic shape factor using an aerodynamic aerosol classifier and a differential mobility analyzer in tandem, J. Aerosol Sci., 75, 35–42, doi:10.1016/j.jaerosci.2014.04.010, 2014.

490 Tavakoli, F., Symonds, J. P. R. and Olfert, J. S.: Generation of a monodisperse size-classified aerosol independent of particle charge, Aerosol Sci. Technol., 48(3), i–iv, doi:10.1080/02786826.2013.877121, 2014.

Tröstl, J., Chuang, W. K., Gordon, H., Heinritzi, M., Yan, C., Molteni, U., Ahlm, L., Frege, C., Bianchi, F., Wagner, R., Simon, M., Lehtipalo, K., Williamson, C., Craven, J. S., Duplissy, J., Adamov, A., Almeida, J., Bernhammer, A. K., Breitenlechner, M., Brilke, S., Dias, A., Ehrhart, S., Flagan, R. C., Franchin, A., Fuchs, C., Guida, R., Gysel, M., Hansel, A., Hoyle, C. R.,

495 Jokinen, T., Junninen, H., Kangasluoma, J., Keskinen, H., Kim, J., Krapf, M., Kürten, A., Laaksonen, A., Lawler, M., Leiminger, M., Mathot, S., Möhler, O., Nieminen, T., Onnela, A., Petäjä, T., Piel, F. M., Miettinen, P., Rissanen, M. P., Rondo,

L., Sarnela, N., Schobesberger, S., Sengupta, K., Sipilä, M., Smith, J. N., Steiner, G., Tomè, A., Virtanen, A., Wagner, A. C., Weingartner, E., Wimmer, D., Winkler, P. M., Ye, P., Carslaw, K. S., Curtius, J., Dommen, J., Kirkby, J., Kulmala, M., Riipinen, I., Worsnop, D. R., Donahue, N. M. and Baltensperger, U.: The role of low-volatility organic compounds in initial particle growth in the atmosphere, Nature, 533(7604), 527–531, doi:10.1038/nature18271, 2016.

Wagner, A. C., Bergen, A., Brilke, S., Fuchs, C., Ernst, M., Hoker, J., Heinritzi, M., Simon, M., Bühner, B., Curtius, J. and Kürten, A.: Size-resolved online chemical analysis of nanoaerosol particles: A thermal desorption differential mobility analyzer coupled to a chemical ionization time-of-flight mass spectrometer, Atmos. Meas. Tech., 11(10), 5489–5506, doi:10.5194/amt-11-5489-2018, 2018.

500

505 Wang, R., Gröhn, A. J., Zhu, L., Dietiker, R., Wegner, K., Günther, D. and Zenobi, R.: On the mechanism of extractive electrospray ionization (EESI) in the dual-spray configuration, Anal. Bioanal. Chem., 402(8), 2633–2643, doi:10.1007/s00216-011-5471-8, 2012.

Wingen, L. M. and Finlayson-Pitts, B. J.: Probing surfaces of atmospherically relevant organic particles by easy ambient sonic-spray ionization mass spectrometry (EASI-MS), Chem. Sci., 10(3), 884–897, doi:10.1039/c8sc03851a, 2019.

510 Yue, D., Hu, M., Wu, Z., Wang, Z., Guo, S., Wehner, B., Nowak, A., Achtert, P., Wiedensohler, A., Jung, J., Kim, Y. J. and Liu, S.: Characteristics of aerosol size distributions and new particle formation in the summer in Beijing, J. Geophys. Res. Atmos., 114(14), doi:10.1029/2008JD010894, 2009.