

Referee #1

General Comment: The paper describes the application of the in house developed upper air simulator (UAS) to investigate the radiation temperature error of the Vaisala RS41 radiosonde. The UAS is used to simulate the conditions, both radiation and ambient temperature, that are encountered in the upper atmosphere during a radiosounding. The main finding of the paper is that lowering the air temperature in the UAS from +20 to -67C leads to an increase of approximately 20% of the radiation temperature error, which the authors postulate is due to changes in the efficiency of the -radiative and conductive- energy transfer mechanisms. This work is very relevant, because it reports on a manufacturer-independent and traceable method to characterize and quantify error sources in radiosonde temperature measurements. This is an essential prerequisite for providing reference quality data that is suitable for e.g. climate monitoring, especially since the corrections applied by Vaisala in the processing of their radiosonde data are not disclosed to the user. However, several improvements to the manuscript are necessary to make it suitable for publication.

→ We thank the Reviewer for valuable comments. We have revised the manuscript to comply with the Reviewer's comment as below.

Major comments: My overall impression is that the manuscript is lacking in interpretation of the findings. The method and the results are presented all right, but there is no real attempt made to come to a deeper understanding of the results. The main message of the paper is the essential role of the ambient temperature during the measurements, and that lowering the air temperature from +20 to -67C leads to an increase of approximately 20% of the observed radiation temperature error. This is a considerable effect, with potentially far-reaching consequences for the uncertainty estimates of radiosonde data products. Therefore this observed temperature effect requires more elaborate discussion than that is currently included in the manuscript, and the authors should make an attempt to provide a quantifiable physical justification and/or explanation for this observed temperature-dependent difference.

→ We have added theoretical discussion on the temperature effect in Section 3.2 with **Figure 3** and **Table 2**.

Before: This phenomenon likely occurs because the long-wave radiation from the temperature sensor reduces as the absolute temperature decreases. In addition, the cooling of the sensor may become less effective as the thermal conductivity of air decreases at low temperatures.

After (Line 177-201): In order to understand the observed temperature effect theoretically, the temperature sensor is modelled as a sphere made of Platinum (Pt) with a diameter (D) of 1 mm. The Pt sphere is placed in the middle of an air flow (v) with varied temperature (T_a) and pressure (P_a) as shown in Fig. 3(a). The sphere is heated by the absorption of the solar irradiance ($S = 1000 \text{ W} \cdot \text{m}^{-2}$) and cooled by the forced air convection ($5 \text{ m} \cdot \text{s}^{-1}$) similarly with the experiment. The radial and angular temperature distribution of the sphere is neglected and assumed to be

uniform. Then, the steady state temperature of the sphere (T_s) is simply decided by the energy balance of the heat transfer exchange as follows:

$$\alpha S = h(T_s - T_a) \quad \text{with} \quad h = \frac{k}{D} \left[2 + (0.4Re^{\frac{1}{2}} + 0.06Re^{\frac{2}{3}}) \left(\frac{\mu C_p}{k} \right)^{\frac{2}{5}} \right] \quad (5)$$

where α is the absorptivity of the metal sphere, S is the solar irradiance and h is the heat transfer coefficient (Incropera and Dewitt, 2002; Luers and Eskridge, 1995). The net heat transfer by longwave radiation from the Pt sphere is not considered because it is negligible ($\sim 10^{-6}$ W) compared to that by the convective heat transfer in Eq. (5). The heat transfer coefficient h is determined by several parameters concerning the diameter of the sphere (D) and the properties of air including thermal conductivity (k), viscosity (μ), heat capacity (C_p) and Reynolds number ($Re = \rho v D / \mu$), in which ρ and v is the density of air and wind speed, respectively.

The radiation correction ($T_s - T_a$) at $T_a = 20$ °C and -70 °C is calculated by Eq. (5) and displayed together with the experimental values as shown in Fig. 3(b). The properties of air (N₂) used for the calculation refer to the NIST Chemistry WebBook (Linstrom and Mallard, 2001). In general, the calculated radiation correction of the Pt sphere is elevated as the pressure decreases as in the case of the experiment. This is because the heat transfer coefficient is reduced by about 35 % as the density of air is decreased with varying pressure from $P_a = 50$ hPa to 5 hPa. Interestingly, the temperature effect on the calculated radiation correction is also observed similarly with the experiment. A decrease of thermal conductivity of air by about 26 % at -70 °C is mainly responsible for the decrease of the heat transfer coefficient and thereby the increase of the radiation correction at low temperature (-70 °C). The thermal conductivity of air plays an important role for the heat transfer at the boundary between the air and the Pt sphere. The same phenomenon was previously observed for thermistors even though there is no apparent air ventilation (Lee et al., 2018a) which may emphasize the role of thermal conductivity of air. The parameters and their values used for the calculation of radiation correction at $T_a = 20$ °C and -70 °C with $P_a = 5$ hPa is summarized in **Table 2**.

Table 2. Parameters and their values in Eq. (5) and the calculation of radiation correction at 20 °C and -70 °C.

Parameter	Symbol (Unit)	Value ($T_a = 20$ °C)	Value ($T_a = -70$ °C)
Diameter	D (m)	0.001	0.001
Air pressure	P_a (hPa)	5	5
Wind speed	v (ms ⁻¹)	5	5
Viscosity	μ (Pa·s)	0.00001754	0.00001307
Density	ρ (kg·m ⁻³)	0.0057466	0.0082925
Thermal conductivity	k (W·m ⁻¹ ·K ⁻¹)	0.025367	0.018869
Heat capacity	C_p (J·kg ⁻¹ ·K ⁻¹)	1039.6	1039.1
Reynolds number	Re	1.64	3.17

Heat transfer coefficient	$h \text{ (W} \cdot \text{m}^{-2} \cdot \text{K}^{-1})$	63.97	51.67
Solar irradiance	$S \text{ (W} \cdot \text{m}^{-2})$	1000	1000
Absorptivity of metal	α	0.2	0.2
Radiation correction	$T_s - T_a \text{ (K)}$	0.78	0.97

Added Statement (Line 231-233): Since the temperature dependency is weak at higher pressures, there is no need to estimate the low temperature effect at 50–500 hPa and the estimation using Eq. (6) is limited within 5–50 hPa.

There should be a clear separation between the description of the measurements and the presentation of the results. Currently these are entangled in section 3. I recommend presenting the description and the results in different sections. Furthermore, provided a more extensive discussion of the assumptions that are made in the data analysis, together with their effect on the quality or uncertainty of the correction.

→ We prefer the current format of the manuscript to explain the effect of each parameter on the radiation correction one by one and show how the correction formula is updated to incorporate each effect.

With regard to the representativeness of the conditions in the UAS compared to the real atmosphere: there are significant differences between the long wave radiation environment in the laboratory setup and in the free atmosphere. In the laboratory setup there is a uniform background emitted by the walls of the measurement chamber, with only a small temperature difference (approximately 1-2K) with respect to the temperature sensor. In the free atmosphere the long-wave radiation background is composed of, amongst others, contributions from the air masses and surface below and from the cold cosmic background above. Therefore, the question is how well the environment inside the UAS is representative for the conditions encountered in the stratosphere, and it cannot be excluded that the observed low-temperature effect is to some extent specific for the conditions inside the measurement chamber. I realize that it is a challenge to recreate the exact atmospheric and radiative conditions of the upper atmosphere in a laboratory set up, but the authors should discuss this. In connection to this it would be very useful to use the dual thermistor that is discussed in another paper by the same authors (Lee et al. Meteorol. Appl. 25: 283–291 (2018), doi: 10.1002/met.1690) to make a connection between tests in the UAS and in situ measurements at high altitude. This should provide an answer to the question whether measurements in the UAS can replicate the temperature difference between both sensors of the dual thermistor radiosonde at high altitude for an ambient temperature of -67°C?

→ We have learned that the net longwave radiative cooling from the sensor is negligible ($\sim 10^{-6}$ W) compared to the convective cooling (200 W) in the above calculation. According to the calculation, the observed temperature effect on ΔT_{rad} can be explained by the change of air properties at low temperatures which is not related with the use of a measurement chamber.

For your information, dual thermistor radiosonde (DTR) was also tested using the UAS and the result was submitted to the same Journal (ID: amt-2021-343).

The introduction should include more discussion of, and references to, other relevant work.

→ Throughout revisions to comply with the specific comments of the Reviewer, the Introduction is reinforced with discussions of relevant works.

The influence of the thermal coupling between sensor boom and temperature sensor for the radiation error should be discussed in more detail.

→ The thermal coupling between sensor boom and temperature sensor is discussed.

Added Statement (Line 205-212): It was previously observed that the temperature rise of RS92 was initially fast due to the small thermal mass of the sensor and subsequently slow (Dirksen et al., 2014). More recently, the temperature of RS41 oscillated when the radiosonde was rotating under irradiation (Von Rohden et al., in review, 2021). These observations are attributed that the heating of the sensor boom with comparably large area is coupled to the heating of the temperature sensor. Since the conductive heat transfer from the sensor boom is missing in the above theoretical calculation, the comparison in Fig. 3(b) may show the effect of the sensor boom on ΔT_{rad} . Interestingly, the growth of ΔT_{rad} of the theoretical calculation is less steep than that of the experiment as the pressure is decreased to 5 hPa. This may imply that the heat transfer from the sensor boom becomes significant especially at low pressures.

The differences with the Vaisala radiation correction are listed in Table 11, but without much further discussion or interpretation of these differences. The UAS-based findings should be presented in a plot, together with Vaisala correction, so that it is easier for the reader to see the differences between both correction methods. At this place a discussion of the observed altitude-dependence of these differences would be appropriate.

→ As suggested, the radiation corrections by the UAS and Vaisala are presented together by plots as shown in Fig.9. To explain the situation, Figure 8 is also added.

Before: The radiation correction of RS41 by the UAS is based on Eqs. (13) and (14) for different pressure ranges. Although the conditions for the UAS correction are different from those considered by the manufacturer, a rough comparison of the radiation corrections is presented in Table 11. For the UAS correction, the solar irradiance is assumed to be $S = 1360$

$\text{W}\cdot\text{m}^{-2}$ at all pressure values. Depending on the effective irradiance (S_{eff}), the UAS correction value should be revised in a proportional manner using Eqs. (13) and (14).

After (Line 397-413): In order to apply the correction formula to actual soundings, the effective irradiance to the sensor should be known. However, radiosondes constantly change positions with respect to the solar irradiation through rotation and pendulum motion, the calculation of effective irradiance resorts on the mean of effective irradiance over the motion of radiosondes. Figure 8(a) shows a schematic diagram of a radiosonde with parameters that affect the effective irradiance S_{eff} on the sensor. Then, the effective irradiance to the sensor can be calculated as follows:

$$S_{\text{eff}} = S_{\text{dir}} \cdot \left| \cos\alpha \cos\theta \cos\varphi - \sin\theta \sin\alpha \right|, \quad (18)$$

S_{dir} is solar direct irradiance, θ is boom tilting angle, α is solar elevation angle and φ is azimuthal angle. The effective irradiation area (A_{eff}/A_0) on the sensor boom is averaged over rotation (φ) with a fixed tilting angle $\theta = 45^\circ$ and plotted as a function of the solar elevation angle as shown in Fig. 8(b). Using this effective irradiance, the radiation correction by the UAS is obtained and compared with that of the manufacturer at two different α (45° and 90°) as shown in Fig. 9. For the UAS correction, the solar direct irradiance is assumed to be $1360 \text{ W}\cdot\text{m}^{-2}$ at all pressure values. To simulate the albedo effect, the radiation correction with additional irradiance of $400 \text{ W}\cdot\text{m}^{-2}$ is also calculated. Consequently, the radiation correction of the UAS is smaller than the Vaisala by about $0.5\text{--}0.7^\circ\text{C}$ at -70°C and 5 hPa when only the solar direct irradiance ($1360 \text{ W}\cdot\text{m}^{-2}$) is considered with the solar elevation angle $\alpha = 45\text{--}90^\circ$. When the albedo effect is additionally included ($400 \text{ W}\cdot\text{m}^{-2}$), the gap between the two corrections is reduced to $0.04\text{--}0.4^\circ\text{C}$ at -70°C and 5 hPa with $\alpha = 45\text{--}90^\circ$. The radiation corrections of the manufacturer and the UAS at some representative conditions are summarized in **Table 11**.

Modified Table (Table 11): Table 11 is modified to include the radiation correction of the UAS obtained by the above method.

In the description of the experiments the following information appears to be missing: the length of the exposure times, and the number of repetitions.

→ The length of the exposure time and the number of repetitions are added.

Added Statement (Line 148): The duration of irradiation is 120 s and the measurement is repeated three times.

Show uncertainties in the plots of figures 2-6

→ The data in Figure 2 represents the mean and the standard deviation of three repeated measurements on a single RS41 unit. Since standard deviations are small (max. 0.014°C), they are not seen clearly. Therefore, the biggest standard deviation is mentioned in the manuscript.

Standard deviations are newly added to Figs. 3 and 4, but not to Figs. 5, 6, and 7 because these experiments were conducted once.

Added Statement (Line 152-153): The data represents the mean and the standard deviation of three repeated measurements on a single RS41 unit. The biggest standard deviation was 0.014 °C.

Please provide a more detailed description of how Delta_T is determined. Is this the difference between the radiosonde temperature and the air temperature in the UAS?

→ It is mentioned in the manuscript.

Before: The temperature rise due to irradiation (ΔT_{rad}) is defined as the difference in the temperatures with irradiation (T_{on}) and without irradiation (T_{off}); $\Delta T_{\text{rad}} = T_{\text{on}} - T_{\text{off}}$.

After (Line 146-148): The temperature rise due to irradiation (ΔT_{rad}) is defined as the difference in the temperatures with irradiation (T_{on}) and without irradiation (T_{off}) as previously reported (Lee et al., 2020); $\Delta T_{\text{rad}} = T_{\text{on}} - T_{\text{off}}$.

Specific comments

l23-24: it should be mentioned that radiosondes perform in situ measurements

→ The sentence is changed to include “in situ measurement”.

Before: Radiosondes are telemetry devices that include various sensors to measure data that are transmitted to a ground receiver while the device is carried by a weather balloon to an altitude of approximately 35 km.

After (Line 25-27): Radiosondes are telemetry devices that include various sensors to perform in situ measurements and transmit the measured data to a ground receiver while the device is carried by a weather balloon to an altitude of approximately 35 km.

l26: add pressure to the list of ECVs

→ Pressure is added to the list of ECVs.

Before: essential climate variables (ECVs) such as the temperature, water vapour, wind speed, and wind direction

After (Line 28): essential climate variables (ECVs) such as the temperature, water vapour, pressure, wind speed, and wind direction

l26: (concerning the high accuracy), provide an estimate how high this accuracy is.

→ The claimed accuracy of Vaisala RS41 is presented.

Before: Owing to their high accuracy

After (Line 29): Owing to their high accuracy of 0.3 to 0.4 K claimed by manufacturers

l28: This sentence needs to be rephrased. I assume you want to say something like...

→ The sentence is changed.

Before: Notably, an effective method to evaluate the measurement accuracy specified by manufacturers remains to be specified.

After (Line 30-33): However, evaluation methods for their sensor accuracy are not fully disclosed to users. Operation principle of laboratory setups, algorithms to correct measurement errors, and corresponding uncertainty evaluations are prerequisites for a reference data product.

l29-30: rephrase: may lead to inhomogeneities in data records due to the use of different radiosonde models.

→ Rephrased.

Before: may lead to inhomogeneity among users, including upper-air observatories that use different radiosonde models.

After (Line 33-34): may lead to inhomogeneities in data records due to the use of different radiosonde models.

l33: of the ECVs -> of selected ECVs

→ Done.

Before: of the ECVs

After (Line 37): of selected ECVs

l37: rephrase: are exposed to solar radiation, which leads to radiative heating of the temperature sensor.

→ Rephrased.

Before: are exposed to solar radiation, and the radiative heating increases the temperature to more than the air temperature

After (Line 41-42): are exposed to solar radiation, which leads to radiative heating of the temperature sensor.

I37: As background info to the reader, provide a typical estimate of the radiation temperature error for current (and previous) radiosondes.

→ Typical estimates of the radiation correction are presented.

Added Sentence (Line 42-45): According to the last intercomparison of high quality radiosonde systems (Nash et al., 2011), radiation correction values applied by manufacturers were distributed from 0.6 to 2.3 K at 10 hPa. More recently, according to the radiation correction of Vaisala RS41, it is increased from 0.83 to 1.16 K at 10 hPa as solar angle is elevated from 10 ° to 90 °.

I40: A proper reference for the need to reduce sensor size would be de Podesta et al 2020.

→ The reference is added.

Added Reference (Line 48): de Podesta *et al.* 2018 "Air temperature sensors: dependence of radiative errors on sensor diameter in precision metrology and meteorology" is added.

I41: This sentence is unclear, please rephrase. What do you mean by the combined effect of solar radiation?

→ The sentence is modified to deliver the meaning more clearly.

Before: However, it is not possible to eliminate the combined effect of solar irradiation

After (Line 49-50): Nevertheless, the effect of solar irradiation cannot be eliminated and thus should be corrected properly.

I48: Add a one-line description of how Schmidlin 1986 corrected the radiation error.

→ A short description on Schmidlin *et al.* 1986 is added.

Added Sentence (Line 57-59): Radiation correction was estimated by using radiosondes equipped with four thermistors having coatings with different spectral responses, i.e., emissivities and absorptivities (Schmidlin et al., 1986).

I50-52: include the quantitative result of Philipona 2013, i.e. the magnitude of the correction they derived.

→ A quantitative result of Philipona *et al.* 2013 is added.

Added Sentence (Line 62-63): As a result, radiation correction was obtained by a linear function of geopotential height which gives 1 K at 32 km.

I57: add proper reference (von Rohden et al. AMT 2021, doi:10.5194/amt-2021-187)

→ The reference is added properly.

Added Reference: von Rohden, C., Sommer, M., Naebert, T., Motuz, V., and Dirksen, R. J.: Laboratory characterisation of the radiation temperature error of radiosondes and its application to the GRUAN data processing for the Vaisala RS41, Atmos. Meas. Tech. Discuss. [preprint], <https://doi.org/10.5194/amt-2021-187>, in review, 2021.

I59-64: The chamber system used by Lee2018, Lee2020, and in this manuscript is one and the same. This should be clear from the text.

→ Two references are separated.

Before: Notably, the existing studies based on other chamber systems reported that the solar-irradiation-induced temperature rise of sensors increases as the air temperature is decreased (Lee et al., 2018a; Lee et al., 2020).

After (Line 70-71): Notably, a previous study based on a chamber system reported that the solar-irradiation-induced temperature rise of sensors increases as the air temperature is decreased (Lee et al., 2018a).

Before: In this study, the uncertainty in the radiation correction of a Vaisala RS41 temperature sensor is evaluated using the UAS at KRISS.

After (Line 76-77): In this study, the uncertainty in the radiation correction of a Vaisala RS41 temperature sensor is evaluated using the UAS developed at KRISS (Lee et al., 2020).

165-66: Add a sentence explaining how the radiation correction is evaluated. This is discussed in detail in the later sections, but a brief mentioning of the principle should be given in the introduction.

→ A sentence explaining the evaluation on the radiation correction is added.

Added Sentence (Line 77-79): It is shown how the uncertainty of each environmental parameter and radiosonde movements in the UAS contributes to the uncertainty of RS41 through a radiation correction formula obtained by a series of radiation experiments.

166: Provide additional details on the UAS. The possibility to tilt and rotate the radiosonde are important innovations that should be mentioned here.

→ The possibility to tilt and rotate the radiosonde is mentioned.

Before: along with the addition of new functions to consider the effect of the rotation and tilting of the sensor.

After (Line 79-80): along with the addition of new functions to consider the effect of the rotation and tilting of the sensor that are an important progress from the previous version of the UAS.

181-82: do you mean the temperature of the test chamber, or the temperature of the air inside the test chamber?

→ The climate chamber is used to control the temperature of both test chamber and air.

Before: The test chamber is inside a climate chamber (Tenney environmental, Model: C64RC) to control the temperature. The working space of the climate chamber is sized 1219 mm × 1219 mm × 1219 mm.

After (Line 91-97): The test chamber is inside a climate chamber (Tenney environmental, Model: C64RC) of which working space is 1219 mm × 1219 mm × 1219 mm. The temperature of the test chamber is controlled by the climate chamber. Air is precooled before entering into the climate chamber by passing through a heat exchanger in a separate bath (Kambic metrology, Model: OB-50/2 ULT) of which temperature is lower than that of the climate chamber by about 5 °C. The temperature of the precooled air is then adjusted to that of the climate chamber while passing through the second heat exchanger (9.3 m in length) in the climate chamber before entering into the test chamber.

192: What are the results of the laser Doppler velocimetry characterisation of the airflow inside the test chamber?

→ Description on the measurement by the laser Doppler velocimetry inside the test chamber is added.

Added sentence (Line 103-104): Thus, the reference value and SI traceability of the ventilation speed are obtained by using the sonic nozzles in the UAS.

Before: The generated air flow is measured through laser Doppler velocimetry to investigate the spatial gradient in the test chamber.

After (Line 108-111): The generated air flow is measured through laser Doppler velocimetry (LDV) (Dantec, Model: BSA F60) to investigate the spatial gradient in the test chamber. Ar-ion laser (3W) having a wavelength of 514.5 nm is used with a focal length of 400.1 mm and nominal beam spacing of 33 mm.

Before: The spatial gradient of the ventilation speed in the test chamber is measured through laser Doppler velocimetry at KRISS.

After (Line 327-332): The spatial gradient of the ventilation speed in the test chamber is measured through the LDV at KRISS. The measurement dimension using the LDV was 35 mm x 35 mm around the sensor (central) location with 5 mm interval (49 points) in the test chamber (50 mm x 50 mm). The measurement was performed at the condition of $v = 4.67 \text{ m}\cdot\text{s}^{-1}$ (reference value), $P = 550 \text{ hPa}$, and room temperature. The LDV value averaged over the measurement area (35 mm x 35 mm) was $4.63 \text{ m}\cdot\text{s}^{-1}$. The difference between the reference and the measurement average is assumed to have a rectangular probability distribution for the calculation of the uncertainty of spatial gradient.

I96: It is not clear whether 980 W/m^2 is the flux at the position of the radiosonde's sensor, so inside the test chamber, or if it denotes the flux at the window of the test chamber.

→ The flux at the position of the radiosonde sensors inside the test chamber is 980 W/m^2 .

Before: A constant irradiance of $980 \text{ W}\cdot\text{m}^{-2}$ is adopted throughout this study.

After (Line 115-116): A constant irradiance of $980 \text{ W}\cdot\text{m}^{-2}$ at the position of the radiosonde sensors inside the test chamber is adopted throughout this study.

I102-103: Should the study still be seen as proof of concept?

→ The phrase “As a proof of concept” is removed.

Removed phrase: As a proof of concept

I100: mention the accuracy at which the pyranometer is calibrated at KRISS.

→ The calibration uncertainty is mentioned.

Before: The pyranometer is calibrated at KRISS.

After (Line 119-120): The pyranometer is calibrated at KRISS and the uncertainty is 1% of the measured value with a coverage factor $k = 1$.

I107: Remove 'it seems that'. The following sentence makes clear that you know the manufacturer's radiation correction is not applied.

→ The phrase is removed as suggested.

Removed phrase: It seems that

I116: considered -> employed

→ The word is changed as suggested.

Before: considered

After (Line 138): employed

I119: Change section number to 3

→ Done.

Before: 2 Experiment Details

After (Line 141): 3 Experiment Details

I128: I don't understand this sentence. Do you mean that the relative uncertainty of Delta T is larger at high pressure because of its smaller value?

→ The relatively larger uncertainties at high pressures indicates that $U(\Delta T_{\text{rad}})/\Delta T_{\text{rad}}$ becomes larger because ΔT_{rad} is decreased at high pressures. Thus, the temperature effect is not clearly observable at high pressures.

Before: This phenomenon can be attributed to the relatively larger uncertainties in ΔT_{rad} at high pressures in the UAS.

After (Line 156-157): This phenomenon can be attributed that the uncertainty of ΔT_{rad} becomes relatively larger with respect to ΔT_{rad} as ΔT_{rad} is decreased at high pressures in the UAS.

I129: Motivation for using exponential functions for fitting?

→ Exponential fittings are replaced by polynomial fittings with $\log_{10} P$ to provide a single fit over entire pressure range of 5–500 hPa (Reviewer 2's suggestion). Consequently, residuals by polynomial fittings (± 0.03 °C) are smaller than those by exponential fittings (± 0.04 °C) as shown in Fig. 2(b). Therefore, polynomial fittings are newly used to obtain radiation correction formula throughout the revised manuscript.

Modified Figure (Figure 2): Figure 2(a) is replotted by using parabola fittings of $\log_{10} P$ and the residual in Fig. 2(b) is newly obtained accordingly.

Modified Equation (Equation 1, 9, 10 & 11): The backbone of the Equations is changed to $\Delta T_{\text{rad}} = A_0(T) + B_0(T) \cdot \log(P) + C_0(T) \cdot [\log(P)]^2$

Modified Table (Table 1): The original **Table 1** is removed and **Table 2** is relabeled to **Table 1**. **Table 1** is modified to include information on new coefficients of $A_0(T)$, $B_0(T)$, and $C_0(T)$.

I144-146: See remark in the major comments above about physical justification for the different radiation error at low temperature. Changing the temperature from +20°C to -67°C has a 20% effect on the observed radiation error. This is a significant difference which merits a thorough discussion of the effects involved. Preferably with appropriate references. For example, what could be the effect of temperature to the convective heat exchange?

→ As suggested, the theoretical discussion on the effect of temperature in terms of convective heat exchange is added in the revised manuscript.

I170-171: This is a bold conclusion that should be reached only after thorough verification and discussion.

→ Both experimental and theoretical works indicate that the radiation correction is dependent on temperature because the properties of air change with temperature. We do not think that the statement is a bold conclusion.

I177: Despite the limited investigated range of ventilation, is it reasonable to assume a simple linear relationship between radiation effect and ventilation speed?

→ Even though the relationship is not perfectly linear, residuals by using a linear function is not significant. The validity of the linear relationship is newly described in the revised manuscript.

Added Statement (Line 249-250): The linear relationship between the ventilation speed and the radiation correction in Eq. (9) is only valid in the range of $4\text{--}7\text{ m}\cdot\text{s}^{-1}$. When v is higher than $7\text{ m}\cdot\text{s}^{-1}$ or lower than $4\text{ m}\cdot\text{s}^{-1}$, the formula underestimates the correction value.

1205-206: please rephrase. Say something along the lines that the exposed surface of the sensor boom depends on the incidence angle, and passes through a maximum twice during a full rotation.

→ The sentence is changed as suggested.

Before: This phenomenon occurs because the sensor boom undergoes similar processes in the first 180° and remaining 180° in a 360° rotation.

After (Line 269-270): The exposed surface of the sensor boom depends on the incidence angle, and passes through a maximum twice during a full rotation.

Table 10: Vaisala provides more detailed information on calibration uncertainty of the temperature sensor in the White Paper ‘Vaisala Radiosonde RS41 Measurement Performance’, B211356EN-B-LOW-v3

→ Since $T_{\text{cor}} = T_{\text{raw}} - \Delta T_{\text{rad}}$ in Eq. (16), the uncertainty of T_{cor} requires that of T_{raw} and ΔT_{rad} . $U(\Delta T_{\text{rad}})$ is obtained by the UAS experiment. $U(T_{\text{raw}})$ is adopted from the White Paper which describes the uncertainty of “repeatability in calibration” is $0.1\text{ }^\circ\text{C}$ at $k = 2$. Although Vaisala provides other uncertainties related with the storage and soundings (as a manufacturer), this work is limited to the evaluation of RS41 by a laboratory facility. The uncertainty due to soundings should be added to the total uncertainty when the correction formula is applied to soundings. The White Paper is cited in **Table 10**.

Added Statement (Line 390-392): Since Vaisala provides additional uncertainty of reproducibility in sounding ($0.15\text{ }^\circ\text{C}$ when $P > 100\text{ hPa}$, $0.3\text{ }^\circ\text{C}$ when $P < 100\text{ hPa}$) (Vaisala), this should be added to the total uncertainty of the corrected temperature when the radiation correction formula in Eq. (11) is applied to soundings.

Figure 5: the plots show the effect of the rotation on the sensorboom on the radiative heating of the temperature sensor. Figure 11 of von Rohden 2021 presents the same effect, however with a considerably larger magnitude (up to 0.3K for $p=7\text{hPa}$ and 16s rotation period). This difference in findings should be addressed by the authors. The amplitude and shape of the oscillations also depend considerably on the incidence angle (equivalent to solar elevation), which is not investigated in the study. This may lead to considerably larger $T_{\text{(on_max)}} - T_{\text{(on_min)}}$ in Fig. 5b, and therefore to higher uncertainty estimates due to sonde rotation. Fig 5 a) shows the results for 100 hPa , where the effect is rather small. It would be more meaningful

showing the results for $p=5$ hPa. The bump around 100 hPa in the pressure dependence of the amplitude for the black and red curve is striking. Does this reflect the uncertainty of the measurement? Finally, mention in the caption the values of v , and T for the data in the plot.

→ The rotation axis is the temperature sensor itself, not the center of the boom. Therefore, the temperature sensor only spins on the spot and thus the distance between the sensor and the light source does not change at all during the rotation. This implies that the irradiance to the sensor is constant whereas the light incident angle (effective irradiance) to the sensor boom changes with rotation. This explains why the maximum temperature peak appears twice during a single cycle. The fixed rotation axis to the sensor may be the reason why the $T_{\text{on_max}} - T_{\text{on_min}}$ in the UAS is smaller than that of von Rohden *et al.* (0.3 °C). In the work of von Rohden *et al.*, although the distance from the light source to the sensor is constant (or slightly changes), that to the sensor boom changes with rotation. This may be the reason why the maximum peak appears once in a single cycle when the sensor boom is close to the light source.

The biggest estimated value of $T_{\text{on_max}} - T_{\text{on_min}}$ (0.06 °C) at $T = -67$ °C and $P = 5$ hPa is used for the uncertainty calculation for all pressures.

The values of v and T are mentioned in the caption.

Added Statement (Line 263-265): The rotation axis is the temperature sensor itself, not the centre of the boom in this work. Therefore, the temperature sensor only spins on the spot and thus the distance between the sensor and the solar simulator does not change during the rotation.

Added Statement (Line 277-282): The maximum value of $(T_{\text{on_max}} - T_{\text{on_min}})$ in the UAS (0.05 °C) is much smaller than that of von Rohden *et al.* (0.3 °C) (Von Rohden *et al.*, in review, 2021). In the work of von Rohden *et al.*, although the distance from the light source to the sensor is constant, that to the sensor boom changes with rotation. This may be the reason why the maximum peak appears once in a full cycle when the sensor boom is close to the light source and the $(T_{\text{on_max}} - T_{\text{on_min}})$ is bigger than this work. The relatively small $(T_{\text{on_max}} - T_{\text{on_min}})$ with respect to ΔT_{rad} observed in this work suggests that the contribution of the thermal conduction to ΔT_{rad} is small compared to that by the direct irradiation of the sensor.

Before: (a) Effect of sensor rotation with varied cycles (5 s, 10 s, and 15 s)

After (Figure 6 Caption): (a) Effect of sensor rotation with varied cycles (5 s, 10 s, and 15 s) at $T = 25$ °C and $v = 5$ m·s⁻¹