

Supplementary Material of the Manuscript Grossi et al.,: 'Characterizing the automatic radon flux Transfer Standard system *Autoflux*: laboratory calibration and field experiments'

Here we present the support material of the Grossi et al., manuscript: 'Characterizing the automatic radon flux Transfer Standard system *Autoflux*: laboratory calibration and field experiments'.

Table S1. Summary of the exhalation bed facilities studied.

Institution	F (mBq m ⁻² s ⁻¹)	Reference
CANMET Elliot lake Laboratory (Canada)	285 ± 41	Stieff et al. (1996)
Radon Laboratory of the University of South China	1480 ± 50	Tan & Xiao (2011)
Oak Ridge Associated Universities (USA)	430 to 80	Altic (2014)
Institute of Industrial Ecology (Russia)	700 ± 80	Onishchenko et al. (2015)
University of Huelva (Spain)	13.3 ± 0.2 23.4 ± 0.3	Gutiérrez-Álvarez et al. (2020)

Table S2. Sensor features used to monitor the environmental conditions.

Sensor	Manufacturer	Model	Range	Declared Accuracy
Temperature	Testo	175T2	(-35 to 55) °C	± 0.5 °C
		Probe	(-40 to 120) °C	± 0.3 °C
Soil moisture	ODYSSEY	Xtreem	(0 to 100) %	$\pm 1\%$
Pressure	ITEFI-CSIC	-	(-600 to 600) Pa	± 3 Pa



Figure S1. Picture of the setup used to empirically determine the exhalation rate reference value of the EB.

Table S3. Main characteristics of radon flux systems available from the literature.

Cod e	Manuf acturer	Volu me (m ³)	Exhalati o area (m ²)	Refr esh*	Chamb er shape	Radon monitor /mode	Enviro nment al sensors within the monito r	Env iron . Sens ors with in the Cha mbe r	Source
1	ANST O	0.019	0.13	Y	Drum	AG/Pump	Y	N	-
2	ANST O	0.018	0.26	N	Shallow conical	AG2 x 1L lucas cells (separated by 6 min flow path)/Pump	Y	N	Zahorowski and Whittlestone, 1996
3	LICO R	0.041	0.03	Y	Hemisp here	-/Pump	-	N	https://www.licor.com/env/products/soil_flux/#chambe r-difference
4	IPSN	0.037	0.21	Y	Cylinde r	AG/Pump	Y	N	Ferry et al., 2001
5	Univ. of Wisco nsin-Madiso n	0.002, 0.018, 0.352	0.02 , 0.07, 2.32	N	Circular , Circular , Square	RAD7/Pump	Y	N	Stefani et al., 2016
6	Helmh oltz Zentru m München	0.044	0.13	Y	Cylinde r	6 Lucas cells/Pump	N	Y	Yang et al., 2017
7	UPC	-	0.01	Only pum p	Cylinde r	DOSEman/Diffusion	N	N	-
8	UC	0.008	^{0.04}	Only pum p	Rectang ular	AlphE/Diffusion	N	N	-




Figure S2. Pictures of enlisted radon chambers, code numbers correspond to those in Table S4.




Figure S3. Schematic representation of the *AutoFlux* system (ANSTO).




Figure S4. *AutoFlux* drum during a typical radon flux measurement: accumulation period (1 hour, on the left side) and ventilation period (2 hours, right side).




Figure S5. Evolution of radon concentration with time in a volume V during the experiment to establish the emanation factor ε of the radon from the soil sample. The coefficients of the exponential fit are presented.




Figure S6. Example of radon concentration over time during the experimental determination of exhalation rate.



Figure S7. Setup of the *AutoFlux* during a typical laboratory measurement at UC.




Figure S8. Conceptual box model of the ANSTO *AutoFlux* system.




Figure S9. Simulated ^{222}Rn concentration behavior within each one of the volumes of the *AutoFlux* system during the hour for which the chamber was closed C_d (light blue line), C_u (blue line) and C_{4G} (red line).