1 Colorimetric derivatization of ambient ammonia (NH₃) for detection by

2 long path absorption photometry

3 Shasha Tian^{a, b, 1}, Kexin Zu^{a, b, 1}, Huabin Dong^{a, b, *}, Limin Zeng ^{a, b}, Keding Lu^{a, b}, Qi Chen ^a

^a State Key Joint Laboratory of Environmental Simulation and Pollution Control, College of
 Environmental Sciences and Engineering, Peking University, Beijing, 100871, China.

^b International Joint laboratory for Regional pollution Control (IJRC), Peking University, Beijing, China

- 7 * Corresponding author: hbdong@pku.edu.cn
- 8

9 Abstract. In the last few decades, various techniques, including spectroscopic, mass spectrometric, 10 chemiluminescence, and wet chemical methods, had been developed and applied for the detection of gaseous ammonia (NH₃). We developed an online NH₃ monitoring system—salicylic acid derivatization 11 12 reaction and long path absorption photometer (SAC-LOPAP)-based on a selective colorimetric reaction 13 to form a highly absorbing reaction product and a LOPAP, which could run stably for a long time and be 14 applied to the continuous online measurement of low concentrations of ambient NH₃ by optimizing the 15 reaction conditions, adding a constant temperature module and liquid flow controller. The detection limit 16 reached with this instrument was 40.5 ppt with a stripping liquid flow rate of 0.49 ml min⁻¹ and a gas 17 sample flow rate of 0.70 L min⁻¹. An inter-comparison of our system with a commercial instrument 18 Picarro G2103 analyzer (Picarro, US) in Beijing was presented, and the results showed that the two instruments had a good correlation with a slope of 1.00 and an R^2 of 0.96, indicating that the SAC-19 20 LOPAP involved in this study could be used for the accurate measurement of NH₃.

21 **1. Introduction**

Gaseous ammonia (NH₃) widely exists in the atmosphere and plays an important role in many atmospheric chemical reactions (Swati and Hait, 2018; Klimczyk et al., 2021; Wang et al., 2018). As the most abundant alkaline gas in the atmosphere, NH₃ easily forms ammonium ions (NH₄⁺) with water and reacts with acidic species to form secondary inorganic particles (Ianniello et al., 2011; Ni et al., 2000). These secondary particles are considered a major source of fine particulate matter (PM), which is harmful to climate, visibility and human health (Bu et al., 2021; Gao et al., 2021). Furthermore, recent studies have shown that NH₃ is necessary to control fine particulate pollution (Wen et al., 2018; Wang et al., 29 2013). Due to those problems, the inventory of NH_3 emissions and concentration in urban air has been 30 highly evaluated. Agriculture, including animal feedlot operations, is considered as the largest emission 31 source of NH₃ with 80.6% of the global anthropogenic emissions followed by 11% from biomass burning 32 and 8.3% from the energy sector, including industries and traffic (Behera et al., 2013). Expert estimate that global annual emissions of NH_3 will increase from 65 Tg N yr⁻¹ in 2008 to 135 Tg N yr⁻¹ in 2100 33 34 (Fowler et al., 2015). However, ambient measurement of NH₃ concentrations is difficult due to several 35 factors: ambient levels vary widely with from 5 pptv to 500 ppbv (Janson et al., 2010; Krupa, 2003; 36 Sutton et al., 1995). Ammonia exists in gaseous, particulate and liquid phases, which add further 37 complicates the measurement (Warneck, 1988). In addition, NH₃ is "sticky" and interacts with surfaces 38 of materials, resulting in slow inlet response times (Yokelson et al., 2003). Finally, the temperature 39 difference between the indoor and outdoor environments and the humidity difference between the inside 40 and outside of the instrument will reduce the accuracy of measurement and calibration. It is therefore 41 essential to accurately measure ambient NH₃ to better quantify concentration and concentration changes 42 and hence to evaluate the impacts of NH₃.

43 In recent years, researchers have developed techniques and methods for detecting NH₃ in the 44 atmosphere, which include spectroscopic, mass spectrometric, chemiluminescence, and wet chemical 45 methods (Von et al., 2009). Spectroscopic methods, such as Cavity Enhanced Absorption Spectroscopy 46 (CEAS) (Gong et al., 2017; Berden et al., 2000) and Cavity Ring-Down Spectroscopy (CRDS) (Martin 47 et al., 2016; Qu et al., 2012), can greatly improve spectral absorption's effective optical path length by 48 using the optical cavity structure. However, the "sticky" of NH₃ will affect background, detection 49 efficiency and detection response time of the instrument (Whitehead et al., 2008; Yokelson et al., 2003). 50 Utilizing a quantum cascade laser (OCL) or a DFB laser in a near-infrared band as the light source can 51 achieve a low detection limit of 0.018ppb (Whitehead et al., 2008; Mcmanus et al., 2002; Von et al., 52 2009), realizing the measurement of low concentrations of NH_3 in ambient air. Mass-spectrography 53 analyzers provide highly sensitive techniques but may be less specific and can be affected by competing 54 ion chemistries. The chemical ionization mass spectrometer (CIMS) technique is based on an ion-55 molecule reaction to selectively ionize and detect trace NH₃ in the atmosphere, which features a fast 56 response and in situ measurement (Benson et al., 2010; Nowak et al., 2007; Yu and Lee, 2012). It has 57 the advantages of small volume and wide measurement range, but its detection limit is very high (Ajay 58 and Beniwal., 2019). Chemiluminescence is an indirect method to measure ammonia. Two catalytic 59 converters of different characteristics catalyze NOx and NO-amine into NO. The NH₃ mixing ratio is 60 calculated by the difference between NOx and NO-amine. This method can realize the simultaneous 61 measurement of NH_3 , NO and NO_2 , but the measurement results are affected by the conversion efficiency 62 (Sharma et al., 2010; Sharma et al., 2012). Wet chemistry methods convert gas-phase NH₃ to aqueous 63 NH_3 (NH_4^+) for online analysis by means of online ion chromatography with a detection limit of 0.05 64 µg m⁻³ (0.72ppb at 25 °C) (Khlystov et al., 1995; Dong et al., 2012; Makkonen et al., 2012). A field inter-65 comparison of NH₃ measurement techniques found that wet chemistry instruments showed better long-66 term stability and agreement than other analyzers (Von et al., 2009), which was due to the wet chemical 67 trapping method and standard calibration solutions, humidity did not affect the measurement, and the 68 standard solution was more stable than standard gases. However, they failed to capture the peak because 69 of lower time resolution. Based on a selective colorimetric reaction to form a highly absorbing reaction 70 product and absorption spectrophotometry collect NH₃ (and ammonium) by aqueous scrubbing in glass 71 frit impactors (Bianchi et al., 2012; Bae et al., 2007) has been used for decades for routine derivatization 72 and colorimetric analysis of NH_4^+ in a wide variety of environmental samples (e.g. soils, environmental 73 waters, etc), which has also been reported by other scholars (Bae et al., 2007). In those studies the product 74 was detected by a long path absorption photometer (LOPAP), in which the absorbance of the solution is 75 amplified in the long path module to reach a lower detection limit (Heland et al., 2001).

76 In this study, we provide an online NH₃ monitoring system based on wet chemistry stripping of 77 atmospheric NH₃, followed by the formation of a highly light-absorbing indophenol after a salicylic acid 78 derivatization reaction to produce the colored reaction product reaction and detected with LOPAP. 79 According to Lambert-Beer's law, the sensitivity of spectrophotometry can be enhanced by increasing 80 the optical path length. This sensitive analytical method has already been successfully applied in different 81 colorimetric detection studies (Yao et al., 1998; Heland et al., 2001; Callahan et al., 2002). In analogy to 82 the original long path absorption photometer (LOPAP) which was developed for HONO measurements 83 (Kleffmann et al., 2002), we call this monitoring system the salicylic acid derivatization reaction and 84 long path absorption photometer (SAC-LOPAP), which features several improvements over versions previously reported by other groups: one is the optimization of reaction conditions, the other 85 86 modification is the use of constant temperature module and flow control system. Secondly, we will


present measurements demonstrating our new system in urban environments in Peking University, with
 measure low concentrations, good stability and low detection limit.


89 **2. SAC-LOPAP instrument**

90 2.1 Measurement principle

91 Our instrument is designed to measure NH₃ in a low-concentration environment (under 20ppb) with the 92 good stability, low detection limit (less than 60 ppt) and small size. There is a brief introduction to the 93 principle of the instrument. The measurement of NH₃ in the SAC-LOPAP instrument is achieved by the selective colorimetric reaction to form a highly absorbing reaction product and absorption 94 95 spectrophotometry. Samples containing dissolved ammonia and ammonium react with a phenolic 96 compound and a chlorine-donating reagent to form indophenol blue during the reaction, with the 97 strongest absorption at a wavelength of 665 nm. (Krom and Michael, 1980; Searle and Phillip, 1984). The 98 reaction mechanism of the chromogenic reactions as shown in (1)-(4). Furthermore, to measure the 99 absorbance of the sample, we used a LOPAP based on liquid-waveguide capillary cell (LWCC) 100 technology to obtain a better detection limit, continuity and stability (Heland et al., 2001).

$$NH_3 + HOCI \implies NH_2CI + H_2O$$
 (1)






101 2.2 Experiment setup

102 We designed our system to consist of four modules: the sampling module, the reacting module, the 103 detecting module, and the control module (Fig. 1). The key component of the sampling module is a glass 104 coil reactor, which is an open glass tube (inner diameter 1.5 mm, 75 cm long) coiled 12 turns. At the 105 beginning of this coil, there is a flow manifold to mix the ambient air flow and the stripping solutions. 106 The air is pumped into the stripping coil under the action of a vacuum diaphragm air pump and a gas 107 flow meter (Horiba, China) (Chen et al., 2004). To protect the gas flowmeter and the air pump, a security 108 bottle is installed in front of the gas flowmeter to prevent the inflow of liquid. At the same time, the 109 stripping solution, regulated by the liquid flow control system, is injected into the stripping coil to capture 110 NH₃ in the air and form a mixture of ammonium-salicylic acid. To achieve higher absorption efficiency, circulating cooling water with a temperature of 10-15 °C is provided outside the stripping coil. The center 111 112 part of the reacting module is a reaction coil and a debubble. The liquid sample is mixed with the alkaline 113 derivatization solution, and a derivatization reaction to produce the colored reaction product reaction 114 occurs in the heated reaction coil. The reaction coil is made of a 90 cm length of Teflon tubes coiled on 115 a heat-conducting metal cylinder, and a PID controller controls the temperature of the reactor at 40-75 °C to accelerate the derivatization reaction. After the derivatization reaction, the sample is sent to the 116 117 detecting module, which comprises a liquid waveguide capillary cell (LWCC-100, World Precision 118 Instruments, USA) with optical path length of 100 cm, an LED light source with the mode at 665 nm (Ocean Optics) and a phototube (S16008-33, HAMAMTSU, Japan) for the long path photometry 119 120 detection. The sample solution to be tested is filtered by a 1.0 µm filter before passing through LWCC to 121 avoid interference from components of the sample matrix/method reagents. Both the fluid propulsion 122 module and detection module can be computer controlled.

123 Eq. (5) can help convert the concentration of NH_{4^+} solution $C_{NH_4^+}$ to the NH₃ concentration in the 124 gasous C_{NH_3} .

125
$$C_{NH_3} = \frac{C_{NH_4^+}F_l RT}{M_{NH_4^+}F_g P\gamma}$$
 (5)

Where C_{NH_3} denotes the content of NH₃ in the air sample (ppb), *P* denotes atmospheric pressure (101.3 kPa), $M_{NH_4^+}$ denotes the molar mass of NH₃ (18.04 g/mol), *R*=8.314 Pa·m³·mol⁻¹·K⁻¹. *T* denotes the room temperature (K), F_l denotes the flow rate of stripping solution, F_g denotes the flow rate of sampling gas, γ denotes the capture efficiency of air NH₃ in the stripping solution (a constant determined by laboratory).



131

Fig. 1. Schematic diagram of SAC-LOPAP. (M: gas flowmeter; S.S.: stripping solution; STA.: standard solution;
D.S.: derivatization solution; Sec. Bottle: security bottle; C.B.T: cooling buffer tube; T.V.: triple valve for switching
stripping solution and standard solution; P1, P2, P3, P4, P5: peristaltic pump for transferring solutions; FT: syringe
filter; IC: integrated circuit. The gas flow rate can be controlled from 0.2-2.0 L min⁻¹, with an optimal gas flow rate
of 0.7 L min⁻¹. The liquid flow rate can be controlled from 0.1-1.0 ml min⁻¹, with an optimal stripping liquid flow
rate of 0.49 ml min⁻¹).

138 2.3 Experiment protocol

The exact recipe of the chemical reactions follows the reactions described by Searle et al (Searle and
Phillip, 1984). We used 0.75g L⁻¹ salicylic acid (TCI, 99.5%, Japan), 0.014 g L⁻¹ sodium nitroferricyanide

(TCI, 99%, Japan), and 0.2 g L⁻¹ NaOH as stripping solution (R1). Then the 0.188ml L⁻¹ Sodium
hypochlorite (Aladdin, active chlorine10%, China) and 1.5 g L⁻¹ NaOH as derivatization solution (R2).
We acknowledge that based on a selective colorimetric reaction to form a highly absorbing reaction
product must be carried out under catalytic and alkaline conditions. Sodium nitroferricyanide is
recognized as a high-efficiency catalytic to increase the sensitivity of the Equation 4 (Krom and Michael,
1980; Searle and Phillip, 1984).

147 Calibrating the setup uses $NH_{4^{+}}$ standard solution produced by the National Institute of Metrology, 148 China. The standards are prepared shortly before use by $NH_{4^{+}}$ standard solution with R1 in volumetric 149 bottle and to use it right after it was ready. The ideal use cycle of R1 and R2 was half a month, after 150 replacement with new R1 and R2 solutions and other instrument fittings, the instrument should be 151 recalibrated to ensure data quality.

152 2.4 Sampling method

153 The inter-comparison experiment was conducted at the College of Environment Sciences and 154 Engineering, Peking University, located within the 4th ring road in northern Beijing, China (39.59° N, 116.18°E). A commercial instrument Picarro G2103 analyzer (Picarro, US) used for atmospheric NH₃ 155 156 measurement based on the CRDS method was deployed concurrently with SAC-LOPAP in the 157 comparison, which could be used to validate other instruments (Twigg et al., 2022). The experiment took 158 place from 15 September 2021 to 15 October 2021, with the instruments installed in a field container. 159 Two instruments shared an inlet and were deployed 2.5 m above the ground. A Polytetrafluoroethylene 160 (PTFE) filter (46.2 mm diameter, 2 µm pore size, Whatman, USA) is used in the front of the sample module to remove ambient aerosols, which is placed into a round filter holder made of perfluoro alkoxy 161 162 (PFA). We changed the filter every day with the aim of avoiding uncertainties. After the filtration of the 163 aerosols, the sample gas flow is delivered into a 3.8 m long 1/4-inch Teflon tube, and a temperaturecontrolled metal heating wire (set at 35 °C ± 0.1 °C) is wrapped around the sample tube and covered with 164 165 thermo-isolation materials. We ran our instrument with an additional drag flow of 1.75 L min⁻¹ with aim to ensure the ambient residence time was about 7.8 msec for all instruments. Data acquisition times were 166 167 different for the above instruments during the inter-comparison. The base reporting periods for Picarro 168 and SAC-LOPAP were 1 s and 30 s. For the purposes of comparison, data from the two instruments

169 presented in this section were averaged to 30 s. In addition, high purity N_2 as zero gas was injected into 170 the sampling tube and carried out every 7 days at the start and end of the campaign as well. The standard 171 air source comes from China Sichuan Zhongce Biaowu Technology Co., LTD. The quality management 172 system of the company conforms to the recognized standard in the Chinese industry (GB/T9001-173 2016/ISO 9001:2015). The composition was ammonia (5.08 ppm) and nitrogen with the uncertainty was 174 2%. In the test, pure N_2 was used as the dilution gas to obtain the required concentration of ammonia 175 standard gas. Calibrations were performed using combinations of concentrations at 1.32, 4.95, 9.59, 176 17.90 and 54.96 ppb from the cylinder. In addition, 4.95 ppb and 54.96 ppb standard gas were injected into the sample tube every 7 days after zero point. The field container was controlled at 25 °C \pm 1 °C to 177 178 reduce the impact of temperature fluctuations on measurement results.

179 **3 Characterization and optimization**

180 **3.1 Sampling efficiency**

181 NH₃ Standard gas of 54.96 ppb was used as the sample to be collected through two identical serial 182 stripping coils, and the concentration of liquid samples collected by the two stripping coils was measured 183 to calculate the capture efficiency. The calculation formula is as below.


184
$$\gamma_1 = \frac{c_1}{c_1 + c_2} \times 100\%$$
 (6)

185 Where, γ_1 denotes the collection efficiency of the first stripping coil, c_1 and c_2 denote the concentration 186 of NH₄⁺ trapped in the first stripping coil and the second stripping coil, respectively.

187 The collection efficiency of NH₃ from the R1 reached more than 99% under different c_{NaOH} , F_l , and

188 F_g . Figure 2a and Figure 2b show that the F_l and the F_g had almost no influence on collection efficiency.

- 189 Figure 2c shows that c_{NaOH} of 1.25 mmol L⁻¹ achieved the greatest collection efficiency in the R1 (99.9%).
- 190 Therefore, the c_{NaOH} of 1.25 mmol L⁻¹ was selected as the R1 of the NH₃. And we selected F_l as 0.49 ml
- 191 min⁻¹ and F_g as 0.7 L min⁻¹ in order to achieve the required detection range in this study.




193 Fig. 2. The absorption efficiency of stripping coil versus (a) gas flow rate ($c_{NaOH} = 4.0 \text{ mmol } \text{L}^{-1}$, $F_l = 0.49 \text{ ml min}^-$ ¹), (b) liquid flow rate ($c_{NaOH} = 4.0 \text{ mmol } \text{L}^{-1}$, $F_g = 0.7 \text{ L min}^{-1}$), (c) concentration of NaOH in R1 ($F_l = 0.49 \text{ ml min}^{-1}$) 194 195 ¹, $F_g = 0.7 \text{ Lmin}^{-1}$).

3.2 Setting reaction conditions 196

197 However, precipitates can attach to the wall of the pipeline and LWCC for on-line instruments, 198 which leads to pipeline blockage and baseline drift. Therefore, we need to optimize reaction conditions, 199 add the constant temperature module and liquid flow controller temperatureto achieve continuous online 200 measurement of low-concentration ammonia in ambient air. The concentration of the R1 we used in the 201 initial reaction conditions (longer optical path and smaller sampling volume) contained 1 g L^{-1} salicylic acid, 0.1 g L⁻¹ sodium nitroprusside, and 1 g L⁻¹ NaOH. 0.5 ml L⁻¹ sodium hypochlorite and 3 g L⁻¹ NaOH 202 203 were used as R2 (Krom and Michael, 1980; Searle and Phillip, 1984). In addition, the syringe filter was 204 introduced to minimize the influence of precipitate (Bianchi et al., 2012), but a large drift of the baseline 205 would still occur during the long time run in our experiment, which will be discussed in detail later. In 206 fact, we tried interrupting the sampling for a few minutes and implementing 5% hydrochloric acid for 207 the system to remove these precipitates. However, the concentration changed greatly before and after 208 each cleaning precipitation. In addition, once the precipitation was formed, it will take a long time to 209 remove the precipitation, which will also increase the risk of contaminating the detector. According to 210 reaction kinetics, reducing the stripping and derivatization concentrations (solution concentration) and 211 [OH] of the system can greatly reduce the formation of precipitates in the solution. Therefore, we need 212 to find the optimal reaction conditions to produce the least amount of precipitate. The maximum absorbance of a 100 μ g L⁻¹ NH₄⁺ standard solution was obtained at 18.75 mmol L⁻¹ OH⁻ and we could 213 obtain a high absorbance of light and a slow speed of precipitate formation, which meant that 1.5 g L^{-1} 214

215 NaOH was added to the derivatization solution, resulted in the precipitate in the solution being too small 216 to cause pipeline blockage and baseline drift. Importantly, we added regular assessment of the system 217 drift through use of online sampling of pure N_2 . The range of blank signal in continuous operation for 48 218 h were 2.856 V ~ 2.848 V and 2.254 V ~ 1.834 V of reduced solution concentration and former high 219 solution concentration, and the maximum offset were 0.3% and 18.6%, respectively, the baseline of low 220 concentration solution has better stability (Fig. 3). In addition, the concentrations of salicylic acid, sodium 221 nitroferricyanide and sodium hypochlorite were 0.04, 0.02 and 0.006 times lower than those in previous 222 research, respectively (Bianchi et al., 2012). In general, the iron-containing precipitate increase the 223 absorbance by scattering or absorbing light, resulting in measurement bias. In this study, the amount of 224 iron-containing precipitation is very small by reducing the content of components and alkali of the 225 solution system, and the voltage of the instrument will not drop significantly due to contamination, which 226 is conducive to better maintenance of the baseline.



227

Fig. 3. The blank time series of the NH₃ detector ran continuously for 48 h.(Low concentration: $0.75g L^{-1}$ salicylic acid, $0.014 g L^{-1}$ sodium nitroferricyanide, and $0.2 g L^{-1}$ NaOH as R1, then the $0.188ml L^{-1}$ Sodium hypochlorite and 1.5 g L⁻¹ NaOH as R2; High concentration: 1g L⁻¹ salicylic acid, 0.1 g L⁻¹ sodium nitroferricyanide, and 1 g L⁻¹ NaOH as R1, then the 0.5ml L⁻¹ Sodium hypochlorite and 3 g L⁻¹ NaOH as R2).


232 **3.3 Stability of liquid flow and temperature**

233 The temperature control module and flow control system were designed because of the sensitivity of 234 molecular absorption spectrophotometry to ambient temperature and residence time. A commercial PID 235 temperature controller was used to control the temperature of the reaction coil with the accuracy of ± 0.1 236 °C. The temperature control module was used to control the constant temperature from the reaction coil 237 to LWCC at 55.0±0.1 °C. At the same time, the flow control system could control the rotational speed of 238 the peristaltic. This system used a commercialized liquid flow meter (SLI-1000, Sensirion, Switzerland) 239 detect the flow rate and feedback to the peristaltic pump control by detecting the flow of tiny bubbles, 240 which further improved the stability of the reaction process. In other words, the flow control system 241 could avoid the flow rate dropping caused by the abrasion of the pump tube and increase the flow rate 242 caused by the replacement of the pump tube, keeping the R1 flow at a constant set point (0.49 ml min⁻¹).

In addition, we designed a buffer tube with a cooling function to further reduce the effects of precipitation. After the derivatization reaction in the reaction coil at 55.0 °C, the mixed solution entered the cooling buffer tube. Most of the precipitation was generated in the buffer tube and attached to the tube wall, while some of the precipitation generated in the downstream pipeline was intercepted by an in-line precipitate filter with a pore size of 1.0 µm before the LWCC, and the filter was changed weekly.

248 Overall, the above work can make the instrument maintain a relatively stable reaction time and 249 temperature, which can promote a relatively stable reaction process, resulting in a high reproducibility 250 to the same concentration of NH₃. In the calibration process, R1 was used as diluent, and the 251 concentrations were 10, 25, 50, 75, 100, 150, and $200\mu g L^{-1}$ of NH₄⁺ standard solution. High purity N₂ 252 was used as blank gas into the sampling tube, and the standard solution entered the solution system 253 instead of the R1. Fig. 4 showed the calibration with the NH₄⁺ concentration gradient of 0, 10, 25, 50 and 100 μ g L⁻¹ (150, and 200 μ g L⁻¹ of NH₄⁺ standard solutions were out of the detection range, which was 254 255 discussed in section 3.4). Each concentration point was run for 40 minutes, and the relative standard 256 deviation (RSD) calculated from four consecutive measurements (the collection of the four replicates 257 were completed during a 4-week of constant instrument operation) ranged from 0.32 % to 2.65 %, with 258 the k varying from 0.0037 to 0.0040. Moreover, the blank experiment tests were automatically made 259 every one or two days, that is, high purity N_2 was used as a blank gas through the sample tube for 40 260 minutes, the RSD of the blank signal in continuous operation for one month was 1.8 %, which indicated 261 good repeatability and stability of the instrument. Seven switching samples were performed with 50 µg 262 L⁻¹ NH₄⁺ standard solution and R1, after calculating 10-90 % of the full signal after a change in

- 263 concentration, the time response was approximately 140 s, which was much quicker than the method
- described by Bianchi et.al (measured to be 10 min) (Bianchi et al., 2012).



266


Fig. 4. Calibration curves of standard solution with the same concentration gradient 4 times

267

Table 1. Linear regression with the same concentration gradient 4 times

3.4 Setup of the temperature

269 High temperature can accelerate the reaction process and achieve better measurement accuracy and 270 precision. The voltage signal decreased with increasing temperature; conversely, the absorbance increased with temperature. According to the flow rate (gas flow rate of 0.70 L min⁻¹, liquid flow rate of 271 272 0.49 ml min⁻¹), the detection limit of our SAC-LOPAP can reduce to less than 50 ppt when the absorbance 273 of 50 μ g L⁻¹ NH₄⁺ standard solution reached 0.15 or more. However, if the temperature is too high, there 274 is a danger that the pipeline interface of the instrument will fall off. Considering the continuous delivery 275 of solutions (the stability of pipeline connections) and the detection limit (lower than 50 ppt), 55 °C was selected as the best reaction operating temperature of the instrument, at which sufficient absorbance could 276 277 be achieved to detect concentrations of ammonia gas. The standard solution entered the solution system 278 instead of the stripping solution, then the measured absorbance values were used as absorbance-standard






Fig. 5. Standard solution and absorbance liner range test, to get a measurement range

F_l , ml min ⁻¹	F_g , L min ⁻¹	$C(NH_3)_{min}$, ppt, $[NH_4^+] = 40.9 \text{ ng } L^{-1}$	$C(NH_3)_{max}$, ppb, $[NH_4^+] = 100 \ \mu g \ L^{-1}$
0.25	1	14.47	35.38
0.35	0.85	23.84	58.29
0.5	0.7	41.35	101.11
0.75	0.4	108.55	265.41
1.1	0.3	212.28	519.02

301 **4. Comparison in urban Beijing**

302 The time series of the concentration of NH₃ during the inter-comparison period of Picarro and SAC-303 LOPAP were presented in Fig. 6a. There were a few data gaps for the above instruments caused by 304 calibration operations and instrument maintenance. Instruments display similar temporal features for 305 NH₃ concentrations over the duration of the study. In this study, the concentration of our instrument 306 ranged from 1.3 ppb to 47.86 ppb with an average of 12.64 ± 8.63 ppb, which was close to the 307 concentrations of Picarro (12.76 \pm 8.57 ppb). The response speed was similar, indicated that SAC-308 LOPAP responded in time to rapid changed in NH₃ concentration. The diurnal variation results showed 309 that the concentrations measured by the two instruments were very similar, with our instrument slightly 310 lower than Picarro by 0.72 ppb (Fig. 6b). Furthermore, relatively good correlations for the NH₃ data 311 observed by these instruments were achieved over a large dynamic range of concentration with a slope 312 of 1.00 and an R^2 of 0.96 (Fig. 6c). We found that most of the time there were good correlations between 313 the two instruments within one day except for the data of 23th and 30th September. The regression slope 314 for all the days with higher and lower slopes are 1.46 and 0.72, respectively. We performed in-situ testing 315 of both systems with a cylinder, we produced NH₃ concentrations of about 1.32, 4.95, 9.59, 17.90 and 316 54.96 ppb. Fig. 6d showed regression analyses of the NH₃ standard gas concentrations obtained with the 317 two instruments. The NH₃ concentrations measured by picarro and our instrument were strongly 318 correlated, with a slope of 1.01 and an R^2 of 0.99.

In general, our instrument run relatively stable with the standard deviation of zero gas during the one month of observations being within 26 ppt (Picarro: 23 ppt), which was far below our detection limit. Furthermore, the drift of SAC-LOPAP and Picarro at 4.95 ppb were 3.5% and 2.8%, while the drifts of 54.96 ppb were 1.5% and 0.7%, which meant that our instrument could keep steady for a long time and 323 it could be used for the continuous online measurement of low concentration of ambient air. More

detailed inter-comparison for these NH₃ instruments will be analyzed in a future publication.




Fig. 6. (a) Time series of NH₃ concentration during the comparison, (b) Diurnal variation of NH₃ concentrations observed by Picarro and SAC-LOPAP, (c) Regression analysis of the NH₃ concentrations observed by Picarro and SAC-LOPAP, and (d) Regression analysis of different concentrations of Picarro and SAC-LOPAP NH₃ standard gases.

5. Conclusions

325

Ammonia (NH_3) in the atmosphere affects the environment and human health and is therefore increasingly recognized by policy makers as an important air pollutant that needs to be mitigated. The accurate and precise detection of ambient NH_3 concentrations is therefore an urgent need for the exploration of secondary pollution at the regional scale in China.

At the present stage, ambient NH₃ measurements at many supersites are still done with spectroscopic, mass spectrometric and wet chemical methods, which are restricted by the high detection limit and lower time resolution. In this study, we provide an online NH₃ monitoring system based on wet chemistry stripping and long path absorption photometer of atmospheric NH₃, our new SAC-LOPAP system has several significant improvements: one is the optimization of reaction conditions. The low concentration but higher flow rate of solutions decreases the precipitate's production, and the cooling buffer tube and 341 the filter trap most of the precipitates. The others are the constant temperature module and liquid flow 342 controller. The constant temperature module in the system reduces the influence of ambient temperature 343 on the reaction process and color degree. Similarly, adding a liquid flow controller is helpful to the 344 stability of the flow rate and further increases the stability of the reaction process. These improvements 345 reduce the system error and significantly increase the sustainability of SAC-LOPAP operation. Our 346 instrument reached a detection limit of about 40.5 ppt with a stripping liquid flow rate of 0.49 ml min⁻¹ and a gas sample flow rate of 0.70 L min⁻¹ in the current condition, and the measuring range of the 347 348 instrument is 40.5 ppt to 99.1 ppb. Our system has also been characterized in a laboratory setting where 349 we can measure low concentrations. SAC-LOPAP and Picarro were compared in urban areas for a month with relatively good agreement ($R^2 = 0.967$). In addition, the diurnal variation results showed that the 350 351 concentrations of the two instruments were very similar. Therefore, we conclude that our update of the 352 ammonia measurement experimental framework has been successful. However, more research about 353 field measurement and comparison is needed to verify the equipment's performance in routine 354 observation, and the influence of particulate ammonium on the results of NH₃ detection also requires 355 further study.

356

357 Data availability. The datasets used in this study are available from the corresponding author upon
 358 request (hbdong@pku.edu.cn).

359

360 Author contributions. H.B.D. designed the study. S.S.T., K.X.Z. set up and characterized the instrument, 361 analyzed the data and wrote the paper with the input of H.B.D. As co-authors, S.S.T and K.X.Z. 362 contributed equally to this paper. All authors contributed to the field measurements, discussed and 363 improved the paper.

364

365 **Competing interests**. The authors declare that they have no conflict of interest.

366

Acknowledgments. This work was supported by special fund of State Key Joint Laboratory of
 Environmental Simulation and Pollution Control (Grants No.22Y04ESPCP)

369

370 **References.**

- 371 Ajay and Beniwal.: Electrospun SnO2/PPy nanocomposite for ultra-low ammonia concentration detection at room
- 372 temperature, Sensors & Actuators B Chemical, 296, 126660.1-126660.9, https://doi.org/10.1016/j.snb.2019.126660, 2019.
- 373
- 374 Bae, M. S., Demerjian, K. L., Schwab, J. J., Weimer, S., Hou, J., Zhou, X., Rhoads, K., and Orsini, D.: 375 Intercomparison of Real Time Ammonium Measurements at Urban and Rural Locations in New York, Aerosol
- 376 Science & Technology, 41, 329-341, https://doi.org/10.1080/02786820701199710, 2007.
- 377 Behera, S. N., Sharma, M., Aneja, V. P., and Balasubramanian, R.: Ammonia in the atmosphere: a review on
- 378 emission sources, atmospheric chemistry and deposition on terrestrial bodies, Environ Sci Pollut Res Int, 20, 8092-
- 379 8131, https://doi.org/10.1007/s11356-013-2051-9 2013.
- 380 Benson, D. R., Markovich, A., Al-Refai, M., and Lee, S. H.: A Chemical Ionization Mass Spectrometer for ambient 381 measurements of Ammonia, Atmos. Meas. Tech., 3, 1075-1087, https://doi.org/10.5194/amt-3-1075-2010, 2010.
- 382 Berden, G., Peeters, R., Meijer, G., and Apituley, A.: Open-path trace gas detection of ammonia based on cavity-
- 383 enhanced absorption spectroscopy, Applied Physics B 71, 231-216, https://doi.org/10.1007/s003400000302, 2000.
- 384 Bianchi, F., Dommen, J., Mathot, S., and Baltensperger, U.: On-line determination of ammonia at low pptv mixing 385 ratios in the CLOUD chamber, Atmospheric Measurement Techniques, 5,7(2012-07-19), 5, 1719-1725, 386 https://doi.org/10.5194/amt-5-1719-2012, 2012.
- 387 Bu, X., Xie, Z., Liu, J., Wei, L., and Ren, H.: Global PM2.5-attributable health burden from 1990 to 2017: estimates 388 the Global Burden of Disease Study 2017, Environmental Research, from 197, 111123, 389 https://10.1016/j.envres.2021.111123, 2021.
- 390 Callahan, M. R., Rose, J. B., and Byrne, R. H.: Long pathlength absorbance spectroscopy: trace copper analysis using a 4.4 m liquid core waveguide, Talanta, 58, 891-898, https://doi.org/10.1016/S0039-9140(02)00403-4 2002. 391
- 392 Chen, X., Oro, Y., Tanaka, K., Takenaka, N., and Bandow, H.: A New Method for Atmospheric Nitrogen Dioxide 393 Measurements Using the Combination of a Stripping Coil and Fluorescence Detection, Analytical Sciences the 394 Society 20, 1019-1023, International Journal of the Japan for Analytical Chemistry,
- 395 https://doi.org/10.2116/analsci.20.1019, 2004.
- 396 Dong, H. B., Zeng, L. M., Hu, M., Wu, Y. S., Zhang, Y. H., Slanina, J., Zheng, M., Wang, Z. F., and Jansen, R.:
- 397 Technical Note: The application of an improved gas and aerosol collector for ambient air pollutants in China, 398 ATMOSPHERIC CHEMISTRY AND PHYSICS, 12, 10519-10533, https://doi.org/10.5194/acp-12-10519-2012, 399 2012.
- 400 Fowler, D., Steadman, C. E., Stevenson, D., Coyle, M., Rees, R. M., Skiba, U. M., Sutton, M. A., Cape, J. N., Dore,
- 401 A. J., and Vieno, M.: Effects of global change during the 21st century on the nitrogen cycle, Atmospheric chemistry
- 402 and physics, 15, 13849-13893, https://doi.org/10.5194/acp-15-13849-2015, 2015.
- 403 Gao, X., Koutrakis, P., Coull, B., Lin, X., Vokonas, P., Schwartz, J., and Baccarelli, A. A.: Short-term exposure to 404 PM2.5 components and renal health: Findings from the Veterans Affairs Normative Aging Study, Journal of
- 405 Hazardous Materials, 420, 126557, https://10.1016/j.jhazmat.2021.126557, 2021.
- 406 Gong, D., Liucheng, L. I., Baozeng, L. I., Liping, D., Wang, Y., Yanhua, M. A., Zhang, Z., and Jin, Y.: NH3 407 measurement based on cavity enhanced absorption spectroscopy, Laser Technology, 5, 664-668, 408 https://doi.org/10.7510/jgjs.issn.1001-3806.2017.05.009, 2017.
- 409 Heland, J., Kleffmann, J., Kurtenbach, R., and Wiesen, P.: A new instrument to measure gaseous nitrous acid (HONO)
- 410 in the atmosphere, Environmental Science & Technology, 35, 3207-3212, 2001.
- 411 Ianniello, A., Spataro, F., Esposito, G., Allegrini, I., Hu, M., and Zhu, T.: Chemical characteristics of inorganic
- 412 ammonium salts in PM2.5 in the atmosphere of Beijing (China), Atmospheric Chemistry & Physics, 11, 10803-
- 413 10822, https://doi.org/10.5194/acp-11-10803-2011, 2011.

- 414 Janson, R., Rosman, K., Karlsson, A., and H.-C.HANSSON: Biogenic emissions and gaseous precursors to forest
- 415 aerosols, Tellus, 53, 423-440, <u>https://doi.org/10.1034/j.1600-0889.2001.d01-30.x</u>, 2010.
- 416 Khlystov, A., Wyers, G. P., Brink, H., and Slanina, J.: The steam-jet aerosol collector (SJAC), Journal of Aerosol
- 417 Science, 26, S111-S112, <u>https://doi.org/10.1016/0021-8502(95)96963-8</u> 1995.
- 418 Kleffmann, J., Heland, J., Kurtenbach, R., Lrzer, J. C., and Wiesen, P.: A new instrument (LOPAP) for the detection
- 419 of nitrous acid (HONO), Environmental Science and Pollution Research, 9, 48-54, 2002.
- 420 Klimczyk, M., Siczek, A., and Schimmelpfennig, L.: Improving the efficiency of urea-based fertilization leading to 421 reduction in ammonia emission, Science of The Total Environment, 771. 145483, 422 https://doi.org/10.1016/j.scitotenv.2021.145483, 2021.
- 423 Krom and Michael, D.: Spectrophotometric determination of ammonia: a study of a modified Berthelot reaction
- 424 using salicylate and dichloroisocyanurate, Analyst, 105, 305-316, <u>https://doi.org/10.1039/an9800500305</u>, 1980.
- Krupa, S. V.: Effects of atmospheric ammonia (NH3) on terrestrial vegetation: a review, Environmental Pollution,
 124, 179-221, <u>https://doi.org/10.1016/S0269-7491(02)00434-7</u> 2003.
- 427 Makkonen, U., Virkkula, A., Mantykentta, J., Hakola, H., Keronen, P., Vakkari, V., and Aalto, P. P.: Semi-
- 428 continuous gas and inorganic aerosol measurements at a Finnish urban site: comparisons with filters, nitrogen in
- 429 aerosol and gas phases, and aerosol acidity, Atmospheric Chemistry and Physics, 12, 5617-5631,
 430 <u>https://doi.org/10.5194/acp-12-5617-2012</u>, 2012.
- Martin, N. A., Ferracci, V., Cassidy, N., and Hoffnagle, J. A.: The application of a cavity ring-down spectrometer to
 measurements of ambient ammonia using traceable primary standard gas mixtures, Applied Physics B, 122, 219,
 https://doi.org/10.1007/s00340-016-6486-9, 2016.
- Mcmanus, J. B., Nelson, D. D., Shorter, J. H., Zahniser, M. S., and Faist, J.: Quantum cascade lasers for open- and
 closed-path measurement of trace gases, Society of Photo-Optical Instrumentation Engineers (SPIE) Conference
 Series, 4817, https://doi.org/10.1117/12.452093 2002.
- 437 Ni, Ji-Qin, Heber, Albert, J., Lim, Teng, T., Diehl, and Claude: Ammonia Emission from a Large Mechanically-
- 438 Ventilated Swine Building during Warm Weather, Journal of Environment Quality, 29, 751-751,
 439 <u>https://doi.org/10.2134/jeq2000.00472425002900030010x</u>, 2000.
- 440 Nowak, J. B., Neuman, J. A., Kozai, K., Huey, L. G., Tanner, D. J., Holloway, J. S., Ryerson, T. B., Frost, G. J.,
- McKeen, S. A., and Fehsenfeld, F. C.: A chemical ionization mass spectrometry technique for airborne
 measurements of ammonia, Journal of Geophysical Research-Atmospheres, 112,
 https://doi.org/10.1029/2006jd007589, 2007.
- 444 Qu, Z. C., Li, B. C., and Han, Y. L.: Cavity ring-down spectroscopy for trace ammonia detection, Journal of Infrared
- 445 & Millimeter Waves, 31, 431-436, <u>https://doi.org/10.3724/SP.J.1010.2012.00431</u>, 2012.
- Searle and Phillip, L.: The berthelot or indophenol reaction and its use in the analytical chemistry of nitrogen. A
 review, Analyst, 109, 549-540, <u>https://doi.org/10.1039/an9840900549</u>, 1984.
- Sharma, S. K., Datta, A., Saud, T., Saxena, M., Mandal, T. K., Ahammed, Y. N., and Arya, B. C.: Seasonal variability
 of ambient NH3, NO, NO2 and SO2 over Delhi, Journal of Environmental Sciences, 22, 1023-1028,
 https://doi.org/10.1016/S1001-0742(09)60213-8 2010.
- 451 Sharma, S. K., Singh, A. K., Saud, T., Mandal, T. K., Saxena, M., Singh, S., Ghosh, S. K., and Raha, S.: Measurement
- of ambient NH3 over Bay of Bengal during W_ICARB Campaign, Annales Geophysicae, 30, 371-377,
 <u>https://doi.org/10.5194/angeo-30-371-2012</u>, 2012.
- 454 Sutton, M. A., Fowler, D., Burkhardt, J. K., and Milford, C.: Vegetation atmosphere exchange of ammonia: canopy
- 455 cycling and the impacts of elevated nitrogen inputs, Water Air & Soil Pollution, 85, 2057-2063,
- 456 <u>https://doi.org/10.1007/BF01186137</u> 1995.

- 457 Swati, A. and Hait, S.: Greenhouse Gas Emission During Composting and Vermicomposting of Organic Wastes –
- 458 A Review, CLEAN Soil Air Water, 46, 1700042.1-1700042. 3, <u>https://doi.org/10.1002/clen.201700042</u>, 2018.
- 459 Twigg, M. M., Berkhout, A. J. C., Cowan, N., Crunaire, S., Dammers, E., Ebert, V., Gaudion, V., Haaima, M., Häni,
- 460 C., John, L., Jones, M. R., Kamps, B., Kentisbeer, J., Kupper, T., Leeson, S. R., Leuenberger, D., Lüttschwager, N.
- 461 O. B., Makkonen, U., Martin, N. A., Missler, D., Mounsor, D., Neftel, A., Nelson, C., Nemitz, E., Oudwater, R.,
- 462 Pascale, C., Petit, J. E., Pogany, A., Redon, N., Sintermann, J., Stephens, A., Sutton, M. A., Tang, Y. S., Zijlmans,
- 463 R., Braban, C. F., and Niederhauser, B.: Intercomparison of in situ measurements of ambient NH3: instrument
- 464 performance and application under field conditions, Atmos. Meas. Tech., 15, 6755-6787, 465 https://doi.org/10.5194/amt-15-6755-2022, 2022.
- 466 Von, B. K., Braban, C. F., Famulari, D., Jones, S. K., Blackall, T., Smith, T., Blom, M., Coe, H., Gallagher, M., and
- 467 Ghalaieny, M.: Field inter-comparison of eleven atmospheric ammonia measurement techniques, Atmospheric
- 468 Measurement Techniques, 3, 91-112, https://doi.org/10.5194/amt-3-91-2010, 2009.
- Wang, Ruyu, Ye, Xingnan, Liu, Yuxuan, Li, Haowen, Yang, and Xin: Characteristics of atmospheric ammonia and
 its relationship with vehicle emissions in a megacity in China, Atmospheric Environment, 182, 97-104,
 https://doi.org/10.1016/j.atmosenv.2018.03.047, 2018.
- 472 Wang, Y., Zhang, Q. Q., He, K., Zhang, Q., and Chai, L.: Sulfate-nitrate-ammonium aerosols over China: response
- to 2000–2015 emission changes of sulfur dioxide, nitrogen oxides, and ammonia, Atmospheric Chemistry and
- 474 Physics, 13, 2635-2652 <u>https://doi.org/10.5194/acp-13-2635-2013</u>, 2013.
- 475 Wen, Liang, Xue, Likun, Wang, Xinfeng, Xu, Caihong, Chen, and Tianshu: Summertime fine particulate nitrate
- 476 pollution in the North China Plain: increasing trends, formation mechanisms and implications for control policy,
- 477 Atmospheric Chemistry & Physics, 18, 11261-11275, <u>https://doi.org/10.5194/acp-2018-89</u>, 2018.
- Whitehead, J. D., Twigg, M., Famulari, D., Nemitz, E., Sutton, M. A., Gallagher, M. W., and Fowler, D.: Evaluation
 of Laser Absorption Spectroscope Techniques for Eddy Covariance Flux Measurements of Ammonia,
 Environmental Science & Technology, 42, 2041-2046, <u>https://doi.org/10.1021/es071596u</u>, 2008.
- 481 Yao, W., Byrne, R. H., and Waterbury, R. D.: Determination of Nanomolar Concentrations of Nitrite and Nitrate in
- 482 Natural Waters Using Long Path Length Absorbance Spectroscopy, Environmental Science and Technology, 32,
- 483 2646-2649, <u>https://doi.org/10.1021/es9709583</u> 1998.
- Yokelson, R., Christian, T., Bertschi, I., and Hao, W.: Evaluation of adsorption effects on measurements of ammonia,
 acetic acid, and methanol, Journal of Geophysical Research Atmospheres, 108,
 https://doi.org/10.1029/2003JD003549, 2003.
- 487 Yu, H. and Lee, S. H.: Chemical ionisation mass spectrometry for the measurement of atmospheric amines,
- 488 Environmental Chemistry, 9, 190-201, <u>https://doi.org/10.1029/2003JD00354910.1071/en12020</u>, 2012.