Local comparisons of tropospheric ozone: Vertical soundings at two neighbouring stations in Southern Bavaria

- Thomas Trickl¹, Martin Adelwart², Dina Khordakova³, Ludwig Ries⁴, Christian Rolf³,
 Michael Sprenger⁵, Wolfgang Steinbrecht² and Hannes Vogelmann¹
- ¹Karlsruher Institut für Technologie, Institut für Meteorologie und Klimaforschung (IMK-IFU), Kreuzeck bahnstr. 19, D-82467 Garmisch-Partenkirchen, Germany
- ⁷ ²Deutscher Wetterdienst, Meteorologisches Observatorium, Albin-Schwaiger-Weg 10, 82383 Hohenpeißenberg,
 ⁸ Germany
- 9 ³Forschungszentrum Jülich, IEK-7, Wilhelm-Johnen-Straße, 52425 Jülich, Germany

⁴Umweltbundesamt II 4.5, Plattform Zugspitze, GAW-Globalobservatorium Zugspitze-Hohenpeißenberg,
 Schneefernerhaus, 82475 Zugspitze, Germany

⁵Eidgenössische Technische Hochschule (ETH) Zürich, Institut für Atmosphäre und Klima, Universitätstraße 16,
 8092 Zürich, Switzerland

14 Correspondence to: Dr. Thomas Trickl, thomas@trickl.de, Thomas-Knorr-Str. 47, D-82467 Garmisch 15 Partenkirchen, Germany; tel. +49-8821-50283; Dr. Hannes Vogelmann, hannes.vogelmann@kit.edu, Karlsruher
 16 Institut für Technologie, IMK-IFU, Kreuzeckbahnstr. 19, D-82467 Garmisch-Partenkirchen, Germany; tel: +49-

- 17 8821-258
- 18 Abstract. In this study ozone profiles of the differential-absorption lidar at Garmisch-Partenkirchen are 19 compared with those of ozone sondes of the Forschungszentrum Jülich and of the Meteorological Observatory 20 Hohenpeißenberg (German Weather Service). The lidar measurements are quality assured by the highly accurate 21 nearby in-situ ozone measurements at the Wank (1780 m a.s.l.) and Zugspitze (2962 m a.s.l.) summits and at the 22 Global Atmosphere Watch station Schneefernerhaus (UFS, 2670 m a.s.l.), at distances of 9 km or less from the 23 lidar. The mixing ratios of the lidar agree with those of the monitoring stations with a standard deviation (SD) of 24 1.5 ppb, and feature a slight positive offset of 0.6 ppb \pm 0.6 ppb (SD) conforming to the known -1.8-% 25 calibration bias of the in-situ instruments. Side-by-side soundings of the lidar and electrochemical (ECC) sonde 26 measurements in February 2019 by a team of the Forschungszentrum Jülich shows small positive ozone offsets 27 for the sonde with respect to the lidar and the mountain stations (0.5 to 3.4 ppb). After applying an altitude-28 independent bias correction to the sonde data an agreement to within just ±2.5 ppb in the troposphere was found, 29 which we regard as the wintertime uncertainty of the lidar. We conclude that the recently published uncertainties 30 of the lidar in the final configuration since 2012 are realistic and rather small for low to moderate ozone 31 concentrations. Comparisons of the lidar with the Hohenpeißenberg routine measurements with Brewer-Mast 32 sondes are more demanding because of the distance of 38 km between both sites implying significant ozone 33 differences in some layers, particularly in summer. Our comparisons cover the three years September 2000 to 34 August 2001, 2009 and 2018. A slight negative average offset (-3.64 ppb ± 3.72 ppb (SD)) of the sondes with 35 respect to the lidar is found. We conclude that most Hohenpeißenberg sonde data could be improved in the 36 troposphere by recalibration with the Zugspitze station data (1978 to 2011 summit, afterwards UFS). This would 37 not only remove the average offset, but also greatly reduce the variability of the individual offsets. The 38 comparison for 2009 suggests a careful partial re-evaluation of the lidar measurements between 2007 and 2011 39 for altitudes above 6 km where occasionally a negative bias occurred.
- 40 Key words: Tropospheric ozone, ozone sonde, ozone lidar, differential absorption

41 **1. Introduction**

46

42 The development of tropospheric ozone has been studied over more than a century (e.g., Gaudel et al., 2018;

- 43 Tarasick et al., 2019). For many decades, balloon-borne ozone sondes have been a primary work horse of ozone
- 44 profiling. Their measurement principle is based on the oxidation of iodide (I^{-}) to iodine (I_2) by ozone in a wet-
- 45 chemical potassium iodide (KI) cell. Between cathode and anode of the wet-chemical cell, the oxidation reaction
- 47 stations have used the so-called ECC (electro-chemical-cell) sonde type (Komhyr 1969; 1995), featuring two

drives an electrical current which can be measured (two electrons per ozone molecule). Recently, nearly all

- 48 cells with different potassium iodide concentrations (anode and cathode cell). Only the Hohenpeißenberg station,
- 49 discussed here, still uses the older-type Brewer-Mast sondes (Brewer and Milford, 1960), which uses one cell
- 50 only (with a platinum cathode and a silver anode), and a less efficient pump design (Steinbrecht et al., 1998).
- 51 Ozone sondes have been characterized in numerous studies, both in flight (e.g., Attmannspacher and Dütsch, 52 1981; De Muer and Malcorps, 1984; Beekmann et al., 1994; Kerr et al., 1994; Jeannet et al., 2007; in recent
- 53 years: Gaudel et al., 2015; Van Malderen et al., 2016; Deshler et al., 2017; Tarasick et al., 2021; Ancellet et al.,
- 54 2022; Stauffer et al., 2022), and in a laboratory simulation chamber (Smit et al., 2007, 2014, 2021). Generally,
- 55 the relative uncertainty of individual ECC soundings for ozone in the mid-latitude troposphere is about 5 to 10%
- 56 (Logan et al., 2012; Smit et al., 2014; Tarasick et al., 2016, 2019). Following rigorous best practices, 5%
- accuracy can be achieved (Vömel et al., 2020; Smit et al., 2021; Tarasick et al., 2019; 2021). For Brewer-Mast
 soundings, the relative uncertainty in the troposphere is slightly higher, about 10 to 15% (Stübi et al., 2008; Smit
- et al. 2014; Tarasick et al., 2016, 2019). For tropospheric ozone from Canadian Brewer-Mast soundings prior to
- 60 1980 Tarasick et al. (2002, 2016) found a negative bias of about 20 % compared to ECC soundings.
- The ozone soundings at the Meteorological Observatory Hohenpeißenberg (MOHp) of the German Weather Service (Deutscher Wetterdienst, DWD) in Southern Bavaria have been routinely carried out since November 1966, yielding one of the longest ozone-sonde time series. Brewer-Mast ozone sonde data tend to have a low bias above about 25 km altitude (Steinbrecht et al., 1998). European Brewer-Mast stations have generally used a much more extensive preparation procedure for their sondes (Claude et al. 1987), and no significant tropospheric bias has been reported for their routine Brewer-Mast soundings (de Backer et al. 1998; Stübi et al. 2008; Logan
- 67 et al., 2012), as well as in chamber experiments (Smit et al., 2014).
- 68 Routine measurements with ozone sondes yield time series free of a fair-weather sampling bias. However, the 69 balloon ascents take place at intervals of several days. Ozone profiles at short intervals (less than one minute to 70 several minutes) can be provided by lidar sounding, but are limited to clear atmospheric conditions. Lidar 71 measurements can generate altitude-time curtain plots and, thus, give much better insight into the impact of
- 72 atmospheric transport (e.g., Browell et al., 1987; Ancellet et al., 1991; Langford et al., 1996).
- 73 At IFU (Fraunhofer-Institut für Atmosphärische Umweltforschung; now: Karlsruher Institut für Technologie,
- 74 IMK-IFU) in Garmisch-Partenkirchen (Germany), a differential-absorption lidar (DIAL) with a particularly wide
- 75 operating range from next to the ground up to the upper troposphere was completed in 1990 in the framework of
- 76 the TESLAS (Tropospheric Environmental Studies by Laser Sounding) subproject of EUROTRAC (TESLAS,
- 1997; EUROTRAC, 1997, Kempfer et al., 1994). Subsequently, the system was applied for a full year (1991)
- 78 within the TOR (Tropospheric Ozone Research; Kley et al., 1997) subproject of EUROTRAC (Carnuth et al.,
- 2002). The operating range of this system was extended upwards to roughly 15 km in 1994 by introducing three-
- 80 wavelength operation (Eisele et al., 1999). Due to its design, the IFU ozone DIAL features particularly low
- 81 uncertainties (Trickl et al., 2020a).

- 82 Until 2003 the system was used for individual research projects. Between 2007 and 2018 routine measurements
- 83 took place, parallel to lidar measurements of water vapour (Vogelmann and Trickl, 2008) and aerosol (Trickl et
- 84 al., 2020b). The complementary information from these instruments has made possible a large number of
- 85 investigations related to atmospheric transport. The IFU ozone DIAL was recently fully described by Trickl et al.
- 86 (2020a).
- 87 The distance between MOHp and IFU is just 38 km which offers a good chance for comparisons. However, such 88 a comparison must be made with care since the atmospheric variability is high on a rather small temporal and 89 spatial scale (Vogelmann et al., 2011; 2015), mostly caused by the advection of air masses from rather different 90 source region and altitudes, with different concentrations (e.g., Stohl and Trickl, 1999; Trickl et al., 2003; Trickl 91 et al., 2011). The variability of the vertical distribution of ozone measurements rarely yields very strong 92 concentration changes, but the concentration changes are extreme for water vapour. Our lidar measurements of 93 water vapour exhibit a concentration span of more than two decades, with minima of the relative humidity (RH) 94 clearly below 1 % in layers descending from the stratosphere (Trickl et al., 2014; 2015; 2016; Klanner et al.,
- 95 2021).
- 96 Comparisons between the MOHp sonde and the IFU ozone lidar were made in the second half of the 1990s and 97 in 2001, after the first upgrading of the lidar A few of these comparisons in 1996 and 1997 were published by 98 Eisele et al. (1997; 1999). For the six cases with supposedly sufficient air-mass matching a principal agreement 99 in the free troposphere to within 5 % prevailed with occasional departures of the order of 10 %. Several 100 unpublished comparisons in 2001 showed principal agreement, but also some structural issues due to focussing
- 101 on stratospheric air intrusions with the STACCATO project (Stohl et al., 2003).
- 102 Afterwards just routine comparisons with the nearby summit stations were made. Until 2010 the lidar results 103 were compared with the long-term measurements at Wank and Zugspitze. Apart from occasional orographically 104 induced deviations an agreement mostly to within ± 2 ppb was found. After these in-situ measurements ended 105 (2011) the lidar measurements were compared with the ozone measurements at the Schneefernerhaus high-106 altitude station (Umweltforschungsstation Schneefernerhaus, UFS, 2671 m a.s.l.). UFS is located just below the
- 107 Zugspitze summit. Mostly a similar agreement was found.
- However, the need for a validation of the lidar also at higher altitudes has been obvious. Such an effort became more and more attractive with the growing technical performance of the system. In addition, hints on ozone differences between the Zugspitze (2962 m a.s.l..) in-situ data and the MOHp values (H. E. Scheel, personal communication around 2010) for 3 km a.s.l. have led to a revived interest in a thorough comparison. There have been speculations about an influence of a different air composition outside the mountains at low altitudes up to a
- 113 few kilometres.
- In this paper we first characterize the lidar by side-by-side ascents of ozone-sondes by a team of the Forschungszentrum Jülich (FZJ). This effort, also based on the measurements at UFS, demonstrates a high performance of the DIAL within the entire free troposphere, at least under winter-time conditions. Then, based on this performance, we give a statistical assessment for the measurements at IFU and MOHp for the year 2018. For this year we achieved the best coverage by DIAL measurements. This allows us to make an air-mass related data selection to improve the comparison. Finally, we also compare lidar and MOHp sonde for two earlier
- 120 development phases of the lidar, for which ozone reference data at the local summit stations Wank (1780 m
- 121 a.s.l.) and Zugspitze exist.
- 122

123 2. Methods

124 2.1 Brewer-Mast sonde system at Hohenpeißenberg

125 MOHp (975 m a.s.l., 47.80 N, 11.00 E) is located on an isolated mountain outside the Alps, 38 km to the north of 126 IMK-IFU and 50 km to the south-west of Munich. Brewer-Mast ozone sondes have been launched on a regular 127 basis since November 1966. The sondes undergo a rigorous preparation procedure (Claude et al. 1987), which 128 has remained essentially unchanged since the early 1970s. From 1995 to 2005, Vaisala RS80 radiosondes and a 129 Vaisala PC-CORA ground station have been used in combination with the ozone sondes. This was changed to 130 Vaisala RS92 radiosondes and DigiCora III MW31 ground equipment in 2005, to MW41 ground station in 2018, 131 and to Vaisala RS41 radiosondes in 2019. The standard processing does not subtract a background current, but 132 ozone sondes with non-negligible background current on the ground (corresponding to more than 2.5 ppb ozone) 133 are not flown. The background of most sondes launched is well below this threshold. The pump temperature is 134 assumed to be constant at 300 K, which compensates to some degree for a too weak pump correction in the 135 stratosphere (Steinbrecht et al., 1998). The time lag is comparable to that of ECC sondes (about 20 s; see Vömel 136 et al., 2020). A time-lag correction is not applied, but this is not critical outside regions with steep ozone 137 gradients since the corresponding vertical shift is just of the order of 0.1 km. Each ozone profile is adjusted by 138 multiplication with an altitude-independent correction factor (typically around 1.08, standard deviation 5 %), so 139 that the total ozone column estimated from the sounding (including an extrapolation above approximately 30 140 km) matches the more accurate total ozone measurement from on-site Dobson or Brewer spectrometers, or from 141 satellite instruments. This so-called "Dobson correction" generally improves that accuracy of the ozone sounding 142 data in the stratosphere, but may introduce a small bias in the tropospheric data of some soundings (e.g., Stübi et

143 al., 2008; Logan et al. 2012).

144 The MOHp ozone-sonde and radiosonde data are stored in the data base of the Network for the Detection of

145 Atmospheric composition change (NDACC), from where they were imported for the study presented here.

146 2.2 ECC sonde system of the Forschungszentrum Jülich (FZJ)

147 A mobile ballon-borne sonde system of FZJ was operated at IMK-IFU (at 730 m a.s.l.), in close vicinity to the 148 ozone DIAL (35 m), during the FIRMOS (Far Infrared Radiation Mobile Observation System) measurement 149 campaign (Klanner et al., 2020; Palchetti et al., 2021; Di Natale, 2021; Belotti et al., 2023). Several balloons 150 with cryogenic frostpoint hygrometers (CFH; Vömel et al., 2007; 2016), standard Vaisala RS-41-SGP 151 radiosondes (Vaisala et al., 2019), En-Sci ECC ozone sondes (Komhyr et al., 1995; Smit et al., 2007) and 152 COBALD backscatter sondes (Brabec, 2011) were launched. The data were transmitted to a ground station 153 installed for this campaign at the Zugspitze summit. The combined balloon payload is well tested and regularly

- 154 also used by the GCOS Reference Upper Air Network (GRUAN) (e.g., Dirksen et al., 2014).
- 155 We followed the standard operating procedures (SOP) of Smit et al. (2014) for the sonde preparation using a 156 solution composition of 1 % and 1/10 (one-tenth) buffer for best results with sondes from the manufacturer En-
- 157 Sci (Thompson et al., 2019).
- 158 For the analysis of the ECC data, the methods described by Vömel et al. (2020) are used, i.e., time lag correction
- 159 and background current correction. The overall uncertainty of the ozone measurements of the ECC sondes is 5%.
- 160 Due to the obstruction of the line of sight between between launch site at IMK-IFU and the ground station at the
- 161 summit by the Waxenstein mountain allowed data recording only from approximately 1500 m altitude upwards.
- 162 Therefore, we used the estimated ECC background current from the sonde preparation one day before a flight as

- 163 starting value for the background correction instead of the actual measured profile from ground up to 1500 m.
- 164 This results in an additional uncertainty in the lower part of the profile (2 to 3 km a.s.l.).

165 2.4 IFU ozone DIAL system

166 The ozone DIAL of IMK-IFU (Garmisch-Partenkirchen), located at 47.477 N, 11.064 E, and 740 m a.s.l., has 167 been developed and optimized since 1988 (Kempfer et al., 1994; Trickl et al., 2020a). It is based on a krypton 168 fluoride excimer laser, operated at 400 mJ per pulse (40 W) of narrowband radiation at 248.5 nm, two 169 Newtonian receiving telescopes (diameter of the primary mirrors: 0.13 m and 0.5 m) and 1.1-m grating 170 spectrographs for wavelength separation. Efficient stimulated Raman shifting in hydrogen and deuterium yields 171 emission at the three operating wavelengths 277.2 nm, 291.8 nm and 313.2 nm. The shorter-wave spectral 172 components are absorbed by ozone ("on" wavelengths), that at 313.2 nm ("off" or reference wavelength) is 173 almost outside the absorption region of O₃. The laser system is operated with a repetition rate of 99 Hz which 174 allows a short data-acquisition time of 41 s for the maximum number of 4096 laser shots accepted by the 24-bit 175 memory of the electronics. More shots are advisable under noisy daytime conditions in summer, but a longer

176 acquisition was prevented by laser issues.

177 The data evaluation is based on differentiating the backscatter signals, which is highly sensitive to the noise and 178 imperfections of the raw data (stored in 7.5-m bins). Therefore, the generated ozone profiles are smoothed with a 179 numerical filter. The noise fraction in the strongly decreasing backscatter signal grows with altitude. Thus, the

smoothing interval must be dynamically enhanced towards the tropopause (yielding a vertical resolution 0.05 to
0.5 km). The entire procedure is described in detail by Trickl et al. (2020a).

182 The shortwave 277.2-nm emission yields particularly accurate measurements, but the strong extinction of this 183 radiation by ozone limits the range to about 8 km. The performance in the two 277.2-nm channels is robust with 184 respect to minor misalignment, with uncertainties of about 2 to 4 ppb up to 5 km (the estimated uncertainties are 185 listed in Table 4 of Trickl et al. (2020a)). This is not the case for 291.8 nm where the optical alignment must be 186 controlled with care because of less tight focussing into the entrance slit of the far-field spectrograph. In 187 addition, the 291.2-nm backscatter signal is three times noisier than that for 277.2-nm which necessitates 188 stronger smoothing of the retrieved ozone profiles (Trickl et al., 2020a) For 5 to 8 km we specify uncertainties of 189 3 to 7 ppb. The noise of the 313.2-nm signal becomes important at large distances. As a consequence, the 190 uncertainty of the ozone mixing ratio can be become rather high in the upper troposphere and the tropopause 191 region, in particular in summer due to the stronger loss of signal caused by the higher levels of ozone. 192 Sometimes the uncertainty just below the tropopause can even exceed 10 ppb.

193 The DIAL data processing is made for different wavelength combinations (Eisele and Trickl, 2005). By 194 comparing the resulting ozone profiles an internal quality control can be achieved. The optical alignment is

195 optimized immediately after detecting an ozone mismatch in the first quicklook data evaluation. Just the laser

196 beam overlap of the different wavelength components (Trickl et al., 2020a) and the beam pointing must be 197 optimized.

- 198 The calibration of the ozone lidar measurements has been based from the very beginning (1991) on the accurate
- 199 temperature-dependent ozone absorption cross sections of the University of Reims (Daumont et al., 1992;
- 200 Malicet et al., 1995). These cross sections were verified for four wavelengths below 300 nm by Viallon et al.
- 201 (2015) to within ± 0.06 %. In the presence of aerosol an aerosol correction is made with the algorithms of Eisele
- and Trickl (2005). This correction is rather robust for the wavelength pair 277 nm 292 nm because of the strong
- 203 absorption at the short "on" wavelength and the moderate wavelength difference (Völger et al., 1996).

- 204 Meteorological data for calculating density and temperature profiles are taken from the Munich radiosonde
- 205 (station 10868). The retrieved 313-nm aerosol backscatter coefficients have been routinely stored in the data 206 base of the European Aerosol Lidar Network (EARLINET) since 2007.
- 207 After repeated system upgrading the final performance of the lidar was reached in late 2012. In the absence of 208 aerosol the far-field ozone could be evaluated with high reliability from the 291.9-nm signal alone, after 209 precisely modelling the air number density from radiosonde data (Trickl et al., 2020). In this way the influence 210 of the daytime noise caused by the high solar background in the 313-nm reference profiles in summer could
- 211 frequently be avoided.
- 212 During the final decade of the lidar operation a fitting procedure was applied in noisy situations in the upper
- 213 troposphere (i.e., under high-ozone conditions in summer). This procedure reduces unrealistic curvature of ozone
- 214 structures caused by enhanced data smoothing, and, thus, abrupt concentration changes (in particular at the 215 tropopause) visible in the raw data are reproduced in the mixing ratio.
- 216 From 1991 to 2003 the DIAL was operated for focussed research projects. Routine measurements took place
- 217 from 2007 to 2018, until 2015 parallel to measurements with a water-vapour DIAL (Trickl et al., 2014, 2015,
- 218 2016, 2020b). In 2012 the highest data quality was finally reached, which included significant improvements for
- 219 the near-field telescope (Trickl et al., 2020a). Thus, the conditions for a meaningful system validation were
- 220 obtained. The operation was discontinued in February 2019, after the retirement of the first author of this paper.

221 2.5 High-elevation surface observations

- 222 Quality-assured ozone measurements at the summit stations Wank (1780 m a.s.l., 7.0 km to the north-east of
- 223 IMK-IFU, 47.511° N, 11.141° E) and Zugspitze (2962 m a.s.l., 8.4 km to the south-west of IMK-IFU, 47.421° N,
- 224 10.986° E) took place from 1978 to 2012. Since the 1990s, two or three TE 49 ozone analysers (Thermo 225

Environmental Instruments, USA) were operated simultaneously at each station. These instruments are based on

- 226 ultraviolet (UV) absorption at 253.65 nm. Several comparisons using transfer standards (O₃ calibrators TE 49
- 227 PS) were made with the World Meteorological Organization (WMO) Global Atmosphere Watch (GAW)
- 228 reference instrument kept at the WMO/GAW calibration centre operated by EMPA, Switzerland (Klausen et al.,
- 229 2003). The most recent comparison was conducted in June 2006 and confirmed that the Zugspitze O_3 data are on
- 230 the GAW scale.
- 231 Apart from the two mountain stations measurements were performed also at IFU at about 740 m a.s.l. (47.477°
- 232 N, 11.064° E). This laboratory was adjacent to that of the ozone DIAL.
- 233 At UFS (0.70 km to the south-east of Zugspitze, 47.417°, 10.980° E) ozone has been continuously measured
- 234 since 2002 by a team of the German Environment Agency (Umweltbundesamt, UBA) using TEI 49i instruments
- 235 (Thermo Electron Corporation). The gas inlet is at 2671 m a.s.l. For weekly and monthly calibration of the ozone
- 236 measurements a TEI 49C-PS station ozone calibrator was applied. This primary standard was annually adjusted
- 237 to the German ozone standard operated by UBA (UBA 204 SRP#29) that was adjusted via BIPM (Bureau
- 238 International des Poids et Mesures) in Paris to the NIST ozone reference standard of GAW. The measurements
- 239 were supported by a second instrument (Horiba APOA-370) which is equivalent to the TEI-49i. GAW
- 240 performance audits at the station for surface ozone took place in 2001, 2006, 2011 and 2020 (Zellweger et al.,
- 241 2001; 2006; 2011; 2020).
- 242 The uncertainty of the in-situ ozone measurements is ± 0.5 ppb with respect to the WMO standard (Hearn et al.,
- 243 1961). This fulfills the GAW requirement.

- 244 The ozone data for all sites are stored at half-hour intervals. The times are specified for the end of the averaging
- interval in Central European Time (CET, = UTC + 1 h). 1-h averages for the Zugspitze stations were made
- available to the World Data Center and the TOAR data base (Schultz et al., 2017). In the present study we use
- 247 data at half-hour time resolution. The ozone series at the two Zugspitze sites have been discussed on two recent
- 248 scientific studies (1970 to 2020; Parrish et al., 2019; Trickl et al., 2023).

249 2.6 LAGRANTO Trajectories

250 Fifteen-day backward trajectories were calculated with the Lagrangian Analysis tool (LAGRANTO; Sprenger 251 and Wernli, 2015; Wernli and Davies 1997). The driving wind fields are obtained from the ERA5 reanalysis 252 dataset (Hersbach et al., 2020), which we interpolated to a 0.5° latitude/longitude grid, and on 137 vertical hybrid 253 levels. The input ERA5 data are available at a one-hour temporal resolution; the output positions of the 254 trajectories are written at 15-min time interval to allow for a more refined analysis. The start coordinates of the 255 backward trajectories are 11.064 E, 47.477 N, and the start altitudes match the altitudes of interest in the 256 soundings (see Sect. 4). The start times of the trajectories correspond to the sounding times within five minutes. 257 Finally, the start times are also shifted by several hours relative to the sounding time to assess the sensitivity of 258 the trajectory calculation on time.

3. Results

260 The main problem in comparing vertical-sounding instruments is illustrated in Fig. 1 which shows several ozone 261 measurements at Garmisch-Partenkirchen and Hohenpeißenberg in the morning of 2 October 2017. The vertical 262 distributions during that period are characterized by a descending stratospheric intrusion layer (indicated by low 263 relative humidity) of rapidly diminishing width and significant changes at all altitudes on a short time scale. This 264 reveals a considerable spatial inhomogeneity of the air mass. The approximate agreement of lidar and 265 Hohenpeißenberg ozone sonde before 6:00 CET is, thus, to some extent fortuitous although a good matching of 266 the peak ozone mixing ratio in intrusion layers at both sites is quite frequently found. Different air masses at 267 different altitudes must be assumed as indicated by matching of the sonde ozone with lidar measurements at 268 different times. The spatial and temporal requirements for comparisons can be even of the order of 1 km and 15 269 min at times (see Introduction).

270 **3.1** Comparisons of the IFU ozone lidar and the Jülich ECC sonde

An optimum lidar validation became possible in early 2019. On 5 and 6 February 2019 a side-by-side instrument comparison took place at Garmisch-Partenkirchen as a contribution to the FIRMOS validation project of the European Space Agency. Two of the three balloons launched on 5 February were equipped with ozone sondes, while both ballons on 6 February carried an ozone sonde. The ascents took place during night-time because of comparisons of the CFH sondes with the water-vapour channel of the UFS Raman lidar that provides humidity

- 276 profiles up to at least 20 km (Klanner et al. 2021).
- 277 The first night of the campaign was clear. The conditions for the comparison were excellent: the sondes rose
- almost vertically up to 8.5 km and then slowly drifted to the south-east (Innsbruck), ideal for the tropospheric
- comparison. The balloons stayed within 20 km distance from IMK-IFU up to the tropopause (12.8 km a.s.l.) and
- remained within 30 km up to 20 km a.s.l. The launch times of the balloons on 5 February were 18:03 CET
- 281 (ascent to 16.147 km), 19:03 CET (29.475 km), and 23:00 CET (29.469 km). During the second night a cirrus
- 282 layer occurred in the upper troposphere which resulted in enhanced uncertainties of the DIAL data evaluation.

- 283 In Fig. 2 we present the results of the four comparisons made. The agreement between UFS and lidar is almost
- 284 perfect, as known from the routine comparisons with the elevated sites between 2007 and 2018 and a number of
- separate comparisons (Trickl et al., 2020a). For the first three sonde ozone profiles, very small, almost altitude-
- independent offsets exist (0.5 to 3.4 ppb). For the fourth sonde ascent at 23:33 CET on 6 February no
- simultaneous lidar measurement was made. Up to 4.8 km a pronounced positive offset of the sonde ozone profile
- with respect to the two earlier lidar measurement (at 18:33 and 19:00 CET) is seen, but the deviation at 2.67 km
- is just 2 ppb if one takes the 23:30-CET measurement at UFS as the reference. We are highly content that the
- 290 difference in ozone between sonde and lidar does not significantly change in the upper troposphere considering
- 291 the low differential absorption for the wavelength pair 292 nm 313 nm typically used above 6 km that implies
- a high sensitivity to potential technical imperfection.
- 293 In addition, we show in Fig. 2 the results of three humidity measurements with the UFS Raman lidar slightly 294 revised with respect to Klanner et al. (2021). For comparison, we added the water-vapour mixing ratios (MRs) 295 for the corresponding CFH sonde ascents of FZJ. The MRs indicate a high variability of the air composition on 296 both days, up to 7 km, with several rapidly changing dry layers. The variability grows with time, as can also be 297 concluded from the differences of Raman lidar and CFH sonde, caused by the 1-h measurement duration of the 298 lidar needed for good stratospheric data quality. Although the vertical concentration change is much less 299 pronounced in the ozone profiles, it is obvious that a good air-mass matching by the side-by-side ozone 300 soundings at IMK-IFU is crucial for the quality of the comparison achieved.
- As mentioned, on 6 February the quality of the lidar retrievals was deteriorated above 9 km by a layer of cirrus clouds, which required an aerosol correction. The increased level of ozone in this layer is remarkable, but is verified by the sonde. By contrast, Reichardt et al. (1996) reported full ozone depletion in a cirrus layer that we traced back to the surface of the Pacific Ocean where ozone destruction can be assumed to prevail (Kley et al., 1996). The fourth comparison shows less perfect agreement because the lidar measurements ended at 19:00 CET, hours before the last sonde ascent. This was the final measurement of the DIAL before its operation was terminated after almost three decades.
- Ozone profiles are also available for the descent of the balloons. The descents took place over Northern Italy and
 intersected different air masses. As a consequence, strong discrepancies are seen, and we do not include these
 data.
- 311 From the comparison of the vertical soundings with the in-situ measurements at UFS we conclude that the ozone
- 312 profiles of the lidar are slightly more quantitative than those of the sonde. The differences are rather constant as a
- 313 function of the altitude. This allows us to derive uncertainties of the ozone from the DIAL measurements after
- 314 subtracting the offsets of the individual sonde ascents. For quantifying the quality of the lidar measurements we
- took just the first three comparisons. We averaged the offset-corrected differences (Fig. 3; altitude grid 52.5 m).
- 316 The averages up to 9.2 km stay within ± 2.5 ppb (about ± 5 %). This approximately matches the performance of
- 317 the lidar at the station altitudes and now characterizes the winter-time specifications of the lidar also in the entire
- 318 free troposphere after 2011. This result justifies to use the lidar as a quality standard in the comparisons with the
- 319 MOHp Brewer-Mast sondes described in the following section.
- 320 The quality of the comparison shown in this section benefits from low to moderate ozone densities during the
- 321 cold season, which ensures limited absorption of the laser radiation within the troposphere. In Sect. 3.2 we assess
- the performance for all seasons.
- 323

324 3.2 Comparison of MOHp ozone soundings with IFU lidar and in-situ measurements for 2018

The routine measurements with the IFU ozone DIAL since 2007 exhibit rather different annual coverages, with gaps due to system damage or upgrading periods. Starting in late 2012 the final technical performance was reached. Retrieval strategies have been further improved. The best coverage of a single year was achieved in 2018 with a total of 587 measurements and 16 (March) to 79 (September) measurements per month. Therefore, we use this year for a thorough comparison with the MOHp ozone sonde. Because of the excellent performance of the lidar verified in Sect. 3.1 we use the lidar as the reference in this comparison, together with the ozone

- 331 mixing ratios from UFS.
- The sonde ascents at MOHp usually take place around 6:00 CET on Monday, Wednesday and Friday, in summer just on Monday and Wednesday. We found a total of 46 of these days on which early-morning lidar measurements exist, not later than around 10:00 CET. On 36 of these days MOHp soundings are available. Thirteen of the days provided particularly good conditions with favourable temporal proximity. In the figures shown in this paper we eliminated ozone profiles for times later than 10:00 CET during a given day.

337 Similar to the comparisons of lidar and ECC sondes the comparisons of the lidar with the MOHp Brewer-Mast 338 sondes reveal altitude-independent offsets of the mixing ratios. The sonde-to-sonde variations of the offsets are 339 larger than those of the ECC sondes, consistent with the considerable uncertainties of the Brewer-Mast sondes 340 specified by the literature (see Introduction). There is clearly an influence of layers with different ozone 341 concentrations at both sites, but also good agreement in wide altitude ranges up to the upper troposphere after 342 subtracting the offset. Because of this agreement it is hard to believe that the offsets are caused by systematic 343 atmospheric differences. It is more reasonable to assume an instrumental issue as an explanation of these shifts. 344 Furthermore, the frequently good matching of the high peak ozone in some of the stratospheric intrusion layers

345 demonstrate the absence of concentration-dependent artefacts.

346 Winter

- 347 During the cold parts of the year the comparisons between the MOHp sondes and the lidar usually exhibit better 348 quality. This is explained by less structure in the ozone vertical distributions and a wider operating range of the 349 lidar due to the low ozone level allowing for a higher, less noisy far-field signal. We found just one example 350 with some deviating structures of the order of ± 10 ppb (10 January 2018). For the 2018 comparison we give one 351 example in Fig. 4 (15 January). The lidar mixing ratio is of the order of 45 ppb, verified by the measurements at 352 UFS (2660 m a.s.l.). The Brewer Mast ozone sonde shows a negative bias of 5.8 ppb relative to the lidar above 353 2.1 km. After removing this bias (cyan curve) the sonde ozone matches the lidar and the UFS values well for 354 altitudes above 2.1 km. This performance almost reaches that in the examples of Sect. 3.1. Just below the 355 tropopause there is a minor discrepancy that could be either due to the higher uncertainty of the lidar 356 measurement at these altitudes or an air-mass difference.
- For the other 2018 winter-time comparisons the constant offsets of the sondes with respect to the lidar and UFS are just +1.5 ppb to -3.0 ppb.

359 Summer

360 During the warm season the ozone distribution in the middle and upper troposphere shows structured maxima 361 caused by long-range transport, in particular STT (stratosphere-to-tropopause transport) layers (Trickl et al.,

- 362 2020b). In this altitude range a summer maximum of STT exists. Usually, these structures do not perfectly match
- 363 for both sites. An example for 9 July 2018 is shown in Fig. 5.

Figure 5 shows good agreement in structure between the soundings at both sites up to 9 km. Again, despite the pronounced ozone layering, the agreement was improved on the absolute scale by adding an altitude-independent correction to the sonde values (6 ppb). The offset is usually determined up to 6 km due to the reliable performance of the 277-nm-313-nm DIAL measurements, but the agreement is mostly reasonable also to higher altitudes. After shifting the sonde mixing ratio we can estimate the uncertainty of the lidar measurements.

The elevated ozone in Fig. 5 between 3.3 km and 4.7 km can be explained by a stratospheric air intrusion, as is verified by the low RH. In the upper troposphere the agreement deteriorates, but at least the increase of ozone with altitude is seen in all profiles up to about 12 km. The ozone minimum around 13 km is just seen in the lidar data, with just a small ozone dip in the sonde profile. It is unreasonable to ascribe this considerable discrepancy to a temporary technical problem in such a limited altitude range. This example documents the difficulty of

- 374 quantitative comparisons of tropospheric ozone even on a horizontal scale of just 38 km.
- 375 In order to clarify the origin of the difference of the ozone mixing ratio in the upper troposphere we calculated

376 backward trajectories with the HYSPLIT model (http://ready.arl.noaa.gov/HYSPLIT.php; Draxler and Hess,

377 1998; Stein et al., 2015). These trajectories reveal northerly advection which implies a southward drift of the
378 sonde towards the lidar during the ascent. In the upper troposphere they did not fully explain the observations
379 within the limited maximum backward time span of 315 h for the few start altitudes selected.

- 380 Therefore, the trajectory calculations were extended to 350 h by using the LAGRANTO model for full-hour start 381 times between 3:00 CET and 8:00 CET, initiated at a large number of altitudes in the low-ozone range in the 382 upper troposphere. Results for start times of 7:00 CET and 8:00 CET are shown in Figs. 6 and 7. Up to a start
- 383 time of 4:00 CET the trajectories stayed almost completely at high altitudes. At 5:00 CET three of the 384 trajectories ended in the lower troposphere above the subtropical Pacific near a longitude of 180°, first sign of an 385 air-mass change. Later (Figs. 6 and 7) we see a clear influence of a Pacific source.
- The low ozone level in the boundary layer above (sub)tropical oceans is well known (Eisele et al., 1999; Grant et
 al., 2000; Trickl et al., 2003; 2010), in particular over the Pacific (Kley et al., 1996; Davies et al., 1998). In this
 way, the lidar observations on 9 July 2018 can be understood. The launch time of the MOHp ozone sonde, 5:42
- 389 CET, is between the two lidar measurements. However, a delay is caused during the ascent which makes a
- 390 quantitative understanding difficult.
- The moderate sonde RH above 12.3 km indicates a potential admixture of aged stratospheric air in this altitude range above MOHp, which would explain the high ozone mixing ratios of more than 120 ppb.

393 Figures 5 and 8 show a rather constant negative ozone offset of the sonde profiles. The ozone profiles can be

brought into much better agreement with the lidar and UFS by upward shifts of 6 ppb and 10 ppb, respectively.

395 In Fig. 9 one sees one of the very rare cases of an ozone mismatch between sonde and lidar up to elevations

- 396 clearly above the mountain sites (1 km above the Zugspitze summit). We did not shift the MOHp profile (e.g.,
- 397 by 3 ppb) to reduce the mismatch since this would deteriorate the agreement above 4 km.

398 *Offsets*

- 399 The offsets of the MOHp data from the DIAL profiles were evaluated for all 36 comparison days. The result of
- 400 the statistical assessment is displayed in Fig. 10 where also the differences between the lidar results for 2671 m
- 401 a.s.l. and the GAW measurements at UFS are shown. Just one case was eliminated in the comparison of lidar and
- 402 UFS: A strong negative shift of -7 ppb can be seen in Fig. 5 where UFS is located in the falling edge of a high-
- 403 ozone range.

- 404 As found for the lidar measurements over many years (examples: Trickl et al., 2014, 2015, 2016, 2020b) the
- 405 lidar ozone agrees with that at UFS to within ± 3 ppb (mostly ± 2 ppb). The agreement would perhaps be better if
- 406 orographic vertical displacements and air flows on the ozone profiles would be considered (Carnuth et al., 2000;
- 407 2002; Yuan et al., 2019; Trickl et al., 2020a). The average difference between lidar and UFS for 2018 (blue
- 408 horizontal line in Fig. 10) is 0.736 ppb \pm 1.46 ppb (standard deviation). A positive offset had also been found for
- 409 an earlier four-day comparison with the Zugspitze summit, but with even higher uncertainty (Trickl et al., 410 2020a). A positive offset of this size could be expected from the highly accurate cross-section measurements of
- 411 Viallon et al. (2015), who determined a negative bias of 1.8 % of in-situ ozone data calibrated with the WMO
- 412 standard. This relative difference becomes more important on the absolute scale in summer than in winter
- 413 because of the higher ozone values. However, the statistical noise of the differences is too high to allow 414 resolving such an effect.
- 415 The offsets between the MOHp sonde and the lidar, again preferentially determined in the range up to 6 km, are
- 416 substantially higher than those between the lidar and UFS (red filled squares in Fig. 10). The offsets of the ozone
- 417 sondes range from -12 ppb to +4 ppb, with an average of -3.77 ppb (red horizontal line in Fig. 10) and a
- 418 standard deviation of 4.22 ppb.
- 419 We exclude the lowest altitudes from the comparison where obvious differences in ozone exist, e.g., due to local
- 420 night-time ozone depletion effects. It is important to note that just in seven cases of the 36 comparisons for 2018
- 421 lower ozone in the sonde profiles reached up to more than 2.67 km (UFS), in three cases to more than 3 km
- 422 (Zugspitze summit). We conclude that differences between the Zugspitze sites and the MOHp sonde are mostly
- 423 related to sonde calibration issues and not to differences in air-composition as suspected earlier.

424 Differences

- In order to determine the quality of the lidar measurements within the free troposphere we show in the three panels of Fig. 11 average differences between lidar and offset-corrected MOHp sonde data as a function of altitude and for three different ozone conditions, roughly below 50 ppb (low ozone; top panel), between 50 to 70 ppb (moderate ozone; second panel) and more than 70 ppb (high ozone; bottom panel), respectively. On a given day, the lidar ozone profiles agreeing best with the MOHp profile was taken. We also give the percentages of the averages with respect to the offset-corrected sonde ozone. At high altitudes the sonde ozone is a more useful reference than the lidar in the case of high ozone because of the considerable absolute uncertainty caused by the
- 432 loss of laser radiation by absorption in ozone.
- For winter-type conditions (top panel) the six examples averaged do not exhibit a significant vertical ozone structure which made the analysis straight forward and yields small average differences between ± 1 ppb and ± 3 ppb, in agreement with the conclusions in Sect. 3.1. For moderate ozone (second panel) and high ozone (bottom panel), mostly during the warm season, the vertical distributions are more complex with changes on a time scale
- 437 of even less than one hour. Here, we eliminated the data for a few pronounced ozone peaks and dips that differed
- 438 at both stations. The six high-ozone cases were restricted to July and August.
- 439 The averaged distributions of the differences exhibit oscillations. These oscillations were analysed for coherency
- 440 (not shown), but no systematic behaviour was identified. Thus, we ascribe the structure to noise. The noise
- 441 contains both an atmospheric and an instrumental component.
- 442 Beyond the days and years of the comparison there are occasionally extreme cases with 100 to 150 ppb in the
- 443 middle and upper troposphere. This can lead to lidar uncertainties even up to more than 20 ppb during daytime,
- 444 also because the raw signal becomes comparable with the additional solar background noise. In the most severe

- 445 cases the stratospheric ozone rise cannot be seen in the lidar data during daytime, and the ozone profile is cut off
- 446 in the upper troposphere for archiving.
- 447 The analyses for 2018 do not reveal significant systematic differences between the lidar values and the offset-
- 448 corrected sonde data in the entire free troposphere (based on the numbers underlying Fig. 10). This confirms the
- 449 conclusion in Sect. 3.1 for the quality of the lidar, now for all seasons. The maximum noise excursions can be
- 450 interpreted as maximum combined uncertainties of lidar and corrected sonde in a given altitude range (slightly
- 451 overestimated due atmospheric differences in ozone between both sites). The results of this analysis confirm the
- 452 estimates in Table 4 of Trickl et al. (2020a).

453 **3.3** Comparisons of MOHp sonde, IFU lidar and in-situ measurements at summits in 2009

- 454 The results in Sect. 3.2 suggested to look also at a few earlier years. We select 2009 from the period of routine
- 455 measurements as another year of comparison. The lidar raw data were noisier than for the period after 2012 and
- 456 a tiny electronic ringing effect had to be removed mathematically. Thus, the uncertainties of the ozone profiles
- 457 above 6 km are higher than after the final system upgrading in 2012, particularly in summer. As a consequence, 458 a lidar validation is desirable at least for the upper troposphere. More importantly, in 2009 high-quality ozone
- 458 a lidar validation is desirable at least for the upper troposphere. More importantly, in 2009 high-quality ozone
- data still exist for the summit stations Wank (1780 m a.s.l.) and Zugspitze (2962 m a.s.l). These stations benefit
- 460 from more frequent direct advection compared with UFS.
- In 2009 the lidar was operated just until October which, nevertheless, allows us to make a reasonable number of comparisons with MOHp. The lidar operation was stopped afterwards since there were more and more cases of single-bit errors in channel 5 of the transient digitizer system which had to be sent for repair. These errors induced unrealistic data in the upper troposphere.
- We identified a total of 23 days suitable for comparisons. On just eight of these days lidar measurements were made in optimum temporal proximity. We find more deviations in the profiles than for 2018. In part, this can be explained by atmospheric variability and insufficient air-mass matching. In addition, as mentioned, the raw data of the lidar are noisier and some weak ringing had to be removed. This caused elevated uncertainties above 6 km. Nevertheless, the data allowed us to determine offsets for the MOHp ozone profiles, after verifying the data
- 470 quality of the lidar with the Zugspitze and Wank in-situ ozone.
- 471 In Fig. 12 we show the results of the analysis for 2009. The difference between IFU DIAL and Zugspitze is
- $472 \qquad -0.165 \text{ ppb} \pm 1.36 \text{ ppb} \text{ (standard deviation), between DIAL and Wank +0.714 ppb} \pm 1.20 \text{ ppb}. \text{ The DIAL ozone}$
- 473 below the Wank altitude is increasingly uncertain because of alignment issues of the near-field telescope.
- In an earlier comparison for May 1999 (Trickl et al., 2020a) we selected a lower altitude in the DIAL data (2786
 m) and found better agreement with the Zugspitze data, but, still, a slight positive offset with respect to the
 station. This is not attempted here, although we can see the effect of orographic lifting in some examples.
- 477 For 2009 the offsets between DIAL and MOHp sondes were determined primarily by between 2 and 5 km. The
- sonde offset obtained in this way is, again, negative on average (-1.500 ppb), with a standard deviation of 2.67
 ppb, both being are less pronounced than in 2018.
- 480 Figure 13 shows a comparison on 12 January 2009, demonstrating excellent agreement between both systems
- 481 after offset correction, except for the upper troposphere and lower stratosphere. In this case, the first lidar
- 482 measurement took place at 9:20 CET, i.e., substantially later than the sonde ascent. Thus, the comparison has its
- 483 limits. In the morning of 12 January westerly advection was revealed by HYSPLIT backward trajectories above
- 484 at 7 km a.s.l. This air mass originated below 2 km over the subtropical Atlantic. This could explain the slightly
- 485 lower ozone level around this altitude in the lidar results.

- 486 Another interesting example is August 17 (Fig. 14). The agreement between lidar and ozone sonde is highly
- 487 satisfactory up to 5.4 km and quite reasonable up to 10 km. However, between 10 km and 14.5 km the lidar
- 488 ozone is extremely low, in contrast to the sonde data. The pronounced ozone increase in the sonde data above 10
- 489 km is difficult to explain since the elevated RH values suggest neither a low tropopause nor the presence of a
- 490 stratospheric intrusion that typically features RH values of a few per cent at most (Trickl et al., 2014; 2015;
- 491 2016). On the other hand, the ozone peak above IMK-IFU descending roughly from 10 to 8 km is attributed by
- 492 HYSPLIT calculations to subsiding air, indicating the presence of an intrusion layer. It is interesting that the
- 493 rather short delay of the lidar measurements (7:00 CET to 9:15 CET) with respect to the sonde ascent (launch
- 494 time 5:57 CET) can result in such a considerable difference.
- 495 Again, 350-h LAGRANTO trajectories were calculated for start times above IMK-IFU between 3:00 CET and
- 496 8:00 CET (interval: 1 h) and start altitudes within the low-ozone layer. Until 6:00 CET the influence of marine
- 497 boundary layers is almost absent. Afterwards, the trajectories reveal a growing import from the first 600 m above
- 498 the subtropical Atlantic Ocean. In Fig. 15 the LAGRANTO results for 8:00 CET are shown.
- 499 In some cases the lidar seems to exhibit a negative bias with respect to the sondes in the upper troposphere. It is
- 500 advisable to re-examine a major part of the data between 2007 and 2011, also including strategies developed
- 501 later. For example, an exponential decay of the analogue signal was identified with the much lower noise of the
- 502 final setup (Trickl et al., 2020a) which must be addressed.

503 3.4 Comparisons of MOHp sonde, IFU lidar and in-situ measurements summits in 2000 and 2001

- The period September 2000 to August 2001 is suitable for another comparison when a large number of STTrelated measurement series were made as a contribution to the STACCATO project (Stohl et al., 2003; examples: Trickl et al., 2003; 2010; 2011; Zanis et al., 2003). These measurements were made with the noisier detection electronics of Eisele et al. (1999), but had the advantage that single-photon counting was used for the "solar blind" "on" detection channels which added linearity above 5 km (starting in spring 1997). The counting system could no longer be computer controlled after 2006. A new one was installed after highly positive results in other IFU lidar systems (Klanner et al., 2021) in autumn 2018, too late for this comparison effort.
- 511 The focus on STT during the STACCATO period made the comparisons a challenge because of the pronounced
- 512 layering. However, on 11 of the useful 20 days of comparison there was reasonable temporal proximity, due to
- 513 running long time series. The agreement between the lidar and the MOHp sonde was much better than expected
- 514 in the entire free troposphere. The agreement (after offset-correcting the MOHp profiles) is almost perfect during
- 515 the cold season. But also under high-ozone conditions the comparisons do not reveal systematic differences
- 516 beyond the sonde offsets.
- 517 Two examples for elevated ozone are shown in Figs. 16 and 17. The good comparisons support our earlier work
- 518 (Trickl et al., 2003, 2010), and we tend to ascribe this to the satisfactory performance of the single-photon 519 counting system.
- 520 For several weeks a strange ozone rise towards the ground was observed in the lidar data below 1.5 km. This
- 521 effect disappeared after realigning the near-field telescope and the normal early-morning ozone drop returned.
- 522 However, the offsets of the MOHp mixing ratios necessary to achieve good agreement are, again, quite
- 523 substantial (Fig. 18). Due to the larger system noise during that period also the differences between lidar and the
- 524 stations are higher than those in the preceding sections, and comparable with those of the mentioned four-day
- 525 comparison for May 1999 (Trickl et al., 2020a). The statistical analysis yields the following average differences
- and standard deviations:

527	IFU DIAL – Zugspitze:	$1.22 \text{ ppb} \pm 1.81 \text{ ppb}$
528	IFU DIAL – Wank	-0.15 ppb \pm 2.26 ppb

529 MOHp – IFU DIAL –5.88 ppb ± 3.35 ppb

530 4. Discussion and Conclusions

For some time tropospheric differential-absorption ozone lidar systems had a bad reputation: The method is highly sensitive to imperfections in the signal acquisition since the ozone number density is obtained by differentiating the backscatter signals (Trickl et al., 2020a). In addition, a lidar covering the entire troposphere and the lowermost stratosphere features a dynamic range of the backscatter signal of about eight decades, which means an extreme challenge for the detection electronics.

Based on continual improvements, starting with the 1994 system upgrading, the IFU ozone DIAL gradually approached a high performance until 2012, but minor potential for improvements remains. Comparison with the nearby mountain stations quite early demonstrated an uncertainty level of ± 3 ppb in the lower troposphere. Occasional comparisons with ozone sondes launched at the Hohenpeißenberg (1996 to 2001, distance 38 km)

540 were rather satisfactory up to the tropopause region.

541 Here, we analyse the lidar performance in three periods during its technical development in a more 542 comprehensive manner. The best agreement was found for the side-by-side comparison with balloon ascents of 543 ECC ozone sondes performed by the FZJ team at IMK-IFU in February 2019. Just small, altitude-independent 544 offsets had to be subtracted from the sonde data to achieve agreement. The lidar itself agreed with the three local 545 summit stations. For all three years and all stations we determined a positive bias of the lidar of just 0.6 ppb \pm 546 0.6 ppb (standard deviation). This value seems to reflect the -1.8-% calibration deficit of the WMO calibration of 547 the in-situ ozone data. Thus, the lidar could be even free of bias in the lower free troposphere, reflecting the high 548 quality of the calibration source (Sect. 2.4).

549 For the more distant MOHp sonde the comparisons are more complex because of the high atmospheric 550 variability (Vogelmann et al., 2011; 2015). This variability is particularly severe in summer when the 551 atmospheric layering is more pronounced. Nevertheless, there was enough agreement in certain altitude ranges 552 for examining the reliability of the ozone profiles obtained from the DIAL, also before the final modifications in 553 2012. Between 2007 and 2011 we suspect an occasional slight negative summertime bias of the lidar of the order 554 of 5 ppb above 6 km. This could be due to interfering structures on the 292-nm analogue signal (requiring 555 mathematical correction) that could not be compensated by photon counting (available until 2003) and the 556 removal of daylight-induced signal distortions at 313 nm (Trickl et al., 2020a). In principle, this calls for a re-557 evaluation of the ozone profiles for the wavelength pair 292 nm - 313 nm over the period 2007 to 2011, based on 558 more recent experience in the signal inversion and the performance of the electronic equipment.

559 Vice versa, the lidar measurements helped us to validate the quality of the sonde measurements. Quite good 560 agreement could be achieved by applying an altitude-independent offset correction to the ozone values that 561 strongly varies from sonde to sonde. Most of the ozone differences the two sites are limited to altitudes below 2 562 km. Thus, the differences between Zugspitze and MOHp (at 3 km) reported earlier by Scheel (see Introduction) 563 are not caused by systematic differences in air composition at both sites. As can be seen from the figures 564 presented in this paper the shifted ozone mixing ratios for the sonde and the Zugspitze ozone mostly agrees to 565 within ± 3 ppb. Given the frequently substantially higher ozone offsets of the MOHp sondes a recalibration of the 566 archived sonde data based on comparisons with the Zugspitze or UFS in-situ data is advisable despite the

- 567 considerable distance between the sites. Such a recalibration should be avoided in the presence of pronounced
- 568 ozone structure around the station altitudes which could account for by elevated uncertainties.
- 569 The comparisons for the three years 2000-2001, 2009 and 2018 reveal just minor performance change of the
- 570 MOHp sonde over the years, with a variation of the annual average offset by about ± 2 ppb. We found a negative
- 571 average offset of -3.64 ppb ± 3.72 ppb (standard deviation) with respect to the IFU ozone DIAL over all three 572 years. It is reasonable to assume that this offset is applicable to the entire tropospheric time series of the MOHp
- 573 sondes. This performance is within the uncertainty range of the literature cited in the Introduction.
- 574 Remaining tasks for the lidar are a substantial reduction of the solar background at 313.2 nm in summer and to 575 enhance the moderate 291.8-nm backscatter signal in the upper troposphere. Further reduction of the residual 576 solar background is difficult since the spectral filtering is already quite narrow. However, replacement of the 577 rather aged (and partly contaminated) primary mirror of the far-field receiver could help by reducing the 578 background radiation reflected into the detection system. As mentioned longer averaging is advisable. By longer 579 averaging, the performance under low-aerosol conditions could almost reach that of in-situ measurements in a 580 major part of the troposphere. Single-photon counting can also be helpful for longer averaging times, as 581 demonstrated for our Raman lidar (Klanner et al., 2021). The noise level for counting is still lower than that of 582 the meanwhile outstanding transient digitizers (Trickl et al., 2020a).

583 5 Data availability

584 Lidar data and information on the lidar systems can be obtained on request from the IMK-IFU authors of this 585 paper (thomas@trickl.de, hannes.vogelmann@kit.edu). The 313-nm aerosol backscatter coefficients are archived 586 in the EARLINET data base, accessible through the ACTRIS data portal http://actris.nilu.no/. The 587 Hohenpeißenberg ozone and humidity data are stored in the NDACC data archive (https://www-588 air.larc.nasa.gov/missions/ndacc/data.html#). The data of the FIRMOS campaign is available via the ESA 589 campaign dataset website https://earth.esa.int/eogateway/campaigns/firmos. The hourly Zugspitze and UFS 590 ozone data are available at the World Data Center for Reactive Gases (WDCRG: https://ebas.nilu.no/) and the 591 TOAR data base (Schultz et al., 2017).

592 6 Author statement

- 593 TT carried out most lidar measurements after spring 1997, following U. Kempfer and H. Eisele. He led the 594 technical development of two ozone DIAL systems since 1990. HV was involved in the system upgrading since 595 2007 and was responsible for the lidar operation during FIRMOS. DC and CW launched several ECC sondes at
- 596 IMK-IFU in February 2019. MA and WS carried out the MOHp ozone sonde measurements. LR performed
- 597 ozone in-situ measurements at UFS. MS provided LAGRANTO backward trajectories.
- 598 7 Competing interests
- 599 The authors declare that they have no conflict of interest.

600 Acknowledgements

The authors thank Wolfgang Seiler and Hans Peter Schmid for their support over so many years. The late Hans Eckhart Scheel provided reference ozone data for the Wank, Zugspitze mountain stations in the vicinity of IMK-

- 603 IFU The different steps of lidar development have been funded by the German Ministry of Research and
- 604 Technology (BMFT), the German Foundation for the Environment (DBU, two projects), and the Bavarian

- 605 Ministry of Economics. Since 2007 the 313-nm aerosol results have contributed to EARLINET (European
- 606 Aerosol Research Lidar Network) that is currently a part of the European Research Infrastructure ACTRIS
- 607 (Aerosol, Clouds and Trace Gases Research Infrastructure). The lidar measurements were also funded by the
- 608 European Union within Vertical Ozone Transport 1 and 2 (e.g., Wotava, G., and Kromp-Kolb, 2000; VOTALP
- 609 2, 2000) and STACCATO (Stohl et al., 2003) and by the German Ministry for Research and Education (BMBF)

610 within ATMOFAST (2005).

- KIT acknowledges support of lidar measurements by the European Space Agency (ESA) under Contract 4000123691/18/NL/NF (FIRMOS validation campaign). Balloon profiles utilized in this paper have been provided within the same ESA project by the Forschungszentrum Jülich via subcontract with KIT. The balloon activities were also partly supported by the Helmholtz Association in the framework of MOSES (Modular Observation Solutions for Earth Systems).
- 616 The service charges for this open access publication have been covered by a Research Centre of the Helmholtz617 Association.

618 References

620

619 ATMOFAST: Atmosphärischer Ferntransport und seine Auswirkungen auf die Spurengaskonzentrationen in der

freien Troposphäre über Mitteleuropa (Atmospheric Long-range Transport and its Impact on the Trace-gas

- 621 Composition of the Free Troposphere over Central Europe), Project Final Report, T. Trickl, co-ordinator, M.
- 622 Kerschgens, A. Stohl, and T. Trickl, subproject co-ordinators, funded by the German Ministry of Education and
- 623 Research within the programme "Atmosphärenforschung 2000", http://www.trickl.de/ATMOFAST.htm, 130
- 624 pp., 2005 (in German), with revised publication list of 2012
- Ancellet, G., Pelon, J., Beekmann, M., Papayannis, A., and Mégie, G.: Ground-Based Lidar Studies of Ozone
 Exchanges Between the Stratosphere and the Troposphere, J. Geophys. Res., 96, 22401-22421, 1991.
- 627 Ancellet, G., Godin-Beekmann, S., Smit, H. G. J., Stauffer, R. M., Van Malderen, R., Bodichon, R., and
- 628 Pazmiño, A.: Homogenization of the Observatoire de Haute Provence electrochemical concentration cell (ECC)
- 629 ozonesonde data record: comparison with lidar and satellite observations, Atmos. Meas. Tech., 15, 3105–3120,
- 630 2022.
- 631 Attmannspacher, W., and Dütsch, H.: 2nd International Ozone Sonde Intercomparison at the Observatory of
- 632 Hohenpeissenberg, Berichte des Deutschen Wetterdienstes 157, 1981.
- 633 Beekmann, M., Ancellet, G., Mégie, G., Snit, H. G. J., and Kley, D.: Intercomparison Campaign of Vertical
- 634 Ozone Profiles Including Electrochemical Sondes of ECC and Brewer-Mast Type and a Ground Based UV-
- 635 Differential Absorption Lidar, J. Atmos. Chem., 19, 259-288, 1994.
- 636 Belotti, C., Barbara, F., Barucci, M., Bianchini, G., D'Amato, F., Del Bianco, S., Di Natale, G., Gai, M.,
- 637 Montori, A.,, Pratesi., F., Rolf, C., Sussmann, R., Trickl, T., Viciani, S., Vogelmann, H., Palchetti, L.: The Far-
- 638 Infrared Radiation Mobile Observation System for spectral characterisation of the atmospheric emission, Atmos.
- 639 Meas. Tech., 16, 2511–2529, https://doi.org/10.5194/amt-16-2511-2023, 2023.
- 640 Brabec, M.: Backscatter and Humidity Measurements in Cirrus and Dust Clouds using Balloon Sondes, Ph.D.
- thesis, Eidgenössische Technische Hochschule, Zürich (Switzerland), 96 pp., 2011.
- 642 Brewer, A. W., and Milford, J. R.: The Oxford-Kew ozone sonde, Proc. R. Soc. Lond. A, 256, 470-495
- 643 [available at http://doi.org/10.1098/rspa.1960.0120], 1960.

- Browell, E. V., Danielsen, E. F., Ismail, S., Gregory, G. L., and Beck, S. M.: Tropopause Fold Structure
 Determined From Airborne Lidar and in Situ Measurements, J. Geophys. Res., 92, 2112-2120, 1987.
- 646 Carnuth, W., and Trickl, T.: Transport studies with the IFU three-wavelength aerosol lidar during the VOTALP
- 647 Mesolcina experiment, Atmos. Environ., 34, 1425-1434, 2000.
- 648 Carnuth, W., Kempfer, U., and Trickl, T.: Highlights of the tropospheric lidar studies at IFU within the TOR
- 649 project, Tellus B, 54, 163-185, 2002.
- 650 Claude, H., Hartmannsgruber, R., and Köhler, U.: Measurement of atmospheric profiles using the Brewer-Mast
- 651 sonde, World Meteorological Organization, Global Ozone Res. and Monit. Proj. Report No. 17, WMO/TD No.
- 652 179, Geneva (Switzerland), [see also https://library.wmo.int/index.php?lvl=notice_display&id=11215], 51 pp.,
 653 1987.
- 654 Deshler., T., Stübi, R., Schmidlin, F. J., Mercer, J. L., Smit, H. G. J., Johnson, B. J., Kivi, K., and Nardi, B.:
- 655 Methods to homogenize electrochemical concentration cell (ECC) ozonesonde measurements across changes in
- sensing solution concentration or ozonesonde manufacturer, Atmos. Meas. Tech., 10, 2021–2043, 2017.
- 657 Daumont, D., Brion, J., Charbonnier, J., and Malicet, J.: Ozone UV Spectroscopy I: Absorption Cross-Sections
- 658 at Room Temperature, J. Atmos. Chem., 15, 145-155, 1992.
- Davies, W. E., Vaughan, G., and O'Connor, F. M.: Observation of near-zero ozone concentrations in the upper
 troposphere at mid-latitudes, Geophys. Res. Lett., 25, 1173-1176, 1998.
- De Backer, H., De Muer, D., and De Sadelaer G.: Comparison of ozone profiles obtained with Brewer-Mast and
 Z-ECC sensors during simultaneous ascents, J. Geophys. Res., 103, 19,641–19,648,
 https://doi.org/10.1029/98JD01711, 1998.
- De Muer, D., and Malcorps, H.: The frequency response of an electrochemical ozone sonde and its application to
 the deconvolution of ozone profiles, J. Geophys. Res., 89, 1361–1372, 1984.
- bi Natale, G., Barucci, M., Belotti, C., Bianchini, G., D'Amato, F., Del Bianco, S. Gai, M., Montori, A.,
- 667 Sussmann, R., Viciani, S., Vogelmann, H., and Palchetti, L.: Comparison of mid-latitude single- and mixed-
- 668 phase cloud optical depth from co-located infrared spectrometer and backscatter lidar measurements, Atmos.
- 669 Meas. Tech., 14, 6749–6758, 2021.
- 670 Dirksen, R. J., Sommer, M., Immler, F. J., Hurst, D. F., Kivi, R., and Vömel, H.: Reference quality upper-air
- 671 measurements: GRUAN data processing for the Vaisala RS92 radiosonde, Atmos. Meas. Tech., 7, 4463-4490,
- 672 doi.org/10.5194/amt-7-4463-2014, 2014.
- Draxler, R., and Hess, G.: An overview of the HYSPLIT_4 modelling system for trajectories, dispersion, anddeposition, Aust. Meteorol. Mag., 47, pp. 295-308, 1998.
- Eisele, H., Trickl, T., and Claude, H.: Lidar als wichtige Ergänzung zur Messung troposphärischen Ozons,
 Ozonbulletin des Deutsches Wetterdiensts, 44, 2 pp., 1997 (in German).
- 677 Eisele, H., Scheel, H. E., Sladkovic, R., and Trickl, T.: High-Resolution Lidar Measurements of Stratosphere-
- 678 Troposphere Exchange, J. Atmos. Sci., 56, 319-330, 1999.
- 679 Eisele, H., and Trickl, T.: Improvements of the aerosol algorithm in ozone-lidar data processing by use of
- 680 evolutionary strategies, Appl. Opt., 44, 2638-2651, 2005.

- 681 EUROTRAC: Transport and Chemical Transformation of Pollutants in the Troposphere, Vol. 1, An Overview of
- 682 the Work of EUROTRAC, P. Borrell and P. M. Borrell, Eds., Springer (Berlin, Heidelberg, New York), ISBN 3-
- 683 540-66775-X, 474 pp., 1997.
- 684 Gaudel, A., Ancellet G., and Godin-Beekmann S.: Analysis of 20 years of tropospheric ozone vertical profiles by
- 685 lidar and ECC at Observatoire de Haute Provence (OHP) at 44° N, 6.7° E, Atmos. Environ., 113, 78-89, 2015.
- 686 Gaudel, A., Cooper, O. R., Ancellet, G., Barret, B., Boynard, A., Burrows, J. P., Clerbaux, J. P., Coheur, P.-F.,
- 687 Cuesta, J., Cuevas, E., Doniki, S., Dufour, G., Ebojie, F., Foret, G., Garcia, O., Granados-Muñoz, M. J.,
- Hannigan, J., Hase, F., Hassler, B., Huang, G., Hurtmans, D., Jaffe, D., Jones, N., Kalabokas, P., Kerridge, B.,
- Kulawik, S., Latter, B., Leblanc, T., Le Flochmoën, E., Lin, W., Liu, J., Liu, X., Mahieu, E., McClure-Begley,
- 690 A., Neu, J., Osman, M., Palm, M., Petetin, H., Petropavlovskikh, I., Querel, R., Rahpoe, N., Rozanov, A.,
- 691 Schultz, M. G., Schwab, J., Siddans, R., Smale, D., Steinbacher, M., Tanimoto, H., Tarasick, D., Thouret, V.,
- 692 Thompson, A., M., Trickl, T., Weatherhead, E., Wespes, C., Worden, H., Vigouroux, C., Xu, X., Zeng, G.,
- 693 Ziemke, J.: Tropospheric Ozone Assessment Report: Present-day distribution and trends of tropospheric ozone
- 694 relevant to climate and global atmospheric chemistry model evaluation, Elem. Sci. Anth., 6, 39, DOI:
- 695 https://doi.org/ 10.1525/elementa.291, 58 pp., 2018.
- 696 Grant, W. B., Browell, E. V., Butler, C. F., Fenn, M. A., Clayton, M. B., Hannan, J. R., Fuelberg, H. E., Blake,
- D. R., Blake, N. J., Gregory, G. L., Heikes, B. G., Sachse, G. W., Singh, H. B., Snow, J., and Talbot, R. W.: A
- 698 case study of transport of tropical marine boundary layer and lower tropospheric air masses to the northern
- 699 midlatitude upper troposphere, J. Geophys. Res., 105, 3757-3769, 2000.
- Hearn, A. G.: The Absorption of Ozone in the Ultra-violet and Visible Regions of the Spectrum, Proc. Phys.Soc., 78, 932-940, 1961.
- Hersbach, H, Bell, B, Berrisford, P, Simmons, A., Berrisford, P., Dahlgren, P., Horanyi, A., Munoz-Sabater, J.,
 Nicolas, J., Radu, R., Schepers, D., Soci, C., Villaume, S., Bidlot, J. R., Haimberger, L.; Woollen, J.,
 Buontempo, C., and Thepaut, J. N.: The ERA5 global reanalysis. Q. J. R. Meteorol Soc., 146, 1999–2049,
 https://doi.org/10.1002/qj.3803, 2020.
- Johnson, B. J., Oltmans, S. J., Vömel, H., Smit, H. G. J., Deshler, T., and Kröger, C.: Electrochemical
 concentration cell (ECC) ozonesonde pump efficiency measurements and tests on the sensitivity to ozone of
 buffered and unbuffered ECC sensor cathode solutions, J. Geophys. Res.-Atmos., 107, ACH 8-1–ACH 8-18,
 https://doi.org/10.1029/2001JD000557, 2002.
- 710 Jeannet, P., Stübi, R., Levrat, G., Viatte, P., and J. Staehelin, J.: Ozone balloon soundings at Payerne
- (Switzerland): Reevaluation of the time series 1967–2002 and trend analysis, J. Geophys. Res., 112, D11302,
 doi:10.1029/2005JD006862, 15 pp., 2007.
- Kempfer, U., Carnuth, W., Lotz, R., and Trickl, T.: A wide range ultraviolet lidar system for tropospheric ozone
 measurements: development and application, Rev. Sci. Instrum., 65, 3145-3164, 1994.
- 715 Kerr, J. B., Fast, H., McElroy, C.T., Oltmans, S.J., Lathrop, J.A., Kyro, E., Paukkunen, A., Claude, H., Köhler,
- 716 U., Sreedharan, C.R., akao T., and Tsukagoshi, Y.: The 1991 WMO International Ozonesonde Intercomparison
- 717 at Vanskoy, Canada. Atmos.–Ocean, 32, 685–716, https://doi.org/10.1080/07055900.1994.9649518, 1994.

- 718 Klanner, L., Höveler, K., Khordakova, D., Perfahl, M., Rolf, C., Trickl, T., and Vogelmann, H.: A powerful lidar
- system capable of 1 h measurements of water vapour in the troposphere and the lower stratosphere as well as the
- temperature in the upper stratosphere and mesosphere, Atmos. Meas. Tech., 14, 531-555, 2021.
- 721 Klausen, J., Zellweger, C., Buchmann, B., and Hofer, P.: Uncertainty and bias of surface ozone measurements at
- selected Global Atmospheric Watch sites, J. Geophys. Res., 108, 4622, doi: 10.1029/2003JD003710, 17 pp.,
 2003.
- Kley, D., Beck, J., Grennfelt, P. I., Hov, O., and Penkett, S. A.: Tropospheric Ozone Research (TOR) A Sub-
- 725 Project of EUROTRAC, J. Atmos. Chem., 28, 1–9, 1997.
- 726 Kley, D., Crutzen, P. J., Smit, H. G. J., Vömel, H., Oltmans, S., Grassl, H., and Ramanathan, V.: Observations of
- 727 Near-Zero Ozone Concentrations Over the Convective Pacific: Effects on Air Chemistry, Science, 274, 230-233,
- 728 1996.
- Komhyr, W.D.: Electrochemical concentration cells for gas analysis, Ann. Geoph., 25, 203–210, 1969.
- 730 Komhyr, W. D., Barnes, R. A., Brothers, G. B., Lathrop, J. A., and Opperman, D. P.: Electrochemical
- 731 concentration cell ozonesonde performance evaluation during STOIC 1989, J. Geophys. Res., 100, 9231–9244,
- 732 https://doi.org/10.1029/94JD02175, 1995.
- 733 Langford, A. O., Masters, C. D., Proffitt, M. H., Hsie, E.-Y., and Tuck, A. F.: Ozone measurements in a
- tropopause fold associated with a cut-off low system, Geophys. Res. Lett., 23, 2501–2504, 1996.
- 735 Logan, J.A., Staehelin, J., Megretskaia, I.A., Cammas, J.-P., Thouret, V., Claude, H., De Backer, H.,
- 736 Steinbacher, M., Scheel, H.-E., Stübi, R., Fröhlich, M., and Derwent, R. (2012), Changes in ozone over Europe:
- 737 Analysis of ozone measurements from sondes, regular aircraft (MOZAIC) and alpine surface sites, J. Geophys.
- 738 Res., 117, D09301, doi:10.1029/2011JD016952.
- Malicet, J., Daumont, D., Charbonnier, J., Parisse, C., Chakir, A., and Brion, J.: Ozone UV Spectroscopy I:
 Absorption Cross-Sections and Temperature Dependence, J. Atmos. Chem., 21, 263-273, 1995.
- 741 Palchetti, L., Barucci, M., Belotti, C., Bianchini, G., Cluzet, B., D'Amato, F., Del Bianco, S., Di Natale, G., Gai,
- 742 M., Khordakova, D., Montori, A., Oetjen, H., Rettinger, M., Rolf, C., Schuettemeyer, D., Sussmann, R., Viciani,
- 743 S., Vogelmann, H., and Wienhold, F. G.: Observations of the downwelling far-infrared atmospheric emission at
- 744 the Zugspitze observatory, Earth Syst. Sci. Data, 13, 4303–4312, https://doi.org/10.5194/essd-13-4303-2021,
- 745 2021.
- 746 Parrish, D. D., Derwent, R. G., Steinbrecht, W., Stübi, R., Van Malderen, R., Steinbacher, M., Trickl, T., Ries,
- 747 L., and Xu, X.: Zonal Similarity of Long-term Changes and Seasonal Cycles of Baseline Ozone at Northern
- 748 Mid-latitudes, J. Geophys. Res., 125, e2019JD031908, https://doi.org/10.1029/2019JD031908, 19 pp., 2020.
- 749 Schultz, M. G., Schröder, S., Lyapina, O., Cooper, O., Galbally, I., Petropavlovskikh, I., von Schneidemesser,
- E., Tanimoto, H., Elshorbany, Y., Naja, M., Seguel, R. J., Dauert, U., Eckhardt, P., Feigenspan, S., Fiebig, M.,
- 751 Hjellbrekke, A.-G., Hong, Y.-D., Kjeld, P. C., Koide, H., Lear, G., Tarasick, D., Ueno, M., Wallasch, M.,
- 752 Baumgardner, D., Chuang, M.-T., Gillett, R., Lee, M., Molloy, S., Moolla, R., Wang, T., Sharps, K., Adame, J.
- A., Ancellet, G., Apadula, F., Artaxo, P., Barlasina, M. E., Bogucka, M., Bonasoni, P., Chang, L., Colomb, A.,
- 754 Cuevas-Agulló, E., Cupeiro, M., Degorska, A., Ding, A., Fröhlich, M., Frolova, M., Gadhavi, H., Gheusi, F.,
- 755 Gilge, S., Gonzalez, M. Y., Gros, V., Hamad, S. H., Helmig, D., Henriques, D., Hermansen, O., Holla, R.,
- Hueber, J., Im, U., Jaffe, D. A., Komala, N., Kubistin, D., Lam, K.-S., Laurila, T., Lee, H., Levy, I., Mazzoleni,

- 757 C., Mazzoleni, L. R., McClure-Begley, A., Mohamad, M., Murovec, M., Navarro-Comas, M., Nicodim, F.,
- Parrish, D., Read, K. A., Reid, N., Ries, L., Saxena, P., Schwab, J. J., Scorgie, Y., Senik, I., Simmonds, P.,
- 759 Sinha, V., Skorokhod, A. I., Spain, G., Spangl, W., Spoor, R., Springston, S. R., Steer, K., Steinbacher, M.,
- 760 Suharguniyawan, E., Torre, P., Trickl, T., Weili, L., Weller, R., Xiaobin, X., Xue, L., and Zhiqiang, M.:
- 761 Tropospheric Ozone Assessment Report: Database and Metrics Data of Global Surface Ozone Observations,
- 762 Elem. Sci. Anth., 5, 58, DOI: https://doi.org/ 10.1525/elementa.244, 25 pp., 2017.
- Reichardt, J., Ansmann, A., Serwazi, M., Weitkamp, C., and Michaelis, W.: Unexpectedly low ozone concentration in midlatitude tropospheric ice clouds: A case study, Geophys. Res.Lett., 23, 1929-1932, 1996.
- 765 Smit, H. G. J, Straeter, W., Johnson, B. J., Oltmans, S. J., Davies, J., Tarasick, D. W., Hoegger, B., Stubi, R.,
- 766 Schmidlin, F. J., Northam, T., Thompson, A. M., Witte, J. C., Boyd, I., and Posny, F.: Assessment of the
- 767 Performance of ECC-ozonesondes under Quasi-flight Conditions in the Environmental Simulation Chamber:
- 768 Insights from the Jülich Ozone Sonde Intercomparison Experiment (JOSIE), J. Geophys. Res., 112, D19306,
- 769 doi:10.1029/2006JD007308, 18 pp., 2007.
- 770 Smit, H.G.J., and ASOPOS panel: Quality assurance and quality control for ozonesonde measurements in GAW,
- 771 World Meteorological Organization, GAW Report No. 201, Geneva (Switzerland). [Available online at
- 772 https://library.wmo.int/doc_num.php?explnum_id=7167], 100 pp., 2014.
- 773 Smit, H.G.J., and Thompson, A.M.: Ozonesonde Measurement Principles and Best Operational Practices: 774 ASOPOS 2.0 (Assessment of Standard Operating Procedures for Ozonesondes), World Meteorological 775 Organization, GAW No. 268, (Switzerland). [Available Report Geneva online at 776 https://library.wmo.int/doc num.php?explnum id=10884], 172 pp., 2021.
- 777 Stauffer, R.M., Thompson, A.M., Kollonige, D.E., Tarasick, D.W., Van Malderen, R., Smit, H.G.J., Vömel, H.,
- 778 Morris, G.A., Johnson, B.J., Cullis, P.D., Stübi, R., Davies, J., and Yan.: An Examination of the Recent Stability
- of Ozonesonde Global Network Data, Earth and Space Science, 9 (10), e2022EA002459, [available online at
- 780 https://doi.org/10.1029/2022EA002459], 2022.
- Stein, A. F., Draxler, R. R, Rolph, G. D., Stunder, B. J. B., Cohen, M. D., and Ngan, F.: NOAA's HYSPLIT
 atmospheric transport and dispersion modeling system, Bull. Amer. Meteor. Soc., 96, 2059-2077, 2015.
- Steinbrecht, W., Schwarz, R., and Claude, H.: New pump correction for the Brewer-Mast ozone sonde:
 determination from experiment and instrument intercomparisons, J. Atmos. Ocean. Tech., 15, 144–156, 1998.
- Stohl, A., and Trickl, T.: A textbook example of long-range transport: Simultaneous observation of ozone
 maxima of stratospheric and North American origin in the free troposphere over Europe, J. Geophys. Res., 104,
 30445-30462, 1999.
- 788 Stohl, A., Bonasoni, P., Cristofanelli, P., Collins, W., Feichter, J., Frank, A., Forster, C., Gerasopoulos, E.,
- 789 Gäggeler, H., James, P., Kentarchos, T., Kromp-Kolb, H., Krüger, B., Land, C., Meloen, J., Papayannis, A.,
- Priller, A., Seibert, P., Sprenger, M., Roelofs, G. J., Scheel, H. E., Schnabel, C., Siegmund, P., Tobler, L., Trickl,
- 791 T., Wernli, H., Wirth, V., Zanis, P., and Zerefos, C.: Stratosphere-troposphere exchange a review, and what we
- 792 have learned from STACCATO, J. Geophys. Res., 108, 8516, doi:10.1029/2002JD002490, STA 1, 15 pp., 2003.
- 793 Stübi, R., Levrat, G., Hoegger, B., Pierre Viatte, P., Staehelin, J., Schmidlin, F.J.: In-flight comparison of
- 794 Brewer-Mast and electrochemical concentration cell ozonesondes, J. Geophys. Res., 113, D13302,
- 795 https://doi.org/10.1029/2007JD009091, 2008.

- 796 Tarasick, D. W., Davies, J., Anlauf, K., Watt, M., Steinbrecht, W., Claude H.-J.: Laboratory investigations of the
- response of Brewer-Mast ozonesondes to tropospheric ozone, J. Geophys. Res., 107, ACH 14-1 14-10,
- 798 https://doi.org/10.1029/2001JD001167, 2002.
- 799 Tarasick, D. W., Davies, J., Smit, H. G. J., and Oltmans, S. J.: A re-evaluated Canadian ozonesonde record:
- 800 measurements of the vertical distribution of ozone over Canada from 1966 to 2013, Atmos. Meas. Tech., 9, 195–
- 801 214, https://doi.org/10.5194/amt-9-195-2016, 2016.
- 802 Tarasick, D., Galbally, I. E., Cooper, O. R., Schultz, G M., Ancellet, G., Leblanc, T., Wallington, T. J., Ziemke,
- 803 J., Liu, X., Steinbacher, M., Staehelin, J., Vigouroux, C., Hannigan, J., García, O., Foret, G., Zanis, P.,
- 804 Weatherhead, E., Petropavlovskikh, I., Worden, H., Osman, M., Liu, J., Chang, K.-L., Gaudel, A., Lin, M.,
- Granados-Muñoz, M., Thompson, A. M., Oltmans, S. J., Cuesta, J., Dufour, G., Thouret, V., Hassler, B., Trickl,
 T., and Neu, J. L.: Tropospheric Ozone Assessment Report: Tropospheric ozone from 1877 to 2016, observed
- T., and Neu, J. L.: Tropospheric Ozone Assessment Report: Tropospheric ozone from 1877 to 2016, observed
 levels, trends and uncertainties, Elem. Sci. Anth., 7, Article 39, DOI: https://doi.org/10.1525/elementa.376, 72
- 808 pp. (plus 56 pp. of supplemental material), 2019.
- 809 Tarasick, D. W., Smit, H. G. J., Thompson, A. M., Morris, G. A., Witte, J. C., Davies, J., Davies, J., Nakano, T.,
- 810 Van Malderen, R., Stauffer, R. M., Johnson, B. J., Stubi1, R., Oltmans, S. J., and Vömel, H.: Improving ECC
- 811 ozonesonde data quality: Assessment of current methods and outstanding issues, Earth and Space Science, 8,
- 812 e2019EA000914. https://doi.org/10.1029/2019EA000914, 27 pp., 2021.
- 813 TESLAS: Tropospheric Environmental Studies by Laser Sounding (TESLAS), in: Transport and Chemical
- 814 Transformation of Pollutants in the Troposphere, Vol. 8, Instrument Development for Atmospheric Research and
- 815 Monitoring, J. Bösenberg, D. Brassington, and P. C. Simon, Eds., Springer (Berlin, Heidelberg, New York),
- 816 ISBN 3-540-62516-X, 1-203, 1997.
- 817 Thompson, A. M., Smit, H. G. J., Witte, J. C., Stauffer, R. M., Johnson, B. J., Morris, G., von der Gathen, P.,
- 818 Van Malderen, R., Davies, J., Piters, A., Allaart, M., Posny, F., Kivi, R., Cullis, P., Hoang Anh, N. T., Corrales,
- 819 E., Machinini, T., da Silva, F. R., Paiman, G., Thiong'o, K., Zainal, Z., Brothers, G. B., Wolff, K. R., Nakano,
- 820 T., Stübi, R., Romanens, G., Coetzee, G. J. R., Diaz, J. A., Mitro, S., Mohamad, M., and Ogino, S.: Ozonesonde
- Quality Assurance: The JOSIE–SHADOZ (2017) Experience, Bulletin of the American Meteorological Society,
 100, 155-171, 2019.
- 823 Trickl, T., Cooper, O. R., Eisele, H., James, P., Mücke, R., and Stohl, A.: Intercontinental transport and its
- 824 influence on the ozone concentrations over central Europe: Three case studies, J. Geophys. Res., 108, D12, 8530,
- 825 10.1029/2002JD002735, STA 15, 23 pp., 2003.
- Trickl, T., Feldmann, H., Kanter, H.-J., Scheel, H. E., Sprenger, M., Stohl, A., and Wernli, H.: Deep
 stratospheric intrusions over Central Europe: case studies and climatological aspects, Atmos. Chem. Phys., 10,
 499-524, 2010.
- 829 Trickl, T., Eisele, H., Bärtsch-Ritter, N., Furger, M., Mücke, R., Sprenger, M., and Stohl, A.: High-ozone layers
- 830 in the middle and upper troposphere above Central Europe: potential import from the stratosphere along the
- 831 subtropical jet stream, Atmos. Chem. Phys., 11, 9343-9366; 5-p. Supplement, 2011.
- 832 Trickl, T., Vogelmann, H., Giehl, H., Scheel, H. E., Sprenger, M., and Stohl, A.: How stratospheric are deep
- 833 stratospheric intrusions? Atmos. Chem. Phys., 14, 9941-9961, 2014.

- 834 Trickl, T., Vogelmann, H., Flentje, H., and Ries, L.: Stratospheric ozone in boreal fire plumes the 2013 smoke
 835 season over Central Europe, Atmos. Chem. Phys., 15, 9631-9649, 2015.
- 836 Trickl, T., Vogelmann, H., Fix, A., Schäfler, A., Wirth, M., Calpini, B., Levrat, G., Romanens, G., Apituley, A.,
- 837 Wilson, K. M., Begbie, R., Reichardt, J., Vömel, H. and Sprenger, M.: How stratospheric are deep stratospheric
- 838 intrusions into the troposphere? LUAMI 2008, Atmos. Chem. Phys, 16, 8791-8815, 2016.
- 839 Trickl, T., Neidl, F., Giehl, H., Perfahl, M., and Vogelmann, H.: Three decades of tropospheric ozone lidar
- development at Garmisch-Partenkirchen, Atmos. Meas. Tech., 13, 6357-6390, 2020a.
- Trickl, T., Vogelmann, H., Ries, L., and Sprenger, M.: Very high stratospheric influence observed in the free
 troposphere over the Northern Alps just a local phenomenon? Atmos. Chem. Phys., 20, 243-266, 2020b.
- 843 Trickl, T., Couret, C., Ries, L., and Vogelmann, H.: Zugspitze ozone 1978 2020: The role of stratosphere-
- troposphere transport, Atmos. Chem.Phys., 23, 8403–8427, 2013; Corrigendum: https://doi.org/10.5194/acp-23-
- 845 8403-2023-corrigendum
- 846 Vaisala: Vaisala Radiosonde RS41 Measurement Performance, White Paper, Vaisala, Helsinki (Finland),
- 847 https://www.vaisala.com/sites/default/files/documents/WEA-MET-RS41-Performance-White-paper-
- 848 B211356EN-B-LOW-v3.pdf, 28 pp. (accessed 7 September 2019), 2017.
- 849 Van Malderen, R., Allaart, M. A. F., De Backer, H., Smit, H. G. J., and De Muer, D.: On instrumental errors and
- 850 related correction strategies of ozonesondes: possible effect on calculated ozone trends for the nearby sites Uccle
- 851 and De Bilt, Atmos. Meas. Tech., 9, 3793–3816, 2016
- Viallon, J., Lee, S., Moussay, P., Tworek, K., Peterson, M., and Wielgosz, R. I.: Accurate measurements of ozone absorption cross-sections in the Hartley band, Atmos. Meas. Tech., 8, 1245-1257, 2015.
- 854 Völger, P., Bösenberg, J., and Schult, I.: Scattering Properties of Selected Model Aerosols Calculated at UV-
- Wavelengths: Implications for DIAL Measurements of Tropospheric Ozone, Beitr. Phys. Atmosph., 69, 177-187, 1996.
- 857 Vömel, H., David, D. E., and Smith, K.: Accuracy of tropospheric and stratospheric water vapor measurements
- by the cryogenic frost point hygrometer: Instrumental details and observations, J. Geophys. Res., 112, D08305,
- 859 doi: 10.1029/2006JD007224, 14 pp., 2007.
- Vömel, H., Naebert, T., Dirksen, R., and Sommer, M.: An update on the uncertainties of water vapor
 measurements using Cryogenic Frostpoint Hygrometers, Atmos. Meas. Tech., 9, 3755-3768, 2016.
- 862 Vömel, H., Smit, H. G. J., Tarasick, D., Johnson, B., Oltmans, S. J., Selkirk, H., Thompson, A. M., Stauffer, R.
- 863 M., Witte, J. C., Davies, J., van Malderen, R., Morris, G. A., Nakano, T., and Stübi, R.: A new method to correct
- 864 the electrochemical concentration cell (ECC) ozonesonde time response and its implications for "background
- 865 current" and pump efficiency, Atmos. Meas. Tech., 13, 5667–5680, 2020.
- 866 Vogelmann, H. and Trickl, T.: Wide-Range Sounding of Free-Tropospheric Water Vapor with a Differential-
- 867 Absorption Lidar (DIAL) at a High-Altitude Station, Appl. Opt., 47, 2116-2132, 2008.
- 868 Vogelmann, H., Sussmann, R., Trickl, T., and Borsdorff, T.: Intercomparison of atmospheric water vapor
- soundings from the differential absorption lidar (DIAL) and the solar FTIR system on Mt. Zugspitze, Atmos.
 Maga Tash, 4, 825, 841, 2011
- 870 Meas. Tech., 4, 835-841, 2011.

- 871 Vogelmann, H., Sussmann, R., Trickl, T., and Reichardt, A.: Spatiotemporal variability of water vapor
- 872 investigated using lidar and FTIR vertical soundings above the Zugspitze, Atmos. Chem. Phys., 14, 3135-3148,
- 873 2015.
- 874 VOTALP II: Vertical Ozone Transport in the Alps II, Final Report for the European Union, Contract Nr.: ENV4
- 875 CT970413, Reporting Period 1/3/1998-29/2/2000, H. Kromp-Kolb, Co-ordinator, Universität für Bodenkultur
- 876 Wien (Austria), Institut für Meteorologie und Physik, 96 pp., 2000.
- 877 Wernli, H., and Davies, H.C.: A Lagrangian-based analysis of extratropical cyclones. I: The method and some
- 878 applications. Q.J.R. Meteorol. Soc., 123: 467-489, https://doi.org/10.1002/qj.49712353811, 1997.
- 879 Sprenger, M., and Wernli, H.: The LAGRANTO Lagrangian analysis tool version 2.0, Geosci. Model Dev., 8,
- 880 2569–2586, https://doi.org/10.5194/gmd-8-2569-2015, 2015.
- Wotava, G., and Kromp-Kolb, H.: The research project VOTALP general objectives and main results, Atmos.
 Environ., 34, 1319-1322, 2000.
- 883 Yuan, Y., Ries, L., Petermeier, H., Trickl, T., Leuchner, M., Couret, C., Sohmer, R., Meinhardt, F., and Menzel,
- 884 A.: On the diurnal, weekly, and seasonal cycles and annual trends in atmospheric CO_2 at Mount Zugspitze,
- 885 Germany, during 1981–2016, Atmos. Chem. Phys., 19, 999–1012, https://doi.org/10.5194/acp-19-999-2019,
- 886 2019.
- 287 Zanis, P., Trickl, T., Stohl, A., Wernli, H., Cooper, O., Zerefos, C., Gaeggeler, H., Priller, A., Schnabel, C.,
- 888 Scheel, H. E., Kanter, H. J., Tobler, L., Kubik, P. W., Cristofanelli, P., Forster, C., James, P., Gerasopoulos, E.,
- 889 Delcloo, A., Papayannis, A., and Claude, H.: Forecast, observation and modelling of a deep stratospheric
- 890 intrusion event over Europe, Atmos. Chem. Phys., 3, 763-777, 2003.
- 891 Zellweger, C., Buchmann, B., Klausen, J., and Hofer, P.: System and Performance Audit of Surface Ozone,
- 892 Carbon Monoxide and Methane at the Global GAW Station Zugspitze/Hohenpeißenberg, Platform Zugspitze,
- 893 Germany, Empa-WCC Report 01/1, submitted to the World Meteorological Organization, 49 pp., February
- 894 2001.
- 895 Zellweger, C., Klausen, J., and Buchmann, B.,: System and Performance Audit of Surface Ozone, Carbon
- 896 Monoxide and Methane at the Global GAW Station Zugspitze/Schneefernerhaus, Germany, Empa-WCC Report
- 897 06/2, submitted to the World Meteorological Organization, 51 pp., June 2006.
- 898 Zellweger, C., Steinbacher, M., and Buchmann, B., and Steinbrecher, R.: System and Performance Audit of
- 899 Surface Ozone, Methane, Carbon Dioxide, Nitrous Oxide and Carbon Monoxide at the Global GAW Station
- 900 Zugspitze-Schneefernerhaus, Germany, submitted to WMO by WMO World Calibration Centre WCC-Empa
- 901 Empa Dübendorf, Switzerland, 46. pp., WCC-Empa Report 11/2, June 2011.
- 902 Zellweger, C., Steinbacher, M., Buchmann, B., and Steinbrecher, R.: System and Performance Audit of Surface
- 903 Ozone, Methane, Carbon Dioxide, Nitrous Oxide and Carbon Monoxide at the Global GAW Station Zugspitze-
- 904 Schneefernerhaus, Germany, submitted to WMO by WMO World Calibration Centre WCC-Empa Empa
- Dübendorf, Switzerland, WCC-Empa Report 20/3, September 2020, GAW report 266, 54. pp., 2021.
- 906
- 907



911 Fig. 1. Ozone measurements at Garmisch-Partenkirchen (IFU, UFS) and Hohenpeißenberg (MOHp) on 2 912 October 2017; the low relative humidity between 5.2 and 8.3 km (RH = 2 %) verifies the presence of a 913 stratospheric air intrusion. The time for MOHp is the launch time of the sonde.



916 Fig. 2. Four ozone measurements on 5 and 6 February 2019 with lidar (IFU), ECC sonde (FZJ) and an in-situ 917 sensor at UFS; for two measurements the FZJ ozone mixing ratios are slightly higher than the lidar results. The 918 fourth FZJ ozone measurement took place much later than the final lidar measurements which resulted in slightly 919 larger differences up to 4.8 km, confirmed by the 23:30-CET measurement at UFS. The lidar results around 10 920 km on 6 February are uncertain due to a cirrus correction. In order to visualize more details on the complex 921 layering we also show water-vapour mixing ratios for roughly coinciding measurements of the UFS Raman lidar 922 and the FZJ CFH sonde. The tropospheric structures are strongly smoothed for the lidar due to the 1-h data-923 acquisition time. At 3.3 km 250 ppm corresponds to roughly 5 % RH.



Fig. 3. Averaged differences between FZJ ozone sonde and IMK-IFU lidar for the first three comparisons after a
slight offset correction of the sonde profiles (see text)



Fig. 4. Ozone measurements on 15 January 2018: The MOHp ozone (red) is also shown shifted by 5.8 ppb to
match the lidar ozone and the UFS value (cyan), in part the black, in part the blue curve. Differences exist in the
tropopause region, which is frequently the case. The sawtooth structure in the MOHp data is due to insufficient
digital resolution in the NDACC data base.



Fig. 5. Summertime ozone measurements (July 9, 2018) with pronounced layering; the sonde ozone (red) is brought to reasonable agreement with the lidar (black curve) above 2.7 km by adding 6 ppb (cyan curve). Above 940 9 km the air masses are no longer comparable. The particularly strong discrepancy of the UFS in-situ ozone can be explained by orographic lifting of the ozone edge at 2.7 km. The low to moderate RH (grey) in parts of the free troposphere indicates that the elevated ozone values could be due to a stratospheric air component.



945 Fig. 6. 350-h LAGRANTO backward trajectories, started above Garmisch-Partenkirchen (G) on 9 July 2018 at
 946 7:00 CET



947 Fig. 7. 350-h LAGRANTO backward trajectories, started above Garmisch-Partenkirchen (G) on 9 July 2018 at
 8:00 CET



Fig. 8. Ozone measurements on 13 August 2018: The agreement of the shifted MOHp ozone profile (cyan) with
the lidar curves is rather good up to 12 km given the high summertime variability. The low to moderate RH
above 4.4 km (grey) indicates that the elevated ozone is partially caused by stratospheric air.



955 Fig. 9. Ozone measurements on 15 October 2018: The MOHp ozone (red) is not shifted. The agreement above 956 4.3 km is better with the earlier lidar measurement (black), above 7 km better with the blue curve. The lidar data 957 are strongly smoothed in the stratosphere, as can be seen from the more detailed ozone structure in the sonde 958 data. This example is one of the two examples with a pronounced low-altitude discrepancy between lidar and 959 sonde extending to more the 3 km.



Fig. 10. Differences between the ozone values of the IFU DIAL at 2670 m and the UFS routine measurements as
well as the offsets of the MOHp profiles with respect to the DIAL for 35 of the 36 measurement days of the 2018
comparison. The blue and red horizontal lines are the arithmetic averages for the full year (for the values see
text).



968 Fig. 11. Average differences between IFU lidar and offset-corrected MOHp sonde in 2018 for low-, moderate 969 and high-conditions (based on six, seven and six comparisons, respectively); the uncertainties may be estimated 970 from the maximum differences around the respective altitudes. We also indicate the approximate altitude ranges 971 of the two wavelength pairs used for the lidar data evaluation.



974 Fig. 12. Differences between the ozone mixing ratios of the lidar (IFU) and the stations Zugspitze (Z), Wank
975 (W) at the summit altitudes, and offsets between lidar and MOHp sonde for 2009



978 Fig. 13. Ozone measurements on 12 January 2009979



Fig. 14. Ozone measurements on 17 August 2009; the structure in the upper troposphere is strongly influencedby smoothing. The bias between 5.5 and 8 km has not been explained.



985 Fig. 15. 350-h LAGRANTO backward trajectories, started above Garmisch-Partenkirchen (G) on 9 July 2018 at

986 7:00 CET



989 Fig. 16. Ozone measurements on 11 September 2000 (see Fig. 13 of Trickl et al., 2003); in this case no offset990 was determined.



Fig. 17. Ozone measurements on 20 June, 2001; the entire temporal development of the stratospheric air
intrusion around 7.3 km is depicted in Fig. 6 of Zanis et al., 2003, and Fig. 3 of Trickl et al., 2010)
995



Fig. 18. Differences between the ozone mixing ratios of the lidar (IFU) and the stations Zugspitze (Z), Wank
(W) at the summit altitudes, and between lidar and MOHp sonde, determined by shifting the sonde profile, for
the period September 2000 to August 2001.