

High precision oxygen isotopes ($\delta^{18}\text{O}$) measurements of atmospheric dioxygen using optical-feedback cavity-enhanced absorption spectroscopy (OF-CEAS)

Commenté [A1]: Corrected in all the manuscript accordingly to technical comment of Referee 3

Commenté [A2]: Modified according to Referee 1 technical comment.

Clément Piel¹, Daniele Romanini², Morgane Farradèche³, Justin Chaillot³, Clémence Paul³, Nicolas Bienville³, Thomas Lauwers³, Joana Sauze¹, Kevin Jaulin⁴, Frédéric Prié³, Amaëlle Landais³

¹ Ecotron Européen de Montpellier (UAR 3248), Centre National de la Recherche Scientifique (CNRS), Université de Montpellier, Campus Baillarguet, Montferrier-sur-Lez, France

² Laboratoire Interdisciplinaire de Physique, Univ Grenoble Alpes, CNRS/UGA, Saint-Martin-d'Hères, France

³ Laboratoire des Sciences du Climat et de l'Environnement, LSCE - IPSL, CEA-CNRS-UVSQ, Université Paris-Saclay, 91191 Gif-sur-Yvette, France

⁴ AP2E, 240, rue Louis de Broglie, Les Méridiens Bâtiment A, CS90537, F-13593 Aix-en-Provence, FRANCE

Correspondence: Clément Piel (clement.piel@cnrs.fr)

Short summary

This paper introduces a new optical gas analyzer based on the Optical-Feedback Cavity-Enhanced Absorption Spectroscopy technique (OF-CEAS) enabling high temporal resolution and high precision measurement of $\delta^{18}\text{O}$ of atmospheric O_2 (0.06 ‰ over 10 minutes integration) and concentration of O_2 in the atmosphere (0.0002 % over 10 minutes integration). The results underscore the good agreement with dual inlet IRMS measurements and the ability of the instrument to monitor biological processes.

Commenté [A3]: Modified according to Referee 1 technical comment.

Commenté [A4]: Modified according to Referee 1 technical comment.

Abstract

Atmospheric dioxygen (O_2) concentration and isotopic composition are closely linked to the carbon cycle through anthropic carbon dioxide (CO_2) emissions and biological processes such as photosynthesis and respiration. Measurement of isotopic ratio of atmospheric dioxygen, trapped in ice core bubbles, brings information about past variation in the hydrological cycle at low latitudes, as

25 well as past productivity. Currently, the interpretation of those variations could be drastically improved with a better (*i.e.* quantitative) knowledge of the oxygen isotopic fractionation that occurs during photosynthesis and respiration processes. This could be achieved, for example, during experiments using closed-biological chambers. In order to estimate the isotopic fractionation coefficient with a good precision, one of the principal limitations is the need for high frequency on-line measurements of $\delta^{18}\text{O}$ of O_2 ($\delta^{18}\text{O}(\text{O}_2)$) and O_2 concentration. To address this issue, we developed a new instrument, based on the optical-feedback cavity-enhanced absorption spectroscopy (OF-CEAS) technique, enabling high temporal resolution and continuous measurements of dioxygen concentration as well as $\delta^{18}\text{O}(\text{O}_2)$, both simultaneously. Minimum of Allan deviation occurred between 10 and 20 minutes while precision reached 0.002 % for O_2 concentration and 0.06 ‰ for $\delta^{18}\text{O}(\text{O}_2)$, which correspond to the optimal integration time and analytical precision before instrumental drift started degrading the measurements. Instrument accuracy was in good agreement with dual-inlet isotope ratio mass spectrometry. Even if humidity did not affect much the measured values, we decided to measure $\delta^{18}\text{O}(\text{O}_2)$ and O_2 concentration after drying the gas. On the other hand, 1 % increase in O_2 concentration increased by 0.53 ‰ the $\delta^{18}\text{O}(\text{O}_2)$. To ensure good quality of O_2 concentration and $\delta^{18}\text{O}(\text{O}_2)$ measurements we eventually proposed to measure calibration standard every 20 minutes.

1. Introduction

Dioxygen (O_2) is the second most important constituent of the atmosphere and the evolution of its atmospheric concentration is closely related to the evolution of the carbon cycle through fossil fuel combustion and biosphere processes (respiration and photosynthesis). The O_2 concentration has been shown to decrease by 0.7 % over the last 800 ka probably because of changes of organic carbon burial and oxidation rates with a stabilizing effect of silicate weathering (Stolper et al., 2016). On shorter timescales, the O_2 concentration is showing clear seasonal variations anticorrelated with atmospheric CO_2 concentration, the combination of both O_2 and CO_2 concentrations enabling to document the variability of the marine and terrestrial biosphere productivity through calculation of the atmospheric potential oxygen (*e.g.* Goto et al., 2017; Keeling and Manning, 2014; Keeling and Shertz, 1992; Stephens et al., 1998).

The isotopic composition of O_2 , expressed as $\delta^{18}\text{O}$ of O_2 ($\delta^{18}\text{O}(\text{O}_2)$), exhibits variations at the orbital (10,000 years) and millennial scale (Landais et al., 2010; Severinghaus et al., 2009) but no appreciable variation at the seasonal scale has been evidenced. Past $\delta^{18}\text{O}(\text{O}_2)$ variability has been linked to

Commenté [A5]: Added accordingly to technical comment referee 1

Commenté [A6]: Modified (Technical comment Referee 1)

Commenté [A7]: Modified (Technical comment Referee 1)

Commenté [A8]: Notation explained (Referred 1 technical comment)

Commenté [A9]: Modified according to Referee 1 technical comments.

Commenté [A10]: Improved explanation according to Referee 1 specific comments.

Commenté [A11]: In response to Referee 3 specific comment 1.

Commenté [A12]: Modified according to Referee 1 technical comment.

Commenté [A13]: Notation explained (Referred 1 technical comment)

Commenté [A14]: Symbol used throughout the manuscript, according to Referee 1 specific comment.

Commenté [A15]: New notation used throughout the manuscript, according to Referee 1 specific comment.

variations in low latitude water cycle and possibly to the variability of the relative proportions of terrestrial and marine productivity (Bender et al., 1994; Extier et al., 2018; Luz and Barkan, 2011). Still, evaluating the relative importance of these two contributions remains difficult since it relies on fractionation factors associated with the biological processes consuming and producing oxygen. These
60 fractionation factors have only been determined for a small number of species and often at the micro-organism scale.

Combining measurements of $\delta^{17}\text{O}$ of O_2 ($\delta^{17}\text{O}(\text{O}_2)$) and $\delta^{18}\text{O}(\text{O}_2)$ permits to have access to gross primary production and is largely used for this purpose in the ocean in combination with the elemental ratio between dioxygen and argon, i.e. O_2/Ar ratio (Jurikova et al., 2022; Luz and Barkan, 2000; Stanley
65 et al., 2010). In addition, when measuring both $\delta^{17}\text{O}(\text{O}_2)$ and $\delta^{18}\text{O}(\text{O}_2)$ in old air trapped in ice cores, we have access to the past variability of the global biosphere productivity although with large uncertainties (Blunier et al., 2002; Brandon et al., 2020; Luz et al., 1999; Yang et al., 2022). Part of this limitation is again linked to our poor knowledge of the isotopic fractionation factors during the biological processes.

Commenté [A16]: New notation used throughout the manuscript, according to Referee 1 specific comment.

70 In order to progress on the quantitative determination of the oxygen fractionation factors in the biosphere, a good approach is to work on closed biological chambers where plants grow in controlled conditions. This approach was already applied in previous studies, mostly at the micro-organism scale with regular sampling of air from the biological set-up (Guy et al., 1993; Helman et al., 2005; Stolper
75 et al., 2018). At a larger scale such as for a macroscopic study on a terrarium the duration of the experiment needs to be larger to have a quantified signal (Luz et al., 1999).

One of the main limitations when working on the experimental determination of the fractionation factors is that we need numerous measurements of $\delta^{18}\text{O}(\text{O}_2)$ and O_2 concentration to make a precise determination (Paul et al., 2023). However, sampling air from the chamber at high resolution is not convenient since it implies to reduce the quantity of atmospheric air in the biological chamber.

80 Development of continuous measurements of dioxygen concentration has permitted to improve the monitoring of O_2 concentration in parallel with greenhouse gases. These methods use different techniques such as fuel cell analyzer (Goto et al., 2013) or gas chromatography coupled with a thermal conductivity detector (Tohjima, 2000). More recently, Berhanu et al. (2019) developed an analyzer based on the cavity-ring-down spectroscopy (CRDS) technique which is able to measure the
85 concentration of O_2 high precision (standard error of 0.0001 % over 1-minute measurement) on one mode as well as the $\delta^{18}\text{O}(\text{O}_2)$ with a 0.01 per mil standard error) on a second mode. However, $\delta^{18}\text{O}(\text{O}_2)$

Commenté [A17]: Modified (Technical comment Referee 1)

and O₂ concentrations are not measured on the same mode and so not at the same time (an instrument software restart is required for switching between modes).

Commenté [A18]: Improved according to Referee 3 specific comment 3.

We present here a new instrument based on the optical-feedback cavity-enhanced absorption spectroscopy (OF-CEAS) technique, primarily designed and optimized to measure $\delta^{18}\text{O}(\text{O}_2)$, but also able to measure the atmospheric O₂ concentration in parallel. After a presentation of the instrumental design, we discuss the performances of the instrument as well as the influences of humidity on the determination of O₂ concentration and isotopic composition and O₂ concentration on $\delta^{18}\text{O}(\text{O}_2)$. Based on our tests, we propose a sequence for calibration of the instrument during series of measurements. This sequence is then used to compare the performance of this instrument to a classical IRMS for measurements of $\delta^{18}\text{O}(\text{O}_2)$ and O₂ concentration.

Commenté [A19]: Modified according to Referee 1 specific comments.

Commenté [A20]: Modified according to Referee 1 specific comments.

2. Material and methods

2.1 Instrument description

Commenté [A21]: Many detailed explanations were added to this section, in answer to the demand of all referees.

We present here a new analyzer for O₂ exploiting the high sensitivity absorption spectroscopy technique OF-CEAS. This technique is based on a high finesse optical cavity and is analogous to the well-known Cavity-Ring-Down Spectroscopy. It was first described in Morville et al. (2005) (see also Morville et al., 2014) and since then successfully exploited in different laboratories and field applications such as multispecies breath analysis (Ventrillard-Courtillot et al., 2009), analyses of geothermal gases (Kassi et al., 2006) or analyses of fossil air extracted from ice cores (Faïn et al., 2014). In addition, the simple and compact optical layout of OF-CEAS led to the development of robust devices now commercialized by the AP2E company [www.AP2E.com]. The instrument used in this study is based on an AP2E analyzer customized to satisfy some specific demands, in our case the possibility to measure simultaneously $\delta^{18}\text{O}(\text{O}_2)$ and O₂ concentration because no such analyzer has been commercialized by the company before. To our knowledge this is the first OF-CEAS implementation using a distributed feed-back (DFB) diode laser in the visible range. For the rest, this instrument represents a standard implementation of the OF-CEAS technique as described in above cited publications, therefore we address the interested reader to those references for an OF-CEAS schematic and details of its working principle and operation (in particular Morville et al. 2014, and in addition Lechevallier et al., 2019). Variations relative to those publications are the proprietary AP2E electronics and software for system and laser control, acquisition and digitization and analysis of signals. In addition, the AP2E instrument opto-mechanical layout is a revisited version of previous OF-

Commenté [A22]: Modified accordingly to Referee 1 technical comments

Commenté [A23]: Examples of applications added, according to Referee 1 technical comment.


Commenté [A24]: Added accordingly to Referee 1 technical comments


Commenté [A25]: Modified accordingly to Referee 1 technical comments. Then only DFB is used

Commenté [A26]: Added accordingly to Referee 1 technical comments

CEAS layouts, of which different versions can be found in the cited works. In any case, the AP2E proprietary implementation follows the previously described OF-CEAS principles of operation, signal analysis, etc., all the way to obtaining absorption spectra like those in Figure 1, which are then fitted by a multi-line spectrum model with polynomial baseline, also previously discussed.

Concerning the most relevant difference relative to previous OF-CEAS implementations, the selected operating wavelength of the DFB diode laser targets a specific region of the dioxygen absorption band around 760 nm. This so-called oxygen “A” band is the strongest absorption band of this molecule in the visible and infrared spectral range, even though it is still a relatively weak “forbidden” magnetic-dipole transition. In this spectral region, several weak lines of the same transition for the ^{18}O - ^{16}O and ^{17}O - ^{16}O isotopologues are present near-by strong lines of the main ^{16}O - ^{16}O isotopologue. The specific spectral window chosen for our analyzer, which can be covered by a single current-driven laser frequency scan, is displayed in Figure 1, top panel. A second novelty of this instrument, discussed further below, is the presence of a saturated absorption line in the measured absorption spectrum, as visible in Figure 1, bottom panel.





Commenté [A27]: Bottom figure added to answer specific comment of the referee 2 and 3

135 Figure 1: Top: Air absorption spectrum from the instrument (green) superposed with HITRAN2021
 140 simulation (black dotted) which results as the sum of absorption from the main O₂ isotopologues and
 the ¹⁸O and ¹⁷O varieties (respectively in red and blue). Simulated water vapor absorption for a 2 %
 molar fraction (corresponding to a high level of ambient relative humidity) is plotted to show that it
 can be neglected in the spectral analysis. Vertical scale is logarithmic in order to allow clearer display
 of weaker absorption features. Bottom: OF-CEAS absorption spectrum (black) superposed with
 145 spectral fit (black dotted), together with baseline polynomial from fit (green) and fit residuals (data-
 fit) multiplied by ten (blue).

The analyzer is a modified prototype of a new series of AP2E devices under development at the time
 of this work, featuring enhanced thermal insulation and stabilization (in particular, sample cell
 145 temperature is 37 °C with mK fluctuations). As is the case for most instruments, improved temperature
 stabilization allows stable measurements on longer time scales, as necessary in this case for averaging
 isotopic ratios and detecting their small variations when switching among samples, and in particular
 for isotopic ratio measurements relative to a reference sample. Since gas sample switching typically
 implies dead times of about 2 minutes, due to limited flow rates and memory effects, and since at
 150 least 5 minutes averaging are additionally required to attain sufficient precision, the shortest time for
 measurement of one sample is in the range of 5 to 10 minutes. We then need to consider the
 instrumental drift (assessed by the Allan deviation presented below) which should remain below the
 desired precision level over the measurement time for two samples (~ 20 minutes) of which one would

155 be a reference. Smaller drift would afford duty cycle gains by flushing several samples through the analyzer before injection of one calibration sample. The chosen measurement strategy will be discussed later.

Commenté [A28]: Modified accordingly to Referee 1 technical comments

160 The most recent description of an OF-CEAS instrument close to the present one is described in Lechevallier et al. (2019). The main difference is the visible DFB diode laser emitting at 760 nm (Toptica LD-0760-0040-DFB) in place of the mid-infrared DFB intraband cascade laser (ICL) of that work. Obviously, optical elements are replaced with ones adapted to 760 nm, including a simple room-temperature Si PIN photodiode to monitor cavity transmission. However, no reference photodiode is installed, with a simulated ramp proficiently replacing the laser power ramp for normalization of cavity transmission signals.

Commenté [A29]: Added accordingly to referee 3 specific comment

165 For this setup we obtained from Layertec cavity mirrors (1 m radius of curvature) with a reflection coefficient as good as 0.999956, allowing for ring-down times of up to 30 μ s (with room air at the working pressure of 150 mbar), from a 40 cm cavity as in Lechevallier et al. (2019). The sample cell has a small volume (less than 20 cc) which, combined with a low sample pressure (150 mbar), results in a transit time of about 10 s with a sample flow of only \sim 15 sccm. Another improvement is a more accurate, stable and fast control of the sample pressure inside the measurement cell, which is also important for low drift. In particular, the fast response allows minimization of flow and pressure perturbations when switching between samples, which contributes to obtaining short commutation times.

Commenté [A30]: Modified accordingly to Referee 1 technical comments

175 As previously described Kerstel and Gianfrani (2008) isotopic ratios may be measured by optical absorption spectroscopy by obtaining spectra containing at the same time the absorption line profiles of an isotopologue and its main variety. In our case, we are considering the dioxygen isotopologues mentioned above, in particular the ^{18}O - ^{16}O variety. A simulated absorption spectrum showing close-by absorption lines for the working spectral region of the instrument is plotted in Figure 1 (based on the well-known HITRAN spectral database <https://hitran.iao.ru/>) together with a real spectrum obtained from the instrument. As visible in the bottom panel in Figure 1, the strong ^{16}O - ^{16}O line is truncated at the top due to saturation of the absorption scale of the instrument (this point will be addressed below). The isotopic ratio is then obtained as the ratio of the line areas for the considered isotopologue normalized to that of the main isotopologue. It is straightforward to demonstrate that, under conditions of constant gas sample temperature and pressure, this ratio is indeed linearly proportional to the ratio of the mole fractions of these species in the sample. The isotopic ratio variation (delta value - δ - in per mil units - ‰) is then calculated after measuring the ratio of areas of

Commenté [A31]: Added accordingly to Referee 1 technical comments

Commenté [A32]: Modified accordingly to Referee 1 technical comments

Commenté [A33]: Modified accordingly to Referee 1 technical comments

the same absorption lines for a reference gas sample relative to which the delta will be defined, by using the Equation 1 (written here for the ^{18}O - ^{16}O case):

$$\delta^{18}\text{O} = \left(\frac{\left(\frac{A_{18}}{A_{16}} \right)_{\text{sample}}}{\left(\frac{A_{18}}{A_{16}} \right)_{\text{reference}}} - 1 \right) \times 1000 \quad \text{Eq. 1}$$

where A_{18} and A_{16} are the area of the absorption line for ^{18}O - ^{16}O and ^{16}O - ^{16}O respectively.

Commenté [A34]: Modified accordingly to Referee 1 technical comments

Commenté [A35]: Modified accordingly to Referee 1 technical comments

Commenté [A36]: Modified accordingly to Referee 1 technical comments

190 Given the low abundances of the ^{18}O - ^{16}O (0.2 %) and ^{17}O - ^{16}O (0.04 %) relative to the main isotopologue ^{16}O - ^{16}O , the corresponding absorption lines appear much less intense, by about the abundance ratio (Figure 1). This imposes that the instrument sensitivity be high, in order to obtain the absorption lines of these minor isotopologues with sufficient signal to noise ratio. On the other hand, the absorption signal of the main variety will be so strong as to actually exceed the dynamic range of the instrument.

195 Indeed, the high finesse optical cavity containing the sample becomes opaque (no light transmitted) when the laser diode is tuned across the center of any of the ^{16}O - ^{16}O lines available near-by the strongest ^{18}O - ^{16}O and ^{17}O - ^{16}O lines in this absorption band. This problem could be avoided by choosing weaker absorption bands at other wavelengths, but no instrument would be capable as of today of providing a sufficiently good signal for the lines of the minor isotopologues. One way to avoid this

200 would be to use a second diode laser to address a sufficiently weak main isotopologue line lying rather far away in the same absorption band or even in another absorption band (each diode laser covers a small spectral region). However, this would double size and complexity of the experimental setup. Therefore, we decided to rely on the measured wings of the main isotopologue line in order to obtain an estimate of the absorption line area. This proved to be a good strategy given the excellent

205 performance, comparable to those obtained by the OF-CEAS technique with non-saturated absorption lines (see for example isotopic measurements on H_2O , Landsberg et al., 2014). Even if this estimate may be affected by a multiplicative form factor, this same factor will affect measurements of the reference sample and cancel out in the above equation. The situation is particularly favorable for measurement of samples presenting small concentration variations from the target species, as is the

210 case here for dioxygen varying from about 18 to 23 % in biological experiments.

In order to obtain the area of the saturated line of the main isotopologue, we exploit the fact that the data analysis software is able to “count” the cavity modes, each corresponding to a data point in the spectrum (Morville et al., 2014). This works perfectly even in the presence of a large region in the laser spectral scan where points are missing. This is possible thanks to the fact that cavity modes appear

215 uniformly spaced over the laser scan. In addition, we had to take care of the fact that, due to temperature drift, the cavity length slowly changes. As a consequence, the absolute frequency

Commenté [A37]: Modified accordingly to Referee 1 technical comments

positions of the cavity modes changes, thus the position of the data points over the absorption line profiles changes too. In particular there are two modes which correspond to the first exploitable data points in the spectrum on the two sides of the strong line, which may go above/below the threshold applied in the analysis software for using them in the spectral fit. Such a change of the number of datapoints in the fit causes a discontinuity in the retrieved line areas used to establish the concentration of the corresponding molecule. In order to fix this problem, we used the position of the data points relative to the absorption lines to continuously adjust (at the mK level) the cavity temperature setpoint and keep the cavity modes at absolute fixed frequency positions (with time fluctuations smaller than 1 % of the cavity mode spacings, or less than 1 MHz in absolute frequency units). This additionally provides more stable retrieval of all line areas from the spectral fit, as the spectral data points keep fixed positions relative to the absorption profiles which are fitted by Rautian profiles. Even if the Rautian model is much better than the more widely used Voigt profile, it still does not perfectly describe real line profiles. Thus, a displacement of the data points induces small changes in the fit results (even in the absence of noise).

Commenté [A38]: Paragraph improved with more information in answer to Referee 2 specific comments.

Over the time span of presented results (18 months), we could test two different instrument configurations, but we obtained similar performances. The main difference between the two configurations is the high reflectivity mirrors installed on the V-shaped optical cavity. For the initial configuration and its corresponding set of mirrors ($R = 0.999979$), the cavity finesse ($\sim 74\ 800$) was higher by about a factor 2. The cavity throughput was also higher, which induced a higher level of scattered light and stronger interference fringes on the recorded spectra. The set of mirrors ($R = 0.999956$), corresponding to the second configuration with lower cavity finesse ($F \sim 35\ 700$), had actually also lower transmission (thus relatively higher absorption losses in the coatings). This produced less signal on the photodetector at cavity output but with fractionally less parasitic fringes, partially compensating for the lower finesse. In both cases the setup was carefully optimized, in particular light traps were placed at all positions possibly causing light scatter by secondary reflected beams. In the following, we will specify which setup was used for which results, accounting for somewhat varying performances. It should be noted that for both configurations, the laser scan time over the monitored spectral window displayed in Figure 1 was the same and about 300 ms, while the photodiodes response and digital acquisition time was around $3\ \mu\text{s}$ in order to resolve the profiles of the transmitted cavity modes in the presence of optical feedback, and also the ringdown decay obtained at the end of each laser scan, as detailed in the cited OF-CEAS references.

Commenté [A39]: Added accordingly to Referee 1 technical comments

Commenté [A40]: Added accordingly to Referee 1 technical comments

Commenté [A41]: Added accordingly to Referee 1 technical comments

Commenté [A42]: Added accordingly to Referee 1 technical comments

Commenté [A43]: Added details in answer to referee 1 and 2 specific comments.

2.2 Air stream selection

250 The air stream allowed to the instrument inlet was selected using a Valco multiposition valve (EUTF-SD8MWE, VICI AG, Switzerland), 1/8 PFA (Perfluoroalkoxy alkanes) tubing (PFA-T2-030-100, Swagelok), and filtered using a Swagelok 7 µm filter (SS-2F-7). Three gas streams, at ambient pressure, could be selected by the Valco valve: (1) synthetic mixture of O₂ and dinitrogen (N₂) with an O₂ concentration similar to ambient (synthetic dry air, Alphagaz 2, Air Liquide France), (2) ambient atmospheric air (sampled at the inlet of the room ventilation duct) dried with a 40 mL magnesium perchlorate filter, 255 (3) natural dry atmospheric air cylinder enriched in oxygen (Natural Air enriched at 22.9 % O₂, Air Liquide Spain) subsequently diluted with N₂ (Alphagaz 2, Air Liquide France) using two mass flow controllers (F200CV and F201CV, Bronkhorst, The Netherlands). All the measurements took place in a temperature-controlled room, with temperature fluctuations within +/- 2 °C around setpoint.

Commenté [A44]: Added accordingly to Referee 1 technical comments

Commenté [A45]: Modified (Technical comment Referee 1)

260 When measuring atmospheric air, humid air was dried using a 20 cm long home-made humidity trap made with 6 mm PFA tubing (Swagelok PFA-T6M-1M-30M) and filled by magnesium perchlorate. This trap was daily replaced. For water vapor dependency measurements, the humidity trap was removed, and a stream of synthetic air (Air Product, < 3 ppm of water), with constant O₂ concentration and δ¹⁸O(O₂), was humidified at a constant setpoint using a vapor generator (see Leroy-Dos Santos et al., 2021 for more details) up to 9,000 ppm.

Commenté [A46]: Modified (Technical comment Referee 1)

265 2.3 Isotope ratio mass spectrometry (IRMS) measurements

IRMS analyses were used as a reference method to validate measurement achieved with OF-CEAS technique. Samples of air were punctually collected in 5 mL glass flasks to be analyzed through IRMS. Operation of the automated sampling system is detailed in Paul et al. (preprint, 2024). Briefly, air from a closed chamber system was circulated through the instrument and the flask for 30 min. Then the 270 flask was automatically isolated thanks to two valves and collected to be analyzed. Results from the flask and the 30 min data acquisition were, in the end, compared.

Commenté [A47]: Modified (Technical comment Referee 1)

Flasks were analyzed for δ¹⁸O(O₂) and δO₂/Ar. δO₂/Ar can be converted into O₂ concentration (or vice versa) following Equation 2:

$$\delta_{\frac{O_2}{Ar}} = \left(\frac{\left(\frac{[O_2]}{[Ar]} \right)_{sample}}{\left(\frac{[O_2]}{[Ar]} \right)_{reference}} - 1 \right) \times 1000$$

Eq. 2

Commenté [A48]: Modified accordingly to Referee 1 technical comments

275 [O₂] and [Ar] are the concentrations of O₂ and Ar respectively and the reference is the present-day atmospheric air.

Before analysis, samples are purified to isolate O₂ and Ar from total air using a gas chromatography column as in Barkan and Luz (2003). Measurements are then performed using a Thermo MAT 253 mass spectrometer (Thermo Scientific) equipped with dual inlet as described in Paul et al. (2023). For 280 each sample, 2 sequences of 16 measurements are performed for measurements of $\delta^{18}\text{O}(\text{O}_2)$ with a final uncertainty of 0.03 ‰ (1 σ). These sequences are followed by two peak jumping sequences to determine $\delta\text{O}_2/\text{Ar}$ which is associated with a final uncertainty of 0.5 ‰ (1 σ).

Commenté [A49]: Modified (Technical comment Referee 1)

3. Results and discussion

3.1 Precision and drift

285 The first characterization of the instrument was done through the measurement of the Allan-Werle deviation (Figure 2), which we will call Allan deviation for simplicity. We introduced in the instrument over 24 hours atmospheric air (dried with a 30 mL magnesium perchlorate trap), which had a constant elemental and isotopic O₂ composition. Allan deviation plots were obtained repeatedly for each different configuration, and the result shown is a typical reproducible result. The minimum of the Allan 290 deviation was reached between 10 and 20 minutes of measurements for both O₂ concentration and $\delta^{18}\text{O}(\text{O}_2)$, and with both configurations (Figure 2). The precision achieved for O₂ concentration measurement was better using the initial configuration (0.002 % at 10 minutes averaging; Figure 2 dark green data) than the second configuration (0.005 % at 3 min; Figure 2 light green data). On the contrary, the $\delta^{18}\text{O}(\text{O}_2)$ minimal Allan deviation was lower with the second configuration than the initial 295 one (50 ppm at 6 min vs 80 ppm at 20 min respectively; Figure 2 grey and dark data).

Commenté [A50]: Modified (Technical comment Referee 1)

Commenté [A51]: Modified (Technical comment Referee 1)

Commenté [A52]: Modified (Technical comment Referee 1)




Figure 2: Top panel shows continuous measurement of $\delta^{18}\text{O}(\text{O}_2)$ (grey and black) and O_2 concentration (light and dark green). Bottom panel shows Allan deviation plots based on continuous measurement of both $\delta^{18}\text{O}(\text{O}_2)$ (grey and black) and O_2 concentration (light and dark dotted green). Black and dark green are used for the initial configuration; grey and light green for the second configuration.

300


After 2 hours, the values of the Allan deviations stayed below 0.2 ‰ for $\delta^{18}\text{O}(\text{O}_2)$ and 0.03 % for O_2 concentration. Considering this drift, in order to maintain measurement repeatability below 0.1 ‰ for $\delta^{18}\text{O}(\text{O}_2)$ we see that calibrations with a measurement of a standard sample are needed over time intervals of at most 2000 s (30 minutes), in particular for the first configuration (for the second, the Allan deviation for the delta appears to reach the 0.1 ‰ limit over longer times).

305

Commenté [A53]: Modified (Technical comment Referee 1)

To assess the stability of the instrument on periods longer than 2 hours, we ran measurements of two standards during 8 consecutive days. The 2 standards were dry atmospheric air and a mixing of O₂ and N₂ in atmospheric proportions (synthetic air from Air Liquide). The evolution of the Allan deviation on Figure 3 was calculated from successive injections of the two standards every 4 minutes. This calibration frequency, higher than required by the Allan deviation discussed above, was chosen as a compromise towards obtaining measurements with high time resolution. Averages of δ¹⁸O(O₂) and O₂ concentration were calculated taking into account the 2 last minutes (the first two minutes after injection were removed from the calculation to take into account memory effect). With this approach, we obtained an Allan deviation for calibrated measurements of δ¹⁸O(O₂) below 0.03 ‰ after 6 hours of measurements. The evolution of O₂ concentration was marked by a diurnal variability which may have been linked with room temperature changes. The Allan deviation however reached values lower than 1.5.10⁻³ ‰ after 15 hours of measurements.

Commenté [A54]: Added (Technical comment Referee 1)



Commenté [A55]: Modified (Technical comment Referee 1)

320

Figure 3: Evolution of difference in mean δ¹⁸O(O₂) (black line) and O₂ concentration (green dotted line) when running successive measurements of two samples (dry atmospheric air and synthetic air) at a periodicity of 8 minutes (4 minutes of atmospheric air followed by 4 minutes of synthetic air). The data used for calculation are the mean values obtained over the last 2 minutes of measurements of each

325 gas. For this series of measurements, the initial configuration has been used but a similar result was
obtained with the second configuration.

3.2 Influence of water vapor concentration

330 We measured the dependence of $\delta^{18}\text{O}(\text{O}_2)$ and O_2 concentration on humidity and found that there
was only a slight dependence. A dependence of $\delta^{18}\text{O}(\text{O}_2)$ and O_2 concentration on humidity could be
expected for two reasons. First, due to weak water vapor absorption lines. However, there are no
water lines with sufficient intensity to be relevant even at high humidity levels, within the wavelength
range analyzed by the instrument. Second, water vapor could affect the linewidth of the oxygen lines
by collisional broadening. Indeed, even for a 1 % mixing ratio corresponding to high atmospheric
humidity, water vapor induces a very slightly enhanced pressure broadening compared to major
335 atmospheric components like N_2 or O_2 . As the end, and even if we could correct for the small measured
humidity effect, we preferred measuring $\delta^{18}\text{O}(\text{O}_2)$ and O_2 concentration after drying the gas flow.

Commenté [A56]: Revised (Technical comment
Referee 1)


3.3 Influence of oxygen concentration

340 An influence of O_2 concentration in the sample on the measured $\delta^{18}\text{O}(\text{O}_2)$ was expected as is usual in
all spectroscopic measurements. Indeed, the parameterized models used for spectrum simulations
are associated with uncertainties so that using the theoretical spectroscopic profile does not enable
us to faithfully reproduce the shape of the experimental spectrum. As a consequence, we cannot
precisely subtract the effect of the amplitude variation of the large ^{16}O line (directly linked to the
abundance of O_2) on the neighboring ^{18}O line (see spectrum in Figure 1). This difficulty can be
overcome by quantifying this effect through the study of the influence of O_2 concentration on the
345 measured $\delta^{18}\text{O}(\text{O}_2)$. Figure 4 showed the linearity effect, *i.e.* the evolution of $\delta^{18}\text{O}(\text{O}_2)$ with increasing
 O_2 concentration. Standard dry atmospheric air enriched with O_2 was diluted with nitrogen to measure
6 different O_2 mixing ratios (Figure 4 x-axis). The difference in $\delta^{18}\text{O}(\text{O}_2)$ between diluted standard and
atmospheric air was then used to correct for any instrument drift (Figure 4 y-axis). The influence of O_2
concentration on $\delta^{18}\text{O}(\text{O}_2)$ was clearly significant. $\delta^{18}\text{O}(\text{O}_2)$ increased by 0.53 ‰ with 1 % increase in
350 O_2 concentration (Figure 4).

Commenté [A57]: Revised (Technical comment
Referee 1)

Commenté [A58]: Reworded (Technical comment
Referee 1)

Commenté [A59]: Corrected (Technical comment
Referee 1)



Commenté [A60]: Regression removed from the graph according to referee 1.

Figure 4: Evolution of the difference in $\delta^{18}\text{O}(\text{O}_2)$ between a standard (atmospheric air enriched with O_2) diluted with N_2 and the atmospheric air, vs the O_2 concentration of the diluted standard. $\delta^{18}\text{O}(\text{O}_2)$ is normalized for the ambient O_2 mixing ratio, and error bars are the standard deviation of individual measurements ($n=27$). Dotted line: linear regression ($y=0.5336x-11.178$, $R^2=0.9782$). This result was obtained with the initial configuration (see end of §2.1).

Commenté [A61]: Added (Technical comment Referee 1)

Commenté [A62]: Added (Technical comment Referee 1)

Finally, note that absorption line profiles are influenced by sample parameters such as temperature and pressure. Because of possible long-term drift in pressure and temperature sensors, it is a good practice to regularly check the evolution of this influence of O_2 on the measured $\delta^{18}\text{O}(\text{O}_2)$ (which we did as a precaution once every 15 days).

Commenté [A63]: Modified (Technical comment Referee 1)

Commenté [A64]: Modified (Technical comment Referee 1)


3.4 Memory effect and response time

Commenté [A65]: Modified (Technical comment Referee 1)

The response time of the analyzer and associated inlet system, when switching between different gas streams, was dependent on the total volume of the system (valve, filter, dryer, tubing, and the instrument internal volume) and the flow rate at the inlet. It was determined that 2 minutes purge

time between samples (with a constant flow rate), were necessary to reach a new steady state for both O₂ concentration and $\delta^{18}\text{O}(\text{O}_2)$ (Figure 5).

Commenté [A66]: Modified (Technical comment Referee 1)



370

Figure 5: Raw data of $\delta^{18}\text{O}(\text{O}_2)$ (3 Hz, gray line) and moving average (over 5 second, black line) measured with two gas streams with contrasted O₂ concentration (green line). The response time is outlined in yellow. These results were exactly the same for the two configurations.

Commenté [A67]: Modified (Technical comment Referee 1)

375 3.5 Calibration strategy

Commenté [A68]: This section was improved, in answer to the demand of referee 1

Based on the detailed results above, we used the following strategy to ensure good quality measurements of the isotopic and elemental composition of O₂ in atmospheric air using an OF-CEAS instrument.

380 First, we always place a magnesium perchlorate trap at the entrance of the instrument to work with dry air, hence easily work with standards made of dry air and avoid the dilution effect observed on the O₂ concentration. Then, for calibration of the signal we need to follow different steps. The

dependency of the $\delta^{18}\text{O}(\text{O}_2)$ on O_2 concentration is rather stable and it can be checked regularly by running successive dilution of a standard with N_2 (as a precaution it was done every 15 days). On top of it, a two-point calibration is needed with different values of $\delta^{18}\text{O}(\text{O}_2)$ and O_2 concentration. In our case, we used dry atmospheric air and a standard made of dry atmospheric air enriched with 2 % of O_2 (hence O_2 concentration of 23 % instead of 21 % in dry atmospheric air). Our tests have shown that the measured difference in $\delta^{18}\text{O}(\text{O}_2)$ and in O_2 concentration between the two standards is stable over a 24 hours period. As a consequence, it is enough to run the two standards only once or twice a day and then to measure only one standard with the measurements in the daily routine. If the two-point calibration is rather stable over the course of one day, the drift of the instrument occurs on a much shorter timescale. The evolution of the Allan deviation suggested that, to avoid any drift, a one-point calibration should be done at least every 20 minutes. Moreover, to obtain a small 1σ value ($< 0.1 \text{ ‰}$ as mentioned in section 3.1), averages over 5-10 minutes should be done for each sample injection.

Commenté [A69]: Clarified (technical comment Referee 1)

Finally, in order to account for the instrument response time, it is important to remove the first 2 minutes from the measurement series (purge time) when we switch from one standard to a sample or from one standard to another standard.

Commenté [A70]: Added (technical comment Referee 1)

Commenté [A71]: Added (Technical comment Referee 1)

Gathering the recommendations listed above, we suggest using the following measurement sequence:

Commenté [A72]: Rewritten (technical comment Referee 1)

- Every 15 days: calibration of the influence of O_2 concentration on $\delta^{18}\text{O}(\text{O}_2)$
- Every morning: 6-minutes measurement of atmospheric air enriched with O_2 and 6-minutes measurement of atmospheric air. Only the last 4 minutes of each plateau were kept for calculation of the average values.
- During the day: 6-minutes measurements of dry atmospheric air, 6-minutes measurements of sample 1, 6-minutes measurements of sample 2, 6-minutes measurements of sample 3, back to 6-minutes measurement of dry atmospheric air. Only the last 4 minutes of each plateau were kept for calculation of the average values.

3.6 Accuracy of measurements: comparison with IRMS measurements

In order to validate the performance of this new analyzer, with the calibration sequence proposed in the previous section, we measured the same samples with our analyzer and with an IRMS. We sampled

410 air in 5 mL flasks from a closed biological chamber while the analyzer was measuring the elemental and isotopic composition of O₂ following the sequence described above. We measured the isotopic value of O₂ from the flasks by IRMS (see section 2.3). These values were compared to the values obtained by the analyzer after corrections of the different influences (cf. previous section) during the same period.

Commenté [A73]: Added

415 The results of this comparison are displayed on Figure 6 for configuration 1 (same result was obtained for configuration 2). The values obtained for the 3 biological samples by the two techniques align well on a line of slope 1 with zero y-intercept within uncertainty range for measurements of $\delta^{18}\text{O}(\text{O}_2)$, and with only a slight discrepancy for O₂ concentration. This validates the use of the spectroscopy analyzer for ambient air and biological chambers measurements when the values for concentration of O₂ do not deviate more than a few % from the current concentration of O₂. If the composition of the analyzed air is too different from the current atmospheric air, nonlinear effects on both concentration of O₂ and $\delta^{18}\text{O}(\text{O}_2)$ are expected through modification of the shape of the absorption lines.

Commenté [A74]: Modified (Technical comment Referee 1)





Figure 6: Comparison of the concentration (a) and isotopic composition (b) of O₂ obtained by IRMS and by our spectroscopic analyzer (SARA) for different biological samples. The dashed line shows the 1:1 relationship between $\delta^{18}\text{O}(\text{O}_2)$ measured by IRMS and $\delta^{18}\text{O}(\text{O}_2)$ measured by the spectroscopic analyzer.

4. Conclusion

We developed a new instrument based on OF-CEAS technique enabling continuous measurements of elemental and isotopic composition of O₂ in the atmosphere. We targeted the region of the dioxygen absorption band around 760 nm to capture a signal for ¹⁶O-¹⁷O, ¹⁶O-¹⁶O and ¹⁸O-¹⁸O and developed a special fitting for the ¹⁶O-¹⁶O peak which is saturated.

The instrument has been characterized and we could reach a precision of 0.002 % on O₂ concentration and 50 ppm on $\delta^{18}\text{O}(\text{O}_2)$ for an integration time of 10 minutes. The drift of the instrument deteriorates the signal but the values for Allan deviation stay below 0.2 ‰ for $\delta^{18}\text{O}(\text{O}_2)$ and 0.03 % for O₂ concentration after 2 hours of continuous measurement. The measurements are not affected by humidity in the atmosphere but O₂ concentration has an influence on $\delta^{18}\text{O}(\text{O}_2)$ which should hence be quantified.

Based on the performances of the instrument, we proposed a procedure for running O₂ measurements for sample or in a continuous way based on a frequent (every 20 minutes) injection of calibration standard. This procedure permitted to obtain performance in good agreement with dual inlet IRMS measurements in a shorter time. It was particularly adapted for monitoring biological processes in a continuous way.

In the future, this instrument may be used in several set-ups where a continuous measurement of O₂ concentration and $\delta^{18}\text{O}(\text{O}_2)$ is needed, such as measurements during biological experiments to follow respiration and photosynthesis evolution, medical set-ups following the evolution of respiration or continuous measurements of the elemental and isotopic composition of fossil air in ice cores. The instrument was also able to measure the $\delta^{17}\text{O}(\text{O}_2)$ but the current precision was not sufficient yet for competing with IRMS measurements.

Data availability

The data supporting the conclusions of this paper is available upon request.

Commenté [A75]: Added (Technical comment Referee 1)

Author contributions

AL and CPi designed the project. DR designed the optical spectrometer, with assistance and expertise from MF and KJ, and most of the hardware was purchased to KJ from AP2E. MF conducted the initial
460 tests and tuning at LiPhy. Final tuning was performed at Liphy by DR and JC. CPi, JS, CPa carried out the experiments at ECOTRON of Montpellier and FP, NB and TL at LSCE. CPi, JS, CPa, NB and AL analyzed the data from the optical spectrometer and CPa and AL analyzed the data from IRMS. CPi, DR, JS and AL prepared the manuscript with contributions from all the authors.

Competing interests

465 The authors declare that they have no conflict of interest.

Acknowledgements

The research leading to these results has received funding from the European Research Council under
470 the European Union H2020 Programme (H2020/20192024)/ERC grant agreement no. 817493 (ERC ICORDA). The authors acknowledge the scientific and technical support of PANOPLY (Plateforme ANalytique géOsciences Paris-sacLaY), Paris-Saclay University, France. This study benefited from the CNRS resources allocated to the French ECOTRONS Research Infrastructure, from the Occitanie Region and FEDER investments as well as from the state allocation 'Investissement d'Avenir' AnaEE- France
475 ANR-11-INBS-0001. We would also like to thank Sébastien Devidal, Abdelaziz Faez, Olivier Ravel and Alex Milcu from ECOTRON of Montpellier for their help.

Bibliography

- Barkan, E. and Luz, B.: High-precision measurements of $^{17}\text{O}/^{16}\text{O}$ and $^{18}\text{O}/^{16}\text{O}$ of O_2 and O_2/Ar ratio in air, *Rapid Commun. Mass Spectrom.*, 17, 2809–2814, <https://doi.org/10.1002/rcm.1267>, 2003.
- 480 Bender, M., Sowers, T., and Labeyrie, L.: The Dole Effect and its variations during the last 130,000 years as measured in the Vostok Ice Core, *Glob. Biogeochem. Cycles*, 8, 363–376, <https://doi.org/10.1029/94GB00724>, 1994.
- 485 Berhanu, T. A., Hoffnagle, J., Rella, C., Kimhak, D., Nyfeler, P., and Leuenberger, M.: High-precision atmospheric oxygen measurement comparisons between a newly built CRDS analyzer and existing measurement techniques, *Atmospheric Meas. Tech.*, 12, 6803–6826, <https://doi.org/10.5194/amt-12-6803-2019>, 2019.

- Blunier, T., Barnett, B., Bender, M. L., and Hendricks, M. B.: Biological oxygen productivity during the last 60,000 years from triple oxygen isotope measurements, *Glob. Biogeochem. Cycles*, 16, <https://doi.org/10.1029/2001GB001460>, 2002.
- 490 Brandon, M., Landais, A., Duchamp-Alphonse, S., Favre, V., Schmitz, L., Abrial, H., Prié, F., Extier, T., and Blunier, T.: Exceptionally high biosphere productivity at the beginning of Marine Isotopic Stage 11, *Nat. Commun.*, 11, 2112, <https://doi.org/10.1038/s41467-020-15739-2>, 2020.
- Extier, T., Landais, A., Bréant, C., Prié, F., Bazin, L., Dreyfus, G., Roche, D. M., and Leuenberger, M.: On the use of $\delta^{18}\text{O}_{\text{atm}}$ for ice core dating, *Quat. Sci. Rev.*, 185, 244–257, <https://doi.org/10.1016/j.quascirev.2018.02.008>, 2018.
- 495 Faïn, X., Chappellaz, J., Rhodes, R. H., Stowasser, C., Blunier, T., McConnell, J. R., Brook, E. J., Preunkert, S., Legrand, M., Debois, T., and Romanini, D.: High resolution measurements of carbon monoxide along a late Holocene Greenland ice core: evidence for in situ production, *Clim. Past*, 10, 987–1000, <https://doi.org/10.5194/cp-10-987-2014>, 2014.
- 500 Goto, D., Morimoto, S., Ishidoya, S., Ogi, A., Aoki, S., and Nakazawa, T.: Development of a High Precision Continuous Measurement System for the Atmospheric O_2/N_2 Ratio and Its Application at Aobayama, Sendai, Japan, *J. Meteorol. Soc. Jpn. Ser. II*, 91, 179–192, <https://doi.org/10.2151/jmsj.2013-206>, 2013.
- 505 Goto, D., Morimoto, S., Aoki, S., Patra, P. K., and Nakazawa, T.: Seasonal and short-term variations in atmospheric potential oxygen at Ny-Ålesund, Svalbard, *Tellus B Chem. Phys. Meteorol.*, 69, 1311767, <https://doi.org/10.1080/16000889.2017.1311767>, 2017.
- Guy, R. D., Fogel, M. L., and Berry, J. A.: Photosynthetic Fractionation of the Stable Isotopes of Oxygen and Carbon, *Plant Physiol.*, 101, 37–47, <https://doi.org/10.1104/pp.101.1.37>, 1993.
- 510 Helman, Y., Barkan, E., Eisenstadt, D., Luz, B., and Kaplan, A.: Fractionation of the Three Stable Oxygen Isotopes by Oxygen-Producing and Oxygen-Consuming Reactions in Photosynthetic Organisms, *Plant Physiol.*, 138, 2292–2298, <https://doi.org/10.1104/pp.105.063768>, 2005.
- Jurikova, H., Abe, O., Shiah, F.-K., and Liang, M.-C.: New constraints on biological production and mixing processes in the South China Sea from triple isotope composition of dissolved oxygen, *Biogeosciences*, 19, 2043–2058, <https://doi.org/10.5194/bg-19-2043-2022>, 2022.
- 515 Kassi, S., Chenevier, M., Gianfrani, L., Salhi, A., Rouillard, Y., Ouvrard, A., and Romanini, D.: Looking into the volcano with a Mid-IR DFB diode laser and Cavity Enhanced Absorption Spectroscopy, *Opt. Express*, 14, 11442, <https://doi.org/10.1364/OE.14.011442>, 2006.
- Keeling, R. F. and Manning, A. C.: Studies of Recent Changes in Atmospheric O_2 Content, in: *Treatise on Geochemistry*, Elsevier, 385–404, <https://doi.org/10.1016/B978-0-08-095975-7.00420-4>, 2014.
- 520 Keeling, R. F. and Shertz, S. R.: Seasonal and interannual variations in atmospheric oxygen and implications for the global carbon cycle, *Nature*, 358, 723–727, <https://doi.org/10.1038/358723a0>, 1992.
- Kerstel, E. and Gianfrani, L.: Advances in laser-based isotope ratio measurements: selected applications, *Appl. Phys. B*, 92, 439–449, <https://doi.org/10.1007/s00340-008-3128-x>, 2008.

- 525 Landais, A., Dreyfus, G., Capron, E., Masson-Delmotte, V., Sanchez-Gofii, M. F., Desprat, S., Hoffmann, G., Jouzel, J., Leuenberger, M., and Johnsen, S.: What drives the millennial and orbital variations of $\delta^{18}\text{O}_{\text{atm}}$?, *Quat. Sci. Rev.*, 29, 235–246, <https://doi.org/10.1016/j.quascirev.2009.07.005>, 2010.
- Landsberg, J., Romanini, D., and Kerstel, E.: Very high finesse optical-feedback cavity-enhanced absorption spectrometer for low concentration water vapor isotope analyses, *Opt Lett*, 39, 1795–1798, <https://doi.org/10.1364/OL.39.001795>, 2014.
- 530 Lechevallier, L., Grilli, R., Kerstel, E., Romanini, D., and Chappellaz, J.: Simultaneous detection of C_2H_6 , CH_4 , and $\delta^{13}\text{C}\text{-CH}_4$ using optical feedback cavity-enhanced absorption spectroscopy in the mid-infrared region: towards application for dissolved gas measurements, *Atmospheric Meas. Tech.*, 12, 3101–3109, <https://doi.org/10.5194/amt-12-3101-2019>, 2019.
- 535 Leroy-Dos Santos, C., Casado, M., Prié, F., Jossoud, O., Kerstel, E., Farradèche, M., Kassi, S., Fourré, E., and Landais, A.: A dedicated robust instrument for water vapor generation at low humidity for use with a laser water isotope analyzer in cold and dry polar regions, *Atmospheric Meas. Tech.*, 14, 2907–2918, <https://doi.org/10.5194/amt-14-2907-2021>, 2021.
- 540 Luz, B. and Barkan, E.: Assessment of Oceanic Productivity with the Triple-Isotope Composition of Dissolved Oxygen, *Science*, 288, 2028–2031, <https://doi.org/10.1126/science.288.5473.2028>, 2000.
- Luz, B. and Barkan, E.: The isotopic composition of atmospheric oxygen, *Glob. Biogeochem. Cycles*, 25, 1–14, <https://doi.org/10.1029/2010GB003883>, 2011.
- 545 Luz, B., Barkan, E., Bender, M. L., Thiemens, M. H., and Boering, K. A.: Triple-isotope composition of atmospheric oxygen as a tracer of biosphere productivity, *Nature*, 400, 547–550, <https://doi.org/10.1038/22987>, 1999.
- Morville, J., Kassi, S., Chenevier, M., and Romanini, D.: Fast, low-noise, mode-by-mode, cavity-enhanced absorption spectroscopy by diode-laser self-locking, *Appl. Phys. B*, 80, 1027–1038, <https://doi.org/10.1007/s00340-005-1828-z>, 2005.
- 550 Morville, J., Romanini, D., and Kerstel, E.: Cavity Enhanced Absorption Spectroscopy with Optical Feedback, in: *Cavity-Enhanced Spectroscopy and Sensing*, vol. 179, edited by: Gagliardi, G. and Looock, H.-P., Springer Berlin Heidelberg, Berlin, Heidelberg, 163–209, https://doi.org/10.1007/978-3-642-40003-2_5, 2014.
- 555 Paul, C., Piel, C., Sauze, J., Pasquier, N., Prié, F., Devidal, S., Jacob, R., Dapoigny, A., Jossoud, O., Milcu, A., and Landais, A.: Determination of respiration and photosynthesis fractionation factors for atmospheric dioxygen inferred from a vegetation–soil–atmosphere analogue of the terrestrial biosphere in closed chambers, *Biogeosciences*, 20, 1047–1062, <https://doi.org/10.5194/bg-20-1047-2023>, 2023.
- 560 Paul, C., Piel, C., Sauze, J., Jossoud, O., Dapoigny, A., Romanini, D., Prié, F., Devidal, S., Jacob, R., Milcu, A., and Landais, A.: A multiplexing system for quantifying oxygen fractionation factors in closed chambers, <https://doi.org/10.5194/egusphere-2024-1755>, In review. 17 June 2024.
- Severinghaus, J. P., Beaudette, R., Headly, M. A., Taylor, K., and Brook, E. J.: Oxygen-18 of O_2 Records the Impact of Abrupt Climate Change on the Terrestrial Biosphere, *Science*, 324, 1431–1434, <https://doi.org/10.1126/science.1169473>, 2009.

- 565 Stanley, R. H. R., Kirkpatrick, J. B., Cassar, N., Barnett, B. A., and Bender, M. L.: Net community production and gross primary production rates in the western equatorial Pacific, *Glob. Biogeochem. Cycles*, 24, 2009GB003651, <https://doi.org/10.1029/2009GB003651>, 2010.
- Stephens, B. B., Keeling, R. F., Heimann, M., Six, K. D., Murnane, R., and Caldeira, K.: Testing global ocean carbon cycle models using measurements of atmospheric O₂ and CO₂ concentration, *Glob. Biogeochem. Cycles*, 12, 213–230, <https://doi.org/10.1029/97GB03500>, 1998.
- 570 Stolper, D. A., Bender, M. L., Dreyfus, G. B., Yan, Y., and Higgins, J. A.: A Pleistocene ice core record of atmospheric O₂ concentrations, *Science*, 353, 1427–1430, <https://doi.org/10.1126/science.aaf5445>, 2016.
- 575 Stolper, D. A., Fischer, W. W., and Bender, M. L.: Effects of temperature and carbon source on the isotopic fractionations associated with O₂ respiration for 17O/16O and 18O/16O ratios in *E. coli*, *Geochim. Cosmochim. Acta*, 240, 152–172, <https://doi.org/10.1016/j.gca.2018.07.039>, 2018.
- Tohjima, Y.: Method for measuring changes in the atmospheric O₂/N₂ ratio by a gas chromatograph equipped with a thermal conductivity detector, *J. Geophys. Res. Atmospheres*, 105, 14575–14584, <https://doi.org/10.1029/2000JD900057>, 2000.
- 580 Ventrillard-Courtillet, I., Gonthiez, T., Clerici, C., and Romanini, D.: Multispecies breath analysis faster than a single respiratory cycle by optical-feedback cavity-enhanced absorption spectroscopy, *J. Biomed. Opt.*, 14, 064026, <https://doi.org/10.1117/1.3269677>, 2009.
- Yang, J.-W., Brandon, M., Landais, A., Duchamp-Alphonse, S., Blunier, T., Prié, F., and Extier, T.: Global biosphere primary productivity changes during the past eight glacial cycles, *Science*, 375, 1145–1151, <https://doi.org/10.1126/science.abj8826>, 2022.
- 585