

1 **A high-accuracy dynamic dilution method for generating reference gas**
2 **mixtures of carbonyl sulfide at sub-nanomole-per-mole levels for long-term**
3 **atmospheric observation**

4

5 Hideki Nara, Takuya Saito, Taku Umezawa, Yasunori Tohjima

6

7 National Institute for Environmental Studies, 16-2 Onogawa, Tsukuba, Ibaraki 305-8506, Japan

8

9 *Correspondence to:* Hideki Nara (nara.hideki@nies.go.jp)

10

1 **Abstract.** Atmospheric carbonyl sulfide (COS) has received increasing attention as a potential tracer
2 for investigating the global carbon cycle. Owing to the irreversible photosynthetic absorption of COS,
3 changes in the atmospheric COS mole fraction can be related to terrestrial gross primary production.
4 However, the instability of COS in high-pressure cylinders has hampered the accurate determination of
5 atmospheric COS. Here, we report a dynamic dilution method for generating reference gas mixtures
6 containing COS at ambient levels (ca. 500 pmol mol⁻¹). Our method combined a dynamic dilution
7 system employing a high-accuracy mass flow measurement system and a dry reference gas mixture
8 prepared gravimetrically as parent gas mixture containing a micromole-per-mole level of COS filled in
9 a high-pressure aluminium cylinder. The storage stability of COS at this level was experimentally
10 validated for three gravimetrically prepared dry reference gas mixtures at least 9 years after their
11 preparation. We evaluated the dilution performance of the developed method using a gravimetric parent
12 gas mixture containing approximately 1 μmol mol⁻¹ of COS and chlorodifluoromethane (HCFC-22).
13 Excellent repeatability (0.2% for COS and 0.4% for HCFC-22 in terms of relative standard deviation),
14 reproducibility (COS: 0.1%, HCFC-22: 0.3%), and dilution linearity ($R^2 > 0.99$, for both COS and
15 HCFC-22) were obtained, which were corroborated by the nearly constant ratio of the normalised gas
16 chromatography–mass spectrometry (GC/MS) response of COS to HCFC-22. The dilution accuracy was
17 examined by comparing the determined HCFC-22 mole fractions in a dynamically diluted parent gas
18 mixture from a mass flow rate measurement system and GC/MS calibrated using a gravimetrically
19 diluted parent gas mixture. The mole fractions of HCFC-22 from these two methods agreed within an
20 acceptable difference of approximately 2 pmol mol⁻¹, validating the dilution accuracy of the developed
21 method. By re-evaluating the experimental data, we determined the mole fractions of COS and HCFC-
22 22 in an ambient-air-based reference gas mixture, with relative standard deviations of 0.1% for COS
23 and 0.3% for HCFC-22. These results demonstrated that the developed method can accurately generate
24 reference gas mixtures containing COS at ambient levels, which we expect will support long-term
25 observations of atmospheric COS.

删除: e

删除: 3

删除: 3

删除: 04

删除: 28

删除:

删除: 3

删除: 4

1 1. Introduction

2 Carbonyl sulfide (COS) is the most abundant sulphur-containing compound in the atmosphere. It acts
3 as a sulphur carrier from the troposphere to the stratosphere, contributing to the distribution of
4 stratospheric sulphate aerosols that influence Earth's radiative balance (Chin and Davis, 1995;
5 Kjellstrom, 1998; Brühl et al., 2012). Atmospheric COS has recently received increasing attention as a
6 potential tracer for examining the global terrestrial carbon cycle (Montzka et al., 2007). Unlike carbon
7 dioxide, COS is irreversibly taken up by terrestrial plants on via photosynthesis (Goldan et al., 1988;
8 Protoschill-Krebs et al., 1996; Sandoval-Soto et al., 2005; Seibt et al., 2010; Stimler et al., 2010, 2012),
9 allowing the estimation of terrestrial gross primary production from the local to the global scale based
10 on atmospheric COS observations (Campbell et al., 2008; Blonquist Jr et al., 2011; Asaf et al., 2013;
11 Commane et al., 2013; Maseyk et al., 2014; Wehr et al., 2016; Yang et al., 2018; Kooijmans et al., 2019).
12 For example, the inter-annual variability in the global annual mean mole fraction of atmospheric COS
13 has been used to estimate the biogeochemical feedback of terrestrial plant ecosystems in response to an
14 increase in atmospheric CO₂ levels (Campbell et al., 2017). The World Meteorological Organization has
15 estimated the inter-annual variability of COS to be only several picomoles per mole in the global
16 background atmosphere over the last two decades (WMO, 2018), highlighting the need for technical
17 methods that allow accurate determination of atmospheric COS levels.

18 Despite promising new directions for estimating global terrestrial gross primary production using
19 COS, atmospheric COS observations are insufficient for understanding its global distribution and budget.
20 One fundamental reason for this is the difficulty in establishing reliable reference gas mixtures
21 containing ambient levels of COS (ca. 500 pmol mol⁻¹). Hall et al. (2014) reported significant
22 differences in the calibration scale of COS among the laboratories in the inter-comparison experiment.
23 This might be attributed to the instability of COS stored in high-pressure cylinders as well as difficulties
24 in accurate COS measurements. Indeed, as demonstrated in this study (see Figure 3 in Section 3), the
25 changes in the mole fraction of COS dry gas mixture stored in the cylinder during storage (hereafter
26 referred to as COS drift) can be greater than the inter-annual variability of COS. To better understand
27 spatial and temporal variations in atmospheric COS, observations on a stable calibration scale that is

删除: due to

1 compatible across multiple laboratories are required. Thus, approaches that address issues related to the
2 storage instability of COS are required.

3 Gravimetric preparation, in which individual gas components are weighed into a cylinder, is the
4 most common approach for producing reference gas mixtures containing long-lived trace gases such as
5 CO₂, CH₄, and N₂O (ISO, 2015). Although this approach is the most accurate, it is time-consuming and
6 generally requires multistep dilutions when preparing a gas mixture at the sub-nanomole-per-mole level.
7 This multistep procedure results in the accumulation of preparation uncertainty from weighing, and in
8 addition, gravimetric dilution into cylinders may introduce uncertainties related to surface effects that
9 can vary by cylinder type.

10 Dynamic dilution is an alternative approach to gravimetric preparation (ISO, 2018). Historically,
11 various dynamic dilution methods have been applied to gas species that exhibit strong adsorption or
12 limited stability at ambient levels in high-pressure cylinders (Goldan et al., 1986; Fried et al., 1990;
13 Wright et al., 1994; Kerwin et al., 1996; Mohamad et al., 1996; Nakao et al., 1999; Goody et al., 2002;
14 Brewer et al., 2011, 2014; Flores et al., 2012; Kim et al., 2016; Guillevic et al., 2018; Macé et al., 2022).
15 A major advantage of dynamic dilution is that the reference gas mixture can be generated from a pristine
16 parent gas mixture immediately before analysis, which eliminates the need for storage and therefore,
17 prevents any possible mole fraction change that could occur over time. However, only a few studies
18 have used dynamic dilution to generate diluted gas mixtures at the sub-nanomole-per-mole level with
19 high accuracy (Kim et al., 2016; Guillevic et al., 2018). Moreover, the application of this method is
20 limited to gas species with a clear stability at high mole fractions in high-pressure cylinders. To the best
21 of our knowledge, there have been no reports on dynamic dilution-based techniques for generating sub-
22 nanomole-per-mole reference gas mixtures of COS.

23 Here, we report a dynamic dilution method for the preparation of reference gas mixtures containing
24 COS at ambient levels, which we expect will support long-term atmospheric COS observations. In
25 Section 2, we present a detailed description of the dynamic dilution system developed in this study,
26 estimate the theoretical dilution uncertainty, and describe the procedure for determining the mole
27 fraction of COS in a generated gas mixture. In Section 3, we examine the long-term stability of a dry

削除:

1 reference gas mixture of COS stored in a high-pressure aluminium cylinder. In Section 4, we evaluate
2 the dilution performance and overall dilution accuracy of our method and provide an example of the
3 mole fraction assignment for COS in an ambient-air-based reference gas mixture. In Section 5, we
4 discuss the possible causes of dilution bias observed for COS in this study. All experiments were
5 designed based on a series of preliminary studies (hereafter referred to as our pilot study), which are
6 detailed in the Supplementary Information.

7

8 **2. Materials and methods**

9 Figure 1 shows the schematic of our dynamic dilution system. Briefly, the system generates the desired
10 gas mixture by blending two gas flows of a gravimetrically prepared parent gas mixture containing a
11 few micromole-per-mole fractions of COS and a COS-free diluent gas. Dilution accuracy is achieved
12 by a high-accuracy flow measurement system that has been used as a dynamic dilution method to
13 generate sub-micromole-per-mole reference gas mixtures and adjust the composition of gas matrices
14 (Brewer et al., 2014; Nara et al., 2012). A detailed description of the main components of the dilution
15 system is provided in Section 2.1; a general explanation of the generation of a dynamically diluted gas
16 mixture is provided in Section 2.2; a theoretical estimation of the dilution uncertainty is discussed in
17 Section 2.3; and the determination of the COS mole fraction in a dynamically generated gas mixture
18 combined with a gas chromatography–mass spectrometry (GC/MS)-based sample determination system
19 (hereafter referred as GC/MS measurement system) is provided in Section 2.4.

20

21 **2.1 Main components of the dilution system**

22 The dilution system consists of two gas cylinders, a high-accuracy flow measurement system, four mass
23 flow controllers (MFCs), a static mixer, and an automatic back-pressure regulator (ABPR). These
24 components are described in detail below.

25

26 **2.1.1 Parent and diluent gases**

27 A dry gas mixture of COS and chlorodifluoromethane (HCFC-22) was used as the parent gas with a


1 high-purity nitrogen balance. The parent gas mixture was prepared gravimetrically by a Japanese gas
2 manufacturer (Japan Fine Products Co., Kanagawa, Japan) and was provided in a Luxfer 9.4-L high-
3 pressure aluminium cylinder which was fitted with a brass bodied SUS316L metal diaphragm valve
4 (G55-L; Hamai Industries Limited, Tokyo, Japan) at a fill pressure of 9.81 MPa (unless otherwise noted,
5 gas pressures are reported on the gauge scale). The inner surface of the cylinder was polished using a
6 proprietary method to reduce adsorption of COS and HCFC-22. The nominal mole fractions of COS
7 and HCFC-22 were 1.01 and 1.00 $\mu\text{mol mol}^{-1}$, respectively. HCFC-22 was added to the parent gas
8 mixture to aid in assessing the presence and extent of the dilution bias (see details in Supplementary
9 Information). HCFC-22 was selected because (1) its atmospheric dry mole fraction is relatively close to
10 that of COS (approximately 260 pmol mol^{-1} in northern hemispheric background air; cf. COS: 400–600
11 pmol mol^{-1}) (WMO, 2022) and (2) it is more stable than COS with respect to adsorption. High-purity
12 nitrogen (>99.99995% purity; Japan Fine Products, Co.) filled in a 47-L high-pressure manganese steel
13 cylinder was used as the diluent. The pressures of the parent and diluent gases supplied from respective
14 cylinders were regulated using single-stage (AEROTRACE; Nissan Tanaka Corp., Saitama, Japan) and
15 two-stage stainless steel pressure regulators (SuperLabo; Nissan Tanaka Corp.), respectively. The former
16 regulator had a SUS316L body with a SUS316L metal diaphragm and the internal gas contact surface
17 was SUS316L and fluorine resin.

18 The presence of COS as a contaminant in the diluent gas was verified before each experiment using
19 a GC/MS measurement system (see Section 2.4). [Repeated measurements of the diluent gas showed that](#)
20 [small amounts of COS \(but not HCFC-22\) were detected initially, but the amounts decreased with the](#)
21 [number of measurements and were finally below the detection limit. This COS decrease was most likely](#)
22 [ascribed to the adsorption/desorption in the GC/MS system, rather than the contamination from the](#)
23 [dilution gas as well as helium gas used as the carrier gas for the GC/MS system.](#) These repeated
24 [measurements of the diluent gas](#) ensured that there was no COS contamination from the diluent or carrier
25 gases, which may have biased the determination of COS and HCFC-22 in our experiment.

削除: Three different COS-free gases filled in high-pressure cylinders were examined: purified air (Japan Fine Products, Co.), in which H₂O and atmospheric trace gases were removed by cryogenic separation (Nara et al., 2012); high-purity nitrogen (>99.99995% purity; Japan Fine Products, Co.); and high-purity helium (>99.99995% purity; Japan Fine Products, Co.). Nitrogen and helium were used as the diluent gas in our dilution system and as the carrier gas for the GC/MS measurement system, respectively.

削除: For the GC/MS measurement of all the three gases, small amounts of COS (but not HCFC-22) were detected, which corresponding to approximately 8 pmol mol^{-1} . However, the comparable COS levels among these cylinder gas measurements suggested that the contamination was likely due to the constant blank of COS in the determination system, rather than in the three gases. It may be due to adsorption/desorption within the GC/MS measurement system....

削除: is blank check



1

2 **Figure 1: Schematic of the dynamic dilution system. The thermal-based mass flow controllers and**
 3 **automatic back-pressure regulator are denoted as MFC and ABPR, respectively. Full-scale**
 4 **volume flow rates (FS) that determine the flow precision are provided for each Molbloc element**
 5 **and thermal-based mass flow controller. The ambient-air-base reference gas mixture was used to**
 6 **calibrate the GC/MS system (see details in Section 2.4).**

削除:

7

8 **2.1.2 High-precision flow measurement system**

9 A Molbloc high-precision flow measurement system (Fluke Co., WA) was used to measure the mass
 10 flow rates of the diluent and parent gases. The system comprised a system controller (Molbox 1+) and
 11 two Molbloc-L laminar mass flow elements (A, model 3E4 VCR-V-Q and B, model 1E1 VCR-V-Q).
 12 The full-scale volume flow rates of the diluent and parent gases were 30 standard liters per minute (slm)
 13 and 10 standard cubic centimetres per minute (sccm), respectively. The Molbloc system accurately
 14 calculates the mass flow rate based on measurements of the temperature and the upstream and

1 downstream pressures of the gas flow passing through the element in a laminar regime according to the
2 Hagen–Poiseuille law (Landau and Lifshitz, 1987). System control and data acquisition were performed
3 using the Tera Term software (version 4.97; Tera Term, 2017). The maximum measurement uncertainty
4 depended on the flow rate and was $\pm 0.2\%$ of the measured value or $\pm 0.02\%$ of the full-scale flow rate
5 at flow rates above and below 10% of the full-scale flow rate. The gas contact area along the entire flow
6 path of the Molbloc elements was SUS316L.

7

8 **2.1.3 Thermal-based mass flow controllers**

9 Our dynamic dilution system used four thermal-based MFCs (A–D). MFC-A and -B (SEC-Z712MGX;
10 HORIBA STEC, Co., Ltd., Kyoto, Japan) were used for automatic flow control of the diluent and parent
11 gases, respectively, with full-scale flow rates of 10 slm and 10 sccm, respectively. MFC-C was the same
12 model as the controller A and B but with a full-scale flow rate of 300 sccm and was used to control the
13 outflow of the dynamically generated gas mixture from the dilution system. The flow precision of these
14 three MFCs was $\pm 1.0\%$ of the set point or $\pm 0.25\%$ of the full-scale flow rate for above and below 25%
15 of their full-scale flow rate. For these MFCs, SUS316L and Co-Ni alloys were used as the wetted part
16 material of which surfaces were mirror-polished to prevent gas adsorption. The fourth controller, MFC-
17 D (SEC-400MK3; HORIBA STEC, Co., Ltd.), was used to control the exhaust of the excess parent gas
18 for dynamic buffering of the parent gas pressure with a precision of 2% for a full-scale flow rate of 100
19 sccm, with a limited operating range of over 5% of the full-scale flow rate. Each MFC was operated
20 individually using a control unit (PD-D20; HORIBA STEC, Co., Ltd.).

削除:

21

22 **2.1.4 Static mixer**

23 A static mixer (G 1/4-18; Noritake, Co., Ltd., Aichi, Japan) was used to facilitate blending of the diluent
24 and parent gas flows. The gas mixer was a 1/4-inch internal diameter SUS316LS stainless steel pipe
25 with a length of 215 mm. The mixing fin assembly consisted of 18 elements and was set within the pipe.
26 The inner surface of the pipe and outer surfaces of the mixing fin assembly were electrically polished
27 to minimise gas adsorption.

1

2 **2.1.5 Automatic back-pressure regulator**

3 An ABPR (UR-Z722M-UC-B; HORIBA STEC, Co., Ltd.) was used to exhaust the excess diluted gas
4 mixture from the system and ensure that the pressure within the system remained constant. The ABPR
5 operated at an absolute operating pressure range of 10–300 kPa and controlled the pressure with a
6 precision of 0.5% for the full-scale operating pressure (corresponding to ± 1500 Pa), resulting in the
7 pressure inside the system being independent of the barometric variations. The ABPR was controlled
8 using the same unit as that used to control the MFCs. The gas contact area along the entire flow path
9 was an all-metal construction made of SUS316L with a mirror polished surface.

10

11 **2.2 General description of the dynamic dilution system**

12 Our dynamic dilution system is similar to that developed by Brewer et al. (2014) but was modified by
13 installing a branch exhaust system to accurately generate gas mixtures containing sub-nanomolar-per-
14 mole levels of COS.

15 When starting the system, we purged it for at least 15 min with the parent and diluent gases supplied
16 from the high-pressure cylinders at the flow rates set in the subsequent dilution experiment. Before each
17 gas dilution, conditioning purge was performed for 15 min (only first gas dilution needed 30 min purge
18 at least). During this 15-min purge, the pressures of the two gases downstream respective Molbloc
19 elements were adjusted to match each other within ± 0.3 kPa based on the pressure readings from the
20 Molbloc elements. Although the mass flow rates of the two gas flows were controlled by MFC-A and -
21 B, using these MFCs alone resulted in insufficient pressure matching between the two flows. Therefore,
22 for more precise pressure control, additional pressure stabilisation techniques were used for each gas
23 flow because their mass flow rates were substantially different; for example, the flow rate of the diluent
24 gas was 2000-times larger than that of the parent gas when generating a diluted gas mixture containing
25 COS at ambient levels of approximately $500 \text{ pmol mol}^{-1}$.

26 The pressure of the diluent gas was controlled using a cylindrical regulator and simple in-line
27 pressure regulator (6600A; Kofloc Corp., Kyoto, Japan). The pressure of the diluent gas was first set at

1 0.24 MPa using the cylinder regulator and was further adjusted using the in-line pressure regulator set
2 at 0.15 MPa to suppress the pressure change in response to regulator cooling owing to adiabatic
3 expansion of the outflowing gas. In contrast, the parent gas was introduced into the flow-mixing
4 chamber along the main flow path at a flow rate usually limited to 10 sccm, with the pressure set by the
5 cylinder regulator at 0.15 MPa. At such a low flow rate, the instability of the cylinder regulator becomes
6 pronounced, leading to pressure fluctuations of several thousand pascals. Because this pressure
7 fluctuation is difficult to remove by passing through a MFC, we installed a branch exhaust system
8 located upstream of the flow-measuring element in the Molbloc system. The system exhausted the parent
9 gas along the branch exhaust flow path via a 70-mL glass buffer tank at a flow rate of more than 10
10 sccm, which was higher than the flow rate of the main flow. This branch exhaust system prevented the
11 dilution biases found in our pilot study (see Supplemental Information: the dilution biases are discussed
12 in detail in Section 5) and dynamically buffered any pressure changes to facilitate the stabilisation of
13 the parent gas pressure, which should match that of the diluent gas during gas dilution. The diluent and
14 parent gas flows were controlled by MFC-A and -B, which were placed in the respective flow paths
15 downstream of the Molbloc elements and converged in a flow-mixing chamber. The flow-mixing
16 chamber was made of a 1/4-inch outer diameter stainless steel T-fitting, in which one port was connected
17 to a bored-through reducer. The diluent gas flowed straight from upstream to downstream, while the
18 parent gas tube (1/16-inch outer diameter) was inserted from the side port of the T-fitting through the
19 reducer and was bent at a right angle in the downstream direction so that the parent gas flow was
20 introduced into the diluent gas flow in the flow direction. After the convergence of the two flows in the
21 chamber, the generated gas mixture was passed through the static mixer to complete gas blending and
22 then into a glass tank (volume: 750 mL) to buffer any pressure changes in the blended gas mixture.

23 Finally, the diluted gas mixture was supplied to the GC/MS measurement system via the MFC-C.
24 Excess flow was exhausted through the ABPR set at an absolute backpressure of 0.15 MPa, which
25 prevented the influence of barometric changes outside the system. The expected mole fractions of COS
26 and HCFC-22 in the generated gas mixture were accurately calculated from the mass flow rates
27 measured using the Molbloc system. To prevent the possible loss of COS by adsorption, a Sulfinert

1 coating (Restek Corporation, PA, USA) was applied to all tubes, fittings, and valves that were exposed
 2 to the parent gas and generated gas mixtures, except the static mixer.

4 2.3 Estimation of uncertainty for the gas dilution

5 The mole fraction of gas x in a dynamically generated gas mixture can be calculated using the mole
 6 fraction of the target gas in the gravimetric parent gas mixture, mass flow rates of the parent and diluent
 7 gases under the assumption of complete mixing of the two gases, and constant mole fraction of gas x in
 8 the gas flow from the parent gas cylinder. The mole fraction of gas x is expressed as follows:

9

$$10 \quad [x]_{diluted} = \frac{[x]_{parent} \cdot f_{parent}}{(f_{parent} + f_{diluent})}, \quad (1)$$

11

12 where $[x]$ indicates the mole fraction of gas x in the gas flow of the dynamically generated gas mixture
 13 (*diluted*) and parent gas (*parent*) and f is the mass flow rate corresponding to the subscripted gas flow.

14 By applying the propagation law to this equation, the uncertainty ($u_{diluted}$) in the mole fraction of the
 15 diluted gas x can be estimated from the uncertainty of each variable associated with the gravimetric
 16 preparation and Molbloc measurement as follows:

17

$$18 \quad u_{diluted} =$$


$$19 \quad \sqrt{\left\{ \frac{1}{(f_{parent} + f_{diluent})^2} \right\}^2 \cdot \left\{ (f_{parent}^2 \cdot \delta[x]_{parent}^2 \cdot (f_{parent} + f_{diluent})^2) + [x]_{parent}^2 \cdot (f_{diluent}^2 \cdot \delta f_{parent}^2 + f_{parent}^2 \cdot \delta f_{diluent}^2) \right\}}$$

$$20 \quad , \quad (2)$$

21 where each δ symbol represents the uncertainty of corresponding variables $[x]_{parent}$, f_{parent} and $f_{diluent}$. The
 22 dilution uncertainty associated only with the Molbloc measurements can be estimated by removing the
 23 uncertainty derived from the gravimetric preparation of the parent gas ($\delta[x]_{parent} = 0$). The mole fraction
 24 distributions of gas x and relevant dilution uncertainties expressed as a relative standard deviation (RSD)
 25 are shown in Figure 2. The dilution system can generate diluted gas mixtures across a wide range of
 26 mole fractions depending on the flow rates of the diluent and parent gases, covering the typical mole

1 fractions of COS in the atmosphere (400–600 pmol mol⁻¹). The dilution uncertainty was determined
 2 using the flow rates and was found to change significantly at flow rates corresponding to 10% full-scale
 3 for the Molbloc elements used. For example, within the typical atmospheric COS mole fraction range,
 4 the uncertainty at parent gas flow rates below 1 sccm, which corresponds to a 10% full-scale flow rate,
 5 increased rapidly from approximately 0.5% with a decreasing flow rate. In contrast, the uncertainty
 6 remained nearly constant at parent gas flow rates above 1 sccm. The parent gas flow rates above 2 sccm
 7 within the typical COS mole fraction range suppressed the dilution uncertainty to below 0.3%.

8



9

10 **Figure 2:** Mole fraction distribution for a dynamically diluted gas produced from a 1 μmol mol⁻¹
 11 parent gas mixture (left) and the corresponding dilution uncertainty (right). The two solid black
 12 lines in the left panel indicate the typical mole fraction range of atmospheric COS (400–600 pmol
 13 mol⁻¹).

14

15 2.4 Determination of COS in a diluted gas mixture

16 To evaluate dilution performance of our method, we measured the mole fractions of COS and HCFC-
 17 22 (hereafter ‘target gases’) in dynamically diluted gas mixtures by using the GC/MS measurement
 18 system. Because the method used for target gas determination [including the basic configuration of the](#)
 19 [measurement system was](#) presented in Saito et al. (2010), only a [brief](#) description is provided here. The
 20 gas mixture generated by the dilution system was introduced via MFC-C into [the GC/MS measurement](#)

削除: details of the measurement system and the

削除: are

削除: general

1 system for dual-stage preconcentration of target gases. The gas mixture passed through a
2 preconcentration trap filled with HayeSep D cooled below $-130\text{ }^{\circ}\text{C}$ at a constant flow rate of 40 sccm
3 for 12 min, while the mass flow rates of the parent and diluent gases were recorded to calculate the mole
4 fractions of the target gases. The preconcentration trap was heated to $-70\text{ }^{\circ}\text{C}$ to release major atmospheric
5 constituents and then up to $100\text{ }^{\circ}\text{C}$ to transfer the trapped gases to a cryofocusing trap filled with
6 HayeSep D cooled at $-85\text{ }^{\circ}\text{C}$. The trap was then heated to $100\text{ }^{\circ}\text{C}$ to inject the trapped gases onto a
7 PoraBOND Q separation column for subsequent MS analysis. The total measurement time for each
8 sample was 1 h.

删除: a

删除: cryogenic

删除: system

删除:

删除: to enrich the target gases in the gas mixture. Concurrently with the enrichment,

删除: concentrated during the initial sample enrichment process...

删除: The cryogenically enriched target gases were thermally desorbed, enriched again, and injected into the GC/MS for analysis. ...

删除:

9 The target gas measurements were referenced against measurements of an ambient air-based
10 reference gas mixture to correct for diurnal variations in the sensitivity of the mass spectrometer. To
11 prepare the reference gas mixture, ambient air was dried to a dew-point temperature of less than $-80\text{ }^{\circ}\text{C}$
12 by passing it through a Nafion Perma Pure dryer (Perma Pure LLC; Toms River, NJ, USA) and a
13 chemical trap packed with phosphorous pentoxide (Sicapent[®]; EMD Millipore, Billerica, MA, USA)
14 and subsequently filled into a 48-L high-pressure aluminium cylinder (Luxfer) fitted with a brass-bodied
15 cylinder valve (G55-L; Hamai Industries Limited). The inner surface of the cylinder was polished and
16 anodized using a proprietary method (Japan Fine Products, Co.). Before filling, the aluminium cylinder
17 was evacuated, and Milli-Q water was added so that the water vapour mole fraction was approximately
18 $500\text{ }\mu\text{mol mol}^{-1}$ at a maximum fill pressure of 14.7 MPa after gas filling to enhance the stability of the
19 trace gases in the compressed air during storage (Montzka et al., 2004; Hall et al., 2014).

删除:

删除: had been

20 The target gases in the reference gas mixture were measured before and after sample gas
21 measurement. The signal responses obtained from the GC/MS measurements were used to estimate the
22 change in GC/MS sensitivity during sample measurements by linear interpolation. The signals from the
23 sample measurements were normalised to the interpolated responses of the reference gas mixture and
24 then multiplied by the known mole fraction of the corresponding target gas in the reference gas mixture
25 as follows:

26

1 $[x]_{sample} = \frac{R_{x,sample}}{R_{x,ref}} \times [x]_{ref},$ (3)

2
3 where the mole fractions of gas x in the sample and the reference gas mixture are indicated in square
4 brackets, and $R_{x,sample}$ and $\hat{R}_{x,ref}$ are the signal responses for gas x in the sample gas mixture and the
5 interpolated response for the reference gas mixture, respectively. The typical repeatability for COS and
6 HCFC-22 determined from repeated measurement of a compressed cylinder gas was RSD of
7 approximately 0.5% during the study period.

8
9 **3 Long-term stability of COS in high-pressure aluminium cylinders**

10 Atmospheric trace gases with dry mole fractions of sub-nanomole-per-mole levels such as COS and
11 halocarbons, can exhibit significant changes during storage in high-pressure cylinders. High-pressure
12 stainless-steel cylinders are generally used for the storage of reference gas mixtures for these trace gases.
13 Previous studies have reported that COS and halocarbons show superior stability, even at ambient mole
14 fractions, in this type of cylinder (Hall et al., 2014; Guillevic et al., 2018). Thus, the use of stainless-
15 steel cylinders can be an effective measure to reduce changes in the mole fractions of COS during
16 storage. However, in this study, we opted to use Luxfer 9.4L aluminium cylinders with a brass-bodied
17 cylinder valve (G55-L; Hamai Industries Limited) because aluminium cylinders with treated inner
18 surfaces are easier for us to obtain than stainless steel cylinders.

19 In the past, we have prepared several dry reference gas mixtures of COS at ambient levels filled in
20 high-pressure aluminium cylinders by the gravimetric method, and almost all of them showed
21 substantial changes in the mole fraction of COS during storage, which hereafter is referred to as ‘drift’.
22 Figure 3 shows an example of drift, in which we observed a substantial decrease in the mole fraction of
23 COS. Although the stability during storage could be enhanced depending on the material of the gas
24 cylinder and inner-surface treatment (Yokohata et al., 1985; Montzka et al., 2004, 2007; Hall et al.,
25 2014), other factors that affect the amount and rate of this drift remain unknown.

削除:




Figure 3: Time-series variation of the COS mole fraction in a dry reference gas mixture balanced with high-purity nitrogen during long-term storage in a non-passivated high-pressure aluminium cylinder.

At a micromole-per-mole level of COS, we found no notable COS drift during the storage in high-pressure aluminium cylinder from a stability check experiment. In the experiment, compressed natural air was measured using the GC/MS measurement system against three different dry reference gas mixtures containing approximately $5 \mu\text{mol mol}^{-1}$ of COS prepared in different years. These reference gas mixtures (balanced in nitrogen) were prepared gravimetrically and filled into Luxfer 9.4L aluminium cylinder with the brass-bodied cylinder valves at a fill pressure of 9.81 MPa by a Japanese manufacturer (Japan Fine Products, Co.) in 2006, 2011, and 2015. The nominal mole fractions were 5.12, 5.02, and $5.05 \mu\text{mol mol}^{-1}$, respectively (Table 1).

For the determination of COS, we used an established automated measurement system combining a commercially available dynamic dilutor (SGGU-7000NS-U6; HORIBA STEC, Co., Ltd.) and the GC/MS measurement system. The GC/MS measurements of the compressed natural air were referenced to those of a gas mixture containing approximately $500 \text{ pmol mol}^{-1}$ of COS, which was dynamically diluted 10,000 times with high-purity nitrogen (>99.99995% purity) from the gravimetrically prepared reference gas mixture using the commercial dilutor. We do not present the details of the dilutor here; only the general features are described. The dilutor employed a two-step dilution method, and the

- 削除: was calibrated using
- 削除: that
- 削除: original
- 削除: a
- 削除: ly available dynamic
- 削除: (SGGU-7000NS-U6; HORIBA STEC, Co., Ltd.)

1 dilution was performed by mixing the parent and diluent gases at gas flow rates regulated by two MFCs.
2 The expected mole fraction from dilution was calculated from the measured flow rates of the two gases.
3 The nominal precision of the dilutions was $\pm 1.0\%$. The COS determination by GC/MS was repeated
4 four to eight times according to the protocol described in Section 2.4, until high repeatability was
5 obtained in the measured values. The COS mole fraction was determined based on the final three
6 measurements.

删除: measurement

删除: stabilised

7 As presented in Table 1, the determined mole fractions of COS can be compared with each other
8 to indirectly evaluate the long-term stability of COS at the micromole-per-mole level in the reference
9 gas mixtures during storage. The mole fractions of COS from each experiment corresponding to
10 different parent gas mixture prepared in different years were determined with standard deviations (1σ)
11 of less than $0.5 \text{ pmol mol}^{-1}$. While the individual average COS values did not agree within the $2\text{-}\sigma$
12 envelope of their corresponding measurement data, they were within 2% of the average values. It is
13 difficult to clarify whether a marked COS drift occurred in the reference gas mixture owing to the
14 unquantified combined uncertainty related to gravimetric preparation and dynamic dilution. However,
15 the observed relationship presented in Table 1 indicates that the nominal COS values provided by the
16 gas manufacturer at the time of preparation were almost consistent for all three cylinders at least at the
17 time of this experiment. These results imply that in our experiments the impact of drift during storage
18 is practically negligible over a period of more than a decade when using a gas mixture containing COS
19 at a micromole-per-mole level. This suggests an alternative approach to maintain the COS calibration
20 scale with a high-accuracy dynamic dilution of the quasi-stable reference gas mixture. Although periodic
21 absolute determination is still required for stability validation, its frequency can be reduced significantly.

删除: shown

删除: COS

删除: one

删除: an

删除: for

删除: ion of

删除: COS

删除: referencing each

删除: T

删除: were

删除: in

删除: ment

删除: with each other

删除: sigma

删除: but agreed

删除: shown

删除: suggests

删除: given

删除: when

删除: ed

删除: , 9 years after the last gas preparation

删除:

22 Note that we used the automated measurement system using the commercial dilutor, which was
23 convenient for detecting significant level of COS drift whereas the dynamic dilution system using the
24 Molbloc system developed in this study required manual dynamic gas dilution and subsequent GC/MS
25 measurements. However, the commercial dilutor needs more time for system conditioning and has
26 inferior dilution accuracy than the developed dilution system. Therefore, dynamic dilution system using

1 [Molbloc system offers significant advantages for the accurate determination of COS in a reference gas](#)
2 [mixture](#)↓

削除: We note that the stability check experiments using the commercial dynamic dilutor gave high precision GC/MS measurements for halocarbons, but not always for COS; similar problem described in Supplemental Information were often observed. Thus, we need to develop a high-accuracy dynamic dilution method for COS determination.

3
4 **Table 1: Mole fractions of COS in compressed natural air determined by GC/MS calibrated**
5 **against three individual dynamically diluted gas mixtures from respective parent reference gas**
6 **mixture prepared in different years.**

Parent reference gas mixture		Determined COS value [†] (n=3)	
Preparation year	Nominal mole fraction ($\mu\text{mol mol}^{-1}$)	Average (pmol mol^{-1})	Standard deviation (1σ)
2006	5.12	270.2	0.1
2011	5.02	274.6	0.4
2015	5.05	272.9	0.5

7 [†]COS determination was performed in March 2023.

8 9 **4 Evaluation of dilution performance**

10 We investigated the dilution performance of the developed dynamic dilution system in terms of precision
11 and accuracy, [comprising](#) repeatability and reproducibility, and trueness and precision, respectively (ISO,
12 2017). The linearity of the developed dynamic dilution system was also investigated. These
13 investigations were conducted based on GC/MS analysis of the target gases, and we report the GC/MS
14 measurements as response values normalised to an ambient-air-based reference gas mixture according
15 to the procedure described in Section 2.4. After validating our dilution method, we assigned mole
16 fractions to the target gases in the ambient-air-based reference gas mixture.

削除: consisting of

削除:

削除:

17 18 **4.1 Repeatability and reproducibility**

19 To evaluate the dilution precision, which consists of repeatability and reproducibility, we conducted a
20 series of experiments in which we generated a diluted parent gas mixture using a dilution system,
21 followed by GC/MS measurements of the target gases. A total of five experiments were conducted, with

1 each experiment consisting of at least four sample measurements. Four of the experiments were
2 performed on different days within a week, and the final experiment was performed approximately two
3 weeks after the previous four experiments. A common dilution ratio was used in these experiments. The
4 expected mole fraction of the diluted target gases was approximately 500 pmol mol⁻¹ for COS and 495
5 pmol mol⁻¹ for HCFC-22, and the flow rates of the parent and diluent gases to be blended were set at
6 2.5 and 5000 sccm, respectively. For each experiment, the average and standard deviation of the
7 response values from the GC/MS measurements were calculated from at least four normalised responses.

8 Figure 4 shows the results of the GC/MS measurements of the diluted target gases. Dilution repeatability
9 was calculated as the RSD of the normalised responses obtained in each experiment. The obtained RSD
10 values for COS and HCFC-22 from each of the five experiments were within 0.3% and 0.6%,
11 respectively, and the average RSD values were 0.2% and 0.4%, respectively. Although these RSD values
12 of the dilution repeatability were determined as the combined uncertainties from the GC/MS
13 measurements and the dynamic dilution, the average RSD values were comparable with those of the
14 typical measurement repeatability of the GC/MS system (approximately 0.5%). These results suggest
15 that the dilution repeatability obtained was primarily subject to the GC/MS measurements rather than
16 from the dynamic dilution. Conversely, the dilution reproducibility can be evaluated through the
17 standard deviation of the average normalised responses obtained from each experiment. The averages
18 for both COS and HCFC-22 from the five experiments were in good agreement. The RSD values for
19 these responses were 0.1 and 0.3% for COS and HCFC-22, respectively, which were reduced compared
20 with those from the repeatability test. These results suggest that the uncertainty in target gas
21 measurements was dominated by random uncertainty from the GC/MS measurements and that there was
22 no significant systematic uncertainty was observed between the experiments. This is also corroborated
23 by the ratio of the normalised response of COS to that of HCFC-22. We observed no clear systematic
24 changes in the normalised response ratios in each experiment and obtained nearly constant average ratios
25 among the five experiments. Based on the evaluation of the repeatability and reproducibility, the dilution
26 system was found to precisely generate a dynamically diluted gas mixture precisely, which can be
27 represented by the repeatability of the GC/MS measurements.

删除: The d

删除: 4

删除: 2

删除: 3

删除: 3

删除: a

删除: s

删除: to

删除:

删除: On the other hand

删除: from

删除: 04%

删除: for COS


删除: 28

删除: to

删除: the

删除: can

1



2


3 **Figure 4: Repeatability and reproducibility of our dynamic dilution method for generating gas**
4 **mixtures containing COS and HCFC-22. Vertical error bars indicate the standard deviation of the**
5 **normalised values obtained in each experiment. The number of measurements and relative**
6 **standard deviation are shown for each average plot.**

7

8 4.2 Dilution linearity

9 The dilution linearity was examined by comparing the measurements of the target gases in the generated
10 gas mixture determined by the Molbloc system with those obtained by the GC/MS measurement
11 determination system using GC/MS. The mole fractions of the target gases in the generated gas mixture
12 were calculated from the mass flow rates of the parent and diluent gases measured using the Molbloc
13 system, whereas the target gases were determined as normalised GC/MS responses based on GC/MS
14 measurements. For the experiment, the dilution ratio was varied by changing the parent gas flow rate
15 using MFC-B such that the expected mole fraction of COS in the generated gas mixture was in the range
16 of 400–600 pmol mol⁻¹, which corresponds to the typical atmospheric mole fraction. We checked the
17 linear response of COS and HCFC-22 in increments of 50 pmol mol⁻¹ in the range defined by the dilution
18 ratio, and then we checked in decrements of 50 pmol mol⁻¹ to investigate the possible memory effect of
19 the previous dilution process. The normalised GC/MS responses were plotted against the mole fractions
20 obtained from the Molbloc system (Figure 5). For the increment and decrement changes, the normalised

1 COS and HCFC-22 values showed excellent linearity with the corresponding mole fractions from the
2 Molbloc system, and linear least-squares fitting yielded determination coefficients greater than 0.99.
3 The relative residuals of the normalised values from the fitting line were also plotted (bottom panels).
4 The values for COS and HCFC-22 were distributed within 0.4%, which can be explained by the typical
5 measurement repeatability (approximately 0.5%) of the GC/MS measurement system. There was no
6 clear dependence of the relative residual values on the Molbloc mole fraction, and similar results were
7 obtained for the ratio of the normalised values (data not shown). These results indicated that there was
8 no significant memory effect from the previous blending process. Overall, our results demonstrate that
9 our dynamic dilution system has excellent linearity in the range of the investigated dilution ratios and
10 shows no bias depending on the mole fraction of the generated target gas.



11
12 **Figure 5: Dilution linearity of the dynamic dilution method for COS and HCFC-22 in response to**
13 **increases or decreases in the generated target gas mole fractions. The normalised GC/MS**
14 **responses (upper) and residuals from the fitting line determined by linear least squares (bottom)**
15 **are shown as a function of the mole fraction calculated from the average mass flow rates during**
16 **the sample concentration process measured by the Molbloc system.**

17
18 **4.3 Validation of dilution accuracy**

1 Subsequently, the dilution accuracy, comprising trueness and precision, of the dynamic dilution method
2 was assessed using a validation experiment. In this experiment, the parent gas mixture was dynamically
3 diluted using the developed dynamic dilution system, and the resulting diluted parent gas mixture was
4 analysed through GC/MS calibrated with the gravimetrically diluted parent gas mixture. The determined
5 mole fractions of HCFC-22 from the GC/MS measurements were compared with those from the
6 Molbloc system based on the measurements of the mass flow rates of the parent gas mixture and diluent
7 gas (Figure 6). The diluted parent gas mixture for GC/MS calibration was prepared by the Japanese gas
8 manufacturer (Japan Fine Products, Co.) by the gravimetric dilution of the parent gas mixture with high-
9 purity nitrogen (>99.99995% purity). This mixture, was filled at a fill pressure of 9.4 MPa into a Luxfer
10 9.4-L high-pressure aluminium cylinder fitted with a brass bodied SUS316L metal diaphragm valve
11 (G55-L; Hamai Industries Limited). The inner surface of the cylinder was, polished using a proprietary
12 method (Japan Fine Products, Co.). The mole fractions of COS and HCFC-22 in the gas mixture were
13 505.0 and 500.4 pmol mol⁻¹, respectively. Owing to this gravimetric dilution, the GC/MS measurements
14 were standardised to those of the parent gas mixture. We expected a significant loss of COS in this
15 gravimetric dry gas preparation process, and a marked decrease in COS was observed (reduced to
16 approximately 1/13 of the initial mole fraction of COS at the time of filling). However, we could analyse
17 the dilution accuracy based on the determination of HCFC-22 as, our experiments demonstrated
18 comparable dilution performances between HCFC-22 and COS, with no significant systematic dilution
19 bias for COS, as presented in the Supplemental Information.

20 The dynamically diluted parent gas mixture was generated by a 2020-fold dilution to produce an
21 expected HCFC-22 mole fraction of approximately 495 pmol mol⁻¹. The generated gas mixture was
22 measured six times in series by GC/MS, and the average HCFC-22 mole fraction in the diluted gas
23 mixture during the preconcentration process of the GC/MS measurement was determined using the
24 Molbloc system (Table 2). We obtained average HCFC-22 values with high precision using both
25 methods. The standard deviations for the GC/MS and Molbloc measurements were 0.8 and 0.1 pmol
26 mol⁻¹ and the corresponding RSD values were 0.2 and 0.1%, respectively. There was a slight difference
27 between the two average values; the GC/MS measurements were approximately 2 pmol mol⁻¹ lower

削除: Next

削除: consists of

削除: e

削除: using

削除: and

削除: had been

削除: 4

削除: 96

削除: 37

削除: COS

削除: determination because

削除: shown

削除: i

削除: 77

削除: 2

削除: 16

削除: 02

1 than the Molbloc measurements. The average values fell outside the 2- σ range of one another, indicating
2 a potential difference in their trueness.

删除: sigma

删除: suggesting that there may be

删除: between

3 Based on the mass flow rates measured by the Molbloc system, we calculated the average mole
4 fraction of HCFC-22 mole fraction in the diluted gas mixture during the preconcentration process of the
5 GC/MS measurements, which was directly traceable to the nominal HCFC-22 value contained in the
6 parent gas mixture. Conversely, the GC/MS was calibrated with the gravimetrically diluted parent gas
7 mixture with traceability to the parent gas mixture to determine the HCFC-22 value. However, this
8 method includes the uncertainty derived from the gravimetric dilution. Although no estimated
9 uncertainty was provided by the manufacturer, the preparation uncertainty can be roughly estimated
10 based on the typical measurement uncertainty of a weighing scale (± 25 mg, 2σ) and was ± 1.2 pmol
11 mol^{-1} for HCFC-22. Since this uncertainty is propagated to the GC/MS measurement, the observed
12 difference of approximately 2 pmol mol^{-1} was acceptable within the expanded uncertainty of 2.8 pmol
13 mol^{-1} ($k=2$) for the GC/MS measurements calculated from the combined uncertainty of the GC/MS
14 measurements and gravimetric preparation. Based on these results, we concluded that dynamic dilution
15 method could accurately generate COS reference gas mixtures at ambient levels from those at
16 micromolar-per-mole levels.

删除: ere

删除: the average mole fraction

删除: On the other hand,


删除: but will

删除: 15

删除: 7

删除: was capable of

删除: ing



1
2 **Figure 6: Scheme for validating the dilution accuracy of our dynamic dilution method.**

3
4 **Table 2: Comparison of mole fractions of HCFC-22 between a dynamically diluted parent gas**
5 **mixture determined by the Molbloc system and GC/MS calibrated with a gravimetrically diluted**
6 **parent gas mixture.**

Determination method	HCFC-22 (pmol mol ⁻¹) (1σ, n=6)
GC/MS	493.7 ± 0.8
Molbloc	495.8 ± 0.1
Difference (GC/MS – Molbloc)	-2.1 ± 0.8

削除: HCFC-22
削除: in

削除: 1
削除: 77
削除: 77
削除: 2
削除: 06
削除: 78

7
8 **4.4 Mole fraction assignment of COS and HCFC-22 in a high-pressure aluminium cylinder**

9 Here, we present an example of mole fraction assignment for target gases in a high-pressure aluminium
10 cylinder using dynamic dilution and the GC/MS measurement systems. By re-evaluating the

1 experimental data described in Section 4.1, we determined the mole fractions of COS and HCFC-22 in
2 the ambient-air-based reference gas mixture, which was used to correct for diurnal variations in the
3 sensitivity of GC/MS. As described in Section 2.4, the mole fraction of gas x was expressed as the
4 product of the normalised GC/MS response and the assigned mole fraction of gas x in a reference gas
5 mixture according to equation (3). Rearranging equation (3) using the mole fraction determined from
6 the Molbloc measurements yields

$$8 \quad [x]_{ref} = \frac{R_{x,ref}}{R_{x,diluted}} \times [x]_{diluted,Molbloc}, \quad (4)$$

9
10 where $R_{x,diluted}$ is the normalised GC/MS response for gas x in the dynamically diluted gas mixture and
11 $[x]_{diluted,Molbloc}$ is the average mole fraction calculated from the mass flow rates of the parent and diluent
12 gases measured by the Molbloc system during sample concentration. The assigned mole fraction values
13 for the target gases in the ambient-air-based reference gas mixture are summarised in Table 3. The
14 assigned values of COS and HCFC-22 were 554.5 ± 0.1 and 289.1 ± 0.3 pmol mol⁻¹, respectively. These
15 results suggest that the determination of COS and HCFC-22 contained in an ambient-air-based reference
16 gas mixture used to calibrate the GC/MS was performed with an accuracy that enables analysis of the
17 spatio-temporal variability of the target gases, in particular, the interannual variability of COS. However,
18 the obtained values were determined against a gravimetrically prepared parent gas mixture. Therefore,
19 the absolute determination of the target gases in the parent gas mixture is required to establish our COS
20 calibration scale for atmospheric COS observations. The major advantage of the developed dynamic
21 dilution method is that COS values determined by our dynamic dilution method can be recalculated on
22 the updated COS calibration scale after the absolute determination, as well as on the scale from other
23 laboratories through an intercomparison experiment, as long as COS in the parent gas mixture is stable.

24
25 **Table 3: Mole fraction assignments of COS and HCFC-22 contained in an ambient-air-based**
26 **reference gas mixture based on GC/MS analysis, calibrated with a Molbloc dynamic dilution**

删除:

删除:

删除: 1

删除: 3

删除: 1

删除: 4

删除:

删除: allows

删除: obtained

删除: It is great

删除: and even

删除:

1 system using a parent gas mixture containing 1.01 and 1.00 $\mu\text{mol mol}^{-1}$ of COS and HCFC-22.

Experiment	COS		HCFC-22	
	(pmol mol ⁻¹)		(pmol mol ⁻¹)	
	Mole fraction	Standard deviation	Mole fraction	Standard deviation
1	554.7 _v	1.2 _v	288.9 _v	1.5 _v
2	554.7 _v	1.0 _v	288.6 _v	0.5 _v
3	554.7 _v	2.0 _v	288.5 _v	1.0 _v
4	554.5 _v	1.9 _v	290.4 _v	1.8 _v
5	554.0 _v	1.2 _v	289.3 _v	1.5 _v
Assigned value[†]	554.5 _v	0.1 _v	289.1 _v	0.3 _v

2 [†]Uncertainty is given as the standard deviation of the determined mole fractions from five experimental
3 data.

5 Discussion

6 We developed a dynamic dilution method for preparing reference gas mixtures containing COS that
7 includes a relative drift-reduction approach to avoid issues related to the change in the COS mole
8 fraction over time; however, outstanding issues still exist that may lead to reduced consumption of the
9 parent gas mixture in our dynamic dilution method.

10 In our pilot study, we found an unexplained dilution bias responsible for the inconsistent
11 determination of the target gases in the dynamically generated gas mixture between the GC/MS and
12 Molbloc measurements (see Supplemental Information). The dilution bias depended on the pressure and
13 flow rates of the parent and diluent gases; even under pressure control of the relevant gas flows, a greater
14 bias was observed for COS than HCFC-22. We attributed this COS-specific dilution bias to factors
15 possibly associated with the adsorption of COS onto the inner surface of the dilution system or with gas
16 fractionation processes occurring at the cylinder and regulator during the passage of the parent gas flow
17 through them. Therefore, we improved our dilution method by increasing the system purge time,

- 削除: 68
- 削除: 0
- 削除: 85
- 削除: 3
- 削除: 65
- 削除: 3
- 削除: 1
- 削除: 3
- 削除: 2
- 削除: 1
- 削除: 1
- 削除: 2
- 削除: 2
- 削除: 87
- 削除: 37
- 削除: 1
- 削除: 0
- 削除: 3
- 削除: 25
- 削除: 48
- 削除: 1
- 削除: 3
- 削除: 1
- 削除: 4

1 introducing a more rigorous pressure matching between the two gas flows, and increasing the parent gas
2 flow rate beyond that required for generation of reference gas mixture, with excess flow exhausted via
3 a branch exhaust system. Consequently, the combination of these strategies worked well for high-
4 accuracy dynamic dilution in this study, although COS-specific bias is not clearly understood.

5 A similar dilution bias has been observed by Brewer et al. (2014). They reported that the
6 dynamically generated mole fractions of SO₂ and CO showed a systematic bias that was inversely
7 proportional to the parent gas flow rate at flow rates of less than 10 mL/min. In our experiments, the
8 dilution accuracy deteriorated at these flow rates before improvements were made to the system;
9 however, no clear inverse dependence was observed, implying that the dilution bias was not limited to
10 COS dilution, and its impact and biasing can vary in a complicated manner, probably depending on the
11 experimental conditions and physical properties of the diluted gases, especially at a lower flow rate of
12 the parent gas. On the other hand, previous studies identified several factors that can cause gas
13 fractionation in gases flowing from a high-pressure cylinder via a pressure regulator. Gas fractionation
14 can be caused by kinetic processes owing to diffusive fractionation, such as pressure diffusion, thermal
15 diffusion, effusion, and gas adsorption to the inner wall of the cylinder (Langenfelds et al., 2005; Schibig
16 et al., 2018; Hall et al., 2019; Aoki et al., 2022). Given that we previously suggested the possible
17 involvement of the cylinder and regulator, these diffusive processes may contribute to the observed
18 dilution bias, with a strong influence on COS dilution. Further investigations are required to identify the
19 mechanisms underlying the observed dilution bias.

20 It is important to note that the precision of the developed dynamic dilution method is inferior to
21 that of gravimetry at this stage; however, the developed method has a practical advantage for
22 atmospheric COS observation because of the limited storage stability of COS in high-pressure cylinders.

23 Another advantage is that the new technical knowledge obtained from this study will enable us to
24 establish an automatic dynamic dilution system that can generate a diluted reference gas mixture with
25 precise pressure control of the parent and diluent gases, linked to the GC/MS system. This will expand
26 the application of the dilution system to other analytical systems for different target gases and reduce
27 the labour and time required to conduct experiments. For example, the international WMO Global

删除: A further

删除: allow

删除: ,

1 Atmosphere Watch program has set a goal of $\pm 2 \text{ nmol mol}^{-1}$ for inter-laboratory compatibility for
2 atmospheric CO measurements (WMO, 2020). In high-pressure aluminium cylinders, CO is generally
3 known to show substantial positive growth (Novelli et al., 1994, 2003; Tanimoto et al., 2007; Nara et
4 al., 2011), but the impact of CO growth can be relatively reduced at the micromolar-per-mole level (Nara
5 et al., 2011), suggesting that our relative drift-reduction approach will most likely maintain the quasi-
6 integrity of CO in the reference gas mixtures at higher CO mole fractions. Because the dilution accuracy
7 of our developed method was estimated to be within 0.1% RSD for the mole fraction of the generated
8 gas, our dynamic dilution approach can contribute greatly to fulfilling the WMO compatibility goal for
9 CO, even without gravimetric preparation. This emphasises the usefulness of our dynamic dilution
10 method with a relative drift-reduction approach for maintaining the reference gas scale and traceability
11 of unstable trace gases at ambient levels.

12

13 **Summary and conclusion**

14 We developed a one-step dynamic dilution method that is capable of accurately generating COS
15 reference gas mixtures at ambient levels (ca. $500 \text{ pmol mol}^{-1}$) from a gravimetrically prepared parent
16 gas mixture containing a high mole fraction of COS (micromole-per-mole level).

17 We began by investigating the stability of a dry reference gas mixture containing ambient levels of
18 COS stored in a high-pressure aluminium cylinder. Although a substantial decrease in the mole fraction
19 of COS was observed over time, at higher mole fractions (approximately $5 \text{ } \mu\text{mol mol}^{-1}$), COS dry gas
20 mixtures were practically stable for at least for 10 years. These results suggested that the impact of COS
21 drift during storage can be reduced to a negligible level using a gas mixture containing a high mole
22 fraction of COS. By applying this drift-reduction approach, we established a dynamic dilution method
23 that uses a gravimetric parent gas mixture containing approximately $1 \text{ } \mu\text{mol mol}^{-1}$ of COS, together with
24 HCFC-22, which was added to aid in diagnosing the cause of the dilution bias of COS observed in our
25 pilot study.

26 In the pilot study, the GC/MS response ratio of COS to HCFC-22 revealed an unexplained marked
27 bias for the dilution of the gravimetric parent gas mixture, with a greater bias for COS than for HCFC-

1 22. Although we could not identify the reason for this COS-specific dilution bias, it was considerably
2 improved by implementing strategies against the possible dilution-biasing factors of adsorption and gas
3 fractionation at the cylinder and regulator. The addition of a branch exhaust system was part of this
4 strategy, which helped rinse the flow path in the dilution system and stabilise the pressure of the
5 gravimetric parent gas while exhausting the excess parent gas flow. After implementing these
6 improvements, excellent repeatability and reproducibility were obtained for the dilution of COS and
7 HCFC-22 and for dilution linearity in our evaluation experiments. RSD values of 0.2% and 0.4% for
8 COS and 0.1% and 0.3% for HCFC-22 were obtained for the repeatability and reproducibility of the
9 method. This excellent dilution performance was corroborated by the nearly constant ratio of the
10 normalised GC/MS response of COS to that of HCFC-22. We obtained a comparable dilution
11 performance between COS and HCFC-22, allowing us to validate the dilution accuracy at ambient levels
12 for HCFC-22 by comparing the flow rate-based determination using the Molbloc system and GC/MS-
13 based determination system using a gravimetrically diluted parent gas mixture. A good agreement was
14 obtained between the two methods, demonstrating that our dynamic dilution method can accurately
15 generate diluted gas mixtures without significant systematic bias. Finally, we determined the mole
16 fractions of COS and HCFC-22 in an ambient-air-based reference gas mixture by reevaluating the
17 experimental data obtained by testing the repeatability and reproducibility of the dilution method. The
18 mole fractions of COS and HCFC-22 were precisely determined, with relative standard deviations of
19 0.1% and 0.3%, respectively.

20 Overall, these results demonstrated that our dynamic dilution method can contribute to accurate
21 long-term observations of atmospheric COS. However, the absolute determination of COS in a
22 gravimetric parent gas mixture is needed to ensure traceability and interlaboratory compatibility; thus,
23 further development is needed to better understand atmospheric COS dynamics. Notably, our dilution
24 method can be applied to the preparation of other reference gas mixtures containing unstable
25 atmospheric trace gases such as CO in high-pressure cylinders, which emphasises the usefulness of our
26 dynamic dilution method.

27

删除: 3
删除: 3
删除: 04
删除: 28

删除:

删除: 3
删除: 4

1 **Acknowledgement**

2 The authors are grateful to three anonymous referees for their insightful comments and
3 suggestions. We wish to thank K. Ooyama for his valuable comments on the pressure regulation of gas
4 flow from a high-pressure cylinder. We also thank N. Aoki for his valuable comments and technical
5 support in the gravimetric preparation of COS standard gas. This work was funded mainly by a research
6 and development project from the National Institute for Environmental Studies and partially by the Air
7 Quality and Climate Change Program of the National Institute for Environmental Studies and a Grant-
8 in-Aid for Scientific Research (S) [22H05006].

9

10 **References**

- 11 Aoki, N., Ishidoya, S., Murayama, S., and Matsumoto, N.: Influence of CO₂ adsorption on cylinders and
12 fractionation of CO₂ and air during the preparation of a standard mixture, *Atmos. Meas. Tech.*, 15,
13 5969–5983. <https://doi.org/10.5194/amt-15-5969-2022>, 2022.
- 14 Asaf, D., Rotenberg, E., Tatarinov, F., Dicken, U., Montzka, S.A., and Yakir, D.: Ecosystem
15 photosynthesis inferred from measurements of carbonyl sulphide flux, *Nat. Geosci.*, 6, 186–190.
16 <https://doi.org/10.1038/NGEO1730>, 2013.
- 17 Blonquist Jr., J.M., Montzka, S.A., Munger, J.W., Yakir, D., Desai, A.R., Dragoni, D., Griffis, T.J.,
18 Monson, R.K., Scott, R.L., and Bowling, D.R.: The potential of carbonyl sulfide as a proxy for
19 gross primary production at flux tower sites, *J. Geophys. Res.*, 116, G04019.
20 <https://doi.org/10.1029/2011JG001723>, 2011.
- 21 Brewer, P.J., Goody, B.A., Woods, P.T., and Milton, M.J.T.: A dynamic gravimetric standard for trace
22 water, *Rev. Sci. Inst.*, 82, 105102. <https://doi.org/10.1063/1.3642660>, 2011.
- 23 Brewer, P.J., Miñarro, M.D., Di Meane, E.A., and Brown, R.J.C.: A high accuracy dilution system for
24 generating low concentration reference standards of reactive gases, *Measurement*, 47, 607–612.
25 <https://doi.org/10.1016/j.measurement.2013.09.045>, 2014.
- 26 Brühl, C., Lelieveld, J., Crutzen, P.J., and Tost, H.: The role of carbonyl sulphide as a source of

1 stratospheric sulphate aerosol and its impact on climate, *Atmos. Chem. Phys.*, 12, 1239–1253.
2 <https://doi.org/10.5194/acp-12-1239-2012>, 2012.

3 Campbell, J.E., Berry, J.A., Seibt, U., Smith, S.J., Montzka, S.A., Launois, T., Belviso, S., Bopp, L., and
4 Laine, M.: Large historical growth in global terrestrial gross primary production, *Nature*, 544, 84–
5 87. <https://doi.org/10.1038/nature22030>, 2017.

6 Campbell, J.E., Carmichael, G.R., Chai, T., Mena-Carrasco, M., Tang, Y., Blake, D.R., Blake, N.J., Vay,
7 S.A., Collatz, G.J., Baker, I., Berry, J.A., Montzka, S.A., Sweeney, C., Schnoor, J.L., and Stanier,
8 C.O.: Photosynthetic control of atmospheric carbonyl sulfide during the growing season, *Science*,
9 322, 1085–1088. <https://doi.org/10.1126/science.1164015>, 2008.

10 Chin, M. and Davis, D.D.: A reanalysis of carbonyl sulfide as a source of stratospheric background
11 sulfur aerosol, *J. Geophys. Res.*, 100, 8993–9005. <https://doi.org/10.1029/95JD00275>, 1995.

12 Commane, R., Herndon, S.C., Zahniser, M.S., Lerner, B.M., McManus, J.B., Munger, J.W., Nelson,
13 D.D., and Wofsy, S.C.: Carbonyl sulfide in the planetary boundary layer: Coastal and continental
14 influences, *J. Geophys. Res.*, 118, 8001–8009. <https://doi.org/10.1002/jgrd.50581>, 2013.

15 Flores, E., Viallon, J., Moussay, P., Idrees, F., and Wielgosz, R.I.: Highly accurate nitrogen dioxide
16 (NO_2) in nitrogen standards based on permeation, *Anal. Chem.*, 84, 10283–10290.
17 <https://doi.org/10.1021/ac3024153>, 2012.

18 Fried, A., Nunnermacker, L., Cadoff, B., Sams, R., Yates, N., Dorko, W., Dickerson, R., and Winstead,
19 E.: Reference NO_2 calibration system for ground-based intercomparisons during NASA's
20 GTE/CITE 2 Mission, *J. Geophys. Res.*, 95, 10139–10146.
21 <https://doi.org/10.1029/JD095iD07p10139>, 1990.

22 Goldan, P.D., Fall, R., Kuster, W., and Fehsenfeld, F.C.: Uptake of COS by growing vegetation: a major
23 tropospheric sink, *J. Geophys. Res.*, 93, 14186–14192. <https://doi.org/10.1029/JD093iD11p1418>,
24 1988.

25 Goldan, P.D. and Kuster, W.C.: A dynamic dilution system for the production of sub-ppb concentrations
26 of reactive and labile species, *Atmos. Environ.*, 20, 1203–1209. [https://doi.org/10.1016/0004-
27 6981\(86\)90154-X](https://doi.org/10.1016/0004-), 1986.

1 Goody, B.A. and Milton, M.J.T.: High-accuracy gas flow dilutor using mass flow controllers with binary
2 weighted flows, *Meas. Sci. Technol.*, 13, 1138–1145. <https://doi.org/10.1088/0957-0233/13/7/323>,
3 2002.

4 Guillevic, M., Vollmer, M.K., Wyss, S.A., Leuenberger, D., Ackermann, A., Pascale, C., Niederhauser,
5 and B., and Reimann, S.: Dynamic–gravimetric preparation of metrologically traceable primary
6 calibration standards for halogenated greenhouse gases, *Atmos. Meas. Tech.*, 11, 3351–3372.
7 <https://doi.org/10.5194/amt-11-3351-2018>, 2018.

8 Hall, B.D., Engel, A., Mühle, J., Elkins, J.W., Artuso, F., Atlas, E., Aydin, M., Blake, D., Brunke, E.-G.,
9 Chiavarini, S., Fraser, P.J., Happell, J., Krummel, P.B., Levin, I., Loewenstein, M., Maione, M.,
10 Montzka, S.A., O’Doherty, S., Reimann, S., Rhoderick, G., Saltzman, E.S., Scheel, H.E., Steele,
11 L.P., Vollmer, M.K., Weiss, R.F., Worthy, D., and Yokouchi, Y.: Results from the International
12 Halocarbons in Air Comparison Experiment (IHALACE), *Atmos. Meas. Tech.*, 7, 469–490.
13 <https://doi.org/10.5194/amt-7-469-2014>, 2014.

14 Hall, B.D., Crotwell, A.M., Miller, B.R., Schibig, M., and Elkins, J.W.: Gravimetrically prepared carbon
15 dioxide standards in support of atmospheric research, *Atmos. Meas. Tech.*, 12, 517–524,
16 <https://doi.org/10.5194/amt-12-517-2019>, 2019.

17 ISO 6142-1: Gas analysis–Preparation of calibration gas mixtures–Part 1: Gravimetric method for Class
18 I mixtures, available at: <https://www.iso.org/standard/59631.html> (last access: Jan 2023), 2015.

19 ISO 6145-7: Gas analysis–Preparation of calibration gas mixtures using dynamic methods–Part 7:
20 Thermal mass-flow, available at: <https://www.iso.org/standard/45471.html> (last access: Jan 2023),
21 2018.

22 ISO 21748: Guidance for the use of repeatability, reproducibility and trueness estimates in measurement
23 uncertainty evaluation, available at: <https://www.iso.org/standard/71615.html> (last access: May
24 2023), 2017.

25 Kerwin, R.A., Crill, P.M., Talbot, R.W., Hines, M.E., Shorter, J.H., Kolb, C.E., and Harriss, R.C.:
26 Determination of atmospheric methyl bromide by cryotrapping-gas chromatography and
27 application to soil kinetic studies using a dynamic dilution system, *Anal. Chem.*, 68, 899–903.

1 <https://doi.org/10.1021/ac950811z>, 1996.

2 Kim, M.E., Kim, Y.D., Kang, J.H., Heo, G.S., Lee, D.S., and Lee, S.: Development of traceable precision
3 dynamic dilution method to generate dimethyl sulphide gas mixtures at sub-nanomole per mole
4 levels for ambient measurement, *Talanta*, 150, 516–524.
5 <https://doi.org/10.1016/j.talanta.2015.12.063>, 2016.

6 Kjellstrom, E.: A three-dimensional global model study of carbonyl sulfide troposphere and the lower
7 stratosphere, *J. Atmos. Chem.*, 29, 151–177. <https://doi.org/10.1023/A:1005976511096>, 1998.

8 Kooijmans, L.M.J., Sun, W., Aalto, J., Erkkilä, K., Maseyk, K., Seibt, U., Vesala, T., Mammarella, I.,
9 and Chen, H.: Influences of light and humidity on carbonyl sulfide-based estimates of
10 photosynthesis, *P. Natl. Acad. Sci. USA*, 116, 2470–2475.
11 <https://doi.org/10.1073/pnas.1807600116>, 2019.

12 Landau, L. and Lifshitz, E.: Fluid mechanics 2nd edition, Course of theoretical Physics, Volume 6, 1987.

13 Langenfelds, R.L., van der Schoot, M.V., Francey, R.J., Steele, L.P., Schmidt, M., and Mukai, H.:
14 Modification of air standard composition by diffusive and surface processes, *J. Geophys. Res.*,
15 110, D13307. <https://doi.org/10.1029/2004JD005482>, 2005.

16 Macé, T., Iturrate-Garcia, M., Pascale, C., Niederhauser, B., Vaslin-Reimann, S., and Sutour, C.: Air
17 pollution monitoring: development of ammonia (NH₃) dynamic reference gas mixtures at
18 nanomoles per mole levels to improve the lack of traceability of measurements, *Atmos. Meas.*
19 *Tech.*, 15, 2703–2718. <https://doi.org/10.5194/amt-15-2703-2022>, 2022.

20 Maseyk, K., Berry, J.A., Billesbach, D., Campbell, J.E., Torn, M.S., Zahniser, M., and Seibt, U.: Sources
21 and sinks of carbonyl sulfide in an agricultural field in the Southern Great Plains, *P. Natl. Acad.*
22 *Sci. USA*, 111, 9064–9069. <https://doi.org/10.1073/pnas.1319132111>, 2014.

23 Mohamad, G.H.P., Coles, G.S.V., and Watson, J.: An automatic low-level gas blender, *Trans. Inst. Meas.*
24 *Cont.*, 18, 62–68. <https://doi.org/10.1177/0142331296018002>, 1996.

25 Montzka, S., Aydin, M., Battle, M., Butler, J., Saltzman, E., Hall, B., Clarke, A., Mondeel, D., and Elkins,
26 J.: A 350-year atmospheric history for carbonyl sulfide inferred from Antarctic firn air and air
27 trapped in ice, *J. Geophys. Res.*, 109, D22302. <https://doi.org/10.1029/2004JD004686>, 2004.

1 Montzka, S.A., Calvert, P., Hall, B.D., Elkins, J.W., Conway, T.J., Tans, P.P., and Sweeney, C.: On the
2 global distribution, seasonality, and budget of atmospheric carbonyl sulfide (COS) and some
3 similarities to CO₂, *J. Geophys. Res. Atmos.*, 112, D09302.
4 <https://doi.org/10.1029/2006JD007665>, 2007.

5 Nakao, S. and Takamoto, M.: Development of the calibration facility for small mass flow rates of gases
6 and the sonic venturi nozzle transfer standard, *JSME Int. J. Ser. B.*, 42, 667–673.
7 <https://doi.org/10.1299/jsmeb.42.667>, 1999.

8 Nara, H., Tanimoto, H., Nojiri, H., Mukai, H., Machida, T., and Tohjima, Y.: Onboard measurement
9 system of atmospheric carbon monoxide in the Pacific by voluntary observing ships, *Atmos. Meas.*
10 *Tech.*, 4, 2495–2507. <https://doi.org/10.5194/amt-4-2495-2011>, 2011.

11 Nara, H., Tanimoto, H., Tohjima, Y., Mukai, H., Nojiri, Y., Katsumata, K., and Rella, C.W.: Effect of air
12 composition (N₂, O₂, Ar, and H₂O) on CO₂ and CH₄ measurement by wavelength-scanned cavity
13 ring-down spectroscopy: calibration and measurement strategy, *Atmos. Meas. Tech.*, 5, 2689–
14 2701. <https://doi.org/10.5194/amt-5-2689-2012>, 2012.

15 Novelli, P.C., Collins Jr., J.E., Myers, R.C., Sachse, G.W., and Scheel, H.E.: Reevaluation of the
16 NOAA/CMDL carbon monoxide reference scale and comparison with CO reference gases at
17 NASA-Langley and Fraunhofer Institute, *J. Geophys. Res.*, 99, 12833–12839.
18 <https://doi.org/10.1029/94JD00314>, 1994.

19 Novelli, P.C., Masarie, K.A., Lang, P.M., Hall, B.D., Myers, R.C., and Elkins, J.W.: Reanalysis of
20 tropospheric CO trends: effects of the 1997–1998 wildfires, *J. Geophys. Res.*, 108, 4464.
21 <https://doi.org/10.1029/2002JD003031>, 2003.

22 Protoschill-Krebs, G., Wilhelm, C., and Kesselmeier, J.: Consumption of carbonyl sulphide (COS) by
23 higher plant carbonic anhydrase (CA), *Atmos. Environ.*, 30, 3151–3156.
24 [https://doi.org/10.1016/1352-2310\(96\)00026-X](https://doi.org/10.1016/1352-2310(96)00026-X), 1996.

25 Saito, T., Yokouchi, Y., Stohl, A., Taguchi, S., and Mukai, H.: Large emissions of perfluorocarbons in
26 East Asia deduced from continuous atmospheric measurements, *Environ. Sci. Technol.*, 44, 4089–

1 4095. <https://doi.org/10.1021/es1001488>, 2010.

2 Sandoval-Soto, L., Stanimirov, M., von Hobe, M., Schmitt, V., Valdes, J., Wild, A., and Kesselmeier, J.:
3 Global uptake of carbonyl sulfide (COS) by terrestrial vegetation: Estimates corrected by
4 deposition velocities normalized to the uptake of carbon dioxide (CO₂), *Biogeosciences*, 2, 125–
5 132. <https://doi.org/10.5194/bg-2-125-2005>, 2005.

6 Schibig, M.F., Kitzis, D., and Tans, P.P.: Experiments with CO₂- in-air reference gases in high-pressure
7 aluminum cylinders, *Atmos. Meas. Tech.*, 11, 5565–5586. [https://doi.org/10.5194/amt-11-5565-](https://doi.org/10.5194/amt-11-5565-2018)
8 2018, 2018.

9 Seibt, U., Kesselmeier, J., Sandoval-Soto, L., Kuhn, U., and Berry, J.A.: A kinetic analysis of leaf uptake
10 of COS and its relation to transpiration, photosynthesis and carbon isotope fractionation,
11 *Biogeosciences*, 7, 333–341. <https://doi.org/10.5194/bg-7-333-2010>, 2010.

12 Stimler, K., Berry, J.A., and Yakir, D.: Effects of carbonyl sulfide and carbonic anhydrase on stomatal
13 conductance^{1[OA]}, *Plant Physiol.*, 158, 524–530. <https://doi.org/10.1104/pp.111.185926>, 2012.

14 Stimler, K., Montzka, S.A., Berry, J.A., Rudich, Y., and Yakir, D.: Relationships between carbonyl
15 sulfide (COS) and CO₂ during leaf gas exchange, *New Phytol.*, 186, 869–878.
16 <https://doi.org/10.1111/j.1469-8137.2010.03218.x>, 2010.

17 Tanimoto, H., Sawa, Y., Matsueda, H., Yonemura, S., Wada, A., Mukai, H., Wang, T., Poon, S., Wong,
18 A., Lee, G., Jung, J.Y., Kim, K.R., Lee, M., Lin, N.H., Wang, J.L., Ou-Yang, C.F., and Wu, C.F.:
19 Evaluation of standards and methods for continuous measurements of carbon monoxide at ground-
20 based sites in Asia, *Pap. Meteorol. Geophys.*, 58, 85–93. <https://doi.org/10.2467/mripapers.58.85> ,
21 2007.

22 Tera Term project team, Tera Term 4.97 (Version 4.97), November 30, 2017.
23 <https://tssh2.osdn.jp/index.html.en>. (last access: Jan 2023).

24 Wehr, R., Munger, J.W., McManus, J.B., Nelson, D.D., Zahniser, M.S., Davidson, E.A., Wofsy, S.C.,
25 and Saleska, S.R.: Seasonality of temperate forest photosynthesis and daytime respiration, *Nature*,
26 534, 680–683. <https://doi.org/10.1038/nature17966>, 2016.

1 WMO: Scientific assessment of ozone depletion, 2018, GAW Report 58, World Meteorological
2 Organization (WMO), Global Ozone Research and Monitoring Project, Geneva, Switzerland,
3 available at: [https://public-old.wmo.int/en/resources/library/scientific-assessment-of-ozone-](https://public-old.wmo.int/en/resources/library/scientific-assessment-of-ozone-depletion-2018)
4 [depletion-2018](https://public-old.wmo.int/en/resources/library/scientific-assessment-of-ozone-depletion-2018) (last access: Jan 2023), 2018.

5 WMO: Scientific assessment of ozone depletion, 2022, GAW Report 278, World Meteorological
6 Organization (WMO), Global Ozone Research and Monitoring Project, Geneva, Switzerland,
7 available at: <https://www.csl.noaa.gov/assessments/ozone/2022/> (last access: Oct 2023), 2022.

8 WMO: Report of the 20th WMO/IAEA Meeting on carbon dioxide, other greenhouse gases and related
9 measurement techniques, 2–5 September 2019, GAW Report No. 255, available at:
10 https://library.wmo.int/index.php?lvl=notice_display&id=21758 (last access: Jan 2023), 2020.

11 Wright, R.S. and Murdoch, R.W.: Laboratory evaluation of gas dilution systems for analyser calibration
12 and calibration gas analysis, *Air Waste*, 44, 428–430.
13 <https://doi.org/10.1080/1073161X.1994.10467265>, 1994.

14 Yang, F., Qubaja, R., Tatarinov, F., Rotenberg, E., and Yakir, D.: Assessing canopy performance using
15 carbonyl sulfide measurements, *Glob. Chang. Biol.*, 24, 3486–3498.
16 <https://doi.org/10.1111/gcb.14145>, 2018.

17 Yokohata, A., Makide, Y., and Tominaga, T.: A new calibration method for the measurement of CCl₄
18 concentration at 10⁻¹⁰ v/v level and the behavior of CCl₄ in the atmosphere, *B. Chem. Soc. Jpn.*,
19 58, 1308–1314. <https://doi.org/10.1246/bcsj.58.1308>, 1985.