

Multiphysical description of atmospheric pressure interface chemical ionisation in MION2 and Eisele type inlets

Henning Finkenzeller^{1,2}, Jyri Mikkilä², Cecilia Righi¹, Paxton Juuti², Mikko Sipilä¹, Matti Rissanen^{3,4}, Douglas Worsnop^{1,5}, Aleksei Shcherbinin², Nina Sarnela¹, and Juha Kangasluoma^{1,2}

¹Institute for Atmospheric and Earth System Research / Department of Physics, Faculty of Science, University of Helsinki, Helsinki, FI-00560, Finland

²Karsa Ltd., FI-00560, Finland

³Aerosol Physics Laboratory, Physics Unit, Faculty of Engineering and Natural Sciences, Tampere University, Tampere, FI-33720, Finland

⁴Department of Chemistry, University of Helsinki, Helsinki, FI-00014, Finland

⁵Aerodyne Research Inc., Billerica, Massachusetts 01821, USA

Correspondence: Henning Finkenzeller (henning.finkenzeller@helsinki.fi)

Abstract. Chemical ionisation inlets are fundamental instrument components in chemical ionisation mass spectrometry (CIMS). However, the sample gas and reagent ion trajectories are often understood only in a general and qualitative manner. Here we evaluate two atmospheric pressure chemical ionisation inlets (MION2 and Eisele type inlet) with computational fluid dynamics 3D physico-chemical models regarding the reagent ion and sample gas trajectories and estimate their efficiencies of reagent ion production, reagent ion delivery from the ion source volume into the ion–molecule mixing region, and the interaction between reagent ions and target molecules. The models are validated by laboratory measurements and quantitatively reproduce observed sensitivities to tuning parameters, including ion currents and changes in mass spectra. The study elucidates how the different transport and chemical reactions proceed within the studied inlets, that space charge can already be relevant at ion concentrations of as low as 10^7 cm^{-3} , and compares the two investigated inlet models. The models provide insights into how to operate the inlets and will help in the development of future inlets that further enhance the capability of CIMS.

1 Introduction

Mass spectrometry (MS) requires target molecules to be electrically charged to determine their mass-to-charge ratios and infer their atomic composition. Electrically neutral sample molecules require ionisation prior to MS. In chemical ionisation (CI) inlets, gas-phase ions created outside the sample inlet are introduced into the ion-molecule mixing region (IMR), where they interact with the sample gases by ion attachment, scattering or recombination (Passananti et al., 2019). The ions are transmitted from the IMR to a mass spectrometer at lower pressure through a suitable aperture.

For accurate and precise mass-spectrometric measurements, chemical ionisation inlets should introduce reagent ions into the IMR at sufficient concentration, enable a suitable reaction time in the IMR, avoid transport losses within the inlet, and minimise other measurement biases, such as contamination. Higher reagent ion concentrations not only allow better detection

20 limits, but also delay reagent ion depletion in conditions with high analyte concentrations, thereby enhancing the dynamic range. Additionally, inlets should be robust and easy to operate.

A variety of CI inlet designs have been developed since the emergence of CIMS, which differ in used reagent ions, supply and generation of the reagent ions (radioactive source, x-ray, VUV lamp, corona discharge, electrospray), reaction time, and IMR pressures. Reduced pressure inlets (IMR at fractions of atmospheric pressure) are used to suppress multiple collisions between the reagent ion and target molecules (required for e.g. proton-transfer-reaction mass spectrometry (PTR-MS, Yuan et al. (2017))) and the formation of reagent ion oligomers. Atmospheric pressure interface (API) CI inlets minimise the introduction of turbulence in the sample flows and reach excellent detection limits due to high reagent ion and sample gas concentrations in the IMR. Prominent examples of API CI inlets used in atmospheric science are the Eisele type (Eisele and Tanner, 1991, 1993; Tanner and Eisele, 1995; Tanner et al., 1997; Mauldin III et al., 1998; Sipilä et al., 2015) and MION inlet (Rissanen et al., 30 2019; Wang et al., 2021; Huang et al., 2021; Rissanen, 2021; Iyer et al., 2021; Shen et al., 2022; Finkenzeller et al., 2023; He et al., 2023; Partovi et al., 2023).

The description of the inlets in the literature is currently limited to schematics or conceptual modelling (Sipilä et al., 2015). While tuning the control parameters of inlets (voltages and gas flow rates) is generally required before measurements, the influence of individual control parameters on the detailed processes within the instruments may not always be straightforward and evident to the users. Examples of well-known but little-understood features are the formation of Br_3^- or I_3^- in Br^- or I^- CIMS, the presence of reagent precursor gas in the IMR of the Eisele inlet, or the sensitivity of the Eisele inlet to voltages and the exact insertion depth of the sample tube. Using physico-chemical modelling together with measurements, this study aims to provide a clear picture of the respective gas and ion trajectories within the inlets, to compare the two inlet designs, and to identify limitations and avenues for the development of improved inlet designs with higher reagent ion yields and other 40 ion-chemistries.

2 Methods

2.1 Physico-chemical processes within the inlets and representation in model

The mechanisms influencing the trajectories and distribution of the reagent ions and target molecules are: (1) the ion generation from a source gas; (2) the gas flow throughout the inlet, i.e., sample and auxiliary gas flows; (3) electric fields from ion optics and ions themselves; (4) diffusion; (5) the transport of molecules and ions in the electro-convective field; (6) the chemical 45 conversion between species in the gas phase; and (7) interaction (loss or conversion) of gas-phase species on surfaces. These processes are quantitatively and spatially represented in a stationary model implemented in COMSOL Multiphysics 6.1. The controlling factors and metrics along the ion trajectories are described in the following.

2.1.1 Ion generation in source region

50 In MION2 and Eisele type inlets, precursor gases are split into positive and negative ions by introducing energy from radioactive sources or x-ray lamps (Anttalainen et al., 2021). The detailed mechanism of ion formation from precursor gases is not trivial – e.g., ionisation radiation may initially ionise the bath gas, not the precursor molecules itself – and is not the focus of this study. Here, the primary production of ions is approximated as splitting up a precursor gas R^cR into the reagent ion R^\pm and a counter ion $R^{c\mp}$ of opposite polarity at a prescribed rate.

55 Ion-ion recombination is extremely fast, with typical bimolecular rate coefficients of $1.7 \cdot 10^{-6} \text{ cm}^3 \text{ s}^{-1}$ (Zauner-Wieczorek et al., 2022), four orders of magnitude faster than radical-radical recombination. As a binary process, it becomes progressively more important at higher ion concentrations. The lifetime of a population against recombination at $[R^\pm] = [R^{c\mp}] = 10^9 \text{ cm}^{-3}$ is only $\sim 1 \text{ ms}$. In absence of efficient charge separation, ion production is near-completely offset by recombination, and a doubling of the initial ion production rate P does not double the established concentrations in the ionisation volume but rather
60 increases them by 40%. The obvious method to separate newly-generated ions of opposite polarity is the application of electric fields. A metric assessing the capability of an inlet to generate reagent ions is the reagent ion concentration c_S downstream of the ion source volume, as it presents the upper limit for the attainable ion concentrations in the IMR.

2.1.2 Electro-convective ion transport to IMR

The ion bulk velocity v_{ion} is the sum of advective and electrophoretic velocity, i.e., the advective flow field and the electric field
65 are additive:

$$v_{\text{ion}} = v_{\text{conv.}} + \mu E \quad (1)$$

Here, v_{ion} is the flow velocity [cm s^{-1}], μ [$\text{cm s}^{-1} \text{ cm V}^{-1}$] the electrical mobility, and E [V cm^{-1}] the electric field strength. The ion transport may occur predominately advectively, convectively (i.e., by advection and diffusion), or electrophoretically in different parts of the inlet.

70 Externally constrained advective and electrophoretic fields (i.e., disregarding space charge induced fields) are divergence free, i.e., the only field sources and sinks are the boundary conditions (gas inlet and outlet, electrodes); the in-flux I into a given volume is equal to its out-flux. If for a given I the out-flux area A changes, the out-flow velocity v changes inversely:

$$\frac{I}{A} = \text{const.} \quad (2)$$

Here, A is the flow cross section area [m^2]. Therefore, ion mixing ratios are conserved along electro-advective streamlines,
75 and ion concentrations c are conserved if the gas density does not change. Analogous to [river-banks narrowing riverbanks](#) that increase the water flow velocity (v) ~~through a narrow~~ [by reducing the cross-section area of the flow](#) (A) but do not change the composition of the water (c), electric fields defined by electrodes affect the ion trajectories without changing their concentration. In absence of collisional focusing (e.g., Kelly et al., 2010), ion concentrations along streamlines do not increase.

The reagent ion concentration at the pinhole c_P , the orifice connecting the IMR and the mass spectrometer, may reach but
80 cannot exceed the upstream ion concentration at the ion source c_S . The theoretical maximum ion current at the pinhole I_{max}

is hence the product of the ion concentration at the ion source region c_S and the flow rate at the pinhole J_P :

$$I_{max} = J_P c_S \quad (3)$$

To achieve a high concentration of reagent ion in the IMR hence requires to (1) create a high initial concentration of ions, and to (2) efficiently deliver the ions to the IMR.

85 Efficient ion transport means that the initially generated ion concentration is maintained along streamlines to the IMR. The ion delivery efficiency η_D is defined here as

$$\eta_D = \frac{c_P}{c_S} \quad (4)$$

Here, c_P is the ion concentration at the pinhole (the aperture to the mass spectrometer), c_S is the ion concentration at the ion source.

90 The space charge of ions distributed in space needs to be considered if ion concentrations are so high that the induced electric fields are comparable in magnitude to the prescribed electro-advective field. Gauss's law describes the creation of electric fields due to charge distributions:

$$\nabla \cdot \mathbf{E} = \frac{\rho}{\varepsilon} \quad (5)$$

Here, ε [F m^{-1}] is the permittivity, $\rho = ce$ [C cm^{-3}] the charge concentration. For a beam of singly charged ions with concentration $c = 10^7 \text{ cm}^{-3}$, the space charge-induced electric field 5 mm off the beam centre axis is $E = 9 \text{ V cm}^{-1}$, corresponding to a radial drift velocity $v = 22 \text{ cm s}^{-1}$ ($\mu(\text{NO}_3^- - \text{N}_2) = 2.4 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$). This is significant when compared to typical advective velocities of 1 m s^{-1} . Space charge matters even at these relatively low concentrations.

Diffusion of the ions perpendicular to the electro-advective streamlines needs to be considered wherever concentration gradients are significant, especially at the edges of ion beams.

100 The electro-convective streamlines that connect the ionisation volume and the pinhole define what part of the ionisation volume actually contributes to the delivery of ions into the pinhole. As the product of flow velocity and area is constant for a given flow rate (eq. 3), the increase of E necessarily requires the area of usable extraction A to become smaller. Faster transport minimises space-charge and diffusional losses, but generally decreases the source concentration c_S . This is analogous to dumping a compound into a river at a constant rate: the faster the river flows, the lower the resulting compound concentration.

105 2.1.3 Reaction time in ion–molecule mixing region

The reaction time t between analyte molecules A and reagent ions R^\pm influences the abundance of analyte–reagent ion clusters, in addition to the concentrations of analyte and reagent ions, and the clustering reaction rate constant k . The model allows to elegantly determine t . Consider a bi-molecular clustering reaction $A + R^\pm \rightarrow AR^\pm$ with reaction rate coefficient k . If the reactants A and R are not significantly consumed in the reaction, then the concentrations $[A]$ and $[R]$ at a given time are representative for the entire reaction time and the cluster concentration at the pinhole is a simple function of t :

$$[AR^\pm] = tk[A][R^\pm] \quad (6)$$

The concentrations and reaction times along different trajectories to the pinhole are generally not the same. The average reaction time t_{avg} that considers different trajectories to the pinhole with different reaction times and concentrations is given by the integral pinhole currents I for a given pinhole flow J_P :

$$115 \quad t_{\text{avg}} = \frac{I_{AR^\pm} J_P}{k I_A I_{R^\pm}} \quad (7)$$

2.1.4 Theoretical calibration factor and detection limit

The calibration factor C_A [$\text{cm}^{-3} \text{cps}^{-1} \text{cps}$] for a target compound is a result of k and t_{avg} :

$$C_A = c_A \frac{I_{R^\pm}}{I_{AR^\pm}} = \frac{I_A}{J_P} \frac{I_{R^\pm}}{I_{AR^\pm}} \quad (8)$$

Here, c_A is the concentration of compound A. C_A depends on the compound- and detector-specific detection sensitivities and
 120 needs to be determined experimentally. The detection limit Λ additionally depends on the ~~magnituded~~ magnitude of I_{R^\pm} and the I_{AR^\pm} baseline.

2.2 Model setup

The inlet geometries are approximated by meshes consisting of multiple million volumes, including surface layers. The symmetry of the inlets is exploited to limit the modelling to half (MION2) or even a quarter (Eisele) of the full geometry. The convective flow field is determined prior to determining the concentrations of chemical compounds in the electro-convective field.
 125 This reduces the complexity of the numerical system and is justified as the convective field is not influenced by the transport of dilute molecules. The model uses a temperature of 293 K and gas reference pressure of 1 atm. The model assumes laminar flow (~~Reynolds number for the inlet $Re \approx 1600$~~ , and uses prescribed rates for the exhaust, pinhole, and auxiliary flows as constraints. For the modelling of the Eisele-type inlet, 10slpm sample, 20slpm sheath, and 1slpm flow to the mass spectrometer are used (Tanner and Eisele, 1995). For the MION2 inlet, 20slpm exhaust flow (Wang et al., 2021) and 0.8slpm flow to the mass spectrometer are used. The auxiliary reagent, purge, and reagent exhaust flow are $J_R = 10\text{smlpm}$, $J_{RE} = 50\text{smlpm}$, and $J_{RP} = 100\text{smlpm}$.

The electric fields are constrained by specified electrode potentials and the space charge of the ions. The model uses the electric mobility constant $\mu = 2.4 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$ ($\mu(\text{NO}_3\text{-air}, T = 24^\circ\text{C})$, Xuemeng Chen, personal communication (compare
 135 Steiner et al., 2014)) for all ions, equivalent to a diffusivity constant $D = 0.062 \text{ cm}^2 \text{ s}^{-1}$. The mobility constant determines what electric field strength magnitude is required. The variability in electrical mobility between different ions is small and not critical in this study (Hwang et al., 1989; Hwang and Su, 1990; De Andrade et al., 1992; Filippov et al., 2017; Cussler, 2009), but could be significant in systems with light reagent ions and large clusters, i.e., proton transfer reaction mass spectrometers.

In this study, simplified chemistry schemes for the operation with either NO_3^- or Br^- as reagent ion were used. The reagent
 140 ions NO_3^- and Br^- are produced from splitting the source gas (nitric acid, $\text{HNO}_3 \rightarrow \text{H}^+ + \text{NO}_3^-$, or bromomethane, $\text{CH}_3\text{Br} \rightarrow \text{CH}_3^+ + \text{Br}^-$). Br^- is in practice often generated from dibromomethane, which could in principle donate two Br^- and has different ionisation properties; in this study, dibromomethane can be considered interchangeable with bromomethane if only

a single dissociation is assumed to take place. As proxy for target molecules, dilute sulphuric acid H_2SO_4 is contained in the sample flow at a mixing ratio of 1 ppt. It reacts kinetically with Br^- and NO_3^- to form $\text{H}_2\text{SO}_4 \cdot \text{Br}^-$ and $\text{H}_2\text{SO}_4 \cdot \text{NO}_3^-$. While the precursor gases are assumed to be in steady state with the surfaces, H_2SO_4 and all ions are assumed to be lost efficiently to the inlet surfaces. Br^- , Br_3^- , and $\text{H}_2\text{SO}_4 \cdot \text{Br}^-$ surface-uptake is assumed to lead to the re-emission of Br_2 (at respective stoichiometric ratios). The same chemistry is used in both inlets, facilitating a direct comparison.

2.3 Laboratory measurements

Measured electrode currents due to the absorption of attracted ions were used to constrain the production rate of ion pairs in the ionisation volume. The currents to the two topmost electrodes of the MION2 inlet ($4 \cdot 10^{-11}$ A, attraction of H^+) and for the ion cage of the Eisele inlet ($6 \cdot 10^{-11}$ A, attraction of H^+ , negligible adsorption of NO_3^-) were determined via the ~~voltage drop~~ voltage-drop across the internal 10. M Ω resistor ~~of a voltmeter dedicated to measure voltages in a simple multimeter~~ (Tenma 72-2595). The voltage-drop of 0.6 mV is measurable by the voltmeter and does not constitute a measurement bias under the test conditions. For both inlets, the model reproduced the measured currents assuming a production rate $P = 6 \cdot 10^7 \text{ cm}^3 \text{ s}^{-1}$, equivalent to an ion-ion recombination determined steady state concentration of $6 \cdot 10^6 \text{ cm}^{-3}$. Bromide spectra with the MION2 inlet were measured with a long time of flight MS (LTOF, Tofwerk AG, Switzerland); the voltage-dependent ion current to the mass spectrometer in the Eisele setup was measured with a high resolution time of flight MS (HTOF, Tofwerk AG, Switzerland). The Eisele inlet was used with a single x-ray source (Hamamatsu L12535).

3 Results

3.1 MION2 inlet

Figure 1 shows the geometry and physical quantities in the MION2 inlet, using nitric acid as reagent gas and sulfuric acid as sample gas. Sample gas is drawn into the IMR (inner diameter 22 mm) and to the pinhole and exhaust advectively, while auxiliary flows in the ion source region are minimal (Fig. 1B). ~~Here, b).~~ Assuming an interface upstream of the MION2 inlet that creates laminar flow (Reynolds number $Re \approx 2100$, using $D = 20 \text{ mm}$, $u = 1.6 \text{ m s}^{-1}$, $\nu = 1.48 \cdot 10^{-5} \text{ m}^2 \text{ s}$), the flow velocity profile is parabolic throughout the sample tube and IMR close up to the pinhole plate, where the flow splits to the exhaust and pinhole. In the ion source, predominately electric fields, generated by 20 electrodes, transport the ions (Fig. 1C). Figure 1D shows the electro-advective field and streamlines for ions. Note that the transfer from electric to convective transport occurs where the electric and convective streamlines are approximately perpendicular to each other. Sample gas (here, H_2SO_4 and H_2O , Fig. 1E and F) is kept out of the ion source volume by a small purge flow J_{RP} , and reagent gas provided in flow J_R is likewise contained to the ionisation volume only (Fig. 1G) with a small exhaust flow J_{RE} . The ionisation of the precursor gas HNO_3 leads to formation of the complementary ions H^+ and NO_3^- (Fig. 1H and J), which are separated by electric fields. Ion-ion recombination is strongest in the centre of the ionisation volume (Fig. 1I). Br^- is transported along the ion-flow streamlines into the IMR. The clustering of NO_3^- and H_2SO_4 leads to buildup of $\text{H}_2\text{SO}_4 \cdot \text{NO}_3^-$ clusters (Fig. 1K).




Figure 1. Physical quantities in the MION2 inlet centre plane using NO_3^- reagent ion. The colour scale ranges from 0 to the maximum described in each panel. Panel **d** shows the electro-advective velocity for anions with $\mu = 2.4 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$. Used settings: Accelerator voltage $U_A = -1500 \text{ V}$, deflector voltage $U_D = -110 \text{ V}$, exhaust flow $J_E = 20 \text{ slpm}$, pinhole flow $J_P = 800 \text{ sccm}$, reagent gas flow $J_R = 15 \text{ sccm}$, reagent exhaust flow $J_{RE} = 50 \text{ sccm}$, reagent purge flow $J_{RP} = 100 \text{ sccm}$.

Figure 2 shows the simplified bromine chemistry in the MION2 inlet when bromomethane is used as the reagent gas. Here, bromomethane (Fig. 2Aa) is split into methylum and bromide (Fig. 2B and Db and d). Br_2 concentrations result from the absorption of Br^- (and Br_3^-) to the surfaces and re-emission as Br_2 (Fig. 2Ee). Br_3^- is formed in the model by the kinetic recombination of Br^- and Br_2 (Fig. 2Ff).

Figure 3 shows the voltage dependent ratio of Br_3^- and Br^- delivery in MION2 in measurements and the model. The model is able to reproduce the observed trend, which is due to the voltage-dependent reaction time within the Br_2 -filled volume: At




Figure 2. Bromine chemistry in the MION2 inlet centre plane. Br_2 forms in the wall uptake of Br^- and Br_3^- . Br_3^- forms from recombining Br^- and Br_2 . The colour scale ranges from 0 to the maximum described in each panel.




Figure 3. Br_3^- sensitivity in measurements and model. At low voltages, the longer residence time in the ion source leads to enhanced relative formation of Br_3^- . Error bars indicate the measurement standard deviation.

180 low electro-convective velocities there is more time for the Br_2 - Br^- clustering to occur. The model-predicted relative Br_3^- abundance is on the order of permil for the studied conditions. Higher Br_3^- abundances would establish if either higher Br^- concentrations led to stronger Br_2 production or dilution in the ionisation volume was reduced by a slower supply of reagent gas. Additionally, the Br_2 - Br^- recombination, a neutral-ion clustering, likely occurs at a rate faster than the neutral-neutral collision rate currently used in the model. As the concentrations of neither Br_3^- nor Br^- were measured quantitatively in the

185 mass spectrometer, due to compound-specific transmission and detection efficiencies, only scaled ratios are shown in Fig. 3. The Br_3^- formation mechanism could be applied to analogously explain the formation of I_3^- in iodide CIMS.




Figure 4. Sensitivities of NO_3^- concentrations in MION2 inlet to different acceleration voltages U_A (A-Ca-c), deflector voltage $U_D = 0$ for deactivation (Dd), and primary ion production rate (Ee). The colour scale ranges from 0 to the maximum described in each panel. Figures A-D-a-d use the same colour scale. The width of the ion beam increases for larger voltages, while the extracted concentrations slightly decrease. At concentrations of 10^7 cm^{-3} space charge leads to a spreading of the ion beam, the concentration at the pinhole is lower than at the ionisation volume.

Figure 4 shows how the accelerator voltage U_A and deflector voltage U_D affect the ion trajectories and concentration. The accelerator voltage U_A influences the concentration of ions extracted from the ionisation volume and controls the width of the reagent ion beam injected into the sample tube and IMR (Figure 4A-Ca-c). The decrease of source concentration c_S for higher U_A is expected as for faster electro-convective transport the ions, produced at a constant rate, are distributed over a wider volume. To maximise the ion delivery to the pinhole, all gas drawn into the pinhole should be illuminated by ions, i.e., the width of the ion beam should be as wide as the pinhole flow collection aperture. Beams slightly larger than geometrically needed help to counter diffusion losses and reduce the sensitivity to choosing the deflector voltage U_D correctly. The deflector voltage U_D controls how deep the ions are pushed radially into the convective sample tube. If U_D is too repellent, ions are pushed to the opposite side of the sample tube. If it is insufficiently repellent, the reagent ions do not penetrate into the pinhole flow. If chosen correctly, the electro-advective streamlines connect the pinhole and the ionisation volume (Figure 1Dd). Figure 4D-d shows that when the deflector voltage U_D is set to ground potential, i.e., minimal repulsion, some ions are still injected. To fully suppress ion injection, the x-ray source should be switched off. Figure 4E-e illustrates how space charge progressively matters at higher ion concentrations. Here, at $[\text{NO}_3^-] = 1.2 \cdot 10^7 \text{ cm}^{-3}$, space charge leads to a widening of the ion beam, and the ion concentration notably decreases from the source to the pinhole.




Figure 5. Ion concentration conservation from the MION2 ionisation volume to the pinhole, as function of the accelerator voltage U_A . The red dashed trace shows the ion concentration respectively the current entering the pinhole. The concentration and current axis are connected by the pinhole flow rate J_P .

Figure 5 shows for the MION2 inlet the NO_3^- ion source concentration c_S , i.e., the concentration of reagent ions just downstream of the ionisation volume, for different accelerating voltages U_A . While stronger electric fields lead to higher fluxes, they do not increase ion concentrations (eq. 2). The extracted ion concentration c_S are in the low 10^6 cm^{-3} range. Figure 5 also shows in red the ion current at the MION2 pinhole for different accelerating voltages, respectively the ion concentration at the pinhole. The factor connecting the source concentration scale and the ion current scale is the flow rate J_P . At $U_A < 750 \text{ V}$ ($|U_A| < 750 \text{ V}$), the deflector voltage $U_D = -210 \text{ V}$ creates an electric barrier, and reagent ions do not enter from the ionisation source into the IMR. The closure between measured and modelled pinhole currents is only qualitative. The onset of transmission occurs at a similar voltage, but it is less sharp in measurements. The maximum pinhole current is measured at a higher voltage ($|U_A|$) than predicted. Although the deflector voltage U_D was chosen in the measurement to maximise the ion delivery, it cannot be ruled out is possible that the ion beam did not always fill was not always axially centred to be contained within or to fully illuminate the pinhole flow collection. This could. In very narrow ion beams (low $|U_A|$, compare Fig. 4a) that are not aligned with the flow going to the pinhole, most ions entering the IMR would be lost to the exhaust flow (J_E). This is a plausible explanation for the gradual onset of the observed measured total pinhole current.

We note that the measured pinhole current in Figure 5 apparently decreases faster towards higher absolute accelerator voltages $|U_A|$ than expected from model simulations. Insufficient centring of the beam does not explain this observation, as the ion concentration within the beam varies only slightly. While shying from pinpointing a specific mechanism, we hypothesise that the effect originates in the volume controlled by U_A , i.e., the ionisation volume or the buffer volume between the ionisation volume and IMR. Actual ion mobilities considerably higher than assumed in the model at high $|U_A|$ would lead to a softening of the voltage sensitivity more rapid decrease. A lower degree of reagent ion hydration and cluster formation between the reagent ion and reagent gas at high field strength would increase the effective electrical mobility. The field strength sensitivity of the reagent ion mobility itself is less likely to be significant because of the still relatively weak field strength (Viehland and Mason, 1995). Space charge losses during transport (especially within the IMR) are found to be not yet

significant for the ion concentrations of few 10^6 cm^{-3} . For $U_A > 3000 \text{ V}$, the modelled ion delivery efficiency η_D (eq. 4) is larger than 90 %, essentially unity, for $|U_A| > 3000 \text{ V}$: The ion concentration is maintained from the ion source to the pinhole.




Figure 6. Model-derived histogram of the reaction times between reagent ion and analyte for the MION2 and Eisele type inlet. Different trajectories exhibit reaction times differing by several percent. t_{geo} is the geometrical interaction, derived from the centre flow velocity and the length of the IMR.

The reaction time t for the standard setup is 22 ms (eq. 7, Fig. 6). This is even shorter than reported values of 30 ms (Rissanen et al., 2019) or 35 ms (He et al., 2023) in the literature, which, however, is not specific about how these values were determined.

3.2 Eisele type inlet

Figure 7 shows the geometry (Fig. 7Aa) and the physical quantities in the Eisele type inlet for voltages that maximise the current to the IMR and pinhole.¹ The coordinated flow rates for sample and sheath gas minimise shear and turbulence in the IMR (inner diameter 44 mm, Fig. 7B)-b). Figure 8 shows the velocity profile of the advective velocity upstream of the flow merging at the x-ray lamp plane and downstream after mixing at the IMR mid plane. Before merging, the flow profiles within the different channels are near-parabolic, as expected for fully developed laminar flows. After merging, the individual flow profiles combine to form a transition composite that maintains a near-parabolic shape in the innermost 5 mm with a pronounced maximum at the centre line and a rather flat shoulder with low velocity. The profile is the result of the relatively little interaction with the IMR surface after merging: The IMR radius (22 mm) is relatively large in comparison to the total IMR length (15 cm). Additionally, the velocity profile at the downstream end of the IMR (close to the pinhole plate) is not parabolic, either. The Reynolds number $Re \approx 1840$ (using $D = 16 \text{ mm}$, $u = 1.7 \text{ m s}^{-1}$, $\nu = 1.48 \cdot 10^{-5} \text{ m}^2 \text{ s}^{-1}$) at the downstream end of the sample tube, the location most prone to cause turbulence, supports assuming laminar flow in the modelling.

¹It is possible to apply a non-ground potential to the pinhole plate. The downstream voltages of the mass spectrometer need to be adjusted accordingly. Obeying volume flow conservation, putting the pinhole plate to another potential does not lead to larger currents.




Figure 7. Physical quantities in Eisele type inlet. Only a quarter cut is shown. The colour scale ranges from 0 to the maximum described in each panel. Panel **D-d** shows the electro-advective velocity for anions with $\mu = 2.4 \text{ cm}^2 \text{ V}^{-1} \text{ s}^{-1}$. Used settings: $U_1 = -110.7 \text{ V}$, $U_2 = -98.9 \text{ V}$, $U_{PP} = 0 \text{ V}$, sheath flow 20slpm, sample flow 10slpm.




Figure 8. Modelled velocity profile within the Eisele-type inlet before and after merging sample and sheath flow, using 10slpm sample flow and 20slpm sheath flow. The composite profile establishing in the IMR has a pronounced maximum in the centre and a rather flat shoulder.

240 The **supplied**-voltages U_1 and U_2 supplied in the Eisele-type inlet lead to an electric field (Fig. 7**Cc**) which is perpendicular to the convective field directly downstream of the ion source, but opposing the flow field closer to the centre line. Figure 7**D-d**

shows the electro-convective field and streamlines for nitrate ions or other ions with a comparable mobility coefficient. At the exit of the sampling tube (inner diameter 22 mm), the ion flow velocity decreases, as electrophoresis counteracts the convective flow. The reagent gas HNO_3 is mixed uniformly into the sheath flow (Fig. 7Ff). HNO_3 ionisation leads to the formation of H^+ (Fig. 7Hh) and NO_3^- (Fig. 7Hi). H^+ and NO_3^- are initially convectively transported out of the ionization volume, while recombination occurs (Fig. 7Jj). H^+ is lost to the ion cage, the least repulsive surface. NO_3^- is first attracted towards the IMR cylinder. Once it has cleared the ion cage, it comes under the influence of the attractive electric field generated by the (electrically grounded) sampling tube. If the electric field is well matched to the convective field, NO_3^- is transported towards the centre and then convectively to the pinhole, without significant losses of NO_3^- to the surfaces. The electric gradient between the IMR cylinder and the pinhole plate leads to a focusing of the ions before entering the pinhole. The clustering of NO_3^- and H_2SO_4 (Fig. 7Gg), a proxy for target species, leads to according buildup of $\text{H}_2\text{SO}_4 \cdot \text{NO}_3^-$ clusters (Fig. 7Kk). Interestingly, diffusion mixing of HNO_3 into the centre axis is predicted to be minimal, centre line concentrations are modelled to be less than 1 permil of the sheath gas concentration (Fig. 7Kk). Likewise, the humidity of a moist sample flow is reduced by a dry sheath flow only marginally (Fig. 7Ee).




Figure 9. Sensitivity of pinhole NO_3^- current towards IMR cylinder voltage U_1 and ion cage voltage U_2 in Eisele type inlet. Ion transport to the IMR and pinhole is only significant for a narrow combination of voltages that cause substantial electrophoretic transport while minimising surface losses.

Figure 9 shows from both measurements and model predictions that significant transport of ions to the IMR and pinhole requires coordinated voltages U_1 and U_2 on the order of -100 V . For lower voltages, the electric transport of ions into the centre is too slow. For higher voltages, the ions are lost to the electrodes. Figure 9 further shows that voltage differences $U_2 - U_1$ of only a few V matter, and that the ion cage needs to be slightly more repulsive than the IMR cylinder. If the

260 repulsion from the ion cage is too high (relative to the IMR cylinder), the ions are lost to the IMR cylinder. Vice versa, if the
 265 repulsive voltage of the ion cage is too low, the ions are lost to the ion cage. The model reproduces the measured general trend.
 The measured band width is slightly larger than model-predicted, likely because of small eddies within the inlet leading to
 transport additional to the idealised laminar flow. The subtly different slope is likely due to the exact insertion depth of the
 sampling tube into the assembly, which is known to affect the transmission.

The average reaction time t_{avg} for the standard setup is modelled to be 113 ms (eq. 7, Fig. 6), compatible with literature
 265 values of 160 ms (He et al., 2023). Interestingly, the average reaction time is about 10 ms longer than the geometrical time,
 derived from the centre flow velocity and length of the IMR. This extra time arises from the transport towards the centre line
 and the slower electro-convective transport at the entrance of the IMR. There is only negligible sensitivity of the reaction time
 to the magnitude of the chosen voltages.

4 Discussion

Table 1. Characteristics and comparison of MION2 and Eisele inlet.

metric		Eisele	MION2
creation of ions		x-ray-irradiation of precursor gas	x-ray-irradiation of precursor gas
charge separation		in peripheries of ionization volume	throughout ionization volume
extracted ion concentration ^a	c_S	$6 \cdot 10^6 \text{ cm}^{-3}$	$2 \cdot 10^6 \text{ cm}^{-3}$
electro-convective coupling / ion injection		counter-convective	perpendicular
ion delivery efficiency	η_D	$\sim 10 - 20\%$	$> 90\%$
ion concentration at pinhole	c_P	$1.1 \cdot 10^6 \text{ cm}^{-3}$	$1.5 \cdot 10^6 \text{ cm}^{-3}$
pinhole ion current	$I_{R\pm}$	$1.5 \cdot 10^7 \text{ s}^{-1}$	$1.7 \cdot 10^7 \text{ s}^{-1}$
flow rate sample/sheath/total		10/20/30 slpm	20/0/20 slpm
average reaction time	t_{avg}	113 ms	$\geq 22 \text{ ms}$
detection limit H_2SO_4 (with NO_3^-)	Λ	$7.6 \cdot 10^{-4} \text{ cm}^{-3}$ $7.6 \cdot 10^4 \text{ cm}^{-3}$ ^b	$6.1 \cdot 10^{-4} \text{ cm}^{-3}$ $6.1 \cdot 10^4 \text{ cm}^{-3}$ ^b
transport to IMR		ions and reagent gas	ions only
rapid reagent switching		incompatible	compatible

a: using a primary production of $6 \cdot 10^7 \text{ cm}^{-3} \text{ s}^{-1}$

b: from He et al. (2023), scaling reported detection limit for Eisele inlet to MION2 inlet using relative IMR ion concentrations

270 Table 1 lists similarities and differences between the MION2 and Eisele inlets. They both typically employ x-ray lamps as
 ionisation source, and use electric fields to transport reagent ions into the IMR. In the Eisele inlet, the ionisation volume is
 essentially electric field free; reagent ions and their complements are transported out of the ionisation volume by convection
 only. In the MION2 inlet, electric fields throughout the ionisation volume already separate ions of opposite charge, suppressing
 ion-ion recombination. The extracted ion concentration c_S in the MION2 inlet is not higher than in the Eisele type inlet, as

275 ion–ion recombination is not yet substantially cancelling the primary ion production rate of $6 \cdot 10^7 \text{ cm}^{-3} \text{ s}^{-1}$. The transfer from electric to advective ion transport is well-defined in the MION2 inlet as electric and advective streamlines are perpendicular to each other, whereas in the Eisele inlet an electro-convective streamline from the ion source volume to the centre line does not even exist due to the rotational symmetry (Fig. 7Dd). Accordingly, MION2 accomplishes a near-ideal ion delivery efficiency η_D whereas Eisele reaches best ion delivery efficiencies η_D of few 10%. Additional mixing (e.g., by eddies) in Eisele would not substantially increase the delivery to the pinhole, but could even dilute ion concentrations in the IMR. The combined differences in ion extraction and delivery lead to the pinhole ion concentration c_P in MION2 to be approximately 40 % larger than in Eisele.

The reaction time in the Eisele inlet is 5 times longer than in the studied version of the MION2 inlet. This is a simple result of the geometry. At an axial advective velocity of 1.5 ms^{-1} , 15 cm extra tubing correspond to a lengthening of the reaction time by 100 ms. MION2 is routinely used with such drift tubes (He et al., 2023). The modelled reaction times and ion concentrations are in line with literature comparing both inlets regarding their detection limits, finding that MION2 inlets enable lower detection limits if a similar reaction time is used (He et al., 2023).

Both Eisele type and MION2 inlets can in principle be used with different reagent ions, but by design only a single Eisele ion source of given reagent ion can be coupled to a mass spectrometer at a time. In contrast, multiple MION2 inlets can be coupled to a mass spectrometer at the same time, either of different reagent ion (Rissanen et al., 2019) or of same reagent ion but different reaction time (He et al., 2023). Rapid switching between ion sources is achieved by controlling the electric fields, i.e., grounding the deflector electrode effectively passivates the source. Field measurements with the Eisele type inlet are arguably not practical with reagent ions other than NO_3^- , requiring large quantities of ultra-pure sheath gas, rather than just filtered air.

295 Overall, the MION2 inlet is more efficient in charge separation, ion extraction and delivery, avoids the contamination of sample gas with reagent gas, enables ion switching, and allows for the adjustment of reaction time down to as little as 24 ms (Rissanen et al. (2019), e.g. via insertion of KF25 drift tubes). MION2 is arguably more complex, as controlling the auxiliary flows in the ionisation part (reagent supply, purge supply, purge exhaust) is required, but it does not require a large sheath flow.

The model closure justifies the assumptions and simplifications made in the model. (1) Laminar gas flows are sufficient in the model to explain observations (Fig. 9), small scale eddies may be present but do not critically influence the operation. (2) The ionisation of the precursor gas is assumed to lead to primary production of ions that is constant throughout the ionisation volume (Fig. 1A); both the geometry of the irradiated volume and the constant rate are a simplification (Anttalainen et al., 2021). The ion creation by photo electrons from the irradiation of the electrodes with x-ray is not considered. For the Eisele inlet, only one x-ray source was used in the laboratory measurements. (3) The electric mobility is assumed to be identical for all studied compounds; the mobility of the reagent ion directly affects what deflector voltage is required to achieve a proper extraction from the ionisation volume and injection into the IMR; the mobility of other compounds matters as they influence ion–ion recombination and secondary chemistry. While the model could still be refined regarding the above points, especially the representation of turbulent flow, it already elucidates the limiting process in the current designs and shows potential for the improvement in new inlet designs.

310 The model suggests that the source concentration c_S is limiting the ion concentration in the IMR. Ion-ion recombination or
space charge are not yet significantly at play under the studied conditions with ion concentrations of few 10^6 cm^{-3} , but cannot
be disregarded for moderately enhanced ion concentrations exceeding $\sim 10^7 \text{ cm}^{-3}$. For the Eisele type inlet, the poor ion
delivery efficiency further reduces the attainable concentrations. Increasing ion concentrations in the IMR in new inlet designs
315 convective field), while not compromising the efficient delivery. At high ion concentrations that lead to space-charge induced
electric fields approaching or exceeding the electrode-prescribed electric field (an ion concentration of ~~10^7 cm^{-3}~~ 10^7 cm^{-3} is
equivalent to an electric field change of $18 \text{ V cm}^{-1} \text{ cm}^{-1}$), space charge ceases to be a minor perturbation to ion trajectories
but becomes their primary driver. Avenues for improvement (e.g., Ewing et al., 2023) will constitute a future study.

5 Conclusions

320 This study elucidates the inlet-internal processes, explains observed sensitivities, and highlights the design differences between
the MION2 and Eisele type inlet. While the Eisele type inlet performs well for a relatively simple setup, MION2 type inlets
extract ions more efficiently because of the electric field within the ion volume, and are near-ideal in delivering the ions to the
IMR. MION2 type inlets also allow ion-switching or the sampling of ambient ions. The finding that the ion delivery of MION2
is already near-ideal is curious and suggests that higher initial production rates are initially needed to substantially enhance the
325 reagent ion concentrations in the IMR. It is clear that the model will prove useful in the development of new inlet designs that
deliver ions at higher concentration or are simpler and more robust.

Code and data availability. The data underlying the figures and information on the model will be provided on reasonable request.

Author contributions. A.S., J.K. and M.S. conceived the study idea. H.F. performed the modelling and MION2 measurements and interpreted
the results with help from all co-authors. C.R. and N.S. acquired data with the Eisele type inlet. H.F. wrote the manuscript with help from all
330 co-authors.

Competing interests. Karsa Ltd. and Aerodyne Research Inc. are involved in the production and distribution of the Eisele and MION2 type
inlets.

Acknowledgements. We thank Osmo Anttalainen for initial help with COMSOL simulations. This work ~~was~~ received support by the
Finnish Research Impact Foundation and Research council of Finland (~~356134, 346370~~). 346373, 356134. This project has received

335 [funding from the European Research Council under the European Union's Horizon 2020 research and innovation programme under Grant No. 101002728.](#)

References

- Anttalainen, O., Lattouf, E., Kotiaho, T., and Eiceman, G.: Ion density of positive and negative ions at ambient pressure in air at 12–136 mm from 4.9 kV soft x-ray source, *Review of Scientific Instruments*, 92, 54 104, <https://doi.org/10.1063/5.0050669>, 2021.
- 340 Cussler, E. L.: *Diffusion: Mass Transfer in Fluid Systems*, Cambridge University Press, New York, 2 edn., 2009.
- De Andrade, J. B., De Aragão, N. M., and Araújo, F. R. J.: Nitric Acid-Air Diffusion Coefficient: Experimental Determination Using a Diffusion Cell, *International Journal of Environmental Analytical Chemistry*, 49, 103–109, <https://doi.org/10.1080/03067319208028130>, 1992.
- Eisele, F. L. and Tanner, D. J.: Ion-assisted tropospheric OH measurements, *Journal of Geophysical Research: Atmospheres*, 96, 9295–9308, 345 <https://doi.org/https://doi.org/10.1029/91JD00198>, 1991.
- Eisele, F. L. and Tanner, D. J.: Measurement of the gas phase concentration of H₂SO₄ and methane sulfonic acid and estimates of H₂SO₄ production and loss in the atmosphere, *Journal of Geophysical Research: Atmospheres*, 98, 9001–9010, <https://doi.org/https://doi.org/10.1029/93JD00031>, 1993.
- Ewing, R. G., Hart, G. L., Nims, M. K., Murphy, S. E., Johnson, S. I., Chun, J., and Denis, E. H.: Reducing ion diffusion at 350 atmospheric pressure through intermingled positive and negative ions, *International Journal of Mass Spectrometry*, 492, 117 115, <https://doi.org/https://doi.org/10.1016/j.ijms.2023.117115>, 2023.
- Filippov, A. V., Derbenev, I. N., Dyatko, N. A., Kurkin, S. A., Lopantseva, G. B., Pal', A. F., and Starostin, A. N.: Ionic composition of a humid air plasma under ionizing radiation, *Journal of Experimental and Theoretical Physics*, 125, 246–267, <https://doi.org/10.1134/S1063776117070020>, 2017.
- 355 Finkenzeller, H., Iyer, S., He, X.-C., Simon, M., Koenig, T. K., Lee, C. F., Valiev, R., Hofbauer, V., Amorim, A., Baalbaki, R., Baccarini, A., Beck, L., Bell, D. M., Caudillo, L., Chen, D., Chiu, R., Chu, B., Dada, L., Duplissy, J., Heinritzi, M., Kempainen, D., Kim, C., Krechmer, J., Kürten, A., Kvashnin, A., Lamkaddam, H., Lee, C. P., Lehtipalo, K., Li, Z., Makhmutov, V., Manninen, H. E., Marie, G., Marten, R., Mauldin, R. L., Mentler, B., Müller, T., Petäjä, T., Philippov, M., Ranjithkumar, A., Rörup, B., Shen, J., Stolzenburg, D., Tauber, C., Tham, Y. J., Tomé, A., Vazquez-Pufleau, M., Wagner, A. C., Wang, D. S., Wang, M., Wang, Y., Weber, S. K., Nie, W., Wu, Y., Xiao, M., Ye, Q., 360 Zauner-Wieczorek, M., Hansel, A., Baltensperger, U., Brioude, J., Curtius, J., Donahue, N. M., Haddad, I. E., Flagan, R. C., Kulmala, M., Kirkby, J., Sipilä, M., Worsnop, D. R., Kurten, T., Rissanen, M., and Volkamer, R.: The gas-phase formation mechanism of iodic acid as an atmospheric aerosol source, *Nature Chemistry*, 15, 129–135, <https://doi.org/10.1038/s41557-022-01067-z>, 2023.
- He, X.-C., Shen, J., Iyer, S., Juuti, P., Zhang, J., Koirala, M., Kytökari, M. M., Worsnop, D. R., Rissanen, M., Kulmala, M., Maier, N. M., Mikkilä, J., Sipilä, M., and Kangasluoma, J.: Characterisation of gaseous iodine species detection using the multi-scheme chemical ionisation inlet 2 with bromide and nitrate chemical ionisation methods, *Atmospheric Measurement Techniques*, 16, 4461–4487, 365 <https://doi.org/10.5194/amt-16-4461-2023>, 2023.
- Huang, W., Li, H., Sarnela, N., Heikkinen, L., Tham, Y. J., Mikkilä, J., Thomas, S. J., Donahue, N. M., Kulmala, M., and Bianchi, F.: Measurement report: Molecular composition and volatility of gaseous organic compounds in a boreal forest – from volatile organic compounds to highly oxygenated organic molecules, *Atmospheric Chemistry and Physics*, 21, 8961–8977, <https://doi.org/10.5194/acp-21-8961-2021>, 2021. 370
- Hwang, M., Jiang, R., and Su, T.: Measurements of the diffusion coefficients of atomic bromine in rare gases, *The Journal of Chemical Physics*, 91, 1626–1630, <https://doi.org/10.1063/1.457122>, 1989.

- Hwang, M.-D. and Su, T.-M.: Diffusion Coefficients of Atomic Halogen, *Journal of the Chinese Chemical Society*, 37, 33–44, <https://doi.org/https://doi.org/10.1002/jccs.199000005>, 1990.
- 375 Iyer, S., Rissanen, M. P., Valiev, R., Barua, S., Krechmer, J. E., Thornton, J., Ehn, M., and Kurtén, T.: Molecular mechanism for rapid autoxidation in α -pinene ozonolysis, *Nature Communications*, 12, 878, <https://doi.org/10.1038/s41467-021-21172-w>, 2021.
- Kelly, R. T., Tolmachev, A. V., Page, J. S., Tang, K., and Smith, R. D.: The ion funnel: Theory, implementations, and applications, *Mass Spectrometry Reviews*, 29, 294–312, <https://doi.org/https://doi.org/10.1002/mas.20232>, 2010.
- Mauldin III, R. L., Tanner, D. J., and Eisele, F. L.: A new chemical ionization mass spectrometer technique for the fast
380 measurement of gas phase nitric acid in the atmosphere, *Journal of Geophysical Research: Atmospheres*, 103, 3361–3367, <https://doi.org/https://doi.org/10.1029/97JD02212>, 1998.
- Partovi, F., Mikkilä, J., Iyer, S., Mikkilä, J., Kontro, J., Ojanperä, S., Juuti, P., Kangasluoma, J., Shcherbinin, A., and Rissanen, M.: Pesticide Residue Fast Screening Using Thermal Desorption Multi-Scheme Chemical Ionization Mass Spectrometry (TD-MION MS) with Selective Chemical Ionization, *ACS Omega*, 8, 25 749–25 757, <https://doi.org/10.1021/acsomega.3c00385>, 2023.
- 385 Passananti, M., Zapadinsky, E., Zanca, T., Kangasluoma, J., Myllyls, N., Rissanen, M. P., Kurtén, T., Ehn, M., Attoui, M., and Vehkamäki, H.: How well can we predict cluster fragmentation inside a mass spectrometer?, *Chem. Commun.*, 55, 5946–5949, <https://doi.org/10.1039/C9CC02896J>, 2019.
- Rissanen, M.: Anthropogenic Volatile Organic Compound (AVOC) Autoxidation as a Source of Highly Oxygenated Organic Molecules (HOM), *The Journal of Physical Chemistry A*, 125, 9027–9039, <https://doi.org/10.1021/acs.jpca.1c06465>, 2021.
- 390 Rissanen, M. P., Mikkilä, J., Iyer, S., and Hakala, J.: Multi-scheme chemical ionization inlet (MION) for fast switching of reagent ion chemistry in atmospheric pressure chemical ionization mass spectrometry (CIMS) applications, *Atmos. Meas. Tech.*, 12, 6635–6646, <https://doi.org/10.5194/amt-12-6635-2019>, 2019.
- Shen, J., Scholz, W., He, X.-C., Zhou, P., Marie, G., Wang, M., Marten, R., Surdu, M., Rörup, B., Baalbaki, R., Amorim, A., Ataei, F., Bell, D. M., Bertozzi, B., Brasseur, Z., Caudillo, L., Chen, D., Chu, B., Dada, L., Duplissy, J., Finkenzeller, H., Granzin, M., Guida,
395 R., Heinritzi, M., Hofbauer, V., Iyer, S., Kempainen, D., Kong, W., Krechmer, J. E., Kürten, A., Lamkaddam, H., Lee, C. P., Lopez, B., Mahfouz, N. G. A., Manninen, H. E., Massabò, D., Mauldin, R. L., Mentler, B., Müller, T., Pfeifer, J., Philippov, M., Piedehierro, A. A., Roldin, P., Schobesberger, S., Simon, M., Stolzenburg, D., Tham, Y. J., Tomé, A., Umo, N. S., Wang, D., Wang, Y., Weber, S. K., Welti, A., Wollesen de Jonge, R., Wu, Y., Zauner-Wieczorek, M., Züst, F., Baltensperger, U., Curtius, J., Flagan, R. C., Hansel, A., Möhler, O., Petäjä, T., Volkamer, R., Kulmala, M., Lehtipalo, K., Rissanen, M., Kirkby, J., El-Haddad, I., Bianchi, F., Sipilä, M., Donahue,
400 N. M., and Worsnop, D. R.: High Gas-Phase Methanesulfonic Acid Production in the OH-Initiated Oxidation of Dimethyl Sulfide at Low Temperatures, *Environ. Sci. Technol.*, <https://doi.org/10.1021/acs.est.2c05154>, 2022.
- Sipilä, M., Sarnela, N., Jokinen, T., Junninen, H., Hakala, J., Rissanen, M. P., Praplan, A., Simon, M., Kürten, A., Bianchi, F., Dommen, J., Curtius, J., Petäjä, T., and Worsnop, D. R.: Bisulfate HSO_4^- cluster based atmospheric pressure chemical ionization mass spectrometer for high-sensitivity (~ 100 ppqV) detection of atmospheric dimethyl amine: proof-of-concept and first ambient data from boreal
405 forest, *Atmospheric Measurement Techniques*, 8, 4001–4011, <https://doi.org/10.5194/amt-8-4001-2015>, 2015.
- Steiner, G., Jokinen, T., Junninen, H., Sipilä, M., Petäjä, T., Worsnop, D., Reischl, G. P., and Kulmala, M.: High-Resolution Mobility and Mass Spectrometry of Negative Ions Produced in a 241Am Aerosol Charger, *Aerosol Science and Technology*, 48, 261–270, <https://doi.org/10.1080/02786826.2013.870327>, 2014.
- Tanner, D. J. and Eisele, F. L.: Present OH measurement limits and associated uncertainties, *Journal of Geophysical Research: Atmospheres*,
410 100, 2883–2892, <https://doi.org/https://doi.org/10.1029/94JD02609>, 1995.

- Tanner, D. J., Jefferson, A., and Eisele, F. L.: Selected ion chemical ionization mass spectrometric measurement of OH, *Journal of Geophysical Research: Atmospheres*, 102, 6415–6425, <https://doi.org/https://doi.org/10.1029/96JD03919>, 1997.
- [Viehland, L. A. and Mason, E. A.: Transport Properties of Gaseous Ions over a Wide Energy Range, IV, Atomic Data and Nuclear Data Tables, 60, 37–95, https://doi.org/https://doi.org/10.1006/adnd.1995.1004, 1995.](https://doi.org/https://doi.org/10.1006/adnd.1995.1004)
- 415 Wang, M., He, X.-C., Finkenzeller, H., Iyer, S., Chen, D., Shen, J., Simon, M., Hofbauer, V., Kirkby, J., Curtius, J., Maier, N., Kurtén, T., Worsnop, D. R., Kulmala, M., Rissanen, M., Volkamer, R., Tham, Y. J., Donahue, N. M., and Sipilä, M.: Measurement of iodine species and sulfuric acid using bromide chemical ionization mass spectrometers, *Atmos. Meas. Tech.*, 14, 4187–4202, <https://doi.org/10.5194/amt-14-4187-2021>, 2021.
- 420 Yuan, B., Koss, A. R., Warneke, C., Coggon, M., Sekimoto, K., and De Gouw, J. A.: Proton-Transfer-Reaction Mass Spectrometry: Applications in Atmospheric Sciences, *Chemical Reviews*, 117, 13 187 – 13 229, <https://doi.org/10.1021/acs.chemrev.7b00325>, 2017.
- Zauner-Wieczorek, M., Curtius, J., and Kürten, A.: The ion–ion recombination coefficient α : comparison of temperature- and pressure-dependent parameterisations for the troposphere and stratosphere, *Atmospheric Chemistry and Physics*, 22, 12 443–12 465, <https://doi.org/10.5194/acp-22-12443-2022>, 2022.